Real-time RT-PCR assays for type and subtype detection of influenza A and B viruses

Luke T. Daum,^{a,b} Linda C. Canas,^c Bernard P. Arulanandam,^{a,b} Debra Niemeyer,^d James J. Valdes,^e James P. Chambers,^{a,b}

^aDepartment of Biology, The University of Texas at San Antonio, San Antonio, TX, USA. ^bThe Center of Excellence in Biotechnology, Bioprocessing, Education, and Research, Brooks City Base, TX, USA. ^cAir Force Institute for Operational Health, Brooks City Base, TX, USA. ^dOffice of the Surgeon General (USAF), Pentagon, Army Navy Drive and Fern Street, Arlington, VA, USA. ^eEdgewood Chemical Biological Center, Aberdeen Proving Ground, MD, USA

Correspondence: James P. Chambers, Department of Biology, The University of Texas at San Antonio, 6900 N. Loop W., San Antonio, Texas 78249, USA. Email: James.chambers@utsa.edu.

Accepted 12 August 2007.

Influenza viruses type A (H3N2 and H1N1 subtypes) and B are the most prevalently circulating human influenza viruses. However, an increase in several confirmed cases of high pathogenic H5N1 in humans has raised concerns of a potential pandemic underscoring the need for rapid, point of contact detection. In this report, we describe development and evaluation of 'type,' i.e., influenza virus A and B, and 'subtype,' i.e., H1, H3, and H5, specific, single-step/reaction vessel format, real-time RT-PCR assays using total RNA from archived reference strains, shellvial cultured and uncultured primary (throat swab/nasal wash) clinical samples. The type A and B specific assays detected all 16 influenza type A viruses and both currently circulating influenza B lineages (Yamagata and Victoria), respectively. 'Type' and 'subtype' specific assays utilize one common set of thermocycling conditions, are specific and highly sensitive (detection threshold of approximately 100 target template molecules). All clinical specimens and samples were evaluated using both the unconventional portable Ruggedized Advanced Pathogen Identification Device (RAPID) and standard laboratory bench LightCycler instruments. These potentially field-deployable assays could offer significant utility for rapid, point of care screening needs arising from a pandemic influenza outbreak.

Keywords H1, H3, H5, influenza A/B, RT-PCR.

Please cite this paper as: Daum et al. (2007) Real-time RT-PCR assays for type and subtype detection of influenza A and B viruses. Influenza and Other Respiratory Viruses 1(4), 167–175.

Background

In humans, Influenza is an acute, highly contagious respiratory disease caused primarily by influenza viruses type A and B. The infectivity of influenza viruses is augmented considerably due to viral alteration of antibody targeted surface glycoproteins, i.e., 'drift,' giving rise to annually occurring epidemics. In contrast to annual epidemics, pandemics occur at 10–50 year intervals, arising by re-assortment, i.e., 'shift' between human and avian/porcine viruses.¹ Of significant concern are human infections arising from avian H5 influenza A virus considered by many as an emerging pandemic threat.^{2–4}

Currently, several conventional 'gold standard' procedures for viral identification are well-established in the clinical reference laboratory.^{5–8} However, these methods require 3–6 days for definitive diagnosis as well as propagation of live (and potentially infectious) influenza viral strains and are of very little 'point of care' value. Rapid ELISA-based 'flu/no flu' assays have been developed and are commercially available. Although fast and easily used, these assays often exhibit specificity and sensitivity short-comings and do not subtype influenza virus type A, e.g., H1, H3, and H5. $^{9-12}$

Traditional RT-PCR has become an effective methodology for viral detection, typing, and subtyping.^{13–15} Although RT-PCR exhibits high sensitivity and considerably shorter detection time than conventional virological methods, drawbacks, such as extraction complexity, post-PCR analysis, and increased potential for cross-contamination have been noted.^{16,17} Several dual step, real-time (rRT-PCR) assays have appeared in the literature but require either nested PCR or use of random hexamers.¹⁸⁻²⁰ Likewise, single-step assays have been developed, but require a variety of reagents of differing concentrations as well as thermocycling parameters.²¹ In 2004, the artus Influenza/H5 kit (Qiagen, Valencia, CA, USA) was made commercially available (for research use only), combining a broad screening capability for detecting type A (15 of 16 HA types) and type B viruses. In 2005, the Laboratory Response Network (LRN) single-step assay was developed and subsequently approved for use by the Food Drug

Administration (FDA) for detection of avian H5 influenza.²²⁻²⁴ However, this assay does not detect nor distinguish commonly circulating strains, i.e., H1, H3, and influenza B viruses. More recently (2006), the TaqMan Influenza A/H5 virus detection kit (Applied Biosystems, Foster City, CA, USA) consisting of two assays; one designed to detect influenza type A (M gene target) and a second specific for the H5 subtype (hemagglutinin gene target), was made commercially available but for research use only. Although the artus Influenza/H5 kit offers a broader screening capability than that of the TaqMan Influenza A/H5 assay, neither of these reagent kits subtype specifically commonly circulating human strains. Recently, Di Trani and coworkers describe a sensitive, one-step, rRT-PCR method for detection of type A influenza viruses (13 of 16 HA types) using a minor groove binding probe;²⁵ however, the method of Di Trani does not differentiate type A and B viruses nor subtype the respective type A viruses.

Objective

The objective of this work was twofold: (i) to develop and (ii) evaluate 'type,' i.e., influenza virus A and B, and 'sub-type,' i.e., H1, H3, and H5, specific, rRT-PCR assays. Fur-thermore, the assays were to be carried out using one identical and common set of thermocycling conditions (single-step/reaction vessel format) and two rRT-PCR plat-forms; the unconventional 50 pound portable Ruggedized Advanced Pathogen Identification Device (RAPID) and standard laboratory bench LightCycler instrumentation.

Patients/methods

Collection of clinical samples and virology

All original primary specimens (throat swab/nasal washes) and cultured samples used in this study were collected over 7 (99/00, 00/01, 01/02, 02/03, 03/04, 04/05, and 05/06) and 3 (03/04; 04/05, and 05/06) influenza seasons, respectively under the auspices of the Department of Defense Global Emerging Infectious Surveillance (DoD-GEIS) network. Primary specimens were collected within the first 48-72 h of symptom onset from patients presenting with a fever ≥100.5°F oral and cough or sore throat. Dacron throat swab specimens were submerged in 3.0 ml viral transport media (M4 MicroTest Multi-Microbe Media, Remel, Lenexa, KS, USA). Submerged throat swabs and saline nasal wash material (5 ml) were shipped on dry ice to Brooks City Base, San Antonio, TX, USA. Propagation of influenza viruses from primary specimens was achieved using the centrifugation-enhanced Shell-Vial culture technique followed by typing using influenza virus A or B specific monoclonal antibodies (Chemicon International,

Temecula, CA, USA) per manufacturer's recommendations and fluorescent microscopy.²⁶ Aliquots (0.25 ml) of primary specimens (before and after culturing) served as source of specimen RNA.

Extraction of RNA

RNA extraction was achieved using the Qiagen M48 automated bead-based extraction robot (Qiagen) with the Mag-Attract Virus Mini M48 kit (Qiagen) per manufacturer's protocols, eluted in 50 μ l of Elution Buffer and stored at -80° C until used.

Genomic primer/probe design

Primer/probe design was based upon sequence data obtained from the Los Alamos National Laboratory and Department of Defense data bases. Type specific (influenza A and B) assays target a highly conserved region of the matrix protein (MP) gene and were designed based on 100 and 50 alignments common to all 16 influenza A virus subtypes and both influenza B lineages (B/Victoria and B/Yamagata), respectively. H1, H3, and H5 influenza A subtype specific assays target conserved regions of the respective hemagglutinins (HA). H1 primer/probe sequence alignment was achieved using 51 geographically diverse strains obtained during the 2005/06 and 2006/07 influenza seasons including the A/New Caledonia/20/99 vaccine strain. H3 primer/probe sequence alignment was achieved using 140 H3N2 field strains collected between 2004 and 2006. H5 primer/probe sequence design was based upon alignment analysis of 22 H5N1 clinical isolates representing clades 1 and 2 (subclades 1, 2, and 3) viruses. All primers and probes were procured from Applied Biosystems.

Real-time RT-PCR platforms

The laboratory-based Lightcycler 2.0 instrument (Roche Molecular Diagnostics, Indianapolis, IN, USA) and its lightweight portable (50 lbs) version, the Ruggedized Advanced Pathogen Identification Device (RAPID, Idaho Technologies, Salt Lake City, UT, USA) are both 32-well capillary, real-time instruments which employ similar components and operational software. The RAPID is configured within a hardened case and can be employed at the point of care.

Real-time RT-PCR amplification

Primer and probe sequences are shown in Table 1. Primer pair melting points are within 2° C and anneal/extend at $58-60^{\circ}$ C. The respective probes anneal/extend $8-10^{\circ}$ C higher than that of the primers. Thermocycling operates in a rapid, two temperature format with annealing and extension (30 s) facilitated by the short nature of the respective amplicons. Due to genetic variability in the influenza viral genome, degenerate nucleotides were placed at specific loci.

Primer	Sequence	Gene target-nucleotide position*	Amplicon (bp)
Real-time RT-PCR			
Туре А		Matrix (Segment 7)	195
Forward	taaccgaggtcgaaacgta	$36 \rightarrow 54$	
Reverse	gcacggtgagcgtgaa	$215 \rightarrow 230$	
Probe	(FAM)-tcaggccccctcaaagc	$74 \rightarrow 90$	
Type B		Matrix (Segment 7)	119
Forward	ggaattgcaaaggatgtaatggaa	673 → 696	
Reverse	agaacaaattgaaagaatctgaaatggt	764 → 791	
Probe	(FAM)-atgggaaattcagctct	721 → 731	
Subtype H1		H1 Hemagglutinin (Segment 4)	89
Forward	agycttcctttccagaatgtaca	948 → 970	
Reverse	agtcctgtarccatycttaattttg	$1012 \rightarrow 1036$	
Probe	(FAM)-tctccaaagtatgtcagg	973 → 988	
Subtype H3	(i, iii) teteeddagaatgeedgg	H3 Hemagglutinin (Segment 4)	113
Foward	aatacgaagtgggaaaagctca	$777 \rightarrow 798$	115
Reverse	gccccrtatgtgatyctgtttac	$889 \rightarrow 911$	
Probe	(FAM)-tgagatcagatgcacccat	803 → 821	
Subtype H5	(FAM) tguguteugutgeucceut	H5 Hemagglutinin (Segment 4)	144
Forward	actayccgcagtattcagaagaagc	$1511 \rightarrow 1535$	144
Reverse	gaccagcyaycatgattgcca	$1634 \rightarrow 1654$	
Probe	(FAM)-agagrggaaataagtgg	1546 → 1562	
cRNA target temp		$1540 \rightarrow 1502$	
	Jales	Matrix (Cogmont 7)	1027
Type A Forward**	actostocastcoctotoagasassacsassacsaatsastatt	Matrix (Segment 7) 1 \rightarrow 20	1027
	gctaatacgactcactatagggagaagcaaaagcaggtagatatt		
Reverse	agtagaaacaaggtagttttttac	$1004 \rightarrow 1027$	1001
Type B		Matrix (Segment 7)	1061
Forward**	gctaatacgactcactatagggagatcgctgtttggagacacaat	$7 \rightarrow 26$	
Reverse	ctccaaaactgtttcaccca	$1048 \rightarrow 1067$	1201
Subtype H1		H1 Hemagglutinin (Segment 4)	1204
Forward**	gctaatacgactcactatagggagaaagcaggggaaaataaaa	$7 \rightarrow 23$	
Reverse	gtaatcccgttaatcgca	1193 → 1210	
Subtype H3		H3 Hemagglutinin (Segment 4)	1192
Forward**	gctaatacgactcactatagggagaactatcattgctttgagc	$7 \rightarrow 24$	
Reverse	atggctgcttgagtgctt	1181→1198	
Subtype H5		H5 Hemagglutinin (Segment 4)	1658
Forward**	gctaatacgactcactatagggagatcatctgtcaaatggagaaaat	$4 \rightarrow 22$	
Reverse	aaggatagaccagctaccatga	$1640 \to 1661$	

Table 1. Primer/probe sequences* used for rRT-PCR amplification and *in vitro* generation of cRNA target templates

Genomic primer/probe design, rRT-PCR and *in vitro* generation of cRNA target templates were carried out as previously described under 'Patients/methods'

*All sequences are listed 5' \rightarrow 3'.

**Inclusive of the 22 nucleotide T7 promoter sequence with the gc linker (5' end).

r = 50% mixture of a and g.

y = 50% mixture of c and t.

Real-time amplification was performed in a singlestep/reaction vessel format. Using the UltraSense Platinum One-Step Quantitative RT-PCR System (Invitrogen, Carlsbad, CA, USA), 2 μ l RNA was added to 18 μ l master mix containing the following components at the indicated final concentrations: 1X reaction buffer, 1X enzyme mixture containing 500 nM of each primer and 300 nM probe labeled at the 5' end with 6-carboxyfluorescein (FAM) reporter dye and at the 3' end with a non-fluorescent quencher and minor groove binder. Thermocycling was carried out as follows: 30 min at 45°C and 2 min at 95°C for reverse transcription (RT) and denaturation, respectively, followed by 40 amplification cycles consisting of 95°C for 5 s (denaturation) and 60°C for 30 s (extension). Amplification efficiency was determined using the C_T slope method (cf. Fig. 1, frame B) according to the equation:

 $E = [10^{(-1/\text{Slope})} - 1] \times 100$. All assays described here exhibited greater than 98.5% amplification efficiency.

For each analysis, 'no template' and 'positive' controls were included. Baseline fluorescence for each analysis was manually adjusted to that of the respective 'no template' control reaction. The 'positive' control (0.1 ng cRNA) gives rise to an increase in fluorescence intensity relative to the no template baseline ($C_{\rm T}$ value between 18 and 22). A 'positive' unknown is defined as amplification exceeding baseline fluorescence with a corresponding $C_{\rm T}$ value not exceeding 36 in a 40 cycle run. All original, i.e., uncultured specimens and cultured samples reported here using both platforms exhibited $C_{\rm T}$ range values of 26–35 (n = 144, mean = 31.5) and 17–27 (n = 407, mean = 23), respectively.

Generation of cRNA target templates

Reverse and forward primers for *in vitro* generation of target complementary RNA (cRNA) templates corresponding to Influenza type (A/B) and Influenza A subtype (H1, H3, and H5) RNA sequences are shown in Table 1.

Briefly, traditional RT-PCR was carried out as follows: 5 µl viral RNA was added to a 45 µl master mix containing the following components at the indicated final concentrations: 1X reaction buffer with 1.6 mM MgSO₄, 1X enzyme mixture containing 400 nM HA or MP primer pairs using the SuperScript III One-Step RT-PCR System (Invitrogen). Reverse transcription was carried out at 50°C for 30 min followed by a 'hot start' step (2 min) at 95°C. Thermocycling (40 amplification cycles) was carried out as follows: 30 s at 95°C, 15 s at 52°C, 1 min at 68°C with final extension for 7 min at 68°C. PCR reaction product (5 µl) was subjected to analytical electrophoresis on 2% pre-cast gels containing ethidium bromide (Invitrogen) and remaining product (45 μ l) purified using the Qiaquick PCR Purification Kit (Qiagen). In vitro transcription was carried out for 4 h using the T7 MegaScript Kit (Ambion, Austin, TX, USA) per manufacturer's recommendations. Reactions were subjected to nuclease digestion using Turbo DNase (Ambion) and subsequently purified using the MegaClear kit (Ambion). RNA was quantitated using a NanoDrop (NanoDrop Technologies, Wilmington, DE, USA) spectrophotometer, aliquoted and served as control cRNA.

Nucelotide sequencing

Purified amplicons were cloned using a Topo 2.0 Cloning Kit (Invitrogen) and sequenced using the Big Dye Terminator v3.1 reagent Kit (Applied Biosystems). Unincorporated fluorescent nucleotides were removed using a Dye Ex 96well plate kit per manufacturer's recommendations (Qiagen). Nucleotide sequencing was performed using an ABI 3100 Genetic Analyzer (ABI Inc.).

Results

As shown in Table 2, influenza A and B virus type specific assays detected all known influenza A hemagglutinin sub-

Table 2.	Detection	of influenza	virus type	(A/B) strains
----------	-----------	--------------	------------	---------------

Subtype	Strain	Flu A Univ*	Flu B Univ*
H1N1	DK/NJ/7717-70/95	+	_
H1N1	CK/NY/21665-73/98	+	-
H2N2	CK/NY/3749-7/96	+	_
H2N4		+	_
H3N8	ENV/NY/19019-6/98	+	-
H4N6		+	_
H4N8	DK/ALB/286/78	+	_
H5N1	TK/ENG/50-92/91	+	-
H5N1	CK/SCTT/59	+	_
H5N1	MA/OH/184/86	+	_
H5N1		+	_
H5N2		+	-
H5N2	CK/PUE/8624-602/94	+	_
H5N2		+	_
H5N3		+	_
H5N3	TK/CA/6878/79	+	_
H5N8		+	_
H5N9		+	-
H6N2	CK/NY/14677-13/99		_
H7N1	SB/IL/35445-136/92	+	-
H7N1		+	_
H7N2	CK/PA/13552-1/98	+	-
H7N2			_
H7N2	TK/PA/7975/97	+ +	-
H7N3		+	_
H7N3	MA/NETHERLANDS/12/00		-
H7N3	Q/AR/16309-2/94	+	-
H7N4	CK/NSW/1688/97	+	-
H7N7	CK/VIC/85	+	_
H7N7	EQ/PRAGUE/E5302/56	+	-
H8N4		+	-
H8N8	DK/VIC/8211-18-1400/92	+	_
H9N2		+	_
H9N2	CK/KOREA/96006/96	+	_
H9N2	CK/NJ/12220/97	+	_
H10N7	TK/VA/31409/91	+	_
	CI/GER/N/49		_
		+	_
H11N1	CK/NJ/15906-6/96		_
	DK/LA/18813/87		_
	DK/ALB/60/76	+	_
H13N6	GULL/6411/MD/704/77	+	_
H14N5	MA/GURJEV/263/82	+	_
H15N9	SW/AUSTR/2576/79	+	_
H16N3	A/BHG/SWEDEN/2/99	+	_
B/Yam	B/Arizona/135/2005	_	+
B/Vic	B/Arizona/140/2005	_	+
B/Vic	B/Nepal/1120/2005	_	+
B/Yam	B/Florida/925/2004		+

Table 2. (Continued)					
Subtype	Strain	Flu A Univ*	Flu B Univ*		
B∕Vic B∕Vic	B/Hawaii/22/2004 B/Brisbane/32/2002	-	++++		

Influenza virus A and B detection by real-time RT-PCR was carried out as previously described under 'Patients/methods'.

Total RNA for influenza virus A H1-15 subtypes was obtained from Dr. David Suarez, Southeast Poultry Research Laboratory, USDA Agricultural Research Service, Athens, Georgia 30605. H16 RNA was provided by Drs. R.A. Fouchier and A.D. Osterhaus of the Department of Virology, Erasmus Medical Center, The Netherlands. Influenza B virus reference strains were obtained from the Department of Defense Global Emerging Infectious Surveillance (DoD-GEIS) network.

*Determined by type A and B specific assays.

B/Yam = Type B Yamagata lineage; B/Vic = Type B Victoria lineage.

types (H1-H16) and both type B (Yamagata and Victoria) viruses, respectively. Importantly, no cross-reactivity was observed between influenza A and B virus specific probes supporting the very specific nature of these assays.

H1, H3, and H5 influenza A subtype specific assays were initially evaluated using 180 archived clinical isolates. As shown in Table 3, 178 samples were correctly typed and subtyped in research blinded fashion; the far greater number of samples identified as type A influenza (91%) with the remainder, i.e., 9% being type B influenza. Of the type A influenza samples, the H3 subtype was the most prevalent (93.2%) with the remaining 6.8% H1 subtype. Two samples (District of Columbia) tested influenza virus negative (using type and subtype assays reported here) and were later confirmed as being Coxsackie B and Adenovirus (data not shown). Consistent with no cross-reactivity of type specific probes, no subtype specific probe cross-reactivity, i.e., H1, H3, and H5 assay cross-reactivity was observed. Furthermore, 40 commonly encountered bacteria/viruses were tested concurrently and did not amplify using either type and subtype specific primer/probes listed in Table 1 (data not shown).

Typing and subtyping influenza viruses at the point of care using uncultured (low viral titer) primary specimens is crucial for expanded surveillance. Shown in Table 4 is analysis of 167 uncultured primary clinical samples. Of the 167 specimens, 100 were correctly identified, i.e., typed (Influenza A or B, 60 and 40, respectively) and type A samples subtyped H1 or H3, 12 (20%) and 48 (80%), respectively. Furthermore, 67 negative influenza samples were subsequently determined to be culture negative for influenza viruses (data not shown). Of the 60 influenza A samples, 16 initially tested influenza negative in contrast to

Table 3. Detection of influenza virus types (A/B) and subtypes (H1, H3, and H5) in cultured clinical isolates by rRT-PCR

Location (samples)	Flu A/B*	H1**	H3**	H5**	Flu A/B***	
United States						
Alabama (2)	2/0	0	2	0	2/0	
Alaska (4)	3/1	0	3	0	3/1	
Arizona (9)	7/2	0	7	0	7/2	
California (4)	4/0	0	4	0	4/0	
Colorado (23)	22/1	0	22	0	22/1	
Connecticut (1)	1/0	0	1	0	1/0	
District of Columbia $(4)^{\dagger}$	1/1	0	1	0	1/1	
Hawaii (11)	9/2	0	9	0	9/2	
Illinois (4)	4/0	0	4	0	4/0	
Maryland (6)	6/0	0	6	0	6/0	
New Jersey (3)	3/0	0	3	0	3/0	
New York (1)	1/0	0	1	0	1/0	
Oklahoma (5)	5/0	0	5	0	5/0	
Texas (24)	24/0	0	24	0	24/0	
Virginia (2)	2/0	0	2	0	2/0	
Country						
Ecuador (1)	1/0	0	1	0	1/0	
England (7)	4/3	1	3	0	4/3	
Guam (2)	2/0	1	1	0	2/0	
Iraq (1)	1/0	0	1	0	1/0	
Italy (1)	1/0	0	1	0	1/0	
Japan (21)	19/2	1	18	0	19/2	
Ketchikan (1)	1/0	0	1	0	1/0	
Kenya (2)	1/1	1	0	0	1/1	
Korea (12)	12/0	5	7	0	12/0	
Kuwait (3)	2/1	1	1	0	2/1	
Peru (21)	19/2	0	19	0	19/2	
Qatar (4)	4/0	1	3	0	4/0	
Turkey (1)	1/0	0	1	0	1/0	
Total Samples = 180	162/16	11	151	0	162/16	

Real-time RT-PCR analyses and traditional virus isolation/identification of shell vial cultured clinical samples were carried out as previously described under 'Patients/methods'

*Determined by type A and B specific assays.

**Determined by subtype (H1, H2, and H3) specific assays.

***Determined by traditional 'gold standard' virus isolation/identification.

[†]Two of 4 District of Columbia samples tested negative for influenza virus using type/subtype specific assays reported here and were subsequently confirmed as being influenza negative by traditional 'gold standard' virus isolation/identification.

original culture data. The inability to detect 16 of the 100 influenza specimens could have arisen from base pair sequence 'drift' in the respective primer/probe binding regions, less than threshold amounts of extracted target RNA arising from prolonged storage/degradation at -80°C and poor sample collection. Therefore, aliquots of all 16 samples were removed and transferred to monolayer primary monkey kidney (PMK) culture tubes and Shell Vials for further analysis. After 48 h, 10 of the 16 Shell Vial

Table 4. Detection of influenza virus types (A/B) and subtypes (H1, H3, and H5) in uncultured primary clinical specimens by rRT-PCR

Location (samples)	Flu A/B*	H1*	* H3**	H5**	Flu A/B***
United States					
Alabama (1)	0/1	0	0	0	0/1
Arkansas (4)	4/0	3	1	0	4/0
California (1)	1/0	0	1	0	1/0
Delaware (1)	1/0	0	1	0	1/0
District of Columbia (2)	0/2	0	0	0	0/2
Georgia (2)	2/0	0	2	0	2/0
Hawaii (14) ^{††(4)}	14/0	1	13	0	14/0
Mississippi (1) ^{††(1)}	1/0	0	1	0	1/0
New Jersey (2)	2/0	1	1	0	2/0
New Mexico (2)	0/2	0	0	0	0/2
Oklahoma (1)	1/0	0	1	0	1/0
South Carolina (1)	1/0	0	1	0	1/0
Texas (3) [†]	1/1	1	0	0	1/1
Virginia (16)	1/15	0	1	0	1/15
Washington (1)	1/0	0	1	0	1/0
Country					
Guam (2) ^{††(2)}	2/0	0	2	0	2/0
lraq (1)	1/0	0	1	0	1/0
Italy (2)	1/1	0	1	0	1/1
Japan (5) ^{††(2)}	3/2	2	1	0	3/2
Kenya (2)	1/1	0	1	0	1/1
Korea (5) ^{††(2)}	5/0	1	4	0	5/0
Kuwait (1)	1/0	1	0	0	1/0
Nepal (69) ^{†, ††(1)}	1/2	0	1	0	1/2
Okinawa (1)	1/0	0	1	0	1/0
Peru (8)	4/4	1	3	0	4/4
Qatar (2)	2/0	1	1	0	2/0
Thailand (13) ^{††(2)}	4/9	0	4	0	4/9
Turkey (2) ^{††(2)}	2/0	0	2	0	2/0
United Kingdom (2)	2/0	0	2	0	2/0
Total Samples = 167	60 ^{††(16)} /40	12	48	0	60/40

Real-time RT-PCR analyses and traditional virus isolation/identification of primary clinical specimens (throat swabs/nasal washes) were carried out as previously described under 'Patients/methods' *Determined by type A and B specific assays.

**Determined by subtype (H1, H2, and H3) specific assays.

***Determined by traditional 'gold standard' virus isolation/identification.

[†]One of 3 Texas samples and 66 of 69 Nepal samples tested negative for influenza virus using type/subtype specific assays reported here and were all subsequently confirmed as being influenza negative by traditional 'gold standard' virus isolation/identification.

⁺⁺⁰Primary uncultured samples requiring culturing for identification. The number of samples is indicated in the parenthesis.

enriched samples tested positive by type specific rRT-PCR and standard immunoreagent fluorescence staining. The remaining six samples were checked/screened daily thereafter and three of six tested positive 5 days post-inoculation and the remaining three testing positive 9 days post-inoculation. All 16 influenza type A amplicons were further validated by sequence analysis (data not shown). The overall specificity of assays using original specimens as source of RNA was 100% (no cross-hybridization) with an overall sensitivity of 90.4% (151 correct positives and negatives out of a total of 167 samples).

Although no H5 influenza virus was observed in our sample collection (Tables 3 and 4), the usefulness of the H5 subtype specific assay is demonstrated by alignment analysis of known H5 influenza A virus with H5 primer/ probe sequences and H5 assay sensitivity by template serial dilution. Shown in Fig. 1 (frame A) is alignment analysis of 22 H5 influenza A hemagglutinin sequences including an isolate from the first human H5 outbreak (1997) and subsequent outbreaks through 2006 with the H5 subtype specific, primer/probe sequences. Complete (100%) primer/probe, H5 virus sequence homology (avian and mammalian sources) was observed. Shown in Fig. 1 (frame B) is a representative profile of serially diluted H5 cRNA template (obtained from a human fatality) over an 8-log dilution range $(10^{-9}-10^{-16} \text{ g})$ corresponding to approximately 100 H5 cRNA target molecules (10⁻¹⁶ g). Serially diluted cRNA targets (type A, subtypes H1 and H3 and type B) exhibited very similar profiles to that shown in Fig. 1, frame B (data not shown).

Conclusions

Influenza type specific assays described in this report detected all 16 known type A viruses including the recently discovered H16 strain ²⁷ as well as both Yamagata and Victoria type B viruses. Using both the laboratory-based Lightcycler 2.0 instrument and its lightweight (50 lbs) version, the Ruggedized Advanced Pathogen Identification Device (RAPID), 347 archived primary clinical samples (throat swab/nasal wash, 180 cultured and 167 uncultured), were typed, i.e., influenza A or B and if A, which subtype. Of the 347 total samples evaluated, 278 were correctly identified as being influenza type A (222) or B (56), and all type A were subsequently subtyped as either H1, H3, or H5. Influenza negative (69) samples (Tables 3 and 4, 2 from the District of Columbia, 66 from Nepal and 1 fromTexas, respectively) were subsequently confirmed as Coxsac kie B, Adenovirus, Parainfluenzas 1, 2, and 3 or virus negative (data not shown). Sixteen of the 100 uncultured primary clinical specimens (Table 4) required subsequent culturing. Although rRT-PCR is capable of detecting the presence of nucleic acid from non-viable virus, this was not the case since all 16 primary samples were successfully cultured and identified suggesting target RNA degradation from prolonged storage at -80°C, poor sample collection or lower sensitivity in clinical samples than that observed for serially diluted cRNA.

Assay specificity is underscored by the absence of any cross-reactivity, i.e., false positives (of particular note H5).

A A/anhui/1/2005 A/anhui/2/2005 A/anhui/2/2005 A/cambodi/3/J952a/2005 A/chicken/cuangxi/12/04 A/cambodi/3/J952a/2005 A/ck/Hong kong/220/97 A/ck/Hong kong/220/97 A/ck/Hong kong/28/97 A/Guangx1/1/2005 A/Hanoi/30408/2005 A/Hanoi/30408/2005 A/Hanoi/30408/2005 A/Iraq/207/2006 A/Thailand/16/2004 A/Turkey/15/2006 A/Vietnam/1203/2003 A/Vietnam/124/2005 A/Womgolia/244/2005* A/Indonesia/S81h/2006 A/Indonesia/S81h/2005






Figure 1. Real-time RT-PCR analysis using serially diluted H5 target cRNA. Real-time RT-PCR analysis was carried out as previously described under 'Patients/methods.' Frame A: primer/probe specificity is representative of H5 human and avian influenza strains from H5 clades 1 and 2 (subclades 1, 2, and 3). *Candidate H5 vaccine reference strains. Forward and reverse primers (sense) are indicated by the green arrows. The red box represents FAM-labeled fluorogenic probe. Frame B: serial dilutions $(10^{-9} \text{ to } 10^{-16} \text{ g})$ of influenza type A H5 standard RNA. An average of 340 Da was used for all four bases (single strand) and 1658 bases for the RNA standard corresponding to 5.78×10^5 Da. Serially diluted RNA standard, i.e., 10^{-16} g corresponds to approximately 100 RNA molecules.

False negatives can arise for a variety of reasons, i.e., presence of RT-PCR inhibitors, low initial viral titer, RNA degradation, poor/low RNA recovery during extraction, user error, or reagent degradation. Although no notable differences in RNA recovery have been observed using manually extracted template compared to robotic extracted template (data not shown), inclusion of an internal positive control could be of value in monitoring the process from extraction through amplification.²⁸

The sensitivity of type and subtype specific assays described in this report using primary, uncultured speci-

mens supports 'point of care' value although detection sensitivity using clinical material could be less than that observed for serially diluted cRNA. However, the robustness of the human H1 and H3 subtype specific assays described in this report is evidenced by successful detection of viruses from several past outbreaks that have subsequently been shown by sequence analysis to have drifted. The majority of H1N1 viruses circulating worldwide since 1999 are genetically and antigenically similar to A/New Caledonia/20/99 vaccine strain.²⁶ During the 2001/02 season and thereafter, reassortment of influenza A H1N2 viruses was observed in humans; however, the H1N2 strain is genetically similar to H1 New Caledonia-like viruses.²⁹ While H1N1 viruses were negligible during the 2004/05 season, an increasing number of human H1N1 viruses (A/New Caledonia-like strains) were detected worldwide during the 2005/06 season. Subsequently, during the 2006/07 season, an H1N1 drift variant constituting a new lineage represented by the A/Solomon Islands/3/2006 strain has emerged prompting a change in the 2007/08 H1 vaccine strain component. In similar fashion, the H3N2 virus co-circulating with H1N1 since 1977 has undergone 4 vaccine strain iterations since 1999.³⁰ The H1 and H3 subtype specific assays reported here were designed and empirically evaluated using 100 plus H1 and H3 vaccine and field strain sequences collected since 2001. Although H1 and H3 lineages exhibit genetic diversity, these changes do not occur within the H1 and H3 specific assay primer/probe binding regions and thus detection of genetically diverse viruses is not compromised.

Since 1997, genetic diversity has also been observed in H5 viruses in humans.³¹⁻³⁴ Human H5 outbreaks during 2004 and 2005 revealed two distinct clades of which the majority of human isolates characterized belonged to Clade 1. However, the number of reported H5 N1 Clade 2 human infections increased in late 2005. Although current H5 viruses have drifted from primer/probe sequences originally described by Spackman,²³ critical sequence analysis of human and avian sequences (22, cf. Fig. 1, panel A) used in this study revealed 100% primer/probe sequence homology supporting detection of both H5 Clades 1 and 2 (subclades 1, 2, and 3).³¹⁻³⁴ Furthermore, of particular note is the absence of H5 false positives observed in 167 original specimens and 180 cultures. However, due to empirical evaluation using a single H5 sample in this study, we cannot rule out the possibility of encountering false negatives using the H5 subtype specific assay. Thus, additional testing is needed to establish firmly the usefulness of the H5 specific assay described in this report.

Although the current FDA-approved LRN assays for influenza virus type A and H5 subtyping have been wellreceived, they do not (i) detect all 16 type A influenza viruses as well as both Yamagata and Victoria type B influenza lineages and (ii) differentiate currently circulating H1 and H3 influenza A viruses. Furthermore, the restricted nature and use of the LRN assay reagents in only a small number of LRN approved laboratories as well as restricted access to commercially available assays underscores the importance of needed additional assays for influenza virus detection in the event of a worldwide influenza outbreak.

Acknowledgments

This work was supported in part from NIH Grant 3S06GM008194-25S1 and funding from The Center of

Excellence in Biotechnology, Bioprocessing, Education, and Research (CEBBER), The University of Texas at San Antonio. Additionally, we wish to acknowledge Drs. Joel Gaydos, Rodney L Coldren, and Forrest Littlebird and the DoD Global Emerging Infectious Surveillance (DoD-GEIS) for their contributions and support of this work.

References

- 1 Yohannes K, Roche P, Spencer J, Hampson A. Annual report of the National Influenza Surveillance Scheme. Commun Dis Intell 2003;27:162–172.
- **2** Horimoto T, Kawaoka Y. Pandemic threat posed by avian influenza A viruses. Clin Microbiol Rev 2001;14:129–149.
- **3** Lin YP, Shaw M, Gregory V *et al.* Avian-to-human transmission of H9N2 subtype influenza A viruses: relationship between H9N2 and H5N1 human isolates. Proc Natl Acad Sci USA 2000;97:9654–9658.
- 4 Webby RJ, Webster RG. Are we ready for pandemic influenza? Science 2003;302:1519–1522.
- 5 Heinen P. Swine influenza: a zoonosis. Veterinary Sciences Tomorrow 2003; 1–11.
- **6** Kendal AP, Pereura MS, Skehel J. Concepts and procedures for laboratory based influenza surveillance. Geneva, Switzerland: World Health Organization, 1982.
- **7** Sato S, Ochiai H, Niwayama S. Application of the single radial complement fixation test for serodiagnosis of influenza, respiratory syncytial, mumps, adeno type 3, and herpes simplex type 1 virus infections. J Med Virol 1988;24:395–404.
- 8 Stamboulian D, Bonvehi PE, Nacinovich FM, Cox N. Influenza. Infect Dis Clin North Am 2000;14:141–166.
- **9** Cattoli G, Drago A, Maniero S *et al.* Comparison of three rapid detection systems for type A influenza virus on tracheal swabs of experimentally and naturally infected birds. Avian Pathol 2004;33:432–437.
- 10 Kaiser L, Briones MS, Hayden FG. Performance of virus isolation and Directigen Flu A to detect influenza A virus in experimental human infection. J Clin Virol 1999;14:191–197.
- **11** Weinberg A, Walker ML. Evaluation of three immunoassay kits for rapid detection of influenza virus A and B. Clin Diagn Immunol 2005;12:367–370.
- **12** Woolcock PR, Cardona CJ. Commercial immunoassay kits for the detection of influenza virus type A: evaluation of their use with poultry. Avian Dis 2005;49:477–481.
- **13** Daum LT, Canas LC, Shadler CA *et al.* A rapid, single-step multiplex reverse transcription-PCR assay for the detection of human H1N1, H3N2, and B influenza viruses. J Clin Virol 2002;25:345–350.
- **14** Poddar SK. Influenza virus types and subtypes detection by single step single tube multiplex reverse transcription-polymerase chain reaction (RT-PCR) and agarose gel electrophoresis. J Virol Methods 2002;99:63–70.
- 15 Zhang WD, Evans DH. Detection and identification of human influenza viruses by the polymerase chain reaction. J Virol Methods 1991;33:165–189.
- 16 Atmar RL, Baxter BD, Dominguez EA, Taber LH. Comparison of reverse transcription-PCR with tissue culture and other rapid diagnostic assays for detection of type A influenza virus. J Clin Microbiol 1996;34:2604–2606.
- 17 Ellis JS, Zambon MC. Molecular diagnosis of influenza. Rev Med Virol 2002;12:375–389.
- 18 Fouchier RA, Bestebroer TM, Herfst S, Van der Kemp L, Rimmelzwaan GF, Osterhaus AD. Detection of influenza A viruses from

different species by PCR amplification of conserved sequences in the matrix gene. J Clin Microbiol 2000;38:4096–4101.

- **19** Schweiger B, Zadow I, Heckler R, Timm H, Pauli G. Application of a fluorogenic PCR assay for typing and subtyping of influenza viruses in respiratory samples. J Clin Microbiol 2000;38:1552–1558.
- **20** Van Elden L, Nijhuis P, Schipper P, Schuurman R, Van Loon AM. Simultaneous detection of influenza viruses A and B using real-time quantitative PCR. J Clin Microbiol 2001;39:196–200.
- **21** Lee CW, Suarez DL. Application of real-time RT-PCR for the quantitation and competitive replication study of H5 and H7 subtype avian influenza virus. J Virol Methods 2004;119:151–158.
- 22 http:// www.hhs.gov/news/press/2006pres/20060203.html.
- **23** Spackman E, Senne DA, Myers TJ *et al.* Development of a real-time reverse transcriptase PCR assay for type A influenza virus and the avian H5 and H7 hemagglutinin subtypes. J Clin Microbiol 2002;40:3256–3260.
- 24 Morbidity and Mortality Weekly Report, New Laboratory Assay for Diagnostic Testing of Avian Influenza A/H5 (Asian Lineage). February 10, 2006; 55(05): 127.
- **25** Di Trani L, Bedini B, Donatelli I *et al.* A sensitive one-step real-time PCR for detection of avian influenza viruses using a MGB probe and an internal positive control. BMC Infect Dis 2006;6:87–100.
- **26** Daum LT, Canas LC, Smith CB *et al.* Genetic and antigenic analysis of the first A/New Caledonia/20/99-like H1N1 influenza isolates reported in the Americas. Emerg Infect Dis 2002;8:408–412.

- **27** Fouchier RA, Munster V, Wallensten A *et al.* Characterization of a novel influenza A virus hemagglutinin subtype (H16) obtained from black-headed gulls. J Virol 2005;79:2814–2822.
- 28 Das A, Spackman E, Senne D, Pedersen J, Suarez DL. Development of an internal positive control for rapid diagnosis of avian influenza virus infections by real-time reverse transcription-PCR with lyophilized reagents. J Clin Microbiol 2006;44:3065–3073.
- **29** Xu X, Smith CB, Mungall BA *et al.* Intercontinental circulation of human influenza A(H1N2) reassortant viruses during the 2001-2002 influenza season. J Infect Dis 2002;186:1490–1493.
- **30** Daum LT, Shaw MW, Klimov Al *et al.* Influenza A (H3N2) outbreak, Nepal. Emerg Infect Dis 2005;11:1186–1191.
- **31** World Health Organization Global Influenza Program Surveillance Network. Evolution of H5N1 avian influenza viruses in Asia. Emerg Infect Dis 2005;11:1515–1521.
- **32** Chan PK. Outbreaks of avian influenza A (H5N1) virus infection in Hong Kong in 1997. Clin Infect Dis 2002;34:558–564.
- **33** Shinya K, Hatta M, Yamada S *et al.* Characterization of a human H5N1 influenza A virus isolated in 2003. J Virol 2005;79:9926–9932.
- **34** Webster RG, Govorkova EA. H5N1 Influenza-Continuing Evolution and Spread. N Engl J Med 2007;355:2174–2177.