


AI-generated B-line counts



12 zone lung ultrasound protocol

Performance of POCUS platform for AI enabled assessment of pulmonary B-lines

- ❖ Single center study with 55 patients using a 12 zone lung ultrasound protocol
- ❖ An automatic B-line tool in a major vendor was utilized
- ❖ AI B-line counts were compared to two expert reviewers using intraclass correlation coefficient (ICC)

ICC between AI and experts



ICC between experts

0.92

