

Available online at www.sciencedirect.com

ScienceDirect

journal homepage: www.jfda-online.com

Original Article

Designing a biochip following multiplex polymerase chain reaction for the detection of *Salmonella* serovars Typhimurium, Enteritidis, Infantis, Hadar, and Virchow in poultry products



Yu-Cheng Chiang^{a,*}, Hsien-Huang Wang^b, Latha Ramireddy^a,
Hsin-Yen Chen^a, Chia-Ming Shih^a, Chien-Ku Lin^a, Hau-Yang Tsen^{a,*}

^a Department of Food Science and Technology, Hung Kuang University, Number 1018, Section 6, Taiwan Boulevard, Shalu District, Taichung City 43302, Taiwan, ROC

^b DR. Chip Biotech, Science-Based Industrial Park, Chu-Nan, Miao-Li County 350, Taiwan, ROC

ARTICLE INFO

Article history:

Received 25 August 2016

Received in revised form

3 November 2016

Accepted 21 November 2016

Available online 7 March 2017

Keywords:

biochip for five *Salmonella* serovars
primers for *Salmonella* Virchow
poultry products

ABSTRACT

Salmonella-contaminated foods, especially poultry-derived foods (eggs, chicken meat), are the major source of salmonellosis. Not only in the European Union (EU), but also in the United States, Japan, and other countries, has salmonellosis been an issue of concern for food safety control agencies. In 2005, EU regulation 1003/2005 set a target for the control and reduction of five target *Salmonella enterica* serovars—*S. Typhimurium*, *S. Enteritidis*, *S. Infantis*, *S. Hadar*, and *S. Virchow*—in breeding flocks. Thus, a simple biochip for the rapid detection of any of these five *Salmonella* serovars in poultry products may be required. The objectives of this study were to design *S. Virchow*-specific primers and to develop a biochip for the simultaneous identification of all or any of these five *Salmonella* serovars in poultry and poultry products. Experimentally, we designed novel polymerase chain reaction (PCR) primers for the specific detection of *S. Virchow*, *S. Infantis*, and *S. Hadar*. The specificity of all these primers and two known primer sets for *S. Typhimurium* and *S. Enteritidis* was then confirmed under the same PCR conditions using 57 target strains and 112 nontarget *Salmonella* strains as well as 103 non-*Salmonella* strains. Following multiplex PCR, strains of any of these five *Salmonella* serovars could be detected by a chromogenic biochip deployed with DNA probes specific to these five *Salmonella* serovars. In comparison with the multiplex PCR methods, the biochip assay could improve the detection limit of each of the *Salmonella* serovars from $N \times 10^3$ cfu/mL to $N \times 10^2$ cfu/mL sample in either the pure culture or the chicken meat samples. With an 8-hour enrichment step, the detection limit could reach up to $N \times 10^0$ cfu/mL.

Copyright © 2017, Food and Drug Administration, Taiwan. Published by Elsevier Taiwan LLC. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

* Corresponding authors. Department of Food Science and Technology, Hung Kuang University, Number 34, Chung-Chi Road, Shalu, Taichung City 433, Taiwan, ROC.

E-mail addresses: honda224@gmail.com (Y.-C. Chiang), hytsen@sunrise.hk.edu.tw (H.-Y. Tsen).

<http://dx.doi.org/10.1016/j.jfda.2016.11.019>

1021-9498/Copyright © 2017, Food and Drug Administration, Taiwan. Published by Elsevier Taiwan LLC. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

1. Introduction

Salmonellosis has been one of the major foodborne diseases worldwide. So far, more than 2600 different *Salmonella* serotypes have been identified [1]. According to a global survey from the World Health Organization, during the years 2001 to 2007, *Salmonella* Enteritidis and *S. Typhimurium* were the most common serotypes found in North America, Australia, New Zealand, and other countries [2]. These *Salmonella* serotypes are responsible for foodborne salmonellosis in humans followed by *S. Infantis*, *S. Hadar*, and *S. Virchow* [2,3]. *S. Infantis* has been ranked as the third most frequently found serovar infecting humans [4]. However, *S. Virchow* has recently become one of the predominant serovars in EU countries, ranking among the top five serovars [5]. This *Salmonella* serovar is known to associate with invasive infection in humans and is resistant to many antibiotics [6,7]. Regarding foodborne poisoning cases and outbreaks, consumption of undercooked beef, poultry, and eggs are most often associated with salmonellosis [8].

Salmonellosis also has been one of the major foodborne outbreaks and/or diseases within EU countries. It causes significant economic losses; for example, the European Food Safety Authority has estimated the economic burden of human salmonellosis to cost 3 billion euros a year [9]. Thus, EU regulation 1003/2005, amending regulation no. EU2160/2003 (European Parliament and European Council, 2005), sets regulatory targets for the control and the reduction of five target *Salmonella* serovars—*S. Typhimurium*, *S. Enteritidis*, *S. Infantis*, *S. Hadar*, and *S. Virchow*—in breeding flocks of *Gallus gallus* [10]. Accordingly, these five *Salmonella* serovars were prioritized by the EU for the control of poultry and poultry products entry, because of the significant risk they pose to public health [11,12]. Good monitoring and screening programs are thus required to prevent *Salmonella* infection.

In addition to the biotyping method for *Salmonella* identification [13], *Salmonella* serovars are determined by lipopolysaccharide (O antigen) and flagellar structures (H antigen) based on the Kauffmann–White scheme. All these steps are labor intensive and time consuming, costly, and complicated [14]. Thus, a rapid and accurate screening for the five *Salmonella* serovars in food samples is required. Over the past decades, polymerase chain reaction (PCR) has been frequently used for the detection of *Salmonella* and other foodborne pathogens [2,15,16]. A number of genus- and serovar-specific genes have been used for designing primers or probes specific for the detection of *Salmonella* serovars Enteritidis, Typhimurium, Typhi, Choleraesuis, Paratyphi, Hadar, and other pathogens [17–25]. For *S. Infantis*, based on *fljB* gene, specific PCR primers also have been designed [26]. However, to our knowledge, no primer set or probe has been reported for the specific detection of *S. Virchow*.

The purpose of this study was thus to design novel PCR primers specific for *S. Virchow* detection. In addition, new PCR primers specific for *S. Infantis* and *S. Hadar* were also designed. Together with the known primers for *S. Typhimurium* and *S. Enteritidis*, target genes for all these five *Salmonella*

serovars could be amplified under the same PCR conditions. After hybridization of multiplex PCR (mPCR) products with species-specific probes deployed on the biochip, all or any of these five *Salmonella* serovars in poultry samples could be rapidly identified.

2. Materials and methods

2.1. Bacterial strains and culture conditions

Bacterial strains used in this study are listed in Table 1. These strains were obtained from the American Type Culture Collection (ATCC; Manassas, VA, USA); United States Department of Agriculture (USDA); Bioresource Collection and Research Center (BCRC; Hsin-Chu, Taiwan); Centers for Disease Control (CDC; Taichung Branch, Taichung, Taiwan); Department of Veterinary, National Ping Tung University (Ping Tung, Taiwan); and Culture Collection of the University of Göteborg (CCUG; Göteborg, Sweden). The strains were stored in Tryptic Soy Broth (TSB; Difco Laboratories, Detroit, MI, USA) containing 50% of glycerol at -80°C .

2.2. DNA preparation

One loopful of *Salmonella* strain was inoculated into 5 mL TSB and incubated at 37°C for 12 hours. Genomic DNA was extracted using Viogene, Blood & Tissue Genomic Mini kit (Viogene, Taipei, Taiwan). One milliliter of the culture was centrifuged for 5 minutes at 7000 rpm (Centrifuge 5424; Eppendorf, Hauppauge, NY, USA), and the pellet was washed by resuspension in 1 mL sterile water. After centrifugation again (5 minutes, 7000 rpm), the cell pellet was suspended in 250 μL lysostaphin buffer (2 mg/mL buffer; AMBI Products LLC, New York, NY, USA), 2.5 μL lysozyme (20 mg/mL; Sigma, St. Louis, MO, USA), and 20 μL RNase (20 mg/mL; Sigma). The mixture was incubated at 37°C for 1.5 hours, followed by addition of 200 μL EX Buffer (Viogene) and 25 μL proteinase K (10 mg/mL, Merck) and incubation at 60°C for 1 hour. After incubation at 70°C for another 30 minutes, total DNA was precipitated with alcohol and extracted according to the manufacturer's manual (Viogene). Afterward, genomic DNA was eluted with 200 μL double deionized water.

2.3. Designing PCR primers and probes

Whole genome sequences of *Salmonella* serovars were obtained from the National Center for Biotechnology Information (<https://ncbi.nlm.nih.gov>). Open reading frames of serovars *S. Infantis* (<ftp://ftp.sanger.ac.uk/pub/pathogens/Salmonella/SIN.dbs>) and *S. Hadar* (<ftp://ftp.sanger.ac.uk/pub/pathogens/Salmonella/HADAR.dbs>) obtained from an FTP site of Sanger Institute were previously reported by Akiba et al [27]. Open reading frames of *S. Hadar* (Hadar-specific genomic region) and *S. Infantis* (Infantis-specific genomic region) were compared using basic local alignment search tool (

Table 1 – Bacteria strains used in this study.

Species	Strain no.
<i>Salmonella</i> Enteritidis	ATCC 13076, CF09.001, CF09.002, CF09.003, CF09.009, CF09.011, CF09.012, CF09.013, CF09.017, CF09.018, CF09.023, US, SE07
<i>Salmonella</i> Typhimurium	BCRC 10747, BCRC 12947, CF09.015, CF09.016, CF09.019, CF09.020, CF09.021, CF09.037, CF09.039, SD09.016, SD09.018, SD09.023
<i>Salmonella</i> Infantis	CC07.003, CC07.016, CC07.063, CC07.090, NQ08.053, SA08.050, SA08.059, SB08.005, SB08.038, SD08.109, US, SI20
<i>Salmonella</i> Virchow	CA08.075, CA08.116, CA08.135, CA08.158, CF08.016, CF08.030, CF08.039, CF08.041, CF08.050, CF08.054
<i>Salmonella</i> Hadar	CA08.102, CA08.106, CA08.124, CA08.129, CA08.157, CA08.159, CA08.160, CB09.011, CF09.010, US, SH12; US, SH13
Other <i>Salmonella</i> serovars	<i>Salmonella abortusequi</i> , <i>S. Adelaide</i> , <i>S. Agona</i> , <i>S. Alachua</i> , <i>S. Albany</i> , <i>S. Amager</i> , <i>S. Anatum</i> , <i>S. Augustenborg</i> , <i>S. Azteca</i> , <i>S. Bareilly</i> , <i>S. Berta</i> , <i>S. Blockley</i> , <i>S. Braenderup</i> , <i>S. Bredeney</i> , <i>S. Boecker</i> , <i>S. Bonn</i> , <i>S. Bouss</i> , <i>S. Bovismorbificans</i> , <i>S. California</i> , <i>S. Cerro</i> , <i>S. Chailey</i> , <i>S. Chester</i> , <i>S. Chittagong</i> , <i>S. Choleraesuis</i> , <i>S. Coley Park</i> , <i>S. Colindale</i> , <i>S. Colorado</i> , <i>S. Crossness</i> , <i>S. Derby</i> , <i>S. Djakarta</i> , <i>S. Dublin</i> , <i>S. Dugbe</i> , <i>S. Emek</i> , <i>S. Essen</i> , <i>S. Florida</i> , <i>S. Gera</i> , <i>S. Goerlitz</i> , <i>S. Haifa</i> , <i>S. Halmstad</i> , <i>S. Havana</i> , <i>S. Houten</i> , <i>S. Hvittingfoss</i> , <i>S. Indiana</i> , <i>S. Isangi</i> , <i>S. Java</i> , <i>S. Javiana</i> , <i>S. Johannesburg</i> , <i>S. Kedougou</i> , <i>S. Kentucky</i> , <i>S. Litchfield</i> , <i>S. Lexington</i> , <i>S. Limete</i> , <i>S. Livingstone</i> var. O14, <i>S. London</i> , <i>S. Massena</i> , <i>S. Mbandaka</i> , <i>S. Meleagridis</i> , <i>S. Menhaden</i> , <i>S. Minnesota</i> , <i>S. Miami</i> , <i>S. Montevideo</i> , <i>S. Muenchen</i> , <i>S. Nchanga</i> , <i>S. Newington</i> , <i>S. Newport</i> , <i>S. Nigeria</i> , <i>S. Ohio</i> , <i>S. Panama</i> , <i>S. Paratyphi A</i> , <i>S. Paratyphi B</i> var. <i>Java</i> , <i>S. Potsdam</i> , <i>S. I6</i> , 7: —, <i>S. Rubislaw</i> , <i>S. Salamae</i> , <i>S. Saintpaul</i> , <i>S. Sandiego</i> , <i>S. Schwarzengrund</i> , <i>S. Senftenberg</i> , <i>S. Seremban</i> , <i>S. Singapore</i> , <i>S. Srkansas</i> , <i>S. Stanley</i> , <i>S. Tambacounda</i> , <i>S. Tennessee</i> , <i>S. Thompson</i> , <i>S. Thomasville</i> , <i>S. Typhi</i> , <i>S. Uganda</i> , <i>S. Victoria</i> , <i>S. Vejle</i> , <i>S. Weltevreden</i>
Non- <i>Salmonella</i> strains	<i>Acinetobacter baumannii</i> , <i>A. calcoaceticus</i> , <i>A. johnsonii</i> , <i>A. sp.</i> , <i>A. ursingii</i> , <i>Alcaligenes faecalis</i> , <i>Bacillus cereus</i> , <i>Brevibacterium linens</i> , <i>Citrobacter freundii</i> , <i>Clostridium acetobutylicum</i> , <i>C. difficile</i> , <i>C. haemolyticum</i> , <i>C. perfringens</i> , <i>Corynebacterium renale</i> , <i>Cronobacter sakazakii</i> , <i>Enterobacter aerogenes</i> , <i>E. cloacae</i> , <i>Erwinia carotovora</i> , <i>Escherichia coli</i> , <i>Hafnia alvei</i> , <i>Listeria innocua</i> , <i>L. monocytogenes</i> , <i>L. grayi</i> , <i>Moraxella catarrhalis</i> , <i>M. osloensis</i> , <i>Proteus vulgaris</i> , <i>Pseudomonas aeruginosa</i> , <i>P. fluorescens</i> , <i>P. mendocina</i> , <i>P. putida</i> , <i>Rahnella aquatilis</i> , <i>Saccharomyces cerevisiae</i> , <i>Streptomyces filipinensis</i> , <i>Serratia ficaria</i> , <i>S. fonticola</i> , <i>S. odorifera</i> , <i>S. quinivorans</i> , <i>Shigella dysenteriae</i> , <i>S. sonnei</i> , <i>Staphylococcus sp.</i> , <i>S. aureus</i> , <i>Streptococcus agalactiae</i> , <i>S. bovis</i> , <i>S. uberis</i> , <i>S. epidermidis</i> , <i>S. haemolyticus</i> , <i>S. hyicus</i> , <i>S. intermedius</i> , <i>S. saprophyticus</i> , <i>S. xylosum</i> , <i>Vibrio alginolyticus</i> , <i>V. parahaemolyticus</i> , <i>Yersinia enterocolitica</i>
ATCC = American Type Culture Collection (USA); BCRC = Bioresources Collection and Research Center (Taiwan); CDC = Centers for Disease Control (R.O.C. Taiwan); CCUG = Culture Collection, University of Göteborg (Sweden); PT = National Ping Tung University (Pingtung, Taiwan); USDA = United States Department of Agriculture; US = City of New York Department of Health (New York, USA).	

www.ncbi.nlm.nih.gov/BLAST). Serovar specific primers for *S. Infantis* and *S. Hadar* were then designed (Table 2).

For *S. Virchow*-specific primers, as *S. Virchow* was relatively common and consists of genes similar to those of *S. Infantis* [26], by comparison of the *S. Infantis* genome sequence with the partial sequences available for *S. Virchow*, primers that allowed the amplification of *S. Virchow* genome were designed using Primer premier 5.0 software (PREMIER Biosoft, Palo Alto, CA, USA). The sequence of the forward primer was 5'-CCTGATGAAAAGCGGACAAAG-3' and the reverse primer was 5'-GCCTTGATAATGCCTGTGCAGAA-3', corresponding to *S. Infantis* (Acc. No: NZ_LN649235.1) positions 3710138 to 3710158 and positions 3713529 to 3713507, respectively. These primers generate a PCR product from *S. Virchow* gene with a molecular size of approximately 3000 bp. This DNA fragment was purified with PCR-M cleanup system (Viogene) and sequenced. We then compared the sequence of this PCR product with the sequence of *S. Infantis* using

MUSCLE (<http://www.ebi.ac.uk/Tools/msa/muscle/>). Species-specific primers for *S. Virchow* were designed (Table 2). For *S. Typhimurium* and *S. Enteritidis*, primers previously reported to be specific to these two serovars were used [28,29]. These primers were synthesized by MDBio Inc. (Taipei, Taiwan). DNAs from all these five *Salmonella* serovars were then PCR assayed using thermal cycler 2720 (Perkin-Elmer Corporation, Norwalk, CT, USA).

2.4. PCR assay

The specificity of all primer sets was tested with DNA isolated from bacterial strains listed in Table 1. The PCR reactions were performed in a total volume of 25 µL mixture containing 200 µM of each deoxynucleoside triphosphate (PRO tech Technology Enterprise Co., Ltd., Taipei, Taiwan), 1 × PCR buffer (PRO tech Technology Ent. Co.), 0.2 µM each primer, 0.6 units of Prozyme (PRO tech Technology Ent. Co.), and 2 µL of

Table 2 – Sequences of primers and probes and the sizes of PCR products.

Strains	Target gene ^a	Primer	Size	Tm	Ref.
S. Enteritidis	Prot6e	ENTE-F: 5'-ATATCGTGTGCTGCTTCG-3' ENTE-R: 5'-CATTGTTCCACCGTCACTTTG-3' Probe/TTTTTTTTTTTTTTTTTTAGGCGCTCATCGGTCCTGCTGT	206	59	Malorny et al [29]
S. Typhimurium	MDH	TYPH-F: 5'-CGCATTCCACCACGCCCTTC-3' TYPH-R: 5'-TGCCAACGGAAGTTGAAGTG-3' Probe/TTTTTTTTTTTTTTTTTTGAAGGGCGTGGTGAATGCCG	261	59	Lin and Tsen [28]
S. Hadar	HSR3	HADA-F: 5'-CCTCAAATTAATGCCAGAGAG-3' HADA-R: 5'-GTAGCAATTTATGGCTTACTACA-3' Probe/AACCGAAACTTTACTGCGAGAGAGGTTTTTTTTTTTTTTTT	427	59	This study
S. Infantis	ISR2–ISR3	INFA1-F: 5'-GACGCTATCAATTCAAGCAGAC-3' In-Vi Uni-R: 5'-ATACGATACTACAATACCCGACG-3' INFA2-F: 5'-TGAGGGGAGAGGTTATTGTTAT-3' In-Vi Uni-R: 5'-ATACGATACTACAATACCCGACG-3' Probe/TACGCTGCTGTAATTGATAGCGTCTTTTTTTTTTTTTTT	268 240	59	This study
S. Virchow	Hypothetical protein	VIRC1-F: 5'-ATTGTTTATGCATAGGCCGAC-3' In-Vi Uni-R: 5'-ATACGATACTACAATACCCGACG-3' VIRC2-F: 5'-GTCAACCGATACTGTAGTAGTGCA-3' In-Vi Uni-R: ATACGATACTACAATACCCGACG-3' Probe/TTTTTTTTTTTTTTTTTTGTGCGCCTATGCATAAACAAT	273 234	59	This study

HSR = Hadar-specific genomic region; ISR = Infantis-specific genomic region; MDH = malate dehydrogenase; PCR = polymerase chain reaction.

^a Prot6e: S. Enteritidis-specific gene (fimbrial biosynthesis).

each target DNA (150–200 ng). The PCR conditions were as follows: an initial denaturation at 94°C for 5 minutes, followed by 35 cycles of 94°C for 30 seconds, 59°C for 30 seconds, 72°C for 30 seconds, and a final extension at 72°C for 7 minutes. PCR products were visualized by 2% agarose gel electrophoresis.

2.5. PCR detection limit for Salmonella

One loop of *Salmonella* Enteritidis ATCC 13076, *S. Typhimurium* BCRC 10747, *S. Hadar* CA08.102, *S. Infantis* CC07.016, and *S. Virchow* CA08.158, respectively, was cultured in 5 mL TSB followed by incubation at 37°C for 18 hours in a rotating shaker at 150 rpm. One hundred microliters of the cell suspension was then 10-fold serially diluted with sterile water. All dilutions (0.1 mL) were then plated on Tryptic Soy Agar, and colony-forming units (cfu/mL) were counted. In addition, DNA extracts from these cell dilutions were PCR assayed according to the conditions described earlier.

2.6. Detection of Salmonella in chicken meat

Chicken meat samples purchased from local food markets (Taichung, Taiwan) were used to determine the detection limit of each *Salmonella* species. The samples were prepared according to Chiang et al [30]. Briefly, 25 g of raw or sterilized (121°C for 15 minutes) chicken meat samples were homogenized with 225 mL of 1% Bacto Peptone water using a Stomacher 400 (Seward, London, UK) at high speed for 1 minute. Then, 1 mL of the homogenized broth was mixed with 8.9 mL TSB, then 0.1 mL of *Salmonella* cells with known cell counts, i.e., 0 cfu/mL and $N \times 10^0$ – 10^7 cfu/mL ($N = 1$ – 9) were inoculated to the mixture followed by incubation at 37°C for 0 hour or 8 hours. DNA was then extracted from each of these samples and assayed by PCR under the same PCR conditions.

2.7. Biochip detection of Salmonella

Five sets of primers including one of the two primer sets for *S. Infantis* and *S. Virchow*, respectively, were used for mPCR followed by biochip assay. These primers were ENTE-F/ENTE-R (206 bp), TYPH-F/TYPH-R (261 bp), INFA1-F/InVi-R (268 bp), VIRC1-F/InVi-R (273 bp), and HADA-F/HADA-R (427 bp). Primers and probe sequences are listed in Table 2. In addition, primers at the 5'-end were labeled with biotin and subjected to PCR assay. These biotin-labeled PCR amplicons were used for the biochip assay.

For biochip construction, 25 µL of each oligonucleotide probe (10–20 µM) was mixed with 25 µL of 2× probe solution (DR. Chip Biotech, Hsin-Chu, Taiwan). It was then spotted onto the plastic chip (DR. Chip Biotech) by using the Micro-Arrayer Ezspot SR-A300 (Shuai Ran Precision, Taoyuan, Taiwan). For positive control, only the biotin-labeled primers were spotted, whereas for negative control, none of the probes (buffer only) were spotted on the chip (Figure 1). The amplified DNA was then cross linked onto the plastic chip by ultraviolet irradiation with UV CROSSLINKER CL-508.S (UVitec, Cambridge, England, UK) at 254 nm/0.6–1.2 J. For hybridization, 25 µL of each biotin-labeled PCR products was heated at 95°C for 10 minutes and immediately cooled in ice-water bath for 5 minutes. Next, 200 µL of DR. Hyb Buffer (DR. Chip Biotech) and the denatured PCR products were mixed and loaded into a chip chamber well. Hybridization was carried out for 30 minutes at 50°C with gentle vibration in the DR. Mini Oven (DR. Chip Biotech). Wash steps and chromogenic reaction were carried out as previously described [30]. Next, the hybridization signals were determined by scanning with DR. AiM Reader (DR. Chip Biotech). To determine the detection limit and specificity of the biochip for *Salmonella* in pure culture or in raw or sterilized chicken meat samples, similar steps as described earlier were followed.




Figure 1 – Detection of the *Salmonella* strains using the biochip. Conditions for the biochip hybridization were described in Materials and methods. (A) Allocation of specific oligonucleotide probes on the biochip. Sequences of probes are listed in Table 2. (B) Simultaneous detection of different combinations of *Salmonella* serovars in chicken meat samples by biochip (1–26). The *Salmonella* counts spiked for each serovar were 10^4 cfu/mL of chicken homogenate. (1) E, T; (2) E, I; (3) E, V; (4) E, H; (5) T; (6) T, H; (7) T, H; (8) I, V; (9) I, H; (10) V, H; (11) E, T, I; (12) E, T, V; (13) E, T, H; (14) E, I, H; (15) E, V, H; (16) E, I, V; (17) T, I, V; (18) T, I, H; (19) T, V, H; (20) I, V, H; (21) E, T, I, V; (22) E, T, I, H; (23) E, T, V, H; (24) T, I, V, H; (25) E, I, V, H; (26) E, T, I, V, H. E = Enteritidis, H = Hadar, I = Infantis; V = Virchow; T = Typhimurium.

3. Results

3.1. Test specificity of the primer sets

The specificity of *S. Typhimurium*, *S. Enteritidis*, *S. Virchow*, *S. Hadar*, and *S. Infantis* primers was tested by PCR assay against 12 *S. Typhimurium*, 12 *S. Enteritidis*, 10 *S. Virchow*, 11 *S. Hadar*, 11 *S. Infantis*, 182 nontarget *Salmonella* bacteria strains, and 103 non-*Salmonella* strains under the same PCR conditions with an annealing temperature of 59°C (Table 3). The amplification size was 206 bp for *S. Enteritidis* and 261 bp for *S. Typhimurium*. Two primer sets for *S. Infantis* and two sets for *S. Virchow* were designed, and their specificity was checked by PCR (Table 2). All these four primer sets share the same reverse primer, i.e., In-Vi Uni-R (Table 2). For *S. Infantis*, two

amplified products, i.e., 268 bp and 240 bp, and for *S. Virchow* two products, i.e., 273 bp and 234 bp, could be generated. For *S. Hadar*-specific primers, *S. Hadar* strains generated PCR products with a size of 427 bp. All primers generated positive results to their target strains, whereas nontarget *Salmonella* strains and strains other than *Salmonella* generated negative results (Table 3).

3.2. Multiplex PCR detection of *Salmonella* serovars

The annealing temperature used for all five primer sets shown in Table 2 was similar, i.e., 59°C. Thus, these five primer sets could be used in a multiplex PCR (mPCR) assay under the same PCR conditions. This mPCR system allowed the amplification of each of the five target *Salmonella* serovars as shown in Figure 2. Because of the closely similar

Table 3 – Specificity of the PCR primers^a for the detection of *Salmonella* Enteritidis, *S. Typhimurium*, *S. Infantis*, *S. Virchow*, and *S. Hadar*.

Bacteria species	Total no. of strains ^b	PCR result with						
		<i>Salmonella</i> Enteritidis	<i>Salmonella</i> Hadar	<i>Salmonella</i> Typhimurium	<i>Salmonella</i> Infantis	<i>Salmonella</i> Virchow		
		ENT-F/ ENT-R	HAD-F/ HAD-R	TYP-F/ TYP-R	INF1-F/ InVi-R	INF2-F/ InVi-R	VIR1-F/ InVi-R	VIR2 F/ InVi-R
<i>Salmonella</i> Enteritidis	12	12+	—	—	—	—	—	—
<i>Salmonella</i> Hadar	11	—	11+	—	—	—	—	—
<i>Salmonella</i> Typhimurium	12	—	—	12+	—	—	—	—
<i>Salmonella</i> Infantis	11	—	—	—	11+	11+	—	—
<i>Salmonella</i> Virchow	10	—	—	—	—	—	10+	10+
Nontarget <i>Salmonella</i> strains	182	—	—	—	—	—	—	—
Non- <i>Salmonella</i> strains	103	—	—	—	—	—	—	—

PCR = polymerase chain reaction.

^a Assay conditions were as those described in Materials and methods. For each assay, single primer sets specific for each *Salmonella* serovar was used.

^b The name and source of the strains are shown in Table 1.

sizes of some PCR products, the mPCR assays were performed with two combination groups, e.g., two serovars (Figure 2A) and multiple serovars (Figure 2B). However, because the molecular sizes of the PCR products for *S. Typhimurium* (261 bp), *S. Infantis* (268 bp), and *S. Virchow* (273 bp) were closely similar, only two or three bands on agarose gel could be observed. For example, in lanes 6–9 of Figure 2B, only three bands among all four or five *Salmonella* serovars could be observed. In addition, some PCR products might generate faint bands. Under such PCR conditions, such PCR system did not allow us to discriminate among these three *Salmonella* serovars. In addition, the detection limit of each of the primer sets to *Salmonella* cells in PBS or in raw or sterilized chicken homogenate samples was $N \times 10^3$ cfu/mL. With an 8-hour preenrichment step, the detection limit for each of these five *Salmonella* serovars could reach $N \times 10^0$ cfu/mL either in PBS or in chicken meat samples (Table 4).

3.3. Use of biochip following mPCR for the discrimination of all five *Salmonella* serovars

Because the mPCR method was unable to discriminate all five *Salmonella* serovars simultaneously, we thus designed a biochip following this mPCR step. The biochip was constructed by spotting the oligonucleotide probes listed in Table 2 for the detection of five *Salmonella* serovars. Allocation of the probes on the biochip is demonstrated in Figure 1. All positive control generated positive signals, whereas all negative control generated negative results. Figure 1 shows the hybridization patterns for multiple *Salmonella* targets artificially spiked in chicken samples. The spiked cells for each *Salmonella* serovar in chicken homogenate was $N \times 10^4$ cfu/g sample. For either the raw or the sterilized sample, similar results could be obtained. All combinations of *Salmonella* serovar generated the expected hybridization patterns, and no false hybridization signal was observed.




Figure 2 – Detection of the *Salmonella* strains by mPCR using ENTE-F/ENTE-R (206 bp), TYPH-F/TYPH-R (261 bp), INFA1-F/InVi-R (268 bp), VIRC1-F/InVi-R (273 bp), HADA-F/HADA-R (427 bp) primers. (A) Lane M: 100 bp ladder; lane 1: E, T; lane 2: E, H; lane 3: E, I; lane 4: E, V; lane 5: T, H; lane 6: T, I; lane 7: T, V; lane 8: H, I; lane 9: V, H; lane 10: I, V. (B) Lane M: 100 bp ladder; lane 1: E, T, H; lane 2: T, I, H; lane 3: I, V, H; lane 4: E, I, V; lane 5: E, V, H; lane 6: E, T, I, H; lane 7: T, I, V, H; lane 8: E, T, I, V; lane 9: E, T, I, V, H. E = Enteritidis; H = Hadar; I = Infantis; T = Typhimurium; V = Virchow.

Table 4 – Detection limit for poultry meat samples by multiplex PCR and biochip.

Strain	Multiplex PCR				Biochip			
	RAW chicken meat (CFU/mL)		Sterilized chicken meat (CFU/mL)		Raw chicken meat (CFU/mL)		Sterilized chicken meat (CFU/mL)	
	Without enrichment	With enrichment ^b	Without enrichment	With enrichment ^b	Without enrichment	With enrichment ^b	Without enrichment	With enrichment ^b
<i>Salmonella</i> Enteritidis	$N \times 10^3$	$N^a \times 10^0$	$N \times 10^3$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$
<i>Salmonella</i> Typhimurium	$N \times 10^3$	$N \times 10^0$	$N \times 10^3$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$
<i>Salmonella</i> Infantis	$N \times 10^3$	$N \times 10^0$	$N \times 10^3$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$
<i>Salmonella</i> Virchow	$N \times 10^3$	$N \times 10^0$	$N \times 10^3$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$
<i>Salmonella</i> Hadar	$N \times 10^3$	$N \times 10^0$	$N \times 10^3$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$	$N \times 10^2$	$N \times 10^0$

Under the multiplex PCR conditions five primer sets were used but single *Salmonella* serovar was used for each assay.

PCR = polymerase chain reaction.

^a $N = 1-9$.

^b Eight-hour enrichment.

When the detection limit of the chip was evaluated with raw and sterilized chicken homogenate spiked with *Salmonella* cells, the detection limit was $N \times 10^2$ cfu/mL for each *Salmonella* serovar. In comparison with the detection limit of mPCR, the detection limit could be improved from $N \times 10^3$ cfu/mL to $N \times 10^2$ cfu/mL by following the use of the biochip. With an 8-hour enrichment step, the detection limit also reached $N \times 10^0$ cfu/mL (Table 4). These results demonstrate that the biochip may serve for specific and highly sensitive detection of these five *Salmonella* serovars.

4. Discussion

For rapid and simultaneous detection of different *Salmonella* serovars, mPCR with multiple species-specific primers has been used [29]. However, in some cases, some PCR products may generate faint signals or may be undiscriminated because of the closely similar molecular sizes. For example, Akiba et al [27] developed an mPCR system to identify seven major serovars of *Salmonella*, i.e., Typhimurium, Choleraesuis, Infantis, Hadar, Enteritidis, Dublin, and Gallinarum. However, *S. Virchow* was not included in the serovars investigated, and few false-positive results were observed in the mPCR assays used to identify Typhimurium, Choleraesuis, Enteritidis, and Dublin. Moreover, because of the closely similar sizes of some PCR products, not all of the seven *Salmonella* serovars could be simultaneously detected. In this study, we developed an mPCR method to amplify the DNAs from five target *Salmonella* serovars. Following mPCR, we used a biochip that was capable of simultaneous and specific detection of all or any of these five *Salmonella* serovars (Figure 1). To the best of our knowledge, this is the first biochip designed for the simultaneous detection of all of these five *Salmonella* serovars. Compared to the use of only mPCR, this biochip allows users to double check the detection specificity of these *Salmonella* serovars because all the primers used for mPCR and the probes deployed on the chip are specific to these five *Salmonella*

serovars. Furthermore, this mPCR–biochip system could further improve the strain discrimination and the detection limit as compared to only the use of mPCR.

Use of DNA biochip following PCR using either universal primers or multiplex primers offers a powerful tool for simultaneous detection of multiple targets in microbial diagnosis. For example, McCabe et al [31] used 16S rRNA-based universal primers to amplify template DNA in filter spots containing boiled bacteria followed by hybridization with species-specific probes and identified 14 clinical isolates. Wang et al [32] also used 16S rRNA gene-based universal primers to amplify the variable region of bacterial 16S rRNA gene, followed by reverse hybridization of the products to species-specific probes on a chip. As for the use of multiplex PCR to amplify the target bacteria genes followed by the use of biochips, such techniques have been used for the rapid detection of different bacteria genus or species based on different genes [19,29,30,33], and different virulence genes, such as the enterotoxin types, of *Staphylococcus aureus* [34]. Ideally, many different pathogenic bacteria species, toxin, or virulence genes could be simultaneously detected as long as the quantity of each target DNA is enough for PCR and the biochip used is highly sensitive. However, because the purpose of this study was to offer a simple and convenient method for the rapid detection of the five *Salmonella* serovars set by EU regulation, only five *Salmonella* serovars were set as targets of this biochip.

In conclusion, in this study, we designed a biochip for five major *Salmonella* serovars. Because salmonellosis has been a global issue of concern among food control agencies, the biochip we developed may offer a rapid and reliable method to monitor these salmonella cells in not only in poultry or poultry products, but also in other food samples as well as clinical samples. Finally, a reminder should be made that before this biochip can be recognized as an official method (e.g., Association of Official Analytical Chemists method in the United States), more studies designed according to the guidelines of official methods may be needed.

Conflicts of interest

All authors declare that they have no conflicts of interest.

Acknowledgments

This study was funded by the grants from National Science Council, Taipei, Taiwan (Project no: NSC 99-2324-B-241-002-CC1 and NSC 102-2632-B-241-001-MY3), and also DR. Chip Biotech Inc. (Miao-Li County, Taiwan).

REFERENCES

- [1] Guibourdenche M, Roggentin P, Mikoleit M, Fields PI, Bockemühl J, Grimont PA, Weill FX. Supplement 2003–2007 (No. 47) to the White–Kauffmann–Le Minor scheme. *Res Microbiol* 2010;161:26–9.
- [2] Hendriksen RS, Vieira AR, Karlsmose S, Lo Fo Wong DM, Jensen AB, Wegener HC, Aarestrup FM. Global monitoring of *Salmonella* serovar distribution from the World Health Organization Global Foodborne Infections Network Country Data Bank: results of quality assured laboratories from 2001 to 2007. *Foodborne Pathog Dis* 2011;8:887–900.
- [3] Abubakar I, Irvine L, Aldus CF, Wyatt GM, Fordham R, Schelenz S, Shepstone L, Howe A, Peck M, Hunter PR. A systematic review of the clinical, public health and cost-effectiveness of rapid diagnostic tests for the detection and identification of bacterial intestinal pathogens in faeces and food. *Health Technol Assess* 2007;11:1–216.
- [4] Galanis E, Lo Fo Wong DM, Patrick ME, Binsztein N, Cieslik A, Chalermchikit T, Aidara-Kane A, Ellis A, Angulo FJ, Wegener HC. Web-based surveillance and global *Salmonella* distribution, 2000–2002. *Emerg Infect Dis* 2006;12:381–8.
- [5] Bonalli M, Stephan R, Kappeli U, Cernela N, Adank L, Hachler H. *Salmonella enterica* serotype Virchow associated with human infections in Switzerland: 2004–2009. *BMC Infect Dis* 2011;11:49.
- [6] Mani V, Brennand J, Mandal BK. Invasive illness with *Salmonella* virchow infection. *Br Med J* 1974;2:143–4.
- [7] Jones TF, Ingram LA, Cieslak PR, Vugia DJ, Tobin-D'Angelo M, Hurd S, Medus C, Cronquist A, Angulo FJ. Salmonellosis outcomes differ substantially by serotype. *J Infect Dis* 2008;198:109–14.
- [8] DuPont HL. The growing threat of foodborne bacterial enteropathogens of animal origin. *Clinic Infect Dis* 2007;45:1353–61.
- [9] EFSA (European Food Safety Authority). EFSA explains zoonotic diseases. 2014.
- [10] European Commission. Regulation (EC) No. 1003/2005 of the Commission of 30 June 2005 implementing Regulation (EC) No. 2160/2003 of the European Parliament and of the Council as regards a community target for the reduction of the prevalence of certain *Salmonella* serotypes in breeding flocks of *Gallus gallus* and amending Regulation (EC) No. 2160/2003. *Official Journal L* 2005;170:12–7.
- [11] Snow LC, Davies RH, Christiansen KH, Carrique-Mas JJ, Wales AD, O'Connor JL, Cook AJ, Evans SJ. Survey of the prevalence of *Salmonella* species on commercial laying farms in the United Kingdom. *Vet Rec* 2007;161:471–6.
- [12] Arnold ME, Papadopoulou C, Davies RH, Carrique-Mas JJ, Evans SJ, Hoinville LJ. Estimation of *Salmonella* prevalence in UK egg-laying holdings. *Prev Vet Med* 2010;94:306–9.
- [13] Andrews WH, Jacobson A, Hammack TS. *Salmonella*. Bacteriological analytical manual (BAM). Silver Spring, MD: US Food and Drug Administration; 2007. p. 1–21.
- [14] Fitzgerald C, Sherwood R, Gheesling LL, Brenner FW, Fields PI. Molecular analysis of the rfb O antigen gene cluster of *Salmonella enterica* serogroup O:6,14 and development of a serogroup-specific PCR assay. *App Environ Microbiol* 2003;69:6099–105.
- [15] Pathmanathan SG, Cardona-Castro N, Sanchez-Jimenez MM, Correa-Ochoa MM, Puthuchery SD, Thong KL. Simple and rapid detection of *Salmonella* strains by direct PCR amplification of the hila gene. *J Med Microbiol* 2003;52:773–6.
- [16] Chiu TH, Chen TR, Hwang WZ, Tsen HY. Sequencing of an internal transcribed spacer region of 16S–23S rRNA gene and designing of PCR primers for the detection of *Salmonella* spp. in food. *Int J Food Microbiol* 2005;97:259–65.
- [17] Wood MW, Mahon J, Lax AJ. Development of a probe and PCR primers specific to the virulence plasmid of *Salmonella* Enteritidis. *Mol Cell Probes* 1994;8:473–9.
- [18] Woods DF, Reen FJ, Gilroy D, Buckley J, Frye JG, Boyd EF. Rapid multiplex PCR and real-time TaqMan PCR assays for detection of *Salmonella enterica* and the highly virulent serovars Choleraesuis and Paratyphi C. *J Clin Microbiol* 2008;46:4018–22.
- [19] Pan TM, Liu YJ. Identification of *Salmonella* Enteritidis isolates by polymerase chain reaction and multiplex polymerase chain reaction. *J Microbiol Immunol Infect* 2002;35:147–51.
- [20] Chiang YC, Yang CY, Li C, Ho YC, Lin CK, Tsen HY. Identification of *Bacillus* spp., *Escherichia coli*, *Salmonella* spp., *Staphylococcus* spp. and *Vibrio* spp. with 16S ribosomal DNA-based oligonucleotide array hybridization. *Int J Food Microbiol* 2006;107:131–7.
- [21] Park SH, Kim HJ, Cho WH, Kim JH, Oh MH, Kim SH, Lee BK, Ricke SC, Kim HY. Identification of *Salmonella enterica* subspecies I, *Salmonella enterica* serovars Typhimurium, Enteritidis and Typhi using multiplex PCR. *FEMS Microbiol Lett* 2009;301:137–46.
- [22] Shanmugasundaram M, Radhika M, Murali HS, Batra HV. Detection of *Salmonella enterica* serovar Typhimurium by selective amplification of fliC, fljB, iroB, invA, rfbJ, STM2755, STM4497 genes by polymerase chain reaction in a monoplex and multiplex format. *World J Microbiol Biotechnol* 2009;25:1385–94.
- [23] Salehi TZ, Tadjbakhsh H, Atashparvar N, Nadalian MG, Mahzounieh MR. Detection and identification of *Salmonella* Typhimurium in bovine diarrhoeic fecal samples by immunomagnetic separation and multiplex PCR assay. *Zoonoses Public Health* 2007;54:231–6.
- [24] Charlton BR, Walker RL, Kinde BH, Bauer CR, Channing-Santiago SE, Farver TB. Comparison of a *Salmonella* Enteritidis-specific polymerase chain reaction assay to delayed secondary enrichment culture for the detection of *Salmonella* Enteritidis in environmental drag swab samples. *Avian Dis* 2005;49:418–22.
- [25] Hong Y, Liu T, Lee MD, Hofacre CL, Maier M, White DG, Ayers S, Wang L, Berghaus R, Maurer JJ. Rapid screening of *Salmonella enterica* serovars Enteritidis, Hadar, Heidelberg and Typhimurium using a serologically-correlative allelotyping PCR targeting the O and H antigen alleles. *BMC Microbiol* 2008;8:178.
- [26] Kardos G, Farkas T, Antal M, Nogrady N, Kiss I. Novel PCR assay for identification of *Salmonella enterica* serovar Infantis. *Lett Appl Microbiol* 2007;45:421–5.
- [27] Akiba M, Kusumoto M, Iwata T. Rapid identification of *Salmonella enterica* serovars, Typhimurium, Choleraesuis, Infantis, Hadar, Enteritidis, Dublin and Gallinarum, by multiplex PCR. *J Microbiol Methods* 2011;85:9–15.

- [28] Lin JS, Tsen HY. Development and use of polymerase chain reaction for the specific detection of *Salmonella* Typhimurium in stool and food samples. *J Food Protect* 1999;62:1103–10.
- [29] Malorny B, Bunge C, Helmuth R. A real-time PCR for the detection of *Salmonella* Enteritidis in poultry meat and consumption eggs. *J Microbiol Methods* 2007;70:245–51.
- [30] Chiang YC, Tsen HY, Chen HY, Chang YH, Lin CK, Chen CY, Pai WY. Multiplex PCR and a chromogenic DNA macroarray for the detection of *Listeria monocytogenes*, *Staphylococcus aureus*, *Streptococcus agalactiae*, *Enterobacter sakazakii*, *Escherichia coli* O157:H7, *Vibrio parahaemolyticus*, *Salmonella* spp. and *Pseudomonas fluorescens* in milk and meat samples. *J Microbiol Methods* 2012;88:110–6.
- [31] McCabe KM, Zhang YH, Huang BL, Wagar EA, McCabe ER. Bacterial species identification after DNA amplification with a universal primer pair. *Mol Genet Metab* 1999;66:205–11.
- [32] Wang XW, Zhang L, Jin LQ, Jin M, Shen ZQ, An S, Chao FH, Li JW. Development and application of an oligonucleotide microarray for the detection of food-borne bacterial pathogens. *Appl Microbiol Biotechnol* 2007;76:225–33.
- [33] Thapa SP, Han AR, Cho JM, Hur JH. Multiplex PCR and DNA array for the detection of *Bacillus cereus*, *Staphylococcus aureus*, *Listeria monocytogenes*, *Escherichia coli* O157:H7, and *Salmonella* spp. targeting virulence-related genes. *Ann Microbiol* 2012;63:725–31.
- [34] Lin CM, Chiang YC, Tsen HY. Development and use of a chromogenic macroarray system for the detection of *Staphylococcus aureus* with enterotoxin A, B, C, D, E, and G genes in food and milk samples. *Foodborne Pathog Dis* 2009;6:445–52.