



Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.



## PULMONOLOGY

[www.journalpulmonology.org](http://www.journalpulmonology.org)


## ORIGINAL ARTICLE

## Comparative bench study evaluation of a modified snorkeling mask used during COVID-19 pandemic and standard interfaces for non-invasive ventilation

Giuliano Ferrone<sup>a,b,\*</sup>, Giorgia Spinazzola<sup>a,b</sup>, Roberta Costa<sup>b</sup>, Antonio Gulli<sup>a,b</sup>,  
Andrea Scapigliati<sup>a,b</sup>, Massimo Antonelli<sup>a,b</sup>, Giorgio Conti<sup>a,b</sup>

<sup>a</sup> Department of Emergency, Intensive Care Medicine and Anesthesia, Fondazione Policlinico Universitario A. Gemelli IRCCS, Rome, Italy

<sup>b</sup> Istituto di Anestesiologia e Rianimazione, Università Cattolica del Sacro Cuore, Rome, Italy

Received 26 January 2021; accepted 23 May 2021

Available online xxx

## KEYWORDS

Covid-19;  
Non-invasive  
ventilation;  
Mechanical  
ventilation;  
ARDS

## Abstract

**Purpose:** The aim of this bench study is to compare the standard NIV and nCPAP devices (Helmet, H; Full face mask, FFM) with a modified full face snorkeling mask used during COVID-19 pandemic.

**Methods:** A mannequin was connected to an active lung simulator. The inspiratory and expiratory variations in airways pressure observed with a high simulated effort, were determined relative to the preset CPAP level. NIV was applied in Pressure Support Mode at two simulated respiratory rates and two cycling-off flow thresholds. During the bench study, we measured the variables defining patient-ventilator interaction and performance.

**Results:** During nCPAP, the tested interfaces did not show significant differences in terms of  $\Delta P_{awi}$  and  $\Delta P_{awe}$ .

During NIV, the snorkeling mask demonstrated a better patient-ventilator interaction compared to FFM, as shown by significantly shorter Pressurization Time and Expiratory Trigger Delay ( $p < 0.01$ ), but no significant differences were found in terms of Inspiratory Trigger Delay and Time of Synchrony between the interfaces tested. At RR 20sim, the snorkeling mask presented the lower  $\Delta P_{trigger}$  ( $p < 0.01$ ), moreover during all the conditions tested the snorkeling mask showed the longer Pressure Time Product at 200, 300, and 500 ms compared to FFM ( $p < 0.01$ ). A

**Abbreviations:** ICUs, Intensive care units; nCPAP, Non-invasive continuous positive airway pressure; NIV, Non-invasive ventilation; Pmus, Simulated effort; PSV, Pressure support ventilation; WHO, World Health Organization; ARDS, Acute respiratory distress syndrome;  $PaO_2/FiO_2$ , Inspired oxygen fraction ratio; RRsim, Simulated respiratory rate; iPS, Inspiratory pressure support;  $\Delta P_{awi}$ , Maximum inspiratory deflection;  $\Delta P_{awe}$ , Expiratory peak; TI, Mechanical inspiratory time; TE, Mechanical expiratory time; VT, Tidal volume;  $VT_{neu}$ , Neural tidal volume; Ttot, Total mechanical breath duration; Paw, Airway opening pressure;  $\Delta P_{trigger}$ , Trigger pressure drop; PTPt, Inspiratory pressure-time product; PTP, Pressure time product;  $Time_{press}$ , Pressurization time;  $Delay_{tr_{insp}}$ , Inspiratory trigger delay;  $Delay_{tr_{exp}}$ , Expiratory trigger delay;  $Time_{sync}$ , Time of synchrony.

\* Corresponding author at: Department of Emergency, Intensive Care Medicine and Anesthesia, Fondazione Policlinico Universitario A. Gemelli IRCCS, Rome, Italy.

E-mail address: [giuliano.ferrone@policlinicogemelli.it](mailto:giuliano.ferrone@policlinicogemelli.it) (G. Ferrone).

<https://doi.org/10.1016/j.pulmoe.2021.05.009>

2531-0437/© 2021 Sociedade Portuguesa de Pneumologia. Published by Elsevier España, S.L.U. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Please cite this article in press as: G. Ferrone, G. Spinazzola, R. Costa et al., Comparative bench study evaluation of a modified snorkeling mask used during COVID-19 pandemic and standard interfaces for non-invasive ventilation, Pulmonology (2021), <https://doi.org/10.1016/j.pulmoe.2021.05.009>

major limitation of snorkeling mask is that during NIV with this interface it is possible to reach maximum 18 cmH<sub>2</sub>O of peak inspiratory pressure.

*Conclusions:* The modified snorkeling mask can be used as an acceptable alternative to other interfaces for both nCPAP and NIV in emergencies.

© 2021 Sociedade Portuguesa de Pneumologia. Published by Elsevier España, S.L.U. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

## Introduction

In March 2020, the World Health Organization (WHO) declared the severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) outbreak a pandemic, due to the increasing number of cases reported worldwide,<sup>1</sup> with high rates of hospitalization and ICU admission.<sup>2</sup> The high number of SARS-CoV-2 cases initially in Lombardy, and subsequently throughout the whole country, made Italy one of the most affected countries in Europe.<sup>3</sup>

As happened in March 2020, wherein Lombardy in few days a total of 1593 patients, affected by severe respiratory failure due to COVID-19, were admitted to the ICU,<sup>4</sup> out of a total of 1202 ICU beds available,<sup>2</sup> the National health care system is now again under pressure from a growing second wave of patients hospitalized or admitted to ICU for acute respiratory failure.

COVID-19 is characterized by a viral interstitial pneumonia<sup>5</sup> with fever, dry cough, dyspnoea, and bilateral ground-glass opacities,<sup>6</sup> with about 67% of patients evolving to severe pneumonia.<sup>7,8</sup>

Preliminary reports described that COVID-19 patients, compared to conventional Acute Respiratory Distress Syndrome (ARDS), are characterized by moderate to severe hypoxemia despite a relatively high pulmonary compliance.<sup>9,10</sup> Due to the enormous number of COVID-19 patients with acute respiratory failure and to the shortage of ICU beds and ventilators, in many Italian hospital, the management of patients with respiratory failure was entrusted to Non-Invasive Ventilation (NIV) or Non-invasive continuous positive airway pressure (nCPAP).

Several respiratory managements were applied to treat ARDS COVID 19 related. The High flow nasal cannula was used, also as first-line therapy, in China, and in USA (although with a high risk of air contamination). The NIV or CPAP were applied in hospitalized patients in China, Italy, and USA with the same proportion (20%, 11% and 19%).<sup>11</sup>

In particular, in a scenario of a discrepancy between facilities and a large number of casualties, as with COVID-19 pandemic, the application of NIV or nCPAP has been useful as a respiratory supportive strategy, especially in patients with mild to moderate ARDS and a PaO<sub>2</sub> to Inspired oxygen fraction ratio (PaO<sub>2</sub>/FiO<sub>2</sub>) > 150. In our country, the Helmet<sup>4,10</sup> has been the most widely used device for non-invasive respiratory support during COVID-19 outbreaks both in general wards and in ICU.<sup>10</sup>

Unfortunately, a major problem of ventilator and device for NIV shortages rapidly emerged because of the further spread of the virus in other regions of Italy. To relieve the pressure on our National Health System, a device converting a full-face snorkeling mask into a mask for CPAP or NIV has

been designed and proposed for clinical use, with the help of 3D printers.

Given the large diffusion of this modified full face snorkeling mask in COVID-19 patients, we designed this bench study to evaluate and compare a helmet, a full-face mask, and a modified full face snorkeling mask in delivering nCPAP and NIV in Pressure Support Ventilation mode (PSV).

## Methods


The study was performed at the Respiratory Mechanics Lab (Ventilab) of the Fondazione Policlinico Universitario A. Gemelli IRCCS, Università Cattolica del Sacro Cuore in Rome, Italy.

## Bench study

Non-invasive CPAP and non-invasive positive pressure ventilation delivered in PSV mode were applied to a mannequin (LaerdalMedical AS, Stavanger, Norway) connected to an active test lung system (ASL 5000; Ingmar Medical, Pittsburgh, PA) set using a single-compartment model, an active inspiration simulated by a semi-sinusoidal pressure waveform (Rise Time 15%, Pause 0% and Release Hold 25%) and the following mechanical properties of the respiratory system: resistance 5 cmH<sub>2</sub>O/l/s and compliance 40 ml/cmH<sub>2</sub>O. nCPAP was applied via Helmet (H) (CPAP-Castar Starmed, Mirandola, Italy), PerforMax Full face mask (Philips Respironics, Murrysville, PA, USA) (FFM), and a modified full face snorkeling mask (SEA VU DRY, Mares Spa, Rapallo, Italy), while non-invasive PSV was delivered through FFM and snorkeling mask. The Helmet used for this bench study is a transparent latex-free polyvinylchloride hood, joined by a rigid plastic ring to a soft collar and secured by two padded armpit braces at four hooks (two in the front and two in the back of the plastic ring). The helmet used was the size Small to attain a good seal and avoid air-leaks.

The snorkeling mask total internal volume is 1350 ml, but the mouth-and-nose pocket internal volume is only 80 ml, while the FFM internal volume is 500 ml and the H internal volume is 15,000 ml (a real pressurized gas reservoir during inspiration).

The snorkeling mask differs from Performax full face mask for shape and design characteristic; it presents a complete separation between inspiratory and expiratory circuits with the following main features: hypoallergenic silicone mouth-and-nose pocket connected to a polycarbonate transparent main body; quick-release buckles for easy doffing and a polycarbonate Charlotte valve with an inspiratory and



**Figure 1** Mares snorkeling mask features: presence of two separate limbs; the inspiratory one is highlighted by blue arrows, the expiratory one by red arrows. The adapter created with 3 d printers features two channels: the central upward-pointing channel is the inspiratory channel, the side channel is the expiratory channel.

an expiratory channel. See Fig. 1 for details. The snorkeling mask presents two parallel connections with a complete separation between inspiratory and expiratory limbs, while Performax Full face mask is characterized by a single limb connected to the Y piece. The measure of masks used was the medium size for FFM and large size for snorkeling mask to attain a good seal and avoid air-leaks. nCPAP (10 cmH<sub>2</sub>O) was applied at a simulated respiratory rate (RR<sub>sim</sub>) of 20 breaths per min (b/min) and a simulated level of inspiratory effort (P<sub>mus</sub>) of 12 cmH<sub>2</sub>O, using a standard CPAP device delivering a flow rate of 60 l/min with reservoir (Drager CF 800 Continuous Flow CPAP System; Dragerwerk AG & Co, Lubeck, Germany). During nCPAP test we applied a PEEP valve in the expiratory limb.

NIV in PSV mode was delivered at 2 RR<sub>sim</sub> (20 and 30 b/min) and a P<sub>mus</sub> of 12 cmH<sub>2</sub>O with the mechanical ventilator (Puritan Bennet 840; Covidien Health-Care, Mansfield, MA) set in inspiratory pressure support (iPS) of 10 cmH<sub>2</sub>O, Positive End-Expiratory Pressure (PEEP) of 8 cmH<sub>2</sub>O, the fastest rate of pressurization, and a cycling-off flow threshold of 25 and 50% of the peak inspiratory flow. We set the inspiratory flow trigger at the lowest value not determining auto-cycling: this threshold was 5 L/min during all conditions tested. This setting was chosen for comparing the performance of these interfaces under the condition of highest pressurization rate and fast or slow cycling-off criteria.<sup>12</sup>

## Measurements

Air flow ( $V'$ ) was measured with a pneumotachograph (Fleish No.1, Metabo, Epalinges, Switzerland), while airway pressure (Paw) was measured by a pressure transducer with a differential pressure of  $\pm 100$  cmH<sub>2</sub>O (Digima Clic-1, ICULab system; KleisTek Engineering, Bari, Italy), placed distally from the pneumotachograph. Airflow ( $V'$ ) and airway pressure (Paw) at the helmet inlet during the inspiratory phase were measured using a pneumotachograph (Fleisch n.2;

Metabo, Epalinges, Switzerland) and a pressure transducer with a differential pressure of  $\pm 100$  cmH<sub>2</sub>O (Digima Clic-1; KleisTEK, ICU-Lab System, Italy) sited at the distal end of the inspiratory limb of the circuit. When the mannequin was ventilated through the FFM, the pneumotachograph and the pressure transducer were positioned at the Y-connection of the ventilator circuit, instead, when we tested the snorkeling mask the pneumotachograph and the pressure transducer were positioned on the inspiratory channel. All these signals were acquired, amplified, filtered, digitized at 100 Hz, recorded on a dedicated personal computer, and analyzed with specific software (ICU lab 2.3; KleisTEK Advanced Electronic System, Italy and Analysis Plus).

Each trial lasted 5 min; the breaths of the last minute (20 or 30 depending on the trial) were recorded and averaged for analysis.

The measured variables assessed during nCPAP were the maximum inspiratory deflection ( $\Delta P_{aw_i}$ , inspiratory drop) and the expiratory peak ( $\Delta P_{aw_e}$ ), calculated as differences from the preset CPAP level.

During the NIV test, we evaluated the following variables: Ventilator inspiratory and expiratory time (mechanical TI and mechanical TE, respectively), and ventilator rate of cycling were all determined on the flow tracing. The inspiratory duty cycle (mechanical TI/T<sub>tot</sub>) was calculated as the ratio between mechanical TI and the total mechanical breath duration (T<sub>tot</sub>). Airflow ( $V'$ ) and tidal volume (VT) delivered to the simulator, airway opening pressure (Paw), and inspiratory muscles effort were displayed online on the computer screen. The signals obtained with the ASL were transmitted to a PC host via 10/100Mbit Ethernet, sampled, and processed in real-time by means of specific software (Lab View, Ingmar Medical). The signals obtained with the ASL were integrated with the signals from the ICULab system by using a specific application of the ICULab (ICULab 2.7, KleisTek). The numerical integration of flow over time determined the mechanical tidal volume (mechanical VT). The amount of tidal volume delivered to the simulator during its

active inspiration (i.e., the neural tidal volume,  $VT_{neu}$ ) was calculated as the volume generated from the onset of inspiratory muscle effort negative deflection to its nadir.

Interfaces performance was evaluated using the following parameters<sup>12–14</sup>:


- 1) Trigger pressure drop ( $\Delta P_{trigger}$ ), defined as the pressure swing generated by the simulator inspiratory effort in the airway during the triggering phase; 2) Inspiratory pressure–time product (PTP<sub>trigger</sub>), defined as the area under the  $P_{aw}$  curve relative to the time between the onset of inspiratory effort and the start of mechanical assistance; 3) pressure-time product at 200 ms from the onset of the ventilator pressurization (PTP200), as the index of pure pressurization performance<sup>15</sup>; 4) Pressure-time product at 300 ms (PTP300) defined as the integration of  $P_{aw}$  over time during the first 300 msec and representing the speediness of the ventilator in reaching the preset level of pressure support; 5) Pressure-time product at 500 ms (PTP500), defined as the integral  $P_{aw}$  area over insufflation time from the simulated effort onset, representing the ventilator capability of maintaining the pressurization; 6) PTP500 ideal index, expressed as a percentage of the ideal PTP, which is unattainable because it would imply a trigger pressure drop and an instantaneous pressurization of the ventilator (Fig. 2).

Patient-ventilator interaction was evaluated by determining:

- 1) Pressurization time ( $Time_{press}$ ), defined as the time necessary to achieve the pre-set level of pressure support from the baseline value; 2) Inspiratory trigger delay ( $Delay_{tr_{insp}}$ ), calculated as the time lag between the onset of inspiratory muscle effort negative swing and the start of the ventilator support (i.e.,  $P_{aw}$  positive deflection); 3) Expiratory trigger delay ( $Delay_{tr_{exp}}$ ), assessed as the delay between the end of the inspiratory effort and the end of the mechanical insufflations (i.e., flow deflection); 4) Time of synchrony ( $Time_{sync}$ ), defined as the time during which inspiratory muscle effort and  $P_{aw}$  are in phase (ideally 100%); 5) Simulator VT/mechanical VT, intended as the percentage of VT delivered during inspiratory muscle effort negative deflection; 6) The time during which simulator respiratory effort and ventilator assistance were synchronous, indexed to simulated inspiratory time ( $Time_{sync}/Ti_{neu}$ ) was also computed<sup>16–18</sup>. 7) Wasted efforts, defined as ineffective inspiratory efforts, not assisted by the ventilator; 8) Auto-triggering, namely a mechanical insufflation in absence of inspiratory effort.

## Statistical analysis

Continuous data were expressed as mean  $\pm$  standard deviation (SD). Categorical data were presented as numbers and percentages in brackets. All variables were compared with each interface used. Comparisons were made by Student's *t*-test and Chi test, as appropriate. The analysis of variance (ANOVA) for repeated measures was used to detect significant differences between the different experimental conditions. When significant differences were detected, a post-



**Figure 2** Inspiratory pressure-time product (PTP), PTP at 300 ms, and PTP at 500 ms on the pressure/time trace.


hoc analysis was performed using the Bonferroni test; *p* values  $< 0.05$  were considered statistically significant. Statistical analysis was performed using MEDcalc version 18.6.

## Results

As shown in Fig. 3, during nCPAP, no significant difference in  $\Delta P_{awi}$  and  $\Delta P_{awe}$  was observed between the three tested interfaces ( $\Delta P_{awi}$   $p = 0.67$ ;  $\Delta P_{awe}$   $p = 0.10$ ).

Patient-ventilator interaction described by the measurement of  $Delay_{tr_{insp}}$ ,  $Time_{press}$ , and  $Delay_{tr_{exp}}$  during NIV delivered through the snorkeling mask and FFM are shown in Figs. 4 and 5. At both RRsim tested and both ventilator settings, the snorkeling mask showed a significantly shorter  $Time_{press}$  and  $Delay_{tr_{exp}}$  compared to FFM ( $p < 0.01$ ). Under all conditions, no significant differences were found in terms of  $Delay_{tr_{insp}}$ ,  $Time_{sync}$  (Fig. 6), and  $Time_{sync}/Ti_{neu}$  between all the interfaces tested. Under all study conditions, the volume delivered with the snorkeling mask was higher than that delivered with the FFM ( $p < 0.01$ ) (Table 1).

The trigger and pressurization performances of the two masks during NIV are shown in Tables 2 and 3. At RRsim 20,




**Figure 3** Maximum inspiratory deflection ( $\Delta P_{awi}$ , inspiratory drop) and the expiratory peak ( $\Delta P_{awe}$ ), during CPAP with Helmet (light gray column), PerforMax Full face mask (FFM, dark gray column) and snorkeling mask (black column) at 20 breaths/min.

the snorkeling mask showed a significantly lower  $\Delta P_{trigger}$  than FFM ( $p < 0.01$ ), while no difference was found at RRsim 30. No significant differences were found in terms of PTP<sub>trigger</sub> between the interfaces during all tested conditions.

Significant differences in PTP 200, PTP 300, and PTP 500 were found between the snorkeling mask and FFM ( $p < 0.01$ ) in all conditions tested. The snorkeling mask demonstrated a significantly higher capacity to maintain the pressurization at 200, 300, and 500 ms after opening the inspiratory valve.


In all settings, no asynchrony phenomena were detected with both interfaces.

During the NIV test, with the snorkeling mask it was possible to reach a maximum of 18 cmH<sub>2</sub>O of peak inspiratory applied pressure. Over this pressure, the presence of air leaks determined important continuous asynchrony phenomena (as continuous auto cycling and double triggering), not modifiable with a change in flow or pressure trigger.



**Figure 4** Inspiratory trigger delay (Delay<sub>tr<sub>insp</sub></sub>), Expiratory Trigger delay (Dealy<sub>tr<sub>exp</sub></sub>), and Pressurization Time (Time<sub>press</sub>) with the snorkeling mask (black column) and the PerforMax Full face mask (FFM) (gray column) at 20 breaths/min, with 2 ventilator settings.






**Figure 5** Inspiratory trigger delay (Delay<sub>trinsp</sub>), Expiratory Trigger delay (Dealy<sub>trexp</sub>), and Pressurization Time (Time<sub>press</sub>) with the snorkeling mask (black column) and the PerforMax Full face mask (FFM) (gray column) at 30 breaths/min, with 2 ventilator settings.

**Discussion**

In this bench study, during nCPAP, the interfaces tested showed a similar performance. Instead, during NIV, the snorkeling mask outperformed the FFM for most of the variables considered and in most of the simulated settings.

During the recent SARS-CoV-2 pandemic, early non-invasive respiratory support allowed for the treatment of a large number of patients with respiratory distress to prevent ICU admission. For several reasons, the CPAP application by Helmet represented the first choice of treatment in patients with mild to moderate COVID-19 related ARDS. The Helmet



**Figure 6** Time of synchrony with the snorkeling mask (black column) and the PerforMax Full face mask (FFM) (gray column) at two respiratory rates (RR 20 and 30 breaths/min), with 2 ventilator settings.

**Table 1** Tidal volume delivered with the different interfaces tested.

	<i>Snorkeling mask</i>		<i>P</i>	<i>FFM</i>		<i>P</i>
	<i>RR 20</i> <i>Timepress100%/</i> <i>Trexp50%</i>	<i>RR20</i> <i>Timepress100%/</i> <i>Trexp50%</i>		<i>Snorkeling mask</i>	<i>FFM</i>	
<b>VT<sub>mech</sub> (ml)</b>	754.33±5.77	676±0.00	0.001	814.33±5.77	694±0.00	0.001
<b>VT<sub>neu</sub> (ml)</b>	606.33± 5.77	459±5.20	0.001	603±0.00	459±5.2	0.001
<b>VT<sub>neu</sub>/VT<sub>mech</sub> (%)</b>	80%	68%	0.001	74%	66%	0.001
	<i>Snorkeling mask</i>		<i>P</i>	<i>FFM</i>		<i>P</i>
	<i>RR 30</i> <i>Timepress100%/</i> <i>Trexp50%</i>	<i>RR30</i> <i>Timepress100%/</i> <i>Trexp50%</i>		<i>Snorkeling mask</i>	<i>FFM</i>	
<b>VT<sub>mech</sub> (ml)</b>	724.33±5.77	586±5.2	0.001	739.67±6.35	566.33±5.77	0.001
<b>VT<sub>neu</sub> (ml)</b>	417.33±5.77	304±9.00	0.001	378±0.00	276±0.00	0.001
<b>VT<sub>neu</sub>/VT<sub>mech</sub> (%)</b>	58%	52%	0.001	51%	49%	0.02

VT<sub>mech</sub>, mechanical tidal volume; VT<sub>neu</sub>, neural tidal volume; VT<sub>neu</sub>/VT<sub>mech</sub>, the percentage of tidal volume delivered during inspiratory simulated muscle effort negative deflection; FFM, respironics PerforMax full face mask; RR, respiratory rates.

**Table 2** Performance of the interfaces during NIV at 20 RRsim.

	<i>Snorkeling mask</i>		<i>P</i>	<i>FFM</i>		<i>P</i>
	<i>RR 20</i> <i>Timepress100%/</i> <i>Trexp50%</i>	<i>RR20</i> <i>Timepress100%/</i> <i>Trexp50%</i>		<i>Snorkeling mask</i>	<i>FFM</i>	
<b>ΔP<sub>trigger</sub> (cmH<sub>2</sub>O)</b>	1.05±0.01	0.72±0.05	0.004	1.03±0.05	0.78±0.04	0.014
<b>PTPt (cmH<sub>2</sub>O/s)</b>	0.04±0.01	0.03±0.01	1	0.05±0.00	0.03±0.00	0.001
<b>PTP200 (cmH<sub>2</sub>O/s)</b>	0.88±0.05	0.57±0.09	0.013	0.87±0.12	0.57±0.05	0.04
<b>PTP300 (cmH<sub>2</sub>O/s)</b>	2.04±0.08	1.22±0.09	0.004	1.97±0.03	1.21±0.05	0.001
<b>PTP500 (cmH<sub>2</sub>O/s)</b>	4.49±0.09	2.66±0.11	0.001	4.35±0.03	2.67±0.03	0.001
<b>PTP500 ideal index (%)</b>	72	56	0.45	71	51	0.5

ΔP<sub>trigger</sub>, trigger pressure drop; PTPt, pressure time product during the triggering phase; Paw, airway pressure; PTP200, PTP300 and PTP500, pressure time product during the initial 200, 300 and 500 ms from the onset of the ventilator pressurization expressed as the absolute value; PTP500 ideal index, pressure time product during the initial 500 ms from the onset of the simulated effort, expressed as the percentage of the area of ideal pressurization, with different ventilator settings (see text); FFM, respironics PerforMax full face mask; RR, respiratory rates.

**Table 3** Performance of the interfaces during NIV at 30 RRsim.

	<i>Snorkeling mask</i>		<i>P</i>	<i>FFM</i>		<i>P</i>
	<i>RR 30</i> <i>Timepress100%/</i> <i>Trexp50%</i>	<i>RR30</i> <i>Timepress100%/</i> <i>Trexp50%</i>		<i>Snorkeling mask</i>	<i>FFM</i>	
<b>ΔP<sub>trigger</sub> (cmH<sub>2</sub>O)</b>	1.28±0.04	1.75±0.47	0.11	1.94±0.23	2.05±0.11	0.31
<b>PTPt (cmH<sub>2</sub>O/s)</b>	0.06±0.01	0.10±0.04	0.12	0.11±0.02	0.13±0.01	0.16
<b>PTP200 (cmH<sub>2</sub>O/s)</b>	0.98±0.07	0.56±0.10	0.02	0.63±0.15	0.32±0.03	0.04
<b>PTP300 (cmH<sub>2</sub>O/s)</b>	2.26±0.07	1.29±0.10	0.002	1.91±0.08	1.07±0.10	0.006
<b>PTP500 (cmH<sub>2</sub>O/s)</b>	5.09±0.03	3.05±0.12	0.007	4.69±0.16	2.77±0.11	0.003
<b>PTP500 ideal index (%)</b>	71	56	0.22	68	48	0.3

ΔP<sub>trigger</sub>, trigger pressure drop; PTPt, pressure time product during the triggering phase; Paw, airway pressure, PTP200, PTP300 and PTP500, pressure time product during the initial 200, 300 and 500 ms from the onset of the ventilator pressurization expressed as the absolute value; PTP500 ideal index, pressure time product during the initial 500 ms from the onset of the simulated effort, expressed as the percentage of the area of ideal pressurization, with different ventilator settings (see text); FFM, respironics PerforMax full face mask; RR, respiratory rates.



was demonstrated to be an effective alternative to a face mask in recruiting alveolar units and improve hypoxemia. It can also limit air-leaks and room contamination, increasing the patient's comfort and is better tolerated than the face mask, requiring fewer discontinuations.<sup>10,19,20</sup> The choice of Helmet as privileged interface applied during treatment of ARDS COVID-19 related is affected by the high diffusibility of SARS-CoV-2 infection. This characteristic of the recent pandemic SARS-CoV-2 infection can increase the level of biological hazard to which healthcare workers are exposed thus requiring the use of personal protective equipment (PPE).<sup>21</sup>

Unfortunately, the enormous demand for helmet CPAP resulted in a rapid lack of supply, so many laboratories, to try to fill this gap, started to readjust snorkeling masks into devices for respiratory support.

We, therefore, decided to compare the modified snorkeling mask with the Helmet and one of the most popular full face masks (PerforMax full face mask), during CPAP. We applied CPAP at 10 cmH<sub>2</sub>O because this is a value most commonly applied in this clinical setting.<sup>22,23</sup>

In our nCPAP bench study evaluation, the snorkeling mask presented a similar performance in terms of  $\Delta P_{awi}$  and  $\Delta P_{awe}$  compared to the other interfaces tested.

The snorkeling mask showed stability in maintaining the PEEP level applied during nCPAP.

The snorkeling mask was initially proposed as an alternative to the helmet for CPAP, but during the emergency, it was often used also for NIV. Therefore, we decided to compare this interface with one of the most popular interfaces used in ICU for NIV (FFM). We limited the comparison only to the FFM because, as previously demonstrated by Chiumello et al.,<sup>24</sup> in pressure support mode, the mask was more efficient than the helmet. In fact, with the helmet, the initial part of the inspiratory pressure applied is dissipated to pressurize its soft wall. Accordingly, Navalesi et al.<sup>25</sup> demonstrated that the helmet significantly worsens patient-ventilator synchrony, when compared to the facemask, as indicated by longer delays between inspiratory muscle effort and support delivery, both at the onset and at the end of inspiration, and by the occurrence of wasted efforts.

Several NIV interfaces, such as Full face mask and Helmet, are characterized by high compliance due to material features that can influence patient-ventilator interaction, and interface performance.

One of the reasons for choosing Respironics mask for this bench study was that this model has relatively low compliance.

For NIV settings, the values of the simulated effort, RR<sub>sim</sub>, resistance, and compliance, were those already utilized in previous investigations.<sup>26,27</sup> The snorkeling mask demonstrated the better simulated patient-ventilator interaction compared to FFM, as shown by shorter Time<sub>press</sub> and Delay<sub>t<sub>rexp</sub></sub>. These results are further validated by interface performance data. Probably the better performance underlined by lower  $\Delta P_{trigger}$ , and higher PTP 200, 300, and 500 can be explained by the reduced snorkeling mask inner volumes and its materials (Fig. 1). A prerogative of this mask is the presence of two separate limbs. The fresh air enters through the snorkel's central channel, passing the lens, and keeping it fog-free. It then enters the oral-nasal pocket through non-return valves and flows towards the nose and

mouth. The oral-nasal pocket's structure guides exhaled CO<sub>2</sub> through 4 no-return valves:

- 2 no-return valves, one each side of the nose, to prevent backward flow,
- 2 no-return valves at the entrance of the exhalation tubes.

Exhaled CO<sub>2</sub> is channeled into the soft silicone side tubes through the second pair of no-return valves. This configuration could explain our results in terms of simulated patient-ventilator interaction and performance devices during NIV settings.

A major limitation of the snorkeling mask must be underlined: during NIV with this interface, it is possible to reach a maximum of 18 cmH<sub>2</sub>O of peak inspiratory applied pressure. Over this pressure, the presence of air leaks determines continuous asynchronies (as auto cycling and double triggering), not modifiable with a change in flow or pressure trigger. For this reason, our test was limited to single pressure support and PEEP level (PS 10 cmH<sub>2</sub>O and PEEP 8 cmH<sub>2</sub>O).

## Conclusion

The results of this comparative bench study suggest that this modified snorkeling mask can be used as an alternative to other interfaces for both CPAP and NIV in emergencies. The snorkeling mask can be proposed in the event of a new pandemic surge or for countries where the number of COVID-19 patients is such that all user interfaces for non-invasive respiratory support must be used. Particular attention should be paid to the use of this modified snorkeling mask during NIV, related to the maximum pressure limit of 18 cmH<sub>2</sub>O.

## Funding

This study was funded solely by institutional/departmental resources.

## Declaration of Competing Interests

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

## Acknowledgments

The authors are grateful to all ICU doctors, residents, and nurses who fight this COVID 19 pandemic with their efforts and passion.

## References

1. WHO Director-General's opening remarks at the media briefing on COVID-19: 11 March 2020. Published March 11, 2020. <https://www.who.int/dg/speeches/detail/who-director-general-s-opening-remarks-at-the-media-briefing-on-covid-19--11-march-2020>. Accessed March 30, 2020.

2. Grasselli G, Pesenti A, Cecconi M. Critical care utilization for the COVID-19 outbreak in Lombardy, Italy: early experience and forecast during an emergency response. *JAMA*. 2020;323(16):1545–6. <https://doi.org/10.1001/jama.2020.4031>.
3. Coronavirus COVID-19 global cases by the Center for Systems Science and Engineering (CSSE) at Johns Hopkins. Up dated March 30, 2020. Accessed March 30, 2020.
4. Grasselli G, Zangrillo A, Zanella A, Antonelli M, Cabrini L, Castelli A, et al. COVID-19 Lombardy ICU network. baseline characteristics and outcomes of 1591 patients infected with SARS-CoV-2 admitted to ICUs of the Lombardy Region, Italy. *JAMA*. 2020;323(16):1574–81. <https://doi.org/10.1001/jama.2020.5394>.
5. Shi Y, Wang Y, Shao C, Huang J, Gan J, Huang X, et al. Covid 19 infection: the perspectives on immune response. *Cell Death Differ*. 2020;27(5):1451–4. <https://doi.org/10.1038/s41418-020-0530-3>.
6. Huang C, Wang Y, Li X, Ren L, Zhao J, Hu Y, et al. Clinical features of patients affected with 2019 novel coronavirus in Wuhan, China. *Lancet*. 2020;395:497–506. [https://doi.org/10.1016/S0140-6736\(20\)30183-5](https://doi.org/10.1016/S0140-6736(20)30183-5).
7. Yang X, Yu Y, Xu J, Shu H, Xia J, Liu H, et al. Clinical course and outcomes of critically ill patients with Sars-CoV2 pneumonia in Whuan, China: a single centered, retrospective, observational study. *Lancet Respir Med*. 2020;8(5):475–81. [https://doi.org/10.1016/S2213-2600\(20\)30079-5](https://doi.org/10.1016/S2213-2600(20)30079-5).
8. Bhatraju PK, Ghassemieh BJ, Nichols M, Kim R, Jerome KR, Nalla AK, et al. Covid-19 in critically ill patients in the seattle region-case series. *N Eng J Med*. 2020;382(21):2012–22. <https://doi.org/10.1056/NEJMoa2004500>.
9. Gattinoni L, Coppola S, Cressoni M, Busana M, Rossi S, Chiumello D. Covid-19 does not lead to “typical” acute respiratory distress syndrome. *Am J Respir Crit Care Med*. 2020;201(10):1299–300. <https://doi.org/10.1164/rccm.202003-0817LE>.
10. Lucchini A, Giani M, Isgro S, Rona R. The “helmet bundle” in COVID-19 patients undergoing non invasive ventilation. *Intensive Crit Care Nurs*. 2020;58:102859. <https://doi.org/10.1016/j.iccn.2020.102859>.
11. Wincka JC, Ambrosino N. COVID-19 pandemic and non invasive respiratory management: every Goliath needs a David. An evidence based evaluation of problems. *Pulmonol*. 2020;26(4):213–20. <https://doi.org/10.1016/j.pulmoe.2020.04.013>.
12. Olivieri C, Costa R, Spinazzola G, Ferrone G, Longhini F, Cammarota G, et al. Bench comparative evaluation of a new generation and standard helmet for delivering non-invasive ventilation. *Intensive Care Med*. 2013;39:734–8. <https://doi.org/10.1007/s00134-012-2765-z>.
13. Vignaux L, Vargas F, Roeseler J, Tassaux D, Thille AW, Kossowsky MP, et al. Patient–ventilator asynchrony during non-invasive ventilation for acute respiratory failure: a multicenter study. *Intensive Care Med*. 2009;35:840.. <https://doi.org/10.1007/s00134-009-1416-5>. 21.
14. Vignaux L, Piquilloud L, Tourneux P, Jollie P, Rimensberger PC. Neonatal and adult ICU ventilators to provide ventilation in neonates, infants, and children: a bench model study. *Respir Care*. 2014;59:1463–75. <https://doi.org/10.4187/respcare.02540>. 22.
15. Olivieri C, Costa R, Conti G, Navalesi P. Bench studies evaluating devices for non-invasive ventilation: critical analysis and future perspectives. *Intensive Care Med*. 2012;38:160–7. <https://doi.org/10.1007/s00134-011-2416-9>.
16. Cammarota G, Longhini F, Perucca R, Ronco C, Colombo D, Messina A, et al. New setting of neurally adjusted ventilatory assist during noninvasive ventilation through a helmet. *Anesthesiology*. 2016;125(6):1181–9. <https://doi.org/10.1097/ALN.0000000000001354>.
17. Olivieri C, Longhini F, Cena T, Cammarota G, Vaschetto R, Messina A, et al. New versus conventional helmet for delivering noninvasive ventilation: a physiologic, crossover randomized study in critically ill patients. *Anesthesiology*. 2016;124(1):101–8. <https://doi.org/10.1097/ALN.0000000000000910>.
18. Longhini F, Pan C, Xie J, Cammarota G, Bruni A, Garofalo E, et al. New setting of neurally adjusted ventilatory assist for noninvasive ventilation by facial mask: a physiologic study. *Crit Care*. 2017;21(1):170. <https://doi.org/10.1186/s13054-017-1761-7>.
19. Antonelli M, Conti G, Pelosi P, Gregoretti C, Pennisi MA, Costa R, et al. New treatment of acute hypoxemic respiratory failure: noninvasive pressure support ventilation delivered by helmet: a pilot controlled trial. *Crit Care Med*. 2002;30(3):602–8. <https://doi.org/10.1097/00003246-200203000-00019>.
20. Antonelli M, Pennisi MA, Pelosi P, Gregoretti C, Squadrone V, Rocco M, et al. Noninvasive positive pressure ventilation using a helmet in patients with acute exacerbation of chronic obstructive pulmonary disease: a feasibility study. *Anesthesiology*. 2004;100(1):16–24. <https://doi.org/10.1097/0000542-200401000-00007>.
21. Ippolito M, Vitale F, Accurso G, Iozzo P, Gregoretti C, Giarratano A, et al. Medical masks and respirators for the protection of healthcare workers from SARS-CoV-2 and other viruses. *Pulmonology*. 2020 Jul-Aug;26(4):204–12. <https://doi.org/10.1016/j.pulmoe.2020.04.009>.
22. Ferrari G, Oliveri F, De Filippi G, Milan A, Aprà F, Boccuzzi A, et al. Non-invasive positive airway pressure and risk of myocardial infarction in acute cardiogenic pulmonary edema: continuous positive airway pressure vs. non-invasive positive pressure ventilation. *Chest*. 2007;132:1804–9. <https://doi.org/10.1378/chest.07-1058>. 19.
23. Cammarota G, Vaschetto R, Turucz E, Dellapiazza F, Colombo D, Blando C, et al. Influence of lung collapse distribution on the physiologic response to recruitment maneuvers during noninvasive continuous positive airway pressure. *Intensive Care Med*. 2011;37:1095–102. <https://doi.org/10.1007/s00134-011-2239-8>.
24. Chiumello D, Pelosi P, Carlesso E, Severgnini P, Aspesi M, Gamberoni C, et al. Noninvasive positive pressure ventilation delivered by helmet vs standard face mask. *Intensive Care Med*. 2003;29:1671–9. <https://doi.org/10.1007/s00134-003-1825-9>.
25. Navalesi P, Costa R, Ceriana P, Carlucci A, Prinianakis G, Antonelli M, et al. Noninvasive ventilation in chronic obstructive pulmonary disease patients: helmet versus facial mask. *Intensive Care Med*. 2006;33:74–81. <https://doi.org/10.1007/s00134-006-0391-3>.
26. Costa R, Navalesi P, Spinazzola G, Ferrone G, Pellegrini A, Cavaliere F, et al. Influence of ventilator settings on patient–ventilator synchrony during pressure support ventilation with different interfaces. *Intensive Care Med*. 2010;36:1363–70. <https://doi.org/10.1007/s00134-010-1915-4>.
27. Grasselli G, Tonetti T, Protti A, Longer T, Girardis M, Bellani G, et al. Pathophysiology of COVID-19-associated acute respiratory distress syndrome: a multicentre prospective observational study. *Lancet Respir Med*. 2020. [https://doi.org/10.1016/S2213-2600\(20\)30370-2](https://doi.org/10.1016/S2213-2600(20)30370-2). Aug 27; S2213-2600(20)30370-2.