

Reply to Goyal et al

TO THE EDITOR—Goyal et al [1] provide additional useful commentary on the article by Patel et al, which investigates the usefulness of real-time fluorescence loop-mediated isothermal amplification (LAMP) for point-of-care testing for malaria [2, 3]. We agree on the issues raised regarding laboratory infrastructure requirements and precautions necessary to prevent contamination. However, specific needs for manual or automated DNA extraction and amplification will vary according to the volume of samples being processed, costs, and the technical capacity of the laboratory. Also, while currently available LAMP technologies require conditions that preclude its use in the most remote settings, the ease of use and speed of LAMP have allowed point-of-care testing in settings where it was previously not available.

Note

Potential conflicts of interest. M. S. H. and B. G. are evaluating LoopAmpTM Malaria Pan/Pf detection kit, a LAMP system developed by Foundation for Innovative New Diagnostics. P. J. R. has no conflict of interest to declare.

All authors have submitted the ICMJE Form for Disclosure of Potential Conflicts of Interest. Conflicts that the editors consider relevant to the content of the manuscript have been disclosed.

Michelle S. Hsiang,^{1,2,3} Bryan Greenhouse,⁴ and Philip J. Rosenthal⁴

¹Department of Pediatrics, University of Texas Southwestern Medical Center, Dallas;

²Malaria Elimination Initiative, Global Health Group,

³Department of Pediatrics, UCSF Benioff

Children's Hospital, and ⁴Department of Medicine,

San Francisco General Hospital,

University of California, San Francisco

References

1. Goyal K, Kaur H, Sehgal A, Sehgal R. Real-Amp loop-mediated isothermal amplification as a point-of-care test for diagnosis of malaria:

neither too close nor too far. *J Infect Dis* 2015; 211:1686.

2. Patel JC, Lucchi NW, Srivastava P, et al. Field evaluation of a real-time fluorescence loop-mediated isothermal amplification assay, Real-Amp, for the diagnosis of malaria in Thailand and India. *J Infect Dis* 2014; 210:1180–7.

3. Hsiang MS, Greenhouse B, Rosenthal PJ. Point of care testing for malaria using LAMP, loop mediated isothermal amplification. *J Infect Dis* 2014; 210:1167–9.

Received 8 December 2014; accepted 9 December 2014; electronically published 15 December 2014.

Correspondence: Michelle S. Hsiang, MD, Department of Pediatrics, University of Texas Southwestern Medical Center, 5323 Harry Hines Blvd, Dallas, TX 75214 (michelle.hsiang@utsouthwestern.edu).

The Journal of Infectious Diseases® 2015;211:1687

© The Author 2014. Published by Oxford University Press on behalf of the Infectious Diseases Society of America. This is an Open Access article distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs licence (<http://creativecommons.org/licenses/by-nc-nd/4.0/>), which permits non-commercial reproduction and distribution of the work, in any medium, provided the original work is not altered or transformed in any way, and that the work is properly cited. For commercial re-use, please contact journals.permissions@oup.com.

DOI: 10.1093/infdis/jiu811