

## Lighting the way forward: using resonance Raman spectroscopy to non-invasively assess tissue perfusion

Chavi Rehani, Rosemary Kozar 

**To cite:** Rehani C, Kozar R. Lighting the way forward: using resonance Raman spectroscopy to non-invasively assess tissue perfusion. *Trauma Surg Acute Care Open* 2024;**9**:e001600. doi:10.1136/tsaco-2024-001600  
Accepted 1 September 2024

Poor tissue perfusion is an important driver of shock in patients after surgery, trauma, and sepsis.<sup>1</sup> Assessment of tissue oxygenation and volume status in these patients involves assimilating multiple endpoints of resuscitation.<sup>2</sup> This commonly includes measuring surrogate markers such as serum lactate and central venous oxygenation (ScVO<sub>2</sub>) through blood sampling. These data points are helpful in guiding treatment, but they can only offer intermittent insights when limited by the frequency of blood draws. In the case of ScVO<sub>2</sub>, they can also require invasive procedures such as central line placement which carries risks of infection, pneumothorax, and thrombosis.

To overcome these challenges, Garcia Mancebo *et al* fabricated a novel device that uses resonance Raman spectroscopy (RRS) to quantify oxyhemoglobin saturation (ShbO<sub>2</sub>) using a small probe placed on the surface of various tissues. Using two different rodent models of hemorrhagic shock, the authors demonstrated that RRS-measured ShbO<sub>2</sub> in the esophagus and tongue strongly correlated with both ScVO<sub>2</sub> and serum lactate. Depending on the site, ShbO<sub>2</sub> also detected near-term death with 72–83% sensitivity.<sup>3</sup>

It is interesting to note that over three decades ago, Scalea *et al* proposed the use of central venous blood oxygen saturation to detect acute blood loss in animals and in trauma patients.<sup>4 5</sup> Although the idea of venous oxygenation monitoring may not be novel, the ability to non-invasively and continuously detect blood loss and hypoperfusion in real time may prove to be. The authors acknowledge limitations of their device, including the small size of the laser which only provides data over a small area of tissue. It can, therefore, be prone to sampling error and not reflect whole organ or systemic perfusion. Although this technology may prove useful in the battlefield or mass casualty setting to detect sequelae of massive hemorrhage

and guide resuscitation, future studies should examine if RRS technology can detect subclinical hemorrhage which would have further application in both civilian and military trauma settings.

**Contributors** CR and RK each contributed to the writing and editing. RK is the guarantor.

**Funding** The authors have not declared a specific grant for this research from any funding agency in the public, commercial or not-for-profit sectors.

**Competing interests** None declared.

**Patient consent for publication** Not applicable.

**Ethics approval** Not applicable.

**Provenance and peer review** Commissioned; internally peer reviewed.

**Open access** This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

**ORCID iD**

Rosemary Kozar <http://orcid.org/0000-0002-9582-8622>

**REFERENCES**

- 1 Vincent JL, De Backer D. Circulatory Shock. *N Engl J Med* 2013;**369**:1726–34.
- 2 Martin ND, Codner P, Greene W, Brasel K, Michetti C, Committee ACC. Contemporary hemodynamic monitoring, fluid responsiveness, volume optimization, and endpoints of resuscitation: an AAST critical care committee clinical consensus. *Trauma Surg Acute Care Open* 2020;**5**:e000411.
- 3 Garcia Mancebo J, Sack K, Romfh P, Peng Y, Kheir J. Esophageal oxyhemoglobin saturation as a resuscitative metric in hemorrhagic shock. *Trauma Surg Acute Care Open* 2024;**9**:e001480.
- 4 Scalea TM, Holman M, Fuortes M, Baron BJ, Phillips TF, Goldstein AS, Sclafani SJ, Shaftan GW. Central venous blood oxygen saturation: an early, accurate measurement of volume during hemorrhage. *J Trauma* 1988;**28**:725–32.
- 5 Scalea TM, Hartnett RW, Duncan AO, Atweh NA, Phillips TF, Sclafani SJ, Fuortes M, Shaftan GW. Central venous oxygen saturation: a useful clinical tool in trauma patients. *J Trauma* 1990;**30**:1539–43.



► <http://dx.doi.org/10.1136/tsaco-2024-001480>

© Author(s) (or their employer(s)) 2024. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

R Adams Cowley Shock Trauma Center, Baltimore, Maryland, USA

**Correspondence to**

Dr Rosemary Kozar; [rkozar@som.umaryland.edu](mailto:rkozar@som.umaryland.edu)