

CORRECTION

Open Access



Correction: Accuracy of guide wire placement for femoral neck stabilization using 3D printed drill guides

Gregory R. Roytman^{1,2,3,4*}, Alim F. Ramji¹, Brian Beitler¹, Brad Yoo¹, Michael P. Leslie¹, Michael Baumgaertner¹, Steven Tommasini^{1,4} and Daniel H. Wiznia^{1,5}

Correction: 3D Printing in Medicine 8, 19 (2022)
<https://doi.org/10.1186/s41205-022-00146-8>

Following publication of the original article [1], the authors reported that there was an issue with figures overlapping in Table 1.

The correct Table 1 has been provided in this Correction.

The original article [1] has been corrected.

Author details

¹Orthopaedics and Rehabilitation, Yale School of Medicine, Yale University, New Haven, CT, USA. ²Yale Center for Medical Informatics, Yale School of Medicine, Yale University, New Haven, CT, USA. ³VA Connecticut Healthcare System, Veterans Health Administration, West Haven, CT, USA. ⁴Biomedical Engineering, Yale School of Engineering and Applied Science, Yale University, New Haven, CT, USA. ⁵Mechanical Engineering & Materials Science, Yale School of Engineering and Applied Science, Yale University, New Haven, CT, USA.

Published online: 09 August 2022

Reference

1. Roytman GR, Ramji AF, Beitler B, et al. Accuracy of guide wire placement for femoral neck stabilization using 3D printed drill guides. *3D Print Med.* 2022;8:19. <https://doi.org/10.1186/s41205-022-00146-8>.

The original article can be found online at <https://doi.org/10.1186/s41205-022-00146-8>.

*Correspondence: gregory.roytman@yale.edu





⁴Biomedical Engineering, Yale School of Engineering and Applied Science, Yale University, New Haven, CT, USA

Full list of author information is available at the end of the article



© The Author(s) 2022. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Table 1 Iterations of 3D printed drill guide being used to drill osteoporotic sawbone femurs

Iteration	Description	Figure
1	All models were created with a cylindrical shell, just covering the lower half of the greater trochanter. Additional coverage of the greater trochanter was needed. Tubes leading from guide wire entry point to sawbone cortex are external.	
2	Tubes were lengthened, however proved too flexible and more prone to breakage.	
3	Internal channels were created to negate the effects of the excessively flexible tubes. However, the channel was too narrow to provide a consistent trajectory (too much toggle).	
4	A guide was made using Formlabs Rigid Resin. However, the highly stiff material did not allow for flexible conformation to the sawbones.	
5	The Percutaneous Screw Guide had a slightly lowered turret to minimize cortical breach in the femoral neck superiorly. Channel length of 2 cm was standardized across all models. The guides are shown left to right: Percutaneous Screw Guide, FNS Guide, DHS Guide.	