

τ FCS: Multi-Method Global Analysis Enhances Resolution and Sensitivity in Fluorescence Fluctuation Measurements

Neil R. Anthony, Keith M. Berland*

Department of Physics, Emory University, Atlanta, Georgia, United States of America

Abstract

Fluorescence fluctuation methods have become invaluable research tools for characterizing the molecular-level physical and chemical properties of complex systems, such as molecular concentrations, dynamics, and the stoichiometry of molecular interactions. However, information recovery *via* curve fitting analysis of fluctuation data is complicated by limited resolution and challenges associated with identifying accurate fit models. We introduce a new approach to fluorescence fluctuation spectroscopy that couples multi-modal fluorescence measurements with multi-modal global curve fitting analysis. This approach yields dramatically enhanced resolution and fitting model discrimination capabilities in fluctuation measurements. The resolution enhancement allows the concentration of a secondary species to be accurately measured even when it constitutes only a few percent of the molecules within a sample mixture, an important new capability that will allow accurate measurements of molecular concentrations and interaction stoichiometry of minor sample species that can be functionally important but difficult to measure experimentally. We demonstrate this capability using τ FCS, a new fluctuation method which uses simultaneous global analysis of fluorescence correlation spectroscopy and fluorescence lifetime data, and show that τ FCS can accurately recover the concentrations, diffusion coefficients, lifetimes, and molecular brightness values for a two component mixture over a wide range of relative concentrations.

Citation: Anthony NR, Berland KM (2014) τ FCS: Multi-Method Global Analysis Enhances Resolution and Sensitivity in Fluorescence Fluctuation Measurements. PLoS ONE 9(2): e90456. doi:10.1371/journal.pone.0090456

Editor: James P Brody, University of California, Irvine, United States of America

Received: October 4, 2013; **Accepted:** February 3, 2014; **Published:** February 28, 2014

Copyright: © 2014 Anthony, Berland. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Funding: This work was supported by National Science Foundation grants MCB0817966 and DMR0907435. The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Competing Interests: The authors have declared that no competing interests exist.

* E-mail: kberlan@emory.edu

Introduction

Modern scientific studies increasingly demand accurate characterization of the spatial and temporal dynamics of specifically identifiable molecules [1–3]. Fluorescence fluctuation spectroscopy (FFS) methods have thus become important research tools as they enable detailed investigations of the chemical and physical properties of molecules or molecular systems in a variety of complex environments [4–9]. When FFS data is analyzed successfully, impressive resolution of sample composition and dynamics is often achievable. This includes the unique capabilities to measure dynamics over a wide range of time-scales, to accurately measure molecular concentrations, and to directly measure the stoichiometric composition of interacting molecular species. On the other hand, there are a number of fundamental challenges that can limit the overall capabilities of FFS methods, notably, limited resolution, parameter stability during curve fitting, and problems with fitting model verification. For example, the second component of a two component sample can be challenging or impossible to resolve if its concentration or molecular brightness is significantly lower than the concentration or brightness of the primary species. Also, in typical FCS measurements, it is generally not possible to resolve two separate sample components by diffusion analysis unless their diffusion coefficients differ by a factor of approximately two [10]. In addition, FCS measurements offer limited capability to discriminate between fitting models

when knowledge of the sample composition or physical dynamics driving fluctuations is not available *a priori*, as is often the case for measurements within living cells or other complex systems. These types of limitations can leave detection of a large number of potentially important molecular phenomena and interactions outside current experimental capabilities. A variety of strategies have been implemented to overcome some of these experimental limitations, including multi-color FCS measurements and the development of numerous molecular brightness based statistical analysis approaches [5,11–21]. In addition to specific advances in fluctuation methods, other approaches have also proved useful for enhancing experimental resolution in fluorescence measurements. Simultaneous analysis of the multiple spectral signatures acquired using multi-parameter fluorescence detection (MFD) [22–25] can greatly reduce false assumptions that commonly occur at the resolution limits of the single techniques alone [26]. Global analysis [27–33], i.e. curve fitting using global parameters 'linked' across multiple data sets, greatly constrains the fitting parameter space that can fit all experimental data simultaneously, enhancing resolution, and also improving model discrimination capabilities in curve fitting routines.

Here we introduce a new conceptual approach to fluorescence fluctuation microscopy that leverages the strengths of fluctuation measurements, MFD, and global analysis, and significantly enhances the broad capabilities of fluorescence fluctuation

measurements. Specifically, we apply simultaneous acquisition of fluorescence correlation spectroscopy (FCS) and fluorescence lifetime data and use multi-modal global analysis to analyze both data types simultaneously with common linked fundamental parameters, an approach we refer to as τ FCS. We demonstrate that τ FCS is remarkably successful in fully resolving the physical properties of a two-component mixture for cases where previously reported FFS methods would be unable to accurately determine the sample composition and dynamics. We also discuss how this new analysis approach improves model discrimination capabilities and improves parameter stability in curve fitting procedures. We note that FCS and lifetime measurements have previously been implemented together [34,35], although the τ FCS approach is fundamentally different from the lifetime-filter based FLCS method and has significantly less demanding requirements for the measurement signal statistics, an important practical advantage. Moreover, the fundamental strategy of combining multi-parameter fluorescence acquisition with global analysis is easily extended to other fluorescence measurement modes (e.g. anisotropy, FRET) with minimal theoretical modifications, and should thus be widely applicable in most experimental systems where FFS measurements are used.

Theory

Simultaneous global fitting of both lifetime and FCS data requires that the theory for each be written in terms of common linkable parameters. Traditionally, lifetime theory is written in terms of the excited state lifetime values and fractional intensities of each molecular species, with no direct link to the concentration or molecular brightness values that are used in FFS theory. We thus introduce a complete theory, derived from basic fluorescence principles, describing fluorescence lifetime decay histograms in terms of the molecular concentrations and molecular brightness parameters used in FFS. Fluorescence lifetime measurements require pulsed laser excitation sources, and thus τ FCS theory describes pulsed laser excitation, spans picosecond to second time scales, and incorporates effects due to finite fluorescence lifetimes [36–40]. To completely describe fluorescence dynamics under pulsed excitation in terms of common fluorescence parameters we must consider both micro- and macro-time scales. Micro-time (ps to ns) is used to describe the excited state dynamics of fluorescent molecules following pulsed excitation, and resets to zero after each incident laser pulse. Macro-time (μ s to s) describes longer time scale behavior, such as fluorescence fluctuations due to physical or chemical dynamics, and is continuous from the start of the experiment. The theory presented below is for two-photon laser excitation [37,38,40–44], as used in the reported experiments, although the details could be easily applied for single photon excitation with only minor modifications.

Two-photon fluorescence measurements

Considering the micro-time photodynamics of a single molecular species treated as a simple two state model, we previously derived an expression describing the steady state probability, $N_{1a}(r)$, that a molecule will occupy the excited state immediately after a laser pulse passes through the sample [38,41,44], with:

$$N_{1a}(\mathbf{r}) = \frac{(1 - e^{-2W(\mathbf{r})\alpha})}{2(1 - e^{-2W(\mathbf{r})\alpha} e^{-\Gamma/f_p})} \quad (1)$$

Here, $W(\mathbf{r})$ represents two-photon molecular excitation rate [44] for pulsed laser excitation that includes saturation and finite lifetime effects, with a spatial dependence determined by the spatial profile of the focused laser excitation source. The laser pulse repetition rate is f_p , and for simplicity the temporal pulse profile is assumed to have constant amplitude with pulse width α , an assumption that has no important consequences for this work. The parameter Γ is the spontaneous relaxation rate of excited state molecules. For a molecular concentration $C(\mathbf{r},t)$, we expect fluorescence emission of $\Gamma e^{-\Gamma t} C(\mathbf{r},t) N_{1a}(\mathbf{r})$ photons per unit volume per unit time following each laser pulse. The measured instantaneous fluorescence signal, $F_{inst}(t',t)$, is the integrated signal from the full sample volume given by:

$$F_{inst}(t',t) = \kappa \Gamma e^{-\Gamma t'} \int C(\mathbf{r},t) N_{1a}(\mathbf{r}) d\mathbf{r} \quad (2)$$

where κ is a sample and equipment dependent constant that incorporates all parameters describing the excitation and optical collection efficiencies, such as absorption cross sections, quantum yields, and detector efficiencies.

Equation (2) contains two distinct time variables that characterize the micro- (t') and macro-time (t) variations in the fluorescence signal. To measure fluorescence lifetimes, we are interested in the micro-time behavior, and in principle, Eq. (2) represents the fluorescence decay curve following each laser pulse. However, in practice, limited photon numbers require that fluorescence lifetime measurements are performed over millions of excitation pulses, with macro-time averaged histograms accumulated to give the micro-time resolved average fluorescence signal:

$$\langle F_{inst}(t') \rangle = \frac{1}{T} \int_0^T F_{inst}(t',t) dt = \kappa \Gamma \langle C \rangle e^{-\Gamma t'} \int N_{1a}(\mathbf{r}) d\mathbf{r} \quad (3)$$

where the total acquisition time, T , is sufficiently long that the spatial and temporal concentration fluctuations are averaged out.

Alternatively, for FFS measurements we are primarily interested in the macro-time variations in fluorescence and can instead average out the micro-time behavior. The macro-time averaged fluorescence signal thus depends explicitly on the time dependent concentration and can be written as:

$$F(t) = f_p \int_0^{1/f_p} F_{inst}(t',t) dt' = \kappa f_p \left(1 - e^{-\Gamma/f_p}\right) \int dr C(r,t) N_{1a}(r) \quad (4)$$

While $F(t)$ averages out the micro-time dynamics we omit the angular brackets indicating temporal averaging to avoid notational confusion when presenting FCS theory below. One can of course average over both micro- and macro-times to find the average fluorescence intensity:

$$\langle F \rangle = \kappa f_p \left(1 - e^{-\Gamma/f_p}\right) \langle C \rangle \int N_{1a}(\mathbf{r}) d\mathbf{r} \quad (5)$$

For purposes of notation consistency with FCS measurements, it is convenient to rewrite the fluorescence signal in terms of an “effective” molecular excitation rate, $\langle W_{eff}(\mathbf{r}) \rangle$, that describes the time-average number of excitation events per second at position \mathbf{r} .

The effective excitation rate has a simple relationship with the average fluorescence intensity given by $\langle F \rangle = \kappa \langle C \rangle \int \langle W_{eff}(\mathbf{r}) \rangle d\mathbf{r}$ [40,41]. Comparison with Eq. (5) indicates that $\langle W_{eff}(\mathbf{r}) \rangle = N_{1a}(\mathbf{r}) f_p (1 - e^{-\Gamma/f_p})$. Using this notation, we can rewrite Eqs. (3) and (4) above as:

$$\langle F_{inst}(t') \rangle = \frac{\kappa \Gamma e^{-\Gamma t'}}{f_p (1 - e^{-\Gamma/f_p})} \langle C \rangle \int \langle W_{eff}(\mathbf{r}) \rangle d\mathbf{r} \quad (6)$$

and

$$F(t) = \kappa \int dr C(r,t) \langle W_{eff}(r) \rangle \quad (7)$$

Writing the fluorescence signals in terms of $\langle W_{eff}(\mathbf{r}) \rangle$ allows for direct comparison with standard FFS notation. In particular, FFS measurements often refer to an ‘‘observation volume’’ [40,45] which is defined as:

$$V = \frac{1}{\langle W_{eff}(0) \rangle} \int \langle W_{eff}(r) \rangle dr \equiv \int \langle \hat{W}_{eff}(r) \rangle \quad (8)$$

where $\langle \hat{W}_{eff}(r) \rangle$ is the normalized fluorescence excitation probability that defines the profile of the observation volume, again including both saturation and finite lifetime effects. Higher order moments of $\langle \hat{W}_{eff}(r) \rangle$ are generally also needed to fully describe FFS theory. For FCS the additional required parameter is called the gamma factor [45–47] given by:

$$\gamma = \frac{\int \langle \hat{W}_{eff}(\mathbf{r}) \rangle^2 d\mathbf{r}}{\int \langle \hat{W}_{eff}(\mathbf{r}) \rangle d\mathbf{r}} \quad (9)$$

This parameter characterizes the uniformity of the fluorescence signal from molecules located at various locations within the volume and the effective steepness of the boundary defining the volume. We note that some authors prefer to incorporate the gamma factor into their definition of the volume, defining an effective detection volume as $V_{eff} = V/\gamma$ [7,47,48].

Using this volume notation, we can write the average fluorescence signal as:

$$\langle F \rangle = \kappa \langle C \rangle \langle W_{eff}(0) \rangle V \equiv \psi \langle C \rangle V \quad (10)$$

where we have introduced the ‘‘molecular brightness’’ parameter $\psi = \kappa \langle W_{eff}(0) \rangle$, so named because it reports the average number of fluorescence photons measured per molecule per second [12,13,49]. We note that molecular brightness is not a fundamental quantity, but depends both molecular properties (cross section and quantum yield), excitation conditions (laser power, pulse width, and beam waist), and the measurement instrumentation (detector and collection optics efficiencies).

Returning to Eqs. (6) and (7) we can write the full time dependence of the fluorescence signals as:

$$\langle F_{inst}(t') \rangle = \frac{\psi \langle C \rangle V}{(1 - e^{-\Gamma/f_p}) f_p} \Gamma e^{-\Gamma t'} \quad (11)$$

and

$$F(t) = \psi \int dr C(r,t) \langle \hat{W}_{eff}(r) \rangle \quad (12)$$

Fluorescence Lifetime and FCS Data

Using the above theory it is straight forward to express the data from both lifetime and FCS measurements in terms of common global parameters. For lifetime data, which is composed of histograms from a total of $f_p T$ laser pulse cycles, and binned into time channels of width $\Delta t'$, the recorded signal can be written as:

$$F_{hist}^{data}(t') = \sum_i \frac{T \Delta t' \psi_i \langle C_i \rangle V}{\tau_i (1 - e^{-1/\tau_i f_p})} e^{-t'/\tau_i} \quad (13)$$

where the subscript i represents each independent fluorescence species within the sample, and emission rates are replaced by fluorescence lifetimes, $\tau_i = 1/\Gamma_i$. Similarly, using well developed FCS theory [4,40,45,46,50] the correlation function for a multi-component sample is described as:

$$G(\tau_{fcs}) = \frac{\langle \delta F(t) \delta F(t + \tau_{fcs}) \rangle}{\langle F \rangle^2} = \frac{\gamma \sum_i \psi_i^2 \langle C_i \rangle A_i(\tau_{fcs})}{\left(\sum_i \psi_i \langle C_i \rangle \right)^2} \quad (14)$$

where $A_i(\tau_{fcs})$ represents the temporal relaxation of the correlation function for species i . For example, for pure diffusion with diffusion coefficient D in a three dimensional Gaussian volume with radial beam waist w_0 , and an axial beam waist aw_0 :

$$A(\tau_{fcs}) = \left(1 - \frac{8D\tau_{fcs}}{w_0^2} \right)^{-1} \left(1 - \frac{8D\tau_{fcs}}{a^2 w_0^2} \right)^{-1/2} \quad (15)$$

The factor, a , in the axial beam waist is sometimes referred to as a structure parameter in FCS literature. Equations(13), (14) & (15) clearly show how simultaneously acquired lifetime and FCS data sets depend on common global parameters, here concentration and molecular brightness, in addition to unique measurement specific parameters such as their diffusion coefficient and fluorescence lifetime. In fact, the theory is also valid for sequentially acquired lifetime and FCS data provided the sample has not changed between measurement modes, although in practice simultaneously acquisition is the best way to ensure that condition. Subsequent data analysis using these linked global parameters facilitates enhanced model discrimination and constrains fitting parameter space to greatly enhance overall experimental resolution, as demonstrated below. We also note the different functional dependencies on these global parameters, whereby the amplitude of the lifetime data scales with $\psi \langle C \rangle$ while the amplitude of the FCS data scales with $\psi^2 \langle C \rangle$. These differences in parameter dependence provide significant constraints for model discrimination in fitting routines.

The theory introduced above accurately describes fluorescence signals originating entirely from the molecular species of interest. In many cases there may be an additional background signal arising from room light, scattered laser light, or background fluorescence from sample contaminants or autofluorescence. If the

background signal is significant compared to the signal of interest, then the background must be accounted for to accurately apply τ FCS theory. The correction depends on the nature of the background. If the background is not constant, with measurable lifetime or non-trivial fluctuation dynamics, then it must be treated as an additional molecular species according to the theory introduced above. For room light or other time-constant background signals the correction for lifetime data is performed simply by subtracting the background from the total signal, while the correction for FCS requires a corrected correlation function amplitude. For an independently measured background signal, $\langle B \rangle$, the corrected correlation function amplitude is given by [45,51]:

$$G(\tau_{fcs})_{corrected} = G(\tau_{fcs})_{measured} \cdot \left(\frac{\langle F \rangle + \langle B \rangle}{\langle F \rangle} \right)^2 \quad (16)$$

Here $\langle F \rangle$ represents the fluorescence signal of interest described in Eq. (10) above, i.e. the measured background signal would need to be subtracted from the measured average signal amplitude, $\langle F \rangle + \langle B \rangle$, to determine the value of the average fluorescence.

Methods

Microscope Setup

Measurements were performed using a home-built two-photon laser-scanning setup built around an inverted microscope (IX71, Olympus) [37,52]. Pulsed excitation, provided by a Titanium:Sapphire Tsunami laser (~100 fs pulses at 80 MHz; Spectra Physics) tuned to 800 nm, was introduced to the optical path *via* a dichroic mirror (675DCSX, Chroma Technology) and focused into the sample using an Olympus 60 \times water objective (UPLSAPO60XW, Olympus; NA = 1.2). Power was controlled using a $\lambda/2$ plate and polarizing cube, and set to 5 mW at the sample. Non-descanned fluorescence signals were measured after a low pass filter (E700SP, Chroma Technology) using a hybrid photodetector (Becker and Hickl HPM 100-40; Boston Electronics). 200 s data points were collected using a time correlated single photon counting (TCSPC) module (PicoHarp300, PicoQuant GmbH). Laser pulse synchronization was provided by a battery powered fast photodiode (New Focus, Model 1621). The TCSPC histograms and autocorrelation curves were calculated using SymPhoTime software and exported for analysis in Igor Pro (Wavemetrics, Inc., OR). TCSPC histograms were binned at 4 ps, and selected to contain approximately 5×10^7 counts. Autocorrelation curves were calculated from 0.0001 to 100 ms with 16 points per time coarsening (Nsub) using the entire 200 s data set.

Binary Dye Titrations

Experiments were performed with mixtures of two Rhodamine dyes, selected because their diffusion coefficients are too similar to be resolved *via* standard FCS analysis. Rhodamine 6G (R6G; Sigma Cat# 252433) and Rhodamine B (RhB; Sigma Cat# R6626) dye solutions were prepared in HBS-EP buffer (10 mM Hepes, 150 mM NaCl, 3 mM EDTA, 0.005% Polysorbate 20; TEKnova Cat# H8020), and samples were loaded into 8-well chamber boxes (Lab-Tek II; No. 1.5 coverglass, Nunc; Thermo Fisher). Concentration titrations were performed to obtain a calibrated series of dye mixtures, starting with an 800 μ l mixture with R6G and RhB at 50 nM each, a 1:1 concentration ratio. R6G/RhB concentrations ratios >1 were achieved by sequentially

replacing 400 μ l of sample volume with an equal volume of 100 nM R6G. Similarly, concentration ratios <1 were prepared by replacing half the sample volume with 100 nM RhB. This procedure provided sample mixtures with concentrations ranging from 4 to 97 nM for R6G and 4 to 89 nM for RhB, leading to a final R6G/RhB concentration ratio ranging from approximately 0.04 to 20.

Data were acquired at each titration step, and the full titration was repeated three times. Error bars reflect the average and standard deviation of these three repeated experiments. To ensure prepared sample concentrations matched their expected values, control measurements were performed in the same sample chamber with solutions of each dye independently (either R6G or RhB), using FCS measurements to determine the actual sample concentration. The same titration procedure described above was repeated, except that HBS-EP buffer or solutions of the same dye being measured were used to replace 400 μ l of solution to respectively decrease or increase concentrations. Dye solutions in HBS-EP exhibit stable and reproducible concentration values, provided the chamber boxes were thoroughly cleaned prior to each use. To clean, a single chamber box well was washed with detergent, followed by repeated rinses with hot water, and finally scrubbed with a cotton swab using ultra-pure water. The cleaning procedure was repeated before each titration. Measurements of ultra-pure water in the cleaned chamber never showed signal above the background level of ~ 1 kHz, measured independently with the laser out of mode-lock.

Data Analysis

TCSPC histograms and FCS autocorrelation curves (ACFs) were analyzed using custom and native global analysis routines in Igor Pro. For τ FCS analysis, a TCSPC histogram and simultaneously acquired ACF are considered together as one independent τ FCS data set (Fig. S1 a). All fits in Igor Pro were performed using a Levenberg-Marquardt non-linear least squares algorithm which minimizes the χ^2 value, for a τ FCS data set defined as:

$$\chi^2 = \left(\sum_{i=0}^{k-1} \frac{(F_{hist}^{data}(x_i) - y_i)^2}{\sigma_{decay}(x_i)} + \sum_{i=k}^{k+l-1} \frac{(G(x_i) - y_i)^2}{\sigma_{fcs}(x_i)} \right) \quad (17)$$

where k and l are the number of data points in the lifetime histogram and the calculated autocorrelation function respectively. Observation volume parameters, w_0 and a , were calibrated using R6G in H_2O solution assuming a diffusion coefficient of $426 \mu m^2 s^{-1}$ [53].

We present two different approaches for data analysis. The first is applicable when a τ FCS data set is available for only a single sample condition, e.g. a single concentration ratio, which is a common experimental scenario. In this case, the analysis is “global” in that the TCSPC and FCS data have common global fitting parameters (molecular concentration and molecular brightness) that are linked across the data types (each has a single value for both types of data), and other local parameters including the lifetimes and diffusion coefficients. A second analysis approach is used when multiple τ FCS data sets are available with one or more experimental variables, such as sample concentration, varied across the data sets (Fig. S1b). For this case, the global fitting routine can be applied to all of the τ FCS data sets simultaneously, with a single set of global parameters that are linked across all of the τ FCS data sets and separate local parameter sets for each τ FCS data set. Data fits to experimental decay curves included a temporal offset such that the peak of the instrument response function (IRF) corresponds to $t' = 0$, but were not reconvolved

with the IRF during fitting. The IRF was recorded using hyper-Rayleigh scattering [54] from colloidal gold solutions (Sigma Cat# G1652) at low excitation powers.

TCSPC histogram data sets were found to contain significant systematic non-Poissonian noise at GHz frequencies (Fig. S2; red line), originating from differential non-linearities within the TCSPC module and other electronic noise. This systematic noise was found to vary linearly with total photon count per time bin in a highly reproducible manner such that this systematic noise could be removed from the signal. A data point of uncorrelated room light containing a comparable or greater number of photon counts as experimental acquisitions was recorded to determine the correction. A 100 point binomial smooth operation was performed to remove Poissonian noise while retaining the lower frequency systematic noise (Fig S2; black line) [55]. The smoothed reference noise data set was then scaled by the photon counts, yielding a multiplicative factor that effectively removes the non-Poissonian systematic noise from the signal while retaining the inherent Poissonian noise [56]. This procedure was followed for all experimental TCSPC data sets.

Simulated Data

Simulated data sets were created using Eqs. (2) & (3) for a given parameter set followed by the addition of random noise. TCSPC histograms were created with a total of 5×10^5 counts followed by the addition of Poissonian noise for a given number of counts per bin. Noise was also added to FCS curves, with noise levels determined using noise levels from experimental FCS curves acquired under comparable conditions [57]. The ACF acquisition time, T , for all calculated data sets was 30 seconds. Each χ^2 surface point is the average of three repetitions with unique simulated noise. All parameters were assumed to be unknown and were not fixed during the analyses, however, we did implement the $\langle F \rangle$ constraint (Eq. 10) to remove fitting problems associated with the covariance of the concentration and molecular brightness.

Results

The primary goal of this work is to demonstrate how experimental resolution and model discrimination capabilities in FFS can be dramatically enhanced by using MFD and multi-method global analysis, here shown through the implementation of τ FCS. We thus demonstrate the capability to resolve the molecular composition of a mixture of two identical molecular weight fluorescence dyes, Rhodamine 6G and Rhodamine B, for which standard FCS experiments would be unable to identify the presence of the two sample components ($D_{R6G} = 390 \mu\text{m}^2\text{s}^{-1}$; $D_{RhB} = 465 \mu\text{m}^2\text{s}^{-1}$) [10] or to accurately recover their concentrations and other physical properties since their diffusion coefficients are too close [10]. The fluorescence lifetimes of these two dyes are easily resolved ($\tau_{R6G} = 3.92$ ns and $\tau_{RhB} = 1.63$ ns), and the use of MFD thus provides an important contrast parameter. However, lifetime fitting alone can only resolve lifetime values and fractional intensities and cannot determine molecular concentrations, diffusion coefficients, or molecular brightness values that report on interactions. Even after determining the presence of two species from the lifetime data, a traditional MFD approach fails to recover this information since FCS analysis of data from this mixture can still not achieve a stable fit for a two-component model. On the other hand, τ FCS analysis does produce stable fits and accurate parameter recovery, thus providing a vastly improved functionality over FCS and/or lifetime measurements alone.

To illustrate the strength of this method we analyze nine individual τ FCS data sets acquired from rhodamine dye mixtures with RhB and R6G dye concentrations ranging from approximately 3 to 97 nM, providing concentration ratios (C_{R6G}/C_{RhB}) spanning almost three orders of magnitude. Comparison of τ FCS fitting results (Fig. 1a, squares) to the known dye concentrations (Fig. 1a, solid lines) demonstrates accurate recovery of molecular concentrations for each species over a fairly wide range of concentration ratios. This achievement is significant considering that FCS or lifetime measurements alone are not able to recover this information at all. These results, fitted one R6G:RhB concentration ratio at a time, do require some *a priori* knowledge for one of the sample components in order to obtain stable data fits. Here we assume that the lifetime, molecular brightness, and diffusion coefficient of the R6G molecules (τ_{R6G} , Ψ_{R6G} & D_{R6G}) can be measured independently, thus serving as fixed fitting parameters during τ FCS analysis (see parameter values in Fig. 1b). We note, however, that prior knowledge of these three parameter values (τ_{R6G} , Ψ_{R6G} & D_{R6G}) is not sufficient to achieve stable fitting results using FCS analysis alone. Only by further assuming prior knowledge of the brightness of the second species can one achieve somewhat stable fits, and even then some of the numerically stable fitting results return inaccurate parameters. Any error in the assumed brightness values also translates directly into errors in the recovered concentrations (Fig. S3). The use of τ FCS avoids these problems without any assumptions about the second brightness value, and thus offers a significant advantage.

R6G has a higher molecular brightness than RhB and is the calibrated species in τ FCS analysis (fixed values for its diffusion coefficient, lifetime and brightness), and as such, its concentration (Fig. 1a, red squares) can be accurately recovered across all points of the titration. Concentrations of the less bright RhB are harder to measure, yet we still see accurate concentration measurements up to a concentration ratio of approximately 3 (Fig. 1a, blue squares). Above the concentration ratio of approximately 3 the recovered value of C_{RhB} becomes unstable in the curve fitting routines and is sensitive to initial parameter guesses due to the covariance of the molecular brightness and concentration parameters. These instabilities can also be seen in the recovery of the diffusion coefficient (Fig. 1c) and molecular brightness (Fig. 1d) of RhB, which are otherwise quite accurate at lower concentration ratios. This observation is not unexpected since at higher concentration ratios the RhB signal becomes an increasingly small fraction of the total fluorescence signal. Also, since each molecular species' contribution to the measured correlation function amplitudes varies with the square of its molecular brightness, the molecule with lower brightness is harder to measure at lower concentrations than the brighter molecular species. Lifetime measurements alone can accurately determine the fluorescence lifetimes of multiple sample components independently of any FCS analysis, reflected by the measured lifetime values which do not exhibit similar instabilities even at high concentration ratios (Fig. 1e).

It is clear that combining lifetime and FCS data for fitting as a single τ FCS data set results in greatly improved capabilities for resolving the molecular composition of a sample. The major limitations of the method as introduced so far are that some independent knowledge of the sample properties (e.g. τ_{R6G} , Ψ_{R6G} & D_{R6G}) was required for stable curve fitting and the RhB concentration was not recoverable at the higher concentration ratios. Each of these limitations can be overcome if one leverages the full power of global analysis. Specifically, if an experimental parameter can be varied over a series of measurements then global analysis allows curve fitting for the entire measurement series with




Figure 1. τ FCS analysis of binary dye mixtures with known R6G and RhB concentrations. (a) Recovery of molecular concentrations across nine known concentration ratios using τ FCS. Measured concentrations are shown as red and blue squares, and the solid lines show the known concentration of each dye mixture. (b) Sample parameters detailing concentrations, diffusion coefficients, molecular brightnesses and fluorescence lifetimes of the prepared R6G and RhB mixtures. Recovery of RhB diffusion coefficient (c), molecular brightness (d), and fluorescence lifetime (e) respectively. Data points and error bars represent the average and standard deviation of three repeated experiments. doi:10.1371/journal.pone.0090456.g001

common experimental parameters ‘linked’ across all data sets. For the example introduced here, the concentration ratios of the two dyes varies from one mixture to the next, but other dye properties including the fluorescent lifetimes, diffusion coefficients, and molecular brightness values must share the same values for all mixtures. Full global analysis of the data thus involves simultaneously fitting all of the data sets with a single set of globally linked fitting parameters for the lifetime, molecular brightness, and diffusion coefficient (τ , ψ & D) of each species in the sample. Two additional ‘local’ fitting parameters, C_{R6G} and C_{RhB} are associated with each individual data set to account for the different molecular concentrations of each rhodamine dye for each concentration ratio.

The results from the Global- τ FCS fit are shown in Fig. 2, and the returned global parameters are shown in Table 1. As can be seen immediately, the full global fits achieve remarkable accuracy across the entire measured concentration range. This includes accurate recovery of the molecular concentration of RhB, the species with lower molecular brightness, even when it constitutes only a few percent of the molecules within the sample. This level of sensitivity for a minor species goes well beyond what has been achievable in fluctuation measurements, demonstrating the greatly enhanced sensitivity and resolution available using the global τ FCS approach. Moreover, these extraordinary global fitting results were obtained *without any constraints* on fitting parameters, and unlike the individually fit τ FCS data sets the full global

fits assume no prior knowledge of molecular brightnesses, diffusion coefficients, lifetime values, or any other parameter for either species. These attributes offer tremendous advantages and flexibility in applications of this method. For complex experimental systems it may be impossible to isolate an individual molecular species for calibration purposes, or calibrations performed under one sample condition (e.g. diffusion coefficients for isolated molecules) may not accurately reflect actual values in a different sample condition (e.g. diffusion coefficients within a living cell). The full global analysis approach completely eliminates the need for any molecule specific calibration measurements, and still returns accurate results across a wide range of sample compositions.

We note that while global analysis on its own offers powerful enhancements in many curve fitting applications, the use of global analysis alone does not explain the success of the τ FCS approach. In particular, one can attempt to fit the same data shown above using a global FCS fitting approach (no lifetime data), but with only minimal success. Achieving numerically stable fitting results with global FCS (alone) first requires holding fixed the diffusion coefficient and molecular brightness parameters for one species, which is not required for global τ FCS, and even then the fit results are highly dependent upon initial guesses for parameter values (see Fig. S4) and can still be unstable. Thus, while one can sometimes get marginally reasonable fits to the data using a global FCS fit (though never close to the accuracy of τ FCS), such fits can also




Figure 2. Comparison of Global- τ FCS analysis with known R6G and RhB sample mixtures. Recovery of molecular concentrations across nine known concentration ratios using Global- τ FCS. Calibrated known concentrations (solid lines) compared to recovered concentrations of R6G (red circles) and RhB (blue circles). Data points and error bars represent the average and standard deviation of three repeated experiments.
doi:10.1371/journal.pone.0090456.g002

return inaccurate results and there is no reliable method to sort the accurate from the inaccurate fit results. The τ FCS fits, which leverage the enhanced capabilities of multi-modal global analysis, do not suffer from any such problems and return stable and accurate fitting results every time, independent of initial parameter guesses and without fixed fitting parameters. As with any global analysis methods, the success of this approach depends fundamentally upon there being common linkable parameters across various measurement conditions, i.e. a given molecular species has the same brightness in each sample condition.

When correct fitting models are identified, τ FCS yields highly accurate curve fitting results, as shown above. Since sample composition is often unknown *a priori*, an important consideration regarding the overall applicability of the τ FCS approach is the extent to which curve fitting procedures are able to successfully discriminate between different fitting models. To test this we employed computationally generated data sets for binary mixtures, and fitted data from multi-component samples to an (incorrect) single component model. We then determined the range of lifetime values and diffusion coefficient ratios for which curve fitting to a single species model indicated a sufficiently poor “goodness of fit” to warrant rejecting the single component model. Figure 3 shows these analyses, plotted as cross sections of χ^2 surfaces, compiled using four different single species models: FCS alone (green lines), standard lifetime alone (yellow lines), lifetime using an intensity constraint (blue lines), and τ FCS, also with an intensity constraint (red lines). Fitting with FCS alone yields a horizontal valley centered at diffusion coefficient ratio of 1, with boundaries that are consistent with Meseth *et al.* [10]. Similarly, the two lifetime models yield vertical valleys, with a slight asymmetry around the ratio of 1. This asymmetry is due to the finite pulse to pulse time window, in which the longer lifetimes of

Table 1. Molecular parameter values recovered using Global- τ FCS compared to known parameter values.

| | | Global- τ FCS | | Parameter Value |
|-----|--------|--------------------|----------|-----------------|
| | | Fit | Rel. Err | |
| R6G | D | 0.39 ± 0.01 | 0.8% | 0.39 ± 0.01 |
| | τ | 3.91 ± 0.01 | 0.2% | 3.92 ± 0.01 |
| | ψ | 22.3 ± 0.2 | 2.9% | 23.0 ± 0.70 |
| RhB | D | 0.46 ± 0.01 | 0.6% | 0.47 ± 0.02 |
| | τ | 1.57 ± 0.01 | 3.7% | 1.63 ± 0.01 |
| | ψ | 12.8 ± 0.1 | 0.5% | 12.8 ± 0.4 |

The accuracy of the recovered information goes well beyond previous experimental capabilities, which is more remarkable given that no fitting constraints or *a priori* assumptions were required for these fitting results.
doi:10.1371/journal.pone.0090456.t001

the second component do not fully decay at larger lifetime ratios. From these plots it is immediately clear that there are many experimental conditions for which either FCS or lifetime measurements alone cannot resolve two sample components. In contrast the τ FCS approach greatly decreases the range of parameter space, centered at $D_2/D_1 = \tau_2/\tau_1 = 1$, for which two species cannot be clearly resolved. This clearly indicates that τ FCS improves the confidence in model selection over standard FFS methods, and also provides experimental guidelines for how different physical parameter values must be to be easily resolved experimentally using τ FCS.

In Fig. 3 the τ FCS chi-squared contours fall slightly outside of the single method curves since they are essentially the average of the chi-squared values for the two different methods. For example, when the two diffusion coefficients are identical and the lifetimes different the FCS fit alone will yield “good fit” with $\chi^2 \sim 1$ while the lifetime fit alone will have a larger chi-squared value since it cannot fit the two component lifetime data as well with a single lifetime model. We note that the “good fit” for the FCS data in this example, as indicated by the chi-squared value, is in this case actually a “poor fit” for failing to resolve the presence of a second species that is contained within the sample. The τ FCS chi-squared, which again basically averages the FCS and lifetime chi-squared values, thus indicates a “less poor” fit than lifetime alone. The difference in lifetime values needed to bring the τ FCS chi-squared value up to the same value as lifetime fitting alone is what pushes the τ FCS contour slightly outside the lifetime or FCS alone curves. This feature of these plots should be understood in terms of the regions of parameter space for which the various methods can distinguish between one- and two-component samples, and does not indicate that τ FCS has lower resolution. To the contrary, as demonstrated above, the τ FCS significantly enhance curve fitting resolution fitting to a two component model when compared with either method used alone.

Discussion/Conclusions

We have introduced a new fluorescence fluctuation analysis technique, τ FCS, and have demonstrated that τ FCS can dramatically enhance sensitivity and resolution in fluctuation measurements, highlighted by the capability to measure concentrations, molecular brightness values, diffusion coefficients and fluorescence lifetimes from multiple molecular species even when their diffusion coefficients are very similar and one of the species makes up a small fraction of the total molecular population within




Figure 3. Identifying a second component from simulated mixture data via incorrect single component analysis. These curves shown represent the chi-squared contours for the different fitting methods when fitting two-component data with a single component model. When the χ^2 value exceeds ~ 1.3 we conclude that the single component analysis fails, indicating the presence of a second component in the sample. Four χ^2 surfaces demonstrate the boundary in parameter space for 1) FCS (green lines); 2) Fluorescence Lifetime (yellow lines); 3) Fluorescence Lifetime with average intensity constraint (blue lines); and 4) τ FCS with average intensity constraint (red lines). doi:10.1371/journal.pone.0090456.g003

the sample. Moreover, Global- τ FCS removes the need to fix any of the fitting parameters during the analysis, and thus removes the need for independent measurements on isolated sample components which are often not possible to measure. The freedom from parameter assumptions and calibrations, together with the marked improvement in accuracy highlight the benefits of Global- τ FCS. τ FCS also has the potential to significantly enhance detection of molecular interactions. When a molecule binds to another molecule of similar molecular weight its change in diffusion coefficient is typically too small to be resolved via FCS analysis alone. However, it is not unusual for the fluorescence to be quenched (or dequenched) upon binding, which has an associated change in the fluorescence lifetime and thus potentially making the interaction resolvable via τ FCS. The sensitivity of τ FCS, with the possibility to resolve even a small fraction of a molecular population involved in interactions as demonstrated above, further enhances the utility of this method for resolving such interactions.

There are two important principles underlying the success of the τ FCS approach. First, by using MFD it is often possible to find a contrast parameter that allows resolution of multiple sample components that are otherwise disguised in a particular measurement. For example, in the case of τ FCS the fluorescence lifetime serves as the contrast parameter which detects the presence of two sample components when the FCS measurement alone would not resolve a second species. Second, and as importantly, global fitting of multiple measurement modes (e.g. FCS and lifetime) simultaneously enhances the resolution of each individual method during curve fitting. The different methods have unique dependencies on the global parameters, such as lifetime data depending linearly on the molecular brightness but the FCS data amplitude depending on the square of the molecular brightness. Thus, an analysis method that forces the fitting parameters to account for total fluorescence signal across independent measurement modes, but with common global parameters describing each, has a greatly

constrained fitting parameter space and produces much better curve fitting results.

Also serving the success of this approach is the strategy to leverage all the information content of the fluorescence signal. This includes, for example, making use of the often ignored amplitude information in the lifetime data as is done in τ FCS. While illustrated here for the specific case of FCS and lifetime measurements, this general principle can and should be further exploited to maximize use of the information content of the fluorescence signal. For example, measurements of molecular rotations via fluorescence anisotropy or spectral signatures through spectrally resolved detection could serve as further extensions of this approach. When coupled with the full global analysis approach, such additional contrast parameters will further provide important ways to enhance sensitivity and resolution for characterizing the molecular scale composition and dynamics of molecules within complex systems.

A less considered, but equally as important topic is that of determining the correct fit model, without assumptions, using data alone. The increased ability to resolve small fractions of different species provides an inherent capacity to assist in model discrimination. We have demonstrated the ability of τ FCS in distinguishing the presence of a second species across a wide range of parameter combinations. The array of χ^2 values demonstrates the increased parameter space for which τ FCS can separate species and provides a metric for experimental design.

Measurements as described in this work are reasonably straightforward to implement. The theory presented is valid as long as the sample does not change in any way from one measurement to the next. The measurements thus need not in principle be acquired simultaneously, although in practice that is likely the best approach to ensure no changes in the sample from one measurement mode to the next. Some additional hardware may be required, such as time correlated single photon counting (TCSPC) electronics for lifetime measurements, but it is generally

very practical to perform multiple modes of measurement at the same time, on the same instrument, for any given sample. Certainly lifetime and FCS data can be acquired simultaneously, and extending such measurements to include molecular rotations, spectra, and other parameters is a straight forward process using modern hardware. Thus, any time fluctuation measurements are contemplated, a multi-modal global analysis approach is probably also accessible – with improved accuracy and resolution as demonstrated above. It is thus likely that methods such as τ FCS will eventually supersede fluctuation methods alone.

Supporting Information

Figure S1 Comparison of τ FCS and global- τ FCS analyses. τ FCS analysis (a) allows brightnesses and concentrations to be treated globally across the two data modalities due to common parameters now describing the amplitudes of both the lifetime decay and autocorrelation function. Global- τ FCS (b) intrinsically retains the pair-global relationship for each pair individually, in addition to treating the lifetimes, diffusion coefficients, and brightnesses as global parameters across all pairs of the titration. Data has been normalized for visual comparison.

(TIF)

Figure S2 Reference data set recorded using uncorrelated light to assess the systematic error in data acquisitions. A 100 point binomial smoothed data set (black line) removes Poissonian noise while retaining the lower frequencies (red line) used for data corrections.

(TIF)

Figure S3 The effects of molecular brightness assumptions in two component FCS analyses compared to τ FCS. Comparison of simulated data sets depicting a binary system with a diffusion coefficient ratio of 3, molecular brightness and lifetime ratios of 2 (a), and titrated across a concentration ratio of three orders of magnitude (b). Here, we have ‘calibrated’ species a and fixed the known values for D_a , ψ_a and τ_a during all subsequent analyses. The covariant autocorrelation amplitudes require that the molecular brightnesses be held for both species using FCS analysis; therefore, we have “guessed” ψ_b in order to attain stable fits. Three different analyses (b) using ψ_b guesses below ($\psi_b = 8$ kcpms; horizontal kites), the same as ($\psi_b = 10$ kcpms; diamonds), and above ($\psi_b = 12.5$ kcpms; horizontal kites) the correct molecular brightness value highlight the potential inaccuracies in two component FCS results. Across concentration ratios C_a/C_b of 0.03 to 1, in which the amount of unknown species is sufficiently large, FCS analysis can distinguish the 2nd

component, albeit with molecular brightness guess dependent errors. Beyond the C_a/C_b of approximately 3, FCS analysis fails to identify two species and transitions into a fit result that finds two identical species of equal concentration, that of half the total. This is corroborated by the transition of the returned diffusion coefficient, D_b , from 0.1 to $0.3 \mu\text{m}^2\text{ms}^{-1}$ (e; all blue data points). τ FCS analysis (b; gold data points) of the same titration data set, in which no assumptions or held parameters are enforced on the 2nd species, returns accurate results across the entire range, even in the case of a very small fraction of the less bright species (gold circles). τ FCS also returns accurate molecular brightnesses (c) and fluorescence lifetimes (d) across the majority of the titration range, and still distinguishes the diffusion coefficient where FCS analysis fails (e; gold triangles). Data points and error bars indicate the average and standard deviation of three independent simulated data sets.

(TIF)

Figure S4 Comparison of initial guesses in global analysis of experimental FCS data only. To demonstrate that multi-method global analysis, here implemented as τ FCS, is the key to the accuracy of the results shown in Fig. 3, we show here that global analysis of FCS data alone does not return comparable accuracy, or even stable fitting results. Shown here are experimental autocorrelation data of binary RhB and R6G mixtures, with individually calibrated parameters (a), subject to global analysis of FCS data only (no lifetime data) incorporating repeated titration data sets. Concentrations are considered local fit parameters while diffusion coefficients and molecular brightness are global. Here, the diffusion coefficient and molecular brightness of R6G have been held fixed at the correct value during fitting. Four different initial guess combinations for the diffusion coefficient and molecular brightness of RhB are shown (b-e). Some initial guesses return somewhat stable fits, albeit with inaccurate results, while other initial guesses can lead to extremely unstable fits. Data points and error bar reflect the average and standard deviation of the three different fits to the independently acquired data sets. These fitting results show that global analysis of FCS data alone cannot accurately fit the data.

(TIF)

Author Contributions

Conceived and designed the experiments: NA KB. Performed the experiments: NA. Analyzed the data: NA. Contributed reagents/materials/analysis tools: NA KB. Wrote the paper: NA KB. Derived theory: KB. Designed software used in analysis: NA.

References

- Lippincott-Schwartz J, Snapp E, Kenworthy A (2001) Studying protein dynamics in living cells. *Nat Rev Mol Cell Bio* 2: 444–456.
- Dehmelt L, Bastiaens PIH (2010) Spatial organization of intracellular communication: insights from imaging. *Nat Rev Mol Cell Bio* 11: 440–452.
- Tsien RY (2003) Imagining imaging's future. *Nat Rev Mol Cell Bio*: SS16.
- Magde D, Webb WW, Elson E (1972) Thermodynamic Fluctuations in a Reacting System - Measurement by Fluorescence Correlation Spectroscopy. *Phys Rev Lett* 29: 705–708.
- Eigen M, Rigler R (1994) Sorting Single Molecules - Application to Diagnostics and Evolutionary Biotechnology. *P Natl Acad Sci USA* 91: 5740–5747.
- Rigler R, Elson E (2001) *Fluorescence Correlation Spectroscopy Theory and Applications*: Springer, Berlin.
- Webb WW (2001) Fluorescence correlation spectroscopy: inception, biophysical experimentations, and prospectus. *Appl Optics* 40: 3969–3983.
- Elson EL (2011) Fluorescence Correlation Spectroscopy: Past, Present, Future. *Biophys J* 101: 2855–2870.
- Schville P, Kim SA, Heinze KG (2007) Fluorescence correlation spectroscopy in living cells. *Nat Methods* 4: 963–973.
- Meseth U, Wohland T, Rigler R, Vogel H (1999) Resolution of fluorescence correlation measurements. *Biophys J* 76: 1619–1631.
- Berland KM, So PTC, Chen Y, Mantulin WW, Gratton E (1996) Scanning two-photon fluctuation correlation spectroscopy: Particle counting measurements for detection of molecular aggregation. *Biophys J* 71: 410–420.
- Chen Y, Muller JD, Berland KM, Gratton E (1999) Fluorescence Fluctuation Spectroscopy. *Methods* 19: 234–252.
- Chen Y, Muller JD, So PTC, Gratton E (1999) The photon counting histogram in fluorescence fluctuation spectroscopy. *Biophys J* 77: 553–567.
- Kask P, Palo K, Fay N, Brand L, Mets U, et al. (2000) Two-dimensional fluorescence intensity distribution analysis: Theory and applications. *Biophys J* 78: 1703–1713.
- Wu B, Chen Y, Muller JD (2006) Dual-color time-integrated fluorescence cumulant analysis. *Biophys J* 91: 2687–2698.
- Muller JD (2004) Cumulant analysis in fluorescence fluctuation spectroscopy. *Biophys J* 86: 3981–3992.
- Wu B, Muller JD (2005) Time-integrated fluorescence cumulant analysis in fluorescence fluctuation spectroscopy. *Biophys J* 89: 2721–2735.

18. Bacia K, Kim SA, Schwille P (2006) Fluorescence cross-correlation spectroscopy in living cells. *Nat Methods* 3: 83–89.
19. Kask P, Palo K, Ullmann D, Gall K (1999) Fluorescence-intensity distribution analysis and its application in biomolecular detection technology. *P Natl Acad Sci USA* 96: 13756–13761.
20. Muller JD, Chen Y, Gratton E (2000) Resolving heterogeneity on the single molecular level with the photon-counting histogram. *Biophys J* 78: 474–486.
21. Schwille P, MeyerAlmes FJ, Rigler R (1997) Dual-color fluorescence cross-correlation spectroscopy for multicomponent diffusional analysis in solution. *Biophys J* 72: 1878–1886.
22. Widengren J, Kudryavtsev V, Antonik M, Berger S, Gerken M, et al. (2006) Single-molecule detection and identification of multiple species by multiparameter fluorescence detection. *Anal Chem* 78: 2039–2050.
23. Brand L, Eggeling C, Zander C, Drexhage KH, Seidel CAM (1997) Single-molecule identification of Coumarin-120 by time-resolved fluorescence detection: Comparison of one- and two-photon excitation in solution. *J Phys Chem A* 101: 4313–4321.
24. Kühnemuth R, Seidel CAM (2001) Principles of Single Molecule Multiparameter Fluorescence Spectroscopy. *Single Molecules* 2: 251–254.
25. Weidtkamp-Peters S, Felekyan S, Bleckmann A, Simon R, Becker W, et al. (2009) Multiparameter fluorescence image spectroscopy to study molecular interactions. *Photoch Photobio Sci* 8: 470–480.
26. Nguyen TA, Sarkar P, Veetil JV, Koushik SV, Vogel SS (2012) Fluorescence Polarization and Fluctuation Analysis Monitors Subunit Proximity, Stoichiometry, and Protein Complex Hydrodynamics. *PLOS ONE* 7: e38209.
27. Verveer PJ, Squire A, Bastiaens PIH (2000) Global analysis of fluorescence lifetime imaging microscopy data. *Biophys J* 78: 2127–2137.
28. Knutson JR, Beechem JM, Brand L (1983) Simultaneous Analysis of Multiple Fluorescence Decay Curves - A Global Approach. *Chem Phys Lett* 102: 501–507.
29. Skakun VV, Hink MA, Digris AV, Engel R, Novikov EG, et al. (2005) Global analysis of fluorescence fluctuation data. *Eur Biophys J* 34: 323–334.
30. Laptinok SP, Borst JW, Mullen KM, van Stokkum IHM, Visser AJWG, et al. (2010) Global analysis of Förster resonance energy transfer in live cells measured by fluorescence lifetime imaging microscopy exploiting the rise time of acceptor fluorescence. *Phys Chem Chem Phys* 12: 7593–7602.
31. Rahim NAA, Pelet S, Kamm RD, So PTC (2012) Methodological considerations for global analysis of cellular FLIM/FRET measurements. *J Biomed Opt* 17: 026013–026013.
32. Previte MJR, Pelet S, Kim KH, Buehler C, So PTC (2008) Spectrally resolved fluorescence correlation spectroscopy based on global analysis. *Anal Chem* 80: 3277–3284.
33. Beechem JM (1992) Global Analysis of Biochemical and Biophysical Data. *Methods in Enzymology*: Academic Press Inc., New York. pp. 37–54.
34. Kapusta P, Wahl M, Benda A, Hof M, Enderlein J (2007) Fluorescence lifetime correlation spectroscopy. *J Fluoresc* 17: 43–48.
35. Böhmer M, Wahl M, Rahn H-J, Erdmann R, Enderlein J (2002) Time-resolved fluorescence correlation spectroscopy. *Chem Phys Lett* 353: 439–445.
36. Mertz J (1998) Molecular photodynamics involved in multi-photon excitation fluorescence microscopy. *Eur Biophys J D* 3: 53–66.
37. Berland K, Shen GQ (2003) Excitation saturation in two-photon fluorescence correlation spectroscopy. *Appl Optics* 42: 5566–5576.
38. Cianci GC, Wu J, Berland KM (2004) Saturation modified point spread functions in two-photon microscopy. *Microsc Res Tech* 64: 135–141.
39. Nagy A, Wu JR, Berland K (2004) Photodynamics in fluorescence correlation spectroscopy. *Biophys J* 86: 320A–320A.
40. Nagy A, Wu J, Berland KM (2005) Observation Volumes and Gamma-Factors in Two-Photon Fluorescence Fluctuation Spectroscopy. *Biophys J* 89: 2077–2090.
41. Wu J, Berland K (2007) Fluorescence intensity is a poor predictor of saturation effects in two-photon microscopy: Artifacts in fluorescence correlation Spectroscopy. *Microsc Res Tech* 70: 682–686.
42. Denk W, Strickler JH, Webb WW (1990) Two-photon laser scanning fluorescence microscopy. *Science* 248: 73–76.
43. So PTC, Dong CY, Masters BR, Berland KM (2000) Two-photon Excitation Fluorescence Microscopy. In: Yarmush ML, editor. *Annual Review of Biomedical Engineering*. Palo Alto: Annual Reviews. pp. 399–429.
44. Xu C, Webb WW (1997) Multiphoton excitation of molecular fluorophores and nonlinear laser microscopy. In: Lakowicz J, editor. *Topics in Fluorescence Spectroscopy*: Plenum. pp. 471–540.
45. Thompson NL (1991) Fluorescence Correlation Spectroscopy. In: Lakowicz JR, editor. *Topics in Fluorescence Spectroscopy*. New York: Plenum, New York. pp. 337–378.
46. Elson EL, Magde D (1974) Fluorescence Correlation Spectroscopy .1. Conceptual Basis and Theory. *Biopolymers* 13: 1–27.
47. Mertz J, Xu C, Webb WW (1995) Single-molecule detection by two-photon-excited fluorescence. *Opt Lett* 20: 2532–2534.
48. Schwille P, Haupts U, Maii S, Webb WW (1999) Molecular dynamics in living cells observed by fluorescence correlation spectroscopy with one- and two-photon excitation. *Biophys J* 77: 2251–2265.
49. Chen Y, Muller JD, Eid JS, Gratton E (2001) Two-photon Fluorescence Fluctuation Spectroscopy. In: Valeur B, Brochon JC, editors. *New Trends in Fluorescence Spectroscopy: Applications to Chemical and Life Sciences*: Springer-Verlag, Berlin.
50. Krichevsky O, Bonnet G (2002) Fluorescence correlation spectroscopy: the technique and its applications. *Rep Prog Phys* 65: 251–297.
51. Korlach J, Schwille P, Webb WW, Feigensohn GW (1999) Characterization of lipid bilayer phases by confocal microscopy and fluorescence correlation spectroscopy. *P Natl Acad Sci USA* 96: 8461–8466.
52. Berland KM (2004) Detection of specific DNA sequences using dual-color two-photon fluorescence correlation spectroscopy. *J Biotechnol* 108: 127–136.
53. Petrasek Z, Schwille P (2008) Precise measurement of diffusion coefficients using scanning fluorescence correlation spectroscopy. *Biophys J* 94: 1437–1448.
54. Mukhtar E, Bergstrom F, Johansson LBA (2002) Hyper Rayleigh scattering yields improved response function in analysing 2-photon excited fluorescence. *J Fluoresc* 12: 481–484.
55. Marchand P, Marmet L (1983) Binomial Smoothing Filter - a Way to Avoid Some Pitfalls of Least-Squares Polynomial Smoothing. *Rev Sci Instrum* 54: 1034–1041.
56. Becker W (2005) *Advanced Time-Correlated Single Photon Counting Techniques*: Springer, Berlin.
57. Starchev K, Ricka J, Buffle J (2001) Noise on fluorescence correlation spectroscopy. *J Colloid Interf Sci* 233: 50–55.