

## A novel approach for endoscopic submucosal dissection: scissors versus electricity



Sergey V. Kantsevoy, MD, PhD,<sup>1,2</sup> Shira Levihim, PA,<sup>1</sup> Deborah Pennington, MSN, RN, CGRN,<sup>1</sup> Amit Raina, MD<sup>1</sup>


EMR and endoscopic submucosal dissection (ESD) are widely used for removal of premalignant and malignant lesions of the GI tract.<sup>1-3</sup> Currently used polypectomy snares and available ESD knives use electric energy for resection of GI tract lesions.<sup>4-6</sup>

SB Knives (Olympus America, Center Valley, Pa, USA) and the Clutch Cutter (Fujifilm, Tokyo, Japan) have recently become available in the United States.<sup>7-11</sup> Although these devices have 2 branches and resemble scissors in appearance, their branches are not sharp and cannot mechanically cut tissues. Instead, both SB Knives and the Clutch Cutter use electric energy similarly to other available electrosurgical ESD knives.


Unfortunately, electric energy causes collateral damage to adjacent tissues, resulting in postprocedural pain, post-polypectomy syndrome, and delayed adverse events (tissue necrosis, bleeding, and perforation).<sup>12-17</sup> To eliminate collateral tissue damage by electric current, previous publications have advocated the use of “cold” snares for removal of small colonic polyps.<sup>18-20</sup>

We describe colonic ESD with recently developed endoscopic monopolar scissors (Ensizor; Slater Endoscopy, Miami Lakes, Fla, USA), which uses “cold” tissue cutting instead of “hot” dissection with electric current.


A 63-year-old man was referred for endoscopic resection of a difficult (sessile, flat, IIa, Kudo IIIs, 20 mm) descending colon polyp that had been partially removed during




**Figure 2.** A circumferential mechanical incision around the polyp started with endoscopic scissors.




**Figure 1.** A sessile, flat (IIa, Kudo IIIs) descending colon polyp partially removed during previous colonic EMR.



**Figure 3.** The polyp is attached to the fore-balloon of the retraction device with endoscopic clip and pulled in an oral direction, exposing extensive fibrosis in the submucosal space. Endoscopic submucosal dissection was performed by cutting fibrotic tissue with scissors.




**Figure 4.** Hemostasis during endoscopic submucosal dissection is achieved by use of the electrocautery function of the scissors. **A**, Active bleeding from a large submucosal vessel. **B**, Tip of the scissors is placed in contact with the bleeding vessel. **C**, Bleeding vessel is coagulated by the scissors. **D**, Bleeding is completely stopped after application of electrocautery through the scissors.



**Figure 5.** The polyp is removed en bloc.


previous colonic EMR (Fig. 1; Video 1; available online at [www.VideoGIE.org](http://www.VideoGIE.org)). After submucosal injection, a circumferential mechanical incision around the polyp was made with endoscopic scissors (Fig. 2). The polyp was attached to the fore-balloon of a DiLumen retraction device (Lumendi, Westport, Conn, USA) with an endoscopic clip (Resolution 360; Boston Scientific, Natick, Mass, USA). The



**Figure 6.** Overstitch endoscopic suturing device is delivered to the endoscopic submucosal dissection site through the DiLumen retractor.

fore-balloon was pushed in an oral direction, exposing extensive fibrosis in the submucosal space. ESD was performed by cutting the fibrotic tissue with scissors (Fig. 3), followed by placement of a second clip for additional traction.





**Figure 7.** Mucosal defect after lesion removal is completely closed with 1 continuous suture.

Hemostasis during ESD was achieved by touching the blood vessels with the tip of the monopolar scissors in soft coagulation mode (50 W, effect 2) (Fig. 4). The polyp was removed en bloc (Fig. 5), and the mucosal defect post lesion removal was closed with 1 continuous suture (Figs. 6 and 7) by use of the Overstitch endoscopic suturing device (Apollo Endosurgery, Austin, Tex, USA). The patient was discharged home after the procedure. Pathologic examination revealed tubular adenoma and confirmed R0 resection with negative margins.

In conclusion, newly designed endoscopic scissors allow ESD and effective endoscopic hemostasis. ESD with mechanical scissors instead of the currently used electrosurgical knives eliminates collateral damage of tissues adjacent to the plane of dissection, preserving the margins of the specimen for histologic assessment and preventing postprocedural pain, postpolypectomy syndrome, and delayed adverse events (tissue necrosis, bleeding, and perforation). The use of a traction device facilitates ESD and serves as a conduit for endoscopic suturing closure of large mucosal defects after lesion removal.

## DISCLOSURE

*Dr Kantsevov is a consultant for Apollo Endosurgery, Aries, Endocages, LumenDi, Medtronic, Olympus, and Vizballoons; is a co-founder of Apollo Endosurgery and Endocages; is a shareholder in Apollo Endosurgery; is on LumenDi advisory board; is in active litigation with LumenR; and an equity holder in Endocages,*

*LumenDi, Slater Endoscopy and Vizballoons. All other authors disclosed no financial relationships relevant to this publication.*

## ACKNOWLEDGEMENT

The authors would like to thank Slater Endoscopy (Miami Lakes, Fla, USA) for providing Ensizor endoscopic monopolar scissors for this procedure.

*Abbreviation: ESD, endoscopic submucosal dissection.*

## REFERENCES

1. Fujiya M, Tanaka K, Dokoshi T, et al. Efficacy and adverse events of EMR and endoscopic submucosal dissection for the treatment of colon neoplasms: a meta-analysis of studies comparing EMR and endoscopic submucosal dissection. *Gastrointest Endosc* 2015;81:583-95.
2. Cao Y, Liao C, Tan A, et al. Meta-analysis of endoscopic submucosal dissection versus endoscopic mucosal resection for tumors of the gastrointestinal tract. *Endoscopy* 2009;41:751-7.
3. Kobayashi N, Yoshitake N, Hirahara Y, et al. Matched case-control study comparing endoscopic submucosal dissection and endoscopic mucosal resection for colorectal tumors. *J Gastroenterol Hepatol* 2011;27:728-33.
4. Kantsevov SV, Adler DG, Conway JD, et al. Endoscopic mucosal resection and endoscopic submucosal dissection. *Gastrointest Endosc* 2008;68:11-8.
5. ASGE Technology Committee; Maple JT, Abu Dayyeh BK, Chauhan SS, et al. Endoscopic submucosal dissection. *Gastrointest Endosc* 2015;81:1311-25.
6. ASGE Technology Committee; Tokar JL, Barth BA, Banerjee S, et al. Electrosurgical generators. *Gastrointest Endosc* 2013;78:197-208.
7. Homma K, Otaki Y, Sugawara M, et al. Efficacy of novel SB knife Jr examined in a multicenter study on colorectal endoscopic submucosal dissection. *Dig Endosc* 2012;24:117-20.
8. Yamashina T, Takeuchi Y, Nagai K, et al. Scissor-type knife significantly improves self-completion rate of colorectal endoscopic submucosal dissection: single-center prospective randomized trial. *Dig Endosc* 2017;29:322-9.
9. Takasago T, Kuwai T, Yamaguchi T, et al. Endoscopic submucosal dissection with a scissors-type knife for post-EMR recurrence tumor involving the colon diverticulum. *VideoGIE* 2017;2:211-2.
10. Ge PS, Thompson CC, Aihara H. Endoscopic submucosal dissection of a large cecal polyp using a scissor-type knife: implications for training in ESD. *VideoGIE* 2018;3:313-5.
11. Yoshida N, Inada Y, Murakami T, et al. Colorectal endoscopic submucosal dissection for a lesion on the dentate line area resected with a scissor-type knife. *VideoGIE* 2018;3:223-5.
12. Harada H, Suehiro S, Murakami D, et al. Clinical impact of prophylactic clip closure of mucosal defects after colorectal endoscopic submucosal dissection. *Endosc Int Open* 2017;5:E1165-71.
13. Ikezawa K, Michida T, Iwahashi K, et al. Delayed perforation occurring after endoscopic submucosal dissection for early gastric cancer. *Gastric Cancer* 2011;15:111-4.
14. Miyahara K, Iwakiri R, Shimoda R, et al. Perforation and postoperative bleeding of endoscopic submucosal dissection in gastric tumors: analysis of 1190 lesions in low- and high-volume centers in Saga, Japan. *Digestion* 2012;86:273-80.

15. Oka S, Tanaka S, Kanao H, et al. Current status in the occurrence of postoperative bleeding, perforation and residual/local recurrence during colonoscopic treatment in Japan. *Dig Endosc* 2010;22: 376-80.
16. Conio M, Repici A, Demarquay JF, et al. EMR of large sessile colorectal polyps. *Gastrointest Endosc* 2004;60:234-41.
17. Piraka C. Cold snare resection of large duodenal and colonic polyps. *Gastroenterol Hepatol (N Y)* 2018;14:539-41.
18. Tachikawa J, Chiba H, Kuwabara H, et al. Successful 2-channel cold snare polypectomy of a colorectal lesion involving the appendiceal orifice. *VideoGIE* 2018;3:279-80.
19. Chang LC, Shun CT, Hsu WF, et al. Risk of delayed bleeding before and after implementation of cold snare polypectomy in a screening colonoscopy setting. *Endosc Int Open* 2019;7:E232-8.
20. Piraka C, Saeed A, Waljee AK, et al. Cold snare polypectomy for non-pedunculated colon polyps greater than 1 cm. *Endosc Int Open* 2017;5:E184-9.

---

Institute for Digestive Health and Liver Diseases, Mercy Medical Center (1); University of Maryland School of Medicine (2), Baltimore, Maryland, USA.

If you would like to chat with an author of this article, you may contact Dr Kantsevov at [skan51@hotmail.com](mailto:skan51@hotmail.com).

Copyright © 2019 American Society for Gastrointestinal Endoscopy. Published by Elsevier Inc. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

<https://doi.org/10.1016/j.vgje.2019.05.010>

---

**Read Articles in Press Online Today!**  
**Visit [www.videogie.org](http://www.videogie.org)**

*VideoGIE* posts in-press articles online in advance of their appearance in a monthly edition of the journal. These articles are available on the *VideoGIE* website by clicking on the “Articles in Press” tab. Articles in Press represent the final edited text of articles that are accepted for publication but not yet scheduled to appear in a specific issue. They are considered officially published as of the date of Web publication, which means readers can access the information and authors can cite the research months prior to its availability in an issue. To cite Articles in Press, include the journal title, year, and the article’s Digital Object Identifier (DOI), located in the article footnote. Visit the website today to stay current on the latest research in the field of gastrointestinal endoscopy.