

Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.

Comment





Published Online September 23, 2020 https://doi.org/10.1016/ S1473-3099(20)30699-X See Articles page 1390

The complexities of SARS-CoV-2 serology

Diagnosing previous infection with respiratory viruses is challenging. Our understanding of individual and population-level immunity to severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) remains incomplete and developing reliable serological assays to detect previous infection has been an intense focus of the global scientific effort. For public health planning we need scalable assays validated against large banks of samples from individuals who had proven seasonal (non-severe acute respiratory syndrome) coronaviruses and those who had well characterised symptomatic and asymptomatic confirmed SARS-CoV-2 infection. False-positive results, due to cross-reactivity with seasonal coronaviruses, are important to avoid, particularly if seropositive-individuals consider themselves immune. In The Lancet Infectious Diseases, the National SARS-CoV-2 Serology Assay Evaluation Group¹ provide the first large comparative investigation of the performance of four widely available commercial assays and a single in-house assay.

Antibody responses to SARS-CoV-2 are predominantly directed at the spike glycoprotein, which the virus requires for entry, and the nucleocapsid protein, which binds the viral RNA genome. The SARS-CoV-2 IgG assay (Abbott, Chicago, IL, USA) and Elecsys Anti-SARS-CoV-2 assay (Roche, Basel, Switzerland) assays detect antibody to the nucleoprotein, whereas the LIAISON SARS-CoV-2 S1/S2 IgG assay (DiaSorin, Saluggia, Italy), and SARS-CoV-2 Total assay (Siemens, Munich, Germany) detect antibodies to the spike glycoprotein. The Abbott and Diasorin assays detect IgG only, whereas Roche and Siemens detect total antibody. The diverse approaches taken by the four commercial assays highlight the challenge of choice posed to laboratories: all manufacturers report similarly high sensitivity and specificity.

The authors compared these four assays and a novel 384-well ELISA detecting total IgG to a trimeric spike protein and used all five assays on 976 pre-pandemic samples presumed to be negative, collected between 2014 and 2016, and 536 serum samples from patients with laboratory-confirmed COVID-19 from research studies in Oxford, UK, or plasma donors. The authors report that all assays had a high sensitivity (92·7-99·1%) and specificity (98·7-99·9%). The most sensitive test assessed was the in-house ELISA. The Abbott, Roche,

and Siemens assays were the most specific. The benefit of the huge sample bank available to these authors was the clearly documented time since PCR positivity, which allowed them to optimise the manufacturers' cut offs and improve sensitivity. Only three cases did not give rise to any detectable antibody responses in all five of the assays, possibly because of a genuine lack of response in infected individuals, or a false-positive quantitative PCR result.

A limitation of this work is the small number of pauci-symptomatic and asymptomatic cases analysed. Antibody responses in these individuals are likely to be lower, and therefore the sensitivity of all assays might be somewhat less than that reported. Also, data on sex, age, and immunocompromise status were incomplete, meaning that the results could be limited in their application to specific patient groups. This limitation could be especially important in children, who are more likely than adults are to have had a recent infection with a seasonal coronavirus.

The expectation is that the best predictor of antibodymediated protection will come from neutralisation assays, in which the ability of patient serum to prevent live virus infecting cell cultures is measured. These assays are impractical to deploy at scale. The presence of antibodies against the spike protein of SARS-CoV-2 correlates well with neutralisation.^{2,3} The DiaSorin, Siemens, and in-house assays measured these potentially protective antibodies, with the in-house ELISA using trimerised spike protein, which shows a high correlation with neutralisation.4.5.6 Further work is required to investigate what titre of neutralising antibodies correlates with protection, how long neutralisation activity persists, and which assay best predicts that. Identifying an appropriate assay will be crucial for assessing vaccine responses, and for assessing potential risk of reinfection, which has been shown with seasonal coronaviruses,⁷ but not so far for SARS-CoV-2. Consistent with this future possibility, the neutralisation potency of serum declines in the months post infection.^{8,9}

As our understanding of immunity and the correlates of protection (both cellular and humoral) increases and the range of immunoassays multiplies, we will probably use different assays to answer specific questions. For example, most vaccine candidates elicit responses to spike rather than nucleocapsid protein. Measuring antibodies to spike will therefore indicate whether there has been a good response, whereas measuring antibodies to nucleocapsid would help identify whether the individual had nonetheless become infected. Measuring the different antibodies might also have prognostic value; a report showed that a predominant humoral response to nucleoprotein is associated with poor outcome in patients admitted to hospital, compared with that of spike.¹⁰ Further investigation is required and the possibility of a one-size-fits-all immunological assay looks less and less likely.

We declare no competing interests.

*Catherine F Houlihan, Rupert Beale c.houlihan@ucl.ac.uk

University College London Hospitals, NHS Foundation Trust, London, UK (CFH); University College London, London, UK (CFH, RB); and The Francis Crick Institute, London, UK (RB)

- 1 The National SARS-CoV-2 Serology Assay Evaluation Group. Performance characteristics of five immunoassays for SARS-CoV-2: a head-to-head benchmark comparison. *Lancet Infect Dis* 2020; published online Sept 23. https://doi.org/10.1016/S1473-3099(20)30634-4.
- 2 Okba N, Müller MA, Li W, et al. Severe acute respiratory syndrome coronavirus 2-specific antibody responses in coronavirus disease patients. *Emerg Infect Dis* 2020; 26: 1478–88.

- 3 Folegatti PM, Ewer KJ, Aley PK, et al. Safety and immunogenicity of the ChAdOx1 nCoV-19 vaccine against SARS-CoV-2: a preliminary report of a phase 1/2, single-blind, randomised controlled trial. Lancet 2020; published online July 20. https://doi.org/10.1016/S0140-6736(20)31604-4.
- Convalescent plasma therapy for the treatment of patients with COVID-19: Assessment of methods available for antibody detection and their correlation with neutralising antibody levels. *medRxiv* 2020; published online May 26. https://doi.org/10.1101/2020.05.20.20091694 (preprint).
- 5 Harvala H, Robb M, Watkins N, et al. Convalescent plasma therapy for the treatment of patients with COVID-19: assessment of methods available for antibody detection and their correlation with neutralising antibody levels. MedRxiv 2020; published online May 26. https://doi. org/10.1101/2020.05.20.20091694 (preprint).
- 6 Wajnberg A, Amanat F, Firpo A, Altman DR. SARS-CoV-2 infection induces robust, neutralizing antibody responses that are stable for at least three months. *MedRxiv* 2020; published online July 17. https://doi.org/10.1101/2020.07.14.20151126 (preprint).
- 7 Kiyuka KP, Agoti CN, Munywoki PK, et al. Human coronavirus NL63 molecular epidemiology and evolutionary patterns in rural coastal Kenya. J Infect Dis 2018; 217: 1728–39
- 8 Muecksch F, Wise H, Batchelor B, et al. Longitudinal analysis of clinical serology assay performance and neutralising antibody levels in COVID19 convalescents. *MedRxiv* 2020; published online Aug 6. https://doi.org/10.1101/2020.08.05.20169128 (preprint).
- Long Q, Tang X, Shi Q, et al. Clinical and immunological assessment of asymptomatic SARS-CoV-2 infections. 2020 Nat Med **26:** 1200–04.
- 10 Atyeo C, Fischinger S, Zohar T, Slein MD, et al. Distinct early serological signatures track with SARS-CoV-2 survival. J Immuni 2020; published online July 30. https://doi.org/ 10.1016/j.immuni.2020.07.020.

Using serological data to understand unobserved SARS-CoV-2 risk in health-care settings

q

During past outbreaks of severe acute respiratory syndrome and Middle East respiratory syndrome, many infections occurred within health-care settings.¹ Since the emergence of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), growing evidence of nosocomial transmission has been observed, but tracking such outbreaks is challenging because a substantial proportion of infected individuals might exhibit mild or no symptoms.2 In The Lancet Infectious Diseases, Kasper Iversen and colleagues³ report results from a large seroprevalence survey of almost 30 000 hospital employees in Denmark.³ The authors found that 1163 (4.04%) of 28792 staff were seropositive overall, which was slightly higher than the 3.04% (142 of 4672) prevalence observed among local blood donors (risk ratio [RR] 1.33 [95% CI 1.12-1.58]). Seroprevalence was also higher among frontline health-care workers than among staff in other hospital roles (1.38 [1.22–1.56]; p<0.001). Staff working in dedicated COVID-19 wards showed substantially higher rates of seropositivity (1.65 [1.34–2.03]; p<0.001) than other frontline health-care workers working in hospitals, reflecting increased risk for this group, a pattern that has also been reported in neighbouring Sweden.⁴ Although lversen and colleagues used a point-of-care lateral flow immunoassay, which is generally considered less conclusive than enzyme-linked immunosorbent assays or similar laboratory-based methods,⁵ the authors did a comprehensive pre-study test assessment and estimated a sensitivity of $82\cdot5-90\cdot6\%$ and specificity of $99\cdot2-99\cdot5\%$. High specificity is essential to minimise high rates of false positives when used in low-prevalence populations, such as the one studied.

The results highlight the risk that SARS-CoV-2 can pose to health-care workers, particularly those in regular contact with patients with COVID-19, and the importance of understanding possible routes of exposure in hospitals. Given the potential for nosocomial transmission to amplify outbreaks,



Published Online August 3, 2020 https://doi.org/10.1016/ S1473-3099(20)30579-X See Articles page 1401