

Determining Transverse Rectus Abdominis Musculo-cutaneous Flap Viability using Fingerstall-type Tissue Oximetry as an Alternative to Indocyanine Green Fluorescence Imaging: A Case of a Patient with Iodine Hypersensitivity

Itaru Tsuge, MD Susumu Saito, MD, PhD

The blood circulation regions of Zone IV and Zone II in the transverse rectus abdominis musculo-cutaneous flap are known to depend on individual patients.¹ Conducting the indocyanine green (ICG) test helps ensure the safety of flap surgery; however, there have been reports of anaphylactic shock due to ICG intravenous injection.^{2,3} Even a single small-dose administration (5 mg) of intravenous ICG can cause anaphylactic shock,⁴ so its administration to patients with iodine hypersensitivity is contraindicated. We developed a noninvasive method using the fingerstall-type tissue oximetry system Toccare (Astem, Japan) to determine the skin perfusion area.⁵ Accumulating results indicated the average tissue oxygen saturation (StO₂) matching the ICG border to be 41.4%. We applied this result to a clinical patient with iodine hypersensitivity for the first time.

CASE

The medical ethics committee of our institution approved this study. A 47-year-old woman underwent right breast reconstruction by a pedicle transverse rectus abdominis musculo-cutaneous flap. She had asthma and atopic dermatitis in her medical history. An asthma attack occurred during a previous session of iodinated contrast computed tomography performed at another hospital. She was therefore diagnosed with iodine hypersensitivity. No subsequent imaging inspections using contrast media had been performed. She also reported food allergies for shrimp and crab. We judged the risk of an ICG test to be very high and therefore contraindicated. As such, the Toccare system was applied as an intraoperative noninvasive evaluation method. For intraoperative use, an echo probe cover was used. Blue fabric was laid under the flap to avoid any influence of underlying structures. We searched for points at which the StO₂ was 45%, 40%, or 30% after elevation of the flap to describe the boundary

lines (Fig. 1). Comprehensive StO₂ measurement of the entire flap and determination of the boundary lines took about 5 minutes. When the flap was cut with a 40% line, dark-red continuous bleeding was observed. After transferring the flap to the right chest, the skin was further




Fig. 1. Measurement of the StO₂ value. Lines for 45%, 40%, and 30% were described.




Fig. 2. Transfer of the flap to the right chest. The skin was trimmed on a 45% line. The fat was trimmed along a 40% line.

From the Department of Plastic and Reconstructive Surgery, Graduate School of Medicine, Kyoto University, Kyoto, Japan.

Received for publication June 16, 2018; accepted August 7, 2018.

Copyright © 2018 The Authors. Published by Wolters Kluwer Health, Inc. on behalf of The American Society of Plastic Surgeons. This is an open-access article distributed under the terms of the Creative Commons Attribution-Non Commercial-No Derivatives License 4.0 (CCBY-NC-ND), where it is permissible to download and share the work provided it is properly cited. The work cannot be changed in any way or used commercially without permission from the journal.

Plast Reconstr Surg Glob Open 2018;6:e1966; doi:10.1097/GOX.0000000000001966; Published online 24 September 2018.

trimmed to a 45% line and showed red bleeding (Fig. 2). The fat layer was excised along a 40% line, and dark-red, continuous bleeding was observed. The surgery resulted in no flap failure.

Anaphylactic shock has been reported even in cases with negative iodine test findings or no history of drug allergies.⁴ The ICG test should be considered as an invasive method, even for patients with no history of allergy. Our novel method can be expected to be widely applicable as a new alternative method for evaluating flap viability.

Itaru Tsuge, MD

Department of Plastic and Reconstructive Surgery
Graduate School of Medicine
Kyoto University
54 Shogoin kawahara-cho, sakyō-ku
Kyoto 606–8507, Japan
E-mail: itsuge@kuhp.kyoto-u.ac.jp

DISCLOSURE

The authors have no financial interest to declare in relation to the content of this article. The Article Processing Charge was paid for by the authors.

REFERENCES

1. Schefflan M, Dinner MI. The transverse abdominal island flap: part I. Indications, contraindications, results, and complications. *Ann Plast Surg.* 1983;10:24–35.
2. Hope-Ross M, Yannuzzi LA, Gragoudas ES, et al. Adverse reactions due to indocyanine green. *Ophthalmology.* 1994;101:529–533.
3. Timothy WO, Jennifer IL, Antonio CJ, et al. Anaphylactic shock following indocyanine green angiography. *Arch Ophthalmol.* 1996; 114: 97.
4. Chu W, Chennamsetty A, Toroussian R, et al. Anaphylactic shock after intravenous administration of indocyanine green during robotic partial nephrectomy. *Urol Case Rep.* 2017;12:37–38.
5. Tsuge I, Enoshiri T, Saito S, et al. A quick evaluation of TRAM flap viability using finger-stall-type tissue oximetry. *Plast Reconstr Surg Glob Open.* 2017;5:e1494.