RESEARCH ARTICLE



Challenges in measuring nitrogen isotope signatures in inorganic nitrogen forms: An interlaboratory comparison of three common measurement approaches

Christina Biasi ¹ Simo Jokinen ¹ Judith Prommer ² Per Ambus ³
Peter Dörsch ⁴ 💿 Longfei Yu ^{4,5} 💿 Steve Granger ⁶ 💿 Pascal Boeckx ⁷
Katja Van Nieuland ⁷ Nicolas Brüggemann ⁸ Holger Wissel ⁸
Andrey Voropaev ⁹ Tami Zilberman ¹⁰ Helena Jäntti ¹ Tatiana Trubnikova ¹
Nina Welti ^{1,11} 💿 Carolina Voigt ^{1,12} 💿 Beata Gebus-Czupyt ¹³ 💿
Zbigniew Czupyt ¹⁴ Wolfgang Wanek ²

Correspondence

C. Biasi, Department of Environmental and Biological Sciences, University of Eastern Finland, Yliopistonranta 1E, 70211, Kuopio, Finland. Email: christina.biasi@uef.fi

Funding information

European Association of National Metrology Institutes, Grant/Award Numbers: 16ENV06 SIRS, 19ENV05 STELLAR **Rationale:** Stable isotope approaches are increasingly applied to better understand the cycling of inorganic nitrogen (N_i) forms, key limiting nutrients in terrestrial and aquatic ecosystems. A systematic comparison of the accuracy and precision of the most commonly used methods to analyze $\delta^{15}N$ in NO₃⁻ and NH₄⁺ and interlaboratory comparison tests to evaluate the comparability of isotope results between laboratories are, however, still lacking.

Methods: Here, we conducted an interlaboratory comparison involving 10 European laboratories to compare different methods and laboratory performance to measure $\delta^{15}N$ in NO₃⁻ and NH₄⁺. The approaches tested were (a) microdiffusion (MD), (b) chemical conversion (CM), which transforms N_i to either N₂O (CM-N₂O) or N₂ (CM-N₂), and (c) the denitrifier (DN) methods.

Results: The study showed that standards in their single forms were reasonably replicated by the different methods and laboratories, with laboratories applying CM- N_2O performing superior for both NO_3^- and NH_4^+ , followed by DN. Laboratories using MD significantly underestimated the "true" values due to incomplete recovery and also those using CM- N_2 showed issues with isotope fractionation. Most methods and laboratories underestimated the at%¹⁵N of N_i of labeled standards in their single forms, but relative errors were within maximal 6% deviation from the real value and therefore acceptable. The results showed further that MD is strongly biased by nonspecificity. The results of the environmental samples were generally highly variable, with standard deviations (SD) of up to $\pm 8.4\%$ for NO_3^- and $\pm 32.9\%$ for NH_4^+ ; SDs within laboratories were found to be considerably lower

For affiliation refer to page 14.

This is an open access article under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made. © 2022 The Authors. *Rapid Communications in Mass Spectrometry* published by John Wiley & Sons Ltd. (on average 3.1‰). The variability could not be connected to any single factor but next to errors due to blank contamination, isotope normalization, and fractionation, and also matrix effects and analytical errors have to be considered.

Conclusions: The inconsistency among all methods and laboratories raises concern about reported δ^{15} N values particularly from environmental samples.

1 | INTRODUCTION

Nitrogen (N), a key limiting nutrient in terrestrial and aquatic ecosystems, is predominantly taken up by plants and phytoplankton in its inorganic forms, nitrate (NO_3^{-1}) and ammonium (NH_4^{+1}). It is very important in regulating ecosystem productivity and carbon (C) sequestration^{1,2} and plays a key role in the internal N cycling in natural and managed systems. To increase yields of cultured plants or organisms, management practices, including N fertilizer application, have been developed. However, the effects of increasing fertilizer application as well as increasing anthropogenic N deposition are farreaching:³ excess inorganic N (N_i) imposes serious environmental risks, by contaminating groundwaters,⁴ causing eutrophication of coastal areas,⁴ promoting soil acidification,⁵ and increasing emissions of reactive N-gases, such as nitrous oxide (N₂O), nitric oxide (NO), and ammonia (NH₃).^{6,7}

The negative effects of excess N on water and air quality and human health have intensified the research efforts to better understand and predict the role and fate of N_i in ecosystems in recent years.^{6,8–10} Traditional methods (e.g., wet chemical methods, spectroscopic methods) quantify concentration changes in NO₃⁻ and NH₄ over time to determine net nitrification and mineralization rates. However, these methods have substantial limitations as they do not assess process rates of these bioavailable N forms, and they do not allow to partition source pools of N species.¹¹ Therefore, stable isotope approaches based on the natural abundance (NA) or experimental isotopic enrichment of ¹⁵N (and ¹⁸O) in NO_3^- and NH_4^+ have become popular. They are increasingly applied to track gross transformations of N, to partition N cycling pathways, and to determine N sources and sinks.¹²⁻¹⁵ Inferences of rate constants and source contributions from both approaches rely on precise measurements of the isotopic composition ($\delta^{15}N$, at% ^{15}N) of N_i.

To date, three commonly used approaches exist to analyze the ¹⁵N isotopic composition of NO_3^- and NH_4^+ : (a) microdiffusion (MD), (b) chemical conversion (CM), and (c) biological conversion, that is, denitrifier (DN) methods (Figure S1 [supporting information]). In addition, there are protocols that combine parts of several of the afore-mentioned methods.¹⁶ In general, all methods are based on the selective separation and purification of N_i from the sample matrix. The MD method relies on a pH increase of the sample to convert NH_4^+ from solution into gaseous NH_3 that is collected on acidified filters. The isotopic signature of NH_4 -N is measured using elemental analyzer-isotope ratio mass spectrometry (EA-IRMS).¹⁷⁻²⁰ The MD

method further enables the analysis of NO₃, which must be reduced first to NH₄ using Devarda's alloy.²¹ In the CM methods, NH_4^+ or NO_3^- is chemically converted to N_2O or N_2 , which are thereafter analyzed for their isotopic composition using gas chromatography (GC)-IRMS, purge-and-trap (PT)-IRMS,^{16,22-24} or laser spectrometric analysis.²² Chemical approaches also comprise ion exchange resins for collection of NO3⁻ and NH4⁺, followed by precipitation as silver, potassium, or barium salts for NO_3^- or conversion to N_2 for $NO_3^$ and $NH_4^{+,23-26}$ In the DN method, NO_3^{-} is converted to N₂O by a denitrifying bacterial pure culture that lacks the N2O reductase enzyme,^{27,28} and the isotope composition of N₂O is then analyzed with the techniques mentioned earlier for the CM method. Generally, in the case of labeled samples, scanning mass spectrometers (e.g., guadrupole mass spectrometers²⁹ or membrane inlet mass spectrometers^{30,31}) can also be used for the analysis of isotope ratios in gaseous reaction products. In this interlaboratory comparison, however, only IRMS-based methods are compared, which are most commonly adopted.

Although the potential bias of each individual method has been examined,³²⁻³⁵ a systematic comparison of the accuracy and precision of the most commonly used IRMS-based methods is still lacking. Moreover, an interlaboratory comparison is lacking where all the possible sources of variation (from sample pretreatment to isotope analysis, including standardization procedures, instrumentation, and data processing) are considered and the validity of the results from laboratories using the same or different method, thus laboratory performance, is assessed. We therefore conducted an interlaboratory comparison on a set of standards and environmental samples containing NO₃⁻ and/or NH₄⁺. All materials were centrally prepared and sent to 10 laboratories, each of which employed at least one of the three recognized techniques for the analysis of isotope ratios $(^{15}N/^{14}N)$ in NO_3^- and/or NH_4^+ . The main aim was to evaluate the performance and practicability of the three techniques relative to one another and to investigate the intercomparability of the results from different laboratories. This knowledge is needed when designing and interpreting experiments that study the fate, origin, and transformation of N_i based on ¹⁵N in NO₃⁻ and NH₄⁺. Accurate results on $\delta^{15}N$ of N_i, which are intercomparable between laboratories, are also needed when laboratories are providing data to a common database, when data synthesis is conducted or when several laboratories are cooperating in joint field surveys. Based on the results, the strengths and weaknesses of each method are discussed, and guidelines are provided, which should aid in choosing the best methods for specific applications and research needs.

2 | EXPERIMENTAL

2.1 | Participating laboratories and organizations in the interlaboratory comparison

The participants in the interlaboratory comparison comprised 10 wellestablished laboratories associated with universities, nonuniversity research institutions, and commercial laboratories with mostly long experience in analyzing δ^{15} N of N_i. Each laboratory used its routine methods to analyze standards and environmental samples. Reported results were assigned lab codes (L1 to L16) to ensure confidentiality regarding the identity of the laboratories (on request, lab IDs and original data are available) and were grouped by method type (CM-N₂O, CM-N₂, MD, and DN; Tables S1A and S1B [supporting information]; Figure S1 [supporting information]). As some laboratories employed more than one method, the total number of lab codes is greater than the number of participants. The participants were given basic information on the origin of the environmental samples, including the approximate range of $\delta^{15}N$ values of the standards, to allow for proper referencing (Tables S2 and S3 [supporting information]), but neither actual δ^{15} N values nor results of other test participants were distributed among the participating laboratories. All standards and environmental samples were centrally prepared at the University of Vienna and dispatched to the participating laboratories immediately after preparation. A range of isotope standards with different $\delta^{15}N$ values was prepared (from -3.04 to 28.1‰), and, in addition to NA, labeled standards (LA) were distributed as well (Table S2 [supporting information]). Participants were asked to deep-freeze the received samples until analysis (mostly done within 6 months after receipt of samples) and to use their routine, in-house methods, and calibration/normalization procedures to provide the raw and final results, along with a clear description of the method used. Each laboratory carried out preparatory steps to isolate the N forms according to their routine methods and analyzed the standards/samples using IRMS (Figure S1 [supporting information]). Therefore, this interlaboratory comparison examines the variability of the whole method chain, from the isolation to the analysis of $\delta^{15}N$ of N_i forms, to the instrumentation used, standardization methods applied, laboratory conditions prevailing, experience level of operators, and the data processing implemented. Statistical analysis, evaluation of the data, and further corrections and calculations (e.g., recovery rates) were carried out either by the individual laboratories or by the main author.

2.2 | Preparation of standards and environmental samples

2.2.1 | Standards

Ten stock solutions of pure standards (S) containing KNO₃ (N) and NH₄Cl (A), respectively, resulting in five NO₃⁻ and five NH₄⁺ standards, were prepared (codes as follows for NO₃⁻: S1N, S2N, S3N,

S4N, S5N; and for NH_4^+ : S1A, S2A, S3A, S4A, S5A). The standards S1–S3 were prepared to cover the natural abundance range of $\delta^{15}N$ of N_i species (NA standards), ranging from -3.04 to 28.1‰ (Table S2 [supporting information]). Standards S4 and S5 were labeled with ¹⁵N (LA) to reach approximately 0.71 and 2.1 at%¹⁵N (Table S2 [supporting information]). All LAs were prepared by mixing KNO₃ and NH₄Cl enriched with ¹⁵N at 98 at% with the same salts at natural ¹⁵N abundance level, respectively, to achieve the isotope ranges. Standards were dissolved in deionized water to achieve a stock concentration of the analyte of 1M in N. Participants were asked to dilute the standards to the concentrations needed by their respective methods and to analyze them in four replicates (n = 4). The "true" δ^{15} N values of NO₃⁻ and NH₄⁺ salts (chemical purity > 99.9%) were derived from analyzing the standards using EA-IRMS by two to three different laboratories (n = 2-3; see below). Participants were asked to include blanks (e.g., the solvent used to dilute the standards) in the analysis and to report recovery rates of the N species.

In addition to the single compound standards, mixed standards were prepared to test for nonspecificity or possible crosscontamination of the target compound by other N forms (S1Nx, S1Ax). The mixed standards contained the target compound at NA level (here, S1) with the addition of the other N forms (NO₃⁻, NH₄⁺, amino acid mix) enriched with ¹⁵N (at 0.71 at% ¹⁵N) (Table S2 [supporting information]). The concentrations of all three N forms were equimolar, that is, at 0.33M in N. All standard solutions were dispatched in 2 ml aliquots.

2.2.2 | Environmental samples

In addition to the standards, five different environmental samples (three soil extracts and two water samples) were collected in March 2012 near Vienna, Austria, to test for consistency of δ^{15} N values of N_i species in real-world samples across the methods (Table S3 [supporting information]). In short, 3 kg of soil was sampled from surface layers (0-20 cm) in two forests and one grassland ecosystem (P1S, P2S, HWS). Soils were sieved through a 2 mm mesh and extracted with 0.5M K₂SO₄ in the lab (1:5 soil:extractant solution [w: v]). Water samples were collected from two adjacent sites, that is, from a floodplain lake and the Danube River channel (DL, DR). The samples were filtered first with a microcloth filter (Calbiochem) and then through glass fiber filters (Whatman GF/F, cat no 1825-047) applying a gentle vacuum. From each site, four replicates were taken, which were then pooled. The concentrations of NO_3^- , NH_4^+ and DON were determined in all environmental samples by routine procedures as detailed by Hood-Novotny et al (2010) (Figure S3 [supporting information]). Accordingly, the NO₃⁻ concentrations ranged between 24 and 212 μ mol L⁻¹ (with one lake water sample having no NO₃⁻), the NH₄⁺ concentration ranged between 2 and 38 µmol L⁻¹, and DON concentrations ranged between 13 and 142 μ mol L⁻¹ (with one river water sample having no DON) (Table S3 [supporting information]). The concentrations of nitrite (NO_2^{-}) were below detection limit in all samples (<0.25 μ mol L⁻¹; data not shown);

4 of 16

II FY-

thus, no measures were taken to remove NO₂⁻. To inhibit microbial growth, both soil and water samples were autoclaved with loosened stoppers at 121°C for 1 h. All sample bottles were stored at -20°C until dispatched to the participants. All participants were asked to freeze the environmental samples until analysis and to analyze them in four replicates (n = 4).

2.3 | Nitrogen isotope analysis

In the following, we introduce briefly the different methods and also variations between individual protocols implemented here. We divide them and treat them as different methods (CM–further divided into CM-N₂O and CM-N₂—MD, and DN methods), although we are aware that we are in fact dealing with different method groups and that the application of individual methods can vary quite significantly. Some methods consist also of a combination of methods (e.g., the analysis of NH₄⁺ with DN and CM-N₂O includes an MD step followed by alkaline persulfate oxidation, APO), as can be seen in Figure S1 (supporting information).

2.3.1 | CM methods

CM methods transform NO_3^- and NH_4^+ to either N₂O (CM-N₂O) or N₂ (CM-N₂) before isotope analysis. The CM methods applied in this interlaboratory comparison differed significantly between the participating laboratories (Table S4 [supporting information]). Briefly, the CM-N₂O approach utilizes VCl₃ (L1)^{16,36} or cadmium (L2, L3)^{36,37} to reduce NO_3^- to nitrite (NO_2^-), which is finally converted to $N_2O_2^$ by azide (L1, L2)^{16,36} or hydroxylamine (L3).^{38,39} If azide is used, one extraneous N is introduced into the analyte, and if the less toxic hydroxylamine is used, the conversion is not quantitative (Table S4 [supporting information]). For isotope analysis of NH_4^+ , the N_i species is either oxidized to NO_3^- by persulfate digestion and further reacted to N₂O as described earlier (L1),¹⁶ or sequentially transformed via NO_2^- to N_2O by hypobromite (BrO⁻) with and without azide (L2, $L3^{40,41}$). In the absence of azide (L3), N₂O is a side product, and the methods can be applied only to labeled samples (L3). Ammonia is best isolated before the conversion to N₂O, for example, via MD (L3; not done by L2). With the CM-N₂ approaches, NH_4^+ is converted to N₂ by BrO⁻ after isolating the N species with MgO and acid trap (L4).⁴² Nitrate is converted to N₂ by pyrolysis, or through thermal conversion (e.g., TC/EA), after isolation from other N-bearing species by ion exchange and/or precipitation (L4, L5).^{26,43} The CM-N₂O approach utilizes continuous flow PT-IRMS techniques, whereas CM-N2 methods can also employ dual-inlet IRMS systems or TC/EA-IRMS techniques (Table S4 [supporting information]). There are basic differences in sample size requirements, which limit quantification and precision. Sample volumes ranging from 2.5 to 40 mL and 230 to 1000 mL were used for CM-N₂O and CM-N₂, respectively (Table S4 [supporting information]). Generally, CM-N₂O methods are more suitable for low-concentration ¹⁵N analysis (in the nmol

range). Average optimal N concentrations as used by the laboratories were 0.312 and 150 μ mol N for CM-N₂O and CM-N₂, respectively (NA analysis of ¹⁵N) (Table 1). In general, about 50 samples can be processed for both NO₃⁻ and NH₄⁺ per week (5 days) with CM-N₂O, and about 30 samples with CM-N₂ (supporting information), though the turnover varies considerably between methods applied.

2.3.2 | MD method

The MD method is based on outgassing dissolved NH₄⁺ as NH₃ under alkaline conditions and subsequent conversion to NH₄⁺ salt under acidic conditions. MD was introduced decades ago¹⁸⁻²⁰ and is still one of the most widely used procedures to analyze ¹⁵N abundance in N_i. Most laboratories that employed the MD method in this interlaboratory comparison (L6-L9) used one of the original methods, with variable modifications (Table S5 [supporting information]).^{16,19} In brief, 10-100 mL of sample was diffused for 3-5 days (L6-L8) or 20 days (L9) after the addition of MgO (L6: 100 mg; L7-L8: 200 mg; L9: 600 mg), and NH₃ was trapped on acid traps (disks of filter paper or glass fiber filter amended with KHSO₄, H₂SO₄, or oxalic acid) wrapped in semipermeable Teflon membrane floating on the solution. For placement within the sample, disks can also be generally fixed to stainless steel wires as, for example, in one of the original methods.¹⁸ After NH_4^+ was guantitatively removed by MD, NO₃⁻ was analyzed by the same method after reduction of NO_3^- to NH_4^+ with Devarda's alloy (L6: 50 mg; 400 mg: L7-L9).²¹ MD samples are analyzed by continuous flow EA-IRMS, that is, combustion without using PT technology, and therefore, the MD method requires a relatively higher sample amount (in the umol range). The average optimal N concentration as used by the laboratories was 7 µmol N for MD (Table 1). The throughput is dependent on the incubation time, and 70 and 100 samples can be processed for $\delta^{15}N$ analysis of NH_4^+ and NO_3^- isotope analysis, respectively, within a week (with a maximum of 100 and 150 samples possible with 3 days' incubation time).

2.3.3 | Biological conversion by the DN method

In the DN method, NO_3^- is selectively converted to N_2O by denitrifying bacterial strains of *Pseudomonas aureofaciens* or *Pseudomonas chlororaphis*, both of which are lacking nosZ, the gene coding for N_2O reductase, which would further reduce N_2O to N_2 .^{27,28} Denitrifying enzymes have high substrate specificity and affinities (K_m values) in the nM range; thus, NO_3^- is quantitatively converted to N_2O , which can then be analyzed using PT-IRMS. Laboratories L10, L11, and L13 used a *P. aureofaciens* strain, and L12 used a *P. chlororaphis* strain for this reaction (Table S6 [supporting information]). All laboratories followed the original protocol by Sigman et al,¹ where the method was first presented, with small adjustments, except laboratory L11, which had adopted



TABLE 1 Overview of characteristics, requirements, and basic performance of the various methods used in this interlaboratory comparison to analyze δ^{15} N signatures of inorganic Nitrogen (N) forms

Method (abbreviation)	Sample volume (mL)	Optimum N concentration (µM)	Optimum amount of N (μmol)	Recovery (%)ª	Recovery (%) ^b	Blank (%)	Blank (μM)	Blank (µmol)	Participating laboratories
Chemical method—N ₂ O (CM-N ₂ O)	9.25 (40 for L3)	32 (300 for L3)	0.312 (12 for L3)	97.0 ± 4.51 (25% for L3)	103 ± 6.0 (n.d. by L3)	9.20 (0.00 for L3)	2.94 (0.00 for L3)	0.029 (0.00 for L3)	3
Chemical method—N ₂ (CM-N ₂)	1810	83	150	Not reported	Not reported	0.00	0.00	0.00	2
Microdiffusion (MD)	42	170	7.00	88.2 ± 3.05	115 ± 8.5	4.30	7.31	0.301	4
Denitrification (DN)	2	50	0.100	101 ± 8.1	106 ± 8.6	4.1	2.05	0.004	4

Note: The average of all reported values is presented method-wise. L3 used a conventional CM-N₂O method where N₂O is a by-product of the reaction between hydroxylamine and NO_2^- and is thus presented separately.

^aAs determined from single standards at natural abundance (S1-S3).

^bAs determined from mixed standards (S1Nx, S1Ax) where target compound was at natural abundance and nontarget compounds were labeled. The average recovery is shown.

the protocol of Morkved et al⁴⁴ for *P. aureofaciens* (more details in supporting information). Ammonium was measured by the DN approach after MD of NH₄⁺ and persulfate digestion, which oxidizes all NH₄⁺ to NO₃⁻ (L13). The optimum amount of N used by the laboratories was with an average of 0.1 µmol N, lowest for DN. The throughput is similar to CM-N₂O: about 45 samples can be processed per week for δ^{15} N analysis of NO₃⁻, and 4 days at a minimum in the case of NH₄⁺ analysis by MD and APO preprocessing. However, bacterial cultures need to be prepared and maintained, which makes this method more labor-intensive compared to the other tested ones.

2.3.4 | Freeze-drying and evaporation followed by EA-IRMS analysis

The relatively simple freeze-drying and evaporation technique (EA) was carried out by L14, L15, and L16. With this technique, an appropriate amount of the standard was freeze-dried, resuspended with distilled water (various amounts; $50-170 \mu$ L), pipetted into tin cups, and, subsequently, placed in an oven at 50° C or under a fume hood at room temperature until dry. The resulting NO₃⁻ and NH₄⁺ salts were analyzed using EA-IRMS. This technique is suitable for pure dissolved target N species only, as it was the case for standards S1-S5 and cannot be applied to natural environmental samples.

2.4 | Data processing and statistics

We calculated $\Delta \delta^{15} N$ values as the difference between raw $\delta^{15} N$ values obtained by each laboratory and the reference ("true") value obtained by the EA method for each standard. The EA results served as a reference in this study, as freeze-drying and/or evaporation followed by EA-IRMS are the most conventional and direct methods of isotope analysis (without any isolation steps). It should be noted that this EA approach does not provide a certified reference value, but one that is sufficiently accurate here to compare the relative precision and accuracy of the other methods. More details on how the reference value was established are given in the supporting information. Accuracy is given as the difference of δ^{15} N values of NA standards from true $\delta^{15}N$ values as $\Delta\delta^{15}N$ and as $\Delta at\%^{15}N$ for LA standards. To better represent the variability in the data set, we report results as mean with SD as well as median values for each method, with upper and lower quartiles (interguartile range: 25th to 75th percentiles expressing reproducibility) in Figure 1. Precision is given as the mean SD (± relative SD) of results received from each laboratory and expresses thus within-laboratory or intralaboratory precision. Interlaboratory precision equals the SD across the means of each method group.

Nonspecificity of each method was calculated using S1Nx and S1Ax results, assuming that a complete conversion of both nontarget N species would result in a 100% contamination, and was computed using the following equation:

Nonspecificity = $\frac{at\%^{15}N_m - at\%^{15}N_T}{(f_T * at\%^{15}N_T + f_{NT1} * at\%^{15}N_{NT1} + f_{NT2} * at\%^{15}N_{NT2}) - at\%^{15}N_T}$

		Method								
Accuracy		CM-N2O		CM-N2		MD	DN			
SN-NA	mean	0.05 [25]	n.s.	1.75 [26]	*	-2.14 ‰ [46]	*	-0.26 ‰ [44]	n.	
	stdev	(0.50)		(1.36)		(2.18)		(2.25)		
	median	0.14 ‰ [25]		2.20 ‰ [26]		-3.06 ‰ [46]		0.73 ‰ [44]		
	interquartile range	(-0.37 to 0.55)		(1.30 to 2.61)		(-3.87 to -0.16)		(-0.59 to1.12)		
SA-NA	mean	-0.54 ‰ [25] (1 13)	n.s.	-2.98 ‰ [12] (2.01)	*	-3.61 ‰ [56]	*	-0.19 ‰ [12]	n.	
	median	-0.73 ‰ [25]		-4 00 % [12]		-3 31 % [56]		-0.21 % [12]		
	interquartile range	(-1.39 to 0.275)		(-4.37 to -0.573)		(-4.14 to -2.00)		(-1.45 to 1.94)		
SN-LA	mean	-0.01 ‰ [17]	n.s.	0.04‰ [11]	*	-0.04 ‰ [40]	*	-0.13 ‰ [10]		
	stdev	(0.06)		(0.05)		(0.05)		(0.15)		
	median	-0.02 at% [17]		0.03 at% [11]		-0.02 at% [40]		-0.08 at% [10]		
	interquartile range	(-0.05 to 0.01)		(0.02-0.09)		(-0.07 to -0.01)		(-0.27 to 0.01)		
SA-LA	mean	0.03 ‰ [16]	*	-0.02‰ [2]	*	-0.03 ‰ [39]	*	not analyzed		
	median	0.04 at% [16]		-0.02 at% [2]		-0.02 at% [39]		not analyzed		
	interquartile range	(-0.00 to 0.06)		(-0.03 to -0.01)		(-0.06 to 0.00)		not analyzed		
Precision	inter quartine range	(,		(,		(,		,		
SN-NA	mean (± STDEV)	0.23 ‰ ± 0.13		0.13 ‰ ± 0.02		0.77 ‰ ± 1.14		0.27 ‰ ± 0.40		
SA-NA	mean (± STDEV)	0.47 ‰ ± 0.21		0.37 ‰ ± 0.11		0.51 ‰ ± 0.40		2.99 ‰ ± 3.03		
SN-LA	mean (± STDEV)	0.01 at% ± 0.02		0.00 at% ± 0.00		0.01 at% ± 0.01		0.03 at% ± 0.02		
SA-LA	mean (± STDEV)	0.01 at% ± 0.03		0.01 at% ± 0.01		0.01 at% ± 0.02		not analyzed		
Non-specificity								-		
S1Nx	Carry-over (in %)	0.15 % ± 0.26 [8]	n.s.	not analvzed		28.9 % ± 25 [27]	*	0.24 % ± 0.19 [9]	n.	

S1Ax

6 of 16 WILEY-

FIGURE 1 A heatmap illustrating the accuracy, precision, and specificity of the methods used by variable laboratories in this interlaboratory comparison to analyze δ^{15} N in inorganic nitrogen (N) forms. Methods tested were the chemical method transforming inorganic N forms to N₂O (CM-N₂O), the chemical method transforming inorganic N forms to N₂ (CM-N₂), the microdiffusion (MD) method, and the bacterial denitrifier (DN) method. Results are presented for natural abundance standards of NO₃⁻ (NA-SN) and NH₄⁺ (NA-SA) and for labeled standards of NO₃⁻ (LA-SN) and NH₄⁺ (LA-SA). Color codes were normalized to the maximum value found for each category, where the dark green color (closer to 0) indicates highest accuracy, precision, and lowest nonspecificity (demonstrating better performance of the method utilized by one lab) and the red color (closer to 1) indicates lowest accuracy, precision, and lowest nonspecificity. Yellow colors indicate intermediate values. Accuracy is given as the difference of δ^{15} N value of NA standard from true δ^{15} N value as $\Delta\delta^{15}$ N [n] and as at% offset for LA standards. Data are mean and median values and standard deviation (SD) including lower (25%) and upper (75%) quartile ranges. Precision is given as the mean SD (± relative SD; in ‰ for NA and in at% ¹⁵N for LA) of results received from each laboratory (intralaboratory precision). Nonspecificity was calculated assuming that a conversion of nontarget N species would result in 100% contamination (more details in the main text). The $\Delta\delta^{15}$ N results of S1, S2, and S3 (NA) and S4 and S5 (LA), which were characterized by different isotope signatures (NA) or isotopic enrichment (LA), respectively, were pooled. Asterisks indicate significant differences at P < 0.05 [Color figure can be viewed at wileyonlinelibrary.com]

Carry-over (in %) 10.9 % ± 13.7 [7] * 3.39 % ± 0.9 [4] * 17. % ± 14.6 [28]

0

where $at\%^{15}N$ is atom-percent ${}^{15}N$, m = measured, T = target compound (NA values of NO_3^- and NH_4^+ in S1Nx and S1Ax, respectively), NT = nontarget compounds (0.71 at% ¹⁵N; combined with other N containing sources), and f represents the added fractions of T and NT (1/3 each).

All statistical analyses were carried out using the IBM SPSS 25 statistical software. For standards, significance of differences between individual lab results and results grouped into the different methods, respectively, and the "true" values based on the EA method were tested using two-sample Student's t-tests (independent sample's t-test). Differences were assumed significant when P < 0.05. For environmental samples, all individual lab results and means of each method were compared with the mean value calculated from the results of all methods used. Deviations from this mean were

evaluated using Student's t-tests as mentioned earlier. Before the use of any statistical method, the data were checked for normality (Kolmogorov-Smirnov's test) and for homogeneity of variances (Levene's test). Further details on statistical analysis are provided in the supporting information.

*

1

9.69 % ± 2.5 [4]

However, it must be noted that this interlaboratory comparison was not designed as a proficiency test, because participants were not judged for the acceptability and traceability of their results, and there was no reference laboratory or standard operating procedures recommended. Outliers (intralaboratory extreme values) were, however, identified based on reported laboratory batch means.45 Participants were asked to perform their routine analysis. Nevertheless, the procedures used by each laboratory were considered appropriate to the objective of the study. Rather, the purpose of the interlaboratory

comparison was to determine whether the most commonly used methods to analyze $\delta^{15}N$ in N_i produce acceptable, concordant, and intercomparable results when performed by different laboratories. As a result, the study has a more descriptive character, and we report and discuss discrepancies in the data from each group of methods rather than following strict metrological protocols.

3 | RESULTS AND DISCUSSION

3.1 | Natural ¹⁵N abundance standards

In the absence of other confounding N_i or organic N species, repeated analysis of NO_3^- and NH_4^+ salts yielded information about the accuracy and precision of isotope ratio measurements for each method while, at the same time, allowing evaluation of the performance of each participating laboratory. For NA nitrate standards (SN-NA), the mean (\pm SD) and median $\Delta \delta^{15}$ N values reported from laboratories using the CM-N₂O technique were $+0.055 \pm 0.50\%$ and +0.138%, respectively, and the results did not differ significantly from the reference method (P = 0.501) (Figure 1; Figure S1 [supporting information]), with no significant differences found between SN1, SN2, and SN3 (Figure S2 [supporting information]). Each individual SN-NA standard measured by the contributing laboratories applying the CM-N₂O method in quadruplicates also exhibited smallest deviations in the actual $\delta^{15}N$ value (always less than 1‰), as compared to the other methods (Figure S4A [supporting information]). Regarding precision, laboratories using the CM-N₂O method reported with 0.229‰ overall intermediate intralaboratory precision for SN-NA (Figure 1: Figure S5 [supporting information]). Laboratories using MD reported the lowest overall precision (Figure 1; Figure S5 [supporting information]) and the lowest accuracy at natural abundance levels; the mean $\Delta \delta^{15}$ N of $-2.14 \pm 2.18\%$ (median -3.06%) was significantly different from the reference method (P = 0.001). The reported accuracy from laboratories applying both DN and CM-N₂ laid between that of CM-N₂O and MD for SN-NA, with mean $\Delta \delta^{15}$ N values of $-0.26 \pm 2.25\%$ for DN and $+1.75 \pm 1.36\%$ for CM-N₂ (median: +0.73‰ and +2.20‰, respectively). The mean $\Delta\delta^{15}$ N value was statistically different for laboratories applying CM-N₂ methods compared to the reference method (P = 0.001) but not for those applying DN (P = 0.477); however, variability for DN was high (Figure 1; Figure S1 [supporting information]). Overall precision was highest for laboratories using CM-N₂ and intermediate for those using DN when analyzing SN-NA (Figure 1; Figure S5 [supporting information]). The precision was for all methods (on average 0.7‰) within the lower range of published values¹⁶ but considerably higher (about 3 times) than interlaboratory precision (Figure 1; on average 1.94‰), indicating that the spread in mean values was largely due to differences in results reported from different laboratories (which was the case for all results reported; see below).

Analysis of NH₄⁺ standards at NA (SA-NA) gave similar results, though differences to the actual values were on average larger Rapid Communications in WILEY 7 of 16 Mass Spectrometry

(Figure 1; Figures S2-S4 [supporting information]), indicating lower accuracy of SA than SN methods with the techniques compared here from variable labs. The accuracy was best among laboratories applying the CM-N₂O methods, followed by those applying DN, and least accurate results were obtained by the laboratories applying MD and CM-N₂ methods (Figure 1; Figures S2 and S3 [supporting information]). Similar to SN-NA, the results reported from laboratories using the CM-N₂O and DN methods were on average not statistically different from the actual δ^{15} N value for SA-NA (P = 0.136 and 0.904, respectively) (Figure 1; Figure S1 [supporting information]). Overall precision was not as good for SA-NA as for SN-NA, with laboratories using DN methods showing lower precision than other methods (Figure 1; Figure S5 [supporting information]); however, only one laboratory employed the DN method for NH_4^+ and used MD in combination with persulfate digestion to collect and oxidize NH4⁺ to NO₃⁻ before DN measurements. Thus, no general conclusion can be drawn on the accuracy and precision of DN for $\delta^{15}N$ analysis of NH_4^+ here.

Laboratories reported recovery rates of >95% (Table 1) for CM-N₂O for both SN and SA, and thus, isotope fractionation during sample preparation was of minor importance,¹⁵ most likely explaining the high accuracy of CM-N₂O methods for pure standards.^{16,37,46} Blanks were significant with the CM-N₂O methods employed by the participating laboratories (on average 9.20% of sample peak or 0.03 µmol for both NO₃⁻ and NH₄⁺ analysis in this ring test; Table 1), though they were higher here than reported elsewhere for the same methods.^{16,36,37,47} However, it seems that proper blank correction could be achieved. The high blank could explain the intermediate precision found for this method by the participating laboratories.

The laboratories utilizing the DN methods also vielded satisfactory results analyzing standards with acceptable intralaboratory precision (at least for SN-NA), though the variability of $\Delta \delta^{15}$ N values was relatively high for this method, with both over- and underestimation of the actual $\delta^{15}N$ value (Figure S4A [supporting information]). Underestimation of $\delta^{15}N$ values can result from isotopic fractionation associated with incomplete conversion.33,48 Recovery rates were quite variable, but on average close to 100% (101% ± 8.1%; Table 1), and not related to the δ^{15} N value deviation found (data not shown), suggesting that isotope fractionation was of minor importance. Previous studies have shown that the conversion of NO₃⁻ to N₂O by DN is quantitative and complete under optimal conditions (e.g., Jantti et al³³ and Casciotti et al⁴⁸) and that results are generally robust.^{25,28,44,49,50} Under- and overestimation of isotope values by DN methods can further result from isotope normalization or blank issues. Because all laboratories applying DN methods also utilized certified standards, not the reference N₂O tank, as a means of calibration of isotope results, along with the samples (similar to CM-N₂O above), normalization issues are less likely. The systematic shift in $\Delta \delta^{15}$ N between S1, S2, and S3, from overestimation in S1 and S2 to underestimation in S3 (Figures S1 and S2 [supporting information]; at least for NO₃⁻ analysis; NH₄⁺ results cannot be properly evaluated due to the scarce dataset), suggests a two-source mixing behavior and a significant impact of the blank. DN methods have manifold blank

ILEY-

issues, including residual NO₃⁻ in the culture medium, extraneous NO₃⁻ used during preculturing not being removed completely, N₂O from denitrification of extraneous NO3⁻ dissolved in the medium or in the bacteria not being removed completely by Helium sparging, and contamination with atmospheric N₂O.^{28,33,50} Blanks were relatively low from the reported values for DN (on average 4.1% of the sample peak, i.e., 4 nmol in this ring test; Table 1). Nevertheless, it seems that proper blank correction would be necessary. Most laboratories utilizing DN, including the ones participating here, conduct the blank correction together with the calibration of the samples utilizing inhouse nitrate reference standards. This is not the most accurate way for a blank correction, because first, the sample-to-blank ratio needs to be kept constant, and second, the blank would need to be precisely guantified for the whole DN procedure.⁵¹ One laboratory also reported contamination problems with some DN analysis (data not shown). It has to be kept in mind that the number of labs employing DN was relatively small in this study.

Although laboratories applying the CM-N₂O and DN results reported on average not significantly different results to the actual δ^{15} N values, laboratories applying MD underestimated the actual δ^{15} N value significantly for both SN and SA by several ‰ (Figure 1; Figure S1 [supporting information]). This is most likely due to the relatively low recovery for this method (88.2% ± 3.05) and the associated isotope fractionation effects with incomplete recovery.^{15,52} MD of NH₃ by volatilization is a slow process that is susceptible to strong isotopic effects (20-30‰⁵³). Indeed, issues with low recoveries for MD have been reported previously,^{20,33,54} especially if the MD fails (e.g., acid traps sticking to the container walls,¹⁹ incubation bottles or acid traps leaking,²⁰ inappropriate headspace/sample ratio⁵²), which causes isotope fractionation resulting in more negative δ^{15} N values.¹⁵ For MD, it is important to include standards to correct for isotope fractionation due to incomplete N_i recovery.

The laboratories using the CM-N₂ methods over- and underestimated the δ^{15} N values for NO₃⁻ and NH₄⁺, respectively. For NO₃⁻, both laboratories used the ion exchange method to collect NO₃⁻, a technique prone to preferential retention/elution of ¹⁵NO₃,²⁶ likely explaining the overestimation of the δ^{15} N values (though this was true only for L5). For NH₄⁺, the traditional steam distillation method was used by L4 before oxidation to N₂ by LiOBr; the method also includes an acid trapping step of NH₄⁺. Thus, there is a potential bias due to isotope fractionation or incomplete recovery, NH₃ distillation, or loss of N during collection,⁵⁵ likely causing the underestimation of the δ^{15} N values. Generally, the low number of participating laboratories makes it difficult to evaluate CM-N₂ methods as one group.

3.2 | Labeled standards

Similar to NA standards, laboratories using the CM- N_2O methods reported the most accurate results when measuring the LA, but only for the SN-LA standards, where the only nonsignificant difference to the actual value was found (P = 0.605) (mean $\Delta at\%^{15}N = -0.012\%$, median $\Delta at\%^{15}N = -0.019\%$). The $\Delta at\%^{15}N$ values were significantly different from those of laboratories using MD (P = 0.001) (mean Δat % $^{15}N = -0.042\%$, median $\Delta at\%^{15}N = -0.021\%$), CM-N $_2$ (P = 0.002) (mean $\Delta at\%^{15}N = +0.041\%$, median $\Delta at\%^{15}N = +0.025\%$). and DN $\Delta at\%^{15} N = -0.131\%$ = 0.002) (mean median (P ∆at $\%^{15}N = -0.073\%$). For individual lab results of each SN-LA standard, most results were statistically different from the actual value, with the exception of two laboratories using MD (25% of the labs participating with MD) (Figure S4B [supporting information]). Intralaboratory precision was best for laboratories applying CM-N₂, followed by MD, CM-N₂O, and DN (Figure S5 [supporting information]). When SA-LA standards were analyzed, all laboratories vielded at%¹⁵N values. which were significantly different from the actual value with their methods used (Figure 1; Figures S1-S3 and S4B [supporting information]), but the mean differences were smallest for CM-N₂ (mean $\Delta at\%^{15}N = -0.018\%$, median $\Delta at\%^{15}N = -0.018$ at%¹⁵N), followed by MD (mean $\Delta at\%^{15}N = -0.026\%$, median Δat $^{15}N = 0.016 \text{ at}\%^{15}N$) and CM-N₂O (mean $\Delta at\%^{15}N = +0.033\%$, median $\Delta at\%^{15}N = +0.035 at\%^{15}N$). Contrary to NA standards, laboratories using MD methods yielded one of the best results at least for SA-LA.

As fractionation effects can be largely neglected in labeled samples, MD methods, which frequently show recovery issues, performed better with LA as compared to NA standards in this interlaboratory comparison. The insensitivity of ¹⁵N analysis to recovery in MD with ¹⁵N-enriched samples was also demonstrated by others.²⁰ An underestimation of the actual δ^{15} N value as found with most methods can be explained by the fact that IRMS is not designed to measure relatively highly labeled samples, as isotope fractionation can occur in the IRMS that does not scale linearly with isotopic (so-called scale compression/expansion).⁵⁶ An composition overestimation of isotope values in ¹⁵N-labeled samples, as observed here from laboratories using CM-N₂O with SA-LA standards and with CM-N₂ (L5) with SN-LA, can only be explained by problems with blank correction, with isotope calibration (CM-N₂O),^{20,26} or with memory effects.⁵⁷ It has to be noted that though the offsets of the LA standards were large, they are of less concern in ¹⁵N tracer studies, where NO_3^- and NH_4^+ are enriched at several at% ¹⁵N. Here, the isotopic offsets in the at% scale ranged between 2 and 6% relative error at 0.7 at%¹⁵N and between <1 and 3% relative error at 2.1 at %¹⁵N and were thus in an acceptable range for all methods employed.

3.3 Mixed standards and nonspecificity

For S1Nx, laboratories applying the CM-N₂O and DN methods showed no carryover and reported accurate results for the target compound (Figure 1). Production of N₂O from NO₃⁻ by denitrifiers is highly specific and does not pose risks for cross-contamination.^{25,28,33,50} Our comparison demonstrates that also the CM-N₂O methods based on the reduction of NO₃⁻ by VCl₃ or Cd via NO₂⁻ to N₂O are unbiased by interference with organic and

other N_i compounds. Only laboratories using MD showed significant cross-contamination and yielded, thus, unsatisfactory results when analyzing S1Nx standards (Figure 1). Three out of four labs obtained results, which were significantly enriched in ¹⁵N and were thus impacted by N from nontarget compounds (Figure S6 [supporting information]). Average nonspecificity was $28.9 \pm 24\%$, with up to 52% of nontarget compounds being carried over (L9). Thus, a variable but potentially large amount (up to half) of non-NO₃⁻, that is, NH₄⁺ or amino acid-N, ended up in the acid traps during MD.

Laboratories applying MD also showed issues with nonspecificity for S1Ax, though the carryover was lower (on average $18.0 \pm 15\%$) compared to S1Nx (Figure 1; Figure S6 [supporting information]). The contamination is of different nature for NH_4^+ as compared to NO_3^- , because both AA and NH_4^+ can end up in the "NO₃⁻ pool" after being reduced to NH_4^+ , whereas it is likely that only AA end up in the NH_4^+ pool, causing lower carryover issues. This is because NO_3^- is relatively stable in solution, whereas AA is easily broken down.⁵⁸ Generally, the high nonspecificity was corroborated by higher yields of S1x standards analyzed by MD as opposed to CM-N₂O and DN (Table 1). The reasons for high nonspecificity by MD can be manifold. First, the addition of MgO to raise the pH to >10 can cause carryover due to alkaline hydrolysis of DON and deamination of amino acids and thereby contamination of the NH4^{+,15,20,33,41,59} Second, the Devarda reaction can cause reductive breakdown of organics or other O-containing substances.^{35,41} This is likely due to the harsh conditions generated by the Devarda's alloy and the high pH. In addition, NH_4^+ needs to be properly removed before isotope analysis of NO₃⁻ via MD.²⁰ Using an optimized method to isolate NH₄⁺ by MD resulted in very low carryover by L1. a lab that otherwise used CM-N₂O to further process the trapped NH₄⁺ by alkaline persulfate oxidation. Correlation analysis did not indicate a significant relationship between the magnitude of nonspecificity and any methodological factor, though we noticed that higher amounts of MgO and longer incubation times resulted in increased carryover, as suggested also by others (e.g., Ros et al³⁵). It has to be noted that carryover due to breakdown of DON or NH_4^+ carryover by MD is an issue for both NA and LA samples. In a method study by Jantti et al³³ where labeled samples were investigated to test proper methods for studying the ¹⁵N isotope pool dilution method to measure gross nitrification rates, MD produced significantly lower at%¹⁵N values of NO₃ as compared to DN. The underestimation of at%¹⁵N values resulted in a significant (up to 10-fold) overestimation of gross nitrification rates by MD.

In addition to MD, also one out of two laboratories using CM-N₂O methods showed high nonspecificity for S1Ax (L2; 21%) (Figure S6 [supporting information]). This method is based on the reaction of NH_4^+ with hypobromite to form NO_2^- and later N₂O and does not—contrary to the method used by L1—isolate NH_4^+ before conversion to N₂O. Other organic compounds broken down by this reaction raise concerns on the use of this method,⁴⁶ an issue that could be solved by pre-isolating NH_4^+ before oxidation.²² L13 measured $\delta^{15}N$ in NH_4^+ by DN after isolating NH_4^+ by MD,

followed by persulfate oxidation to NO₃⁻, a method that yielded significant carryover by the laboratories using it (9.7%; Figure S6 [supporting information]), likely due to issues with MD. Similarly, carryover was found to be significant for L4 (3.4%), using the steam distillation method after isolating NH₄⁺ by acid traps before oxidation to N₂ with LiOBr. This method is afflicted with a high risk of nonspecificity, due to the aforementioned problems with acid traps, high distillation temperatures, and subsequent decomposition of organic N.⁴¹

3.4 | Environmental samples

Although the analysis of pure NO_3^- or NH_4^+ salts produced acceptable results for most methods applied by the participating laboratories, more divergent $\delta^{15} N$ values were reported when the isotopic composition of NO_3^- and NH_4^+ in the environmental samples was analyzed (Figure 2). Concentrations of NO_3^- and NH_4^+ were low to intermediate, ranging from 24 to 212 µmol l⁻¹ and from 2 to 38 μ mol l⁻¹, respectively, but within the range typically found for water samples and soil extracts (Table S3 [supporting information]). Concentrations of DON were from less than 10 µmol l⁻¹ µp to 142 µmol l⁻¹ also in the low to high range (Table S3 [supporting information]). Because the actual δ^{15} N values from the environmental samples were not known, we calculated the mean deviation and SD across all results. SD across all laboratories and methods (n = 8-10for each sample) reported for the environmental samples was 2.01% for PS1, 4.86‰ for PS2, 1.62‰ for HWS, and 8.43‰ for DR for NO_3^{-} analysis, with all laboratories and methods showing a similarly high variability. The maximum difference between two reported values was as high as 32.5% for DR. If SDs were calculated only for laboratories L1, L2, and L10 (laboratories that provided accurate and unbiased results for the standards; Figure 1), the SDs were on average smaller but still substantial for NO_3^- (0.65% for PS1, 3.83% for PS2, 1.91‰ for HWS, and 1.62‰ for DR), rendering high uncertainties in the reported mean values. SD across laboratories was even higher for NH_4^+ isotope results in the environmental samples (n = 5-7): 32.9‰ for PS1, 20.5‰ for PS2, 14.4‰ for HWS, 16.6‰ for DR, and 16.2‰ for DL. SDs within laboratories were for both NO_3^- and NH_4^+ found to be considerably lower (on average 3.1%), indicating again that the spread in mean values was largely due to differences in results reported from different laboratories. Given the high deviations between reported values for standards, we could not narrow down the "real" isotope value of NH_4^+ for the environmental samples based on the interlaboratory comparison. Although interlaboratory variability was very large and not connected to any methodological or sample factor (e.g., DON content, content of target compound), the intralaboratory variability or precision of $\delta^{15}\mathsf{N}$ of environmental samples correlated negatively with the target concentration for NO3⁻ analysis but not for NH_4^+ analysis (Figure S7 [supporting information]). The unsatisfying performance of each method or laboratory for environmental samples is, for one, connected to the shortcomings of each method (as discussed earlier) and related





10 of 16

FIGURE 2 The δ^{15} N results from isotopic analysis of environmental samples by different laboratories (A, δ^{15} N of NO₃⁻; B, δ^{15} N of NH₄⁺). The solid line indicates the average of all the results, plus minus standard deviation (±SD; shaded area). Red arrows indicate results from laboratories showing high precision and selectivity when analyzing standards, and the dotted line indicates the SD when calculating the average from those results. P1S, P2S, and HWS are soil extracts, and DR and DL are river samples (for more information, refer to the text and Table S3 [supporting information]). Black circles (L1–L3) = results from CM-N₂O; semi-filled diamonds (L4–L5) = results from CM-N₂; gray circles (L6–L9) = results from MD; crossed squares (L10–L13) = results from DN. For abbreviations of the methods, see the legend of Figure 1. Note that not all laboratories participating in this intercomparison analyzed ¹⁵N from environmental samples (but only from standards, e.g., L3, L12) [Color figure can be viewed at wileyonlinelibrary.com]

to issues with precision, accuracy, blank contamination, and nonspecificity. However, as the laboratories that did not report such problems yielded highly variable results, additional factors like matrix effects (e.g., high salt concentration in soil extracts or marine samples) or the impact of third compounds must be considered.^{20,26,60,61} Nitrite was not a problem here, because NO₂⁻ concentrations were below detection limit (<0.25 μ mol l⁻¹) in all environmental samples. However, if NO₂⁻ is detectable in the samples, it needs to be removed by sulfamic acid or ascorbic acid pretreatment.^{51,62,63} However, high salt concentrations should not be a problem for the N₂O-based methods, as supported by published data sets on reference materials prepared in seawater and freshwater (e.g., ^{references 28 and 51)} but it can be a problem for DN and CM-N₂.^{64,65}

Other method comparisons for $\delta^{15}N$ of NO₃⁻ in environmental samples showed much better (though also not optimal) agreement between different methods tested, for example, groundwater and surface water samples differed only by 0.4-0.5‰ between MD and CM-N₂ in a study carried out by Sebilo et al.⁶⁶ Thus, it shows that high accuracy can be achieved with MD if sufficient care is taken to optimize the MD procedure (Figure 3). Similarly, there was sufficient agreement ($\Delta \delta^{15}$ N of 1.5%) between δ^{15} N of NO₃⁻ of surface waters measured by a CM-N₂ (AgNO₃) method and DN, though much larger differences were found for specific samples.⁶⁵ Moreover, δ^{15} N results of NH4⁺ from environmental samples (lake water) measured with the BrO⁻ oxidation method also agreed sufficiently well with the MD method.⁴⁶ However, in all cases, the analyses were always carried out by one laboratory, suggesting that imprecise standardization procedures and erogenous application of methods by some laboratories are eventually responsible for the large spread in the mean values in our case. In agreement with that, Aly et al⁶⁷ demonstrated high variability of δ^{15} N of NO₃⁻ (up to appr. 5%) by conducting an interlaboratory comparison of various CM-N₂ and DN methods on standards.

Thus, the variability found here can also result from poor performance and/or weaknesses in one or another laboratory protocol, which were often modified from the original protocol by participants. Though it can be assumed that the participating laboratories are well experienced, the "schooling" and level of training obviously vary. The cause of the problem can also be found in the performance of the instrumentation, laboratory conditions, and other factors, which are difficult to trace back but which contribute to a loss in analytical accuracy and precision achieved by any method in any laboratory. Importantly, however, the high variability between the methods is also due to the unsuitability of specific methods for a particular sample type. For example, samples that contain low amounts of N are not suitable for MD, and so are samples with high amounts of DON. The limit of DON concentration acceptable for analysis with MD is 10 μ mol I^{-1} .³³ above which DON needs to be removed before analysis of δ^{15} N of N_i, for example, by microdialysis or using a sodium-acetone mixture to precipitate DON.⁶⁸ In case of too low amounts of N, there is the possibility of spiking with known quantities and isotope compositions of N_i, but this works only with labeled samples.⁶⁹ The pyrolysis-based AgNO₃ method (CM-N₂) is also not suitable for samples with high salt content, and there are also limitations at high salt content with DN. Still these methods are frequently used for such samples and work in soil extracts up to 2 M KCI.^{44,70} In Figure 3, we first provide guidelines that should aid in the decision-making process to find the best suitable method for specific sample types, variables, and parameters to optimize the protocols (best-practice guidelines) and further caution notes, to aid in producing the most reliable results. Assuming that our test is representative of laboratories routinely carrying out δ^{15} N analysis of N_i, our study shows that the current reproducibility is not sufficient for meta-analysis of δ^{15} N of N_i. We call for improved standardization and better validation of methods used by laboratories routinely conducting $\delta^{15}N$ analysis of N_i, and for more harmonized analytical protocols.

3.5 | Method comparison and recommendations on method improvements

Overall, for standards (both in single form and mixed), laboratories utilizing the CM-N₂O methods performed best in this interlaboratory comparison in terms of precision, accuracy, and specificity, followed by laboratories utilizing DN. CM-N₂O and DN methods have also been previously reported to be accurate and quantitative. 16,36,37,46,47,71,72 Both methods convert N_i forms to N₂O and analyze it using PT-IRMS, thereby pre-concentrating N₂O from a larger gas volume (10-30 mL) before IRMS analysis, and are thus highly sensitive, being able to analyze Ni at low concentrations (<30 nmol).^{16,33,71} One reason for the high accuracy is the complete conversion of CM methods, as also confirmed by our study. Another reason may be associated with the fact that certified standards are added through the entire run, which increases sample analysis time but simultaneously eliminates IRMS calibration issues because the standards, and not the N2O reference tanks, are used for normalization of stable isotope signals. The main disadvantages of the



recommendation for samples with high and low nitrogen content and for natural abundance (NA) and labeled samples for analysis of $\delta^{15}N$ of inorganic nitrogen forms (NO₃⁻, NH_4^+). Best-practice guidelines (standard operating procedures and optimization procedures) are included. Note that all methods require proper calibration and normalization procedures. Note also that for labeled samples, it is recommended to use labeled standards and to include "dummies" to avoid memory effects. The check mark and green color mean "recommended." the cross and red color mean "not suitable." and the exclamation mark and vellowish color mean "not recommended, only if." For more details, read the information in the boxes [Color figure can be

FIGURE 3 Method

viewed at wileyonlinelibrary.com]

CM-N₂O methods are the safety precautions required for handling the highly toxic azide reagent and its subsequent disposal (Table S7 [supporting information]). New methods are currently being developed, which hold promise to achieve similar precision and accuracy while utilizing less toxic reagents.²² Moreover, more suitable N₂O reference materials are currently being developed⁷³ to implement the recommended two- or three-point calibration approach for CM-N₂O and DN methods also in the future. For CM-N₂O, reliable results can be achieved, even though the blanks are relatively high, and the methods require isotopic reference standards to account for the offset of δ^{15} N deriving from the incorporation of one N atom in N₂O from azide,^{16,36,37,47} which involves extensive correction of the raw results in CM-N₂O methods. Some blank issues were noted from laboratories using DN, though the blank was very low, likely due to improper blank correction; direct measurements of the blank or employing indirect calculation methods/regression equations to account for δ^{15} N and the size of blanks is generally recommended for all methods, even when blanks are low.^{25,28,44} For DN, further improvements such as the one recently published by Zhu et al,⁵⁰ which involves the removal of background NO₃⁻ in the media by anoxic preincubation with *Pseudomonas denitrificans*, need to be achieved to reduce the blank size and to increase the accuracy of

 δ^{15} N analysis of NO₃^{-.51} The main weakness of DN is, however, the requirement and maintenance of special denitrifier cultures, which may be impractical for some laboratories (Table S7 [supporting information]).

Laboratories applying the MD methods, on the contrary, performed inferior for NA standards, both in single and in mixed forms. The main problem with this method is the low recovery, on the one side, and the high risk for carryover, on the other side (Table S7 [supporting information]). Improvements may be achieved by using NaOH instead of MgO, by keeping the incubation times short, by adopting cleaning techniques to properly remove NH₄⁺ before analysis of NO_3^{-} , by avoiding oversaturation of the acid traps, by using vessels with high surface-headspace ratios during MD, by using high-purity KCl and Devarda's alloy, and/or by removing DON before the analysis of Ni.^{20,21,35,52,74,75} Measurements of blanks and of standards within appropriate concentration ranges of the samples by MD are highly recommended to correct for the blank contribution and isotope fractionation due to incomplete recoveries in NA studies. Another disadvantage of the MD approach is the relatively large sample requirement and the high N concentrations/amounts needed (> 1 µmol N). CM-N₂ methods have similar disadvantages but were too variable to be reliably evaluated here.

When LAs were analyzed, laboratories using MD and CM-N₂ methods performed better than those with NA standards, because the isotope fractionation effects can be ignored. Most of the laboratories underestimated the true isotope values with LA, though the relative error was never more than 6%, which is in an acceptable range. To match the true at%¹⁵N values of labeled materials, laboratories would need to use certified, ¹⁵N-enriched international reference materials in a suitable range, ideally within the range of the samples, and utilize two- or more-point calibrations, an approach rarely adopted, even in well-known stable isotope laboratories for labeled samples.

Nearly all laboratories failed to provide comparable data for the isotopic composition of NO₃⁻ and NH₄⁺ in the environmental samples with the variable methods utilized. Matrix effects and effects of "contaminating" compounds can be a problem for all methods, but also analytical errors and improper standardization procedures need to be considered. We conclude that the current reproducibility is not sufficient for meta-analysis of δ^{15} N of N_i at NA and provide guidelines to aid choosing the right method for particular samples and steps for optimization of standard operation procedures (SOP) (Figure 3). This interlaboratory comparison is a first step for future endeavors to develop harmonized protocols and best-practice guidelines for δ^{15} N measurements of N_i.

Though all the methods studied here have never been comparatively evaluated in one systematic interlaboratory test, the main differences between them are in agreement with prior tests of the individual methods³²⁻³⁵ and reflect the historical development of the individual methods. The CM-N₂ methods (e.g., hypobromite oxidation) were the first in the field, and other methods were developed subsequently, to overcome their weaknesses, which are mostly related to their low precision and accuracy and the high

sample need. The MD methods are known to have two overarching challenges, as discussed here and in other methods papers: 15,20,33,52,54 (a) the completeness of NH₄⁺ collection from the sample into the acid traps, coupled with the strong isotope effect of this step, and (b) cross-reactivity by dissolved organic nitrogen in natural samples. These challenges were part of the motivation for the development of N₂O-based methods for NA isotopic analyses, the DN and the CM-N₂O methods, alongside with great improvements in the sensitivity for N_i analysis. Accordingly, the better performance of the N₂O-based methods is not surprising and indeed has been demonstrated previously, explaining why so much of the field has adopted these methods. This interlaboratory comparison was nevertheless novel and comprehensive, because (a) never before all the three method groups were compared simultaneously on the same set of samples, (b) both NO_3^- and NH_4^+ isotope analyses were performed in one study, (c) both NA and ¹⁵N-enriched materials were investigated, (d) both reference materials and environmental samples were included in the interlaboratory comparison, and (e) by conducting the test as an the interlaboratory comparison, we examined the variability of the whole method chain and thus also the performance of laboratories. This allowed us to test thus whether δ^{15} N results from N_i produced by different laboratories using the same or a different method are intercomparable. Particularly, the latter two points revealed major, novel challenges and important new insights into the capability of commonly used IRMS methods to reproduce δ^{15} N of N_i in natural samples.

4 | CONCLUSIONS

Most laboratories using common methods to analyze ¹⁵N in N_i produced acceptable results of $\delta^{15}N$ of N_i for standards in single forms. The accuracy of the results depends not only on the method used but also on the performance of the laboratory and the application and suitability of the method for the particular sample. The highly variable results of the environmental samples tested in this interlaboratory comparison suggest, however, large uncertainties in any $\delta^{15}N$ data reported from environmental samples analyzed by any of the widely used methods. It is difficult to relate the overall unsatisfying performance to singular factors, but given that many participating laboratories routinely analyze $\delta^{15}N$ of N_i compounds in environmental samples, these differences are of concern. Further systematic tests are immediately needed to test, evaluate, and optimize the methods compared in this interlaboratory comparison. In addition, the performance of laboratories conducting routinely $\delta^{15}N$ analysis of Ni must be improved. Before this careful examination, caution should be exercised when interpreting isotope data of N_i forms. The development of environmentally relevant standards, where sample matrices (e.g., river water, soil extracts) are first stripped from native NH_4^+ and NO_3^- by ion exchange or DN approaches, followed by addition of Ni of known isotopic composition, would greatly help to assess and compare the methods. In addition, it is highly recommended that laboratories include known

14 of 16

WILEY-Withow

reference materials in their analysis and validate their measurements before sending out results. Regular participation in relevant proficiency tests and interlaboratory comparisons is strongly recommended. This should enable laboratories to detect poor performance as a result of mistakes during analytical runs.

ACKNOWLEDGMENTS

This work was supported by EURAMET (European Association of National Metrology Institutes) through the projects 16ENV06 SIRS and 19ENV05, *STELLAR*. The authors further thank the isotope programs SIBAE (ESF) and NORD-SIR for their support in coordinating and initializing the interlaboratory comparison.

AFFILIATIONS

¹Department of Environmental and Biological Sciences, University of Eastern Finland, Kuopio, Finland

²Division of Terrestrial Ecosystem Research, Department of

Microbiology and Ecosystem Science, University of Vienna, Vienna, Austria

³Department of Geosciences and Natural Resource Management, University of Copenhagen, Copenhagen K, Denmark

⁴Faculty of Environmental Sciences and Natural Resource Management, Norwegian University of Life Sciences, Ås, Norway ⁵Laboratory for Air Pollution & Environmental Technology, Swiss Federal Laboratories for Materials Science and Technology, Empa, Dübendorf, Switzerland

⁶Rothamsted Research, North Wyke, Okehampton, Devon, UK ⁷Isotope Bioscience Laboratory-ISOFYS, Department of Green Chemistry and Technology, Ghent University, Ghent, Belgium

⁸Forschungszentrum Jülich GmbH, Institute of Bio- and

Geosciences—Agrosphere (IBG-3), Jülich, Germany

⁹Hydroisotop GmbH, Woelkestr, Schweitenkirchen, Germany
¹⁰Geological Survey of Israel, Jerusalem, Israel

¹¹Agriculture and Food CSIRO, Urrbrae, South Australia, Australia
 ¹²Department of Geography, Université de Montréal, Québec,
 Canada

¹³Stable Isotope Laboratory, Institute of Geological Sciences, Polish Academy of Sciences, Warszawa, Poland

¹⁴Micro-area Analysis Laboratory, Polish Geological Institute– National Research Institute, Warszawa, Poland

PEER REVIEW

The peer review history for this article is available at https://publons. com/publon/10.1002/rcm.9370.

DATA AVAILABILITY STATEMENT

Data available on request from the authors

ORCID

Christina Biasi b https://orcid.org/0000-0002-7413-3354 Per Ambus https://orcid.org/0000-0001-7580-524X Peter Dörsch https://orcid.org/0000-0002-4916-1839 Longfei Yu https://orcid.org/0000-0002-2127-6343 Steve Granger b https://orcid.org/0000-0003-0183-0244 Helena Jäntti b https://orcid.org/0000-0002-6400-9114 Nina Welti b https://orcid.org/0000-0001-9966-5915 Carolina Voigt b https://orcid.org/0000-0001-8589-1428 Beata Gebus-Czupyt b https://orcid.org/0000-0002-0146-3627 Wolfgang Wanek b https://orcid.org/0000-0003-2178-8258

REFERENCES

- Sigman DM, Casciotti KL, Andreani M, Barford C, Galanter M, Bohlke JK. A bacterial method for the nitrogen isotopic analysis of nitrate in seawater and freshwater. *Anal Chem.* 2001;73(17):4145-4153. doi:10.1021/ac010088e
- LeBauer DS, Treseder KK. Nitrogen limitation of net primary productivity in terrestrial ecosystems is globally distributed. *Ecology*. 2008;89(2):371-379. doi:10.1890/06-2057.1
- Reich PB, Hungate BA, Luo Y. Carbon-nitrogen interactions in terrestrial ecosystems in response to rising atmospheric carbon dioxide. Annu Rev Ecol Evol Syst. 2006;37(1):611-636. doi:10.1146/ annurev.ecolsys.37.091305.110039
- Schlesinger WH. On the fate of anthropogenic nitrogen. Proc Natl Acad Sci U S A. 2009;106(1):203-208. doi:10.1073/pnas.0810193105
- Bowen JL, Kroeger KD, Tomasky G, et al. A review of land-sea coupling by groundwater discharge of nitrogen to New England estuaries: Mechanisms and effects. *Appl Geochem*. 2007;22(1):175-191. doi:10.1016/j.apgeochem.2006.09.002
- Gao W, Yang H, Kou L, Li S. Effects of nitrogen deposition and fertilization on N transformations in forest soils: A review. J Soil Sediment. 2015;15(4):863-879. doi:10.1007/s11368-015-1064-z
- Fowler D, Coyle M, Skiba U, et al. The global nitrogen cycle in the twenty-first century. *Philos Trans R Soc Lond B Biol Sci.* 2013; 368(1621):20130165. doi:10.1098/rstb.2013.0165
- Mosier A, Kroeze C, Nevison C, Oenema O, Seitzinger S, van Cleemput O. Closing the global N2O budget: Nitrous oxide emissions through the agricultural nitrogen cycle. Nutr Cycl Agroecosyst. 1998; 52(2):225-248. doi:10.1023/A:1009740530221
- Galloway JN, Townsend AR, Erisman JW, et al. Transformation of the nitrogen cycle: Recent trends, questions, and potential solutions. *Science*. 2008;320(5878):889-892. doi:10.1126/science. 1136674
- Gruber N, Galloway JN. An earth-system perspective of the global nitrogen cycle. *Nature*. 2008;451(7176):293-296. doi:10.1038/ nature06592
- van Groenigen JW, Huygens D, Boeckx P, et al. The soil N cycle: New insights and key challenges. Soil. 2015;1(1):235-256. doi:10.5194/ soil-1-235-2015
- 12. Butterbach-Bahl K, Baggs EM, Dannenmann M, Kiese R, Zechmeister-Boltenstern S. Nitrous oxide emissions from soils: how well do we understand the processes and their controls? *Philos Trans R Soc Lond, B, Biol Sci.* 2013;368(1621).
- Booth MS, Stark JM, Rastetter E. Controls on nitrogen cycling in terrestrial ecosystems: A synthetic analysis of literature data. *Ecol Monogr.* 2005;75(2):139-157. doi:10.1890/04-0988
- Denk TRA, Mohn J, Decock C, et al. The nitrogen cycle: A review of isotope effects and isotope modeling approaches. *Soil Biol Biochem*. 2017;105:121-137. doi:10.1016/j.soilbio.2016.11.015
- Matiatos I, Wassenaar LI, Monteiro LR, et al. Global patterns of nitrate isotope composition in rivers and adjacent aquifers reveal reactive nitrogen cascading. *Commun Earth Environ*. 2021;2(1): doi:10. 1038/s43247-021-00121-x
- Robinson D. Delta N-15 as an integrator of the nitrogen cycle. *Trends Ecol Evol*. 2001;16(3):153-162. doi:10.1016/S0169-5347(00) 02098-X

- Lachouani P, Frank AH, Wanek W. A suite of sensitive chemical methods to determine the delta N-15 of ammonium, nitrate and total dissolved N in soil extracts. *Rapid Commun Mass Spectrom.* 2010; 24(24):3615-3623. doi:10.1002/rcm.4798
- Barrett JE, Burke IC. Potential nitrogen immobilization in grassland soils across a soil organic matter gradient. Soil Biol Biochem. 2000; 32(11–12):1707-1716. doi:10.1016/S0038-0717(00)00089-4
- Brooks PD, Stark JM, McInteer BB, Preston T. Diffusion method to prepare soil extracts for automated Nitrogen-15 analysis. *Soil Sci Soc Am J.* 1989;53(6):1707-1711. doi:10.2136/sssaj1989. 03615995005300060016x
- Sorensen P, Jensen ES. S sequential diffusion of ammonium and nitrate from soil extracts to a polytetrafluoroethylene trap for N-15 determination. Anal Chim Acta. 1991;252(1-2):201-203. doi:10. 1016/0003-2670(91)87215-S
- Stark JM, Hart SC. Diffusion technique for preparing salt solutions, Kjeldahl digests, and persulfate digests for nitrogen-15 analysis. *Soil Sci Soc Am J.* 1996;60(6):1846-1855. doi:10.2136/sssaj1996. 03615995006000060033x
- Sigman DM, Altabet MA, Michener R, McCorkle DC, Fry B, Holmes RM. Natural abundance-level measurement of the nitrogen isotopic composition of oceanic nitrate: An adaptation of the ammonia diffusion method. *Mar Chem.* 1997;57(3–4):227-242. doi: 10.1016/S0304-4203(97)00009-1
- Altabet MA, Wassenaar LI, Douence C, Roy R. A Ti (III) reduction method for one-step conversion of seawater and freshwater nitrate into N2O for stable isotopic analysis of N-15/N-14, O-18/O-16 and O-17/O-16. *Rapid Commun Mass Spectrom*. 2019;33(15):1227-1239. doi:10.1002/rcm.8454
- Huber B, Bernasconi SM, Luster J, Pannatier EG. A new isolation procedure of nitrate from freshwater for nitrogen and oxygen isotope analysis. *Rapid Commun Mass Spectrom*. 2011;25(20):3056-3062. doi: 10.1002/rcm.5199
- Lehmann MF, Bernasconi SM, McKenzie JA. A method for the extraction of ammonium from freshwaters for nitrogen isotope analysis. *Anal Chem.* 2001;73(19):4717-4721. doi:10.1021/ ac010212u
- Rock L, Ellen BH. Nitrogen-15 and oxygen-18 natural abundance of potassium chloride extractable soil nitrate using the denitrifier method. Soil Sci Soc Am J. 2007;71(2):355-361. doi:10.2136/ sssaj2006.0266
- Silva SR, Kendall C, Wilkison DH, Ziegler AC, Chang CCY, Avanzino RJ. A new method for collection of nitrate from fresh water and the analysis of nitrogen and oxygen isotope ratios. *J Hydrol.* 2000;228(1–2):22-36. doi:10.1016/S0022-1694(99) 00205-X
- Christensen S, Tiedje JM. S sub-parts-per-billion nitrate method: Use of an N2O-producing denitrifier to convert NO3– or 15NO3– to N2O. Appl Environ Microbiol. 1988;54(6):1409-1413. doi:10.1128/ aem.54.6.1409-1413.1988
- Stange CF, Spott O, Apelt B, Russow RWB. Automated and rapid online determination of N-15 abundance and concentration of ammonium, nitrite, or nitrate in aqueous samples by the SPINMAS technique. *Isotopes Environ Health S.* 2007;43(3):227-236. doi:10. 1080/10256010701550658
- Eschenbach W, Lewicka-Szczebak D, Stange CF, Dyckmans J, Well R. Measuring N-15 abundance and concentration of aqueous nitrate, nitrite, and ammonium by membrane inlet quadrupole mass spectrometry. *Anal Chem.* 2017;89(11):6077-6082. doi:10.1021/acs. analchem.7b00724
- Yin GY, Hou LJ, Liu M, Liu ZF, Gardner WS. A novel membrane inlet mass spectrometer method to measure (NH4+)-N-15 for isotopeenrichment experiments in aquatic ecosystems. *Environ Sci Technol.* 2014;48(16):9555-9562. doi:10.1021/es501261s

 Cooney KS. A stable isotopic investigation of precipitation nitrate: Faculty of the Graduate School. Md.), Maryland: University of Maryland (College Park; 2005.

Rapid

15 of 16

WILEY

- Jantti H, Leskinen E, Stange CF, Hietanen S. Measuring nitrification in sediments - comparison of two techniques and three (NO3-)-N-15 measurement methods. *Isotopes Environ Health Stud.* 2012;48(2): 313-326. doi:10.1080/10256016.2012.641543
- Jokinen S. Analyzing stable isotope signatures of inorganic nitrogen forms a method comparison. Kuopio: University of Eastern Finland; 2013.
- Ros GH, Temminghoff EJM, van Groenigen JW. Isotopic analysis of dissolved organic nitrogen in soils. *Anal Chem.* 2010;82(18):7814-7820. doi:10.1021/ac1018183
- McIlvin MR, Altabet MA. Chemical conversion of nitrate and nitrite to nitrous oxide for nitrogen and oxygen isotopic analysis in freshwater and seawater. Anal Chem. 2005;77(17):5589-5595. doi:10.1021/ ac050528s
- Schilman B, Teplyakov N, Gavrieli I, Guttman Y. Identification of the Sources of Nitrate Contamination in the Western Galilee Aquifers, Northern Israel, Using a Multi-proxy Stable Isotope and Chemical Approach. Geological Survey of Israel: Ministry of National Infrastructures; 2007.
- Keeney DR, Nelson DW. Nitrogen-Inorganic Forms. In: Page AL, ed. Methods of Soil Analysis, Agronomy Monograph 9, Part 2. 2nd ed. Madison, WI: ASA, SSSA; 1982:643-698.
- Stevens RJ, Laughlin RJ. Determining Nitrogen-15 in nitrite or nitrate by producing nitrous oxide. *Soil Sci Soc Am J*. 1994;58(4):1108-1116. doi:10.2136/sssaj1994.03615995005800040015x
- Hauck RD. Nitrogen-isotope-ratio analysis. In: Page AL, ed. Methods of Soil Analysis, Agronomy Monograph 9. Part 2. 2nd ed. Madison, WI: ASA, SSSA; 1982:735-779.
- Saghir NS, Mulvaney RL, Azam F. Determination of nitrogen by microdiffusion in mason jars: Inorganic nitrogen in soil extracts. *Commun Soil Sci Plant Anal.* 1993;24(13–14):1745-1762. doi:10. 1080/00103629309368912
- Amberger A, Schmidt HL. Natürliche Isotopengehalte von Nitrat als Indikatoren für dessen Herkunft. *Geochim Cosmochim Acta*. 1987; 51(10):2699-2705. doi:10.1016/0016-7037(87)90150-5
- Bottcher J, Strebel O, Voerkelius S, Schmidt HL. Using isotope fractionation of nitrate-nitrogen and nitrate-oxygen for evaluation of microbial denitrification in a sandy aquifer. J Hydrol. 1990;114(3-4): 413-424. doi:10.1016/0022-1694(90)90068-9
- Morkved PT, Dorsch P, Sovik AK, Bakken LR. Simplified preparation for the delta N-15-analysis in soil NO3 by the denitrifier method. *Soil Biol Biochem*. 2007;39(8):1907-1915. doi:10.1016/j.soilbio.2007. 02.004
- Bahlmann E, Bernasconi SM, Bouillon S, et al. Performance evaluation of nitrogen isotope ratio determination in marine and lacustrine sediments: An inter-laboratory comparison. Org Geochem. 2010;41(1): 3-12. doi:10.1016/j.orggeochem.2009.05.008
- Zhang L, Altabet MA, Wu TX, Hadas O. Sensitive measurement of (NH4+N)-N-15/N-14 (delta (NH4+)-N-15) at natural abundance levels in fresh and saltwaters. *Anal Chem.* 2007;79(14):5297-5303. doi:10.1021/ac070106d
- Ti CP, Wang X, Yan XY. Determining delta N-15-NO3 (–) values in soil, water, and air samples by chemical methods. *Environ Monit* Assess. 2018;190(6):341. doi:10.1007/s10661-018-6712-5
- Casciotti KL, Sigman DM, Hastings MG, Bohlke JK, Hilkert A. Measurement of the oxygen isotopic composition of nitrate in seawater and freshwater using the denitrifier method. *Anal Chem.* 2002;74(19):4905-4912. doi:10.1021/ac020113w
- 49. Isobe K, Suwa Y, Ikutani J, et al. Analytical techniques for quantifying N-15/N-14 of nitrate, nitrite, Total dissolved nitrogen and ammonium in environmental samples using a gas chromatograph equipped with a

quadrupole mass spectrometer. *Microbes Environ*. 2011;26(1):46-53. doi:10.1264/jsme2.ME10159

- Zhu J, Yu LF, Bakken LR, Morkved PT, Mulder J, Dorsch P. Controlled induction of denitrification in pseudomonas aureofaciens: A simplified denitrifier method for dual isotope analysis in NO3. *Sci Total Environ*. 2018;633:1370-1378. doi:10.1016/j.scitotenv.2018. 03.236
- Weigand MA, Foriel J, Barnett B, Oleynik S, Sigman DM. Updates to instrumentation and protocols for isotopic analysis of nitrate by the denitrifier method. *Rapid Commun Mass Spectrom*. 2016;30(12):1365-1383. doi:10.1002/rcm.7570
- Stephan K, Kavanagh KL. Suitability of the diffusion method for natural abundance Nitrogen-15 analysis. *Soil Sci Soc Am J.* 2009;73(1): 293-302. doi:10.2136/sssaj2007.0079
- Cejudo E, Schiff SL. Nitrogen isotope fractionation factors (alpha) measured and estimated from the volatilisation of ammonia from water at pH 9.2 and pH 8.5. *Isot Environ Health Stud.* 2018;54(6): 642-655.
- Jensen ES. Evaluation of automated-analysis of N-15 and total N in plant-material and soil. *Plant and Soil*. 1991;133(1):83-92. doi:10. 1007/BF00011902
- Mulvaney RI. Mass spectrometry. In: Blackburn TH, Knowles R, eds. Nitrogen Isotope Techniques. San Diego, CA: Academic Press; 1993: 11-57. doi:10.1016/B978-0-08-092407-6.50007-9.
- Werner RA, Brand WA. Referencing strategies and techniques in stable isotope ratio analysis. *Rapid Commun Mass Spectrom*. 2001; 15(7):501-519. doi:10.1002/rcm.258
- Petzke KJ, Metges CC. Practical recommendations for the reduction of memory effects in compound-specific N-15/N-14-ratio analysis of enriched amino acids by gas chromatography/combustion/isotope ratio mass spectrometry. *Rapid Commun Mass Spectrom*. 2012;26(2): 195-204. doi:10.1002/rcm.5319
- Mulvaney RL, Khan SA. Diffusion methods to determine different forms of nitrogen in soil hydrolysates. *Soil Sci Soc Am J.* 2001;65(4): 1284-1292. doi:10.2136/sssaj2001.6541284x
- Liu KK, Su MJ, Hsueh CR, Gong GC. The nitrogen isotopic composition of nitrate in the Kuroshio water northeast of Taiwan: Evidence for nitrogen fixation as a source of isotopically light nitrate. *Mar Chem.* 1996;54(3-4):273-292. doi:10.1016/0304-4203(96) 00034-5
- Morkved PT, Dorsch P, Bakken LR. The N2O product ratio of nitrification and its dependence on long-term changes in soil pH. *Soil Biol Biochem*. 2007;39(8):2048-2057. doi:10.1016/j.soilbio.2007. 03.006
- Rutting T, Huygens D, Muller C, Cleemput O, Godoy R, Boeckx P. Functional role of DNRA and nitrite reduction in a pristine south Chilean Nothofagus forest. *Biogeochemistry*. 2008;90(3):243-258. doi: 10.1007/s10533-008-9250-3
- Granger J, Sigman DM. Removal of nitrite with sulfamic acid for nitrate N and O isotope analysis with the denitrifier method. *Rapid Commun Mass Spectrom*. 2009;23(23):3753-3762. doi:10.1002/rcm. 4307
- Granger J, Sigman DM, Prokopenko MG, Lehmann MF, Tortell PD. A method for nitrite removal in nitrate N and O isotope analyses. *Limnol Oceanogr-Meth.* 2006;4(7):205-212. doi:10.4319/lom.2006. 4.205
- Stock P, Roder S, Burghardt D. Further optimisation of the denitrifier method for the rapid N-15 and O-18 analysis of nitrate in natural water samples. *Rapid Commun Mass Spectrom*. 2021;35(1):e8931. doi: 10.1002/rcm.8931

- 65. Xue DM, De Baets B, Botte J, Vermeulen J, Van Cleemput O, Boeckx P. Comparison of the silver nitrate and bacterial denitrification methods for the determination of nitrogen and oxygen isotope ratios of nitrate in surface water. *Rapid Commun Mass Spectrom.* 2010;24(6):833-840. doi:10.1002/rcm.4445
- Sebilo M, Mayer B, Grably M, Billiou D, Mariotti A. The use of the 'Ammonium Diffusion' method for delta N-15-NH4+ and delta N-15-NO3- measurements: Comparison with other techniques. *Environ Chem.* 2004;1(2):99-103. doi:10.1071/EN04037
- Aly AIM, Ahmed MA, Gomaa HE, Abdel Monem N, Hanafy M. Rapid and accurate technique for measuring N-15 of nitrate method development, Calibration and Validation. *J Radiat Res.* 2010;42: 39-56.
- Huber B, Bernasconi SM, Pannatier EG, Luster J. A simple method for the removal of dissolved organic matter and d15N analysis of NO3-from freshwater. *Rapid Commun Mass Spectrom*. 2012;26(12): 1475-1480. doi:10.1002/rcm.6243
- Mannerheim N, Werner RA, Buchmann N. Measurement precision and accuracy of high artificial enrichment N-15 and C-13 tracer samples. *Rapid Commun Mass Spectrom*. 2019;33(13):1153-1163. doi: 10.1002/rcm.8451
- Fenech C, Rock L, Nolan K, Tobin J, Morrissey A. The potential for a suite of isotope and chemical markers to differentiate sources of nitrate contamination: A review. *Water Res.* 2012;46(7):2023-2041. doi:10.1016/j.watres.2012.01.044
- Tu Y, Fang YT, Liu DW, Pan YP. Modifications to the azide method for nitrate isotope analysis. *Rapid Commun Mass Spectrom*. 2016; 30(10):1213-1222. doi:10.1002/rcm.7551
- Wassenaar LI, Douence C, Altabet MA, Aggarwal PK. N and O isotope (N-15, N-15, O-18, O-17) analyses of dissolved NO3- and NO2- by the cd-azide reduction method and N2O laser spectrometry. *Rapid Commun Mass Spectrom*. 2018;32(3):184-194. doi:10.1002/ rcm.8029
- Hill-Pearce RE, Hillier A, Webber EM, et al. Characterisation of gas reference materials for underpinning atmospheric measurements of stable isotopes of nitrous oxide. *Atmos Meas Tech.* 2021;14(8):5447-5458. doi:10.5194/amt-14-5447-2021
- Griesheim KL, Mulvaney RL. Improving the accuracy of diffusion for inorganic N-15 analyses of soil extracts and water. *Commun Soil Sci Plant Anal.* 2019;50(9):1161-1169. doi:10.1080/00103624.2019. 1604734
- Schleppi P, Bucher-Wallin I, Saurer M, Jaggi M, Landolt W. Citric acid traps to replace sulphuric acid in the ammonia diffusion of dilute water samples for N-15 analysis. *Rapid Commun Mass Spectrom*. 2006;20(4):629-634. doi:10.1002/rcm.2351

SUPPORTING INFORMATION

Additional supporting information can be found online in the Supporting Information section at the end of this article.

How to cite this article: Biasi C, Jokinen S, Prommer J, et al. Challenges in measuring nitrogen isotope signatures in inorganic nitrogen forms: An interlaboratory comparison of three common measurement approaches. *Rapid Commun Mass Spectrom*. 2022;36(22):e9370. doi:10.1002/rcm.9370