

CORRECTION

Open Access



Correction to: Spasticity assessment based on the Hilbert–Huang transform marginal spectrum entropy and the root mean square of surface electromyography signals: a preliminary study

Baohua Hu¹, Xiufeng Zhang¹, Jingsong Mu², Ming Wu² and Yong Wang^{1*}

*Correspondence:

simenkouwang@sina.com

¹ School of Mechanical Engineering, Hefei University of Technology, No. 193 Tunxi Road, Hefei 230009, China
Full list of author information is available at the end of the article

Correction to: *Biomed Eng OnLine* (2018) 17:27

<https://doi.org/10.1186/s12938-018-0460-1>

After the publication of the original article work [1], it was highlighted that the circuit information and the A/D converter used for acquiring sEMG signal were not correctly reported in “Methods” section. The circuit included a 10 Hz high-pass filter instead of 10 Hz notch filter as stated in the original article. The A/D converter was ADS7818, instead of ADS1198 as stated in the original paper. The authors apologize to the readers for the inconvenience.

Author details

¹ School of Mechanical Engineering, Hefei University of Technology, No. 193 Tunxi Road, Hefei 230009, China. ² Department of Rehabilitation Medicine, Anhui Provincial Hospital, No. 1 Swan Lake Road, Hefei 230001, China.

The original article can be found online at <https://doi.org/10.1186/s12938-018-0460-1>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Received: 8 March 2019 Accepted: 8 March 2019

Published online: 20 March 2019

Reference

1. Hu B, Zhang X, Mu J, Wu M, Wang Y. Spasticity assessment based on the Hilbert–Huang transform marginal spectrum entropy and the root mean square of surface electromyography signals: a preliminary study. *BioMed Eng Online*. 2018;17:27. <https://doi.org/10.1186/s12938-018-0460-1>.

