

RESEARCH ARTICLE

Mono-Exponential Fitting in T2-Relaxometry: Relevance of *Offset* and First Echo

David Milford*, Nicolas Rosbach, Martin Bendszus, Sabine Heiland

Department of Neuroradiology, University Hospital Heidelberg, Heidelberg, Germany

* David.Milford@med.uni-heidelberg.de

Abstract

Introduction

T2 relaxometry has become an important tool in quantitative MRI. Little focus has been put on the effect of the refocusing flip angle upon the *offset* parameter, which was introduced to account for a signal floor due to noise or to long T2 components. The aim of this study was to show that B1 imperfections contribute significantly to the *offset*. We further introduce a simple method to reduce the systematic error in T2 by discarding the first echo and using the *offset* fitting approach.

Materials and Methods

Signal curves of T2 relaxometry were simulated based on extended phase graph theory and evaluated for 4 different methods (inclusion and exclusion of the first echo, while fitting with and without the *offset*). We further performed T2 relaxometry in a phantom at 9.4T magnetic resonance imaging scanner and used the same methods for post-processing as in the extended phase graph simulated data. Single spin echo sequences were used to determine the correct T2 time.

Results

The simulation data showed that the systematic error in T2 and the *offset* depends on the refocusing pulse, the echo spacing and the echo train length. The systematic error could be reduced by discarding the first echo. Further reduction of the systematic T2 error was reached by using the *offset* as fitting parameter. The phantom experiments confirmed these findings.

Conclusion

The fitted offset parameter in T2 relaxometry is influenced by imperfect refocusing pulses. Using the *offset* as a fitting parameter and discarding the first echo is a fast and easy method to minimize the error in T2, particularly for low to intermediate echo train length.



OPEN ACCESS

Citation: Milford D, Rosbach N, Bendszus M, Heiland S (2015) Mono-Exponential Fitting in T2-Relaxometry: Relevance of *Offset* and First Echo. PLoS ONE 10(12): e0145255. doi:10.1371/journal.pone.0145255

Editor: Xiaobing Fan, University of Chicago, UNITED STATES

Received: June 12, 2015

Accepted: November 30, 2015

Published: December 17, 2015

Copyright: © 2015 Milford et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: All relevant data are available within the paper and its Supporting Information files.

Funding: This research was supported by grants of the Deutsche Forschungsgemeinschaft (DFG, SFB 1118) and of the Dietmar-Hopp-Stiftung.

Competing Interests: The authors have declared that no competing interests exist.

Introduction

T2 relaxometry is a frequently used method of magnetic resonance imaging (MRI), particularly in preclinical and clinical research. Ever since the first publication in 1971 by Damadian [1], researchers and clinicians alike tried to determine the T2 as bio markers for various diseases and as a parameter for prognosis and therapy control.

The “gold standard” method for acquiring T2 relaxometry data is the use of multiple single Spin Echo (SE) sequences with different echo times (TE)[2]. Due to the time constraints in clinical routine, however, Multi-Spin Echo (MSE) sequences [3] are generally used. MSE allows for multiple echoes within one acquisition depending on the number of 180° refocusing pulses. The number of echoes is given by the so called echo train length (ETL) and is usually constructed as a CPMG sequence [4]. Major reasons for incorrect T2 times are imperfect slice excitation profiles and issues with B1 inhomogeneities yielding low refocusing flip angles (FA) [5–7]. Multiple groups have tried to compensate and correct for these inhomogeneities. These techniques, however, are usually computationally intensive, complicated to implement or are restricted to a certain set of sequence parameters [8–12]. Further reasons for inaccurate T2 measurements are long superimposing T2 components either due to partial volume effects or due to several proton pools [13, 14]; furthermore incorrect sampling of the signal decay can contribute to errors in T2 [15–17].

Although those potential sources of systematic errors in T2 calculation are known, more often than not, post-processing and data-fitting techniques do not account for them. Data from T2 relaxometry are most often fitted to a simple exponential curve:

$$S(TE) = kS_0 \cdot \exp(-TE/T2) \tag{1}$$

Where k is a proportionality constant subsuming signal gain or attenuation by the scanner’s hard-/software, S_0 is the proton density and TE is the echo time. In most cases k and S_0 are merged together to a single factor, because true proton density is hard to separate from the signal gain caused by the measurement process itself.

Besides this simple mono-exponential fit, sometimes an offset or baseline is introduced:

$$S(TE) = kS_0 \cdot \exp\left(-\frac{TE}{T2}\right) + offset, \tag{2}$$

where *offset* is thought to represent a non-zero baseline taking into account signal that may not have converged towards zero. While this approach is sometimes used without explanation for the non-zero baseline [18], some groups use it for compensation of long T2 components such as CSF [14, 19], and some use it to compensate for offset signal originating from the system [20] or Rician noise, particularly at low signal-to-noise (SNR) level [21, 22].

The aim of this study is to show, that B1 inhomogeneities and imperfect refocusing pulses contribute significantly to the *offset*. By using extended phase graph (EPG) theory [23–26] we aim to show that errors in the early echoes caused by B1 inhomogeneities increase the *offset* depending on FA, echo spacing (ESP) and echo train length (ETL) and the T2 of the tissue. We further introduce a simple method to reduce the systematic error on T2 due to B1 imperfection by discarding the very first echo and using the *offset* within the fitting approach.

Materials and Methods

Data simulation

T2 relaxation curves were simulated based on the EPG method [26]. Before one can answer which equation provides with the most stable solution, one has to know what parameters will

change the signal and cause discrepancies in the evaluation of T2, S_o and offset. When simulating the data with EPG theory we have the ability to alter the following parameters; T2, T1, ESP, ETL, FA and S_o . It has been shown in previous work that T1 does not alter the curves to a discernible degree [11]. Thus, we kept T1 constant at 3000ms. Unless otherwise stated the following parameters were used: $S_o = 1000a.u.$, T2 = 100ms, ETL = 20–50ms in steps of 2ms, ESP = 10–40 in steps of 2 and FA = 120°–180° in steps of 20°. To monitor the accuracy of the four fitting methods for different T2, simulation were undertaken with T2 times of 20ms, 60ms, and 100ms. For these simulations we used an ESP time of 5ms with an ETL of 32 (to ensure that the complete T2 decay is covered by the echoes) and a FA of 120°. Furthermore, to monitor the effect of noise, the simulations were rerun with added Rician noise. The used algorithm to include the noise was $S_r = F * Ra + S_c$, where S_r is the resulting signal with noise, F the noise factor (signal to noise ratio of the first echo, set to 10), Ra random generated number and S_c is the noise-free signal from the EPG simulation. These simulation were ran 1000 times to monitor not only the systematic error but the statistical error as well.

It is important to note that the EPG algorithm itself does not account for B1 inhomogeneity within the slice profile. One could follow the approach of Lebel et. al. [12] and break up the slice into partitions with quasi-homogeneous B1 and then perform EPG simulation for each partition. However, as the scope of this study is to identify the link between B1 inhomogeneity and *offset*, there is no need to distinguish between different sources of B1 inhomogeneities. Therefore EPG was used to simulate the measured signal for imperfect refocusing pulses without considering the origin of this imperfection.

Phantom measurements

To test the results from the EPG based simulations, phantom measurements were performed using a 50ml tube filled with a 2.5% agarose to water mixture (to ensure that $3 * T1 < TR$). Images were performed on a Bruker 9.4T horizontal bore NMR scanner (BioSpec 94/20 USR, Bruker BioSpin GmbH, Ettlingen, Germany) with a four channel phased array surface coil. Firstly, spin echo sequences with single refocusing pulse were acquired to obtain the T2 time without systematic errors by B1 inhomogeneities. Following this, MSE sequences were used at different pulse angles (120°, 140°, 160° and 180°). The following parameters were the same in all spin echo and MSE sequences: TR = 3000ms, matrix = 150 x 150, FoV = 30mm x 30mm, slice thickness/number = 2mm/1, slice selective pulses, acquisition time = 7min 30sec). We performed 45 different spin echo sequences with TE ranging from 10ms to 450ms with 10ms spacing. For the MSE, we used ESP = 10ms and ETL = 45 to obtain the same echo times as used for the spin echo sequences.

Calculation of T2

Curve fitting was undertaken in MATLAB, release 2014a® (MathWorks Inc.), using the Levenberg-Marquardt nonlinear least squares algorithm [27] provided by levmar [28]. Four different techniques were used for determining T2, S_o and the offset:

1. all echoes were fitted with Eq 1,
2. all echoes were fitted with Eq 2,
3. the first echo was discarded and the remaining echoes were fitted with Eq 1,
4. the first echo was discarded and the remaining echoes were fitted with Eq 2.

To determine the systematic error coming from curve fitting the relative deviation was calculated for T2 ($dT2 = |(T2_{fit} - T2_{in}) / T2_{in}| * 100$) and S_o ($dS_o = |(S_{ofit} - S_{oin}) / S_{oin}| * 100$).

Results

Influence of B1 Inhomogeneities upon *offset*

The offset parameter in Eq 2 has been introduced to cover cases where the T2 decay does not tend towards zero, but to an asymptote > 0 . Due to the fitting process, however, *offset* is also influenced by errors caused by B1 inhomogeneities. More specifically *offset* equals the mean vertical offset from the measured data point to the fitted function:

$$offset_{calc} = \frac{\sum_{i=1}^{i=ETL} (m_i - r_i)}{N} \tag{3}$$

where r_i is the signal from the fitted curve (using Eq 1) at the echo i , m_i is the measured or simulated signal, and N is the number of echoes.

To illustrate this, we generated a MSE signal curve based on EPG using the following parameters: $T_2 = 100\text{ms}$, $ESP = 20\text{ms}$, $ETL = 24$ (Fig 1). FA is varied from 100° to 180° in steps of 10. The curve is then fitted as described above using Eq 2 to produce $offset_{fitted}$. In a next step, $offset_{calc}$ is calculated according to Eq 3 using the previously determined T_2 and S_0 .

Table 1 shows the results of these calculations. It can be seen, that although the simulated data converges to 0 for all FAs (Fig 1), the offset increases with decreasing FA. Furthermore $offset_{fitted}$ equals $offset_{calc}$ for all FA. This shows that the mean of the vertical offsets from the determined signal to the actual signal over all echoes substantially influences the offset. The main contribution to the offset, when it is of a higher magnitude, comes from the errors in the early echoes if the pulse angle is not a perfect 180° . This can be seen from Fig 2, where the first echo shows by far the largest difference from the ideal curve (at 180°) if measured at 120° . The error, which is also seen as an oscillation between odd and even echoes, decreases with increasing TE. It should also be noted that if noise was included in this signal the offset itself would not result as zero even for perfect refocusing due to the noise floor.




Fig 1. Illustration of a EPG derived curve with an FA of 120° (x) and correct 180° FA (line) for a $T_2 = 100\text{ms}$, $T_1 = 3000\text{ms}$, $ESP = 20\text{ms}$ and ETL of 24. It can be seen that due to the incorrect FA the first echo point is lower than that of the second and the signal seems to oscillate between odd and even echoes.

doi:10.1371/journal.pone.0145255.g001

Table 1. Fitted results ($T2_{fitted}$, $S0_{fitted}$, $Offset_{fitted}$) from EPG simulated data of different FAs ($T2 = 100ms$, $ESP = 20ms$, $ETL = 24$, $S_0 = 1000$ a.u.). $Offset_{calc}$ is the back calculated vertical offset (Eq 3).

FA (°)	$T2_{fitted}$ (ms)	$S0_{fitted}$ (a.u)	$offset_{fitted}$ (a.u)	$offset_{calc}$ (a.u)
100	156.45	689.03	-20.61	-20.61
110	139.67	753.11	-14.86	-14.86
120	127.23	812.74	-10.41	-10.41
130	117.90	866.34	-6.92	-6.92
140	110.97	912.47	-4.25	-4.25
150	105.95	949.86	-2.26	-2.26
160	102.56	977.42	-0.89	-0.89
170	100.61	994.31	-0.16	-0.16
180	100.00	1000.00	0.00	0.00

doi:10.1371/journal.pone.0145255.t001

Comparison of different fitting methods: Simulation based on EPG

For EPG simulations using 180° as refocusing angle, $T2$ and S_0 determined after fitting matched the inputs for $T2$ and S_0 exactly for all fitting methods. With method (2) and (4), where the offset is a fitting parameter, offset was 0 for all ETL an ESP. Thus, the plots for $FA = 180^\circ$ have not been included in Figs 3–5.

Fig 3 presents the relative deviation $\delta T2$ of the fitted $T2$ for the different fitting methods. There is a distinct increase of $\delta T2$ with decreasing FA; $\delta T2$ at 120° is roughly a factor of 10 higher than $\delta T2$ at 160°. Furthermore, $\delta T2$ decreases for all FA, if the first echo is discarded (fitting methods (3) and (4)). The degree of reduction in $\delta T2$, however, depends on ETL and ESP: The lowest reduction in $\delta T2$ is seen for low ESP (method 3) and for low ESP and high ETL (method 4). With method 1, $\delta T2$ is highest for high ESP, whereas it is lowest for high ESP with method 3. In contrast, $\delta T2$ is lowest for high ETL and low ESP when using method 2, whereas it is highest at for high ETL and low ESP with method 4. While the introduction of the offset as fitting parameter does even slightly increase $\delta T2$ when using all echoes (method 2 vs. method 1), it leads to a gross decrease of $\delta T2$ over the majority of ESP/ETL combinations (method 4 vs. method 3). This decrease is most pronounced for low to intermediate ETL.

Fig 4 presents the relative deviation δS_0 of the fitted S_0 for the different fitting methods. As with $\delta T2$, δS_0 increases markedly with decreasing FA. Discarding the first echo leads to a more accurate determination of S_0 (i.e. δS_0 decreases) for high to intermediate ESP, whereas there is little or no reduction of δS_0 for low ESP. Other than with $\delta T2$, neither ETL nor the use of the offset as fitting parameter have a distinct influence on the accuracy of S_0 .




Fig 2. EPG simulated curve for $T2 = 50ms$, $T1 = 3000ms$, $ESP = 20$, $ETL = 24$, and $FA = 120^\circ$. Fig 2A illustrates the difference (blue solid line) between the curve simulated at 120° (black dashed line) an optimal 180° pulse (red solid line). Fig 2B illustrates the envelope (black lines) calculated from the odd and even echoes of the signal (blue line).

doi:10.1371/journal.pone.0145255.g002




Fig 3. Relative T2 deviation, dT2 (in %), for method 1 (first row), method 2 (second row), method 3 (third row) and method 4 (fourth row). These results are presented for 3 different FA. It is seen that T2 becomes longer as the FA decreases. Closer approximation to the actual T2 are seen when Eq 2 is used and the first point is excluded. Please note that the scales of dT2 are not uniform provide maximum dynamic range for the different ETL and ESP.

doi:10.1371/journal.pone.0145255.g003

Fig 5 presents the determined offsets for the fittings with and without the first echo. Again, it is clear to see that the offset is heavily influenced by the FA: The more the FA deviates from 180°, the higher the offset. The highest offset is seen for low ETL and low ESP. Discarding the first echo has little effect on the absolute value of the offset other than sign reversal: Offset is negative, if the first echo is used (Fig 5, upper row), whereas it is positive, if the first echo is discarded (Fig 5., lower row). This shows, that the large error coming from the relative signal increase of the second echo with respect to the first echo (cf. Fig 2A), which is highest for combination of low ETL and ESP, has a huge influence on the offset and leads to an overcompensation, if method (2) is used for fitting and the first echo is not discarded. If the first echo is discarded, however, the offset helps to compensate for the oscillation between odd and even echoes. Therefore, using the offset as fitting parameter and discarding the first echo is particularly helpful for accurate determination of T2 if short echo trains are used (i.e. low ETL and low ESP).




Fig 4. Relative S_0 deviation, dS_0 (in %), for methods 1–4 corresponding to rows 1–4 respectively. Closer approximations with little difference are seen for both equations with the first point excluded from the fit. Please note that the scales of dS_0 are not uniform and therefore provide maximum dynamic range for the different ETL and ESP.

doi:10.1371/journal.pone.0145255.g004

Comparison of different fitting methods: Influence of different T2 times and noise

Method four was shown to have the closest approximation to the actual T2 time for $T_2 = 100\text{ms}$. In order to evaluate whether this result is still holds for different T2s the simulation where rerun. Table 2 presents the results for each method for the different T2. The results show again that method four, i.e. discarding the first echo and including the offset as fitting parameter, yield a closer approximation in comparison to the other techniques.

When adding Rican noise to the signal prior to T2 fitting, the offset (method 4) increased for all T2, as both imperfect RF pulses and noise floor contribute to the offset in this case (Table 3). However, the performance of the four fitting methods (measured by δT_2) was in the same order as for the noise-free simulations with method 4 performing most accurately and being superior to the other methods1 (Table 3). While the statistical error of all fitting methods is in the same range for low T2, methods 2 and 4 are more prone to statistical errors at high T2.




Fig 5. Determined offset values for different FAs for method 2 (top row) and method 4 (bottom row). Distributions remain relatively similar with decreasing values as FA tends to 180°. Note the decreasing offset with increasing FA.

doi:10.1371/journal.pone.0145255.g005

This can be explained by the fact, that for long T2 the noise floor is not reached within the echo train, which increases the uncertainty in estimating offset and T2.

Comparison of different fitting methods: Phantom measurements

Signal decay in the phantom experiments (Fig 6) corresponded closely to the findings of the simulated work (Fig 2). Fitting of the data from the spin echo sequences (see Supporting Information, S1 Dataset: SE mapping) yielded a mean T2 of 67.7 ± 0.60 ms over three regions of interest (ROI) and over the four methods. Table 4 presents the mean T2 values and the relative deviation from T2, of three ROIs, determined from the spin echo sequences for all four fitting methods. Furthermore, for methods (2) and (4) the offset parameter is given.

Corresponding to our findings in the EPG simulations the relative deviation of T2 of the MSE from the reference value of the spin echo sequences increased with decreasing FA. Discarding the first echo reduced the relative T2 deviation for all FA. Relative T2 deviation was further reduced, if the offset was used as a fitting parameter (method 4). However, even for an 180° pulse as flip angle there still was a systematic error between 3.5% (method 1) and 3.6% (method 4) which is most likely due to an imperfect RF setup, B1 variations within the slice

Table 2. Mean results of T2 fitting for each method per T2 time for an ESP time of 5ms (ETL = 32, FA = 140°). Simulations were ran 1000 times.

Method 1			Method 2		
T2(ms)	T2 _{fitted} (ms)	dT2 (%)	T2 _{fitted} (ms)	dT2 (%)	Offset
20	21.98	9.90	22.16	10.79	-1.67
60	64.54	7.57	66.22	10.37	-8.95
100	107.29	7.29	112.92	12.92	-25.04
Method 3			Method 4		
T2(ms)	T2 _{fitted} (ms)	dT2 (%)	T2 _{fitted} (ms)	dT2 (%)	Offset
20	20.52	2.60	20.28	1.41	1.91
60	63.03	5.05	61.10	1.83	9.95
100	105.07	5.07	99.42	0.58	24.85

doi:10.1371/journal.pone.0145255.t002

Table 3. Mean results of simulated signal with added noise fittings for each method per T2 time for an ESP of 5ms (ETL = 32, FA = 140°). Simulations were ran 1000 times.

Method 1			Method 2		
T2(ms)	T2 _{fitted} (ms)	dT2 (%)	T2 _{fitted} (ms)	dT2 (%)	Offset
20	22.06 ± 0.37	10.29 ± 1.82	21.70 ± 0.41	8.49 ± 2.05	3.40
60	64.51 ± 0.63	7.51 ± 1.05	66.31 ± 1.93	10.51 ± 3.21	-9.61
100	107.32 ± 1.01	7.32 ± 1.01	113.19 ± 5.58	13.1 ± 5.559	-26.11
Method 3			Method 4		
T2(ms)	T2 _{fitted} (ms)	dT2 (%)	T2 _{fitteds} (ms)	dT2 (%)	Offset
20	20.61 ± 0.45	3.05 ± 1.96	19.74 ± 0.5	1.30 ± 1.68	7.04
60	63.01 ± 0.66	5.01 ± 1.10	61.2 ± 1.943	2.05 ± 2.30	9.13
100	105.10 ± 1.05	5.10 ± 1.04	99.6 ± 4.884	0.36 ± 2.92	23.97

doi:10.1371/journal.pone.0145255.t003

profile, or gross B1 inhomogeneity, which is known to occur particularly at high field strength [29].

Discussion

Mono-exponential fitting is not the proper method for data fitting in T2 relaxometry due to its known inaccuracies in presence of B1 inhomogeneities. As mono-exponential fitting methods, however, are used in the majority of clinical and preclinical studies, our aim was to minimize the error in T2 by two modifications of mono-exponential fitting, that can be easily performed, are not time-intensive and have already been proposed: introducing an offset and/or discarding the first echo. By simulations based on EPG theory we could show, that B1 inhomogeneities and imperfect refocusing pulse angles provide major contributions to the *offset* fitting parameter in T2 relaxometry, while past published work pointed to the *offset* as being a compensation




Fig 6. Example of the signal decay curves of the phantom measurements for a defined pixel (shown by the cross section of the lines): Curves are shown for the single spin echo sequence (top right) and the MSE sequences for different FAs (bottom row). The X-axis represents TE and y-axis the signal ($\times 10^8$ a.u.). It can be noticed that as the FA reduces the first point, in particular, deviates from the expected exponential decay curve. The variation of the refocusing FA was performed by variation of the FA in the sequence protocol. The actual FA at the respective position might even differ from this value due to B1 inhomogeneities and imperfect slice profiles.

doi:10.1371/journal.pone.0145255.g006

Table 4. Fitted results from phantom measurements a MSE sequence with different FAs as used in the sequence protocol (see Supporting Information, S1 Dataset: MSE datasets) and different fitting methods. Single spin echo sequence yielded a mean T2 of 67.7 ± 0.60ms. T2 and dT2 relate to the mean and standard deviation over 3 ROIs within the phantom for each measurement.

FA	Method 1		Method 2		
	T2 (ms)	dT2 (%)	T2 (ms)	dT2 (%)	Offset (a.u)
180°	70.7 ± 1.0	3.5 ± 0.8	69.4 ± 1.2	1.9 ± 0.8	33949.6
160°	72.2 ± 0.9	5.6 ± 1.1	70.9 ± 1.1	4.1 ± 0.9	28239.8
140°	77.0 ± 0.7	12.7 ± 1.4	76.0 ± 0.9	11.6 ± 1.4	36127.6
120°	84.7 ± 0.6	24.0 ± 1.9	84.5 ± 0.8	24.1 ± 1.9	37770.0
FA	Method 3		Method 4		
	T2 (ms)	dT2 (%)	T2 (ms)	dT2 (%)	Offset (a.u)
180°	70.6 ± 1.0	4.6 ± 1.0	69.3 ± 1.1	3.6 ± 0.7	27230.2
160°	71.3 ± 0.9	5.5 ± 1.0	68.9 ± 1.2	2.8 ± 0.8	33127.8
140°	73.7 ± 0.8	9.1 ± 1.0	71.1 ± 1.1	6.2 ± 0.8	30207.4
120°	78.2 ± 0.8	15.7 ± 1.3	74.5 ± 1.1	11.3 ± 1.0	21556.9

doi:10.1371/journal.pone.0145255.t004

factor for either long T2 time within a compartment [13, 14, 19] or a baseline for noise and other system related signal [8, 20]. As the refocusing flip angle deviates from an optimal 180° pulse, the offset value increases. This can be explained by the fact, that the offset as a fitting parameter does only equal the asymptote of the signal curve, if the early echoes, i.e. the echoes at low TE, match the exponential decay curve exactly. If this is not the case, as for the example of oscillating between odd and even echoes, the vertical displacement from the perfect exponential decay curve contributes to the *offset*. The *offset* is particularly high for low ESP and low ETL, a setting which is often used particularly in T2 relaxometry in humans, as clinical MRI scanners usually limit the number of echoes and the user often reduces ESP to keep TR and thus, the acquisition time, as short as possible, and maintain the integrity of the curve for shorter T2 times (as measured in brain, muscle etc.).

Although we show that the *offset* is predominately due to an error in the refocusing FA, this does not mean, that noise and long T2 components do not play a role. Long T2 components, due to several proton pools with different T2 times or due to partial volume effects, always influence the determination of T2 and the *offset*, in addition to the effects of imperfect B1. If the longer T2 component is not covered by the entire ETL, then the *offset* would be higher. One could check for possible bi-exponential curves using a statistical Fχ test [30] and then fit the data as a bi-exponential function including the *offset* as an additional parameter [31].

Noise may also play a role in the final value of the offset, particularly in the lower values where there is a Rican and or Rayleigh distribution [32, 33]. It would be of interest to investigate the effect of SNR with the offset as a free parameter in the fitting.

Although the fitting offset parameter is strongly influenced by imperfections of the refocusing pulse angle, we found that using the *offset*, within method 2, as an additional fitting parameter does not lead to less systematic error in T2. Quite contrary, the systematic error in T2 increases by adding *offset* as fitting parameter, particularly for low ESP and low ETL, where the *offset* is high.

We have identified the signal oscillations in the early echoes as a major contribution to the *offset* and as a major source of error in T2 quantification. One easy way to account for large portions of this error is to discard the first echo for curve fitting [34–36]. We, therefore, included this procedure as one possible post-processing method (method 3 in this publication) and found, that this method actually reduces the systematic error in T2, but most efficiently for high ESP. With low ESP, however, there is only a slight reduction in error, as the deviation of

the first echo from the exponential signal decay curve is most prominent for high ESP as this allows for more mixing of longitudinal and transverse components at imperfect refocusing pulses.

It is important to note that the method of discarding the first echo will not be able to compensate for the systematic error in T2 completely, as all echoes are affected by imperfect refocusing pulses. Several authors have put forward that if there is a discrepancy in the refocusing FA, all odd echoes will not be completely refocused and all even echoes will refocus all isochromats into the transverse plane [37, 38]. From this they deduced the method to use only even echoes to determine the T2 time. Although this method allows one to get rid of large parts of the signal error, it does not eliminate the systematic error in T2 completely: As can be seen from the signal decay calculated with EPG theory in Fig 1, it is clear to see that although the curve with a true 180° pulse starts by following the even echoes, the curve matches closer to the odd echoes at the end. There is a mixing of the transverse and longitudinal components of the preceding echoes due to incorrect refocusing. That is, if the isochromats are not correctly refocused they will influence the subsequent echoes. This was further pointed out by Maudsley et al. [39].

We could show, that most of the remaining systematic error in T2 after discarding the first echo can be eliminated by using the offset as a *fitting* parameter (method 4). This method works exceptionally well with low to intermediate ETL: δT_2 was less than 6% for all ESP and all $ETL \leq 32$, even for a refocusing pulse of 120°. As in clinical as well as in experimental settings, most often relatively short echo train lengths are used, method 4 provides an easy-to-use, fast and reliable method to correct for B1 imperfections. Although it is thought that long ETL should be used to gain the most accurate determination of T2 [17] it is seen from Fig 3 that using long echo trains can be detrimental to the accuracy of T2, particularly when using method 4.

Although using the *offset* as a fitting parameter and discarding the first echo can minimize the error in T2, it should be pointed out that this method does not compensate errors in the refocusing pulse completely and further correction of the T2 may be of interest [12, 40]. In the first place, B1 inhomogeneities should always be reduced by advanced acquisition techniques, e.g. by using a larger spatial width of the refocusing pulse compared to the excitation pulse [14] or by parallel transmission [41]. A more recent example of a postprocessing algorithm to correct for the remaining B1 inhomogeneity would be the method introduced by Neumann et al. [11]. In this method the authors provide a means of correcting T2 with the use of a heuristic formula based on a lookup table derived from simulated EPG data. Although this method allows one to significantly reduce the systematic error in T2, there are several drawbacks: It needs much more computing time than the method introduced by us and is restricted to a limited choice of ETL (16, 24, and 32).

In conclusion, we have shown that the *offset* in T2 relaxometry is influenced by imperfect refocusing pulses. Using the *offset* as a fitting parameter and discarding the first echo is a fast and easy method to minimize the error in T2, particularly for low to intermediate echo train length. We expect that by identifying the optimal method for T2 determination, variations in reported times for specific tissue types and pathologies will be minimized, thus achieving better results for T2 as a quantitative measure.

Supporting Information

S1 Dataset. DICOM data sets from MRI investigations (Fig 6 and Table 4).
(ZIP)

Author Contributions

Conceived and designed the experiments: DM NR SH. Performed the experiments: DM NR. Analyzed the data: DM. Contributed reagents/materials/analysis tools: DM NR SH. Wrote the paper: DM NR MB.

References

1. Damadian R. Tumor detection by nuclear magnetic resonance. *Science*. 1971; 171(3976):1151–3. PMID: [5544870](#)
2. Hahn EL. Spin Echoes. *Physical Review*. 1950; 80(4):580–94.
3. Harms SE, Siemers PT, Hildenbrand P, Plum G. Multiple spin echo magnetic resonance imaging of the brain. *Radiographics*. 1986; 6(1):117–34. PMID: [3685481](#)
4. Meiboom S, Gill D. Modified Spin Echo Method for Measuring Nuclear Relaxation Times. *Review of Scientific Instruments*. 1958; 29(8):688–91.
5. Majumdar S, Orphanoudakis SC, Gmitro A, O'Donnell M, Gore JC. Errors in the measurements of T2 using multiple-echo MRI techniques. I. Effects of radiofrequency pulse imperfections. *Magn Reson Med*. 1986; 3(3):397–417. PMID: [3724419](#)
6. Zelaya FO, Roffmann WU, Crozier S, Teed S, Gross D, Doddrell DM. Direct visualisation of B1 inhomogeneity by flip angle dependency. *Magn Reson Imaging*. 1997; 15(4):497–504. PMID: [9223051](#)
7. Dietrich O, Reiser MF, Schoenberg SO. Artifacts in 3-T MRI: physical background and reduction strategies. *Eur J Radiol*. 2008; 65(1):29–35. PMID: [18162353](#)
8. Petrovic A, Scheurer E, Stollberger R. Closed-form solution for T2 mapping with nonideal refocusing of slice selective CPMG sequences. *Magn Reson Med*. 2015; 73(2):818–27. doi: [10.1002/mrm.25170](#) PMID: [24634257](#)
9. Akhondi-Asl A, Afacan O, Mulkern RV, Warfield SK. T2-relaxometry for myelin water fraction extraction using wald distribution and extended phase graph. *Med Image Comput Assist Interv*. 2014; 17(Pt 3):145–52. PMID: [25320793](#)
10. Majumdar S, Gmitro A, Orphanoudakis SC, Reddy D, Gore JC. An estimation and correction scheme for system imperfections in multiple-echo magnetic resonance imaging. *Magn Reson Med*. 1987; 4(3):203–20. PMID: [3574056](#)
11. Neumann D, Blaimer M, Jakob PM, Breuer FA. Simple recipe for accurate T2 quantification with multi spin-echo acquisitions. *MAGMA*. 2014; 27(6):567–77. doi: [10.1007/s10334-014-0438-3](#) PMID: [24643838](#)
12. Lebel RM, Wilman AH. Transverse relaxometry with stimulated echo compensation. *Magn Reson Med*. 2010; 64(4):1005–14. doi: [10.1002/mrm.22487](#) PMID: [20564587](#)
13. Pell GS, Briellmann RS, Waites AB, Abbott DF, Jackson GD. Voxel-based relaxometry: a new approach for analysis of T2 relaxometry changes in epilepsy. *Neuroimage*. 2004; 21(2):707–13. PMID: [14980573](#)
14. Pell GS, Briellmann RS, Waites AB, Abbott DF, Lewis DP, Jackson GD. Optimized clinical T2 relaxometry with a standard CPMG sequence. *J Magn Reson Imaging*. 2006; 23(2):248–52. PMID: [16416434](#)
15. Duncan JS, Bartlett P, Barker GJ. Technique for measuring hippocampal T2 relaxation time. *AJNR Am J Neuroradiol*. 1996; 17(10):1805–10. PMID: [8933861](#)
16. Duncan JS, Bartlett P, Barker GJ. Measurement of hippocampal T2 in epilepsy. *AJNR Am J Neuroradiol*. 1997; 18(9):1791–2. PMID: [9367334](#)
17. Whittall KP, MacKay AL, Li DK. Are mono-exponential fits to a few echoes sufficient to determine T2 relaxation for in vivo human brain? *Magn Reson Med*. 1999; 41(6):1255–7. PMID: [10371459](#)
18. Kosior RK, Lauzon ML, Federico P, Frayne R. Algebraic T2 estimation improves detection of right temporal lobe epilepsy by MR T2 relaxometry. *Neuroimage*. 2011; 58(1):189–97. doi: [10.1016/j.neuroimage.2011.06.002](#) PMID: [21689766](#)
19. Pell GS, Waites A, Briellmann RS, Jackson GD, editors. Voxel-based relaxometry. *Proceedings of the ISMRM, 11th Annual Meeting; 2003; Toronto, Canada*.
20. Carneiro AAO, Vilela GR, de Araujo DB, Baffa O. MRI relaxometry: methods and applications. *Brazilian Journal of Physics*. 2006; 36(1a):9–15.
21. Feng Y, He T, Gatehouse PD, Li X, Harith Alam M, Pennell DJ, et al. Improved MRI R2* relaxometry of iron-loaded liver with noise correction. *Magn Reson Med*. 2013; 70(6):1765–74. doi: [10.1002/mrm.24607](#) PMID: [23359410](#)

22. Wood JC, Enriquez C, Ghugre N, Tyzka JM, Carson S, Nelson MD, et al. MRI R2 and R2* mapping accurately estimates hepatic iron concentration in transfusion-dependent thalassemia and sickle cell disease patients. *Blood*. 2005; 106(4):1460–5. PMID: [15860670](#)
23. Hennig J. Multiecho imaging sequences with low refocusing flip angles. *Journal of Magnetic Resonance (1969)*. 1988; 78(3):397–407.
24. Hennig J. Echoes—how to generate, recognize, use or avoid them in MR-imaging sequences. Part I: Fundamental and not so fundamental properties of spin echoes. *Concepts in Magnetic Resonance*. 1991; 3(3):125–43.
25. Hennig J. Echoes—how to generate, recognize, use or avoid them in MR-imaging sequences. Part II: Echoes in imaging sequences. *Concepts in Magnetic Resonance*. 1991; 3(4):179–92.
26. Weigel M. Extended phase graphs: Dephasing, RF pulses, and echoes—pure and simple. *J Magn Reson Imaging*. 2014.
27. Kanzow C, Yamashita N, Fukushima M. Levenberg-Marquardt methods for constrained nonlinear equations with strong local convergence properties. *Journal of Computational and Applied Mathematics* 2004; 170:375–97.
28. Lourakis MIA. Levenberg-Marquardt Nonlinear Least Squares Algorithms in C/C++ www.ics.forth.gr/~lourakis/levmar/. Available: www.ics.forth.gr/~lourakis/levmar/.
29. Andreychenko A, Bluemink JJ, Raaijmakers AJ, Lagendijk JJ, Luijten PR, van den Berg CA. Improved RF performance of travelling wave MR with a high permittivity dielectric lining of the bore. *Magn Reson Med*. 2013; 70(3):885–94. doi: [10.1002/mrm.24512](#) PMID: [23044511](#)
30. Bevington PD, Robinson DK, editors. *Data reduction and error analysis for the physical sciences*. Third Edition ed: McGraw-Hill; 2003.
31. Armspach JP, Gounot D, Rumbach L, Chambron J. In vivo determination of multiexponential T2 relaxation in the brain of patients with multiple sclerosis. *Magn Reson Imaging*. 1991; 9(1):107–13. PMID: [2056848](#)
32. Gudbjartsson H, Patz S. The Rician distribution of noisy MRI data. *Magn Reson Med*. 1995; 34(6):910–4. PMID: [8598820](#)
33. St Pierre TG, Clark PR, Chua-Anusorn W. Single spin-echo proton transverse relaxometry of iron-loaded liver. *NMR Biomed*. 2004; 17(7):446–58. PMID: [15523601](#)
34. Maier CF, Tan SG, Hariharan H, Potter HG. T2 quantitation of articular cartilage at 1.5 T. *J Magn Reson Imaging*. 2003; 17(3):358–64. PMID: [12594727](#)
35. Mosher TJ, Dardzinski BJ, Smith MB. Human articular cartilage: influence of aging and early symptomatic degeneration on the spatial variation of T2—preliminary findings at 3 T. *Radiology*. 2000; 214(1):259–66. PMID: [10644134](#)
36. Smith HE, Mosher TJ, Dardzinski BJ, Collins BG, Collins CM, Yang QX, et al. Spatial variation in cartilage T2 of the knee. *J Magn Reson Imaging*. 2001; 14(1):50–5. PMID: [11436214](#)
37. Thorpe JW, Barker GJ, Jones SJ, Moseley I, Losseff N, MacManus DG, et al. Magnetisation transfer ratios and transverse magnetisation decay curves in optic neuritis: correlation with clinical findings and electrophysiology. *J Neurol Neurosurg Psychiatry*. 1995; 59(5):487–92. PMID: [8530932](#)
38. Tofts P. *Quantitative MRI of the Brain: Measuring Changes Caused by Disease*. 1 ed. Tofts P, editor: Wiley and sons Ltd; 2005. p. 143–201 p.
39. Maudsley AA. Modified Carr-Purcell-Meiboom-Gill sequence for NMR Fourier imaging applications. *Journal of Magnetic Resonance (1969)*. 1986; 69(3):488–91.
40. Prasloski T, Madler B, Xiang QS, MacKay A, Jones C. Applications of stimulated echo correction to multicomponent T2 analysis. *Magn Reson Med*. 2012; 67(6):1803–14. doi: [10.1002/mrm.23157](#) PMID: [22012743](#)
41. Schar M, Ding H, Herzka DA. Improvement in B1+ Homogeneity and Average Flip Angle Using Dual-Source Parallel RF Excitation for Cardiac MRI in Swine Hearts. *PLoS One*. 2015; 10(10):e0139859. doi: [10.1371/journal.pone.0139859](#) PMID: [26436658](#)