

Axially decoupled photo-stimulation and two photon readout (ADePT) for mapping functional connectivity of neural circuits

Matthew Koh^{1,2,#}, Francesca Anselmi^{2,#}, Sanjeev K. Kaushalya^{2,#}, Diego E. Hernandez², Walter Germán Bast², Pablo S. Villar², Honggoo Chae², Martin B. Davis², Sadhu Sai Teja², Zhe Qu³, Viviana Gradinaru³, Priyanka Gupta^{2,§}, Arkarup Banerjee^{1,2,§} & Dinu F. Albeanu^{1,2,§}

1. CSHL School for Biological Sciences
2. Cold Spring Harbor Laboratory, Cold Spring Harbor, NY, 11724, USA
3. California Institute of Technology, Pasadena, CA, 91125

- equal contribution

§ - Correspondence to: albeanu@cschl.edu; banerjee@cschl.edu; pgupta@cschl.edu

Key words: *patterned photo-stimulation, digital micro-mirror device, holographic diffuser, two photon microscopy, all-optical physiology, 3-D functional connectivity mapping, olfactory bulb, sister mitral/tufted cells, short axon cells*

Abstract

All optical physiology *in vivo* provides a conduit for investigating the function of neural circuits in 3-D. Here, we report a new strategy for flexible, axially-decoupled photo-stimulation and two-photon readout (*ADePT*) of neuronal activity. To achieve axially-contained widefield optogenetic patterned stimulation, we couple a digital micro-mirror device illuminated by a solid-state laser with a motorized holographic diffuser. In parallel, we use multiphoton imaging of neural activity across different z-planes. We use *ADePT* to analyze the excitatory and inhibitory functional connectivity of the mouse early olfactory system. Specifically, we control the activity of individual input glomeruli on the olfactory bulb surface, and map the ensuing responses of output mitral and tufted cell bodies in deeper layers. This approach identifies cohorts of *sister* mitral and tufted cells, whose firing is driven by the same parent glomerulus, and also reveals their differential inhibition by other glomeruli. In addition, selective optogenetic activation of glomerular GABAergic/dopaminergic (DAT+) interneurons triggers dense, but spatially heterogeneous suppression of mitral and tufted cell baseline activity and odor responses, further demonstrating specificity in the inhibitory olfactory bulb connectivity. In summary, *ADePT* enables high-throughput functional connectivity mapping in optically accessible brain regions.

Introduction

Manipulating select neural circuit elements is critical for understanding the circuit mechanisms underlying brain computation, especially when coupled with activity monitoring and behavioral analysis. In recent years, optogenetic control of specific cell-types has been widely adopted as a method of choice^{1–5}, and various combinations of optical imaging and spatial light patterning techniques^{6–19} have achieved concurrent stimulation and activity recording, typically in the same optical z-plane. This is primarily constrained by the microscope anatomy, as the photo-stimulation and imaging light paths converge onto the sample through a shared microscope objective which focuses both the stimulation and imaging light patterns in its fixed focal plane. Neuronal circuits are however distributed in 3-D and investigating the functional connectivity of such neural circuits requires flexible control and readout of neural activity in independent optical planes (e.g. stimulating layer 1 dendrites and monitoring neuronal activity in layers 5, **Fig. 1a**).

Several strategies have been developed over the past two decades to target multiple optical z-planes simultaneously, or in quick succession. These range from dual scanning head implementations¹⁹ to using tunable focusing elements (e.g. electrically-tunable lenses, ETL) whose focal length can be dynamically controlled^{20–22}, acousto-optic deflector (AOD)-based arbitrary access imaging systems^{14,23–28} and digital holography^{8,9,13,27,29–33}. These methods have varying trade-offs, and to date, monitoring and photo-activating neurons in three dimensions remains challenging. When choosing an adequate photo-stimulation and imaging approach, several aspects are worth considering: the required spatial resolution in 3-D (stimulating individual cell bodies versus extended patches of tissue), the size of the field of stimulation, the axial separation between the stimulation and imaging planes (relative locations of the input-output nodes

of the circuit), the cost and simplicity of implementation and its robustness. Early scanning-based methods enabled photo-stimulation over large fields of view (FOVs), but conferred poor axial resolution^{34,35}, potentially leading to off target photo-activation in 3-D. In contrast, modern methods achieved flexible and precise 3-D confinement of photo-stimulation using AOD-based approaches, or digital holography in combination with temporal focusing^{8,9,13,27–33}. AOD-based and holographic control of activity, however, can be technically challenging and cost-prohibitive, and the size of the field of photo-stimulation is limited by steep power requirements. Axial decoupling between the stimulation and imaging planes has also been achieved by simpler methods such as using ETLs, but these implementations are constrained by spherical aberrations which deteriorate the point spread function and limit the size of the field of stimulation.

Here we introduce *ADePT* (Axially Decoupled Photo-stimulation and Two-Photon readout), an easy implementable, cost-effective and high-throughput technique that combines one-photon widefield patterned illumination with multiphoton imaging in axially decoupled planes (**Fig. 1b**). Our approach enables precise axially-confined photo-stimulation, with arbitrarily shaped light patterns, across an extended region on the brain surface, while imaging neuronal activity in other optically accessible z-planes of choice. We use a digital micro-mirror device (DMD)^{17,36–46} to impose customizable 2-D spatial patterns onto a Gaussian laser beam that are subsequently optically relayed onto the sample. Due to the coherent nature of the illumination source (laser), the spatial patterning is maintained throughout the optical path between the DMD chip and the sample (infinite focus). Placing a (holographic) diffuser in the optical path breaks the coherence and disrupts this infinite focus. The desired 2-D pattern comes into focus only in an axially confined optical-plane in the sample that is conjugate to the diffuser. To flexibly target photo-stimulation

across different depths in the sample, independent of the multiphoton imaging plane (objective's focal plane), we simply translate the diffuser along the optical path using a motorized stage. Akin to remote focusing imaging approaches^{16,20,47–52}, translating the diffuser displaces the optical z-plane within the sample where the stimulation pattern is in focus independent of the objective's focal plane (**Fig. 1c**). In our implementation, we report ~20 μm lateral (17.1 μm x-y-FWHM) and ~40 μm axial (36.3 μm z-FWHM) resolution of the patterned photo-stimulation across a 1.2 x 1.5 mm field of stimulation (20X, 1.0 NA, Olympus objective). We calibrated this method across up to 500 μm axial displacement between the photo-stimulation plane and the focal plane of the objective. Note that these parameters are not strictly rigid, and *ADePT* can accommodate many different operating configurations depending upon the specific biological application.

We subsequently assessed the effectiveness of *ADePT* for investigating the logic of circuit connectivity by leveraging the 3-D spatial layout of the mouse olfactory bulb (OB)^{53,54}. In the OB, the input nodes called glomeruli (balls of neuropil sorted by odorant receptor identity) and the principal output neurons (mitral and tufted cells) are separated by approximately two hundred microns in depth. We investigated the excitatory and inhibitory logic of this early olfactory circuit by photo-stimulating either the glomeruli or local inhibitory interneurons while monitoring responses of mitral/tufted cells (MT cells). We found that *sister* OB output neurons^{36,55–65} – those that receive primary input from the *same* 'parent' glomerulus – differ substantially in their lateral inhibitory connectivity with other glomeruli with the OB network, potentially explaining their non-redundant responses to odors^{36,55}. Further, we found that local GABAergic/dopaminergic interneurons^{42,43,66–74} associated with different glomeruli provide dense, but selective inhibition to regulate odor responses of OB output neurons. In summary, *ADePT* is readily-deployable and well-

suited for high-throughput mapping of both excitatory and inhibitory functional connectivity in optically accessible brain regions.

Results

Design of the axially decoupled patterned photo-stimulation and imaging setup

We couple a custom laser scanning two photon microscope (Methods) with a modular add-on for axially-contained widefield (one-photon) patterned photo-stimulation (**Fig. 1b**). Patterned photo-stimulation is achieved using a digital micro-mirror device (VX4100 DMD, 0.7" XGA VIS 1024 x 768 pixels, 13.68 μm micro-mirror pitch, Vialux) illuminated by a solid-state CW laser (Sapphire, 488 nm, 200 mW, Coherent). The photo-stimulation and two-photon excitation paths are combined using a dichroic mirror (680dcxr, 50 mm diameter, Chroma) placed above the objective (20X, 1.0 NA, Olympus). A diffuser mounted on a motorized translation stage enables flexible control of the axial location of the photo-stimulation plane in the sample (**Figs. 1a-c**). Without the diffuser, the photo-stimulation 2-D pattern is in focus throughout the optical path as pencils of coherent light (**Fig. 1c, Top**). By placing the diffuser in the optical path, the coherence of the beam is disrupted and the 2-D spatial pattern imposed by the DMD is accurately reconstructed in the sample only in a thin (axially-contained) image plane. This image plane is optically conjugated with both the diffuser and the DMD chip. In the default 4f configuration, the DMD pattern thus comes into focus in the focal plane of the objective such that the photo-stimulation and multiphoton imaging planes overlap axially (**Fig. 1c, Middle**). Importantly, translating the diffuser along the optical path causes its conjugate image plane in the sample to move as well, therefore axially decoupling the photo-stimulation plane from the focal (sample) plane of the imaging objective (**Figs. 1b-c, Bottom, S1a-b**).

Introducing a diffuser in the optical path has the added benefit of improving the axial confinement of the photo-stimulation pattern. The diffuser breaks the coherence of the DMD-modulated laser pattern, dispersing light across a solid angle (Methods). As the photo-stimulation pattern in the sample plane is a conjugate image of the diffuser, it is expected to be restricted to a thin optical z-plane and its axial extent to scale linearly with the lateral (x-y) spot size^{6,12,75}. Note that this differs from the focusing regime of a Gaussian laser beam where the axial extent of the focal spot increases quadratically with the lateral extent of the spot. In *ADePT*, the axial confinement of the photo-stimulation pattern depends directly on the available light angles emerging from the diffuser (effective numerical aperture): the wider the cone of angles, the tighter the axial confinement (higher axial resolution) albeit with increased loss of excitation light.

Figure 1




Figure 1: Axially decoupled patterned photo-stimulation and multiphoton imaging. (a) (Top) Cartoon schematics of functional connectivity mapping of a neural circuit (e.g. olfactory bulb);

photo-stimulation (2-D light mask from a digital micro-mirror device illuminated with a solid-state CW laser) and two photon imaging are confined to different optical z-planes that can be flexibly and independently adjusted by translating the diffuser and respectively the primary objective; (Bottom) Alternating (strobing) between photo-stimulation and imaging periods. Each red bar represents a single frame of multiphoton imaging. Photo-stimulation and imaging periods are interleaved. (b) Microscope schematics. DM, dichroic mirrors. DMD, digital micro-mirror device. I, iris diaphragm. L1-L12, lenses. O, primary objective. PMT, photomultiplier tube. PS, periscope. S, shutter. SM, scan mirrors. (c) (Top) Illustration of using a movable diffuser to decouple the patterned photo-stimulation and multiphoton imaging planes. The diffuser is imaged into the sample, in a 4f lens configuration; translating the diffuser along the optical path causes the corresponding projection plane to shift axially. OFP, objective focal plane. PSP, photo-stimulation plane. (d) DMD chip-to-CCD camera-to-2p microscope registration. (i) two photon micrograph of 10 μm fluorescence microbeads; arrows mark two microbeads, part of the larger DMD-modulated projection target pattern (8 microbeads), which were taken as fiduciary points; (ii) widefield fluorescence image (full field illumination) of a larger field of view including the target microbeads; (iii) ROIs selected from the 2p image were used to generate DMD-chip light masks; these were further projected at the primary objective focal plane and imaged using the primary CCD camera (CCD 1); (iv) overlay of DMD-generated photo-stimulation masks and widefield fluorescence image of 10 μm microbeads from (ii); note that fluorescence is restricted to the microbeads targeted by the DMD photo-stimulation masks with minimal spillover to adjacent (off-target) microbeads (see fiduciary marks).

We registered the DMD field of stimulation (DMD pixel size = 2.4 μm at the sample, **Fig. 1d**) to the primary CCD camera image (CCD1 pixel size = 1.1 μm at sample; CCD 1300-QF, Vosskuhler) and the multiphoton field of imaging into a common reference frame (Methods). We verified that the targeting of DMD-modulated spatial light patterns was precise (< 5 μm error, note that fluorescence is restricted to the microbeads targeted by the DMD photo-stimulation masks with minimal spillover to adjacent, off-target microbeads, **Fig. 1d, 2b, S1d**). To ensure flat photo-stimulation probability across the field, we measured and further corrected light power inhomogeneities arising due to the Gaussian profile of the photo-stimulation laser beam and aberrations along the optical path (**Fig. S1e**). In control experiments (no fluorescent sample), we observed unwanted phosphorescence upon photo-stimulation, depending on the lens material (BK7 vs. fused silica). Consequently, we compared different objective and lens combinations and

chose the primary objective (20X, 1.0 NA, Olympus) and collection tube lens (L5; 100 mm FL, fused-silica, LA4545-A, Thorlabs) such as to minimize these artifacts (**Fig. S1f**, Methods).

Lateral and axial resolution of photo-stimulation

To determine the lateral and axial resolution of photo-stimulation setup, we projected light spots of increasing nominal diameters onto a glass coverslip spin-coated with 1 μ m thick rhodamine B layer (**Figs. 2,3**). We measured the resulting fluorescence signals by imaging the rhodamine-coated coverslip from below with an inverted (secondary) microscope. The secondary microscope was set up so as to be confocal with the primary microscope objective (**Figs. 2a-c, 3a**)⁶. Each spot was sampled through the fluorescent coverslip by moving the primary objective up or down in steps of 1-2 μ m. We used the images acquired by the primary (CCD1) and secondary (CCD2) cameras (**Fig. 3a**; CCD1300QF, Vosskuhler) to obtain the lateral (x-y) and axial (z) extent of the light intensity profile at full width half maximum (FWHM, **Figs. 2b-c, S2a-b**). Using this strategy, the smallest resolvable photo-stimulation spot we achieved was ~ 20 μ m in lateral diameter (17.1 μ m FWHM, **Figs. 2b, S2a-b**). Below this size, decreasing the nominal spot size did not change the measured x-y-FWHM of the spot, potentially due to light scattering from the extended spot, light loss and imperfections in the setup's optical transfer function.

We further investigated the effect of the choice of diffuser on the z-FWHM for projected light spots of different sizes. As reported previously for light patterning techniques such as digital holography^{6,12,75}, we observed a linear dependence between the axial and lateral extent of the light patterns projected (z-FWHM vs. x-y-FWHM of photo-stimulation spots, **Fig. 2c**). As expected, the ratio between the axial vs. lateral resolution was inversely proportional to the diffuser

divergence angle (slope = 1.52 when using a *grit ground* diffuser, #38-786 vs. 1.62 for a 10° *holographic* diffuser, #48-506 vs. 2.12 for a 5° *holographic* diffuser, #55-849, Edmund Optics, **Fig. 2c**). In other words, the axial spread of the projected light pattern increased as the diverging angle of the *holographic* diffuser got narrower (lower NA). To minimize excitation light loss while maintaining ~40 μm axial resolution, we settled on a 5° *holographic* diffuser, a diffractive optical element designed to limit the spread of transmitted light to a specified solid angle (e.g., 5°, #55-849, Edmund Optics, **Fig. 2c**). Using a 5° *holographic* diffuser only slightly under-filled the angular aperture of the primary objective and did not affect the lateral resolution of the system. Using a 10° *holographic* diffuser, or a *grit ground* diffuser would further increase the axial specificity (~30 μm) at the expense of light loss.

Figure 2




Figure 2. Axial confinement and resolution of the patterned stimulation and multiphoton imaging setup. (a) Axial confinement of light intensity of a 2-D projection pattern using a holographic diffuser conjugated to the focal plane of the principal objective (Left) vs. same optical configuration without a diffuser (Right); spot diameter = 50 μm . (b) Measured spot x-y FWHM versus expected nominal spot size calculated by multiplying the pattern size in pixels by the expected size of a pixel given the DMD chip and the optical magnification of the system. Circular spots of different sizes (10-100 μm in diameter) were projected on a glass slide coated with a 1 μm thin layer of rhodamine B. (c) Measured axial (z) FWHM versus lateral (x-y) FWHM of the projected spots from (b) for three different types of diffusers; black circles = 5° *holographic* diffuser; red circles = 10° *holographic* diffuser; blue circles = *grit ground* diffuser.

Decoupling imaging and photo-stimulation across different z-planes

To flexibly manipulate the relative positioning of the photo-stimulation plane with respect to the multiphoton imaging plane, we physically translate the diffuser on a motorized stage (**Fig. 1b**). Theoretically, we expect a linear relationship between the axial displacement of the objective and the corresponding compensatory displacement of the diffuser that is proportional to the square of the magnification of the optical system (Eqs. 1-5, Methods). To verify and experimentally calibrate the decoupling of photo-stimulation and imaging planes across a range of depths in our setup, we projected a checker-board light pattern on a fluorescent coverslip coated with a $\sim 1\ \mu\text{m}$ thin layer of rhodamine B. We imaged the resulting fluorescence through both the primary microscope and a secondary inverted widefield microscope placed below (**Fig. 3a**)⁶. In the default configuration, the diffuser is conjugate to the sample (objective focal plane), the secondary objective is confocal with the primary objective, and the rhodamine coverslip is placed in their shared focal plane (also the sample plane). The photo-stimulated pattern thus appears in focus on both CCD1 (primary objective) and CCD2 (secondary objective). As we lowered the primary objective, as one would to image deeper structures in a sample, the projected pattern also moved lower (below the coverslip). This can be observed as a blurred image of the photo-stimulation pattern on CCD2 as its imaging plane is still conjugate to the rhodamine coverslip. We determined the extent of the diffuser displacement necessary to re-focus the projected pattern back onto the original plane of interest in the sample (the rhodamine coverslip) by translating the diffuser away from the objective until a sharp image of the checker-board reappeared on CCD2 (**Fig. 3a**). Moving down the primary objective by incremental amounts ($100\ \mu\text{m}$) and monitoring the required diffuser displacement via the secondary microscope, we obtained a calibration curve, describing the effective decoupling of

the imaging and photo-stimulation planes up to 500 μm (i.e. spanning the 50 mm range of a Newport 433 translation stage we used to mount the diffuser, Methods, **Fig. 3b**).

As the photo-stimulation pattern was moved away from the focal plane of the objective, we measured a slight increase in lateral (<10%) and axial extent (<20%, **Fig. 3c**), which we further accounted for by applying corrections in the control software (Methods). Over the axial displacement interval sampled, we observed a linear relationship between translating the objective axially and the compensatory displacement of the diffuser required to bring the pattern back into focus, proportional to the square of the effective optical magnification in our setup (theoretical vs. effective magnification $\sim 11.1 \times$ vs. $\sim 9.3 \times$, **Fig. 3b**, Methods).

Figure 3




Figure 3. Decoupling the photo-stimulation and imaging z-planes using a translatable diffuser. (a) Schematic cartoon describing using of the primary and secondary (auxiliary) microscopes to calibrate the decoupling of the photo-stimulation and optical-readout planes. A thin (1 μm) rhodamine B fluorescent coverslip sample is placed in the focal plane of both primary and

secondary microscope objectives, where a checker-board pattern is projected using the DMD-chip (Step 1). Lowering the primary objective by a given amount translates the pattern axially below the focal plane of the secondary objective (which remained stationary), and produces a blurred image of the checker-board pattern on the secondary (CCD2) camera (Step 2). The projected light pattern is brought back into focus on CCD2 by compensating with translating the diffuser along the optical path (i.e. moving the diffuser away from the back-aperture of the primary objective, which defocuses the image on the primary (CCD1) camera (Step 3). **(b)** Measured calibration curve of diffuser translation required to shift photo-stimulation plane by a given amount with respect to the focal plane of the objective. Each point represents Mean $\pm$ SEM across several months, indicating the robustness of the setup and linear relationship between objective and compensatory diffuser displacements for the range of photo-stimulation and imaging plane decoupling sampled (500 μ m). **(c)** Normalized axial light intensity profile of spots. Measurements for spots of three different diameters (30, 50 and 90 μ m), indicated by the three clusters of color traces. Each line color represents a different diffuser offset, expressed as degree of axial displacement between the photo-stimulation and focal plane of the primary objective in 100 μ m steps (0 to 400 μ m). (Inset) Measured change in axial (z) FWHM of the light profiles in **c.** as a function of axial displacement between photo-stimulation and imaging planes relative to no-offset (z-overlapping) condition, which we further accounted for by applying corrections in the control software.

Quantifying photo-stimulation pattern propagation through brain tissue

One-photon patterned stimulation is constrained to surface applications because of limited penetration of shorter wavelengths into tissue due to scattering. To characterize the propagation of DMD-generated light patterns through brain slices of different thickness^{76,77}, we placed fixed olfactory bulb slices (Methods) atop a fluorescent coverslip, and recorded images of the projected patterns from underneath with the auxiliary microscope. To reduce the effect of variability in slicing angle and local neuronal architecture, we acquired images at two axial locations in each slice (Methods). Patterns remained distinct in slices up to $\sim$ 150 μ m thick (**Figs. S2c-f**), albeit substantial fluorescence loss (**Fig. S2f**). We note that using the diffuser, besides confining the axial spread of photo-stimulation, also increases the homogeneity of the projected light patterns in scattering media (**Figs. S2f-g**). Without the diffuser in the optical path, patterns appeared speckled: standard deviation of pixel intensities within the pattern was higher compared to same thickness samples with the diffuser in place (**Figs. S2e-g**). This effect can be explained intuitively by taking

a ray-tracing point of view on image formation. With the diffuser in the optical path, due to angular dispersion, light emitted from a single point in the pattern travels along more paths through the tissue compared to the no-diffuser condition to reach at a given camera pixel (**Fig. S2e**). As such, using the diffuser increases homogeneity in light intensity across the field of stimulation, as this configuration is less sensitive to variations in scattering from local structure in the tissue.

Mapping functional connectivity in the intact brain

ADePT combines precise, patterned optogenetic stimulation of the brain with high-resolution monitoring of resultant neuronal activity. This method is ideal for investigating the logic of functional connectivity in layered neuronal circuits where input and output circuit elements do not necessarily lie in the same plane. In addition, our one-photon DMD based implementation is particularly well-suited for understanding information flow in areas where sensory stimuli evoke structured spatiotemporal patterns of activity that extend over several 10s of microns (e.g. glomeruli in the olfactory, whisker barrels in barrel cortex, etc.). As a proof of principle, we used *ADePT* to investigate the role of different local inhibitory and excitatory connections in shaping the output of the mouse olfactory bulb. Briefly, in the OB, the input nodes called glomeruli (balls of neuropil sorted by odorant receptor identity) and the cell bodies of the principal output neurons (mitral and tufted cells) are separated by a couple hundred micrometers in depth^{53,54}. Individual odors activate distinct spatial-temporal combinations of glomeruli. This odor-evoked activity is heavily reformatted within the OB by numerous excitatory and inhibitory interneurons and relayed to other brain areas in the form of activity patterns of the mitral/tufted cells. To date, understanding the OB computations has been challenging due to constraints in precisely and flexibly controlling the activity of individual input nodes (glomeruli) using traditional odor stimuli. Using *ADePT* we

can bypass these constraints by direct photo-stimulation while simultaneously monitoring the OB outputs at high throughput and cellular resolution.

Application 1: Mapping glomerular receptive fields of sister mitral/tufted cells

We used *ADePT* to first identify cohorts of *sister* output cells in the OB – mitral and tufted cells that receive excitatory input from the same *parent* glomerulus^{36,55–64} – and then assess whether they integrate lateral inputs from different sets of glomeruli in the OB. Indeed, *sister* mitral and tufted cells extend multiple secondary (lateral) dendrites, which may each contact different sets of inhibitory interneurons^{36,55,56,78}. Previous work has shown that while *sister* cells are generally responsive to the same set of odors, they are not redundant in their odor responses: pairs of *sister* cells are as different as pairs of *non-sisters* in their response time with respect to inhalation onset^{36,55}. A potential mechanism for temporal differences in the odor-evoked responses of *sister* cells is differential inhibitory inputs from other glomeruli in the OB. However, to date, mapping glomerular receptive fields of individual mitral/tufted cells has been difficult given technical limitations in precise control of single-glomerulus activity in conjunction with monitoring responses at a high-enough yield to identify and further compare cohorts of *sister* mitral/tufted cells. Using *ADePT* enabled us to bypass the constraints imposed by odor stimulation and directly photo-stimulate individual glomeruli on the OB surface. In parallel, it enabled monitoring the responses of many mitral/tufted cells simultaneously at cellular resolution using multiphoton calcium imaging. To this end, we used mice expressing both an opsin in the mature olfactory sensory neurons (OSN), OMP-Cre x ReaChR-mCitrine⁷⁹ and the genetically encoded calcium indicator (GCaMP6f/s⁸⁰) in the mitral and tufted cells (Thy1-GCaMP6s⁸¹; **Fig. S3a**, Methods).

Figure 4




Figure 4. Targeting individual glomeruli by photo-stimulation of olfactory sensory neuron terminals and monitoring glomerular responses of sister mitral and tufted cell dendrites. (a) (i) Glomerular resting fluorescence in OMP-Cre x ReaChR-citrine x Thy1-GCaMP6s mice; (b) Fluorescence changes (two dimensional glomerular light activation maps, 2DLAMs) evoked by photo-stimulation of 2 example glomeruli (#3, #4) out of 15 glomeruli targeted in the example field of view (5.0 mW/mm^2 for 50 ms). Red ellipses mark the location and contour of the projected light masks. Color scale in heat map indicates $\Delta F/F$, comparing fluorescence after light stimulation to baseline; (c) Response matrix of fluorescence traces (individual repeats in gray, and mean signal as thicker lines) corresponding to 15 glomerular regions of interest versus the 15 (matching) glomerular light masks projected on the OB surface. Red traces mark responses of the targeted glomerulus in a given trial. (d) (Left) Fluorescence changes (two dimensional glomerular light activation maps, 2DLAMs; 5.0 mW/mm^2) evoked by photo-stimulation of the 15 glomeruli targeted in the example field of view (a). (Right) Significant 2DLAMs. Non-significant responses were thresholded to blue background. Out of 15 targeted glomeruli, 5 glomeruli were considered non-responsive, given our signal threshold criteria (Methods). Of the 10 responsive light maps, 8 showed specific responses (pertaining to the target glomerulus), and 2 light maps were not specific (responses were observed across multiple glomeruli). (e) Example target glomeruli from a different example FOV across a range of three light intensities (2.5 , 5.0 and 7.0 mW/mm^2). Projected light masks match the red ellipses in (i) and (ii). Color scale in heat map indicates $\Delta F/F$, comparing fluorescence triggered by light stimulation to baseline; (iii) Same as (ii) for statistically significant glomerular photo-stimulation responses (rectified 2DLAMs); (iv) Fluorescence traces (mean signal for each light intensity) corresponding to three regions of interest (ROI) in each heat map, whose position is indicated by red ellipses in i-iii. Different intensities are represented by different color traces (red - 2.5 mW/mm^2 , green - 5.0 mW/mm^2 , blue - 7.5 mW/mm^2). (f) Specificity of glomerular responses across light intensity. Blue trace: responsive pixels within the target glomerulus normalized by all responsive pixels located within glomerular boundaries across the FOV. Red trace: responsive pixels within the target glomerulus normalized by all responsive pixels within the FOV, including those spanning the juxtglomerular space, surrounding glomeruli (Methods); $N=7$ mice, 21 FOVs; responsive vs. photo-stimulated glomeruli: 2.5 mW/mm^2 , 12/74; 5.0 mW/mm^2 , 48/122; 7.5 mW/mm^2 , 27/57; 10.0 mW/mm^2 , 48/98; 15.0 mW/mm^2 , 48/77.

We stimulated glomeruli by projecting DMD-modulated light masks on top of individual glomeruli, identified based on their baseline fluorescence (Methods). We used photo-stimulation spot sizes smaller ($30\text{-}50 \text{ }\mu\text{m}$) than the targeted glomeruli (Avg. $\sim 75 \text{ }\mu\text{m}$)^{36,82}. To assess the resultant responses in the mitral/tufted cells, we strobed between stimulating OSN axon terminals ($50\text{-}150 \text{ ms}$ stimulation) and monitoring calcium responses ($5\text{-}10 \text{ Hz}$ imaging rate) in either the postsynaptic mitral/tufted dendritic tufts in the glomerular layer (same z-plane) or the cell-bodies ($100\text{-}270 \text{ }\mu\text{m}$ below the photo-stimulation plane). First, we calibrated the stimulation light intensity to ensure single glomerulus specificity – photo-stimulation of OSNs only within the targeted

glomerulus while avoiding spurious excitation of passing-by OSN axons that terminate in neighboring glomeruli. To assess the efficacy of our approach, we set the photo-stimulation and multiphoton activity readout planes to overlap axially in the glomerular layer, such as to stimulate the OSNs and image the resultant postsynaptic glomerular response in the dendrites of MT cells. Photo-stimulation induced rapid and reliable changes in fluorescence (**Fig. 4**). For each photo-stimulation mask (**Figs. 4b-e, S3b**), we constructed a two-dimensional light activation map (2DLAM) at sub-glomerular resolution³⁶ (e.g. 40 x 40 grid of 7.2 μm pixels, Methods). To ensure we operated in a regime of specific glomerular activation, we successively lowered the light intensity until no light-triggered activity was observed in neighboring glomeruli (15-to-2 mW / mm², 122 glomeruli, 21 FOVs, 7 mice, **Figs. 4f, S3c**). Across mice, we identified a range of intensities that on average resulted in specific and reliable activation of targeted glomeruli and negligible activation of adjacent glomeruli (**Fig. 4f**).

Next, we axially decoupled the photo-stimulation and imaging z-planes to monitor light-triggered responses in the cell bodies of the corresponding *daughter* M/T cells (receiving excitatory input from the targeted glomeruli, and 100-270 μm deep from surface), while still photo-stimulating individual glomeruli on the bulb surface (**Figs. 5a-e, S4a-f**). Photo-stimulation of individual glomeruli evoked robust excitatory responses in ~19 % of the cells (162/856). The majority of these responsive cells (141/162) were activated by only one unique glomerulus that we ascribed as their corresponding parent glomerulus. A small proportion of cells were activated by more than one glomerulus and excluded from further analysis (21/162 cells, **Fig. 5b iii**, cell 30). By comparing the identity of parent glomeruli across the imaged M/T cells, we identified cohorts of *sisters* as the cells that were activated by the same parent glomerulus.

Figure 5




Figure 5. Identifying sister mitral and tufted cells via axially decoupled glomerular stimulation and two photon fluorescence imaging. (a-b) anaesthetized, (c-d) awake. (a) (i) Resting glomerular fluorescence (ReaChR-mCitrine and GCaMP6s); (ii) Fluorescence changes (raw, Top and significant pixels glomerular 2DLAMs, Bottom) evoked by photo-stimulation of four target glomeruli (7.5 mW/mm^2); (iii) Response matrix of fluorescence change traces (individual repeats in gray, and mean signal as thicker lines) corresponding to four glomerular regions of interest versus the four corresponding (matching) glomerular light masks projected on the OB surface. Red shaded areas mark responses of the targeted glomerulus in a given trial. (b) (i) Resting fluorescence (GCaMP6s) of mitral cell bodies at the same x-y coordinates, $210 \mu\text{m}$ below the glomerular field of view in (a); (ii) Significant cell body responses evoked by photo-stimulation of the four target glomeruli (a, 7.5 mW/mm^2); (iii) Average fluorescence response traces corresponding to six example mitral cells in each heat map, whose positions are indicated by numbers in (ii); shaded area marks SEM. Two example pairs of sister cells (47, 7 with respect to parent glomerulus #1; and 37, 45 with respect to parent glomerulus #3) are shown. Note that cell #30 is non-specifically responding to photo-stimulation of multiple glomeruli (#2,#3,#4). In addition, glomerular photo-stimulation triggered sparse suppressed responses (e.g. cells 7,11) intermingled with the sister cells. (c) (i-iii) Same as (a) for an example awake experiment: photo-stimulation of three target glomeruli (2.5 mW/mm^2 ; 50 ms light on/trial); (iii) Response matrix of fluorescence change traces (individual repeats in gray, and mean signal as thicker lines) corresponding to three glomerular regions of interest (#3,#4,#5) versus the three (matching) glomerular light masks projected on the OB surface. Stimulation of glomeruli #1 and #2 resulted in non-specific responses. (d) (i-iii) same as (b) for an example awake experiment sampling mitral cell bodies at the same x-y coordinates, $190 \mu\text{m}$ below the glomerular field of view shown in (c); (ii) Significant cell body responses evoked by photo-stimulation of three target glomeruli (c,ii); (iii) Average fluorescence time traces corresponding to nine example mitral cells in each heat map, whose positions are indicated by numbers in (ii). Three example pairs of sister cells (11, 50 with respect to parent glomerulus #3; 65, 73 with respect to parent glomerulus #4; and 7 and 42 with respect to parent glomerulus #5) are shown. Glomerular photo-stimulation also triggered spatially distributed suppressed responses (e.g. cells 13, 40, 42, 50, 73, 74), denser than in anaesthetized mice. (e) (i) (Top) Specificity of enhanced responding mitral and tufted cells as a function of the targeted glomerulus ($N = 6$ mice; 27 responsive and specific glomeruli / 46 photo-stimulated target glomeruli; 856 mitral & tufted cells). Note that most cells were non-responsive, or responded specifically to the photo-stimulation of the target glomerulus (13 % of cells had off-target responses to more than one glomerulus; see Fig. S4a, Methods); (Bottom) Same for suppressed responses. (ii) (Top) Number of observed sister cells (enhanced responses) per glomerulus photo-stimulation across fields of view (Left; 6.9 ± 6.0 , $N = 27$ FOVs); (Bottom) Number of cells whose baseline fluorescence was suppressed upon photo-stimulation of the same glomeruli (Right; 3.2 ± 2.3 , $N = 20$ FOVs); (iii) (Top) Pairwise distance between sister cells vs. between any cells in the field of view ($142.6 \pm 82.1 \mu\text{m}$ vs. $137.7 \pm 71.8 \mu\text{m}$; 725 vs. 29,850 pairs). (Bottom) Pairwise distance between suppressed cells vs. any cells in the field of view ($116.2 \pm 58.6 \mu\text{m}$ vs. $138.2 \pm 70.4 \mu\text{m}$, 83 vs. 19,607 pairs). Unless specified Avg. $\pm$ SD values are shown. (f) Photo-stimulation of individual (non-parent) glomeruli triggered inhibitory responses in only a small subset of the sister cells of a given identified cohort (Avg. $\pm$ SEM, $14.0 \pm 2.7 \%$, $N=6$).

On average, across experiments (10 FOVs, 6 mice, 27/46 response evoking glomeruli) we identified 6.9 ± 6.0 (Avg. $\pm$ SD) *sister* mitral and tufted cells per glomerulus (4 SD response threshold, see Methods). Consistent with previous literature^{36,56,58,63,83}, *sister* cells were spread across the field of view and spatially intermingled with cells receiving inputs from other glomeruli (*non-sisters*) (Avg. $\pm$ SD: 142.6 ± 82.1 μ m pairwise distance between *sister* cells vs. 137.7 ± 71.8 μ m between any cells in the FOV). Note that the statistics reported here underestimate the number of *sister* cells per glomerulus⁵⁸ simply because of the limited number of different depths (z-optical planes) and the size of fields of view sampled for this experiment.

In addition to light-triggered excitatory responses, we also observed sparse light-induced suppression of the baseline fluorescence in 3.2 ± 2.3 cells per glomerulus. These suppressed cells were intermingled with the identified *sister* cells, and peppered the fields of view (3 SD response-to-baseline signal threshold; 116.2 ± 58.6 μ m pairwise distance of suppressed cells vs. 138.2 ± 70.4 μ m between any cells in the field). We interpret the observed decrease in M/T resting fluorescence as the suppression of respiration-locked baseline firing patterns^{84–86} due to lateral inhibition from the photo-stimulated glomeruli acting via local bulbar circuits. Interestingly, we could also observe such light-triggered suppression for a subset of the *sister* cell cohorts (6/27 *sister* cell cohorts) identified via ADePT. This allowed us to assess whether *sister* cells receive lateral inhibitory inputs from the same or different glomeruli. Photo-stimulation of individual, *non-parent* glomeruli inhibited a small fraction of *sisters* within a cohort (1-2 *sisters* per cohort; Avg. $\pm$ SEM, 14.0 ± 2.7 %, N=6, **Fig. 5f**, Methods). In rare instances, where we observed inhibition in 2 or more *sister* cells, we found that different *sister* cells were suppressed by photo-stimulation of different individual glomeruli (**Fig. S4f**). These data indicate that indeed *sister* cells, while sharing excitatory drive from the same parent glomerulus, receive sparse inhibitory inputs from different

sets of non-parent glomeruli. This is in agreement with previous reports of differences in *sister* cells spike-timing across the respiration cycle^{36,55,78}. While further investigation is needed to understand the logic of these lateral inhibitory interactions, these experiments show that *ADePT* is well-suited for direct mapping of excitatory/inhibitory connections in layered neuronal circuits.

Application 2: Assessing the spatial heterogeneity of long-range olfactory bulb inhibition

Next, we used *ADePT* to assess the spatial heterogeneity in the impact of GABAergic/dopaminergic (DAT+ cells, historically known as superficial short axon cells), a class of long-range inhibitory interneurons in the glomerular layer, on the activity of mitral and tufted cells. DAT+ neurons integrate excitatory inputs within individual glomeruli and establish chemical and electrical synapses with other interneurons and mitral/tufted cells within other glomeruli located as far as 1.5 mm away^{43,66,67,71–73,87}. Given these long-range processes, DAT+ cells are thought to re-distribute focal glomerular odor inputs and convey long-range signals across the glomerular layer. Work from several groups including ours^{42,43,66–73,88} suggested that DAT+ neurons enable olfactory circuits to negotiate wide variations in odor concentration by implementing a form of global gain-control (divisive normalization)^{42,43,68,89}. However, it remains unclear whether DAT+ cells broadcast inhibitory inputs uniformly within a local neighborhood or act selectively on particular cells that share specific features such as similarity in odor tuning^{43,67,88,90}. This is due to previous technical constraints in controlling the activity of DAT+ cells in conjunction with simultaneous imaging of activity in the mitral and tufted cells. We used *ADePT* to photo-stimulate DAT+ cells associated with individual glomeruli while monitoring their impact on either the spontaneous or odor-evoked responses of mitral/tufted cells.

We first tested the specificity of targeting DAT⁺ interneurons across glomerular regions of interest using mice co-expressing both an opsin (ChR2-mCherry⁹¹) and a calcium indicator (GCaMP6f⁸⁰) in the DAT⁺ cells (**Fig. S5a**). We selected individual adjacent glomeruli for photo-activation, triggering and reading out DAT⁺ cell activity in the same optical z-plane. Through these experiments, we identified a range of light intensities across which the stimulus-responsive area remained confined to the glomerulus targeted for photo-stimulation (**Figs. S5b-e**, Methods).

To assess the impact of glomerular DAT⁺ interneurons on the activity of mitral and tufted cells, we used mice expressing an excitatory opsin in DAT⁺ cells and the calcium indicator GCaMP6f in mitral/tufted cells (DAT-Cre x Thy1GCaMP6f^{80,92} mice injected retro-orbitally with ChrimsonR-tdtomato AAV-PHP.eB:Syn-FLEX-rc[ChrimsonR-tdtomato]^{93,94}, **Fig.S6a**, Methods). We optogenetically activated DAT⁺ neurons and monitored the ensuing tufted and mitral cell responses 120-275 μ m below the photo-stimulated glomeruli (**Fig. 6**). On average, glomerular DAT⁺ photo-stimulation targeted to *single* glomeruli resulted in dense, but heterogeneous suppression of the *baseline fluorescence* in mitral cells across the field of view (fraction of mitral cells modulated in example field of view: 0.40 ± 0.19 ; 200 x 200 μ m, 294 cell-mask pairs, **Figs. 6a-d, S6b-d**). Stimulation of different DAT⁺ glomerular masks resulted in ensemble mitral cell responses that were significantly different from each other, but more correlated than the average pairwise responses for controls where cell identities were shuffled (**Fig. 6e**; 0.42 ± 0.23 vs. 0.29 ± 0.19 , $p < 0.001$, Wilcoxon signed-rank test) or across odors (**Fig. 6e**; 0.42 ± 0.23 vs. 0.25 ± 0.35 , $p < 0.001$, Wilcoxon signed-rank test). As reference, we also calculated a ‘self’ correlation to assess the variability of photo-stimulation responses to the *same* DAT⁺ glomerular mask across trials (Methods). Responses of mitral cell ensembles across different DAT⁺ masks were substantially

381 lower than the ‘self’ control (0.42 ± 0.23 vs. 0.81 ± 0.14 , $p < 0.001$, Wilcoxon signed-rank test,
382 **Figs. S6b**), demonstrating significant impact of DAT+ neurons on OB outputs.

Figure 6




Figure 6. Glomerular photo-activation of DAT+ interneurons triggers dense, but spatial heterogeneous inhibition of mitral and tufted cell activity. (a) (i) Cartoon schematics: photo-stimulation of opsin expressing DAT+ cells while monitoring mitral cell activity (GCaMP6f) by axial decoupling of the photo-stimulation and imaging planes (awake). (ii) Example glomerular field of stimulation; mice express GCaMP6f in mitral and tufted cell glomerular dendritic tufts and ChrimsonR-tdtomato in DAT+ cells (not shown). (iii) Example field of view of mitral cells imaged while photo-stimulating DAT+ cells (200 ms, 7.5 mW/mm²) in the glomerular layer. (b) (i) (Top) Dense, but heterogeneous inhibition of baseline fluorescence in mitral cells across the field of view evoked by photo-stimulation of DAT+ cells across six target glomerular light masks (G1-G6); (Bottom) Odor responses of the same example mitral cells to six odors (isoamyl acetate, IAA; acetaldehyde, ADE; valeraldehyde, PEN; allyl tiglate, AT; heptanal, HEP; acetophenone, ACP). (c) (i) (Top) Fluorescence changes (two dimensional light suppression maps) in mitral cells evoked by photo-stimulation of DAT+ cells across six individual glomerular light masks (G1-G6); (Bottom) Diversity of same mitral cell responses to six example odors (b); each cell in the field of view is displayed on a blue-to-red color scale indicating its response amplitude to light/odor stimulation; non-significant responses were thresholded to white (Methods). (d) Peak fluorescence changes (dF/F) of same mitral cells (49) across the field of view in response to glomerular DAT+ cells photo-stimulation (Left) and to a panel of eleven odors (Right). (e) (Left) Correlation distribution of mitral cell ensemble responses across different pairs of DAT+ cell glomerular photo-stimulation masks (G1-G6). (Right) Same across pairs of odors in the panel (1-11). Shuffled mitral cell index control distributions are shown in red; ‘self’ control (see Fig. S6b-d). (f-h) anaesthetized, also see Fig. S7a-f; (f) Fluorescence change (dF/F) elicited by odor presentation in one example tufted cell, and its subsequent modulation by projecting of a glomerular light mask to optogenetically augment DAT+ cell activity (500 ms, 10mW/mm²; 2s from odor onset). (g) Mitral cell peak odor responses (ADE, PEN, 2-HEX) and their modulation by 4 glomerular DAT+ cell glomerular light masks in an example field of view (Fig. S7). (h) Stimulus response pairwise correlations of mitral cell ensembles: distribution of pairwise correlations for different glomerular DAT+ cells photo-stimulation masks imposed on top of the same odor aggregated across the three odors in (g) (0.21 ± 0.25 ; 27 pairs, 2 FOVs, 39 MC cells).

Stimulation of DAT+ neurons results predominantly in the suppression of mitral and tufted cells from their baseline activity. To boost the baseline M/T cell fluorescence and better characterize the impact of DAT+ neurons, we presented odors (4s) in conjunction with DAT+ cell photo-activation in the middle of the odor presentation period. Similar to the experiments described above (Figs. 6a-e), stimulating DAT+ cells across different glomeruli during odor presentation also resulted overall in differential suppression of the OB outputs (Figs. 6f-h, S7a-f, TC: 0.44 ± 0.33 ; MC: 0.21 ± 0.25 ; pairwise correlation, $p < 0.001$, Wilcoxon signed-rank test; modulation by light on top of odor: 78.0 % unmodulated, 18.2% suppressed vs. 3.8% enhanced MC-odor-pairs; N=2,214

pairs; and 54% unmodulated, 42.6% suppressed vs. 3.4% enhanced TC-odor pairs, N = 945 pairs). Thus, we find that the impact of DAT+ cells on the OB output is dense, but locally heterogeneous with different cells exhibiting diverse sensitivities to DAT+ inhibition originating from a given glomerulus (**Fig. 6**). Further investigation will determine whether DAT+ cell action shapes the MC and TC cell odor responses in a systematic manner as a function of similarity in their odor responses.

In summary, we show that *ADePT* can be flexibly combined with genetic strategies and/or sensory stimuli to probe the contribution of both excitatory and inhibitory neuronal types in shaping the output of optically accessible neuronal circuits.

Discussion

We demonstrate a readily implementable method (*ADePT*) for optical control and recording of activity in three-dimensional neural structures by decoupling the focal planes for optogenetic patterned stimulation and multiphoton imaging. When referenced to other comparable techniques^{8,9,13,20–22,29–33}, *ADePT* enables us to control neuronal activity across a wide field of view, does not suffer from aberrated imaging, and is easy to implement. We confirmed that we were able to maintain precise spatial control over a range of activation powers and demonstrated the usefulness of the system by mapping *both excitatory and inhibitory* functional connectivity in the mammalian olfactory bulb.

ADePT allows arbitrary axial shifting of 2D photo-stimulation light patterns of choice relative to the focal plane of the imaging objective by translating a motorized diffuser along the photo-

stimulation optical path (**Fig. 1**). As a function of its angular divergence, the choice of the diffuser determines the degree of axial confinement of the photo-stimulation pattern, enabling optical sectioning in the widefield illumination configuration, and also increasing homogeneity in the photo-stimulation patterns (**Figs. 2,3**). The larger the angular divergence of the diffuser, the tighter the axial extent of the pattern, which comes at the expense of excitation light loss. In our implementation, we controlled glomerular activity optogenetically over a 1.5 x 1.2 mm field of stimulation with ~ 20 μ m lateral and ~ 40 μ m axial resolution. Different magnification choices in the photo-stimulation path, however, can be used to customize the extent of the field of stimulation with inherent resolution trade-offs. In its current implementation, *ADePT* does not afford the resolution to optogenetically target individual cells, but is effective in controlling the activity of larger input units (e.g. single olfactory glomeruli, whisker barrels) in neural circuits of interest, while monitoring their impact across different cell types at cellular resolution via multiphoton imaging of changes in activity. Furthermore, different choices of the diffuser enable tighter axial resolution at the expense of photo-stimulation light loss (grit ground diffuser vs. holographic diffusers, **Fig. 2**). While achieving axial-confinement of extended 2D photo-stimulation patterns (**Fig. 2**), *ADePT* is effective for optical control of activity largely in the superficial brain layers, since it is constrained by light scattering-related degradation (**Fig. S2**).

In proof-of-principle experiments, we used *ADePT* to investigate the functional connectivity of olfactory bulb circuits in the mouse. We controlled optogenetically the activity of select excitatory and inhibitory cell types in the input layer (OSNs vs. GABAergic/dopaminergic DAT+ interneurons) at *sub-glomerular* resolution, while monitoring their impact on the responses of the principal neurons (mitral and tufted cells). By targeting optically addressable glomeruli (**Figs. 4,5**),

we identified multiple cohorts of daughter *sister* mitral and tufted cells driven by *parent* glomeruli and revealed sparseness and differential action in the olfactory bulb inhibitory lateral interactions (**Fig. 5**). Our approach further enables discriminating between different inhibitory inter-glomerular connectivity schemas (e.g. uniform vs. heterogeneous) implemented by different cell populations (e.g. GABAergic/dopaminergic DAT⁺ interneurons)^{43,67,88}, and testing hypotheses on how connectivity relates to the similarity in odor response tuning^{67,90}. We found that DAT⁺ cells associated with different glomeruli mediate dense, but heterogeneous (non-uniform) suppression of MC and TC baseline and odor responses (**Fig. 6**).

Compared to previous work on mapping individual glomerular outputs^{36,55}, *ADePT* affords one-to-two orders of magnitude higher throughput and cell type specificity in the activity readout. Identifying multiple cohorts of *sister* cells is key for determining the specificity of inhibitory inputs that principal OB neurons integrate to compute relevant stimulus features. As previously proposed^{36,55,56}, we found that *sister* cells, driven by the same *parent* glomerulus, receive differential inhibitory input from other glomeruli with implications for odor coding. As such, many more output information channels may leave the bulb than the number of input sensory afferent types. Increasing the number of *sister cell cohorts* mapped to their *parent* glomeruli, and systematic photo-stimulating spatial-temporal combinations of multiple (*non-parent*) glomeruli will enable further testing hypotheses on how inhibitory connectivity relates, for example, to the degree of similarity in the odor responses of the corresponding input glomeruli. In recent sequencing-based single-cell projection mapping, we identified structure in the projections of individual mitral cells at different locations along the A-P axis of the piriform cortex with respect to their co-projections to extra-piriform target regions^{95,96}. By combining functional mapping and

anatomical DNA barcoding projection mapping approaches, future work will determine whether the *sister* outputs of individual glomeruli segregate into parallel functional streams that may support different computations related to perception versus valence versus action⁹⁶ during odor-guided behaviors.

More broadly, simultaneous patterned photo-stimulation and multiphoton imaging in different axial planes enables mapping functional connectivity not only in the olfactory bulb, but can be implemented between any optically accessible brain regions (across layers) including the neocortex. For example, *ADePT* is well positioned for mapping functional connectivity in the barrel cortex within and across barrels^{34,97}, as well as bidirectional interactions between cortical visual areas (V1 vs. PM, AL, etc.)^{98,99}, visuomotor (M2-V1)^{100–102} or audiomotor circuits (M2-A1)^{103,104}, etc. Given its relative simplicity, cost-effectiveness, and ease-of-use, compared to other axially decoupling optical imaging and photo-stimulation strategies, *ADePT* is readily applicable for investigating the logic of neural connectivity.

Acknowledgements: The authors acknowledge R. Eifert and D. Cowan for technical support. This work was supported by NIH 5R01DC012853-05, NSF IOS-1656830 to D.F.A. and Beckman Institute CLOVER (CLARITY, Optogenetics, and Vector Engineering Research) to V.G.

Data and code availability: All data matrices included in the analyses presented here representing mitral/tufted and GABAergic/dopaminergic DAT+ interneuron GCaMP responses triggered by

odor and optogenetic light stimuli, as well as the code used for the instrument control and analysis are available upon request.

Declaration of interests: The authors declare no competing interests.

Authors contributions: conceptualization, investigation, methodology, formal analysis, visualization, software, data curation and writing, M.K., F.A.; methodology, formal analyses, visualization, data curation, validation, software, S.K.K., S.S.T; methodology, visualization, validation, D.E.H., W.G.B., P.S.V., H.C, M.B.D.; reagents development, Z.Q., V.G.; conceptualization, methodology, formal analysis, visualization, and writing, A.B.; conceptualization, methodology, formal analysis, visualization, software and writing, P.G.; conceptualization, methodology, investigation, methodology, writing, supervision, project administration, and funding acquisition, D.F.A.

Methods

Combined two-photon imaging and photo-stimulation microscope: For *two photon imaging* (**Fig. 1b**), the output from a Ti:Sapphire laser (Coherent Chameleon Ultra II) was magnified by a 1.5x beam expander (L1, L2, focal lengths, FL: 40, 60 mm), then reflected off two mirrors (10D20ER.4, Newport) and a periscope onto a galvanometric mirror scanning system (GM, 6215HB, Cambridge Technologies). The semi-stable stationary point between the GM was expanded and imaged to fill the back aperture of the microscope objective by a telescope (L3, AC300-050-B, L4, AC508-300-B, FL 50, 300 mm, Thorlabs). The objective and microscope stage

505 were mounted on motorized linear translators (x-y stage: 462-XY-M; z-stage: CMA-25CCCL,
506 Newport) connected to a three-axis motion controller (Newport ESP301). Fluorescence light from
507 the sample was collected by the objective, then reflected by a dichroic mirror (DM1, 680dcxr, 50
508 mm diameter, Chroma) placed atop of the imaging objective (Olympus 20x 1.00 NA
509 XLUMPlanFLN) and a second short-pass dichroic mirror (DM2, 505 nm, FF505-SDi01-50.8-D,
510 50 mm diameter, Semrock) towards a GaAsP PMT (H11526-20-NF, Hamamatsu) for green or
511 red fluorescent detection (sliding holder, LCFH2 green, FF03-525/50-25 and red, FF01-607/70-
512 25 emission filters, Semrock). The back aperture of the microscope objective was imaged into the
513 PMT aperture by a telescope (L5/L6, FL: 100/32 mm; L5 – LA4545-A (UV fused silica); L6 -
514 AL4532-A, Thorlabs). The current output of the PMT was transformed to voltage, amplified
515 (SR570, Stanford Instruments) and digitized using a data acquisition board (NI-PCI-6115,
516 National Instruments). The same board was also used to generate signals for controlling the
517 galvano-metric mirrors. By rotating DM2 by 90° on a custom holder (CSHL, Machine Shop), light
518 was directed either to the PMT or to a CCD camera (Vosskuhler 1300-QF, **Fig. 1**). The image of
519 the sample was further minified by a telescope (L7/L8, focal lengths: 100/40 mm) and captured on
520 the CCD chip (CCD 1), allowing for a larger field of view of approximately 1,500 x 1,200 µm.
521 Additional short pass filters (FF01-750/SP-50, 50 mm diameter, Semrock) were used in front of
522 light detectors (CCD and PMTs) to block stray light from the infrared multiphoton excitation laser.
523 For *patterned photo-stimulation*, the beam from a 488-nm laser (Coherent Sapphire 488, 200mW)
524 was magnified using a beam expander (L12/L11, FL: 30/300 mm) and reflected by a mirror and a
525 periscope (RS99, Thorlabs) onto the surface of a DMD chip (DMD 0.7" XGA VIS, Vialux GmbH,
526 1024 x 768 pixels, 13.68 µm micro-mirror pitch). The DMD chip was controlled using a Vialux
527 DLP-Discovery 4100 board and the corresponding application programming interface (Vialux

GmbH). To obtain direct access to the chip, we removed the projecting lens and prism installed in front of the DMD by the manufacturer. The steep angle of incidence of the photo-stimulation beam onto the DMD was carefully tuned to ensure maximal diffraction efficiency. The patterned beam reflected off the DMD was further magnified by a L10/L9 telescope (FL 100/200 mm), and coupled into the microscope by DM1. An iris (SM2D25, Thorlabs) was placed at the front focal plane of L10 to block the higher diffraction order replicas of the DMD-generated spatial pattern. The photo-stimulation pattern was further projected into the sample by the tube lens (L5, FL 100 mm) and the principal objective (Olympus 20x 1.00 NA XLUMPlanFLN).

Diffuser: A diffuser was used to axially confine the photo-stimulation pattern. The diffuser was mounted on a translational stage (Newport, 433-series, #BZA647202) connected to a stepper motor. The motor was controlled using an open source Arduino-based system¹⁰⁵. In the home position (imaging and photo-stimulation planes superimposed), the diffuser sits at the focal plane of the tube lens L5. Three different round diffusers (25 mm diameter) were tested, as detailed in the text: two holographic diffusers with 5° and 10° divergence angle (5° - #55-849, 10° - #48-506, Edmund Optics), and a grit ground glass diffuser (#38-786, Edmund Optics). Two fast shutters (Uniblitz VS14) were placed in front of the PMT and the blue photo-stimulation laser respectively, and operated in anti-phase so that only one shutter was open at a time such as to alternate between periods of photo-stimulation and photon collection.

Excitation power control and lenses: The excitation two photon laser power was controlled by a Pockels cell (Conoptics 350-80 BK) placed in the imaging path. An additional filter (775 nm long pass, 25 mm diameter, #86-069, Edmund Optics) was added to the excitation path to block a


residual red component from the Ti:Sapphire laser, that would otherwise saturate the red PMT. When required, additional neutral density filters (Thorlabs, NDK01) ranging from ND 0.2 to ND 2.0 were added to the photo-stimulation path to further reduce the power. All lenses (Thorlabs) had anti-reflective coating for either infrared (imaging path) or visible (photo-stimulation and detection paths) light. A UV fused silica lens was used instead of the more common N-BK7 lenses in position L5 (LA4545-A) to avoid spurious phosphorescence signals following illumination with intense blue light pulses (see *Phosphorescence Artifacts*).

Auxiliary microscope: An auxiliary microscope was used during characterization experiments, as described in *Lutz et al. (2008)*⁶. We placed a secondary objective, tube lens and green emission filter (HQ510DCLP, Chroma) below the main imaging objective to image the focal plane of the main objective onto a CCD camera (Vosskuhler 1300-QF, **Fig. 3a**). Depending on the experiment, different objectives (Olympus 4X, 0.13 NA UPlanFLN or Olympus 20X, 0.50 NA UPlanFLN) were used in combination with a 150 mm focal length tube lens to obtain either a minified image of the entire field of photo-stimulation or a higher resolution image of a smaller field. The microscope framework was constructed with a Thorlabs 2" cage system. The secondary objective was held either by a kinematic mount for fine tilt adjustment (KC2-T, Thorlabs), or by a laterally-translatable platform. The photo-stimulation patterns were imaged onto a coverslip spin-coated with a thin (~1 μ m) rhodamine B layer (courtesy of V. Emiliani, Paris Descartes University).

Calibrating decoupling of photo-stimulation and imaging planes: We used the auxiliary microscope to calibrate the decoupling of photo-stimulation and imaging planes (**Fig. 3a**). Initially, the diffuser was positioned at a plane conjugated to the focal plane of the principal and auxiliary

objectives, and the photo-stimulation pattern was in focus on the coverslip, appearing sharp both on the principal (primary) and the auxiliary (secondary) CCD cameras (**Fig. 3a, Left**). The principal objective was then moved closer to the coverslip to simulate shifting the imaging plane to deeper layers in a biological preparation. At this point, the photo-stimulation pattern was still imaged (projected into sharp focus) on the focal plane of the principal microscope objective (now below the coverslip), and no longer in focus on the secondary CCD camera. To move the photo-stimulation pattern back to the initial position, we moved the diffuser away from the tube lens (L5) until the projected pattern was again in focus on the fluorescent coverslip and on the secondary CCD camera (**Fig. 3a, Right**). Using this strategy, we iteratively constructed an empirical calibration curve that relates the diffuser displacement to the position of the photo-stimulation plane in the sample with respect to the focal/image plane of the principal objective (**Fig. 3b**).

Relating movement of the objective to the adjusting displacement of the diffuser: Within the thin lens approximation framework, let u be the distance between the objective (O) and the image of the projected light pattern, $s_i + s_o$ the distance between the objective and the tube lens (L5), v the distance from the diffuser to the tube lens, and f_1 and f_2 the focal lengths of the objective and the tube lens.



How does the displacement of the objective (d) relate to the displacement of the diffuser (D) needed to bring the projected pattern in focus in the desired z -plane in the sample, as a function of

magnification (M) of the system given by the ratio of focal lengths of the tube lens and objective lens ($M=f_2/f_1$)?

$$v = \frac{u * f}{u - f} \quad \text{Eq. 1}$$

Calculating image distance ($u = f_1 - d$) for the objective lens (O), when d is small:

$$s_i = \frac{(f_1 - d) * f_1}{(f_1 - d - f_1)}$$

$$s_i = -\frac{(f_1 - d) * f_1}{d} \quad \text{Eq. 2}$$

For the tube lens, the object distance is: $s_o = (f_1 + f_2 + d) - s_i$

Replacing s_i from Eq. 2:

$$s_o = (f_1 + f_2 + d) + \frac{(f_1 - d) * f_1}{d}$$

Since d is small, ignoring higher order terms of d , we obtain:

$$s_o = \frac{f_1^2 + f_2 * d}{d} \quad \text{Eq. 3}$$

For the tube lens, image distance (v) is given by:

$$v = f_2 + D = \frac{s_o * f_2}{s_o - f_2}$$

Solving for D :

$$D = \frac{s_o * f_2}{s_o - f_2} - f_2$$

$$D = \frac{f_2^2}{s_o - f_2} \quad \text{Eq. 4}$$

Replacing s_o and solving for D :

$$D = \left(\frac{f_2}{f_1} \right)^2 * d \quad \text{Eq. 5}$$

Therefore, we find that the axial displacement of the photo-stimulation plane at the sample (d) is linearly proportional to the translation of the diffuser in the optical path (D).

Photo-stimulation and imaging calibrations

DMD-chip to CCD-chip to multiphoton image registration: To photo-stimulate specific locations in the sample, we determined where a given *pixel on the DMD* falls on the *specimen* using the primary camera (CCD1) as a reference. After registering pixels on the DMD-chip to CCD1 chip pixels, a point-to-point correspondence between the DMD-chip and the focal plane of the principal objective was obtained. Further, the auxiliary microscope was used to map the correspondence between pixels on the *DMD* and the *specimen* in z-planes out of the focal plane of the principal objective. To map pixels in the *two photon field of view* to the *DMD*, we also used CCD1 as intermediary reference. Within this frame of reference, we targeted ROIs that were axially (z) displaced from the principal objective focal plane, and imaged either using the primary CCD camera or the two photon microscope. Pixels of the DMD-chip were mapped onto pixels of CCD1 by projecting a series of light spots generated from known DMD locations. Routinely, we placed 20 μm sized spots in a 4 x 5 grid spanning the field of stimulation. After recording the position of all these fiduciary light spots, we computed the 3 x 3 matrix $\mathbf{T} = \mathbf{D}\mathbf{C}^T(\mathbf{C}\mathbf{C}^T)^{-1}$, where $\mathbf{C}$ and $\mathbf{D}$ are $N \times 2$ matrices whose rows contain the light spot coordinates on the CCD1 and DMD chips respectively. The matrix $\mathbf{T}$, when left-multiplied with CCD1 chip coordinates, transforms CCD1 coordinates into DMD-chip coordinates. To verify this procedure, a high-density sample with 15 μm fluorescent latex microbeads was imaged using first the widefield fluorescence path (CCD1), and subsequently the two photon microscope. Individual microbeads were matched between the acquired images manually (*cpselect* tool in the MATLAB image processing toolbox). We obtained

a look-up table of two photon / CCD1 images point pairs. Further, by minimizing the sum of squared errors, we calculated the corresponding transform matrix that maps two photon image pixels onto CCD1 image pixels. The absolute position of a particular location in a two photon image is dependent on both the field of view and the resolution settings (number of pixels) of the image. Several FOV and pixel resolution settings were chosen for registration, and further used accordingly when targeting ROIs within a two photon reference image. Post-registration, we verified that the targeting of the DMD-chip generated patterns was precise. The control software reproduced the desired target light spots with high accuracy within the field of stimulation, using only the two photon image or only the CCD1 image as guide; error between projected centroids light spots vs. fluorescence microbeads was less than 5 μm , **Figs. 1d, S1d**.

Correcting the field of photo-stimulation power inhomogeneity: Due to the Gaussian intensity profile of the blue laser beam illuminating the DMD chip, the power towards the borders of the field of photo-stimulation is lower than in the center. To correct for this inhomogeneity, and ensure flatness of photo-stimulation across the field, we calculated a power distribution map (PDM) across the field of stimulation, obtained by sequentially projecting 25 μm spots in an 11 x 11 grid pattern covering 121 pixel coordinates across the DMD-chip, and by linearly interpolating such as to estimate the power at the remaining positions. We projected the spot pattern on the DMD chip and placed a power sensor (Thorlabs, 50mW, S120B) under the primary objective with the pattern in focus. For simple patterns, such as spatially-localized single spots, the laser power can be adjusted accordingly using the PDM look-up table to compensate for power loss at the borders. For more complex light patterns, where different regions of the pattern have different local light

intensities, pulse-width modulation can be used to control the average light intensity experienced by each spot independently.

Off-focal plane aberration correction: We often presented a spatial photo-stimulation pattern based on glomerular shapes on the brain surface, while imaging deeper structures, such as mitral and tufted cells, which resulted in small, but significant changes in the size of the projected pattern (i.e. DMD-generated optogenetic stimulation pattern is in focus above the focal plane of the principal imaging objective). We compensated for this change in magnification by another linear transformation that we measured using the auxiliary microscope mounted beneath the principal microscope. The same procedure described above for mapping DMD-chip pixels to the primary camera (CCD1) was repeated using the secondary camera (CCD2) at several focal plane displacements in steps of 100 μm (e.g. 0, 100, 200...500 μm) and a corresponding set of transformations $T(0), \dots, T(N)$ from the DMD-chip pixels to the secondary camera (CCD2) pixels was obtained. For each such transformation, we isolated the effect of changing displacement by multiplying the inverse of the zero-axial displacement transformation to obtain a set of “depth transformations” $B(0), \dots, B(N)$. The depth transformation for arbitrary depths was calculated by linearly interpolating between the measured transformations.

Dichroic mirror offset: When registering the primary camera (CCD1), the dichroic mirror (DM2) must be positioned to direct light from the sample onto the primary camera. However, light from the DMD-chip is refracted as it passes through the dichroic mirror, causing the photo-stimulation pattern to be translated when DM2 is rotated by 90 degrees towards the PMT. We measured the displacement by projecting a square light pattern and recording an image with the auxiliary

microscope with the dichroic positioned towards the CCD1 or the PMT. Identifying a corner of the projected square light pattern across the two images allowed using the secondary camera CCD2-to-DMD-chip registration to determine the dichroic induced offset in projection.

Lateral spot size measurement: To calculate the full-width at half maximum (FWHM) in the lateral direction (**Figs. 2a-c**), we projected a circular spot onto a rhodamine B-coated coverslip and acquired images with the auxiliary microscope. We thresholded each image and found the center of the spot by computing a weighted centroid of the thresholded image:

$$c = \operatorname{argmax}_x \sum_x I(x)x$$

where x is a vector containing the pixel coordinates and $I(x)$ is the brightness of the image at point x . The threshold was chosen heuristically by visual inspection of a single image and finding a value which isolated the light spot, and which, when changed by small amounts, did not significantly change the position of the centroid. We generated and averaged a set of line profiles passing through the centroid rotated in 1° increments over 360° . The resulting symmetric line profile was further used to compute the FWHM.

Axial spot size measurement: To measure the axial (z) profile of the projected light spots, we brought different axial sections of the spots into the focal plane of CCD2 by stepping the principal objective up and down over an interval of $\pm 100 \mu\text{m}$ (for spots of diameter $\leq 40 \mu\text{m}$) or over $\pm 200 \mu\text{m}$ range (for spots of diameter $> 40 \mu\text{m}$) in z increments of 1 and $2 \mu\text{m}$ respectively. We defined a region of interest (ROI) for analysis by taking all points within the FWHM of the spot in the central section. The axial profile was defined as the intensity within the ROI across all sections (**Figs. 2a-c**). To estimate the effect of spherical and other types of aberrations, we measured the

axial profile of our projected spots for photo-stimulation plane z offsets of $\Delta z = 0 - 400 \mu\text{m}$ with respect to the focal plane of the principal objective (**Fig. 3c**). For these measurements, we used spot diameters of 30, 60 and 90 μm respectively. Since in these configurations the diffuser is not positioned at a plane conjugate to the focal plane of the principal objective, any relative movement of the principal objective with respect to the tube lens L5 results in a slight change in magnification. The projected light spot size was corrected accordingly for the predicted increase in magnification.

Optical alignment. Alignment of the two-photon microscope followed the typical procedure¹⁰⁶. The photo-stimulation beam was further aligned on the DMD chip. In our implementation, the DMD chip was mounted onto an articulating base-ball stage (SL20, Thorlabs) that can be raised and lowered. We used a periscope (RS99, Thorlabs) formed by two mirrors placed on kinematic mounts to center the photo-stimulation laser beam on the DMD chip. By applying an all-white pattern to the DMD-chip, a coarse alignment was obtained with the light reflected off the DMD-chip into the microscope and the photo-stimulation beam centered on the chip. One point of potential difficulty in the alignment procedure is given by the diffractive behavior of the DMD-chip when illuminated with coherent light. The face of each individual micro-mirror is offset with respect to the plane of the entire array of mirrors. Because of this, the DMD-chip behaves like a blazed diffraction grating. We aligned the blue laser beam such that most of the power is sent into a particular diffraction order. For a one-dimensional grating, the angle θ_m of diffraction order m is given by:

$$m\lambda = d(\sin \theta_i + \sin \theta_m)$$

while θ_r , the angle at which the reflected laser power leaves the DMD chip is given by:

$$\theta_r = 2\alpha - \theta_i$$

where θ_i is the incident angle with respect to the normal of the grating, λ is the wavelength, d is the grating spacing, and α is the offset or “blaze angle” of the mirror face relative to the grating. Since the micro-mirrors rotate along one diagonal of the mirror, the projection of the correct incident angle onto the mirror face must lie orthogonal to this diagonal. This line was determined by toggling all the mirrors between the on and off positions (setting an all-white or all-black image) and observing where the light is redirected.

To make final adjustments to the position, we compared the power in the brightest diffraction order to the four adjacent ones. We then adjusted the incident angle of the beam by tilting or translating the DMD-chip mount and compensating with the periscope mirrors to keep the photo-stimulation beam centered until the power in the brightest spot was maximally bright relatively to its neighbors. After optimizing the incident angle, the brightest spot was $\sim 60\%$ as powerful as before the DMD chip. A circular iris (SM2D25, Thorlabs) was used to block out light from other diffraction orders. Applying a cross-shaped or small circular pattern to the DMD chip is useful for determining the center of the field of stimulation. Finally, the dichroic mirror (DM1, **Fig. 1**) was used to align the center of the photo-stimulation pattern to the same vertical line as the two-photon beam, and the primary camera (CCD1) and PMTs were placed and aligned as described above.

Propagation of patterns through brain tissue: We used the auxiliary microscope to characterize how the lateral profile of the one photon extended (2D) photo-stimulation patterns deteriorates after propagation through brain slices of different thicknesses, following the procedure described in *Papagiakoumou et al., 2013*⁷⁷. Adult mouse brains were fixed with paraformaldehyde and slices cut to various thicknesses. After focusing the photo-stimulation pattern on the fluorescent coverslip so that it was in focus on both the primary and the secondary cameras, we placed a slice on the top

of the coverslip with the pattern illuminating the olfactory bulb. The photo-stimulation light travelled through the slice and was scattered by the tissue before the pattern was imaged on the coverslip and the secondary camera (CCD2). We took a total of six measurements for each slice thickness (three positions, two slices). We used as reference for normalization the configuration when no slice was present (0 thickness) and defined the ROI for analysis as the set of points within the FWHM when no slice was present. We then calculated the light intensity and the standard deviation of pixel values within the ROI and compared it to values obtained for different slice thicknesses (**Figs. S2c-g**).

Phosphorescence artifacts. We observed a small slowly-decaying PMT signal upon illumination with the blue photo-stimulation laser in the absence of a fluorescence specimen. By progressively blocking off different parts of the optical path, we determined that the main contributors to this signal were the objective and tube lens (O and L5 in **Fig. 1**). Upon replacing the N-BK7 glass tube lens (LA1384-A) with a fused silica lens (LA4545-A) with the same focal length, the signal dropped substantially (**Fig. S1e**). The artifact was not affected by the type of anti-reflective coating used (**Fig. S1e**). We tested several objectives (Olympus 20x 1.00 NA XLUMPlanFLN, Olympus 20x 0.95 XLUMPlanFL, Olympus 25x 1.05NA XLPlanN) and determined that they contributed differentially to this signal. Of these, the 25x fared best followed by the 1.0 NA 20x and the 0.95 NA 20x (data not shown). Following this optimization, when comparing identified cells to background areas in our imaging experiments, we could not detect any spurious signals in background selected areas.

772 **Mice:** For the *in vivo* experiments described here, we included data from 12 OMP-Cre x ReChr x
773 Thy1GCaMP6s mice, 3 DAT-Cre mice and 4 DAT-Cre x Thy1-GCaMP6f mice. For experiments
774 to map mitral/tufted cell parent glomeruli and identify *sister* cell cohorts, triple mouse crosses
775 (OMP-Cre x Rosa26 CAG-LSL-ReaChR-mCitrine⁷⁹ x Thy1GCaMP6s_{4.3}⁸¹) were used to
776 express ReaChR-mCitrine and GCaMP6f/s in the OSNs and respectively mitral and tufted cells
777 (**Figs. 4,5**). For experiments expressing ChR2 and GCaMP6f in DAT+ cells (**Fig. S5a-e**), we used
778 mice expressing Cre recombinase in DAT+ neurons and approximately 300 nl of a mixture of
779 AAV (serotype 2.9, UNC Vector Core) virus carrying DIO-ChR2-mCherry and DIO-GCaMP6f
780 was injected into the olfactory bulb at a depth of 300 μm as previously described⁴³. For glomerular
781 stimulation, for each light stimulus other than control stimuli, we fit a function of the form $f(d) =$
782 $Ae^{-d/b}$ where f represents the light-evoked fluorescence change in a region of interest and d is the
783 distance from that region to the center of the projected light mask. The width constant b of this
784 exponential fit gives a measure of the spread of activity evoked by that stimulus. For each light
785 intensity sampled, we repeated the analysis in two different fields of view, and the relationship
786 between activation width and light intensity is shown in **Figs. S5b-e**. As in the case of OSN
787 terminals photo-stimulation experiments, we flashed single 100 ms sub-glomerular light masks
788 (20 μm diameter) between the imaging frames and recorded calcium transients in the DAT+ cells
789 via multiphoton imaging. For analysis, the field of view was divided into a 20 by 20 grid (15 x 15
790 μm grid units), and the response was defined as the average change in fluorescence in each unit of
791 the grid during the two frames (10 Hz) immediately following the light stimulus. We tested a range
792 of photo-stimulation light intensities ranging from 0.5 to 10.0 mW/mm². We tuned up the intensity
793 of the fluorescence response, while aiming to maintain spatial specificity, as indicated by the
794 photo-activated area remaining largely confined to the chosen glomerulus. We quantified the

spread of activity by characterizing the relationship between the DAT⁺ cell response strength and the distance from the centroid of the light stimulus, which was well-fit by exponential decay. Within the range of light intensity used, the characteristic width constant varied from 25 μ m for the lowest to 62 μ m for the highest.

For experiments investigating the impact on mitral/tufted cell activity of glomerular GABAergic/DAT⁺-neurons (superficial short axon cells), DAT-Cre mice were crossed to Thy1GCaMP6f_{5.11}⁸¹ mice (**Figs. 6, S6, S7**). DAT-Cre x Thy1-GCaMP6f mice were retro-orbitally injected¹⁰⁷ with 5×10^{12} vp of AAV-PHP.eB:Syn-FLEX-rc[ChrimsonR-tdtomato], which was prepared as previously described^{93,108}. Viral particles were diluted in saline to achieve a final volume of 100 μ L per injected mouse. All animal procedures conformed to NIH guidelines and were approved by the Animal Care and Use Committee of Cold Spring Harbor Laboratory.

Surgical procedures: Mice were injected with NSAID Meloxicam 0.5 mg/Kg (Metacam®, Boehringer Ingelheim, Ingelheim, Germany) 24 hours prior the surgical procedure, at the onset of surgery, and for 2 days *post* each surgical procedure. Depending on the recovery progression, the NSAID treatment was maintained until mice showed alert, and responsive behavior. Before each stereotaxic surgery, mice were anesthetized with 10% v/v Isoflurane (Cat# 029405, Covetrus, Portland, ME, US). For the chronic window and headbar implantation procedures, mice were anesthetized with a ketamine/xylazine (125 mg/Kg - 12.5 mg/Kg) cocktail. During surgery, the animal's eyes were protected with an ophthalmic ointment (Puralube®, Dechra, Northwich, England, UK). Temperature was maintained at 37 °C using a heating pad (FST TR-200, Fine Science Tools, Foster City, CA, USA). Respiratory rate and lack of pain reflexes were monitored

throughout the procedure. Chronic window implant surgeries were supplemented with dexamethasone (4 mg/Kg) to prevent swelling, enrofloxacin (5 mg/Kg) to prevent bacterial infection, and carprofen (5 mg/Kg) to reduce inflammation.

Chronic implantations: After recovery from the stereotaxic surgery, mice were implanted with a custom titanium head bar attached with C&B Metabond Quick adhesive luting cement (Cat# S380. Parkell, Edgewood, NY, USA), followed by black Ortho-JetTM dental acrylic application (Cat# 1520BLK. Lang. Chicago, IL, USA) and with a cranial window on top of the olfactory bulb as previously described^{109,110}. Special care was taken to remove the bone under the inter-frontal suture and remove small bone pieces at the edges. During surgery, the exposed olfactory bulb was continuously protected and cleaned of blood excess using artificial cerebrospinal fluid (aCSF) and aCSF-soaked gelfoam. Once both hemibulbs were exposed and clean, a fresh drop of aCSF was placed on top, followed by a 3 mm round cover glass, which was gently pushed onto the OB surface to minimize motion artifacts in further experiments. Once in place, the coverslip was sealed along the edges with a combination of VitrebondTM, Crazy-GlueTM, and dental acrylic to cover the exposed skull. Mice recovered for ~ 7 days before imaging experiments.

Odor delivery: Odors were presented using a custom-built odor machine as described elsewhere^{85,111}. Undiluted or oil diluted (1:10 or 1:100) odorants were placed in glass vials and airflow through the vials controlled by solenoid valves. The output odor stream was further mixed with clean air in a 50:1 ratio, then this diluted stream was further mixed with clean air at a 5:1 ratio before being sent through short teflon coated tubing placed in front of the mouse's snout. The 12 vials in the machine were split into three banks of four vials. Each bank could be independently

switched between sending odorized air and clean air to the mouse, in principle allowing combinations of up to three odors. For assessing the effect of modulating GABA-ergic/DAT+ interneurons onto mitral and tufted cells responses, we used the following odors: isoamyl acetate, acetaldehyde, valeraldehyde, allyl tiglate, ethyl heptanoate, heptanal, hexanol, 2-hexanone, ethyl tiglate, 1,4 cineole, acetophenone.

Data pre-processing

Movement correction and ROI selection: Rigid registration in Fiji/ImageJ (template matching) was applied to the acquired fluorescence time-lapse stacks. Images were visually inspected to select a motion-free sequence of frames and create a mean reference image to which we registered each image stack corresponding to a given trial. ROI selection was performed manually (ImageJ): we used both the average and standard deviation projections of the registered images to draw ROIs around the glomeruli and respectively mitral and tufted cell bodies in the FOV.

Bleaching correction: Post ROI extraction (glomeruli, mitral/tufted cells), we applied fluorescence bleaching correction, assuming that: (1) bleaching follows a first order exponential decay across trials for each individual ROI; (2) during the baseline period, any measured activity is random. Given these assumptions, by averaging the fluorescence traces across trials of an ROI, the only robust trend potentially present in the baseline is the characteristic bleaching for that ROI. We checked that the baseline fluorescence of a given ROI is comparable across trials within a session. The procedure used for each ROI was as follows: (1) we fitted an exponential decay function to the fluorescence traces of individual trials, (2) subtracted the fitted function from the

ROI trace, (3) averaged the fluorescence traces across all trials for that ROI (aligned to stimulus presentation).

Normalization: For each ROI and trial, we normalized the traces individually, using a fluorescence baseline period (average of ten frames just before the stimulus onset as baseline fluorescence, F_0). For each time point, fractional changes over the baseline fluorescence for the average trace were calculated as dF/F_0 .

ROI responsiveness and classification: On average, stimuli were presented across 4-5 repeats when imaging glomerular responses to optogenetic stimulation, and across 6-10 repeats when sampling the responses of mitral and tufted cells. To evaluate the responsiveness of each glomerulus, we obtained a distribution of average dF/F_0 values from baseline periods, and used 95 percentile value of the distribution as response (significant signal) threshold. The baseline reference distribution was obtained from averaging dF/F_0 values quantified over 10-frames intervals extracted from the baseline period and accumulated across all trials. For each ROI, we also obtained peak dF/F_0 values during the optogenetic stimulation frames (1-2 frames). For each ROI, we compared the average response across trials with the baseline reference distribution. If this value crossed the 95th percentile of the baseline distribution, the ROI was classified as responsive (*enhanced*).

To evaluate the responsiveness of each mitral and tufted cell, we searched for the peak of a monotonous response (signal) within 5 frames following optogenetic stimulation onset. Similarly, we obtained a distribution of average dF/F_0 values from baseline periods using 4 frames just before the stimulation and accumulating across all trials. We compared the peak response value to the

standard deviation (SD) of the baseline. For calling *enhanced* responses, we used 4 SD of the baseline as response (significant signal) threshold; for calling *suppressed* responses, we used 3 SD of the baseline as response threshold to account for the intrinsic lower range of negative deflections from baseline. Cells that had slow and prolonged negative slope ($<15^\circ$) with respect to baseline were not included in further analyses. To assess response reproducibility across trials, for each optogenetic stimulation mask, for each trial, we built a *trial response vector* (using 5-frames from the stimulus onset) and computed the Pearson correlation with respect to the response vectors of the other trials. If for a given trial the average correlation with respect to the response vectors of the other trials, the trial was rejected. For an ROI to be considered for further analysis, at least 2 repeats had to pass the criterion. Cells that had slow responses, which occurred only after offset of light stimulation were not included for further analysis to minimize the contribution of disinhibitory effects across the bulb network.

Self-response correlations: To obtain a measure of reliability for ensemble mitral cell responses across different trials to the same DAT+ glomerular photo-stimulation mask, and respectively to the same odorant, using bootstrapping (50 times) in each iteration split the trials into two random subsets of trials, and calculated ‘self’ response correlations, whose average values were 0.81 ± 0.14 for light masks and 0.87 ± 0.10 for odor responses of the same MC ensembles (**Fig. S5b**).

For each mitral and tufted cells, the *odor responses* (**Fig. 6**) were calculated as described above using 10-frames periods to build a distribution of baseline values and accumulating across trials. We also used 10-frames periods starting from stimulus onset to search for the response peak. To calculate the *light modulation of the odor response* to DAT+ glomerular photo-stimulation masks

(**Figs. 6f-h, S7**), we used as reference the frame just before light stimulation onset. We asked whether in the 2 frames following light onset, we could identify a significant increase/dip in signal compared to this reference dF/F value. A change in signal (defined delta between reference and signal during light stimulation) was considered significant if it amounted to more than 10% of the average peak odor response amplitude (preceding light onset), and was also larger than 3 SDs of the baseline fluorescence (preceding odor onset). To assess response reproducibility across trials, for each stimulus, we performed correlation analysis across trials as described above.

Assessing glomerular response spatial extent and specificity: In **Fig. S3c i** (Right), for each photo-stimulation mask, we normalized the sum total area of the significantly responsive glomerular pixels. We multiplied the sum total area of these responsive pixels by the ratio between the area of the photo-stimulation mask and the surface area of target glomerulus surface. We defined the anatomical outline of each glomerulus in the field of view manually using its baseline fluorescence. To determine the specificity of glomerular responses across light intensities we computed two metrics. First (blue trace, **Fig. 4f**), we normalized responsive pixels within the boundaries of the target glomerulus by all responsive pixels located within the boundaries of any glomeruli across the FOV. Second (red trace, **Fig. 4f**), we normalized the significantly responsive pixels within the target glomerulus by the sum area of all responsive pixels within the FOV, including those spanning the juxtaglomerular space, surrounding the glomeruli.

Correlation analysis and shuffled controls: For each pairwise correlation analysis (**Figs. 6e,h, S5b, S7e-f**), we ran 15 repetitions of the shuffling controls and compared the experimental data with the shuffled data (Wilcoxon signed-rank test).

932

933 **Statistical tests:** Depending on the properties of the analyzed data, different statistical tests were
 934 used to evaluate the differences between data groups. The information for each statistical
 935 comparison is detailed in the legends of each figure panel. ‘Two-sample unpaired Student’s t-test’
 936 was used to evaluate significant differences in the means of two normally distributed data sets.
 937 ‘The normality of the data distribution was checked using a ‘One-sample Kolmogorov-Smirnov
 938 test’ for each data group before the testing differences with the ‘Student’s t-test’. A two-tailed 95%
 939 confidence interval was used for both Student’s t-test. Significant differences between two non-
 940 normally distributed data sets were evaluated with a nonparametric ‘Wilcoxon signed-rank test.’
 941 The linear correlation between two data sets was assessed by computing the Pearson correlation
 942 coefficient. When needed, the coefficient of determination (R^2) was used to evaluate how well the
 943 compared data sets fit a linear regression model. All the statistical comparisons were tested using
 944 MATLAB® using the following functions for each test: ‘Two-sample unpaired Student’s t-test’:
 945 ttest; ‘One-sample Kolmogorov-Smirnov test’: kstest; ‘Multiple comparisons of means test’:
 946 multcompare; ‘Two-sample Kolmogorov-Smirnov test’: kstest2; ‘Pearson correlation coefficient’:
 947 corrcoef. ‘Wilcoxon signed-Rank’: signrank.

References

1. Scanziani, M. & Häusser, M. Electrophysiology in the age of light. *Nature* **461**, 930–939 (2009).
2. Emiliani, V., Cohen, A. E., Deisseroth, K. & Häusser, M. All-Optical Interrogation of Neural Circuits. *J. Neurosci.* **35**, 13917–13926 (2015).
3. Peron, S. & Svoboda, K. From cudgel to scalpel: toward precise neural control with optogenetics. *Nat. Methods* **8**, 30–34 (2011).
4. Prakash, R. *et al.* Two-photon optogenetic toolbox for fast inhibition, excitation and bistable modulation. *Nat. Methods* **9**, 1171–1179 (2012).
5. Adesnik, H. & Abdeladim, L. Probing neural codes with two-photon holographic optogenetics. *Nat. Neurosci.* **24**, 1356–1366 (2021).
6. Lutz, C. *et al.* Holographic photolysis of caged neurotransmitters. *Nat. Methods* **5**, 821–827 (2008).
7. Nikolenko, V. *et al.* SLM Microscopy: Scanless Two-Photon Imaging and Photostimulation with Spatial Light Modulators. *Front. Neural Circuits* **2**, 5 (2008).
8. Packer, A. M. *et al.* Two-photon optogenetics of dendritic spines and neural circuits. *Nat. Methods* **9**, 1202–1205 (2012).
9. Packer, A. M., Russell, L. E., Dagleish, H. W. P. & Häusser, M. Simultaneous all-optical manipulation and recording of neural circuit activity with cellular resolution in vivo. *Nat. Methods* **12**, 140–146 (2015).
10. Carrillo-Reid, L., Yang, W., Bando, Y., Peterka, D. S. & Yuste, R. Imprinting and recalling cortical ensembles. *Science* **353**, 691–694 (2016).

11. Yang, W., Carrillo-Reid, L., Bando, Y., Peterka, D. S. & Yuste, R. Simultaneous two-photon imaging and two-photon optogenetics of cortical circuits in three dimensions. *eLife* **7**, e32671 (2018).
12. Papagiakoumou, E. *et al.* Scanless two-photon excitation of channelrhodopsin-2. *Nat. Methods* **7**, 848–854 (2010).
13. Papagiakoumou, E., Ronzitti, E. & Emiliani, V. Scanless two-photon excitation with temporal focusing. *Nat. Methods* **17**, 571–581 (2020).
14. Villette, V. *et al.* Ultrafast Two-Photon Imaging of a High-Gain Voltage Indicator in Awake Behaving Mice. *Cell* **179**, 1590-1608.e23 (2019).
15. Fan, L. Z. *et al.* All-optical physiology resolves a synaptic basis for behavioral timescale plasticity. *Cell* **186**, 543-559.e19 (2023).
16. Anselmi, F., Ventalon, C., Bègue, A., Ogden, D. & Emiliani, V. Three-dimensional imaging and photostimulation by remote-focusing and holographic light patterning. *Proc. Natl. Acad. Sci. U. S. A.* **108**, 19504–19509 (2011).
17. Anselmi, F., Banerjee, A. & Albeanu, D. F. Patterned Photostimulation in the Brain. in *New Techniques in Systems Neuroscience* (ed. Douglass, A. D.) 235–270 (Springer International Publishing, Cham, 2015).
18. Andrasfalvy, B. K., Zemelman, B. V., Tang, J. & Vaziri, A. Two-photon single-cell optogenetic control of neuronal activity by sculpted light. *Proc. Natl. Acad. Sci. U. S. A.* **107**, 11981–11986 (2010).
19. Rickgauer, J. P., Deisseroth, K. & Tank, D. W. Simultaneous cellular-resolution optical perturbation and imaging of place cell firing fields. *Nat. Neurosci.* **17**, 1816–1824 (2014).

991 20. Grewe, B. F., Voigt, F. F., van 't Hoff, M. & Helmchen, F. Fast two-layer two-photon imaging
992 of neuronal cell populations using an electrically tunable lens. *Biomed. Opt. Express* **2**, 2035–
993 2046 (2011).

994 21. Sheffield, M. E. J. & Dombeck, D. A. Calcium transient prevalence across the dendritic arbour
995 predicts place field properties. *Nature* **517**, 200–204 (2015).

996 22. Iwai, D., Izawa, H., Kashima, K., Ueda, T. & Sato, K. Speeded-Up Focus Control of
997 Electrically Tunable Lens by Sparse Optimization. *Sci. Rep.* **9**, 12365 (2019).

998 23. Katona, G. *et al.* Fast two-photon in vivo imaging with three-dimensional random-access
999 scanning in large tissue volumes. *Nat. Methods* **9**, 201–208 (2012).

1000 24. Iyer, V., Losavio, B. E. & Saggau, P. Compensation of spatial and temporal dispersion for
1001 acousto-optic multiphoton laser-scanning microscopy. *J. Biomed. Opt.* **8**, 460 (2003).

1002 25. Duemani Reddy, G., Kelleher, K., Fink, R. & Saggau, P. Three-dimensional random access
1003 multiphoton microscopy for functional imaging of neuronal activity. *Nat. Neurosci.* **11**, 713–
1004 720 (2008).

1005 26. Salomé, R. *et al.* Ultrafast random-access scanning in two-photon microscopy using acousto-
1006 optic deflectors. *J. Neurosci. Methods* **154**, 161–174 (2006).

1007 27. Akemann, W. *et al.* Fast optical recording of neuronal activity by three-dimensional custom-
1008 access serial holography. *Nat. Methods* **19**, 100–110 (2022).

1009 28. Hernandez, O., Pietrajtis, K., Mathieu, B. & Dieudonné, S. Optogenetic stimulation of
1010 complex spatio-temporal activity patterns by acousto-optic light steering probes cerebellar
1011 granular layer integrative properties. *Sci. Rep.* **8**, 13768 (2018).

1012 29. Hernandez, O. *et al.* Three-dimensional spatiotemporal focusing of holographic patterns. *Nat.*
1013 *Commun.* **7**, 11928 (2016).

30. Oron, D., Papagiakoumou, E., Anselmi, F. & Emiliani, V. Two-photon optogenetics. *Prog. Brain Res.* **196**, 119–143 (2012).
31. Russell, L. E. *et al.* All-optical interrogation of neural circuits in behaving mice. *Nat. Protoc.* **17**, 1579–1620 (2022).
32. Ronzitti, E., Emiliani, V. & Papagiakoumou, E. Methods for Three-Dimensional All-Optical Manipulation of Neural Circuits. *Front. Cell. Neurosci.* **12**, 469 (2018).
33. Pégard, N. C. *et al.* Three-dimensional scanless holographic optogenetics with temporal focusing (3D-SHOT). *Nat. Commun.* **8**, 1228 (2017).
34. Petreanu, L., Huber, D., Sobczyk, A. & Svoboda, K. Channelrhodopsin-2–assisted circuit mapping of long-range callosal projections. *Nat. Neurosci.* **10**, 663–668 (2007).
35. Katz, L. C. & Dalva, M. B. Scanning laser photostimulation: a new approach for analyzing brain circuits. *J. Neurosci. Methods* **54**, 205–218 (1994).
36. Dhawale, A. K., Hagiwara, A., Bhalla, U. S., Murthy, V. N. & Albeanu, D. F. Non-redundant odor coding by sister mitral cells revealed by light addressable glomeruli in the mouse. *Nat Neurosci* **13**, 1404–1412 (2010).
37. Chong, E. *et al.* Manipulating synthetic optogenetic odors reveals the coding logic of olfactory perception. *Science* **368**, (2020).
38. Guo, Z. V., Hart, A. C. & Ramanathan, S. Optical interrogation of neural circuits in *Caenorhabditis elegans*. *Nat. Methods* **6**, 891–896 (2009).
39. Blumhagen, F. *et al.* Neuronal filtering of multiplexed odour representations. *Nature* **479**, 493–498 (2011).

40. Zhu, P., Fajardo, O., Shum, J., Zhang Schärer, Y.-P. & Friedrich, R. W. High-resolution optical control of spatiotemporal neuronal activity patterns in zebrafish using a digital micromirror device. *Nat. Protoc.* **7**, 1410–1425 (2012).
41. Wyart, C. *et al.* Optogenetic dissection of a behavioural module in the vertebrate spinal cord. *Nature* **461**, 407–410 (2009).
42. Zhu, P., Frank, T. & Friedrich, R. W. Equalization of odor representations by a network of electrically coupled inhibitory interneurons. *Nat. Neurosci.* **16**, 1678–1686 (2013).
43. Banerjee, A. *et al.* An Interglomerular Circuit Gates Glomerular Output and Implements Gain Control in the Mouse Olfactory Bulb. *Neuron* **87**, 193–207 (2015).
44. Leifer, A. M., Fang-Yen, C., Gershow, M., Alkema, M. J. & Samuel, A. D. T. Optogenetic manipulation of neural activity in freely moving *Caenorhabditis elegans*. *Nat. Methods* **8**, 147–152 (2011).
45. Arrenberg, A. B., Stainier, D. Y. R., Baier, H. & Huisken, J. Optogenetic control of cardiac function. *Science* **330**, 971–974 (2010).
46. Farah, N., Reutsky, I. & Shoham, S. Patterned optical activation of retinal ganglion cells. *Conf. Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. IEEE Eng. Med. Biol. Soc. Conf.* **2007**, 6368–6370 (2007).
47. Fast and light-efficient remote focusing for volumetric voltage imaging | Nature Communications. <https://www.nature.com/articles/s41467-024-53685-5#Bib1>.
48. Botcherby, E. J., Juškaitis, R., Booth, M. J. & Wilson, T. An optical technique for remote focusing in microscopy. *Opt. Commun.* **281**, 880–887 (2008).
49. Botcherby, E. J. *et al.* Aberration-free three-dimensional multiphoton imaging of neuronal activity at kHz rates. *Proc. Natl. Acad. Sci. U. S. A.* **109**, 2919–2924 (2012).

1058 50. Rapid 3D light-sheet microscopy with a tunable lens.
1059 <https://opg.optica.org/oe/fulltext.cfm?uri=oe-21-18-21010&id=260811>.

1060 51. Shain, W. J., Vickers, N. A., Goldberg, B. B., Bifano, T. & Mertz, J. Extended depth-of-field
1061 microscopy with a high-speed deformable mirror. *Opt. Lett.* **42**, 995–998 (2017).

1062 52. Remote z-scanning with a macroscopic voice coil motor for fast 3D multiphoton laser scanning
1063 microscopy. <https://opg.optica.org/boe/fulltext.cfm?uri=boe-7-5-1656&id=338741>.

1064 53. Shepherd, G. M. Synaptic organization of the mammalian olfactory bulb. *Physiol. Rev.* **52**,
1065 864–917 (1972).

1066 54. Wilson, R. I. & Mainen, Z. F. Early events in olfactory processing. *Annu. Rev. Neurosci.* **29**,
1067 163–201 (2006).

1068 55. Arneodo, E. M. *et al.* Stimulus dependent diversity and stereotypy in the output of an olfactory
1069 functional unit. *Nat. Commun.* **9**, 1347 (2018).

1070 56. Kikuta, S., Fletcher, M. L., Homma, R., Yamasoba, T. & Nagayama, S. Odorant response
1071 properties of individual neurons in an olfactory glomerular module. *Neuron* **77**, 1122–1135
1072 (2013).

1073 57. Padmanabhan, K. & Urban, N. N. Intrinsic biophysical diversity decorrelates neuronal firing
1074 while increasing information content. *Nat. Neurosci.* **13**, 1276–1282 (2010).

1075 58. Schwarz, D. *et al.* Architecture of a mammalian glomerular domain revealed by novel volume
1076 electroporation using nanoengineered microelectrodes. *Nat. Commun.* **9**, 1–14 (2018).

1077 59. Odour responses of sister projection neurons in the mouse olfactory bulb, revealed by
1078 correlative two-photon imaging and synchrotron X-ray holographic nano-tomography.
1079 <https://www.abstractsonline.com/pp8/#!/10892/presentation/40650>.

60. Functional interactions between olfactory circuits shapes the timing and structure of odor representations. <https://www.abstractsonline.com/pp8/#!/10892/presentation/40671>.
61. Tootoonian, S., Schaefer, A. T. & Latham, P. E. Sparse connectivity for MAP inference in linear models using sister mitral cells. *PLoS Comput. Biol.* **18**, e1009808 (2022).
62. Tan, J., Savigner, A., Ma, M. & Luo, M. Odor information processing by the olfactory bulb analyzed in gene-targeted mice. *Neuron* **65**, 912–926 (2010).
63. Buonviso, N. & Chaput, M. A. Response similarity to odors in olfactory bulb output cells presumed to be connected to the same glomerulus: electrophysiological study using simultaneous single-unit recordings. *J. Neurophysiol.* **63**, 447–454 (1990).
64. Schoppa, N. E. & Westbrook, G. L. Glomerulus-specific synchronization of mitral cells in the olfactory bulb. *Neuron* **31**, 639–651 (2001).
65. Chen, T.-W., Lin, B.-J. & Schild, D. Odor coding by modules of coherent mitral/tufted cells in the vertebrate olfactory bulb. *Proc. Natl. Acad. Sci. U. S. A.* **106**, 2401–2406 (2009).
66. Aungst, J. L. *et al.* Centre-surround inhibition among olfactory bulb glomeruli. *Nature* **426**, 623–629 (2003).
67. Zavitz, D., Youngstrom, I. A., Borisyuk, A. & Wachowiak, M. Effect of Interglomerular Inhibitory Networks on Olfactory Bulb Odor Representations. *J. Neurosci. Off. J. Soc. Neurosci.* **40**, 5954–5969 (2020).
68. Olsen, S. R., Bhandawat, V. & Wilson, R. I. Divisive normalization in olfactory population codes. *Neuron* **66**, 287–299 (2010).
69. Cleland, T. A. Early transformations in odor representation. *Trends Neurosci.* **33**, 130–139 (2010).

- 1102 70. Lyons-Warren, A. M. *et al.* Co-transmitting interneurons in the mouse olfactory bulb regulate
1103 olfactory detection and discrimination. *Cell Rep.* **42**, 113471 (2023).
- 1104 71. Kiyokage, E. *et al.* Molecular identity of periglomerular and short axon cells. *J. Neurosci. Off.*
1105 *J. Soc. Neurosci.* **30**, 1185–1196 (2010).
- 1106 72. Whitesell, J. D., Sorensen, K. A., Jarvie, B. C., Hentges, S. T. & Schoppa, N. E.
1107 Interglomerular lateral inhibition targeted on external tufted cells in the olfactory bulb. *J.*
1108 *Neurosci. Off. J. Soc. Neurosci.* **33**, 1552–1563 (2013).
- 1109 73. Liu, S., Plachez, C., Shao, Z., Puche, A. & Shipley, M. T. Olfactory bulb short axon cell release
1110 of GABA and dopamine produces a temporally biphasic inhibition-excitation response in
1111 external tufted cells. *J. Neurosci. Off. J. Soc. Neurosci.* **33**, 2916–2926 (2013).
- 1112 74. Liu, S., Puche, A. C. & Shipley, M. T. The Interglomerular Circuit Potently Inhibits Olfactory
1113 Bulb Output Neurons by Both Direct and Indirect Pathways. *J. Neurosci. Off. J. Soc. Neurosci.*
1114 **36**, 9604–9617 (2016).
- 1115 75. Papagiakoumou, E., de Sars, V., Oron, D. & Emiliani, V. Patterned two-photon illumination
1116 by spatiotemporal shaping of ultrashort pulses. *Opt. Express* **16**, 22039–22047 (2008).
- 1117 76. Bègue, A. *et al.* Two-photon excitation in scattering media by spatiotemporally shaped beams
1118 and their application in optogenetic stimulation. *Biomed. Opt. Express* **4**, 2869–2879 (2013).
- 1119 77. Papagiakoumou, E. *et al.* Functional patterned multiphoton excitation deep inside scattering
1120 tissue. *Nat. Photonics* **7**, 274–278 (2013).
- 1121 78. Fukunaga, I., Berning, M., Kollo, M., Schmaltz, A. & Schaefer, A. T. Two distinct channels
1122 of olfactory bulb output. *Neuron* **75**, 320–329 (2012).
- 1123 79. Hooks, B. M., Lin, J. Y., Guo, C. & Svoboda, K. Dual-Channel Circuit Mapping Reveals
1124 Sensorimotor Convergence in the Primary Motor Cortex. *J. Neurosci.* **35**, 4418–4426 (2015).

1125 80. Chen, T.-W. *et al.* Ultrasensitive fluorescent proteins for imaging neuronal activity. *Nature*
1126 **499**, 295–300 (2013).

1127 81. Dana, H. *et al.* Thy1 transgenic mice expressing the red fluorescent calcium indicator
1128 jRGECO1a for neuronal population imaging in vivo. *PloS One* **13**, e0205444 (2018).

1129 82. Mombaerts, P. Axonal wiring in the mouse olfactory system. *Annu. Rev. Cell Dev. Biol.* **22**,
1130 713–737 (2006).

1131 83. Chae, H. *et al.* Mosaic representations of odors in the input and output layers of the mouse
1132 olfactory bulb. *Nat. Neurosci.* **22**, 1306–1317 (2019).

1133 84. Fantana, A. L., Soucy, E. R. & Meister, M. Rat olfactory bulb mitral cells receive sparse
1134 glomerular inputs. *Neuron* **59**, 802–814 (2008).

1135 85. Gupta, P., Albeanu, D. F. & Bhalla, U. S. Olfactory bulb coding of odors, mixtures and sniffs
1136 is a linear sum of odor time profiles. *Nat. Neurosci.* **18**, 272–281 (2015).

1137 86. Adrian, E. D. Olfactory reactions in the brain of the hedgehog. *J. Physiol.* **100**, 459–473
1138 (1942).

1139 87. Galliano, E. *et al.* Embryonic and postnatal neurogenesis produce functionally distinct
1140 subclasses of dopaminergic neuron. *eLife* **7**, e32373 (2018).

1141 88. Economo, M. N., Hansen, K. R. & Wachowiak, M. Control of Mitral/Tufted Cell Output by
1142 Selective Inhibition among Olfactory Bulb Glomeruli. *Neuron* **91**, 397–411 (2016).

1143 89. Niessing, J. & Friedrich, R. W. Olfactory pattern classification by discrete neuronal network
1144 states. *Nature* **465**, 47–52 (2010).

1145 90. Wanner, A. A. & Friedrich, R. W. Whitening of odor representations by the wiring diagram
1146 of the olfactory bulb. *Nat. Neurosci.* **23**, 433–442 (2020).

91. Boyden, E. S., Zhang, F., Bamberg, E., Nagel, G. & Deisseroth, K. Millisecond-timescale, genetically targeted optical control of neural activity. *Nat Neurosci* **8**, 1263–1268 (2005).
92. Tian, L. *et al.* Imaging neural activity in worms, flies and mice with improved GCaMP calcium indicators. *Nat Meth* **6**, 875–881 (2009).
93. Chan, K. Y. *et al.* Engineered AAVs for efficient noninvasive gene delivery to the central and peripheral nervous systems. *Nat. Neurosci.* **20**, 1172–1179 (2017).
94. Klapoetke, N. C. *et al.* Independent optical excitation of distinct neural populations. *Nat. Methods* **11**, 338–346 (2014).
95. Kebschull, J. M. *et al.* High-Throughput Mapping of Single-Neuron Projections by Sequencing of Barcoded RNA. *Neuron* **91**, 975–987 (2016).
96. Chen, Y. *et al.* High-throughput sequencing of single neuron projections reveals spatial organization in the olfactory cortex. *Cell* **185**, 4117–4134.e28 (2022).
97. Petersen, C. C. H. Sensorimotor processing in the rodent barrel cortex. *Nat. Rev. Neurosci.* **20**, 533–546 (2019).
98. Han, Y. *et al.* The logic of single-cell projections from visual cortex. *Nature* **556**, 51–56 (2018).
99. Glickfeld, L. L., Andermann, M. L., Bonin, V. & Reid, R. C. Cortico-cortical projections in mouse visual cortex are functionally target specific. *Nat. Neurosci.* **16**, 219–226 (2013).
100. Orsolic, I., Rio, M., Masic-Flogel, T. D. & Znamenskiy, P. Mesoscale cortical dynamics reflect the interaction of sensory evidence and temporal expectation during perceptual decision-making. *Neuron* **109**, 1861–1875.e10 (2021).
101. Leinweber, M., Ward, D. R., Sobczak, J. M., Attinger, A. & Keller, G. B. A Sensorimotor Circuit in Mouse Cortex for Visual Flow Predictions. *Neuron* **96**, 1204 (2017).

102. Attinger, A., Wang, B. & Keller, G. B. Visuomotor Coupling Shapes the Functional Development of Mouse Visual Cortex. *Cell* **169**, 1291-1302.e14 (2017).
103. Schneider, D. M., Nelson, A. & Mooney, R. A synaptic and circuit basis for corollary discharge in the auditory cortex. *Nature* **513**, 189–194 (2014).
104. Schneider, D. M., Sundararajan, J. & Mooney, R. A cortical filter that learns to suppress the acoustic consequences of movement. *Nature* **561**, 391–395 (2018).
105. Campbell, R. A. A., Eifert, R. W. & Turner, G. C. Openstage: a low-cost motorized microscope stage with sub-micron positioning accuracy. *PloS One* **9**, e88977 (2014).
106. (PDF) Principles, Design, and Construction of a Two-Photon Laser Scanning Microscope for In Vitro and In Vivo Brain Imaging. https://www.researchgate.net/publication/233870537_Principles_Design_and_Construction_of_a_Two-Photon_Laser_Scanning_Microscope_for_In_Vitro_and_In_Vivo_Brain_Imaging.
107. Yardeni, T., Eckhaus, M., Morris, H. D., Huizing, M. & Hoogstraten-Miller, S. Retro-orbital injections in mice. *Lab Anim.* **40**, 155–160 (2011).
108. Challis, R. C. *et al.* Systemic AAV vectors for widespread and targeted gene delivery in rodents. *Nat. Protoc.* **14**, 379–414 (2019).
109. Otazu, G. H., Chae, H., Davis, M. B. & Albeanu, D. F. Cortical Feedback Decorrelates Olfactory Bulb Output in Awake Mice. *Neuron* **86**, 1461–1477 (2015).
110. Chae, H., Banerjee, A., Dussauze, M. & Albeanu, D. F. Long-range functional loops in the mouse olfactory system and their roles in computing odor identity. *Neuron* **0**, (2022).

1191 111. Gupta, P., Albeanu, D. F. & Bhalla, U. S. An odor delivery system for arbitrary time-
 1192 varying patterns of odors, mixtures and concentrations. *bioRxiv* 077875 (2016)
 1193 doi:10.1101/077875.