

RESEARCH ARTICLE

Stability of serum ferritin measured by immunoturbidimetric assay after storage at -80°C for several years

Anne-Sylvia Sacri^{1,2,3*}, Daniela Ferreira⁴, Babak Khoshnood¹, Laurent Gouya⁵, Henrique Barros⁴, Martin Chalumeau^{1,2}

1 INSERM UMR1153, Centre de Recherche en Épidémiologie et Statistique Sorbonne Paris Cité (CRESS), équipe Épidémiologie Périnatale, Obstétricale et Pédiatrique (ÉPOPé); Labex GR-Ex; Université Paris Descartes; Paris, France, **2** Department of General Pediatrics and Pediatric Infectious Diseases, Necker-Enfants malades hospital, AP-HP; Université Paris Descartes, Paris, France, **3** Université Paris Diderot, Sorbonne Paris Cité, Paris, France, **4** EPIUnit—Instituto de Saúde Pública, Universidade do Porto, Porto, Portugal, **5** Centre de recherche sur l'inflammation, INSERM UMR 1149, Université Paris Diderot; ERL CNRS 8252, Faculté de Médecine site Bichat, Labex GR-Ex, Paris, France

* anne-sylvia.sacri@inserm.fr



OPEN ACCESS

Citation: Sacri A-S, Ferreira D, Khoshnood B, Gouya L, Barros H, Chalumeau M (2017) Stability of serum ferritin measured by immunoturbidimetric assay after storage at -80°C for several years. PLoS ONE 12(12): e0188332. <https://doi.org/10.1371/journal.pone.0188332>

Editor: James R. Connor, Pennsylvania State University College of Medicine, UNITED STATES

Received: August 24, 2017

Accepted: November 3, 2017

Published: December 11, 2017

Copyright: © 2017 Sacri et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: All relevant data are within the paper and its Supporting Information files.

Funding: This work was supported by the Secteur Français des Aliments de l'Enfance <https://www.secteurfrancaisdesalimentsdelenfance.com/> (ASS, MC); French Ministry of Health "DGOS PHRC régional 2014 n° AOR14053" (<http://solidarites-sante.gouv.fr/>) to ASS, MC and the Paris Diderot University - Sorbonne Paris Cité (<https://www.univ-paris-diderot.fr/>) to ASS. The funders had no role in

Abstract

Background

Iron deficiency (ID) may impair long-term neurological development when it occurs in young infants. In cohort studies, it is sometimes necessary to evaluate ID with sera kept frozen for several years. To assess ID, learned societies recommend measuring serum ferritin (SF) level combined with C-reactive protein level. The long-term stability of C-reactive protein in frozen samples is well established but not ferritin.

Methods

We measured SF level (immunoturbidimetric assay; in micrograms per liter) immediately after collection from 53 young adults recruited and followed-up in Porto, Portugal, from 2011 to 2013 (SF₁), and then, in 2016 in two aliquots kept frozen at -80°C for 3 to 5 years: one without (SF_{2A}) and one with (SF_{2B}) intermediate thawing in 2014. We compared SF₁ to SF_{2A} then SF_{2B}; statistical agreement was evaluated by the Bland and Altman method and the effect of intermediate thawing by regression modelling.

Results

Mean SF_{2A}–SF₁ and SF_{2B}–SF₁ differences were -2.1 (SD 7.0) and 48.9 (SD 66.9). Values for Bland and Altman 95% limits of agreement were higher for the comparison of SF_{2B} and SF₁ than SF_{2A} and SF₁: -82.2 to 179.9 and -15.8 to 11.8 , respectively; the effect of thawing was highly significant ($p < 0.001$).

Conclusions

Agreement between SF values before and after 3 to 5 years of constant freezing at -80°C was in a generally accepted range, which supports the hypothesis of ferritin's stability at this

study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Competing interests: Two authors (Anne-Sylvia Sacri and Martin Chalumeau) declare financial competing interest by receiving research grant from the commercial source "Secteur Français des Aliments de l'Enfance" for another research project. This funder has no role in present study design, data collection and analysis, preparation of the manuscript, and decision to submit for publication. This competing interest does not alter our adherence to PLoS ONE policies on sharing data and materials.

temperature for a long period. In long-term storage by freezing, intermediate thawing induced a major increase in values.

Introduction

Iron deficiency (ID) is considered the most frequent micronutrient deficiency worldwide, including in industrialized countries [1], and is suspected to be associated with adverse short- and long-term neurocognitive sequelae when it occurs in young people [2–5]. ID is a target of various primary prevention strategies and the subject of many studies [6–9]. In cohort studies, including birth ones, it is sometimes necessary to evaluate ID with sera stored for several years, with or without intermediate thawing. The World Health Organization, American Academy of Pediatrics and European Food Safety Authority recommend using serum ferritin (SF) level to measure ID [1, 6, 9]. SF well reflects body iron stores [10], and its level decreases in the first “pre-latent” stage of ID, which involves a reduction in the body’s iron stores without any impairment of erythropoiesis [8, 10]. SF level increases with inflammation, and recommendations suggest coupling its measurement with that of C-reactive protein (CRP), an acute-phase inflammation protein [9, 11].

Comparability of CRP levels after freezing (-80°C to -70°C) during long periods (7 to 14 years [12–14]) is well established, but few data are available for SF levels. SF levels reported after storage at positive temperature varies widely during short periods (from 4 hrs to 2 weeks) [15–17] but SF comparability is considered acceptable after storage at -20°C , -70°C , -80°C and -196°C for 4 days to 12 months [18–20] (S1 Table). A few studies have compared SF levels after a freezing–thawing procedure or after a storage at positive temperature during long periods to those after immediate freezing. They reported comparable SF levels measures except if stored at $+32^{\circ}\text{C}$ [21–23]. However, in these studies, the time interval between the two measures was 13 months maximum, a short period compared to the times commonly encountered in birth-cohort or large-scale cohort studies [24]. Only one study compared SF levels after storage at -25°C during 2 years and 25 years with 1-month-old samples [25] and found notable percentage differences in mean SF level values: -12.1% and -18.5% after 2 and 25 years, respectively. Such alarming results could discourage the use of ferritin to evaluate ID in sera kept frozen on the long-term (>12 months). However, the data were obtained from non-paired samples. Furthermore, no data have been published on long-term stability at -80°C nor on the effect of intermediate thawing, which may occur during long storage. Finally, in studies evaluating the long-term stability of ferritin, comparability was assessed by correlation, coefficient of variation (CV) or mean differences and rarely with recommended agreement measures such as the Bland and Altman method [26].

The objective of this work was to evaluate the comparability of SF levels after storage for several years by freezing at -80°C and the impact of intermediate thawing on these conditions, using recommended agreement evaluation methods.

Methods

General methodology and participants

The present study is an ancillary analysis of an adolescent cohort (Epidemiological Health Investigation of Teenagers in Porto [EPITeen]) performed in Porto, Portugal in 2003–2004 [27, 28]. It is based on frozen serum samples from 53 participants who were young adults in 2011–2013. Parents of included adolescents gave their written consent after receiving

information from the investigator, and written informed consent was obtained from all adult participants. The study was approved by the Ethics Committee of the University Hospital of São João (Porto, Portugal).

Samples and measurements

For the present ancillary study, serum samples from the initial population in EPITeen were selected after considering the following criteria regarding the participants and the samples: a measurement of SF performed at the 2011–2013 follow-up and at least four frozen aliquots kept at -80°C . Then, samples were randomly selected among each of five groups of the initial SF levels— <10 ; 10–29, 30–99, 100–199, and ≥ 200 $\mu\text{g/L}$ —to allow for representation of extreme values. SF levels were measured by immunoturbidimetric assay (Beckman Coulter, Krefeld, Germany) and were reported in micrograms per liter. According to the manufacturer's information, the ferritin procedure is linear from 8 to 450 $\mu\text{g/L}$ with recovery within 10% or 3 $\mu\text{g/L}$; the “within-run” precision and “total” precision SD are 1.15 to 1.89 and 1.21 to 4.36, respectively [29]; the limit of detection is < 4.6 $\mu\text{g/L}$ and limit of quantification < 7.8 $\mu\text{g/L}$. In 2011–2013, blood samples were collected in VACUMED 16x100 mm tubes with an inert separator gel and a clot activator and centrifuged within 2 hrs after collection. Several serum aliquots were separated for immediate storage at -80°C (including aliquots A and B), and one aliquot (named 1) was immediately sent to the laboratory for analysis, as defined by the project protocol. With this aliquot 1, SF level was measured within 6 hrs after blood collection (SF_1).

For the present study (S1 Fig), we selected a sample of aliquots, one that had not been frozen corresponding to aliquots 1 ($n = 53$), then the matched sample of aliquots A, frozen constantly until 2016 ($n = 53$), and another matched sample of aliquots B that had undergone intermediate thawing within 24 hrs, in 2014 ($n = 28$). This thawing was justified by the need for another laboratory assay and was followed by a new cycle of storage at -80°C that took place no longer than 24 hrs after thawing. Finally, in 2016, after 3 to 5 years of storage, aliquots A and B were thawed on the day of the current assay, at ambient temperature, before a second measure of SF (subsequently named SF_{2A} and SF_{2B} for aliquots A and B) immediately performed with the same method as for SF_1 .

Statistical analyses

First, we analyzed the comparability of SF levels before and after long-term freezing without intermediate thawing by describing SF_1 and SF_{2A} ($n = 53$ pairs); calculating their mean difference ($\text{SF}_{2A} - \text{SF}_1$), its SD, and the CV between SF_{2A} and SF_1 ; and by building an equality line graph. SF_1 and SF_{2A} distributions were compared by Student t test for paired samples. Agreement was evaluated by calculating the intraclass correlation coefficient (ICC) in a two-way mixed-effect model [30] and by the Bland and Altman method (calculating limits of agreement [LOA] with a Bland-Altman difference plot, as recommended [26]). Then, we compared the proportion of ID among SF_1 and SF_{2A} distributions after dichotomization around the classical 15- $\mu\text{g/L}$ threshold [1] by using the McNemar test for paired samples.

Second, we analyzed the comparability of SF levels before and after long-term freezing with intermediate thawing by comparing SF_1 and SF_{2B} ($n = 28$ pairs) using the same statistical approach described above.

Third, the quantitative effect of thawing on the overall comparability of SF levels after long-term freezing was evaluated by using a linear regression model after testing the deviance to linearity. For this, we defined a new variable, SF_2 , that corresponded to SF_{2B} if available ($n = 28$ sera) or to SF_{2A} otherwise ($n = 25$ sera). We built a linear regression model with SF_1 as the dependent variable and the independent variables SF_2 , thawing (binary variable “yes/no”) and

an interaction term between “thawing” and SF₂. The ICC between SF₁, SF_{2A} and SF_{2B} in a two-way mixed-effect model was also calculated for the 28 participants who had 3 SF level measures available.

The analyses involved use of Stata/SE 13.1 (StataCorp, USA).

Post-hoc supplementary analyses

Given the results on the effect of thawing in the case of long-term storage by freezing, we performed supplementary experiments to explore its effect during short-term storage by freezing. We conducted the experiment on a convenience sample of 6 participants. SF dosages were performed in the same laboratory and with the same techniques as the main study. For each participant, 4 SF dosages were made on three aliquots: one immediately after blood sampling (SF_A), one after continuous freezing at -80°C during 7 days (SF_B) and two on the same aliquot on day 3 (SF_C) and day 7 (SF_D) after freezing at -80°C with an intermediate thawing at day 3. We performed similar statistical analyses as the ones performed for the main study.

Results

The mean age of the 53 initial participants was 22.1 years (SD 0.4) and 49% were males.

There were 53 paired values SF₁ and SF_{2A}. The mean (SD, range) values for SF₁ and SF_{2A} were 108.6 µg/L (117.6, 6.6–429.3) and 104.6 µg/L (117.2, 7.1–443), respectively. The mean (SD) SF_{2A}–SF₁ difference was -2.1 µg/L (7.0, CV 2%). The equality line graph for SF_{2A} versus SF₁ is in [S2 Fig](#). The P-value from the *t* test for paired samples was 0.04; the ICC was 0.998 (95% confidence interval [CI] 0.997–0.999); and the Bland and Altman 95% LOA was -15.8 to 11.8 ([Fig 1A](#)). The ID proportions were 25% (95% CI 15–38) and 32% (95% CI 21–46) for SF₁ and SF_{2A}, respectively (mean difference in proportion SF_{2A}–SF₁: 8%, 95% CI 3–19; P_{McNemar test} = 0.046).

There were 28 paired values for SF₁ and SF_{2B}. The mean (SD, range) value for SF_{2B} was 135.2 µg/L (166.6, 7.5–710.9). The mean (SD) SF_{2B}–SF₁ difference was 48.9 µg/L (66.9, CV 30%). The equality line graph of SF_{2B} vs SF₁ is in [S2 Fig](#). The P-value of the *t* test for paired samples was < 0.001; the ICC was 0.88 (95% CI 0.76–0.94); and the Bland and Altman 95% LOA was -82.2 to 179.9 ([Fig 1B](#)). Among these 28 paired values, ID proportions were 32% (95% CI 17–52) and 25% (95% CI 12–45) for SF₁ and SF_{2B}, respectively (mean difference in proportion SF_{2B}–SF₁: -7%, 95% CI -26 to -2; P_{McNemar test} = 0.16).

The regression model found a strong interaction of thawing with SF₂ in the relationship between SF₁ and SF₂ (P-value for the interaction term <0.001). Estimates of the model were SF₁ = 4.38 + 0.99xSF₂ without intermediate thawing and SF₁ = 3.96 + 0.61xSF₂ with intermediate thawing. The ICC between SF₁, SF_{2A} and SF_{2B} in a two-way mixed-effects model (n = 28) was 0.91 (95% CI 0.84–0.95).

In post-hoc supplementary analyses, the participants' mean age was 27.0 years (SD 0.1). The mean (SD, range) values for SF_A, SF_B, SF_C and SF_D were 124.8 µg/L (121.1, 34.9–363.2), 124.0 µg/L (121.9, 31.9–363.4), 122.6 µg/L (120.8, 31–359.9), and 123.1 µg/L (121.7, 30–362.5), respectively. The mean (SD) differences values for SF_B–SF_A, SF_C–SF_A and SF_D–SF_A were -0.8 µg/L (1.8), -2.2 µg/L (1.7), and -1.6 µg/L (2.3), respectively, and the Bland and Altman 95% LOA were -4.3 to 2.7, -5.5 to 1.1, and -6.1 to 2.8, respectively.

Discussion

In this first evaluation of the comparability of SF levels in the long term, the agreement between immediate measures and those after freezing at -80°C varied to a limited extent, with a mean (SD) difference of -2.1 µg/L (7.0, CV 2%), ICC of 0.998 (95% CI 0.997–0.999), and




Fig 1. A. Bland-Altman graph: mean of serum ferritin (SF) at the first (SF₁) and second measure without intermediate thawing (SF_{2A}). Dotted line: mean difference between values of SF₁ and SF_{2A}. Grey zone: 95% limits of agreement for mean difference between SF₁ and SF_{2A}. **B. Bland-Altman graph: mean of serum ferritin (SF) at the first (SF₁) and second measure with intermediate thawing (SF_{2B}).** Dotted line: mean difference between values of SF₁ and SF_{2B}. Grey zone: 95% limits of agreement for mean difference between SF₁ and SF_{2B}.

<https://doi.org/10.1371/journal.pone.0188332.g001>

Bland and Altman 95% LOA from -15.8 to 11.8 µg/L. Our results are consistent with previous reports of the comparability of SF levels after storage at -20°C, -70°C, -80°C and -196°C during shorter periods [18–20]. The mean difference observed in our study is much smaller than variations observed in the only available study reporting comparability after storage for 2 years [25] (S1 Table). The lower comparability observed in this latter study may be due to differences in storage temperatures (-25°C vs -80°C in our study) and the study’s non-paired design [25].

The variation in SF levels reported here was statistically significant but may be considered acceptable as compared with the repeatability precision and the within-laboratory precision CV provided by the manufacturer (0.4% to 4.7% and 1.0% to 4.9%, respectively) and the CV observed in a national quality insurance study that analyzed 2391 French private laboratories: 12.7% and 8.9% around 50 and 320 µg/L, respectively [31]. The absolute value of the mean difference, -2.1 µg/L, could have more impact when ferritin is dichotomized around low thresholds to study iron depletion than in studies of iron overload. Indeed, in our study, the iron depletion/deficiency proportion after long-term constant freezing was modified from 25%

(13/53) to 32% (17/53) without intermediate thawing vs. from 32% (9/28) to 25% (7/28) with intermediate thawing. In the first case, 4 participants were reclassified as having ID according to SF_{2A} dosage and in the second case, 2 participants were reclassified as not having ID according to SF_{2B} dosage. These results suggest that when using SF measures from biobanks stored by freezing during several years, one should be cautious about diagnosis of ID.

Intermediate thawing had a major impact on the SF levels comparability after long-term storage by freezing, with a mean difference of 48.9 µg/L. The effect of thawing was greater with higher values of SF, as highlighted by the strong interaction identified in the regression model. Thus, based on these results, the use of SF levels measures after long-term freezing at -80°C and intermediate thawing should be discouraged.

The strengths of the present work are that all analyses were performed with the same technique in the same laboratory, with standardized protocols for storage, freezing and thawing, with a large range of SF values usually encountered in clinical practice (from 6.6 to 429.3 µg/L). We used several recommended statistical approaches to study the reproducibility of measurements [26, 32]. The design used did not allow for comparing variations in SF levels after long-term freezing to those observed during a short-time freezing period. However, in our post-hoc supplementary analyses, we compared SF measures on a convenience sample of sera of 6 participants after short-term freezing (seven days) with and without intermediate thawing. The variations observed in SF levels were not clinically significant, with or without intermediate thawing (for initial SF > 30 µg/L). This suggests that intermediate thawing in the case of short-term freezing may not affect ferritin measures. This is consistent with the results reported by Gonzales-Gross et al. and Brinc et al. [22, 23]. Gonzales-Gross found that the mean difference in SF measures on aliquots frozen at -20°C continuously during 20 days after blood extraction vs with discontinuous freezing due to transport was -0.82 µg/L (95% CI -8.5 to +7.5) [22]. Brinc found that the variation in SF measures in aliquots kept frozen at -80°C during 13 months with a monthly freeze-thaw cycle before analysis were not significant (percentage change from baseline: ±9) [23].

An important limitation of our study is its retrospective design that did not allow for evaluating the variations in SF by repeated measures without freezing or by a replication study (with storage by short-term and long-term freezing, with or without intermediate thawing). Some hypotheses can be proposed to explain SF variations. Precipitation and desiccation are possible but unlikely, given the use of a standard procedure such as the Vortex system and the sealing of tube joints.

In conclusion, the agreement between SF values before and after 3 to 5 years of constant freezing at -80°C was in a generally accepted range, which supports the hypothesis of ferritin's stability at this temperature for a long period. In long-term storage by freezing, intermediate thawing induced a major increase in values.

Supporting information

S1 Dataset. Dataset for our main analyses.
(XLSX)

S2 Dataset. Dataset for our post-hoc supplementary analyses.
(XLSX)

S1 Fig. Flow-chart.
(TIF)

S2 Fig. Equality line between the first serum ferritin measure (SF₁) and the second measure without intermediate thawing (SF_{2A}) or with intermediate thawing (SF_{2B}). Full circles:

values for SF_{2A} according to SF₁ values (n = 53) with the equality line (full line). Hollow circles: values for SF_{2B} according to SF₁ values (n = 28) with the equality line (dashed line). (TIF)

S1 Table. Comparability of biomarker levels used for iron deficiency diagnosis in the literature after various storage modalities.

(DOCX)

Author Contributions

Conceptualization: Anne-Sylvia Sacri, Babak Khoshnood, Laurent Gouya, Henrique Barros, Martin Chalumeau.

Data curation: Daniela Ferreira, Henrique Barros.

Formal analysis: Anne-Sylvia Sacri, Babak Khoshnood, Martin Chalumeau.

Funding acquisition: Anne-Sylvia Sacri, Henrique Barros, Martin Chalumeau.

Investigation: Daniela Ferreira, Henrique Barros.

Methodology: Anne-Sylvia Sacri, Babak Khoshnood, Laurent Gouya, Henrique Barros, Martin Chalumeau.

Project administration: Henrique Barros, Martin Chalumeau.

Resources: Daniela Ferreira, Henrique Barros.

Supervision: Martin Chalumeau.

Validation: Laurent Gouya, Henrique Barros, Martin Chalumeau.

Visualization: Anne-Sylvia Sacri.

Writing – original draft: Anne-Sylvia Sacri, Martin Chalumeau.

Writing – review & editing: Daniela Ferreira, Babak Khoshnood, Laurent Gouya, Henrique Barros.

References

1. World Health Organization. Iron deficiency anaemia. Assessment, prevention and control. A guide for programme managers. Geneva: World Health Organization (WHO); 2001 [cited May 2017 2016]. World Health Organization (WHO):[Available from: http://www.who.int/nutrition/publications/en/ida_assessment_prevention_control.pdf.
2. Lozoff B, Jimenez E, Wolf AW. Long-term developmental outcome of infants with iron deficiency. *N Engl J Med* 1991; 325:687–694. <https://doi.org/10.1056/NEJM199109053251004> PMID: 1870641
3. Wang B, Zhan S, Gong T, Lee L. Iron therapy for improving psychomotor development and cognitive function in children under the age of three with iron deficiency anaemia. *Cochrane Database Syst Rev* 2013; 6:CD001444.
4. Roncagliolo M, Garrido M, Walter T, Peirano P, Lozoff B. Evidence of altered central nervous system development in infants with iron deficiency anemia at 6 mo: delayed maturation of auditory brainstem responses. *Am J Clin Nutr* 1998; 68:683–690. PMID: 9734748
5. Moffatt ME, Longstaffe S, Besant J, Dureski C. Prevention of iron deficiency and psychomotor decline in high-risk infants through use of iron-fortified infant formula: a randomized clinical trial. *J Pediatr* 1994; 125:527–534. PMID: 7523647
6. EFSA. European Food Security Authority. Panel on Dietetic Products, Nutrition and Allergies (NDA). Scientific Opinion on nutrient requirements and dietary intakes of infants and young children in the European Union. *European Food Security Authority (EFSA) Journal [serial online]* 2013; 11:3408–3511.
7. Hercberg S, Preziosi P, Galan P. Iron deficiency in Europe. *Public Health Nutr* 2001; 4:537–545. PMID: 11683548

8. Domellof M, Braegger C, Campoy C, Colomb V, Decsi T, Fewtrell M, et al. Iron requirements of infants and toddlers. *J Pediatr Gastroenterol Nutr* 2014; 58:119–129. <https://doi.org/10.1097/MPG.000000000000206> PMID: 24135983
9. Baker RD, Greer FR. Diagnosis and prevention of iron deficiency and iron-deficiency anemia in infants and young children (0–3 years of age). *Pediatrics* 2010; 126:1040–1050. <https://doi.org/10.1542/peds.2010-2576> PMID: 20923825
10. Zimmermann MB, Hurrell RF. Nutritional iron deficiency. *Lancet* 2007; 370:511–520. [https://doi.org/10.1016/S0140-6736\(07\)61235-5](https://doi.org/10.1016/S0140-6736(07)61235-5) PMID: 17693180
11. Assessing the iron status of populations. Joint World Health Organization/Centers for Disease Control and Prevention technical consultation on the assessment of iron status at the population level. Geneva, Switzerland: World Health Organization (WHO) Press; 2007 [cited March 2017 2016]. 2nd:[Available from: http://www.who.int/nutrition/publications/micronutrients/anaemia_iron_deficiency/9789241596107/en/].
12. Doumatey AP, Zhou J, Adeyemo A, Rotimi C. High sensitivity C-reactive protein (Hs-CRP) remains highly stable in long-term archived human serum. *Clin Biochem* 2014; 47:315–318. <https://doi.org/10.1016/j.clinbiochem.2013.12.014> PMID: 24373927
13. Nilsson TK, Boman K, Jansson JH, Thogersen AM, Berggren M, Broberg A, et al. Comparison of soluble thrombomodulin, von Willebrand factor, tPA/PAI-1 complex, and high-sensitivity CRP concentrations in serum, EDTA plasma, citrated plasma, and acidified citrated plasma (Stabilyte) stored at -70 degrees C for 8–11 years. *Thromb Res* 2005; 116:249–254. <https://doi.org/10.1016/j.thromres.2004.12.005> PMID: 15935834
14. Ishikawa S, Kayaba K, Gotoh T, Nakamura Y, Kario K, Ito Y, et al. Comparison of C-reactive protein levels between serum and plasma samples on long-term frozen storage after a 13.8 year interval: the JMS Cohort Study. *J Epidemiol* 2007; 17:120–124. PMID: 17641447
15. Kubasik NP, Ricotta M, Hunter T, Sine HE. Effect of duration and temperature of storage on serum analyte stability: examination of 14 selected radioimmunoassay procedures. *Clin Chem* 1982; 28:164–165. PMID: 7034999
16. Henriksen LO, Faber NR, Moller MF, Nexø E, Hansen AB. Stability of 35 biochemical and immunological routine tests after 10 hours storage and transport of human whole blood at 21 degrees C. *Scand J Clin Lab Invest* 2014; 74:603–610. <https://doi.org/10.3109/00365513.2014.928940> PMID: 24988314
17. Tanner M, Kent N, Smith B, Fletcher S, Lewer M. Stability of common biochemical analytes in serum gel tubes subjected to various storage temperatures and times pre-centrifugation. *Ann Clin Biochem* 2008; 45:375–379. <https://doi.org/10.1258/acb.2007.007183> PMID: 18583622
18. van Eijnsden M, van der Wal MF, Hornstra G, Bonsel GJ. Can whole-blood samples be stored over 24 hours without compromising stability of C-reactive protein, retinol, ferritin, folic acid, and fatty acids in epidemiologic research? *Clin Chem* 2005; 51:230–232. <https://doi.org/10.1373/clinchem.2004.042234> PMID: 15613719
19. Jansen EH, Beekhof PK, Schenk E. Long-term stability of biomarkers of the iron status in human serum and plasma. *Biomarkers* 2013; 18:365–368. <https://doi.org/10.3109/1354750X.2013.781223> PMID: 23627617
20. Mathew G, Zwart SR, Smith SM. Stability of blood analytes after storage in BD SST tubes for 12 mo. *Clin Biochem* 2009; 42:1732–1734. <https://doi.org/10.1016/j.clinbiochem.2009.07.015> PMID: 19631634
21. Drammeh BS, Schleicher RL, Pfeiffer CM, Jain RB, Zhang M, Nguyen PH. Effects of delayed sample processing and freezing on serum concentrations of selected nutritional indicators. *Clin Chem* 2008; 54:1883–1891. <https://doi.org/10.1373/clinchem.2008.108761> PMID: 18757584
22. Gonzalez-Gross M, Breidenassel C, Gomez-Martinez S, Ferrari M, Beghin L, Spinneker A, et al. Sampling and processing of fresh blood samples within a European multicenter nutritional study: evaluation of biomarker stability during transport and storage. *Int J Obes (Lond)* 2008; 32 Suppl 5:S66–75.
23. Brinc D, Chan MK, Venner AA, Pasic MD, Colantonio D, Kyriakopoulou L, et al. Long-term stability of biochemical markers in pediatric serum specimens stored at -80 degrees C: a CALIPER Substudy. *Clin Biochem* 2012; 45:816–826. <https://doi.org/10.1016/j.clinbiochem.2012.03.029> PMID: 22510430
24. Sacri AS, Hercberg S, Gouya L, Levy C, Bocquet A, Blondel B, et al. Very low prevalence of iron deficiency among young French children: A national cross-sectional hospital-based survey. *Matern Child Nutr* 2017 May 03. <https://doi.org/10.1111/mcn.12460> [Epub ahead of print]. PMID: 28466606
25. Gislefoss RE, Grimsrud TK, Morkrid L. Stability of selected serum proteins after long-term storage in the Janus Serum Bank. *Clin Chem Lab Med* 2009; 47:596–603. <https://doi.org/10.1515/CCLM.2009.121> PMID: 19290843
26. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986; 1:307–310. PMID: 2868172

27. Lourenco S, Costa L, Rodrigues AM, Carnide F, Lucas R. Gender and psychosocial context as determinants of fibromyalgia symptoms (fibromyalgia research criteria) in young adults from the general population. *Rheumatology (Oxford)* 2015; 54:1806–1815.
28. Ramos E, Barros H. Family and school determinants of overweight in 13-year-old Portuguese adolescents. *Acta Paediatr* 2007; 96:281–286. PMID: [17429921](#)
29. Chesher D. Evaluating assay precision. *Clin Biochem Rev* 2008; 29 Suppl 1:S23–26.
30. Shrout PE, Fleiss JL. Intraclass correlations: uses in assessing rater reliability. *Psychol Bull* 1979; 86:420–428. PMID: [18839484](#)
31. [Annals of national quality control of medical biology analyzes. Biochemistry 05BIO1. Uric acid, Glucose, Urea, Creatinin, Iron, Ferritin, CRP, Total bilirubin, Total calcium, Sodium, Potassium, Bicarbonates]: French National Agency for the Drugs Safety (ANSM). 2005 [cited March 2017]. Available from: http://ansm.sante.fr/var/ansm_site/storage/original/application/595c4e8256d5d4abf758c281b62c506f.pdf.
32. Bland JM, Altman DG. Transformations, means, and confidence intervals. *BMJ* 1996; 312:1079. PMID: [8616417](#)