BMC Microbiology



Research article Open Access

Optimized molecular resolution of cross-contamination alerts in clinical mycobacteriology laboratories

Ana Martín¹, Marta Herranz¹, Miguel Martínez Lirola², Rosa Fernández Fernández³, INDAL-TB group, Emilio Bouza¹ and Darío García de Viedma*¹

Address: ¹Servicio de Microbiología Clínica y Enfermedades Infecciosas. Hospital Gregorio Marañón, Universidad Complutense, Madrid, CIBER-Enfermedades Respiratorias CIBERES, Spain, ²Servicio de Microbiología, Complejo Hospitalario Torrecárdenas, Almería, Spain and ³Unidad de Tuberculosis del Distrito Poniente, Almería, Spain

Email: Ana Martín - amartsan@yahoo.es; Marta Herranz - m_herranz01@hotmail.com; Miguel Martínez Lirola - manezlirola@ono.com; Rosa Fernández - manezlirola@ono.com; INDAL-TB group - manezlirola@ono.es; Emilio Bouza - ebouza@microb.net; Darío García de Viedma* - dgviedma@microb.net

Published: 14 February 2008

BMC Microbiology 2008, 8:30 doi:10.1186/1471-2180-8-30

This article is available from: http://www.biomedcentral.com/1471-2180/8/30

© 2008 Martín et al; licensee BioMed Central Ltd.

This article is available from http://www.siomedecina.ai.com// 177 2700/0/50

This is an Open Access article distributed under the terms of the Creative Commons Attribution License (http://creativecommons.org/licenses/by/2.0), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Received: 17 July 2007 Accepted: 14 February 2008

Abstract

Background: The phenomenon of misdiagnosing tuberculosis (TB) by laboratory cross-contamination when culturing *Mycobacterium tuberculosis* (MTB) has been widely reported and it has an obvious clinical, therapeutic and social impact. The final confirmation of a cross-contamination event requires the molecular identification of the same MTB strain cultured from both the potential source of the contamination and from the false-positive candidate. The molecular tool usually applied in this context is IS6110-RFLP which takes a long time to provide an answer, usually longer than is acceptable for microbiologists and clinicians to make decisions. Our purpose in this study is to evaluate a novel PCR-based method, MIRU-VNTR as an alternative to assure a rapid and optimized analysis of cross-contamination alerts.

Results: MIRU-VNTR was prospectively compared with IS6110-RFLP for clarifying 19 alerts of false positivity from other laboratories. MIRU-VNTR highly correlated with IS6110-RFLP, reduced the response time by 27 days and clarified six alerts unresolved by RFLP. Additionally, MIRU-VNTR revealed complex situations such as contamination events involving polyclonal isolates and a false-positive case due to the simultaneous cross-contamination from two independent sources.

Conclusion: Unlike standard RFLP-based genotyping, MIRU-VNTR i) could help reduce the impact of a false positive diagnosis of TB, ii) increased the number of events that could be solved and iii) revealed the complexity of some cross-contamination events that could not be dissected by IS6110-RFLP.

Background

The false diagnosis of tuberculosis (TB) due to laboratory cross-contamination is a well-known phenomenon and

has been reported to occur in 0.1–3% of cases [1-5]. It has an obvious epidemiological, clinical and therapeutical impact -each misdiagnosis of tuberculosis due to labora-

^{*} Corresponding author

tory cross-contamination has been estimated to cost on average 10,872 dollars [6]. We recently presented data indicating that laboratory cross-contamination events are more frequent than expected [7] and that these alerts require a faster clarification. False positivity due to laboratory cross-contamination is suspected when i) Mycobacterium tuberculosis (MTB) is cultured from only one of the serial specimens of a patient, ii) the bacterial yield in the culture is low and iii) the suspected sample has been processed together (or in a short period of time apart) with at least one other from a patient with a high bacterial load. The final confirmation of false positivity requires the application of molecular tools to prove that the MTB isolates from the co-processed specimens share identical genotypic patterns (after having ruled out epidemiological links between the cases involved). Unfortunately, the reference MTB genotyping method, IS6110-RFLP, requires well-grown cultures and takes a long time to provide an answer, usually longer than is acceptable for microbiologists and clinicians to make decisions.

A rapid PCR-based MTB genotyping tool, MIRU-VNTR (Mycobacterial-interspersed-repetitive-units-Variable-number-tandem-repeats) [8], has recently been developed and has proved useful in different epidemiological studies [9,10]. It could allow quicker resolution of cross-contamination alerts although studies evaluating the efficiency of MIRU-VNTR in this context in a prospective design are lacking, and only a few isolated examples of its potential in identifying false-positive cases have been reported [11]. Our purpose in this study is to evaluate MIRU-VNTR as an alternative to assure a rapid analysis of cross-contamination alerts in reference laboratories.

Results and Discussion

We prospectively evaluated whether MIRU-VNTR could be an alternative to RFLP for the fast resolution of laboratory cross-contamination alerts in reference genotyping centres. Therefore, we applied both IS6110-RFLP and MIRU-VNTR in a pilot study to analyze all the alerts received from laboratories in Almería, Spain. We compared the response time and the correlation between the diagnosis of either true positivity or laboratory cross-contamination using both techniques. The response time was measured from the moment the culture was received until MIRUtype for all the 12 loci assayed, or an RFLPtype were obtained. We decided to accept only those results obtained within a reasonable time frame (below 90 days), because longer times were not considered useful for resolving the alerts.

Between November 2005 and February 2007, 19 crosscontamination alerts were received in our laboratory. If the potential sources (candidates of true-positive cases) and receptors (candidates of false-positive cases) in con-

tamination events are taken together, a total of 48 cultures were involved (Figures 1 &2). In all cases MIRU-VNTR resolved the alert before RFLP. The response time range for MIRU-VNTR from the reception of the sample was 3-28 days (median: 13 days) whereas RFLP required 24-77 days (median 45 days) (Figure 1). This means that MIRU-VNTR reduced the standard response time by 16–60 days (median: 27 days), which could easily have been increased if priority had been given to performing MIRU-VNTR analysis. Instead, we preferred to perform this evaluation in the real situation of a genotyping laboratory with many other routines. The response time for MIRU-VNTR was longer than could be expected because the primary culture had to be sent to another reference laboratory to perform susceptibility assays, and this meant the reception in many cases of fresh subcultures that required preincubation before analysis. The application of capillary electrophoresis to analyze MIRU-VNTR products [8] had also reduced the response time. With regard to the correlation in the answers, RFLP detected false-positivity in eight of the thirteen events in which it was able to provide a response, indicating that the cross-contamination alerts were justified. MIRU-VNTR, even having applied the 12-loci set which is less discriminatory than the 15 or 24loci sets [12,13], correlated in all cases except in one alert (alert 4) considered as true positivity by RFLP whereas MIRU-VNTR found identical genotypes for the involved cases. To try to clarify this discrepancy, MIRU-VNTR with the whole set of 24 loci was applied and, again, the involved cases shared identical MIRUtypes. A detailed analysis of the RFLPtypes indicated that most of the bands in alert 4 were shared and that differences could result from the acquisition of an additional PvuII site responsible for the digestion of a long band into two smaller bands (Figure 1). One possible explanation for this potential discrepancy would be that the potential source case was infected by two IS6110 RFLP variants and only one was involved in the cross-contamination event. Consistent with this hypothesis, an independent specimen from the source case was genotyped and the same IS6110 profile as the one displayed in the contamination event was observed (data not shown).

As indicated above, RFLP could not offer an answer in six alerts, either because some of the cultures (in three alerts) did not lead to the bacterial yield required for RFLP or because the 90-day limit was exceeded. In all the cases that remained unsolved by RFLP, MIRU-VNTR provided a result and, in two of them, it identified a cross-contamination that would have gone undetected if only RFLP analysis had been available. Each one of these two cross-contamination events revealed only by MIRU-VNTR showed interesting features (Figure 2). The first one (alert 15) involved a polyclonal isolate, with two variants in one loci, similar to others previously reported [14]. The sec-

	VNTR-MIRU ANALYSIS			RFLP ANALYSIS			
ALERT	MIRU-type	TIME	RESULT	RFLP-type	TIME	RESULT	TIME DIFFERENCES
1	PS: 225325153323 CCA:225325153323	8	FP		51	FP	43
2	PS: 223415153323 CCA:223326153325	21	TP		48	TP	27
3	PS: 225325153323 CCA:225325153323	28	FP		49	FP	21
4	PS: 225323153423 CCA: 225323153423	13	FP		29	TP	16
5	PS: 224213153323 CCA: 224213153323	13	FP		29	FP	16
6	PS: 225325153323 CCA:225325153323	8	FP		51	FP	43
7	PS: 224226153313 CCA:223415143323	13	TP		29	TP	16
8	PS: 225313153323 CCA:225313153323	17	FP		77	FP	60
9	PS1: 224325154323 CCA:224325154323 PS2: 224426163321	5	FP		45	FP	40
10	PS: 323325153325 CCA: 323325153325	3	FP	1 1 11 1 1 1 1	24	FP	21
11	PS: 126325153324 CCA:126325153324	11	FP		62	FP	51
12	PS: 224325163221 CCA:225325154324	13	TP		29	TP	16
13	PS1: 224325154323 CCA:224325154323 PS2: 224426163321	5	FP		45	FP	40

Figure I

The numbers for each alert were assigned chronologically. For each alert the potential source/s (PS) and the cross-contamination alerts (CCA) are indicated. The MIRUtypes and RFLPtypes and the time (days) for the obtention of results, the result (TP:True positivity, FP: False positivity) after analyzing the MIRU-VNTR and RFLP fingerprints, and the differences between the time for solving the alerts by MIRU-VNTR and RFLP are shown. The asterisks in alert 4 indicate the differential bands between the RFLP patterns. The MIRU-VNTR alleles which are different for the isolates from an alert are highlighted in bold.

ond (alert 19) was a complex situation in which two crosscontamination events occurred simultaneously. Each of two different sources (PS1 and PS2) contaminated a specimen from two independent patients (CCA1 and CCA2) and both (PS1 and PS2) were also involved in the simultaneous contamination of an additional specimen from another case (CCA3). This double contaminated case (CCA3) could be detected by MIRU-VNTR because its pattern was the combination of the MIRUtypes of the sources (PS1 and PS2), with two different alleles in three of the loci (Figure 2). In this alert, the 90-day limit was exceeded by RFLP and the profiles are not shown; however, the complex patterns of the isolates involved (14 bands and 8 bands) led to a 22-band pattern in the double-contaminated false-positive case that prevented identification of false-positivity by RFLP but not by MIRU-VNTR.

Conclusion

Our data mean that MIRU-VNTR is more adequate than RFLP for analyzing cross-contamination alerts. It was

faster than RFLP, the correlation with RFLP diagnosis was high and it succeded in resolving alerts even under circumstances that were not appropriate for the RFLP analysis requirements. A permanently suspicious attitude on the part of the clinical mycobacteriologist together with access to a fast resolution of cross-contamination alerts could enable more rapid management of suspected false-positive cases, because culture-results would only need to be retained for a short time before clarification. Unlike standard RFLP-based genotyping, MIRU-VNTR could help reduce the impact of a false positive diagnosis of TB.

Methods Samples

Clinical specimens were processed according to standard methods and grown in Lowenstein-Jensen slants and in MGIT (Becton Dickinson, Sparks, Md) liquid media. Analysis of the cross-contamination alerts by MIRU-VNTR and RFLP

	VNTR-MIRU ANALYSIS				
ALERT	MIRU-type	RESULT			
14	CCA:333315153222 PS: 22_2_153324	TP			
15	CCA : $224226163\frac{3}{2}21$ PS : $224226163\frac{3}{2}21$	FP			
16	CCA: 225313153321 PS1: 2225163324 PS2: 224326133324	TP			
17	CCA:223_25153322 PS1: 2434_42532_2 PS2: 22635_1533_3 PS3: 224226163 ³ / ₂ 21	TP			
18	CCA:22 532613 332 4 PS: 22 4213 1 5 332 3	TP			
19	PS1: 225325153323 CCA1:225325153323 CCA3: $22\frac{3}{5}\frac{2}{6}2\frac{5}{6}4\frac{4}{5}332\frac{1}{3}$ CCA2:223226143321 PS2: 223226143321	FP			

Figure 2

Analysis by MIRU-VNTR of the cross-contamination alerts that could not be solved by RFLP. The numbers for each alert were assigned chronologically. For each alert the potential source/s (PS) and the cross-contamination alerts (CCA) are indicated. The MIRUtypes and the results (TP:True positivity, FP: False positivity) after analyzing the MIRU-VNTR finger-prints are shown. The MIRU-VNTR alleles which are different for the isolates from an alert are highlighted in bold. The loci which appear underscored mean that no amplification product was obtained. The loci with two values indicate that two alleles were simultaneously detected for that locus.

Molecular analysis

For IS6110-RFLP we followed the standard procedures [15], and for MIRU-VNTR we applied the 12-loci set [8], trying to apply the simplest and fastest MIRU format and also attempting to obtain a result by directly amplifying a crude extract of the culture (after boiling and sonicating for ten and five minutes, respectively). MIRU-VNTR products were separated by electrophoresis at 45 V for 17 h 30 min, using MS8 2% agarose gels (Pronadisa, Madrid, Spain). Fragment sizes were calculated with the Chemi-Doc system (BioRad, CA, USA) and the Diversity database (BioRad), using a 100-bp ladder (Invitrogen, CA, USA) as a molecular weight marker. The number of repeats in each

locus was calculated by applying the corresponding conversion tables (P. Supply, personal communication)

Molecular patterns were analyzed using Bionumerics 4.6 (Applied Maths, Sint-Martens Laten, Belgium). Results were interpreted as false positivity if both the potential source and the cross-contamination alert had identical MIRU-VNTR and IS6110 RFLP and the converse for true positivity.

Competing interests

The author(s) declare that they have no competing interests.

Authors' contributions

AM performed all the VNTR-MIRU assays, analyzed the results and was involved in the first version of the MS. MH performed the IS6110-RFLP assays, analyzed the results and was involved in the first version of the MS. MML performed all the microbiological procedures and identified the cross-contamination alerts. He has been involved in the compilation of all microbiological and clinical data essential for the analysis of the alerts RFF was involved in the analysis of all the clinical and epidemiological data required for the analysis of the alerts. EB revised critically the final version of the MS. DGV* designed the study, supervised all experimental work, analyzed the results, corrected and produced the final version of the MS. INDAL-TB group was involved in the compilation and analysis of microbiological, clinical and epidemiological data relevant for the analysis of the cross-contamination alerts and critically reviewed the final version of the MS. All the authors read and approved the final MS.

Acknowledgements

We are indebted to Thomas O'Boyle for proofreading and editing the manuscript.

This study was partially financed by Fondo de Investigaciones Sanitarias (FIS030654; FIS030986; FIS060882; FIS061467; 06/90490; 06/90357), Junta de Andalucía (0453/06, 151/05) and by the Instituto de Salud Carlos III (CIBER Enfermedades Respiratorias CB06/06/0058). A.M is receptor of a grant from Comunidad de Madrid cofinanced by the European Social Fund (Order no 5297/2006). INDAL-TB group: MI Sánchez, MC Rogado, T Cabezas, W SánchezYebra, J Martínez, MA Lucerna, P Barroso, I Cabeza-Barrera, LF Díez, M Rodríguez, M Escámez, P Marín, A Lazo, J Gamir, J Vázquez, C Gutiérrez, A Reyes and T Peñafiel.

References

- Ruddy M, McHugh TD, Dale JW, Banerjee D, Maguire H, Wilson P, Drobniewski F, Butcher P, Gillespie SH: Estimation of the rate of unrecognized cross-contamination with mycobacterium tuberculosis in London microbiology laboratories. J Clin Microbiol 2002, 40(11):4100-4104.
- Small PM, McClenny NB, Singh SP, Schoolnik GK, Tompkins LS, Mickelsen PA: Molecular strain typing of Mycobacterium tuberculosis to confirm cross-contamination in the mycobacteriology laboratory and modification of procedures to minimize occurrence of false-positive cultures. J Clin Microbiol 1993, 31(7):1677-1682.

- Multiple misdiagnoses of tuberculosis resulting from laboratory error--Wisconsin, 1996. MMWR Morb Mortal Wkly Rep 1997, 46(34):797-801.
- de Boer AS, Blommerde B, de Haas PE, Sebek MM, Lambregts-van Weezenbeek KS, Dessens M, van Soolingen D: False-positive mycobacterium tuberculosis cultures in 44 laboratories in The Netherlands (1993 to 2000): incidence, risk factors, and consequences. J Clin Microbiol 2002, 40(11):4004-4009.
- de C Ramos M, Soini H, Roscanni GC, Jaques M, Villares MC, Musser JM: Extensive cross-contamination of specimens with Mycobacterium tuberculosis in a reference laboratory. J Clin Microbiol 1999, 37(4):916-919.
- Northrup JM, Miller AC, Nardell E, Sharnprapai S, Etkind S, Driscoll J, McGarry M, Taber HW, Elvin P, Qualls NL, Braden CR: Estimated costs of false laboratory diagnoses of tuberculosis in three patients. Emerg Infect Dis 2002, 8(11):1264-1270.
- Martinez M, Garcia de Viedma D, Alonso M, Andres S, Bouza E, Cabezas T, Cabeza I, Reyes A, Sanchez-Yebra W, Rodriguez M, Sanchez MI, Rogado MC, Fernandez R, Penafiel T, Martinez J, Barroso P, Lucerna MA, Diez LF, Gutierrez C: Impact of laboratory cross-contamination on molecular epidemiology studies of tuberculosis. J Clin Microbiol 2006, 44(8):296-2969.
- Supply P, Lesjean S, Savine E, Kremer K, van Soolingen D, Locht C: Automated high-throughput genotyping for study of global epidemiology of Mycobacterium tuberculosis based on mycobacterial interspersed repetitive units. J Clin Microbiol 2001, 39(10):3563-3571.
- Blackwood KS, Wolfe JN, Kabani AM: Application of mycobacterial interspersed repetitive unit typing to Manitoba tuberculosis cases: can restriction fragment length polymorphism be forgotten? J Clin Microbiol 2004, 42(11):5001-5006.
- Hawkey PM, Smith EG, Evans JT, Monk P, Bryan G, Mohamed HH, Bardhan M, Pugh RN: Mycobacterial interspersed repetitive unit typing of Mycobacterium tuberculosis compared to IS6110-based restriction fragment length polymorphism analysis for investigation of apparently clustered cases of tuberculosis. J Clin Microbiol 2003, 41(8):3514-3520.
- Allix C, Supply P, Fauville-Dufaux M: Utility of fast mycobacterial interspersed repetitive unit-variable number tandem repeat genotyping in clinical mycobacteriological analysis. Clin Infect Dis 2004, 39(6):783-789.
- Supply P, Allix C, Lesjean S, Cardoso-Oelemann M, Rusch-Gerdes S, Willery E, Savine E, de Haas P, van Deutekom H, Roring S, Bifani P, Kurepina N, Kreiswirth B, Sola C, Rastogi N, Vatin V, Gutierrez MC, Fauville M, Niemann S, Skuce R, Kremer K, Locht C, van Soolingen D: Proposal for standardization of optimized mycobacterial interspersed repetitive unit-variable-number tandem repeat typing of Mycobacterium tuberculosis. J Clin Microbiol 2006, 44(12):4498-4510.
- Oelemann MC, Diel R, Vatin V, Haas W, Rusch-Gerdes S, Locht C, Niemann S, Supply P: Assessment of an optimized mycobacterial interspersed repetitive- unit-variable-number tandemrepeat typing system combined with spoligotyping for population-based molecular epidemiology studies of tuberculosis. J Clin Microbiol 2007, 45(3):691-697.
- 14. Shamputa IC, Jugheli L, Sadradze N, Willery E, Portaels F, Supply P, Rigouts L: Mixed infection and clonal representativeness of a single sputum sample in tuberculosis patients from a penitentiary hospital in Georgia. Respir Res 2006, 7:99.
- van Embden JD, Cave MD, Crawford JT, Dale JW, Eisenach KD, Gicquel B, Hermans P, Martin C, McAdam R, Shinnick TM, et al.: Strain identification of Mycobacterium tuberculosis by DNA finger-printing: recommendations for a standardized methodology. J Clin Microbiol 1993, 31(2):406-409.

Publish with **Bio Med Central** and every scientist can read your work free of charge

"BioMed Central will be the most significant development for disseminating the results of biomedical research in our lifetime."

Sir Paul Nurse, Cancer Research UK

Your research papers will be:

- available free of charge to the entire biomedical community
- peer reviewed and published immediately upon acceptance
- cited in PubMed and archived on PubMed Central
- ullet yours you keep the copyright

Submit your manuscript here: http://www.biomedcentral.com/info/publishing_adv.asp

