

# Comparison of Leukosan SkinLink with surgical suture for traumatic laceration repair

## A randomized controlled trial

Hyeongtae Kim, MD<sup>a</sup>, Wonhee Kim, MD<sup>a,b</sup>, Gu Hyun Kang, PhD<sup>a,\*</sup>, Yong Soo Jang, PhD<sup>a</sup>, Hyun Young Choi, MD<sup>a</sup>, Jae Guk Kim, MD<sup>a</sup>, In Young Kim, PhD<sup>b</sup>, Minji Kim, BN<sup>a</sup>

### Abstract

**Background:** Leukosan SkinLink (LS), which combines non-woven textile strips and tissue adhesive, offers the advantage of atraumatic treatment while effectively shortening the procedure time. We hypothesized that wound closure would be faster with LS than with surgical suture (SS) and the wound infection and dehiscence would be similar.

**Methods:** A prospective, open label, single-center, randomized controlled trial was performed. Between November 2014 and April 2016, 49 patients with traumatic lacerations who presented to the emergency department were eligible for study inclusion.

**Results:** The mean wound closure time was significantly lower in the LS group than in the SS group ( $1.48 \pm 0.2$  seconds vs  $8.8 \pm 3.6$  minutes,  $P < .001$ ). After adjusting the wound closure time according to the lacerations length, the time remained significantly lower in the LS group than in the SS group ( $1.0 \pm 0.8$  seconds vs  $5.03 \pm 2.5$  minutes,  $P < .001$ ). During follow-up for 14 days, no significant differences in dehiscence and wound infection were observed between the 2 groups.

**Conclusion:** Wound closure was approximately 4minutes faster with LS and there were no differences in wound infection and dehiscence rates. Thus, the LS could be used as a timesaving suture technique for acute traumatic lacerations in emergency department (ED).

**Trial registration:** ClinicalTrials.gov NCT02333877

**Abbreviations:** ED = emergency department, LS = Leukosan SkinLink, SD = standard deviation, SS = surgical suture.

**Keywords:** lacerations, randomized controlled trial, surgical tape, tissue adhesive

## 1. Introduction

Traumatic lacerations are common acute wounds in patients visiting the emergency department (ED). Surgical suture (SS) has been widely used to repair traumatic lacerations as it promotes wound healing and reduces scars.<sup>[1-3]</sup> Although SS is safe and effective, the procedure is time-consuming and operator-dependent. Therefore, several alternative materials for wound closure, such as tissue adhesives (skin glue) and surgical tapes,

have been recently used because the methods are easy to learn and are time-saving.<sup>[4]</sup> The currently used tissue adhesives are 2-octylcyanoacrylate (Dermabond; Ethicon, Inc., Somerville, NJ) and n-butyl-2 cyanoacrylate (Histoacryl; B. Braun, Ann Arbor, MI). However, these adhesives can seep into the wound and cause dehiscence or hematoma in improperly everted wounds. The use of surgical tapes, such as Steri-Strips (3M Deutschland GmbH, Neuss, Germany), is another easy alternative approach for wound closure. Nevertheless, the theoretical disadvantages of surgical tapes are that they frequently fall off, have lower tensile strength than SS, have the highest rate of dehiscence and cannot get wet.<sup>[5-7]</sup>

Leukosan SkinLink (LS; BSN medical GmbH, Hamburg, Germany) is a non-invasive, 2-component skin closure system, combining non-woven textile strips and a tissue adhesive (n-butyl-2-cyanoacrylate). One previous study reported that the tissue adhesive proved superior to Steri-Strips but inferior to stapled closure.<sup>[8]</sup> The tissue adhesive of the LS has therefore benefits to supplement these disadvantages of Steri-Strips, because the tissue adhesive reinforces the structure and strength of the textile strips (Fig. 1).

No previous randomized controlled trial has compared LS with SS for acute traumatic lacerations. The present study aimed to compare LS with SS for acute traumatic lacerations in the setting of a randomized controlled trial. We hypothesized that wound closure would be faster with LS than with SS and the wound infection and dehiscence would be similar.

Editor: Jianxun Ding.

HK and WK have contributed equally to this study.

This research was supported by Hallym University Research Fund 2015 (HURF-2015-09) and Ministry of Public Safety and Security (MPSS-Fire Safety-2015-76).

All authors declare that they do not have any potential conflicts of interest.

<sup>a</sup> Department of Emergency Medicine, College of Medicine, Hallym University, Seoul, Republic of Korea. <sup>b</sup> Department of Biomedical Engineering, College of Medicine, Hanyang University, Seoul, Republic of Korea.

\* Correspondence: Gu Hyun Kang, Department of Emergency Medicine, College of Medicine, Hallym University, Kangnam Sacred Heart Hospital, 1, Singil-ro, Yeongdeungpo-gu, Seoul 07441, Korea (e-mail: drkang9@gmail.com).

Copyright © 2018 the Author(s). Published by Wolters Kluwer Health, Inc. This is an open access article distributed under the Creative Commons Attribution-NoDerivatives License 4.0, which allows for redistribution, commercial and non-commercial, as long as it is passed along unchanged and in whole, with credit to the author.

Medicine (2018) 97:25(e10918)

Received: 17 November 2017 / Accepted: 7 May 2018

<http://dx.doi.org/10.1097/MD.0000000000010918>



**Figure 1.** Leukosan SkinLink. The tissue adhesive was applied on the strips rather than directly on the wound.

## 2. Materials and methods

### 2.1. Study design

We performed a randomized controlled trial to evaluate the outcomes of LS and SS. The study was performed at Hallym University Kangnam Sacred Heart Hospital between Jan 2015 and April 2016. Written informed consent was obtained for all participants. The local ethics committee approved this trial on November 27, 2014 (IRB No.: 2014-11-152). This trial was registered in Clinical Trials before initiation. (Clinicaltrials.gov: NCT02333877).

### 2.2. Participants

Between Jan 2015 and April 2016, 49 patients with at least 1 traumatic laceration (SS group:  $n=18$ , LS group:  $n=31$ ) were eligible for enrollment at Hallym University Kangnam Sacred Heart Hospital in South Korea.

The inclusion criteria were age 18 to 75 years, presence of a partial thickness laceration (defined as “laceration limited to the epidermis and superficial dermis”), generally good health without systemic abnormalities, agreement to return for a 14-day follow-up. The exclusion criteria were previous diagnosis of diabetes mellitus, peripheral vascular disease, or bleeding disorders; history of keloid formation or scar hypertrophy; and allergy to cyanoacrylate compounds or formaldehyde (Fig. 2).

### 2.3. Interventions

Random numbers were generated by a sequence generator (<http://www.random.org>). These numbers were printed on slips of paper and were placed in translucent boxes. Suture sets were numbered with numbers corresponding to the slip numbers, and the sets contained SS or LS material. The study nurses were asked to select a translucent box, and the suture set with the corresponding number was selected. The physician opened the suture set just before beginning the procedure.

In the SS group, after rinsing the wound, lidocaine was applied near the wound. After injection of a local anesthetic, simple interrupted skin suture with nylon (4-0, 5-0, 6-0) was performed. Deep suture was not used.

In the LS group, after rinsing and drying the wound, adhesive textile strips were applied perpendicular to the wound. The

wound edges were everted with fingers assisted by forceps. Adhesive adjunct was not used. The tissue adhesive was applied on the strips rather than directly on the wound.

After skin closure, semi-occlusive dressing using foam was applied to the wound. Prophylactic oral antibiotics were used for all patients for 5 days. The suture and LS were removed at postoperative day 14.

At short-term follow-ups, patients were evaluated for wound outcomes (postoperative days 2, 7, 14). The observers (senior resident of emergency medicine or emergency physician) were blinded to the identity of the physicians who performed the wound closures.

### 2.4. Equipment and materials

In the SS group, 10% povidone-iodine (Q&Q, Pharm Co., Ltd., Seoul, South Korea; Huons Co., Ltd., Seongnam-si, South Korea; Ailee Co., Ltd., Busan, South Korea.) was used for wound sterilization on the skin edges and 2% lidocaine (Lidocaine HCL injection; Huons Co., Ltd., South Korea) was used for local anesthesia. Non-absorbable suture material (3-0 to 6-0 Nylon, Blue, Ailee Co., Ltd., South Korea) was used for wound closure. In the LS group, LS was used for wound closure. In both the groups, Allevyn adhesive (Smith & Nephew, London, UK) was used for semi-occlusive dressing.

### 2.5. Outcomes

The primary outcome was the wound closure time which could compare timesaving efficiency between both the groups.<sup>[9,10]</sup> The wound closure time was defined as the time required only for primary closure, excluding the time of general wound care, including cleansing and local anesthesia. In the LS group, the wound closure time was defined as the time from everting the wound edges to the physician saying, “finished” after applying the tissue adhesive. In the SS group, the wound closure time was defined as the time from applying local anesthetics to the physician saying, “finished” after completing the simple interrupted suture.

The secondary outcomes were wound dehiscence and infection. Wound dehiscence was assessed as <1, 1–2, and >2 mm. Wound infection was assessed according to erythema, edema, pain, and temperature using a 4-point scale for each.<sup>[11]</sup> One or more points defined as wound infection.




Figure 2. Flow diagram.

## 2.6. Statistical analysis

The sample size was calculated based on our informal study on the wound closure time. The wound closure time (mean  $\pm$  SD) in that study was as follows: LS group,  $4.0 \pm 1.7$  minutes and SS group,  $5.9 \pm 2.0$  minutes. To detect 2 minutes difference in wound closure time, at least 17 patients for each group were needed. Hence, we estimated that 24 patients would be adequate for each group, with a 30% dropout rate and a power of 0.8.

We generated descriptive statistics and presented them as frequencies and percentages for categorical data and means with standard deviations for continuous data.

To compare the wound closure time between the LS and SS groups, the Mann-Whitney *U* test (nonparametric data) or independent *t* test (parametric data) was used for continuous variables. The Fisher exact test was used to compare categorical variables, such as dehiscence. The data were compiled using a standard spreadsheet program (Excel; Microsoft, Redmond, WA) and analyzed using the Statistical Package for the Social Sciences (SPSS) 20.0 KO for Windows (SPSS Inc., Chicago, IL). A *P*-value  $< .05$  was considered statistically significant.

## 3. Results

### 3.1. Characteristics

One thousand nine hundred thirty-six patients were assessed for eligibility and 1888 patients were excluded (not meeting inclusion criteria: 1081, declined to participate: 807). Forty-nine patients with traumatic laceration (SS group:  $n=18$ , LS group:  $n=31$ ) were eligible for enrollment (Fig. 2).

The most common mechanism of injury was knife. The mean laceration length was significantly greater in the SS group than in the LS group ( $1.9 \pm 0.7$  cm vs  $1.2 \pm 0.5$  cm,  $P = .001$ ). All other baseline variables were similar between the 2 groups (Table 1).

### 3.2. Initial outcomes

The mean wound closure time was significantly lower in the LS group than in the SS group ( $1.4 \pm 0.2$  minutes vs  $8.8 \pm 3.6$  minutes,  $P < .001$ ). However, wound length was about 0.7 cm longer in the SS group as shorter wound length could contribute to shorter wound closure time, we adjusted wound closure time according to the wound length. After adjusting the wound closure time according to the laceration length, the time remained significantly lower in the LS group than in the SS group ( $1.0 \pm 0.8$  seconds vs  $5.0 \pm 2.5$  minutes,  $P < .001$ ) (Table 2).

### 3.3. Follow-up

Over the 14-day follow-up, no significant differences in dehiscence and wound infection were observed between the 2 groups (Table 2).

## 4. Discussion

To our knowledge, this is the first randomized controlled trial comparing LS with SS for acute traumatic lacerations. In this study, we found that LS was faster than SS and that the adverse events were similar between LS and SS.

**Table 1****Baseline patient characteristics.**

	Surgical suture (n = 18)	Leukosan SkinLink (n = 31)	P-value
Age, y	41.7 ± 15.3	36.2 ± 12.7	.18
Female	9 (50.0)	16 (51.6)	.91
Mechanism of injury			N/A
Knife	8 (44.4)	16 (51.6)	
Glass	6 (33.3)	4 (12.9)	
Can	0 (0)	3 (9.7)	
Scissors	2 (11.1)	2 (6.5)	
Others	2 (11.1)	6 (19.4)	
Location of laceration			N/A
Finger	6 (33.3)	27 (87.1)	
Hand, except fingers	7 (38.9)	1 (3.2)	
Upper arm	0 (0)	0 (0)	
Lower arm	1 (5.6)	0 (0)	
Upper leg	2 (11.1)	2 (6.5)	
Lower leg	2 (11.1)	0 (0)	
Unknown	0 (0)	1 (3.2)	
Laceration length, cm	1.9 ± 0.7	1.2 ± 0.5	.001*
Depth of laceration			N/A
Superficial	0 (0)	2 (6.5)	
Partial thickness	18 (100)	29 (93.5)	
Gross contamination	0 (0)	0 (0)	N/A
Foreign body identified	0 (0)	0 (0)	N/A
Initial tenderness (VAS)	3.1 ± 2.6	3.1 ± 2.1	.84*

Categorical values are presented as number (%).

Continuous variables are presented as mean ± SD. These variables were assessed with the independent *t* test. N/A = not available.

\* Calculated with the Mann–Whitney *U* test.

The ideal method of wound closure should be simple, safe, rapid, and painless; prevent bacterial growth; and result in optimal cosmetic appearance of the scar.<sup>[12]</sup> The LS procedure does not require the use of anesthesia and is painless. Additionally, the procedure is easy to learn. Furthermore, a reinforced tension-free effect is obtained with topical adhesives and there is no effect on epithelialization. These characteristics of LS are consistent with the characteristics of an ideal method of wound closure. Nevertheless, there have been no clinical studies on LS, except for 1 prospective observational study in the ED and 1 randomized study in the operating room.<sup>[13,14]</sup>

Published studies for tissue adhesives and surgical tapes concentrate solely on the repair of surgical incision wounds in the operating room and not on acute traumatic wounds in the ED.<sup>[15,16]</sup> Bleeding and infection are more likely to occur in traumatic wounds than surgical incision wounds; therefore, the efficacy and safety of LS identified in previous studies on surgical

incision wounds cannot be extended to traumatic wounds.<sup>[17–19]</sup> In the limited small study, the rate of short-term adverse events was similar to sutures.

A previous study reported that the cosmetic outcome of wounds closed with the SS was dependent on the level of training.<sup>[12]</sup> In contrast, the experience of practitioners with tissue adhesives was found to not affect the cosmetic outcome of wounds closed with octylcyanoacrylate.<sup>[12]</sup> In a previous study, the wound closure time was lower with surgical tapes, such as Steri-Strip S, than with the SS, and the complication rates were equivalent.<sup>[20]</sup> Hence, we believe that less experienced practitioners could use LS easily and safely as the procedure is similar to that for surgical tapes and octylcyanoacrylate.

#### 4.1. The present study had some limitations

First, the patient follow-up was limited to 14 days, although cosmetic outcomes, such as dehiscence, were evaluated after 3 months or 1 year in several studies. Therefore, the outcomes of our study do not reflect long-term cosmetic outcomes.

Second, the patients were from South Korea; therefore, it might not be possible to generalize the findings to patients from other countries/regions. Further studies with a long-term follow-up and wide patient representation are needed to overcome these limitations and obtain robust results.

Third, lines of minimal tension were not considered in this study. We also think that lines of minimal tension could affect cosmetic outcome. However, the cosmetic outcomes of LS were similar with those of SS. Hence, minimal tension lines have minimal effect on cosmetic outcomes in this study.

Fourth, the study was under-powered to detect meaningful differences in safety outcomes and had sample imbalance between both the groups. We assumed that this imbalance between both the groups was caused by considering high dropout rate in calculating sample size. Nevertheless, we believed that this imbalance did not significantly affect the results. It was because most baseline variables were similar between both the groups as shown in Table 1.

In conclusion, wound closure was approximately 4 minutes faster with LS and there were no differences in wound infection and dehiscence rates. Thus, the LS could be used as a timesaving suture technique for acute traumatic lacerations in ED.

#### Author contributions

**Conceptualization:** Wonhee Kim, Gu Hyun Kang, Yong Soo Jang, Hyun Young Choi, Jae Guk Kim, In Young Kim.

**Table 2****Outcomes.**

	Surgical suture (n = 18)	Leukosan SkinLink (n = 27)	P-value
Initial outcome			
Wound closure time (min)	8.8 ± 3.6	1.4 ± 0.2	<.001
Wound closure time after adjusting for laceration length (s/cm) <sup>‡</sup>	5.0 ± 2.5	1.0 ± 0.8	<.001*
Follow-up (14 days)			
Dehiscence	2 (11.1)	0 (0)	.15 <sup>†</sup>
Wound infection	0 (0)	0 (0)	N/A

Categorical values are presented as number (%).

Continuous variables are presented as mean ± SD. The variables were assessed with the independent *t* test. N/A = not available.

\* Calculated with the Mann–Whitney *U* test.

<sup>†</sup> Calculated with the Fisher exact test.

<sup>‡</sup> Calculated using the following formula: wound closure time (s)/laceration length (cm).



**Data curation:** Wonhee Kim, Hyun Young Choi, Jae Guk Kim, In Young Kim.

**Formal analysis:** Yong Soo Jang, Hyun Young Choi, Jae Guk Kim.

**Funding acquisition:** Wonhee Kim, Gu Hyun Kang, Hyun Young Choi, Minji Kim.

**Investigation:** Hyeongtae Kim, Wonhee Kim, Gu Hyun Kang, Yong Soo Jang, Jae Guk Kim, Minji Kim.

**Methodology:** Hyeongtae Kim, Wonhee Kim, Yong Soo Jang, Jae Guk Kim.

**Project administration:** In Young Kim, Minji Kim.

**Software:** Wonhee Kim, Hyun Young Choi, In Young Kim, Minji Kim.

**Supervision:** Gu Hyun Kang, In Young Kim.

**Validation:** Wonhee Kim, Gu Hyun Kang, Yong Soo Jang, Hyun Young Choi, Jae Guk Kim, In Young Kim, Minji Kim.

**Visualization:** Wonhee Kim.

**Writing – original draft:** Hyeongtae Kim, Wonhee Kim.

**Writing – review and editing:** Hyeongtae Kim, Wonhee Kim, Yong Soo Jang, In Young Kim, Minji Kim.

## References

- [1] Hollander JE, Singer AJ. Laceration management. *Ann Emerg Med* 1999;34:356–67.
- [2] Colovic R. Wound treatment in ancient Greece. *Acta Chir Iugosl* 2001;48:6–8.
- [3] Takayama S, Yamamoto T, Tsuchiya C, et al. Comparing Steri-Strip and surgical staple wound closures after primary total knee arthroplasties. *Eur J Orthop Surg Traumatol* 2017;27:113–8.
- [4] Singer AJ, Quinn JV, Clark RE, et al. “TraumaSeal Study Group”- Closure of lacerations and incisions with octylcyanoacrylate: a multicenter randomized controlled trial. *Surgery* 2002;131:270–6.
- [5] Woodward SC, Herrmann JB, Cameron JL, et al. Histotoxicity of cyanoacrylate tissue adhesive in the rat. *Ann Surg* 1965;162: 113–22.
- [6] Zempsky WT, Parrotti D, Grem C, et al. Randomized controlled comparison of cosmetic outcomes of simple facial lacerations closed with Steri Strip Skin Closures or Dermabond tissue adhesive. *Pediatr Emerg Care* 2004;20:519–24.
- [7] Singer AJ, Quinn JV, Hollander JE. The cyanoacrylate topical skin adhesives. *Am J Emerg Med* 2008;26:490–6.
- [8] Shapiro AJ, Dinsmore RC, North JH Jr. Tensile strength of wound closure with cyanoacrylate glue. *Am Surg* 2001;67:1113–5.
- [9] van de Gevel DF, Hamad MA, Elenbaas TW, et al. Is the use of Steri-Strip™ S for wound closure after coronary artery bypass grafting better than intracuticular suture? *Interact Cardiovasc Thorac Surg* 2010;10:561–4.
- [10] Duscher D, Apfelterer S, Garner S, et al. Time is money: a faster way of Steri-Strip application (the 3-finger technique). *Aesthet Surg J* 2013;33:924–5.
- [11] Hollander JE, Singer AJ, Valentine S, et al. Wound registry: development and validation. *Ann Emerg Med* 1995;25:675–85.
- [12] Singer AJ, Hollander JE, Quinn JV. Evaluation and management of traumatic lacerations. *N Engl J Med* 1997;337:1142–8.
- [13] Juergens S, Maune C, Kezze F, et al. A randomized, controlled study comparing the cosmetic outcome of a new wound closure device with Prolene suture closing caesarean wounds. *Int Wound J* 2011;8:329–35.
- [14] Kim H, Kim J, Choi J, et al. The usefulness of Leukosan SkinLink for simple facial laceration repair in the emergency department. *Arch Plast Surg* 2015;42:431–7.
- [15] Coulthard P, Esposito M, Worthington HV, et al. Tissue adhesives for closure of surgical incisions. *Cochrane Database Syst Rev* 2010; CD004287.
- [16] Singer AJ, Thode HC Jr. A review of the literature on octylcyanoacrylate tissue adhesive. *Am J Surg* 2004;187:238–48.
- [17] Howell JM, Chisholm CD. Outpatient wound preparation and care: a national survey. *Ann Emerg Med* 1992;21:976–81.
- [18] Berk WA, Osbourne DD, Taylor DD. Evaluation of the ‘golden period’ for wound repair: 204 cases from a Third World emergency department. *Ann Emerg Med* 1988;17:496–500.
- [19] Edlich RF, Rodeheaver GT, Thacker JG, et al. Revolutionary advances in the management of traumatic wounds in the emergency department during the last 40 years: part I. *J Emerg Med* 2010;38:40–50.
- [20] Kerrigan CL, Homa K. Evaluation of a new wound closure device for linear surgical incisions: 3M Steri-Strip S surgical skin closure versus subcuticular closure. *Plast Reconstr Surg* 2010;125:186–94.