

ORAL PRESENTATION**Open Access**

Study of the corrective forces applied by a dynamic derotation brace (DDB)

I Loukos^{1*}, C Nicolopoulos², C Zachariou³

From 8th International Conference on Conservative Management of Spinal Deformities and SOSORT 2011 Annual Meeting
Barcelona, Spain. 19-21 May 2011

Background and purpose

The direct forces exerted by the brace pad were analysed, on 44 idiopathic scoliotic patients (38 girls, 6 boys) treated with a DDB [1]. Twenty seven patients had a single right thoracic curve and 17 had a single left or right thoracolumbar curve. The aim was the analysis of the corrective forces, exerted at the skin-brace interface, by altering the posture, activity and strap tension [2-7].

Materials and methods

We used the F-Socket 9801 pressure sensor and the MatScan Research BETA STAM v. 6.30 software (TekScan, Boston MA, USA), and measurements were taken in nine body postures. The patients were divided into three groups: those who were wearing for first time the brace, those who changed their brace with a new one and those who made corrections on their brace.

Results

Changes in strap tension, body posture, resulted in statistically significant alterations of the interface pressure on the pads and thus of the resultant forces exerted on the patient's body by the pads. Comparing the three groups in relation with the magnitude of the mean exerted force, the correction of the brace caused the highest mean exerted force.

Conclusions

Even though the TekScan system does not provide direct information on the correction of spinal curvature, it appears to be a useful tool in the treatment of scoliotic patients. Also, the analysis of the corrective forces seems

to be very helpful trying to achieve brace's optimal fit and the same time the best therapeutic result for the patient.

Author details

¹Medical School, National & Capodistrian University of Athens Anoixi, Greece.
²ORTHO-FOOT Center, Athens, Greece. ³KAT Hospital of Athens, Greece.

Published: 27 January 2012

References

- Grivas TB, Bountis A, Vrasami I, Bardakos NV: **Brace technology thematic series: the dynamic derotation brace.** *Scoliosis* 2010, 5:20.
- Aubin CE, Labelle H, Ruszkowski A, et al: Variability of strap tension in brace treatment for adolescent idiopathic scoliosis. *Spine* 1999, 24(4):349-54.
- Mac-Thiong JM, Petit Y, Aubin CE, et al: Biomechanical evaluation of the Boston brace system for the treatment of adolescent idiopathic scoliosis. *Spine* 2003, 29(1):26-32.
- Nicolopoulos CS, Anderson EG, Solomonidis SE, et al: Evaluation of the gait analysis F-Scan pressure system: clinical tool or toy? *The Foot* 2000, 10:124-130.
- Pham VM, Houilliez A, Schill A, et al: Study of the pressures applied by a Chêneau brace for correction of adolescent idiopathic scoliosis. *Prosthet Orthot Int* 2008, 32(3):345-355.
- van den Hout JAAM, van Rhijn LW, van den Munckhof RJH, et al: Interface corrective force measurements in Boston brace treatment. *Eur Spine J* 2002, 11:332-335.
- Wong MS, Evans JH: Biomechanical evaluation of the Milwaukee brace. *Prosthet Orthot Int* 1998, 22:54-67.

doi:10.1186/1748-7161-7-S1-O31

Cite this article as: Loukos et al: Study of the corrective forces applied by a dynamic derotation brace (DDB). *Scoliosis* 2012 7(Suppl 1):O31.

¹Medical School, National & Capodistrian University of Athens Anoixi, Greece
Full list of author information is available at the end of the article