

CORRECTION

Correction: Particle removal from air by face masks made from Sterilization Wraps: Effectiveness and Reusability

Sachin Walawalkar, Manish Joshi, Navin Khattry, Balvinder Kaur Sapra, Arshad Khan, Pradeep Kumar Pujari, Lalit Mohan, Sushil Prasad Srivastava, Chital Naresh, Rajendra Badwe, Sudeep Gupta

There is an error in [Table 1](#). In the second row, the values in columns four, five, and six are incorrect. Please see the correct [Table 1](#) here.


Table 1. Aerosol capture efficiency for test specimens.

S. No.	Mask type (n = 5)	Aerosol capture Efficiency (%) for particle size			
		0.3 μm	0.5 μm	1.0 μm	Total ($>0.3 \mu\text{m}$)
1.	45 GSM SW	82.65 \pm 3.08	88.61 \pm 4.29	94.01 \pm 8.57	84.88 \pm 3.50
2.	60 GSM SW	92.10 \pm 3.69	98.19 \pm 5.64	99.13 \pm 7.44	94.48 \pm 4.18
3.	Surgical mask (commercial)	40.08 \pm 3.85	56.35 \pm 4.41	84.12 \pm 6.76	42.44 \pm 3.91
4.	N-95 respirator (commercial)	96.19 \pm 2.55	96.67 \pm 2.18	100.00 \pm 0.00	96.64 \pm 2.21

<https://doi.org/10.1371/journal.pone.0252693.t001>

Reference

- Walawalkar S, Joshi M, Khattry N, Sapra BK, Khan A, Pujari PK, et al. (2020) Particle removal from air by face masks made from Sterilization Wraps: Effectiveness and Reusability. PLoS ONE 15(10): e0240398. <https://doi.org/10.1371/journal.pone.0240398> PMID: 33052962



OPEN ACCESS

Citation: Walawalkar S, Joshi M, Khattry N, Sapra BK, Khan A, Pujari PK, et al. (2021) Correction: Particle removal from air by face masks made from Sterilization Wraps: Effectiveness and Reusability. PLoS ONE 16(5): e0252693. <https://doi.org/10.1371/journal.pone.0252693>

Published: May 27, 2021

Copyright: © 2021 Walawalkar et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.