

Effect of Specific Over Nonspecific VR-Based Rehabilitation on Poststroke Motor Recovery: A Systematic Meta-analysis

Neurorehabilitation and
Neural Repair
2019, Vol. 33(2) 112–129
© The Author(s) 2019



Article reuse guidelines:
sagepub.com/journals-permissions
DOI: 10.1177/1545968318820169
journals.sagepub.com/home/nnr



Martina Maier, MSc¹, Belén Rubio Ballester, PhD¹, Armin Duff, PhD¹,
Esther Duarte Oller, PhD², and Paul F. M. J. Verschure, PhD^{1,3}

Abstract

Background. Despite the rise of virtual reality (VR)-based interventions in stroke rehabilitation over the past decade, no consensus has been reached on its efficacy. This ostensibly puzzling outcome might not be that surprising given that VR is intrinsically neutral to its use—that is, an intervention is effective because of its ability to mobilize recovery mechanisms, not its technology. As VR systems specifically built for rehabilitation might capitalize better on the advantages of technology to implement neuroscientifically grounded protocols, they might be more effective than those designed for recreational gaming. **Objective.** We evaluate the efficacy of specific VR (SVR) and nonspecific VR (NSVR) systems for rehabilitating upper-limb function and activity after stroke. **Methods.** We conducted a systematic search for randomized controlled trials with adult stroke patients to analyze the effect of SVR or NSVR systems versus conventional therapy (CT). **Results.** We identified 30 studies including 1473 patients. SVR showed a significant impact on body function (standardized mean difference [SMD] = 0.23; 95% CI = 0.10 to 0.36; $P = .0007$) versus CT, whereas NSVR did not (SMD = 0.16; 95% CI = -0.14 to 0.47; $P = .30$). This result was replicated in activity measures. **Conclusions.** Our results suggest that SVR systems are more beneficial than CT for upper-limb recovery, whereas NSVR systems are not. Additionally, we identified 6 principles of neurorehabilitation that are shared across SVR systems and are possibly responsible for their positive effect. These findings may disambiguate the contradictory results found in the current literature.

Keywords

stroke, paresis, virtual reality, rehabilitation, occupational therapy, review

Introduction

Better medical treatments in the acute phase after stroke have increased survival and with that the number of patients needing rehabilitation with an associated increased burden on the health care system.¹ Novel technologies have sought to meet this increased rehabilitation demand and to potentially allow patients to continue rehabilitation at home after they leave the hospital.² Also, technology has the potential to gather massive and detailed data (eg, kinematic and performance data) that might be useful in understanding recovery after stroke better, improving the quality of diagnostic tools and developing more successful treatment approaches.³ Given these promises, several studies and meta-analyses have evaluated the effectiveness of technologies that use virtual reality (VR) in stroke rehabilitation. In a first review, Crosbie et al⁴ analyzed 6 studies that used VR to provide upper-limb rehabilitation. Although they found a positive effect, they concluded that the evidence was only weak to moderate given the low quality of the research. A later meta-analysis analyzing 5 randomized controlled trials (RCTs) and 7 observational studies

suggested a positive effect on a patient's upper-limb function after training.⁵ Another meta-analysis of 26 studies by Lohse et al,⁶ which compared specific VR (SVR) systems with commercial VR games, found a significant benefit for SVR

¹Laboratory of Synthetic, Perceptive, Emotive and Cognitive Systems (SPECS), Institute for Bioengineering of Catalonia (IBEC), The Barcelona Institute of Science and Technology, Barcelona, Spain

²Rehabilitation Research Group, Institut Hospital del Mar d'Investigacions Mèdiques (IMIM), Physical Medicine and Rehabilitation Department Parc de Salut Mar (Hospital del Mar, Hospital de l'Esperança), Barcelona, Spain

³Institució Catalana de Recerca i Estudis Avançats (ICREA), Barcelona, Spain

Supplementary material for this article is available on the *Neurorehabilitation & Neural Repair* website at <http://nnr.sagepub.com/content/by/supplemental-data>.

Corresponding Author:

Paul F. M. J. Verschure, PhD, Laboratory of Synthetic, Perceptive, Emotive and Cognitive Systems (SPECS), Institute for Bioengineering of Catalonia (IBEC), The Barcelona Institute of Science and Technology, Av d'Eduard Maristany 10-14, 08019 Barcelona, Spain.
Email: pverschure@ibecbarcelona.eu

systems as compared with conventional therapy (CT) in both body function and activity but not between the 2 types of systems. This study, however, included a variety of systems that would treat upper-limb, lower-limb, and cognitive deficits. Saywell et al⁷ analyzed 30 “play-based” interventions, such as VR systems including commercial gaming consoles, rehabilitation tools, and robot-assisted systems. They found a significant effect of play-based versus control interventions in dose-matched studies in the Fugl-Meyer Assessment of the Upper Extremity (FM-UE).⁷ In contrast, a more recent large-scale analysis of a study with Nintendo Wii-based video games, including 121 patients concluded that recreational activities are as effective as VR.⁸ A later review evaluated 22 randomized and quasi-randomized controlled studies and concluded that there is no evidence that the use of VR and interactive video gaming is more beneficial in improving arm function than CT.⁹ In all, 31% of the included studies tested nonspecific VR (NSVR) systems (Nintendo Wii, Microsoft Xbox Kinect, Sony PlayStation EyeToy). Hence, although VR-based interventions have been in use for almost 2 decades, their benefit for functional recovery, especially for the upper limb, remains unknown. Possibly, these contradictory results indicate that, at present, studies are too few or too small and/or the recruited participants too variable to be conclusive.¹⁰ However, alternative conclusions can be drawn. First, VR is an umbrella term. Studies comparing its impact often include heterogeneous systems or technologies, customized or noncustomized for stroke treatment, addressing a broad range of disabilities. However, effectiveness can only be investigated if similar systems that rehabilitate the same impairment are contrasted. This has been achieved by meta-analyses that investigated VR-based interventions for the lower limb, concluding that VR systems are more effective in improving balance or gait than CT.¹¹ Second, a clear understanding of the “active ingredients”³ that should make VR interventions effective in promoting recovery is missing. Therapeutic advantages of VR identified in current meta-analyses are that it might apply principles relevant to neuroplasticity,^{5,9} such as providing goal-oriented tasks,^{5,9} increasing repetition and dosage,^{5,9} providing therapists and patients with additional feedback,^{5,6,9} and allowing to adjust task difficulty.⁶ In addition, it has been suggested that the use of VR increases patient motivation,⁶ enjoyment,^{8,9} and engagement⁷; makes intensive task-relevant training more interesting^{4,7}; and offers enriched environments.⁹ Although motivational aspects are important in the rehabilitation process because they possibly increase adherence,³ their contribution to recovery is difficult to quantify because it relies on patients’ subjective evaluation.^{7,12-15} Rehabilitation methods, whether VR or not, however, need to be objectively beneficial in increasing the patient’s functional ability. Hence, an enormous effort has been expended to identify principles of neurorehabilitation that enhance motor learning and recovery.¹⁶⁻²⁴ Consequently, an effective VR system should besides

be motivating, also augment CT by applying these principles in the design.²³ Following this argument, we advance the hypothesis that custom-made VR rehabilitation systems might have incorporated these principles, unlike off-the-shelf VR tools, because they were created for recreational purposes. Combining the effects of both approaches in one analysis might, thus, mask their real impact on recovery. Again, in the rehabilitation of the lower limb, this effect has been observed. Two meta-analyses investigating the effect of using commercial VR systems for gait and balance training did not find a superior effect, which contradicts the conclusions of the other systematic reviews.¹¹ In upper-limb rehabilitation, this question has not been properly addressed until the most recent review by Aminov et al.²⁵ However, there are several flaws in the method applied that could invalidate the results they found. Specifically, studies were included regardless of their quality, and it is not clear which outcome measurements were taken for the analysis according to the World Health Organization’s International Classification of Function, Disability, and Health (ICF-WHO).²⁶ In addition, a specifically designed rehabilitation system (Interactive Rehabilitation Exercise [IREX])²⁷ was misclassified as an off-the-shelf VR tool. Because their search concluded in June 2017, the more recent evidence is missing. We decided to address these issues by conducting a well-controlled meta-analysis that focuses only on RCTs that use VR technologies for the recovery of the upper limb after stroke. We analyze the effect of VR systems specifically built for rehabilitation (ie, SVR systems) and off-the-shelf systems (ie, NSVR commercial systems) against CT according to the ICF-WHO categories. Also, we extracted 11 principles of motor learning and recovery from established literature that could act as “active ingredients” in the protocols of effective VR systems. Through a content analysis, we identified which principles are present in the included studies and compared their presence between SVR and NSVR systems. We hypothesized, first, that SVR systems might be more effective than NSVR systems as compared with CT in the recovery of upper-limb movement and, second, that this superior effect might be a result of the specific principles included in SVR systems.

Methods

This meta-analysis was conducted following the Preferred Reporting Items for Systematic Reviews and Meta-Analysis (PRISMA) guidelines.²⁸

Identification of RCTs

We define VR as a computer-based technology that provides the user with a sense of presence in a virtual environment,²⁹ which is induced by exposing the user to computer-generated sources of sensory stimulation that satisfy their perceptual predictions and expected

sensorimotor contingencies.³⁰ The studies included aimed at training the upper extremity of stroke patients through active participation, without assistive robotic devices (eg, exoskeleton, end-effector devices) or exogenous stimulation. We compared the impact on body function and activity of 2 kinds of VR systems with CT: SVR and NSVR systems. SVR systems were developed exclusively for neurorehabilitation purposes. NSVR systems, on the other hand, are recreational and/or off-the-shelf video games (eg, Nintendo Wii, Microsoft Xbox). As CT, we considered occupational therapy and physical therapy. To identify all RCTs in these 2 categories, we performed a computerized search in the bibliographic databases MEDLINE (OVID), Cochrane Library Plus (including EMBASE), CINAHL, APA PsycNET, DARE, and PEDro for studies that were published in English from inception until August 7, 2018, the day of the conclusion of the search. The search strategy (Supplementary Table 1) included only RCTs that tested the efficacy of SVR or NSVR systems in recovering the upper limbs of stroke patients who were either in the acute (up to 21 days poststroke), subacute (between 3 weeks and 3 months poststroke), or chronic (after 3 months poststroke) stage. We combined the effects of various chronicity bands because the current literature suggests that principles of motor learning interact constantly with the biological processes of recovery,³¹ and therefore, no differential effect between SVR and NSVR systems resulting from chronicity should be expected. This notion has also been confirmed by the latest meta-analysis.²⁵ In addition, splitting the identified literature into VR type, ICF-WHO category, and chronicity reduces statistical power because of the small number of studies remaining in each band. Two reviewers (BRB and MM) assessed the studies for eligibility. We excluded studies that were not carried out on humans, lacked a control group, included less than 5 participants per experimental condition, did not target upper-extremity rehabilitation, used exoskeletons as interfaces, used exogenous stimulation (such as transcranial stimulation), or did not provide information on standard clinical scales (Figure 1). Exoskeletons and exogenous stimulation protocol were excluded for the passive or active support provided in the rehabilitation process that might lead to different outcomes.

Outcome Measurements

Two reviewers (BRB and MM) cross-analyzed the content of the included studies and extracted the relevant data into a separate database. In general, published articles were used. If information in the articles was missing, the respective authors were contacted by mail. To classify the impact of VR on upper-extremity function and activity at the end of therapy according to the ICF-WHO framework, we followed the recommendations given by the Stroke

Rehabilitation Evidence-Based Review³² and considered the following outcome measurements in the respective order. For body function, we considered the FM-UE,³³ Modified Ashworth Scale, Motricity Index,³⁴ Brunnstrom Motor Recovery Stage,³⁵ and Stroke Impact Scale (SIS, only hand items).³⁶ For activity, we considered the (Modified) Barthel Index,³⁷ the Functional Independence Measure,³⁸ the Action Research Arm Test,³⁹ the Box and Block Test,⁴⁰ and the Wolf Motor Function Test (WMFT).⁴¹ We did not conduct a comparison for the ICF-WHO category participation because of the 4 studies^{8,42-44} that had a corresponding outcome measurement (SIS and Medical Outcomes Study Short Form 36 Health Survey⁴⁵), only one classified as SVR intervention.⁴² For each study, we identified 1 measurement in each category and took the absolute score (mean and SD) at the end of the treatment for intervention and control group. When the SD of the mean was not available,^{14,44} we requested it from the corresponding authors. When only the median and first/third quartile^{42,46,47} or minimum/maximum^{48,49} was reported, we estimated the mean and the SD using the method proposed by Wan et al.⁵⁰

Quality Appraisal and Risk-of-Bias Assessment

We used the established PEDro checklist to assess the quality of the RCTs.⁵¹ In this review, we only included RCTs with a PEDro score of 5 or greater, which we considered to be high-quality studies. We then used The Cochrane Collaboration's Risk of Bias tool to evaluate the methodological quality of the included studies (Supplementary Figures 1 and 2).

Content Analysis of Included Principles of Neurorehabilitation

To see whether SVR and NSVR systems are different according to their therapeutic specifications, 2 reviewers (MM and BR) reviewed the existing literature on principles of motor learning and recovery for neurorehabilitation. We extracted a list of 11 principles that have been shown to be effective for motor recovery because they enhance neural plasticity and, therefore, optimize acquisition, retention, and generalization of motor skills: massed practice (training that is repetitive),^{17,19,22-24} dosage (training that is intensive),^{17,19,20,22} structured practice (training that is spaced in time),^{17,19,52} task-specific practice (skill training that is relevant for activities of daily living [ADL]),^{17,19,23} variable practice (training that is randomized and variable),^{16,23} multisensory stimulation (training that provides not only visual feedback),^{19,23} increasing difficulty (training that is individualized),^{22,23} explicit feedback (training that provides knowledge about results),^{18,23} implicit feedback (training that delivers implicit task-relevant cues),^{18,23} avatar representation (training that is embodied and immersive),^{21,23}




Figure 1. Study flow diagram (PRISMA). The selection process of identified randomized controlled trials. Abbreviations: NSVR, nonspecific VR; SVR, specific VR; VR, virtual reality.

and promoting the use of the paretic limb (training that counteracts compensation and learned nonuse).^{17,19,22} Each principle was then assigned key descriptors. One of us (MM) then performed a qualitative content analysis in the included studies using the key descriptors as an indicator of whether a given principle was present or not (deductive category application). Only if the key descriptors were explicitly explained or mentioned in the text, the principle was defined to be present in the study. In Table 1, we present the 11 principles that were extracted from the literature together with their definitions, their ascribed effect on recovery, and the assigned key descriptors for encoding. We performed a pure content analysis without following up with the authors to examine the reporting pattern of the principles they thought were relevant for their results. Finally, we calcu-

lated the presence as a percentage for each principle, separately for SVR and NSVR studies.

Statistical Analysis

We performed a subgroup analysis using RevMan 5.1. Outcome measures were included in absolute terms as provided by the authors or estimated from raw data. Heterogeneity was assessed using the χ^2 test, and I^2 and was considered significant when the probability value of χ^2 was $<.05$ or when I^2 was $>40\%$.⁵³ The pooled treatment effect (inverse variance) was evaluated using random-effect models to avoid a heterogeneity bias.⁵³ Because a direct comparison between the effects of SVR and NSVR on outcomes for body function and activity is not possible, we conducted

Table 1. Qualitative Content Analysis: Description, Definition, and Effect of Identified Principles and Their Key Descriptors.

Name	Definition	Effect	Key Descriptors
Massed practice	The number of repetitions performed	Small effects on improvement and retention ^{17,19,22,24}	<ul style="list-style-type: none"> - Number of repetitions was counted - Tasks were aimed at increasing number of repetitions of a movement
Dosage	Training of more than 5 hours a week	Can speed up functional recovery ^{17,19,20,22}	<ul style="list-style-type: none"> - Training is more than 60 minutes of therapy per session and week day
Structured practice	Training schedule with frequent and longer breaks	Better retention than massed protocols ^{17,19,52}	<ul style="list-style-type: none"> - Rests were given during during the session
Task-specific practice	Movements performed are relevant for ADL and goal oriented	Learning is maximal if the task trained is specific ^{17,19}	<ul style="list-style-type: none"> - Tasks incorporated movements that are functionally meaningful (reaching, lifting, grasping pronation, supination, pinching, etc) and were goal oriented - Tasks or movements were relevant for ADL
Variable practice	Several tasks that require different movements	Better retention and enhances generalization ¹⁶	<ul style="list-style-type: none"> - Training included various tasks that require a variety of movements
Multisensory stimulation	Providing feedback through multiple senses	Restoration of sensorimotor contingencies ¹⁹	<ul style="list-style-type: none"> - Besides visual, other types of feedback were provided (auditory, tactile, etc)
Increasing difficulty	Progressively increase the difficulty of the task or the involved movements	Augment task-specific use of the impaired limb ²²	<ul style="list-style-type: none"> - Difficulty or complexity of tasks or movement is changing depending on ability, performance, or time
Explicit feedback	Knowledge about results (task success or failure, or movement outcome)	Retain an adapted movement better ¹⁸	<ul style="list-style-type: none"> - Providing cues on task completion with regard to success or failure, or movement outcome (trajectory errors, average completion time, or exactness) - Feedback can also be provided through a therapist
Implicit feedback	Knowledge about performance that is obtained from tracking, analyzing, and visualizing kinematic movement data	Reduce the sensorimotor prediction error and promote learning ¹⁸	<ul style="list-style-type: none"> - Real-time visualization of arm/hand movement and other kinematic properties (speed, rotations, synergies compensations) - Display of correct trajectory to follow
Avatar representation	Active execution and observation of movement through an avatar	Degree of agency aids learning from sensorimotor prediction error ²¹	<ul style="list-style-type: none"> - Virtual movement is represented as a human- or body part-like avatar (whole body, arm, or hand)
Promote use of affected limb	Tasks that are forcing or reinforcing the use of the affected arm	Counteracting learned nonuse ^{17,19,22}	<ul style="list-style-type: none"> - Tasks were designed or required to be performed with the paretic limb - Tasks cannot be accomplished by the healthy arm only

Abbreviation: ADL, activities of daily living.

an indirect comparison in which each VR type was compared with CT at each ICF level through a subgroup analysis. Because SVR and NSVR studies reported the continuous outcomes in different psychometric scales, the standardized mean difference (SMD) and 95% CI to represent the magnitude of the reported improvement were used. It should be noted that the SMD method does not correct for differences in the direction of the scale. Because WMFT is measured in seconds to complete the task (and, therefore, decreases with better performance), its mean value was multiplied by -1 to

ensure that all the scales point in the same direction. For all analyses, the statistical significance level was set at $P < .05$. Risk of publication bias across studies was estimated visually by inspecting the funnel plots. We used GRADEpro to assess the overall quality of the evidence found.

Results

We wanted to assess whether VR-based systems that are purposefully designed for stroke rehabilitation (SVR)

render rehabilitation outcomes different from systems that are NSVR. Our prediction is that SVR systems should outperform NSVR because the former are designed around distinct principles for neurorehabilitation, whereas the latter are not.

Study Identification

We identified 1751 articles that matched the search strategy. Ten additional studies were identified through other sources (eg, meta-analyses). Of the 1164 records screened, 30 articles that were published between January 2002 and August 2018 satisfied the inclusion criteria and were included in this review. The study's characteristics can be found in Table 2; the aim, the selected outcome measurements per ICF-WHO category and the main finding are reported in Table 3. A total of 1137 records and articles were removed, of which 22 were after qualitative full-text analysis (Supplementary Table 2). One of the articles included 3 experimental subgroups,⁵⁴ which were considered as separate trials, resulting in a total of 32 outcomes that were included in the analysis. A total of 22 RCTs qualified as SVR systems* and 8 as NSVR systems.^{8,43,44,68-72} Of the 30 articles included, 13 evaluated motor function at follow-up after a period of no treatment (SVR = 8, NSVR = 5). Interventions were delivered from 2 to 12 weeks (mean SVR = 4.4 weeks, mean NSVR = 4.3 weeks) across all studies. The duration of the rehabilitation sessions varied in SVR studies from 20 to 158.3 mins (mean 23.9 hours total intervention time), and in the NSVR studies from 60 mins to 135 mins (mean 21.9 hours total intervention time). Overall, the most frequently used outcome measure was the FM-UE (SVR = 16, NSVR = 3).

Assessment of Risk of Bias

We assessed the methodological quality of the included studies, by analyzing each dimension in the risk-of-bias analysis. The detailed analysis per study and the summary plot can be found in Supplementary Figures 1 and 2, respectively.

Allocation. Random sequence generation was adequately reported by 18 SVR and all NSVR studies. One SVR study⁵⁴ stated that no random allocation was performed, and therefore, also no allocation concealment was applied. In the other studies, allocation concealment was adequately reported by 9 SVR and 4 NSVR studies.

Blinding. A total of 18 SVR and all NSVR studies adequately reported that the outcome assessor was blinded. Because of the nature of the interventions, only a few studies could

blind participants and therapists. We evaluated studies at a low risk if either of the 2 groups was blind or if they tried to limit the impact of nonblindness (3 SVR and 2 NSVR). Therefore, the nonblinding of personnel and patients could be a high risk of bias.

Incomplete Outcome Data. In all, 19 SVR and 5 NSVR studies adequately reported how missing data points were handled. Two NSVR studies reported inconsistent information about how the missing data were handled.

Selective Reporting. Except for 1 SVR study, all included studies reported the outcomes for all measurements taken.

Effects of SVR and NSVR Interventions

When analyzing the outcome of the subgroup analysis, SVR studies showed a significant impact on the recovery of the upper-limb function (SMD = 0.23; 95% CI = 0.10 to 0.36; $P = .0007$) and activity (SMD = 0.31; 95% CI = 0.15 to 0.47; $P = .0001$) that is superior in comparison to CT (Figures 2 and 3, upper panel). NSVR studies showed no significant effect, neither on body function (SMD = 0.16; 95% CI = -0.14 to 0.47; $P = .30$) nor on activity (SMD = 0.15; 95% CI = -0.15 to 0.45; $P = .33$); see Figures 2 and 3, lower panel. No significant heterogeneity was present in any comparison. Also, there were no significant differences between the subgroups, neither in body function ($P = .70$) nor in activity ($P = .36$) because the CIs overlapped substantially. According to GRADE (Figure 4), there is moderate confidence in the effect estimates for the results found in SVR studies.

Assessment of Reporting Bias

Funnel plot asymmetry might point to a possible publication bias because of a lack of small studies with nonsignificant or unfavorable results (Supplementary Figure 3). Because of our exclusion criteria, only one study had a small sample size⁵⁷ ($n = 10$). Together with other smaller studies, it skews the plot slightly to the right. However, other explanations are possible. Many SVR systems have become commercially available to clinics after the treatment effect was confirmed through experiments. It, therefore, cannot be ruled out that the confounding factor of conflict of interest could have biased the result described above. Within the included SVR studies, we identified 3 groups of systems called IREX,^{57,58,61} Virtual Reality Rehabilitation System (VRRS),[†] and other commercial systems^{14,15,55,66,67} that qualified as commercially available devices for clinics. We then separated the funnel plots by these groups and contrasted them with systems that

*References 2, 14, 15, 42, 46-49, 54-67

†References 2, 47, 54, 59, 60, 63, 65

Table 2. Characteristics of Included Studies.^a

Author	Intervention	n	Age	DSS	Phase	Type of VR	PEdro
SVR studies							
Aşkin et al, 2018 ⁴⁹	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours	18/38	55, 10.4	603.33 (151.33)	Chronic	VR environment on TV and motion tracking through Microsoft Kinect	6
Brunner et al, 2017 ⁶⁷	VR + CT vs CT + CT; 4 × 4.1 × 5.1.1 (107.2) = 43.7 hours ^b	57/112	62 (32-88)	34.5 (20)	Subacute	VR environment on computer and motion tracking through data gloves	9
da Silva Cameirão et al, 2011 ⁵⁵	VR + OT vs intensive OT; 12 × 3 × 20 = 12 hours ^b	10/19	61.4 (11.6)	13.2 (5.2)	Acute	VR environment on computer and motion tracking through computer vision and data gloves	7
Crosbie et al, 2012 ⁵⁶	VR vs CT; 3 × 3 × 30-45 = 4.5-6.8 hours ^b	9/18	60.3 (10.9)	329 (216)	Chronic	VR environment in head-mounted display and motion tracking through sensors	9
Duff et al, 2012 ⁴²	VR vs PT; 4 × 3 × 60 = 12 hours ^b	11/21	68.8 (8.2)	392 (316)	Chronic	Mixed VR environment and motion tracking through computer vision	6
Jang et al, 2005 ⁵⁷	VR vs passive control; 4 × 5 × 60 = 20 hours	5/10	57.1 (4.5)	414 (88)	Chronic	VR environment on screen and motion tracking through a video camera	5
Jo et al, 2012 ⁵⁸	VR + CT vs CT; 4 × 5 × 60 (18) = 26 hours	15/29	63.85 (7.95)	NA	NA	VR environment on screen and motion tracking through a video camera	6
Kiper et al, 2011 ⁵⁹	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours ^b	40/80	64.0 (16.4)	173.4 (106.5)	Chronic	VR environment on screen and motion tracking through video camera	6
Kiper et al, 2014 ⁶⁰	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours ^b	23/44	64.3 (12.6)	127.8 (94.3)	Chronic	VR environment on screen and motion tracking through video camera	7
Kiper et al, 2018 ⁶⁵	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours ^b	68/136	63.9 (14.1)	127.75 (91.25)	Chronic	VR environment on screen and motion tracking through video camera	6
Kottink et al, 2014 ⁶⁶	VR vs CT; 6 × 3 × 30 = 9 hours ^b	8/18	61.85 (10.65)	1196.9 (743.69)	Chronic	VR environment on horizontal screen and motion tracking through webcam	6
Kwon et al, 2012 ⁶¹	VR + CT vs CT; 4 × 5 × 30 (70) = 33 hours	13/26	57.5 (13.7)	24.3 (18.1)	Subacute	VR environment on screen and motion tracking through video camera	9
Lee et al, 2016 ⁶²	VR + OT vs TV + OT; 6 × 3 × 30 (50) = 24 hours ^b	10/18	71.2 (7.2)	504.9 (196.4)	Chronic	Mixed VR environment on computer and motion tracking through video camera	8
Levin et al, 2012 ¹⁵	VR vs OT; 3 × 3 × 45 = 6.75 ^b	6/12	58.95 (14.85)	1168 (383.25)	Chronic	VR environment on screen and motion tracking through video camera	6
Piron et al, 2009 ²	VR vs CT; 4 × 5 × 60 = 20 hours ^b	18/36	65.2 (7.8)	405 (158)	Chronic	VR environment on computer and motion tracking through sensors	7
Piron et al, 2010 ⁶³	VR vs CT; 4 × 5 × 60 = 20 hours ^b	27/47	60.5 (9)	464 (374)	Chronic	VR environment on screen and motion tracking through sensors	8
Shin et al, 2014 ⁶⁴	VR + OT vs OT; 2 × 5 × 20 (20) = 6.6 hours	9/16	49.3 (8.9)	71.9 (36.9)	Subacute	VR environment on screen and motion tracking through depth sensor	8
Standen et al, 2016 ⁴⁸	VR vs passive control; 8 × 5 × 60 = 40 hours (maximum), actual ~7 hours	9/18	61 (13.1)	119 (83-279)	Subacute	VR environment on screen and motion tracking through light-emitting diodes	5
Turolla et al, 2013 ⁵⁴	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours ^b	68/100	62.8 (13.4)	<91	Subacute	VR environment on screen and motion tracking through sensors	5

(continued)

Table 2. (continued)

Author	Intervention	n	Age	DSS	Phase	Type of VR	PEDro
Turolla et al, 2013 ⁵⁴	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours	113/170	62.8 (13.4)	91-365	Subacute	VR environment on screen and motion tracking through sensors	5
Turolla et al, 2013 ⁵⁴	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours	82/106	62.8 (13.4)	>365	Chronic	VR environment on screen and motion tracking through sensors	5
Yin et al, 2014 ⁶⁶	VR + PT/OT vs PT/OT; 2 × 4.5 × 30 (90) = 18 hours	11/23	58.3 (13.5)	16.3 (7.4)	Acute	VR environment on screen and motion tracking through hand-held sensors	6
Zondervan et al, 2016 ¹⁴	VR vs standard at home training; 3 × 3 × 60 = 9 hours ^b	9/17	59.5 (40-74)	1551.3 (1058.5)	Chronic	VR environment on laptop and motion tracking through sensors	8
Zucconi et al, 2011 ⁴⁷	VR vs PT; 4 × 5 × 60 = 20 hours ^b	11/22	62.25 (56-73)	236.5 (88-544)	Chronic	VR environment on screen and motion tracking through sensors	8
Mean	4.4 × 4.4 × 49.7 (55) = 23.9 hours	27.1/43.6	61.4	370.4			6.7
NSVR studies							
da Silva, Ribeiro et al, 2015 ⁴³	VR vs PT; 2 × 2 × 60 = 4 hours ^b	15/30	53.3 (7.4)	1559 (1080)	Chronic	Nintendo Wii	5
Kong et al, 2016 ⁶⁸	VR + PT/OT vs CT + PT/OT; 3 × 4 × 60 (75) = 27 hours ^b	33/67	57.5 (9.8)	13.7 (8.9)	Acute	Nintendo Wii	9
Rand et al, 2017 ⁶⁹	VR vs standard at home therapy; 5 × 6 × 37.6 = 18.8 hours ^b	13/24	62 (8.7)	495.8 (263.1)	Chronic	Microsoft Xbox Kinect or Sony PlayStation EyeToy	7
Saposnik et al, 2010 ⁴⁴	VR + CT vs recreational therapy + CT; 2 × 4 × 60 (60) = 16 hours ^b	9/18	61.3 (13)	24.7 (12.5)	Subacute	Nintendo Wii	5
Saposnik et al, 2016 ⁸	VR + CT vs recreational therapy + CT; 2 × 5 × 60 (37.3) = 16 hours ^b	59/121	62 (12.5)	25.8 (9.5-46.75)	Subacute	Nintendo Wii	6
Sin and Lee, 2013 ⁷⁰	VR + OT vs OT; 6 × 3 × 30 (30) = 18 hours	18/35	73.7 (7.5)	239 (64)	Chronic	Microsoft Xbox Kinect	6
Türkbey et al, 2017 ⁷¹	VR + CT vs CT; 4 × 5 × 60 (60) = 40 hours	10/19	62 (38-79)	47 (13-125)	Subacute	Microsoft Xbox Kinect	9
Yavuzer et al, 2008 ⁷²	VR + CT vs CT + watching VR; 4 × 5 × 30 (60) = 30 hours	10/20	61.1 (8)	118.7 (70)	Subacute	Sony PlayStation EyeToy	8
Mean	3.5 × 4.3 × 52.5 (57) = 21.9 hours	20.9/41.8	61.6	315.4			6.9

Abbreviations: CT, conventional therapy; DSS, days since stroke; NSVR, nonspecific VR; OT, occupational therapy; PT, physical therapy; SVR, specific VR; VR, virtual reality.

^aIntervention: intervention (VR) versus control group (CT, OT, PT), Weeks × Sessions per week × Minutes (if additional CT was given) = Total amount of intervention in hours; n = Number of patients in intervention/Total number of patients. Age: mean years (SD or range). Phase: acute, 1 day to 3 weeks; subacute, 3 weeks to 3 months; chronic, more than 3 months after stroke.

^bDose matched between groups.

Table 3. Aim, Outcome Measurements, Main Finding, and Assigned Principles of Included Studies.^a

Author	Aim	ICF-WHO category				Follow-up	Main Finding	Principles
		BF	AC	PP	Other Scales			
SVR studies								
Aşkin et al, 2018 ⁴⁹	Effect of VR on upper-limb recovery	FM-UE	BBT		MAS, BS, MI	No	FM-UE significantly higher for VR than control after treatment	- Dosage - Task-specific practice - Variable practice - Multisensory stimulation
Brunner et al, 2017 ⁶⁷	Compare effectiveness of VR to CT		FIM		BBT, ARAT, Abilhand, PGIC	3 Months	No significant difference after treatment, both groups improved	- Dosage - Task-specific practice - Variable practice - Increasing difficulty - Implicit feedback
da Silva Cameirão et al, 2011 ⁵⁵	Clinical impact of VR on recovery time course	FM-UE	BI		MRC, MI CAHAI	24 Weeks	FM-UE significantly higher for VR than control after treatment	- Task-specific practice - Variable practice - Avatar representation - Increasing difficulty - Implicit feedback - Promote use of affected limb
Crosbie et al, 2012 ⁵⁶	Effectiveness of VR to CT on motor rehabilitation	MI	ARAT			6 Weeks	VR maintained improvement in MI at follow-up	- Variable practice - Increasing difficulty - Promote use of affected limb
Duff et al, 2012 ⁴²	Compare VR and PT	FM-UE	WMFT	SIS	MAL QOM/AOU	No	FM-UE significantly higher for control than VR after treatment	- Variable practice - Multisensory stimulation - Explicit feedback - Implicit feedback - Promote use of affected limb
Jang et al, 2005 ⁵⁷	Effect of VR on cortical reorganization and motor recovery	FM-UE	BBT		MAL QOM/AOU, MFT	No	FM-UE significantly higher for VR than control after treatment	- Task-specific practice - Variable practice - Increasing difficulty - Avatar representation - Implicit feedback - Promote use of affected limb
Jo et al, 2012 ⁵⁸	Changes in upper-extremity function and visual perception using VR		WMFT		MVPT	No	No significant difference after treatment, both groups improved significantly in WMFT	- Dosage - Structured practice - Variable practice - Increasing difficulty - Explicit feedback - Promote use of affected limb
Kiper et al, 2011 ⁵⁹	Impact of VR versus CT on treatment of upper extremity	FM-UE	FIM		MAS	No	FM-UE significantly higher for VR than control after treatment	- Dosage - Task-specific practice - Variable practice - Increasing difficulty - Implicit feedback - Promote use of affected limb
Kiper et al, 2014 ⁶⁰	Is VR more effective than CT on treatment of upper-limb motor function	FM-UE	FIM			No	FM-UE significantly higher for VR than control after treatment	- Dosage - Variable practice - Increasing difficulty - Implicit feedback - Promote use of affected limb
Kiper et al, 2018 ⁶⁵	Effectiveness of reinforced feedback in VR vs CT	FM-UE	FIM		NIHSS, ESAS	No	FM-UE significantly higher for VR than control after treatment	- Dosage - Task-specific practice - Variable practice - Multisensory stimulation - Explicit feedback - Implicit feedback - Promote the use of affected limb
Kottink et al, 2014 ⁶⁶	Compare effect of VR to CT on arm function	FM-UE	ARAT			1 Month	No significant difference after treatment, both groups improved significantly in FM-UE	- Task-specific practice - Increasing difficulty - Explicit feedback - Promote the use of affected limb

(continued)

Table 3. (continued)

Author	Aim	ICF-WHO category				Follow-up	Main Finding	Principles
		BF	AC	PP	Other Scales			
Kwon et al, 2012 ⁶¹	Impact of VR with CT on upper-extremity function and ADL in acute stage	FM-UE	BI		MFT	No	No significant difference after treatment, both groups improved significantly in FM-UE	- Dosage - Task-specific practice - Variable practice - Avatar representation
Lee et al, 2016 ⁶²	Effect of VR on upper-limb function and muscle strength		BBT		JTHFT, GPT	No	BBT significantly higher for VR than control after treatment	- Structured practice - Variable practice - Implicit feedback - Promote use of affected limb
Levin et al, 2012 ¹⁵	Potential of VR to improve upper-limb motor ability	FM-UE	BBT		CSI, RPSS, WMFT, MAL QOM/AOU	1 Month	More patients improved in FM-UE in VR than control	- Task-specific practice - Variable practice - Avatar representation - Increasing difficulty - Explicit feedback - Promote use of affected limb
Piron et al, 2009 ²	Impact of VR on treating motor deficits	FM-UE			Abilhand, MAS	2 And 3 months	FM-UE significantly higher for VR than control after treatment	- Variable practice - Explicit feedback - Implicit feedback - Promote use of affected limb
Piron et al, 2010 ⁶³	Impact of VR versus CT	FM-UE	FIM			No	FM-UE was systematically lower in control than VR	- Variable practice - Increasing difficulty - Explicit feedback - Implicit feedback - Promote use of affected limb
Shin et al, 2014 ⁶⁴	Assessment of usability and clinical efficacy of VR	FM-UE	BI		MRC	No	FM-UE higher after treatment but not significant for VR	- Task-specific practice - Variable practice - Avatar representation - Increasing difficulty - Explicit feedback - Implicit feedback - Promote use of affected limb
Standen et al, 2016 ⁴⁸	Feasibility of home-based VR for arm rehabilitation		WMFT		9 Peg hole, MAL QOM/AOU	No	WMFT grip strength at midpoint significantly higher improvement for VR	- Massed practice - Task-specific practice - Variable practice - Increasing difficulty - Explicit feedback - Promote use of affected limb
Turolla et al, 2013 ⁵⁴	Effectiveness of VR on restoration of upper-limb function and ADL	FM-UE	FIM			No	FM-UE significantly higher for VR than control after treatment	- Dosage - Task-specific practice - Variable practice - Increasing difficulty - Explicit feedback - Implicit feedback
Yin et al, 2014 ⁴⁶	Effect of VR on rehabilitation of upper-limb motor performance	FM-UE	FIM		ARAT, MAL QOM/AOU	1 Month	No significant difference between groups in FM-UE	- Dosage - Structured practice - Task-specific practice - Multisensory stimulation - Avatar representation - Explicit feedback - Implicit feedback - Promote use of affected limb
Zondervan et al, 2016 ⁴⁴	Feasibility and efficacy of VR at patient's home		ARAT		BBT, MAL QOM/AOU 9 Peg Hole	1 Month	MAL QOM change from baseline significant for VR	- Massed practice - Task-specific practice - Multisensory stimulation - Explicit feedback - Promote use of affected limb
Zucconi et al, 2011 ⁴⁷	Effect of VR on motor impairment	FM-UE	FIM		MAS, RPS	No	Only VR improved significantly after treatment in FM-UE	- Variable practice - Increasing difficulty - Implicit feedback - Promote use of affected limb

(continued)

Table 3. (continued)

Author	Aim	ICF-WHO category				Follow-up	Main Finding	Principles
		BF	AC	PP	Other Scales			
NSVR studies								
da Silva Ribeiro et al, 2015 ⁴³	Effect of VR vs CT on sensorimotor function and quality of life	FM-UE		SF-36		No	No significant difference after treatment, both groups improved significantly in FM-UE	- Structured practice - Variable practice - Increasing difficulty
Kong et al, 2016 ⁶⁸	Efficacy of VR with CT on upper-limb recovery	FM-UE	FIM		ARAT, SIS-UL, VAS	7 And 15 weeks	No significant difference after treatment, both groups improved significantly in FM-UE	- Dosage - Variable practice - Explicit feedback - Promote use of affected limb
Rand et al, 2017 ⁶⁹	Effectiveness of self-training programs on upper-limb function		ARAT		MAL QOM/AOU, BBT	4 Weeks	No significant difference or improvement in MAL QOM after treatment	- Variable practice - Promote use of affected limb
Saposnik et al, 2010 ⁴⁴	Efficacy of VR for stroke rehabilitation	SIS grip strength	WMFT	SIS	BBT	4 Weeks	VR had significant improvement in WMFT, but only at follow-up	- Dosage - Variable practice - Multisensory stimulation - Avatar representation - Implicit feedback
Saposnik et al, 2016 ⁹	Compare safety and efficacy of VR with recreational therapy on motor recovery	SIS grip strength	BI	SIS	WMFT, BBT, FIM, MRS	4 Weeks	No significant difference after treatment, both groups significantly improved in WMFT	- Dosage - Task-specific practice - Variable practice - Promote use of affected limb
Sin and Lee, 2013 ⁷⁰	Effects of additional VR on upper-extremity function	FM-UE	BBT			No	FM-UE significantly higher for VR than control after treatment	- Task-specific practice - Variable practice - Multisensory stimulation - Explicit feedback - Implicit feedback - Promote use of affected limb
Türkbeý et al, 2017 ⁷¹	Feasibility and safety of VR on upper-limb recovery	BS	BBT		WMFT, FIM	No	No significant difference after treatment, both groups significantly improved in WMFT	- Dosage - Task-specific practice - Variable practice - Multisensory stimulation - Avatar representation - Promote use of affected limb
Yavuzer et al, 2008 ⁷²	Effect of VR on upper-limb motor recovery	BS	FIM			3 Months	BS UE significantly higher in VR than control after treatment	- Dosage - Task-specific practice - Variable practice - Increasing difficulty - Promote use of affected limb

Abbreviations: AC, Activity; ADL, activities of daily living; AOU, amount of use; ARAT, Action Research Arm Test; BBT, Box and Block Test; BF, body function; BI, Barthel Index; BS, Brunstrom Motor Recovery Stage; CAHAI, Chedoke Arm and Hand Inventory; CSI, Composite Spaticity Index; CT, conventional therapy; ESAS, Edmonton Symptom Assessment Scale; FIM, Functional Independence Measure; FM-UE, Fugl-Meyer Assessment Upper Extremity; GPT, Grooved Pegboard Test; ICF-WHO, World Health Organization's International Classification of Function, Disability, and Health; JTHFT, Jepsen-Taylor Hand Function Test; MAL, Motor Activity Log; MAS, Modified Ashworth Scale; MFT, Manual Function Test; MI, Motricity Index; MRC, Medical Research Council Grade; MVPT, Motor-Free Visual Perception Test; NIHSS, National Institutes of Health Stroke Scale; NSVR, nonspecific VR; PGIC, Patient Global Impression; PP, Participation; PT, physical therapy; QOM, quality of movement; RPSS, Performance Reaching Scale for Stroke; SF-36, Short-Form Health Survey; SIS, Stroke Impact Scale; SIS-UL, SIS upper limb items; SVR, specific VR; VAS, Visual Analogue Scale; VR, virtual reality; WMFT, Wolf Motor Function Test.
[†]It explains that BF, AC and PP are the ICF-WHO categories.

remained experimental set-ups only (Supplementary Figure 4).[‡] The studies using VRRS are large sized, and therefore, cluster at the top of the effect, both in body function outcomes (SMD = 0.21; 95% CI = 0.06 to 0.36; $P = .007$) and activity outcomes (SMD = 0.38; 95% CI = 0.19 to 0.56; $P < .0001$). Neither of the other groups reached significance. Therefore, the presence of a bias resulting from commercialization cannot be confirmed.

[‡]References 42, 46, 48, 49, 56, 62, 64, 66

Evaluation of Included Principles of Neurorehabilitation

We identified relevant differences between SVR and NSVR studies (Figure 5) with respect to the included principles. In Table 3, the assigned principles for each study can be found and the full data set used for the analysis is provided as a supplemental material. First, the spectrum of the principles that are mentioned in more than 50% of the studies is broader in SVR than NSVR interventions. NSVR studies focused on 3 principles—variable practice,^{8,43,44,68-72} promoting of the use of




Figure 2. Forest plot of functional outcomes: SVR versus NSVR studies on upper-limb function as measured by the selected outcome. Abbreviations: SVR, specific VR; NSVR, nonspecific VR; VR, virtual reality.

the paretic limb,^{8,68-72} and dosage^{8,44,68,71,72}—that were present in 100%, 75%, and 63% of the studies, respectively. SVR studies did not share 1 specific principle in common, but more than 50% of the studies in this category included the same 6 principles: variable practice (86%),[§] promoting the use of the paretic limb (86%),^{||} implicit feedback (64%),[¶] increasing difficulty (64%),[#] task-specific practice (64%),^{**} and explicit feedback (59%).^{††} We conducted a follow-up analysis to evaluate the effect of dosage because NSVR studies seem to have more intense intervention regimes. We compared the outcomes of those studies that provided more than 60 minutes of therapy per session per week day. We identified 14 studies, 9 SVR^{††} and 5 NSVR,^{8,44,68,71,72} that fulfill this criterion. Comparing this

subset of SVR and NSVR studies with their respective controls, we still observe a significant superior impact of SVR studies on body function (SMD = 0.23; 95% CI = 0.07 to 0.38, *P* = .004; see Supplementary Figure 5, upper panel) and activity (SMD = 0.27; 95% CI = 0.13 to 0.41, *P* = .0002, see Supplementary Figure 6, upper panel), whereas the total number of hours of intervention (SVR: mean [SD] = 35.6 [8] hours; NSVR: mean [SD] = 25.8 [9.1] hours) and the number of weeks (SVR: mean [SD] = 3.8 [0.6] weeks; NSVR: mean [SD] = 3 [0.9] weeks) were not significantly different.

Discussion

The use of VR is increasing in neurorehabilitation. However, so far, it is unclear whether VR is effective in enhancing recovery after stroke. We proposed to distinguish between VR systems specifically built for rehabilitation (SVR) and off-the-shelf recreational VR systems (NSVR), based on the assumption that SVR systems incorporate principles of neurorehabilitation that potentially enhance learning and recovery, whereas NSVR systems do not. Our results demonstrate

[§]References 2, 15, 42, 47-49, 54-65, 67

^{||}References 2, 14, 15, 42, 46-49, 55-60, 62-66

[¶]References 2, 42, 46, 47, 54, 55, 57, 59, 60, 62-65, 67

[#]References 15, 47, 48, 54-60, 63, 64, 66, 67

^{**}References 14, 15, 46, 48, 49, 54, 55, 57, 59, 61, 64-67

^{††}References 2, 14, 15, 42, 46, 48, 49, 54, 58, 63-66

^{†††}References 46, 49, 54, 58-61, 65, 67




Figure 3. Forest plot of activity outcomes. SVR versus NSVR studies on upper-limb activity as measured by the selected outcome. Abbreviations: NSVR, nonspecific VR; SVR, specific VR; VR, virtual reality.

that SVR systems show a higher impact on recovery, on body function, and on activity than CT and that NSVR systems do not. This is in line with evidence found for the use of VR interventions to train balance and gait¹¹ and the most recent meta-analysis on VR interventions.²⁵ The difference between our results and previous analyses is our focus on rehabilitation tools in VR for enhancing upper-limb function and activity only. Hence, the recategorization in SVR and NSVR systems provides a valid basis for the reinterpretation of effects reported in previous reviews.

We propose that the overall positive effect of SVR protocols is a result of the incorporation of principles of neurorehabilitation that enhance motor learning and recovery. Of the 11 principles identified through the literature, we found 6 to be present in more than 50% of the SVR studies. In NSVR interventions, however, only 3 principles surpassed this level. Variable practice ranked high in both SVR and NSVR studies. In VR systems, variable practice can be easily achieved by including a variety of tasks with different goals, movement

requirements, and stimuli²³ to enhance learning¹⁶ and retention.⁷³ In addition, the variety can make repetitive training more engaging and enjoyable for the patient and counteract boredom, which has been associated with low adherence to standard training protocols.^{3,7} However, variable practice alone is possibly not sufficient to lead to a noticeable effect on recovery. If applied, the 5 additional principles that were present in SVR systems would generate a VR training that challenges the patient optimally through adaptive difficulty^{22,23} while providing information on success (results)^{18,23} and optimizing implicit error-based learning (performance)^{18,23} through tasks that are relevant for ADL (task specific)^{17,19,23} besides promoting the use of the paretic arm.^{17,19,22} Only 1 SVR study included all 6 principles⁶⁴ and showed a large positive effect for the experimental group in recovery of body function (SMD = 1.13; 95% CI = 0.04 to 2.21) and activity (SMD = 1.47; 95% CI = 0.32 to 2.62).

However, besides known methodological issues,⁹ we note that many protocols relied on the therapists to individualize

Specific and nonspecific VR-based systems compared to conventional therapy in upper limb recovery of function and activity						
Patient or population: Stroke patients Setting: Clinic or at home Intervention: Specific or nonspecific VR-based rehabilitation systems Comparison: Conventional therapy						
Outcomes	Anticipated absolute effects [*] (95% CI)		Relative effect (95% CI)	No of participants (studies)	Certainty of the evidence (GRADE)	Comments
	Risk with conventional therapy	Risk with VR-based systems				
Function - SVR	-	The mean score in the intervention group was SMD 0.23 higher (0.1 higher to 0.36 higher)	-	945 (19 RCTs)	⊕⊕⊕⊕ MODERATE ¹	
Function - NSVR	-	The mean score in the intervention group was SMD 0.16 higher (0.14 lower to 0.47 higher)	-	310 (7 RCTs)	⊕⊕⊕⊕ LOW ^{1,2}	
Activity - SVR	-	The mean score in the intervention group was SMD 0.31 higher (0.15 higher to 0.47 higher)	-	1103 (23 RCTs)	⊕⊕⊕⊕ MODERATE ¹	
Activity - NSVR	-	The mean score in the intervention group was SMD 0.15 higher (0.15 lower to 0.45 higher)	-	304 (7 RCTs)	⊕⊕⊕⊕ LOW ^{1,2}	
[*] The risk in the intervention group (and its 95% confidence interval) is based on the assumed risk in the comparison group and the relative effect of the intervention (and its 95% CI). CI: Confidence interval; RR: Risk ratio; OR: Odds ratio;						
GRADE Working Group grades of evidence High certainty: We are very confident that the true effect lies close to that of the estimate of the effect Moderate certainty: We are moderately confident in the effect estimate: The true effect is likely to be close to the estimate of the effect, but there is a possibility that it is substantially different Low certainty: Our confidence in the effect estimate is limited: The true effect may be substantially different from the estimate of the effect Very low certainty: We have very little confidence in the effect estimate: The true effect is likely to be substantially different from the estimate of effect						
Footnotes ¹ Directness was downgraded because studies were only available for specific versus control or non-specific versus control. ² Precision was downgraded because confidence intervals include null effect and effect of SVR.						

Figure 4. Summary of findings for the main comparisons. The quality of evidence for this review was evaluated using GRADEpro, finding a moderate certainty of the effects observed in SVR studies. Abbreviations: SVR, specific VR; NSVR, nonspecific VR; VR, virtual reality.

the practice to the patient’s needs by selecting the training task or movement requirements or adjusting the difficulty parameters.^{§§} This might have biased the outcomes and could compromise the internal and external validity of these studies. Computerized systems have the advantage that every principle could be customized to the patient’s individual ability and necessity automatically.⁷⁴ Whereas NSVR systems are typically not similarly adaptive and accessible for modification, this is a unique opportunity for SVR systems.

The results presented in this study do require further investigation for several reasons. First, it must be noted that the included studies may not have published all the details of their intervention. We, therefore, cannot rule out the possibility that VR systems in this analysis might have incorporated principles that were not detected and reported. To conclusively identify the “active ingredients”³³ of effective VR systems, a structured interview with the study authors

might be the best approach. We see our analysis, however, as a first attempt to shift awareness from form (VR) to content (principles). Second, we recognize that the content analysis could gain validity if it was performed by an independent rater. However, given the relatively small set of indicators and the availability of the full data set with this article, we believe that this risk is sufficiently mitigated. Furthermore, the number of studies included in the NSVR category is relatively small, and therefore, the nonsignificant effect may be a result of low statistical power. However, individual studies do report sufficient sample sizes. In addition, besides the exclusion of small studies, a source of reporting bias may relate to SVR systems that are commercially available to clinics. However, the system with the largest populations clustered well around the mean effect magnitude and the slight skewness is a result of commercial and noncommercial systems. Hence, a bias resulting from financial interest cannot be confirmed. Another potential limitation of our meta-analysis is the high heterogeneity

^{§§}References 54, 56-58, 60, 63-65, 67, 68




Figure 5. Distribution of included principles in SVR versus NSVR studies. Blue indicates SVR and red NSVR studies. Abbreviations: AR, avatar representation; D, dosage; EF, explicit feedback; ID, increasing difficulty; IF, implicit feedback; MP, massed practice; MS, multisensory stimulation; NSVR, nonspecific VR; PUA, promote the use of the affected limb; SP, structured practice; SVR, specific VR; TSP, task-specific practice; VP, variable practice; VR, virtual reality.

across studies in terms of intervention protocols (eg, training intensity, type of task, movement patterns addressed, etc) and the measurement tools used (eg, the clinical scales). This also made it impossible to provide proof for the clinical relevance of our finding. Values of clinically important differences are not available for all clinical scales and chronicity bands established. For instance, for FM-UE, clinically important differences are available for the subacute⁷⁵ and chronic⁷⁶ phases but not for the acute phase. Despite these limitations, we are confident about the higher impact of SVR systems on motor recovery because the groups were narrowly defined. Our results may also aid researchers in selecting the appropriate principles that drive the desired outcome and then identify the technology that can best implement and deliver these principles. This could be VR alone or coupled with other technologies (eg, robotics or exogenous stimulation), potentially further enhancing recovery.

Conclusion

Overall, our findings suggest that tailor-made VR systems for neurorehabilitation may be valid tools to deliver effective motor rehabilitation poststroke. Future studies should, therefore, not ask if VR should be used or not. Instead, they should investigate which technology, including VR, is most appropriate to facilitate the implementation of principles of

neurorehabilitation in a more effective way than CT. We believe that VR is well suited for rehabilitation because it allows the patient to interact in a safe and ecologically valid environment, where the exposure to sensorimotor contingencies can be controlled and modulated in a goal-oriented and autonomous fashion. In our analysis, the superiority of specific VR systems is associated with the following combination of principles that might possibly lead to a greater effect on recovery: task-specific practice, explicit feedback, increasing difficulty, implicit feedback, variable practice, and mechanisms to promote the use of the paretic limb. We are confident that dedicated VR-based systems are well suited for exploiting these principles, and we expect that future technologies will contribute to an even more advantageous implementation of this set of principles underlying recovery and brain repair.

Declaration of Conflicting Interests

The authors declared the following potential conflicts of interest with respect to the research, authorship, and/or publication of this article: PFMJV is founder and interim CEO of Eodyne S L, which aims at bringing scientifically validated neurorehabilitation technology to society. The rest of the authors have nothing to disclose.

Funding

The authors disclosed receipt of the following financial support for the research, authorship, and/or publication of this article: This

study was supported by SANAR (MINECO, TIN2013- 44200), cDAC (ERC 2013 ADG 341196), and socSMCs (Grant Number EC, H2020-641321).

References

1. Feigin VL, Lawes CMM, Bennett DA, Anderson CS. Stroke epidemiology: a review of population-based studies of incidence, prevalence, and case-fatality in the late 20th century. *Lancet Neurol.* 2003;2:43-53. doi:10.1016/S1474-4422(03)00266-7
2. Piron L, Turolla A, Agostini M, et al. Exercises for paretic upper limb after stroke: a combined virtual-reality and telemedicine approach. *J Rehabil Med.* 2009;41:1016-1020. doi:10.2340/16501977-0459
3. Proffitt R, Lange B. Considerations in the efficacy and effectiveness of virtual reality interventions for stroke rehabilitation: moving the field forward. *Phys Ther.* 2015;95:441-448. doi:10.2522/ptj.20130571
4. Crosbie JH, Lennon S, Basford JR, McDonough SM. Virtual reality in stroke rehabilitation: still more virtual than real. *Disabil Rehabil.* 2007;29:1139-1146. doi:10.1080/09638280600960909
5. Saposnik G, Levin M; Outcome Research Canada (SORCan) Working Group. Virtual reality in stroke rehabilitation: a meta-analysis and implications for clinicians. *Stroke.* 2011;42:1380-1386. doi:10.1161/STROKEAHA.110.605451
6. Lohse KR, Hilderman CGE, Cheung KL, Tatla S, Van Der Loos HFM. Virtual reality therapy for adults post-stroke: a systematic review and meta-analysis exploring virtual environments and commercial games in therapy. *PLoS One.* 2014;9(3):e93318. doi:10.1371/journal.pone.0093318
7. Saywell N, Taylor N, Rodgers E, Skinner L, Boocock M. Play-based interventions improve physical function for people with adult-acquired brain injury: a systematic review and meta-analysis of randomised controlled trials. *Clin Rehabil.* 2017;31:145-157. doi:10.1177/0269215516631384
8. Saposnik G, Cohen LG, Mamdani M, et al; Stroke Outcomes Research Canada. Efficacy and safety of non-immersive virtual reality exercising in stroke rehabilitation (EVREST): a randomised, multicentre, single-blind, controlled trial. *Lancet Neurol.* 2016;15:1019-1027. doi:10.1016/S1474-4422(16)30121-1
9. Laver KE, Lange B, George S, Deutsch JE, Saposnik G, Crotty M. Virtual reality for stroke rehabilitation. *Cochrane Database Syst Rev.* 2017;(11):CD008349. doi:10.1002/14651858.CD008349.pub4
10. Shrier I, Platt RW, Steele RJ. Mega-trials vs meta-analysis: precision vs heterogeneity? *Contemp Clin Trials.* 2007;28:324-328. doi:10.1016/j.cct.2006.11.007
11. de Rooij IJM, van de Port IGL, Meijer JWG. Effect of virtual reality training on balance and gait ability in patients with stroke: systematic review and meta-analysis. *Phys Ther.* 2016;96:1905-1918. doi:10.2522/ptj.20160054
12. Jack D, Boian R, Merians AS, et al. Virtual reality-enhanced stroke rehabilitation. *IEEE Trans Neural Syst Rehabil Eng.* 2001;9:308-318. doi:10.1109/7333.948460
13. Kizony R, Katz N, Weiss PL. Adapting an immersive virtual reality system for rehabilitation. *J Vis Comput Animat.* 2003;14:261-268. doi:10.1002/vis.323
14. Zondervan DK, Friedman N, Chang E, et al. Home-based hand rehabilitation after chronic stroke: randomized, controlled single-blind trial comparing the MusicGlove with a conventional exercise program. *J Rehabil Res Dev.* 2016;53:457-472.
15. Levin MF, Snir O, Liebermann DG, Weingarden H, Weiss PL. Virtual reality versus conventional treatment of reaching ability in chronic stroke: clinical feasibility study. *Neurol Ther.* 2012;1:3. doi:10.1007/s40120-012-0003-9
16. Hanlon RE. Motor learning following unilateral stroke. *Arch Phys Med Rehabil.* 1996;77:811-815. doi:10.1016/S0003-9993(96)90262-2
17. Dobkin BH. Strategies for stroke rehabilitation. *Lancet Neurol.* 2004;3:528-536. doi:10.1016/S1474-4422(04)00851-8
18. Cirstea MC, Levin MF. Improvement of arm movement patterns and endpoint control depends on type of feedback during practice in stroke survivors. *Neurorehabil Neural Repair.* 2007;21:398-411. doi:10.1177/1545968306298414
19. Kleim JA, Jones TA. Principles of experience-dependent neural plasticity: implications for rehabilitation after brain damage. *J Speech Lang Hear Res.* 2008;51:S225-S239. doi:10.1044/1092-4388(2008)018
20. Kwakkel G. Intensity of practice after stroke: more is better. *Schweiz Arch Neurol Psychiatr.* 2009;160:295-298. doi:10.1080/09638280500534861
21. Christ O, Reiner M. Perspectives and possible applications of the rubber hand and virtual hand illusion in non-invasive rehabilitation: technological improvements and their consequences. *Neurosci Biobehav Rev.* 2014;44:33-44. doi:10.1016/j.neubiorev.2014.02.013
22. Kwakkel G, Veerbeek JM, van Wegen EEH, Wolf SL. Constraint-induced movement therapy after stroke. *Lancet Neurol.* 2015;14:224-234. doi:10.1016/S1474-4422(14)70160-7
23. Levin MF, Weiss PL, Keshner EA. Emergence of virtual reality as a tool for upper limb rehabilitation: incorporation of motor control and motor learning principles. *Phys Ther.* 2015;95:415-425. doi:10.2522/ptj.20130579
24. Thomas LH, French B, Coupe J, et al. Repetitive task training for improving functional ability after stroke. *Stroke.* 2017;48:102-104. doi:10.1161/STROKEAHA.117.016503
25. Aminov A, Rogers JM, Middleton S, Caeyenberghs K, Wilson PH. What do randomized controlled trials say about virtual rehabilitation in stroke? A systematic literature review and meta-analysis of upper-limb and cognitive outcomes. *J Neuroeng Rehabil.* 2018;15:29. doi:10.1186/s12984-018-0370-2
26. World Health Organization. Classifications: International Classification of Functioning, Disability and Health (ICF). <http://www.who.int/classifications/icf/en/>. Accessed September 21, 2018.
27. GestureTek.IREX.<http://www.gesturetekhealth.com/products/irex>. Accessed September 24, 2018.
28. Moher D, Liberati A, Tetzlaff J, Altman DG; PRISMA Group. Preferred reporting items for systematic reviews

- and meta-analyses: the PRISMA Statement. *PLoS Med.* 2009;6:e1000097. doi:10.1371/journal.pmed.1000097
29. Lombard M, Ditton T. At the heart of it all: the concept of presence. *J Comput Mediat Commun.* 2006;3(2):JCMC321. doi:10.1111/j.1083-6101.1997.tb00072.x
 30. Steuer J, Breitrose H, Cool J, et al. Defining virtual reality: dimensions determining telepresence. *J Commun.* 1992;42:73-93.
 31. Zeiler SR, Hubbard RB, Gibson EM, et al. Paradoxical motor recovery from a first stroke by re-opening a sensitive period with a second stroke: reopening a postischemic sensitive period. *Neurorehabil Neural Repair.* 2015;46(suppl 1):ATP86. doi:10.1177/1545968315624783
 32. Duncan PW. Outcome measures in stroke rehabilitation. *Handb Clin Neurol.* 2013;110:105-111.
 33. Fugl-Meyer AR, Jääskö L, Leyman I, Olsson S, Stegling S. The post-stroke hemiplegic patient: 1. A method for evaluation of physical performance. *Scand J Rehabil Med.* 1975;7:13-31.
 34. Collin C, Wade D. Assessing motor impairment after stroke: a pilot reliability study. *J Neurol Neurosurg Psychiatry.* 1990;53:576-579. doi:10.1136/JNPN.53.7.576
 35. Brunnstrom S. Motor testing procedures in hemiplegia: based on sequential recovery stages. *Phys Ther.* 1966;46:357-375. doi:10.1093/ptj/46.4.357
 36. Duncan PW, Wallace D, Lai SM, Johnson D, Embretson S, Laster LJ. The stroke impact scale version 2.0: evaluation of reliability, validity, and sensitivity to change. *Stroke.* 1999;30:2131-2140. doi:10.1161/01.STR.30.10.2131
 37. Mahoney FI, Barthel DW. Functional evaluation: the Barthel Index. *Md State Med J.* 1965;14:61-65.
 38. Keith RA, Granger CV, Hamilton BB, Sherwin FS. The functional independence measure: a new tool for rehabilitation. *Adv Clin Rehabil.* 1987;1:6-18.
 39. Carroll D. A Quantitative test of upper extremity function. *J Chronic Dis.* 1965;18:479-491. doi:10.1016/0021-9681(65)90030-5
 40. Mathiowetz V, Volland G, Kashman N, Weber K. Adult norms for the Box and Block Test of Manual Dexterity. *Am J Occup Ther.* 1985;39:386-391. doi:10.5014/ajot.39.6.386
 41. Wolf SL, Catlin PA, Ellis M, Archer AL, Morgan B, Piacentino A. Assessing Wolf Motor Function Test as outcome measure for research in patients after stroke. *Stroke.* 2001;32:1635-1639. doi:10.1161/01.STR.32.7.1635
 42. Duff M, Chen Y, Cheng L, et al. Adaptive mixed reality rehabilitation improves quality of reaching movements more than traditional reaching therapy following stroke. *Neurorehabil Neural Repair.* 2012;27:306-315. doi:10.1177/1545968312465195
 43. da Silva Ribeiro NM, Ferraz DD, Pedreira É, et al. Virtual rehabilitation via Nintendo Wii® and conventional physical therapy effectively treat post-stroke hemiparetic patients. *Top Stroke Rehabil.* 2015;22:299-305. doi:10.1179/1074935714Z.0000000017
 44. Saposnik G, Teasell R, Mamdani M, et al; Stroke Outcome Research Canada (SORCan) Working Group. Effectiveness of virtual reality using wii gaming technology in stroke rehabilitation: a pilot randomized clinical trial and proof of principle. *Stroke.* 2010;41:1477-1484. doi:10.1161/STROKEAHA.110.584979
 45. Ware JE Jr, Sherbourne CD. The MOS 36-item short-form health survey (SF-36): I. Conceptual framework and item selection. *Med Care.* 1992;30:473-483.
 46. Yin CW, Sien NY, Ying LA, Chung FM, Leng DTM. Virtual reality for upper extremity rehabilitation in early stroke: a pilot randomized controlled trial. *Clin Rehabil.* 2014;28:1107-1114. doi:10.1177/0269215514532851
 47. Zucconi C, Valt V, Agostini M, Turolla A, Tonin P, Piron L. Assessment of a virtual teacher feedback for the recovery of the upper limb after stroke. *Ital J Physiother.* 2011;1:101-106. doi:http://dx.doi.org/10.1177/1545968312447071
 48. Standen PJ, Threapleton K, Richardson A, et al. A low cost virtual reality system for home based rehabilitation of the arm following stroke: a randomised controlled feasibility trial. *Clin Rehabil.* 2016;31:340-350. doi:10.1177/0269215516640320
 49. Aşkın A, Atar E, Koçyiğit H, Tosun A. Effects of Kinect-based virtual reality game training on upper extremity motor recovery in chronic stroke. *Somatosens Mot Res.* 2018;35:25-32. doi:10.1080/08990220.2018.1444599
 50. Wan X, Wang W, Liu J, Tong T. Estimating the sample mean and standard deviation from the sample size, median, range and/or interquartile range. *BMC Med Res Methodol.* 2014;14:135. doi:10.1186/1471-2288-14-135
 51. Maher CG, Sherrington C, Herbert RD, Moseley AM, Elkins M. Reliability of the PEDro scale for rating quality of randomized controlled trials. *Phys Ther.* 2003;83:713-721. doi:10.1093/ptj/83.8.713
 52. Yamazaki T, Nagao S, Lennon W, Tanaka S. Modeling memory consolidation during posttraining periods in cerebellar vestibular learning. *Proc Natl Acad Sci U S A.* 2015;112:3541-3546. doi:10.1073/pnas.1413798112
 53. Higgins JPT, Thompson SG. Quantifying heterogeneity in a meta-analysis. *Stat Med.* 2002;21:1539-1558. doi:10.1002/sim.1186
 54. Turolla A, Dam M, Ventura L, et al. Virtual reality for the rehabilitation of the upper limb motor function after stroke: a prospective controlled trial. *J Neuroeng Rehabil.* 2013;10:85. doi:10.1186/1743-0003-10-85
 55. da Silva Cameirão M, I Badia SB, Duarte E, Verschure PFMJ. Virtual reality based rehabilitation speeds up functional recovery of the upper extremities after stroke: a randomized controlled pilot study in the acute phase of stroke using the Rehabilitation Gaming System. *Restor Neurol Neurosci.* 2011;29:287-298. doi:10.3233/RNN-2011-0599
 56. Crosbie JH, Lennon S, McGoldrick MC, McNeill MD, McDonough SM. Virtual reality in the rehabilitation of the arm after hemiplegic stroke: a randomized controlled pilot study. *Clin Rehabil.* 2012;26:798-806. doi:10.1177/0269215511434575
 57. Jang SH, You SH, Hallett M, et al. Cortical reorganization and associated functional motor recovery after virtual reality in patients with chronic stroke: an experimenter-blind preliminary study. *Arch Phys Med Rehabil.* 2005;86:2218-2223. doi:10.1016/j.apmr.2005.04.015
 58. Jo K, Jung J, Yu J. Effects of virtual reality-based rehabilitation on upper extremity function and visual perception in

- stroke patients: a randomized control trial. *J Phys Ther Sci*. 2012;24:1205-1208. doi:10.1589/jpts.24.1205
59. Kiper P, Piron L, Turolla A, Stozek J, Tonin P. The effectiveness of reinforced feedback in virtual environment in the first 12 months after stroke. *Neurol Neurochir Pol*. 2011;45:436-444. doi:10.1016/S0028-3843(14)60311-X
60. Kiper P, Agostini M, Luque-Moreno C, Tonin P, Turolla A. Reinforced feedback in virtual environment for rehabilitation of upper extremity dysfunction after stroke: preliminary data from a randomized controlled trial. *Biomed Res Int*. 2014;2014:752128. doi:10.1155/2014/752128
61. Kwon JS, Park MJ, Yoon IJ, Park SH. Effects of virtual reality on upper extremity function and activities of daily living performance in acute stroke: a double-blind randomized clinical trial. *NeuroRehabilitation*. 2012;31:379-385. doi:10.3233/NRE-2012-00807
62. Lee S, Kim Y, Lee BH. Effect of virtual reality-based bilateral upper extremity training on upper extremity function after stroke: a randomized controlled clinical trial. *Occup Ther Int*. 2016;23:357-368. doi:10.1002/oti.1437
63. Piron L, Turolla A, Agostini M, et al. Motor learning principles for rehabilitation: a pilot randomized controlled study in poststroke patients. *Neurorehabil Neural Repair*. 2010;24:501-508. doi:10.1177/1545968310362672
64. Shin JH, Ryu H, Jang S. A task-specific interactive game-based virtual reality rehabilitation system for patients with stroke: a usability test and two clinical experiments. *J Neuroeng Rehabil*. 2014;11:32. doi:10.1186/1743-0003-11-32
65. Kiper P, Szczudlik A, Agostini M, et al. Virtual reality for upper limb rehabilitation in subacute and chronic stroke: a randomized controlled trial. *Arch Phys Med Rehabil*. 2018;99:834-842.e4. doi:10.1016/j.apmr.2018.01.023
66. Kottink AIR, Prange GB, Krabben T, Rietman JS, Burke JH. Gaming and conventional exercises for improvement of arm function after stroke: a randomized controlled pilot study. *Games Health J*. 2014;3:184-191. doi:10.1089/g4h.2014.0026
67. Brunner I, Skouen JS, Hofstad H, et al. Virtual reality training for upper extremity in subacute stroke (VIRTUES): a multicenter RCT. *Neurology*. 2017;89:2413-2421. doi:10.1212/WNL.0000000000004744
68. Kong KH, Loh YJ, Thia E, et al. Efficacy of a virtual reality commercial gaming device in upper limb recovery after stroke: a randomized, controlled study. *Top Stroke Rehabil*. 2016;23:333-340. doi:10.1080/10749357.2016.1139796
69. Rand D, Weingarden H, Weiss R, et al. Self-training to improve UE function at the chronic stage post-stroke: a pilot randomized controlled trial. *Disabil Rehabil*. 2017;39:1541-1548. doi:10.1080/09638288.2016.1239766
70. Sin H, Lee G. Additional virtual reality training using Xbox Kinect in stroke survivors with hemiplegia. *Am J Phys Med Rehabil*. 2013;92:871-880. doi:10.1097/PHM.0b013e3182a38e40
71. Türkbey T, Kutlay S, Gök H. Clinical feasibility of Xbox Kinect™ training for stroke rehabilitation: a single-blind randomized controlled pilot study. *J Rehabil Med*. 2017;49:22-29. doi:10.2340/16501977-2183
72. Yavuzer G, Senel A, Atay MB, Stam HJ. "Playstation eyetoy games" improve upper extremity-related motor functioning in subacute stroke: a randomized controlled clinical trial. *Eur J Phys Rehabil Med*. 2008;44:237-244.
73. Krakauer JW. Motor learning: its relevance to stroke recovery and neurorehabilitation. *Curr Opin Neurol*. 2006;19:84-90.
74. Nirme J, Duff A, Verschure PFMJ. Adaptive rehabilitation gaming system: on-line individualization of stroke rehabilitation. *Conf Proc IEEE Eng Med Biol Soc*. 2011;2011:6749-6752. doi:10.1109/IEMBS.2011.6091665
75. Arya KN, Verma R, Garg RK. Estimating the minimal clinically important difference of an upper extremity recovery measure in subacute stroke patients. *Top Stroke Rehabil*. 2011;18(suppl 1):599-610. doi:10.1310/tsr18s01-599
76. Page SJ, Fulk GD, Boyne P. Clinically important differences for the upper-extremity Fugl-Meyer scale in people with minimal to moderate impairment due to chronic stroke. *Phys Ther*. 2012;92:791-798. doi:10.2522/ptj.20110009