

SCIENTIFIC REPORTS



OPEN

Multiplexed temporally focused light shaping through a gradient index lens for precise in-depth optogenetic photostimulation

Nicolò Accanto^{1,2}, I-Wen Chen^{1,2}, Emiliano Ronzitti^{1,2}, Clément Molinier^{1,2}, Christophe Tourain^{1,2}, Eirini Papagiakoumou^{1,2} & Valentina Emiliani^{1,2}

In the past 10 years, the use of light has become irreplaceable for the optogenetic study and control of neurons and neural circuits. Optical techniques are however limited by scattering and can only see through a depth of few hundreds μm in living tissues. GRIN lens based micro-endoscopes represent a powerful solution to reach deeper regions. In this work we demonstrate that cutting edge optical methods for the precise photostimulation of multiple neurons in three dimensions can be performed through a GRIN lens. By spatio-temporally shaping a laser beam in the two-photon regime we project several tens of spatially confined targets in a volume of at least $100 \times 150 \times 300 \mu\text{m}^3$. We then apply such approach to the optogenetic stimulation of multiple neurons simultaneously *in vivo* in mice. Our work paves the way for an all-optical investigation of neural circuits in previously inaccessible brain areas.

Understanding communication among neurons and how they coordinate and integrate multiple signals is essential for discovering how neural circuits determine brain function and dysfunction. With a still growing toolbox of optogenetic reporters^{1,2} and actuators^{3,4}, and the parallel development of advanced optical techniques, it is today possible to activate and inhibit neuronal activity with light while optically recording the evoked response. Ultimately this will enable all-optical neural circuit interrogation with single-cell and single-action potential precision, even in deep brain regions⁵.

One-photon (1P) wide field illumination using single optical fibers, enables the simultaneous illumination of large brain regions up to few mm deep and has already permitted to establish the links between neural activity and behaviour in different areas^{6–11}. This light-delivery approach however lacks spatial selectivity. More sophisticated multi-point photonic probes^{12–14} achieved selective excitation of single brain layers but still lacked cellular resolution and did not permit concurrent optical recording of functional activity. By using 1P-computer generated holography (CGH) for patterned illumination through a fiber bundle coupled with a small objective, our group previously demonstrated single- and multi-neuron activation together with functional calcium imaging in freely moving mice¹⁵. However, the large dimension of the micro-objective ($<2.5 \text{ mm}$ diameter) limited this approach to superficial brain layers and the use of the fiber bundle did not preserve the holographic phase, thus constraining multi-spot generation to a single plane.

Two-photon (2P) excitation^{16,17} combined with wavefront shaping and opsin engineering allows millisecond manipulation of brain circuits with near-single cell resolution at multiple planes in three dimensions (3Ds). This was done either by generating multiple diffraction-limited spots that were scanned simultaneously across multiple cell somata^{18–20}, or by using computer-generated holography (CGH) to produce light patterns covering multiple cell somata at once²¹, thus optimizing the temporal precision of the photostimulation²². Recently, several research groups^{23–27} have shown that using a two-step wavefront shaping combined with temporal focusing (TF)^{28–31} it is possible to generate multiple high resolution extended light patterns in 3D, a technique we named multiplexed

¹Wavefront-Engineering Microscopy group, Neurophotonics Laboratory, CNRS UMR8250, Paris Descartes University, 45 rue des Saints-Pères, 75270, Paris 06, Cedex, France. ²Institut de la Vision, Sorbonne Université, Inserm S968, CNRS UMR7210, 17 Rue Moreau, 75012, Paris, France. Nicolò Accanto, I-Wen Chen and Emiliano Ronzitti contributed equally. Correspondence and requests for materials should be addressed to V.E. (email: valentina.emiliani@inserm.fr)




Figure 1. Schematics of the optical setup used in the experiment, comprising a 10 times beam expander, two SLMs, the diffraction grating (G) for TF, the appropriate lenses (L), the 10X, 0.3 NA air objective and the GRIN lens. For optical characterization the sample was a thin layer of rhodamine that was scanned in z for axial resolution measurements.

temporally focused light shaping (MTF-LS). These approaches led to the first demonstrations of neural circuit manipulation in 3D^{25,26}, yet the need of using conventional high numerical aperture (NA) objectives limited their use to circuits in superficial ($\leq 300\ \mu\text{m}$) cortical areas or to *in-vitro* applications.

Micro-endoscopes (MEs), i.e. small optical probes that can be inserted into living tissues, represent a promising solution to extend optical brain manipulation to deeper brain structures.

Most MEs are based on the use of gradient index (GRIN) lenses, which are small cylinders of glass of diameter $< 1\ \text{mm}$ and length of several centimetres, characterized by a gradual variation of refractive index in the radial direction³². In the last 15 years, GRIN lenses were studied and optimized for the 2P imaging of living tissues, e.g. implanted till a few mm deep inside the rodent brain^{33–41}. Recently⁴², 2P volumetric imaging through a GRIN lens was achieved by using extended Bessel beams. *In vivo* functional calcium imaging was demonstrated through GRIN lenses both in 1P⁴³ and in 2P excitation^{44–47}, in one case also combined with optogenetic point scanning activation⁴⁸. In ref.⁴⁴, 2P CGH was demonstrated through a GRIN lens (NA = 0.5, diameter 0.5 mm) and a GRIN objective (NA = 0.8, diameter 1 mm). Multiple diffraction-limited spots allowed the activity from several neurons to be imaged simultaneously through the GRIN lens. Extended light patterns were also generated through the GRIN objective ($10\ \mu\text{m}$ spots with axial resolution of $\sim 22\ \mu\text{m}$). However, CGH was limited to a single plane and extended holographic patterns with high axial resolution were only shown through the larger and more invasive GRIN objective.

In this work we demonstrate an optical system combining our recent MTF-LS technique²⁷ with the use of a GRIN lens (NA = 0.5, diameter 0.5 mm), and we show that a small ME is suitable for the generation of multiple axially-confined 2P excitation spots matching target somata in 3D. Successively, using this system we demonstrate all-optical control of single or multiple neurons through the ME, by performing *in vivo* 2P optogenetic stimulation of target cells and reading out their induced activity with 2P calcium imaging.

Results

Optical setup. The optical setup coupling the MTF-LS²⁷ to a GRIN lens ME is shown in Fig. 1. In the optical characterization experiments we used a femtosecond fiber laser (Fidelity 10, Coherent), emitting at 1040 nm. The MTF-LS setup consisted of (1) a first SLM that determined the size and shape of the spot(s) to be generated at the sample plane; (2) a diffraction grating for TF; (3) a second SLM that spatially multiplexed the axially confined spot(s) at arbitrary positions in 3D. The two SLMs were also used to correct for the aberrations of the system, in particular of the GRIN lens. To do so we maximized the 2P signal from a diffraction-limited spot made at the centre of the field of view (FOV) by introducing the appropriate first order Zernike aberration corrections. Aberration correction at coordinates different from the centre was not carried out in these experiments. We used an air objective (Olympus, UPLFLN 10 \times) with NA of 0.3 to focus into the GRIN lens (GRINTECH GmbH, model NEM-050-25-10-860-DM), which had a diameter of 0.5 mm and a total length of 9.86 mm, with NA of 0.19 on the objective side and of 0.5 on the sample side (in water), resulting in an internal magnification of 2.6 and a total field of view (FOV) of $\sim 150\ \mu\text{m}$ diameter⁴⁷. The combination of the 0.3 NA (10 \times) objective and GRIN lens resulted in a total magnification of 26⁴⁴. For the optical characterization, the spots produced using MTF-LS through the GRIN lens were focused onto a thin ($\sim 1\ \mu\text{m}$) spin-coated fluorescent layer of rhodamine-6G in




Figure 2. Spatial resolution of the MTF-CGH spots throughout the FOV of the GRIN lens. **(a)** Axial profiles (solid curves) and simulations (scattered curve) for a 15 μm holographic spot at the center of the FOV when using the GRIN lens with TF (red), without TF (green) and the 10X objective alone with TF (blue). **(b)** Position dependent FWHM of one holographic spot displaced along the TF (x) or the perpendicular (y) direction. The data are the average over 6 different repetitions of the same experiment and are given with an error bar calculated as the standard deviation over all the repetitions. **(c)** Lateral (x - z) view of 28 holographic spots distributed in 3D. **(d)** In plane (x - y) view of 3 holographic spots at 3 different depths. **(e)** Comparison of the axial resolutions with TF (28 spots) and without TF (16 spots) as a function of z . For the corresponding plots as a function of x and y see Supplementary Fig. 2.

polymethyl methacrylate 2% w/v in chloroform, and imaged with a second objective (40x-NA 0.8 objective LUM PLAN FI/IR, Olympus) in transmission on a CCD camera (CoolSNAP HQ2, Photometrics). To characterize the axial resolutions, the rhodamine sample was vertically scanned together with the imaging objective with a piezoelectric z scanner (PI N-725.2 A PIFOC). As discussed in ref.²⁷, the MTF-LS technique is compatible with several different beam-shaping strategies that one can perform with the first SLM. In the following we show the results of (i) multiplexed temporally focused computer-generated holography (MTF-CGH) and (ii) multiplexed temporally focused multi shapes (MTF-MS) through the GRIN lens.

MTF-CGH through the GRIN lens. In a first experiment we used the first SLM (see Fig. 1) to generate a simple holographic shape (a 15- μm diameter round spot) that was focused onto the diffraction grating for TF. The second SLM spatially multiplexed the axially confined spot at arbitrary positions in 3D. In Fig. 2a we compare the measured axial resolution (solid curves) of a 15- μm spot generated through the GRIN lens at the centre of the FOV with (red) and without (green) TF and of a 15- μm spot generated through the 0.3 NA air objective alone (with no GRIN lens) with TF (blue). The scattered curves in the same plot correspond to simulations for the three cases and are in agreement with the experimental results. We modelled the total system 0.3 NA objective + GRIN lens as a single water-immersion microscope objective with 0.5 NA and magnification of 26. As the green curves clearly suggests, when TF was not used, the axial resolution, calculated as the full width at half maximum (FWHM) of the curves, was excessively large ($\sim 80\mu\text{m}$ from experiments, $\sim 70\mu\text{m}$ from simulations), as a consequence of the relatively small NA of the GRIN lens. When using TF through the air objective alone, the axial resolution at the centre of the FOV improved to $\sim 29\mu\text{m}$ ($\sim 29\mu\text{m}$ in the simulations) and down to $\sim 20\mu\text{m}$ after adding the GRIN lens ($\sim 17\mu\text{m}$ in the simulations). The better axial resolution obtained when using the GRIN lens with respect to the 0.3 NA objective with TF mainly stems from the higher NA of the GRIN lens (0.5 NA in water, with respect to 0.3 NA in air for the objective).

In Fig. 2b we compare how the axial resolution of the holographic spots changed as a function of the lateral displacement both in the x (direction of dispersion of the diffraction grating) and in the y (perpendicular to the dispersion) direction (Fig. 2b). In Supplementary Fig. 1 we show similar data as a function of z . From the graph one sees that the axial resolution worsened in the TF direction (i.e. the x direction) when moving away from the centre much faster than for the y direction. This asymmetry, as detailed in the discussion, was not observed in previous experiments²⁷, and could be due to a loss of colours through the GRIN lens when displacing the

holographic spot in the TF direction. Taking this into account, as better shown in Supplementary Fig. 1, the effective FOV in the x - y plane, resulting in an axial resolution for holographic spots below $35\ \mu\text{m}$ is $\sim 100 \times 150\ \mu\text{m}^2$.

We next used the MTF-CGH system to produce 28 holographic spots in a large volume through the GRIN lens. Figures 2c,d show respectively a lateral projection (xz plane, for the yz projection see Supplementary Fig. 3) of the spots and the xy view at three different z positions, illustrating how different spots are focused at different depths. The plot in Fig. 2e compares the axial resolutions of the 28 temporally focused spots as a function of z (for similar plots as a function of x,y see Supplementary Fig. 2) with a similar experiment in which we generated a distribution of 16 spots without using TF. The difference between the two cases is striking; the mean axial resolution was $28 \pm 7\ \mu\text{m}$ in the former and $82 \pm 7\ \mu\text{m}$ in the latter case (the error was calculated as the standard deviation over all the experimental repetitions). From these results one can clearly see that advanced optical methods, here CGH and TF, can be implemented through a GRIN lens ME. Moreover, TF is essential to preserve high axial resolution when using low-NA GRIN lenses to create extended 2P excitation spots. The results of Fig. 2 and of Supplementary Fig. 1–3 demonstrate our ability of producing MTF-CGH spots in 3D on a FOV of $\sim 100 \times 150 \times 300\ \mu\text{m}^3$.

MTF-MS through the GRIN lens. As described in ref.²⁷, the MTF-MS method relies on a mixed phase/amplitude shaping approach, capable of simultaneously producing different speckle-free shapes, with an overall better axial resolution with respect to MTF-CGH, and to independently multiplex them at the sample plane. Another possibility to obtain speckle-free patterns could be to use the generalized phase contrast (GPC) method^{49,50}. As shown in ref.²⁷ however, MTF-MS gives the possibility to independently multiplex different shapes, and to optimize the illumination of the second SLM. Here we used MTF-MS in two different configurations (Fig. 3): in the first one we generated a mixture of round and stars (for a total of 20 spots) (Fig. 3a,b); in the second one we generated 16 identical round spots (Fig. 3c). In Fig. 3a we show the lateral view (x - z plane, for the yz projection see Supplementary Fig. 5) corresponding to the case of two different shapes whereas in Fig. 3b and c we show the projections along the z direction in the case of two and one shapes respectively. In both cases, we generated a 3D distribution of shapes with improved axial resolution and intensity homogeneity with respect to Fig. 2. As summarized in Fig. 3d, the axial resolution was around $22 \pm 6\ \mu\text{m}$ in the case of a single shape and $26 \pm 7\ \mu\text{m}$ when using two different shapes. As discussed later, contrary to previous results through a normal objective²⁷, when using the GRIN lens we could not easily generate more than two shapes, probably due to more important aberrations of the system.

***In vivo* setup.** We next used the GRIN lens based ME to perform concurrent holographic stimulation of single or multiple cells and optical readout of population activity *in vivo* in mice. As previously reported, a subset of neurons can be reliably co-labelled with opsins and calcium sensors by infecting cortical neurons of GCaMP6s transgenic mouse line GP4.3^{51,52} with ReaChR⁵³ viral vectors⁵⁴.

In the *in vivo* experiments, we mounted the GRIN lens ME on a microscope already equipped for 2P optogenetic stimulation and 2P scanning imaging, described in ref.⁵⁴. In this case, we used a more basic version of the TF-holographic setup shown in Fig. 1, comprising only one SLM and a diffraction grating, which could therefore only generate excitation spots on a single plane. In this setup, aberration correction was therefore only performed, at the centre of the FOV, using a single SLM. The laser sources were an amplified fiber laser (Satsuma HP, Amplitude Systemes) emitting at 1030 nm for the photostimulation and a tuneable Ti-Sapphire laser (Coherent Chameleon Vision II) for imaging. The imaging laser beam was raster scanned on the sample with xy galvanometric mirrors (6215 H series, Cambridge Technology) and the emitted fluorescence was collected back through the GRIN lens and objective, separated from the excitation path with a dichroic mirror, and directed to two photomultiplier tubes. In Supplementary Fig. 6 we show the comparison of the axial profile for a temporally focused holographic spot at the centre of the FOV generated with the setup used for *in vivo* experiments and the MTF-CGH setup, showing comparable performances.

After viral infection of ReaChR at the depths of $\sim 250\ \mu\text{m}$ and $\sim 500\ \mu\text{m}$ in the mouse primary visual cortex (V1), we performed acute craniotomy and removed the dura mater to expose the brain in anesthetized mice. All the *in vivo* experiments were performed in the anesthetized mouse V1 using the GRIN lens to image and photostimulate through the craniotomy. By moving the objective and the GRIN lens together with a z motor we could change the focus of the system from the brain surface down to $\sim 250\ \mu\text{m}$ deep (when the GRIN lens was in contact with the brain surface), possibly deeper if the GRIN lens slightly penetrated the brain.

All-optical control of neuronal activity through GRIN lens *in vivo*. Through the GRIN lens, we were able to resolve on average 34 ± 3 individual cortical neurons in a FOV of $150 \times 150\ \mu\text{m}^2$ (mean \pm s.e.m., 9 FOVs) at different depths (between 110 μm and 250 μm). As Fig. 4a shows, we observed GCaMP6s fluorescence changes in single (5 FOVs) or multiple (5.2 ± 0.7 cell, 5 FOVs) ReaChR-expressing target neurons by stimulating with 10 light-pulses at a repetition rate of 11.84 Hz, through 12- μm diameter circular holographic spots at a threshold illumination intensity (defined as the intensity resulting in ≥ 0.5 activation probability in target cells, see Methods) of $0.39 \pm 0.09\ \text{mW}/\mu\text{m}^2$. The imaging power was kept at $90 \pm 14\ \text{mW}$, using a scanning speed of 5.92 or 11.84 Hz. Higher imaging powers led to an increased GCaMP6s fluorescence in cytosol, probably due to the slow channel kinetics of ReaChR⁵³.

To examine the spatial selectivity of photostimulation, we determined the relationship between the activation probability (see methods) of target and non-target cells and their distance to the target cells⁵⁴. The red curves in the graphs of Fig. 4b are fits to the data with a single exponential decay function. In both cases, the probability of photo-activating a non-target cell decreases with increasing distance from the target cells. In the multi-cell experiment the decay rate was slower compared to that of single-cell activation. As explained later, out-of-target activation could result from a combination of neurite activation, smaller FOV and the relative low NA of the




Figure 3. Results obtained using the MTF-MS technique through the GRIN lens. **(a)** Lateral (x - z) view of 20 spots distributed in 3D when using the first SLM to produce two shapes. **(b,c)** Projections along the z direction in the case of two (rounds and stars) and one (rounds) shapes produced by the first SLM respectively. **(d)** Comparison of the axial resolutions for the two cases as a function of z . For similar plots as a function of x and y , see Supplementary Fig. 4.

GRIN lens. Of note, for 2 FOVs in multi-cell activation experiments, non-target cells displayed >0.9 activation probability at the threshold stimulation intensity up to $100\ \mu\text{m}$ away from a target cell, likely due to neuronal network activation, thus reducing the spatial selectivity of photoactivation.

Discussion

In this work we developed a ME, based on the combination of MTF-LS with a GRIN lens, for the 2P excitation of multiple targets in 3D and we combined it with functional imaging to enable cell-targeted all-optical interrogations of deep brain regions *in vivo*.

Using the TF technique and holographic spatial multiplexing we could generate 3D distributions of extended excitation spots over a total FOV of $\sim 100 \times 150 \times 300\ \mu\text{m}^3$, keeping at the same time an axial resolution better than $35\ \mu\text{m}$. Importantly, our results demonstrate that the TF effect is maintained through the GRIN lens, despite its larger aberrations⁵⁵. As shown in Fig. 2, this is crucial to keep a good axial confinement when generating large excitation spots, as the relatively small NA of the GRIN lens (0.5 on the sample side) would lead to a 3 times worse axial confinement if TF were not used therefore preventing targeted excitation. We showed that the GRIN lens is compatible with different MTF-LS techniques, such as MTF-MS. This latter configuration has the advantage of generating uniform speckle-free shapes with better axial resolution and gives the flexibility to separately multiplex different shapes. The increased light uniformity could be advantageous for parallel multi target imaging, while




Figure 4. All-optical control through the GRIN lens in mouse V1 *in vivo*. (a) Upper panels show a case of single-cell activation experiment at the depth of $\sim 110\mu\text{m}$ below the brain surface. The red circle, in the average projection profile of 2P image intensity from 2 channels, represents the target cell and the 5 white circles are examples of non-target cells. The corresponding calcium traces are displayed at the right, with red vertical bars denoting the photostimulation epochs. Lower panels show a case of multi-cell activation experiment with 6 target cells (red circles) and 6 examples of non-target cells (white circles) at the depth of $\sim 250\mu\text{m}$. Of note, the consecutive lines of bright pixels in the 2P images are photostimulation artefacts. Calcium signal in imaging frames from ROIs that were affected by stimulation artefacts is not shown. An additional case of *in vivo* photostimulation showing cellular selectivity is presented in Supplementary Fig. 7. (b) Color lines (except red) indicate activation probability of target and non-target cells in relation to the distance to target cell for 4 FOVs of single-cell activation (left) and 5 FOVs of multi-cell activation (right). Target cells' activation probabilities are plotted at Distance 0 with jitters for visualization. Legends denote the depth of each FOV. Red triangles with error bars are mean \pm s.e.m. Red solid lines represent the exponential fits of activation probability ($y = Ae^{-\frac{x}{k}}$ as fit; $A = 1.06$, $k = 27.89$, adjusted $R^2 = 0.97$ for single-cell activation and $A = 0.91$, $k = 126.10$, adjusted $R^2 = 0.81$ for multi-cell activation). The distance between non-target and target cell is distributed in 20- μm bins. For activation probability in single-cell activation as a function of both lateral (x) and vertical (y) directions see Supplementary Fig. 8.

the possibility of generating distinct shapes could be used in experiments requiring activating different cellular compartments or different neuronal populations with variable cell size.

Compared to a MTF-LS conventional microscope, the use of a GRIN lens imposes few limitations (Figs 2 and 3). First, we observed an asymmetric deterioration of the axial resolution when moving holographic spots in the x, y plane. Precisely, the axial resolution worsened more in the TF (x) than in its perpendicular direction (y) (Fig. 2b and Supplementary Fig. 1), a feature we did not observe when using a conventional high NA objective²⁷. As shown in Supplementary Fig. 1, this results in an elliptical FOV in the x, y plane extending to $\sim 100 \times 150 \mu\text{m}^2$.

One can think of the GRIN lens as a doublet, which collimates the light it receives at the input and focuses it back at the output. There is therefore inside the GRIN lens an effective Fourier plane, namely a plane at which the wavelength components (or colours) of the laser pulse are separated in space. Shifting the spots at the entrance of the GRIN lens results in some portion of the beam (at some intermediate plane inside the GRIN lens) being closer to the edges of the GRIN lens, which are more prone to effective loss of light and aberrations. Intuitively, one can think that moving the spots in the x direction away from the centre, results in some of the wavelength components at an intermediate plane inside the GRIN lens being closer to the edge of the GRIN lens, producing a net loss of colours and hence of axial resolution. Moving the spots in the y direction instead, causes a similar loss of light for all the colours, leading to a net decrease of light, but not of axial resolution.

The second limitation was in the maximum number of different shapes we could create when using MTF-MS. While using a conventional high NA objective allowed us to simultaneously generate up to 4 different shapes²⁷, using the GRIN lens we could only generate 2 different shapes. MTF-LS requires illuminating different portions of the second SLM with each shape, which in turns means entering into the microscope objective at different positions of its back entrance. This results in a tilted propagation at the sample plane for those shapes entering in the objective away from the centre, as also explained in²³. If this problem had negligible consequences in the case of a high NA objective²⁷, when using the GRIN lens it resulted in a loss of axial resolution and spot quality as the number of different shapes created by the first SLM increased. Also this second effect was probably due to the larger aberrations of the GRIN lens compared to a normal objective⁵⁵.

A deeper study of the aberrations, including a method to use the two SLMs to correct for aberrations at multiple planes and/or lateral positions (and not only at the centre of the FOV as demonstrated here), or the future availability of optimized GRIN lenses with lower aberrations could help to overcome these limitations. A promising possibility in this sense was recently shown in ref.⁵⁶, in which customized aspheric lenses were fabricated to extend the FOV of different types of GRIN lenses. A FOV as large as $\sim 400 \mu\text{m}$ for a singlet GRIN lens of 0.5 mm in diameter could in such a way be achieved. Coupling a similar strategy with MTF-LS is therefore highly promising for the photo-stimulation of tens of neurons in a large volume.

Low-NA GRIN lenses are currently the preferential choice for in-depth functional imaging in living animals as they permit minimal tissue damage compared to larger-diameter high-NA versions. We demonstrated the possibility to couple them with patterned illumination and temporal focusing enabling all-optical neuronal investigations *in vivo*. We were able to evoke calcium transients via soma-targeted illumination in single and multiple cells co-expressing GCaMP6s and ReaChR in the mouse visual cortex *in vivo*. Even if such study was limited to two dimensions (as one single SLM and 2D imaging were used), we were able to reliably activate target neurons at different z depths from the brain surface down to $\sim 250 \mu\text{m}$ deep, with the GRIN lens in actual contact with the brain surface (see Fig. 4 and Supplementary Fig. 7). This is consistent with previous studies^{57,58} in which it was shown that the transverse shape as well as the axial confinement of CGH and GPC spots with TF is substantially maintained across brain tissue up to $500 \mu\text{m}$ depth. To further support this, in Supplementary Fig. 9 we plot the simulated axial confinement for a holographic spot with TF at the center of the GRIN lens when travelling through different depths of scattering tissue. Indeed, the axial confinement is found to increase only from $17 \mu\text{m}$ to $18 \mu\text{m}$ through $300 \mu\text{m}$ of scattering tissue, thus confirming that the optical performances of our system are kept *in vivo* within the FOV of $100 \times 150 \times 300 \mu\text{m}^3$.

In Fig. 4b we studied the lateral spatial selectivity of the photostimulation *in vivo* for single and multi-cell holographic activation. Different factors, concerning both the preparation and the optical system, may result in a broadening of the lateral selectivity. On the one hand, as ReaChR expression was not confined to the soma, opsin-positive neurites crossing the photostimulation volume could result in accidental out-of-target activation. On the other hand, due to the overlap of the 2P absorption spectra of ReaChR and GCaMP6 as well as ReaChR's slow channel closing time, spurious activation during 2P scanning imaging could be induced⁵. Both such factors might be more relevant when using a GRIN lens, due to its relatively low NA and intrinsic aberrations, which result in an enlarged photostimulation volume (the average axial resolution is $28 \mu\text{m}$ with MTF-CGH) and make it necessary to use a higher imaging power to generate the same fluorescence signal compared to a higher NA objective. The combined use of a more performing GRIN lens, such as a better aberration corrected system with increased FOV, and of somatic and fast opsins^{59,60} will allow the current ME to be further improved for unbiased single-cell activation.

Finally, extending the use of MTF-LS to *in vivo* photostimulation with a volumetric imaging method, such as upstream divergence control^{20,21}, remote focusing⁶¹ or extended depth-of-field scan⁶² will enable to extend these results to multiple planes. The use of chronic implantation of GRIN lenses (as shown for example in^{44,47}) will also extend this approach to the all-optical manipulation of deeper brain regions.

Methods

Animal preparations. All animal experiments were performed in accordance with the Directive 2010/63/EU of the European Parliament and of the Council of 22 September 2010. The protocols were approved by the Paris Descartes Ethics Committee for Animal Research with the registered number CEEA34.EV.118.12. Mice were anesthetized with intraperitoneal injection of a ketamine-xylazine mixture (0.1 mg ketamine and 0.01 mg xylazine/g body weight) during viral injection and with isoflurane (2% for induction and 0.5–1% for

experiment) during photostimulation experiments. To express both opsin and calcium indicator in the same neuronal sub-population, stereotaxic injections of the viral vector AAV1-EF1 α -ReaChR-dTomato were performed in 4-week-old male or female mice of transgenic line GP4.3 (The Jackson Laboratory), which constitutively expresses the calcium indicator GCaMP6s⁵². Viral vectors of 1–1.5 μ L were infused at either 250 μ m or in combination with 500 μ m to target cortical neurons in the right primary visual cortex at a speed of 80–100 nL/min. Acute experiments of holographic stimulation were performed 6–12 weeks after viral infection.

Holographic stimulation and calcium imaging through GRIN lens *in vivo*. The *in vivo* endoscope system was integrated to an existing custom-made 2P all-optical system thoroughly described in⁵⁴. Specifically, a 0.5 NA GRIN lens was positioned by means of a custom-designed holder underneath a 10 \times microscope objective coupled with a SLM-based 2P patterned photostimulation and a 2P galvo-based scan imaging system. The GRIN lens holder was composed of kinematic mounts allowing X,Y,Z translation and rotation of the pitch and yaw axes of the GRIN lens, thus ensuring GRIN lens alignment to imaging and photostimulation paths. GRIN lens holder and objective were jointly connected to an axial motor allowing beam refocusing in the sample.

Initially, GRIN lens and objective were axially translated towards the head of the mouse till the image of brain surface appeared. That set the zero position of the GRIN lens, where its tip was at a distance of 250 μ m (i.e., the optical system's working distance) away from the brain surface. To perform all-optical experiments on a sub-population of cortical neurons co-expressing the opsin ReaChR and the calcium sensor GCaMP6s located up to 250 μ m below the brain surface (i.e. layer 2/3), we axially refocused the GRIN lens/objective system like with an ordinary 2P objective microscope until at the brain surface.

For acute *in vivo* experiments, mice were anesthetized with isoflurane (as described above) and the skull was exposed after subcutaneous application of lidocaine. A custom-made head-plate was attached to the skull using dental cement (Unifast Trad, GC). A circular craniotomy of 2–3 mm diameter was made over the injection site and the dura mater was removed.

Two-photon imaging of GCaMP6s calcium signal and the fluorophore labelling of dTomato was carried out by exciting with a scan beam provided by a Ti:Sapphire laser emitting at 920 nm (Coherent Chameleon Vision II). SLM-based patterned photostimulation was provided by using high-energy fiber laser-pulses at 1030 nm (Satsuma HP, Amplitude Systemes).

Target cells were selected based on a high-resolution reference image (512 \times 512 pixels) of the red channel collecting signal from ReaChR-dTomato. Through single or multiple circular holographic spots of 12- μ m diameter, the target somata were simultaneously illuminated with 10 holographic light-pulses of 5 or 10 ms at 11.84 Hz. Concurrently, the neuronal population activity of GCaMP6s calcium signal in a 150 \times 150 μ m² FOV was monitored by scanning at a frame rate of 5.92 or 11.84 Hz with 128 \times 128 pixel resolution.

During all-optical experiments, the photostimulation laser induced artefactual excitation of fluorophores, which appeared as consecutive lines of bright pixels. The image frames of GCaMP6s signal from regions-of-interest (ROIs) that were affected by stimulation artefacts were removed in post-processing.

Data analysis. Image analysis was performed by using ImageJ and MATLAB (Mathworks). ROIs covering individual target and non-target cell somata were manually selected in ImageJ based on both red (ReaChR-dTomato) and green (GCaMP6s) channels. The time-lapse fluorescent signal of GCaMP6s from all ROIs was exported in MATLAB. Image frames from ROIs that are affected by the photostimulation artefacts were removed in analysis. For each ROI, the relative percentage change of GCaMP6s fluorescence was computed as $\Delta F/F = (F - F_0)/F_0$, where F_0 was the average raw fluorescence signal 3–0.2 s before photostimulation started. A cell was considered activated if the average $\Delta F/F$ 1 s after the last illumination onset was significantly larger compared to that 1 s before the first illumination onset (right-tailed paired t-test with a significance level of 0.05). The relationship between activation probability and distance-to-target was determined similarly as previously described⁵⁴. Specifically, the threshold stimulation intensity was determined when the activation probability of target cells > 0.5 . The average values of activation probability at different distance-to-target were described by fitting with an exponential function.

Data Availability

The datasets generated during and/or analysed during the current study are available from the corresponding author on reasonable request.

References

- Knöpfel, T. Genetically encoded optical indicators for the analysis of neuronal circuits. *Nat. Rev. Neurosci.* **13**, 1–14 (2012).
- Lin, M. Z. & Schnitzer, M. J. Genetically encoded indicators of neuronal activity. *Nature Neuroscience* **19**, 1142–1153 (2016).
- Deisseroth, K. Optogenetics. *Nature Methods* **8**, 26–29 (2011).
- Deisseroth, K. Optogenetics: 10 years of microbial opsins in neuroscience. *Nature Neuroscience* **18**, 1213–1225 (2015).
- Emiliani, V., Cohen, A. E., Deisseroth, K. & Häusser, M. All-Optical Interrogation of Neural Circuits. *J. Neurosci.* **35**, 13917–13926 (2015).
- Jennings, J. H. *et al.* Distinct extended amygdala circuits for divergent motivational states. *Nature* **496**, 224–228 (2013).
- Zhang, F. *et al.* Optogenetic interrogation of neural circuits: Technology for probing mammalian brain structures. *Nature Protocols* **5**, 439–456 (2010).
- Sparta, D. R. *et al.* Construction of implantable optical fibers for long-term optogenetic manipulation of neural circuits. *Nat. Protoc.* **7**, 12–23 (2012).
- Jennings, J. H. *et al.* Visualizing hypothalamic network dynamics for appetitive and consummatory behaviors. *Cell* **160**, 516–527 (2015).
- Adamantidis, A. R., Zhang, F., Aravanis, A. M., Deisseroth, K. & De Lecea, L. Neural substrates of awakening probed with optogenetic control of hypocretin neurons. *Nature* **450**, 420–424 (2007).
- Aravanis, A. M. *et al.* An optical neural interface: *in vivo* control of rodent motor cortex with integrated fiberoptic and optogenetic technology. *J. Neural Eng.* **4**, (2007).

12. Pisanello, F. *et al.* Multipoint-emitting optical fibers for spatially addressable *in vivo* optogenetics. *Neuron* **82**, 1245–1254 (2014).
13. Zorzos, A. N., Scholvin, J., Boyden, E. S. & Fonstad, C. G. Three-dimensional multiwaveguide probe array for light delivery to distributed brain circuits. *Opt. Lett.* **37**, 4841 (2012).
14. Segev, E. *et al.* Patterned photostimulation via visible-wavelength photonic probes for deep brain optogenetics. *NeuroPhotonics* **4**, 011002 (2016).
15. Szabo, V., Ventalon, C., De Sars, V., Bradley, J. & Emiliani, V. Spatially Selective Holographic Photoactivation and Functional Fluorescence Imaging in Freely Behaving Mice with a Fiberscope. *Neuron* **84**, 1157–1169 (2014).
16. Helmchen, F. & Denk, W. Deep tissue two-photon microscopy. *Nature Methods* **2**, 932–940 (2005).
17. Grienberger, C. & Konnerth, A. Imaging Calcium in Neurons. *Neuron* **73**, 862–885 (2012).
18. Packer, A. M., Russell, L. E., Dagleish, H. W. P. & Häusser, M. Simultaneous all-optical manipulation and recording of neural circuit activity with cellular resolution *in vivo*. *Nat. Methods* **12**, 140–146 (2015).
19. Carrillo-Reid, L., Yang, W., Bando, Y., Peterka, D. S. & Yuste, R. Imprinting and recalling cortical ensembles. *Science (80-)*. **353**, 691–694 (2016).
20. Yang, W., Carrillo-Reid, L., Bando, Y., Peterka, D. S. & Yuste, R. Simultaneous two-photon imaging and two-photon optogenetics of cortical circuits in three dimensions. *Elife* **7**, e32671 (2018).
21. dal Maschio, M., Donovan, J. C., Helmbrecht, T. O. & Baier, H. Linking Neurons to Network Function and Behavior by Two-Photon Holographic Optogenetics and Volumetric Imaging. *Neuron* **94**, 774–789.e5 (2017).
22. Ronzitti, E. *et al.* Sub-millisecond optogenetic control of neuronal firing with two-photon holographic photoactivation of Chronos. *J. Neurosci.* **37**, 1246–17 (2017).
23. Hernandez, O. *et al.* Three-dimensional spatiotemporal focusing of holographic patterns. *Nat. Commun.* **7**, 11928 (2016).
24. Sun, B. *et al.* Four-dimensional light shaping: Manipulating ultrafast spatiotemporal foci in space and time. *Light Sci. Appl.* **7**, 17117 (2018).
25. Pégard, N. C. *et al.* Three-dimensional scanless holographic optogenetics with temporal focusing (3D-SHOT). *Nat. Commun.* **8**, 1228 (2017).
26. Mardinly, A. R. *et al.* Precise multimodal optical control of neural ensemble activity. *Nat. Neurosci.* **21**, 881–893 (2018).
27. Accanto, N. *et al.* Multiplexed temporally focused light shaping for high-resolution multi-cell targeting. *Optica* **5**, 1478 (2018).
28. Oron, D., Tal, E. & Silberberg, Y. Scanningless depth-resolved microscopy. *Opt. Express* **13**, 1468 (2005).
29. Zhu, G., van Howe, J., Durst, M., Zipfel, W. & Xu, C. Simultaneous spatial and temporal focusing of femtosecond pulses. *Opt. Express* **13**, 2153–2159 (2005).
30. Papagiakoumou, E., de Sars, V., Oron, D. & Emiliani, V. Patterned two-photon illumination by spatiotemporal shaping of ultrashort pulses. *Opt. Express* **16**, 22039–22047 (2008).
31. Andrasfalvy, B. K., Zemelman, B. V., Tang, J. & Vaziri, A. Two-photon single-cell optogenetic control of neuronal activity by sculpted light. *Proc. Natl. Acad. Sci.* **107**, 11981–11986 (2010).
32. Gomez-Reino, C., Perez, M. V. (Maria V. & Bao, C. (Carmen). *Gradient-Index Optics: Fundamentals and Applications*. (Springer Berlin Heidelberg 2002).
33. Jung, J. C. & Schnitzer, M. J. Multiphoton endoscopy. *Opt. Lett.* **28**, 902–904 (2003).
34. Levene, M. J., Dombbeck, D. A., Kasischke, K. A., Molloy, R. P. & Webb, W. W. *In Vivo* Multiphoton Microscopy of Deep Brain Tissue. *J. Neurophysiol.* **91**, 1908–1912 (2004).
35. Mehta, A. D., Jung, J. C., Flusberg, B. A. & Schnitzer, M. J. Fiber optic *in vivo* imaging in the mammalian nervous system. *Current Opinion in Neurobiology* **14**, 617–628 (2004).
36. Göbel, W., Kerr, J. N. D., Nimmerjahn, A. & Helmchen, F. Miniaturized two-photon microscope based on a flexible coherent fiber bundle and a gradient-index lens objective. *Opt. Lett.* **29**, 2521 (2004).
37. Fu, L. & Gu, M. Fibre-optic nonlinear optical microscopy and endoscopy. *Journal of Microscopy* **226**, 195–206 (2007).
38. Barretto, R. P. J., Messerschmidt, B. & Schnitzer, M. J. *In vivo* fluorescence imaging with high-resolution microlenses. *Nat. Methods* **6**, 511–512 (2009).
39. Murray, T. A. & Levene, M. J. Singlet gradient index lens for deep *in vivo* multiphoton microscopy. *J. Biomed. Opt.* **17**, 021106 (2012).
40. Attardo, A., Fitzgerald, J. E. & Schnitzer, M. J. Impermanence of dendritic spines in live adult CA1 hippocampus. *Nature* **523**, 592–596 (2015).
41. Miyamoto, D. & Murayama, M. The fiber-optic imaging and manipulation of neural activity during animal behavior. *Neuroscience Research* **103**, 1–9 (2016).
42. Meng, G. *et al.* High-throughput synapse-resolving two-photon fluorescence microendoscopy for deep-brain volumetric imaging *in vivo*. *Elife* **8** (2019).
43. Murayama, M., Pérez-Garci, E., Lüscher, H.-R. & Larkum, M. E. Fiberoptic System for Recording Dendritic Calcium Signals in Layer 5 Neocortical Pyramidal Cells in Freely Moving Rats. *J. Neurophysiol.* **98**, 1791–1805 (2007).
44. Moretti, C., Antonini, A., Bovetti, S., Liberale, C. & Fellin, T. Scanless functional imaging of hippocampal networks using patterned two-photon illumination through GRIN lenses. *Biomed. Opt. Express* **7**, 3958 (2016).
45. Andermann, M. L. *et al.* Chronic Cellular Imaging of Entire Cortical Columns in Awake Mice Using Microprisms. *Neuron* **80**, 900–913 (2013).
46. Sato, M. *et al.* Fast varifocal two-photon microendoscope for imaging neuronal activity in the deep brain. *Biomed. Opt. Express* **8**, 4049 (2017).
47. Bocarsly, M. E. *et al.* Minimally invasive microendoscopy system for *in vivo* functional imaging of deep nuclei in the mouse brain. *Biomed. Opt. Express* **6**, 4546 (2015).
48. Jennings, J. H. *et al.* Interacting neural ensembles in orbitofrontal cortex for social and feeding behaviour. *Nat.* **2019**, **1**, <https://doi.org/10.1038/s41586-018-0866-8> (2019).
49. Glückstad, J. Phase contrast image synthesis. *Opt. Commun.* **130**, 225–230 (1996).
50. Papagiakoumou, E. *et al.* Scanless two-photon excitation of channelrhodopsin-2. *Nat. Methods* **7**, 848–854 (2010).
51. Chen, T. W. *et al.* Ultrasensitive fluorescent proteins for imaging neuronal activity. *Nature* **499**, 295–300 (2013).
52. Dana, H. *et al.* Thy1-GCaMP6 Transgenic Mice for Neuronal Population Imaging *In Vivo*. *PLoS One* **9**, e108697 (2014).
53. Lin, J. Y., Knutsen, P. M., Muller, A., Kleinfeld, D. & Tsien, R. Y. ReaChR: A red-shifted variant of channelrhodopsin enables deep transcranial optogenetic excitation. *Nat. Neurosci.* **16**, 1499–1508 (2013).
54. Chen, I.-W. *et al.* *In Vivo* Submillisecond Two-Photon Optogenetics with Temporally Focused Patterned Light. *J. Neurosci.* **39**(18), 3484–3497 (2019).
55. Wang, C. & Ji, N. Characterization and improvement of three-dimensional imaging performance of GRIN-lens-based two-photon fluorescence endomicroscopes with adaptive optics. *Opt. Express* **21**, 27142–27154 (2013).
56. Antonini, A. *et al.* Extended field-of-view ultrathin microendoscopes with built-in aberration correction for high-resolution imaging with minimal invasiveness. *bioRxiv* 504472, <https://doi.org/10.1101/504472> (2018).
57. Papagiakoumou, E. *et al.* Functional patterned multiphoton excitation deep inside scattering tissue. *Nature Photonics* **7**, 274–278 (2013).
58. Bègue, A. *et al.* Two-photon excitation in scattering media by spatiotemporally shaped beams and their application in optogenetic stimulation. *Biomed. Opt. Express* **4**, 2869–2879 (2013).

59. Baker, C. A., Elyada, Y. M., Parra, A. & Bolton, M. M. Cellular resolution circuit mapping with temporal-focused excitation of somatargeted channelrhodopsin. *Elife* **5** (2016).
60. Shemesh, O. A. *et al.* Temporally precise single-cell-resolution optogenetics. *Nat. Neurosci.* **20**, 1796–1806 (2017).
61. Botcherby, E. J. *et al.* Aberration-free three-dimensional multiphoton imaging of neuronal activity at kHz rates. *Proc. Natl. Acad. Sci.* **109**, 2919–2924 (2012).
62. Lu, R. *et al.* Video-rate volumetric functional imaging of the brain at synaptic resolution. *Nat. Neurosci.* **20**, 620–628 (2017).

Acknowledgements

We thank Vincent de Sars for software developing, Coherent Inc. for the loan of the Fidelity laser. This research received support from the 'Agence Nationale de la Recherche' (grant ANR-15-CE19-0001-01, 3DHoloPac), the Human Frontiers Science Program (Grant RGP0015/2016), the Fondation Bettencourt Schueller (Prix Coups d'élan pour la recherche française), the Getty Lab, the National Institute of Health (Grant NIH 1UF1NS107574 - 01) and Axa research funding. N.A. received funding from the European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement no. 746173. E.R. received funding from the European Research Council SYNERGY Grant scheme (HELMHOLTZ, ERC Grant Agreement # 610110). IWC received funding from the European Union's Horizon 2020 research and innovation program under the Marie Skłodowska-Curie grant agreement no. 747598.

Author Contributions

N.A., I.W.C. and E.R. designed the experiments. N.A. and C.M. built the 3D holographic micro-endoscope system, performed the optical experiments and analysed the optical data. E.P. participated in the development of the optical system. I.W.C. performed animal surgery. E.R. and E.P. built the scanning imaging and 2D stimulation microscope. C.T. designed and built the GRIN lens holders to align the endoscope both for the optical and the *in vivo* setups. I.W.C., E.R. and N.A. performed *in vivo* experiments. I.W.C. analysed the *in vivo* data. N.A., I.W.C., E.R. and V.E. wrote the manuscript with the contribution of all the authors. V.E. conceived and supervised the project.

Additional Information

Supplementary information accompanies this paper at <https://doi.org/10.1038/s41598-019-43933-w>.

Competing Interests: The authors declare no competing interests.

Publisher's note: Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2019