

Validation of Monte Carlo dose calculation algorithm for CyberKnife multileaf collimator

Maude Gondré¹ | Fanny Marsolat¹ | Jean Bourhis² | François Bochud¹ |
Raphaël Moeckli¹ 

¹ Institute of Radiation Physics, Lausanne University Hospital and Lausanne University, Lausanne, Switzerland

² Radio-Oncology Department, Lausanne University Hospital and Lausanne University, Lausanne, Switzerland

Correspondence

Raphaël Moeckli, Institute of Radiation Physics, Lausanne University Hospital and Lausanne University, Rue du Grand Pré 1, 1007 Lausanne, Switzerland.
Email: Raphael.moeckli@chuv.ch

Abstract

Purpose: To commission and evaluate the Monte Carlo (MC) dose calculation algorithm for the CyberKnife equipped with a multileaf collimator (MLC).

Methods: We created a MC model for the MLC using an integrated module of the CyberKnife treatment planning software (TPS). Two parameters could be optimized: the maximum energy and the source full width at half-maximum (FWHM). The optimization was performed by minimizing the differences between the measured and the MC calculated tissue phantom ratios and profiles. MLC plans were calculated in the TPS with the MC algorithm and irradiated on different phantoms. The dose was measured using an A1SL ionization chamber and EBT3 Gafchromic films, and then compared to the TPS dose to obtain dose differences (ΔD). Finally, patient-specific quality assurances (QA) were performed with global gamma index criteria of 3%/1 mm.

Results: The maximum energy and source FWHM showing the best agreement with measurements were 6.4 MeV and 1.8 mm. The output factors calculated with these parameters gave an agreement within $\pm 1\%$ with measurements. The ΔD showed that MC model systematically underestimated the dose with an average of -1.5% over all configurations tested. For depths deeper than 12 cm, the ΔD increased, up to -3.0% (maximum at 15.5 cm depth).

Conclusions: The MC model for MLC of CyberKnife is clinically acceptable but underestimates the delivered dose by an average of -1.5% . Therefore, we recommend using the MC algorithm with the MLC only in heterogeneous regions and for shallow-seated tumors.

KEYWORDS

commissioning, CyberKnife, Monte Carlo dose algorithm, multileaf collimator

1 | INTRODUCTION

The CyberKnife (CK) is designed to deliver stereotactic radiosurgery (SRS) treatments, as well as stereotactic body radiation therapy (SBRT) using a robotic arm and real-time adaptive delivery. Three types of collimators are available. The fixed collimators provide fixed diameters of 7.5–60 mm, the Iris collimator provides variable circular apertures of 7.5–60 mm, and the

multileaf collimator (MLC) is for treating larger irregularly shaped lesions with a maximum treatment size of $115.0 \times 100.1 \text{ mm}^2$. Three dose calculation algorithms are available in the Precision TPS of CK. The RayTracing algorithm is available for the fixed and Iris collimators and uses effective path length to account for density variations. This algorithm ignores scatter changes due to local heterogeneities and thus, the calculated dose may result in inaccuracies near density interfaces.¹ This

This is an open access article under the terms of the [Creative Commons Attribution](https://creativecommons.org/licenses/by/4.0/) License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2021 The Authors. *Journal of Applied Clinical Medical Physics* published by Wiley Periodicals, LLC on behalf of The American Association of Physicists in Medicine

means that it overestimates the calculated dose in low-density tissue, such as the lungs, making the RayTracing algorithm clinically unsuitable for lung treatments. The Finite Size Pencil Beam (FSPB) algorithm is available for the MLC only. The beam is divided into small rectangular pencil beams and the dose distribution of each beam is calculated by the convolution of energy fluence and dose deposition kernel. The sum of all pencil beams contributions gives the dose at a given point.^{2,3} Because it uses effective path length to calculate doses, the FSPB algorithm is also not suited in case of heterogeneity, like the RayTracing algorithm. The Monte Carlo (MC) algorithm was first available for the fixed and Iris collimators and is now for the MLC. The MC algorithm accounts for lateral electronic scatter and lateral electronic disequilibrium. This makes the dose calculations more accurate, especially in a heterogeneous medium.⁴ Its accuracy has been evaluated for the fixed collimator and the difference between measured and MC calculated doses was within $\pm 3\%$.⁵ The objective of this work was to commission and validate the MC calculation algorithm for the MLC with a special emphasis on lung treatments.

2 | MATERIALS AND METHODS

2.1 | Beam modeling for MC algorithm

We used the Precision TPS version 2.0.1.1 to create the beam model for RayTracing (fixed and Iris collimators), FSPB (MLC), and MC (all collimators) algorithms. To create the MC model, we measured tissue phantom ratios (TPRs), profiles, and output factors for 11 MLC square field sizes ranging from 7.6×7.7 to 115.0×100.1 mm². The acquisition of these data (referred hereafter as “measurements”) was performed with a PTW 60018 stereotactic diode (PTW, Germany) in a water tank. Commissioning measurements of TPRs, profiles in the X and Y-directions, and output factors were acquired in 2015 in order to implement FSPB algorithm for the MLC. We were able to use the same measurements for implementing the MC model. However, at the time of these initial measurements, Accuray recommended the use of stereotactic diode, despite its inaccurate dose measurements in tails of profiles due to its nonwater equivalence. This constitutes a limitation of our study that will be discussed further below. The model was iteratively built by reducing the differences between the measured and the MC calculated TPRs and profiles. The maximum energy (E_{\max}), which represents the maximum value of the energy spectra, and the source size (S) defined by its full width at half-maximum (FWHM) were the two parameters that could be optimized. The TPRs were used to optimize E_{\max} and the profiles were used to optimize S . In fact, although S has a minor effect on the TPRs, E_{\max} has a consequent impact on profiles⁶ and therefore TPRs for all field sizes and depths should be calcu-

lated with the chosen E_{\max} in order to calculate profiles. This is a limitation because if evaluating the influence of different E_{\max} on profiles is needed, all TPRs should be calculated several times with different E_{\max} . The operation would need to be repeated for different S . This process would have been very time consuming and would not have been compatible with clinical time constraints, and this forced us to optimize E_{\max} and S independently.

2.1.1 | Determination of the optimized E_{\max}

Energy spectra that represent the distribution of initial photon energies and that are characterized by E_{\max} are pre-calculated in the TPS. They serve as an input to calculate the TPRs. Accuray recommended starting the optimization by calculating the TPRs for small (15.4×15.4 mm²) and large (84.6×84.7 mm²) field sizes, and for five depths (10, 15, 100, 200, and 300 mm).³ The MC calculation statistical uncertainty was set to 0.5% and S to 1.8 mm, as recommended by Accuray.³ We performed the optimization in two successive phases. *First phase*: TPRs were calculated with E_{\max} ranging from 6.3 to 6.8 MeV with intervals of 0.1 MeV for field sizes and depths defined above. Measured and MC calculated TPRs were compared using the metrics defined hereafter. *Second phase*: from the differences obtained between MC calculated and measured TPRs in the first phase, two E_{\max} (6.4 and 6.5 MeV) were considered as potentially optimal. These two E_{\max} were therefore used to calculate TPRs for all field sizes and depths available in the Precision TPS. Additionally, TPR differences for a reduced set of available field sizes and depths, considered as more clinically relevant, were also computed (field sizes from 7.6×7.7 mm² to 38.4×38.5 mm² and depths of 10, 15, 20, 50, 100, and 200 mm). For this second phase, we used the same metrics as for the first phase in order to define the optimal E_{\max} between 6.4 and 6.5 MeV. The metrics used to guide the discrimination of the optimal E_{\max} were (1) the mean difference between measured and MC calculated TPRs, and (2) the percentage of calculated points of the TPRs with a difference from measurements below $\pm 1\%$ ($\Delta\text{TPR}\%$). These two metrics were used as follows: if an E_{\max} gave the lowest mean difference as well as the highest $\Delta\text{TPR}\%$ among all tested E_{\max} , it was considered optimal. We preferred a higher $\Delta\text{TPR}\%$ so that if the mean difference obtained with an $E_{\max,1}$ was lower than with an $E_{\max,2}$, but the $\Delta\text{TPR}\%$ was higher with $E_{\max,2}$, then $E_{\max,2}$ was preferred to $E_{\max,1}$. We favored a higher $\Delta\text{TPR}\%$ because both the Swiss Society of Radiology and Medical Physics (SSRMP) recommendation for quality control of medical accelerators⁷ and the American Association of Physicists in Medicine (AAPM) guidelines on linear accelerator performance tests⁸ require a tolerance of 1% on TPRs dose differences (DDs).

Additionally, if the optimal E_{\max} was not the same between the field sizes $15.4 \times 15.4 \text{ mm}^2$ and $84.6 \times 84.7 \text{ mm}^2$, we chose the optimal E_{\max} for the lowest field size, because small field sizes are most predominantly used in CK treatments.

2.1.2 | Source FWHM optimization

The source distribution represents the distribution of photons' direction from the target. It is modeled as a Gaussian function³ and has one user input that is the photon source FWHM (S). From the results obtained using the method described in Section 2.1.1, an optimal E_{\max} was found and used to calculate the profiles. As recommended by Accuray, we calculated the profiles for two field sizes ($15.4 \times 15.4 \text{ mm}^2$ and $84.6 \times 84.7 \text{ mm}^2$) at 100 mm depth. As for the optimization of the E_{\max} , we performed the S optimization in two successive phases. *First phase*: the profiles were calculated for S of 1, 2, and 3 mm and compared to the measured ones with metrics defined hereafter. *Second phase*: from the first phase, two S were potentially optimal (1 and 2 mm). The profiles for S of 1.4, 1.6, and 1.8 mm were calculated and compared to the measured ones with the same metrics used in the first phase. The metrics used to compare the profiles were (1) the field size difference (measured at the FWHM of the profiles) obtained by the average of the FWHM of left–right and head–feet profiles, and (2) a gamma index (GI) between the measured and MC calculated profiles. The GI calculations were performed with a local 2% of maximum DD, 2 mm of maximum distance to agreement (DTA) and a dose threshold of 10% of maximum dose. These metrics were used as follows. If both the field size difference and the GI were better for one S , it was considered optimal. A higher GI was preferred so that if the field size difference obtained with a S_1 was lower than with a S_2 , but accompanied by a lower GI, then S_2 was considered as optimal compared to S_1 . Finally, for the same reason as for the E_{\max} optimization, the results obtained for small field size were dominant in the final decision.

2.2 | Model validation

2.2.1 | Output factors

After having optimized E_{\max} and S , we calculated the output factors for all field sizes with no possibility of further optimization and then compared them to the ones measured at source-axis distance (SAD) of 800 mm. The output factors were considered acceptable if the difference was lower than 1%, as recommended by the SSRMP⁷ and AAPM⁸ in their recommendations for the quality control of medical linear accelerators.

2.2.2 | Phantom measurements

We validated the MC model by comparing measured and MC calculated (with TPS) doses. The measurements were performed in four different phantoms (see Figure 1). Ionization chamber measurements (for phantoms shown in Figure 1a,c,d) were performed with an A1SL ionization chamber (Sun Nuclear, USA) metrologically traceable to the Federal Institute of metrology (METAS) and with a valid calibration for the MLC, and we used EBT3 GafChromic films (Ashland Inc., Wayne, NJ, USA) for the phantom shown in Figure 1b. For each measurement, the measured dose was corrected by the daily output variation of the CK. To calculate the dose distribution in these phantoms with the TPS, Accuray advised using the relative electron density (RED) of the phantom to set the mass density, because although the RED of the soft tissues in a human body are within 1% of their mass densities, plastic phantoms are not.⁹ The model validation consisted of six steps. From steps 1 to 5, the dose differences (ΔD) between the MC model calculation and the measurement were evaluated for different configurations in order to estimate the accuracy of the model. We fixed a maximum ΔD of $\pm 2\%$ to consider the model as clinically acceptable. However, a maximum ΔD of $\pm 1\%$ was expected to consider the MC model as really accurate. Step 6 consisted of calculating real treatment plans. The following sections describe each step.

Step 1: Single beam in homogeneous phantom (ionization chamber measurements)

The goal of this step was to compare the MC calculated and measured doses in a homogeneous phantom (Figure 1a), at the center of the beam and at 5 cm depth. For that purpose, we calculated nine plans with different equivalent square field sizes (different field sizes and/or shapes) ranging from 20.0 to 53.2 mm with a single beam (beam entrance normal to the phantom surface) in the TPS. These plans were then exported to the CK and delivered to the phantom. The dose was measured with a A1SL ionization chamber and compared to the TPS dose to obtain the dose difference ΔD . The closer the ΔD was to 0, the more accurate the MC algorithm. Additionally, to determine if the MC algorithm should be used for all clinical situations, even for homogeneous regions where the FSPB algorithm is available, we also measured the ΔD for four plans (among the nine previous plans) calculated with the FSPB algorithm and compared the dose differences with the ones obtained with the MC algorithm.

Step 2: Single beam in homogeneous phantom (film measurements)

EBT3 GafChromic films were used to verify the accuracy of the model in high-dose gradient regions. The




FIGURE 1 Phantoms used for validating the Monte Carlo (MC) model (a) homogeneous phantom with ionization chamber insert, (b) homogeneous phantom with films insert, (c) heterogeneous phantom with lung slabs and ionization chamber insert, and (d) homogeneous phantom with different possible depths of measurements with ionization chamber inserts

films were calibrated with an Elekta Synergy linear accelerator (Elekta AB, Stockholm, Sweden) against a farmer ionization chamber (Nuclear Enterprise, USA) with a 6 MV beam energy. The energy independence of the Gafchromic films¹⁰ made possible their use in the CyberKnife beam. The uncertainty related to film dosimetry is estimated to be $\pm 2\%$.¹⁰ We used five plans already calculated in step 1 (with equivalent square field sizes of 14.2, 20.0, 10.1, 30.8, and 17.6 mm) to obtain fields with dimensions (in one direction) of, respectively, 10.9, 16.5, 21.2, 30.9, and 45.0 mm. The dose profiles were measured with EBT3 Gafchromic films in the homogeneous phantom (Figure 1(b)) at 5.1 cm depth. A GI between the films and the MC plans was performed for each film. The calculations were performed with a local 2% of maximum DD, 2 mm of maximum DTA, and a dose threshold of 10% of the maximum dose.

Step 3: Single beam in heterogeneous phantom

This step followed the same principle as step 1, but we performed the calculations and the measurements using a heterogeneous phantom with a lung insert (Figure 1(c)). Nine plans were created with different equivalent square field sizes ranging from 20.4 to 45.2 mm with a single beam. The dose was measured at the center of the beam at 10 cm depth with A1SL ionization chamber.

Step 4: Multiple beams in homogenous and heterogeneous phantoms

Four plans were calculated in the homogeneous phantom (Figure 1(a)) with 6, 10, 12, and 21 beams with several different entry angles. The dose measurements were performed at the center of the beam and at 5 cm in the homogeneous phantom. The same procedure was applied using a heterogeneous phantom with 6, 10, 14, and 20 beams with several different entry angles. The dose measurements were performed at the center of the beam at 10 cm depth.

Step 5: Different depths in homogeneous phantom

A plan with multiple beams of field sizes larger than 28.5 mm equivalent field size was calculated in the phantom showed in Figure 1(d) and the ΔD was evaluated at 3.0, 6.0, 9.0, 12.0, 15.5, and 19.5 cm depth. The phantom used for those measurements was not specific to the CK. Therefore, we inserted a fiducial on the phantom to help the positioning, but a higher uncertainty was however observed due to the impossibility to correct for rotations. To mitigate this, we performed each measurement four times, with phantom repositioning between each measurement, and compared the mean difference to the MC calculated dose.

Step 6: Patient-specific QAs

Step 6 consisted on the creation of five MLC plans using the MC algorithm and the irradiation of patient-specific QAs following our routine procedure to accept a treatment plan. The patient-specific QA of each MLC plan was performed with the Octavius detector 1000 SRS (PTW, Germany) and a global GI of 3% of DD and 1 mm of DTA was applied. We assumed these QAs too be clinically acceptable if at least 95% of the points fulfilled the GI criteria.

3 | RESULTS

3.1 | Beam modeling for Monte Carlo algorithm

3.1.1 | Determination of the optimized E_{\max}

Table 1 presents the results obtained for the two metrics used to determine the optimal E_{\max} . According to these results, the optimal E_{\max} was 6.4 MeV.

TABLE 1 Mean difference (%) between measured and Monte Carlo (MC) calculated tissue phantom ratios (TPRs), and Δ TPR% (%) for different E_{\max} for different sets of field sizes and depths

		E_{\max} (MeV)	6.3	6.4	6.5	6.6	6.7	6.8
First phase	15.4×15.4 mm ² and depths of 10, 15, 100, 200 and 300 mm	Mean difference (%)	−0.3	−0.2	0.4	0.9	1.0	1.1
		Δ TPR% (%)	40	80	80	40	40	40
	84.6×84.7 mm ² and depths of 10, 15, 100, 200, and 300 mm	Mean difference (%)	0.1	−0.1	0.4	0.8	0.6	1.3
		Δ TPR% (%)	80	100	100	40	60	40
Second phase	All field sizes and depths	Mean difference (%)	−*	0.5	0.6	−	−	−
		Δ TPR% (%)	−	82	72	−	−	−
	Field sizes from 7.6×7.7 mm ² to 38.4×38.5 mm ² and depths of 10, 15, 20, 50, 100, 200 mm	Mean difference (%)	−	0.2	0.3	−	−	−
		Δ TPR% (%)	−	93	87	−	−	−

*Measurements for the second phase were only performed with 6.4 and 6.5 MeV.

TABLE 2 Field size difference and 2%/2 mm local gamma index (GI) comparison for 15.4×15.4 mm² and 84.6×84.7 mm² field sizes for S values of 1, 2, and 3 mm (first phase) and for S of 1.4, 1.6, and 1.8 mm (second phase)

	S (mm)	15.4 × 15.4 mm ² field size		84.6 × 84.7 mm ² field size	
		Field size difference (%)	GI (%)	Field size difference (%)	GI (%)
First phase	1	−0.1	76	−0.5	98
	2	0.2	76	−0.5	98
	3	0.2	71	−0.5	100
Second phase	1.4	0.0	76	−0.4	96
	1.6	0.2	81	−0.5	98
	1.8	0.0	81	−0.5	98

3.1.2 | Determination of the optimized source FWHM

Table 2 summarizes the different metrics used to compare the measured and MC calculated profiles. From those results, the optimized S was 1.8 mm. Large local DDs between the measured and the MC calculated profiles could be observed in the tail (overestimation of measured doses) and shoulder regions (underestimation of MC model), especially for large field sizes and depths. These differences became less significant for smaller field sizes and shallower depths, as seen in Figure 2.

3.2 | Model validation

3.2.1 | Output factors

Output factors were calculated with $E_{\max} = 6.4$ MeV and $S = 1.8$ mm and showed an agreement within $\pm 0.5\%$ when compared to the measured ones.

3.2.2 | Phantom measurements

Step 1: Single beam in homogeneous phantom (ionization chamber measurements)

The ΔD obtained fell between -2.2% and -1.3% . The average ΔD was -1.7% with a standard deviation of 0.3% . For plans calculated with FSPB algorithm, the ΔD fell between -1.3% and 0.4% , with an average of -0.8% .

Step 2: Single beam in homogeneous phantom (film measurements)

Figure 3 presents the profiles obtained with films (markers) compared to the ones calculated with the TPS (filled lines) for 16.5 mm and 21.2 mm field sizes. Profiles for field sizes of 10.9, 30.9, and 45.0 mm can be found in Figure S1. For all field sizes, the calculated dose in the plateau region (or on-axis for smallest field sizes) were within the $\pm 2\%$ of the film uncertainty, with a systematic under-estimation (except for the 10.9 mm field size) in the MC profiles. For all profiles, the correspondence on the tails and the gradients were correct. For the field sizes of 10.9 mm and 16.5 mm, the MC profiles were




FIGURE 2 Monte Carlo (MC) calculated profiles (dashed lines) and measured profiles (filled lines) for (a) $7.6 \times 7.7 \text{ mm}^2$ field size at 15 mm depth and (b) $115 \times 100.1 \text{ mm}^2$ field size at 300 mm depth




FIGURE 3 Profiles measured with EBT3 Gafchromic films (markers with error bars representing $\pm 2\%$ uncertainty) and calculated with the Monte Carlo (MC) model (filled line) for field sizes of (a) 16.5 mm and (b) 21.2 mm

slightly narrower than the measured ones; this was the contrary for the 21.2 mm field size.

Step 3: Single beam in heterogeneous phantom

The ΔD obtained for measurements with single beam in heterogeneous phantom fell between -1.7% and -0.8% , with an average of -1.2% and a standard deviation of 0.3% .

Step 4: Multiple beams in homogenous and heterogeneous phantoms

In the homogeneous phantom, the ΔD was between -1.8% and -1.2% with an average of -1.6% and a standard deviation of 0.3% . In the heterogeneous phantom, the average ΔD was -0.7% with a standard deviation of 0.5% . The differences were comprised between -1.3% and -0.1% . Three out of four measurements had a ΔD less than 1% .

TABLE 3 ΔD between measurement and TPS for different depths in homogeneous phantoms

		ΔD (%)
Depths (cm)	3.0	-1.7 ± 0.6
	6.0	-1.8 ± 0.5
	9.0	-1.7 ± 0.3
	12.0	-1.8 ± 0.2
	15.5	-3.0 ± 0.4
	19.5	-2.3 ± 0.3

Step 5: Different depths in homogeneous phantom

Table 3 shows the ΔD obtained at different depths of measurement. The average ΔD was -2.3% with a maximum difference of -3.0% at 15.5 cm depth.

Step 6: Re-calculation of patient treatment plans of lung tumors

All the patient-specific QAs performed passed our clinical criteria of acceptability with GI comprised between 95.5% and 100% with a global 1 mm/3% criteria. Four out of five QAs had GI above 98%.

4 | DISCUSSION

4.1 | Beam modeling for Monte Carlo algorithm

The two metrics chosen for E_{\max} optimization and our two-step process enabled us to unambiguously define the optimal E_{\max} . The other energies were discarded due to their higher $\Delta\text{TPR}\%$ or mean difference. However, even with the optimal E_{\max} , only 82% of the points (with all field sizes and depths) of the TPRs had a difference less than $\pm 1\%$, which suggested a lack of accuracy of the MC model. Optimizing the source FWHM was difficult due to two major problems. The first was caused by a lower dose than expected at the profiles' shoulders (Figure 2b), which may have been the result of a too simple source model and the absence of a particles' scatter model.³ It is a serious limitation of the MC model. Second, there were large discrepancies in the tails of the profiles due to the use of a diode for the measurements, which led to an overestimation of the measured dose.¹¹ In an addendum of 2017 in the Physics Essential Guide, Accuray introduced the use of synthetic microdiamond for profile measurements in order to reduce the overestimation observed with the diode. They advised either to be aware of profile overestimation in the tails of larger beams when diodes are used or to use a synthetic microdiamond.³ In our case, since a diode was used to measure profiles, the correct correlation at the tail regions was verified using EBT3 Gafchromic films when we validated the model (see Section 4.2). These differences, both in the tails and shoulders, were less significant when the field size and depth were reduced, as can be seen by comparing Figure 2a,b. Another limitation of the profile optimization was a result of the user-limited TPS module for the MC model, which prevented optimizing the profiles with both S and E_{\max} , although E_{\max} had an impact on profiles.⁶ Despite these limitations, the source FWHM obtained allowed us to achieve a good compromise on gradients for different field sizes (see Section 4.2).

4.2 | Model validation

For all field sizes, the output factors were within usual tolerances for a beam model validation.

Measurements made with a single beam in a homogeneous phantom led to a clinically acceptable ΔD ($< \pm 2\%$), although it was higher than expected for a

MC algorithm in such a simple configuration. Despite the correct output factors obtained with the MC model, we did obtain a systematic dose underestimation compared to the measurements. This result highlighted the MC model's lack of accuracy. This behavior has already been reported in a previous study, where a similar dose underestimation of 1%–2% was observed between the MC model and the measurements.¹² The ΔD obtained with the MC model was not affected by the equivalent square field sizes used. Therefore, we do not recommend any restrictions on the field sizes used for MLC plans. For plans calculated with the FSPB algorithm, the average ΔD was lower compared to those obtained with the MC algorithm. Therefore, it seems that the FSPB algorithm is more accurate in homogeneous situations compared to the MC algorithm. Clinically, this means that for treatments in homogeneous regions, the FSPB algorithm is preferred. During step 2, profiles were acquired with EBT3 Gafchromic films. For all field sizes, the correspondence on the tails agreed between the profiles obtained from Precision with the MC model and the measured ones. This result confirmed that the large discrepancies observed during MC modeling between measured and MC calculated profiles were due to the use of a diode for the measurements. By comparing the measured and MC calculated profiles, it seems that we found a good compromise for S between too narrow or too large MC profiles. The film dosimetry confirmed the underestimation of the MC model seen with ionization chamber measurements. There was an exception for the smaller field size of 10.9 mm, where the dose obtained was similar to the MC dose. The average ΔD obtained in the heterogeneous phantom with a single beam (step 3) was slightly better than that obtained with the homogeneous phantom. Therefore, it seems that the MC algorithm has a slightly better accuracy when used in heterogeneous conditions, which is quite surprising. As for the homogeneous phantom, the ΔD obtained was not dependent on the equivalent square field sizes used. In step 4, multiple beam plans were created on homogeneous and heterogeneous phantoms. In the homogeneous phantom, the average ΔD obtained with multiple beams was the same as for the single beam. Therefore, in a homogeneous situation, the number of beams used to create a plan does not influence the accuracy of the MC model. However, in the heterogeneous phantom, we obtained a lower ΔD with multiple beams. This behavior has already been seen in the previous work for the fixed collimator.⁵ Therefore, the use of multiple beams in heterogeneous conditions, which is the typical situation encountered in lung treatments, led to the higher accuracy of the MC algorithm. Our step 5 investigated the accuracy of the MC model regarding the depth of measurement. For depths up to 12.0 cm, the ΔD obtained were similar to steps 1 and 4. However, ΔD increased for deeper depths. Therefore, for the treatment of tumors located at a position deeper

than 12 cm, using the MC algorithm with MLC has to be used with caution. In step 6, five patient-specific QAs of MLC plans calculated with the MC algorithm were performed. All MLC plans were clinically acceptable in term of the gamma passing rate for patient-specific QA. Overall, the MC model could be created despite limitations related to the use of a diode, to a user-limited interface and to an overly simplistic model. These limitations probably led to a MC model that was less accurate than could be expected, but still clinically acceptable. Our results demonstrated a better accuracy in heterogeneous conditions with multiple beams, which is clinically the most representative situation for lung treatments. Therefore, we recommend the use of the MC model only in heterogeneous conditions and we advise users to be careful regarding the depth of the target. As mentioned before, although the MC model lacks accuracy, all patient-specific QAs largely fulfilled the criteria of acceptability, and therefore, the model is considered as clinically usable.

5 | CONCLUSION

We created a beam model for the MC algorithm used with the MLC in the Precision CyberKnife TPS. Our study results demonstrated important differences between the MC calculations and the measurements of TPRs and profiles. However, the different configurations we tested led to dose differences acceptable enough to consider the MC algorithm as clinically adequate, despite a lack of accuracy. All patient-specific QAs performed fulfilled the criteria applied for clinical acceptance of treatment plans. Therefore, we conclude that using the MC model with the MLC has a benefit but should be used carefully due to a lack of accuracy.

ACKNOWLEDGMENTS

We would like to thank Dr. Till Böhlen for his help in Python programming, Dr. Patrik Gonçalves Jorge for his advice about film dosimetry, Valery Taranenko from Accuray for guidance, and the entire medical physics team (Dr. Veronique Vallet, Michele Zeverino, Dr. Anais Viry, Dr. David Patin, Dr. Maud Marguet, and Natacha Ruiz Lopez) from CHUV for their advice and support.

CONFLICT OF INTEREST

The author declares that there is no conflict of interest that could be perceived as prejudicing the impartiality of the research reported.

AUTHOR CONTRIBUTIONS

All authors have made substantial contributions to conception and design, or acquisition of data, or analysis and interpretation of data, have been involved in drafting the manuscript or revising it critically for important intellectual content, have given final approval of the ver-

sion to be published, have participated sufficiently in the work to take public responsibility for appropriate portions of the content, and have agreed to be accountable for all aspects of the work in ensuring that questions related to the accuracy or integrity of any part of the work are appropriately investigated and resolved.

ORCID

Raphaël Moeckli 

<https://orcid.org/0000-0002-5885-934X>

REFERENCES

1. Lacornerie T, Lisbona A, Mirabel X, Lartigau E, Reynaert N. GTV-based prescription in SBRT for lung lesions using advanced dose calculation algorithms. *Radiat Oncol*. 2014;9:223.
2. Jelen U, Sohn M, Alber M. A finite size pencil beam for IMRT dose optimization. *Phys Med Biol*. 2005;50(8):1747-1766.
3. Accuray, CyberKnife Treatment Delivery System Physics Essentials Guide. 2018.
4. Murali V, Gopalakrishna Kurup PG, Bhuvanewari N, Sudahar H, Muthukumaran M. Monte Carlo and ray tracing algorithms in the cyberknife treatment planning for lung tumours—comparison and validation. *J Radiosurg SBRT*. 2013;2(2):85-98.
5. Pan Y, Yang R, Li J, Zhang X, Liu Lu, Wang J. Film-based dose validation of Monte Carlo algorithm for Cyberknife system with a CIRS thorax phantom. *Med Phys*. 2018;45(6):E325-E325.
6. Wagner A, Brou Boni K, Rault E, et al. Integration of the M6 Cyberknife in the Moderato Monte Carlo platform and prediction of beam parameters using machine learning. *Phys Med*. 2020;70:123-132.
7. Swiss Society of Radiobiology and Medical Physics. Quality control of medical accelerators. Recommendation No 11. 2014.
8. Smith K, Balter P, Duhon J, et al. AAPM Medical Physics Practice Guideline 8.a.: linear accelerator performance tests. *J Appl Clin Med Phys*. 2017;18(4):23-39.
9. Seco J, Evans PM. Assessing the effect of electron density in photon dose calculations. *Med Phys*. 2006;33(2):540-552.
10. Jaccard M, Petersson K, Buchillier T, et al. High dose-per-pulse electron beam dosimetry: usability and dose-rate independence of EBT3 Gafchromic films. *Med Phys*. 2017;44(2):725-735.
11. Francescon P, Kilby W, Noll JM, Masi L, Satariano N, Russo S. Monte Carlo simulated corrections for beam commissioning measurements with circular and MLC shaped fields on the CyberKnife M6 System: a study including diode, microchamber, point scintillator, and synthetic microdiamond detectors. *Phys Med Biol*. 2017;62(3):1076-1095.
12. Mackeprang PH, Vuong D, Volken W, et al. Benchmarking Monte-Carlo dose calculation for MLC CyberKnife treatments. *Radiat Oncol*. 2019;14(1):172.

SUPPORTING INFORMATION

Additional supporting information may be found in the online version of the article at the publisher's website.

How to cite this article: Gondré M, Marsolat F, Bourhis J, Bochud F, Moeckli R. Validation of Monte Carlo dose calculation algorithm for CyberKnife multileaf collimator. *J Appl Clin Med Phys*. 2022;23:e13481.
<https://doi.org/10.1002/acm2.13481>