Therapeutic Advances in Gastrointestinal Endoscopy

Safety of different electrocautery modes for endoscopic sphincterotomy: a Bayesian network meta-analysis

Abdellah Hedjoudje^(D), Chérifa Cheurfa, Jad Farha, Bénédicte Jaïs, Alain Aubert, Diane Lorenzo, Frédérique Maire, Dilhana Badurdeen, Vivek Kumbhari and Frédéric Prat

Abstract

Background and aims: Post-endoscopic retrograde cholangiopancreatography acute pancreatitis (PAP) and post-sphincterotomy hemorrhage are known adverse events of post-endoscopic retrograde cholangiopancreatography. Various electrosurgical currents can be used for endoscopic sphincterotomy. The extent to which this influences adverse events remains unclear. We assessed the comparative safety of different electrosurgical currents, through a Bayesian network meta-analysis of published studies merging direct and indirect comparison of trials.

Methods: We performed a Bayesian random-effects network meta-analysis of randomized controlled trials that compared the safety of different electrocautery modes for endoscopic sphincterotomy.

Results: Nine studies comparing four electrocautery modes (blended cut, pure cut, endocut, and pure cut followed by blended cut) with a combined enrollment of 1615 patients were included. The pooled results of the network meta-analysis did not show a significant difference in preventing post-sphincterotomy pancreatitis when comparing electrocautery modes. However, pure cut was associated with a statistically significant increased risk of bleeding compared with endocut [relative risk = 4.30; 95% confidence interval (1.53–12.87)]. On the other hand, the pooled results of the network meta-analysis showed no significant difference in prevention of bleeding when comparing blended cut *versus* endocut, pure cut followed by blended cut *versus* blended cut, pure cut followed by blended cut, and pure cut *versus* pure cut followed by blended cut. The results of rank probability found that endocut was most likely to be ranked the best.

Conclusion: No electrocautery mode was superior to another with regard to preventing PAP. Endocut was superior with respect to preventing bleeding. Therefore, we suggest performing endoscopic sphincterotomy with endocut.

Keywords: electrocautery, electrocoagulation, endoscopic sphincterotomy, endoscopic retrograde cholangiopancreatography, pancreatitis

Received: 12 January 2021; revised manuscript accepted: 10 November 2021.

Introduction

Since its first description in 1974, endoscopic biliary sphincterotomy has become the keystone of therapeutic endoscopic retrograde cholangiopancreatography (ERCP).¹ During this procedure, the sphincter of Oddi is dissected by cutting and/ or causing the coagulation of the tissue using a high-frequency electrosurgical current. Adverse events following sphincterotomy include pancreatitis, bleeding, cholangitis, and adverse cardiopulmonary events which occur in approximately 10% of patients prospective studies.^{2–6} In a large multicenter trial of more than 2000 patients undergoing biliary sphincterotomy, the incidence Ther Adv Gastrointest Endosc

2021, Vol. 14: 1–11 DOI: 10.1177/ 26317745211062983

© The Author(s), 2021. Article reuse guidelines: sagepub.com/journalspermissions

Correspondence to: Abdellah Hedjoudje

Service d'endoscopie digestive, DMU Digestif. Hôpital Beaujon, Assistance Publique des Hôpitaux de Paris, 100 boulevard du Général Leclerc, 92110 Clichy, France. Division of Gastroenterology. Johns Hopkins Medical Institutions, Baltimore, MD USA abdellah.hedioudie@ aphp.fr; Abdellah.hedjoudje@ gmail.com

Chérifa Cheurfa

Université de Paris, Epidemiology and Statistics, Sorbonne Paris Cité Research Center, (CRESS-UMR1153), INSERM, Cochrane France, Paris, France

Jad Farha

Dilhana Badurdeen Vivek Kumbhari Division of

Gastroenterology, Johns Hopkins Medical Institutions, Baltimore, MD, USA

Bénédicte Jaïs Alain Aubert Diane Lorenzo Frédérique Maire Frédéric Prat Service d'endoscopie digestive, DMU Digestif, Hôpital Beaujon, Assistance Publique des

Hôpitaux de Paris, Clichy,

France

journals.sagepub.com/home/cmg



Creative Commons Non Commercial CC BY-NC: This article is distributed under the terms of the Creative Commons Attribution-NonCommercial 4.0 License (https://creativecommons.org/licenses/by-nc/4.0/) which permits non-commercial use, reproduction and distribution of the work without further permission provided the original work is attributed as specified on the SAGE and Open Access pages (https://us.sagepub.com/en-us/nam/open-access-at-sage).

of pancreatitis and bleeding were 5.4% and 2%, respectively.³ Acute pancreatitis remains the most frequent adverse event associated with the procedures, despite innumerable efforts at preventing its development with drugs and/or pancreatic stents, but papillary bleeding also remains serious, particularly in patients under anticoagulant and/or antiplatelet agents.

The characteristics of the electrosurgical current can affect the risk of complications of endoscopic sphincterotomy (ES) because of the different nature of thermal tissue injury produced. A highvoltage, pure cutting current develops high tissue cutting ability and therefore achieves rapid and precise sphincter section, whereas a low-voltage coagulating current provides better hemostasis. A blend of both pure cutting and coagulating currents is available on high frequency (HF) generators designed for endoscopic procedures under the blended current label, while generator manufacturers, initially Erbe Medizin in the 1990s, later followed by others, developed new settings for sphincterotomy, polypectomy, and other techniques under the 'Endocut' label, which deliver sequential short bursts of pure or blended current followed by coagulation in a feedback loop between the electrical impedance of the tissue and the amount of energy delivered.7 Compared with the conventional blended mode, the endocut (EC) can provide with a precise and continuous degree of coagulation of the incision margins for hemostasis by its automatic voltage regulation and controlled cutting speed. This automatically fractionated cut aims to avoid perforation of the upper part of the papilla which can result from an uncontrolled cutting speed and may thus theoretically reduce complications of sphincterotomy.

There are, however, contradictory reports as to the respective risks of the development of acute pancreatitis and bleeding in multiple clinical trials comparing the pure cut (PC), EC, and the blended modes.^{7,8} Underpowerment and retrospective analysis have seriously limited the value of previous studies, making it difficult to gain definitive conclusions on the efficacy and safety of the different HF currents. To overcome those limitations, we made use of the direct and indirect evidence through a Bayesian network meta-analysis (NMA).

Methods

This systematic review is reported in accordance with the Preferred Reporting Items for Systematic Reviews and Meta-Analyses for Network Metaanalysis (PRISMA-NMA) guidelines.^{9–11}

Search strategy

A literature search was conducted in four databases including MEDLINE through PubMed, Embase, Cochrane Central Register of Controlled Trials (CENTRAL) in the Cochrane Library, and Web of Science from inception to April 2020 to identify potentially relevant studies. We did not limit searching by publication date or language. MEDLINE search strategy is detailed in Supplemental File 1. Search equation for the different databases was developed from PubMed search equation. We also evaluated the reference lists of the relevant clinical trials to identify additional studies.

Study selection

In a first step, titles and abstracts of articles were screened independently by two reviewers to exclude irrelevant articles. Then, the full text of all selected studies was screened according to predefined inclusion and exclusion criteria. Included studies were exclusively randomized controlled with the following criteria: (1) adult patients undergoing sphincterotomy, (2) intervention: sphincterotomy with different electrical modes, (3) comparator: another electrical current mode, and (4) outcome: pancreatitis or bleeding. We excluded (1) nonrandomized controlled studies, (2) trials not comparing different electrical modes such as prophylactic stent or nonsteroidal antiinflammatory drugs (NSAIDs), (3) case reports, meta-analyses, and nonrandomized trials, (4) language other than English, or (5) not involving human beings. Two reviewers (AH and CC) independently evaluated the studies for eligibility. Any disagreements were solved through discussion with the third investigator (FP).

Data extraction and quality assessment

Two reviewers (AH and CC) independently extracted data using a predefined extraction form. Extracted data were the following: the first author name, publication year, study period and design, and country. Patient's baseline characteristics were age and sex, indication for ERCP, and adverse events including pancreatitis, bleeding, cholangitis, or perforation. The qualities of randomized clinical trials were assessed using the risk of bias tool recommended by the Cochrane Collaboration on several dimensions: (1) random sequence generation; (2) allocation concealment; (3) blinding of participants and personnel; (4) blinding of outcome assessment; and (5) incomplete outcome data, as recommended by the Cochrane handbook.¹² Any disagreement in data extraction was resolved through consensus and discussion with a third reviewer (FP).

Statistical analysis

In the first step, multiple direct pairwise metaanalysis was performed using a fixed randomeffects model to estimate pooled relative risk (RR) and 95% confidence intervals (CIs).13 Statistical heterogeneity was assessed using I^2 statistic, with values less than 25% indicating no heterogeneity, between 25% and 50% indicating moderate heterogeneity, and more than 50% indicating substantial heterogeneity and the Cochran's Q test.¹⁴ In the second step, a random-effects Bayesian NMA using the Markov chain Monte Carlo (MCMC) methods was conducted. NMA allows the comparison of evidence from clinical studies where directly comparative data are not available. The posterior distribution of all parameters was estimated using informative priors to limit inference to data derived from the trials at hand. NMA also enables the ranking of treatments in order of efficacy. Electrocautery method mode was ranked in each iteration according to their preventive effect for pancreatitis or bleeding, respectively. We assessed the probability that each intervention was the most efficacious in decreasing the rate of pancreatitis or bleeding, the second best, the third best, and so on, by calculating the surface area under the curve (SUCRA) for each current mode compared with an arbitrary common control group. A node-splitting method was used to check network consistency by comparing direct and indirect estimates.¹⁵ Pairwise and network meta-analyses were conducted using R software (version 3.6.3; R Foundation for Statistical Computing, Vienna, Austria) with the meta, GeMTC, and BUGSnet packages.^{16,17} Finally publication bias was explored quantitatively with Egger's test and graphically with funnel plot.¹⁸

Results

Study selection

From a total of 50 unique studies identified using the search strategy, we included nine randomized controlled trials (RCTs)^{7,19–24} in this NMA. Seven studies were published article and two were published as abstracts.^{25,26} The detailed steps of the study selection process are shown in Figure 1. The reasons for the exclusion of studies during the final review were as follows: not randomized (n = 3) or review or meta-analysis (n = 4). Figure 2 displays the available direct comparisons and network of trials and shows the network plot of relevant studies. Pure cut (PC) mode was the most widely used with 533 patients followed by blended cut (BC) mode with 522 patients, Endocut (EC) with 403 patients, and pure cut followed by blended (PC/BC) with 157.

Characteristics and quality of included studies

Table 1 summarizes the main characteristics of the included trials. Overall, the nine trials had 1615 participants. Seven of them^{7,19-21,23-25} were two-armed studies and two were three-armed.22,26 The age of patients ranged from 55 to 73 years. The main indication for ERCP was cholelithiasis in 69.7% (1025/1471) of patients, sphincter of Oddi dysfunction (SOD) in 7.3% (108/1471), and other indications in the remaining 30% (338/1471). These trials were published between 1998 and 2015 with an enrollment period from June 1994 to May 2011. All studies were at high risk of bias for at least one domain of the Cochrane risk of bias tool (Supplemental File 2). Devices used and outcome definitions used by authors are summarized in Supplemental File 3.

Acute pancreatitis

Supplemental File 4 shows direct comparisons using a traditional pair-wise meta-analysis. There was no statistical difference between the different electrocautery modes in rates of post-sphincterotomy pancreatitis. The pooled results of the NMA showed no significant difference when comparing EC versus BC [RR = 1.24; 95% CI (0.41-4.21)] PC versus BC [RR = 0.48; 95% CI (0.18-1.09)], PC/BC versus BC [RR = 0.84; 95% CI (0.25-2.82)], PC versus EC [RR = 0.28; 95% CI (0.06-1.32)], PC/BC versus EC [RR = 0.68; 95% CI (0.13-3.18)], or PC/BC versus PC [RR = 1.76; 95% CI (0.64-5.73)].

Combined direct and indirect estimates from the Bayesian NMA are demonstrated in Figure 3 and Table 2. The results of rank probability showed that PC was most likely to be ranked the best (rank probability: 63%) (Supplemental File 5A).




Figure 1. Flow chart. PRISMA flow diagram. PRISMA, Preferred Reporting Items for Systematic Reviews and Meta-Analyses.

Bleeding

Supplemental File 6 shows direct comparisons using a traditional pair-wise meta-analysis. PC was associated with an increased risk of bleeding compared with EC [RR = 4.34; 95% CI (2.09–9.01)]. There was no statistically significant difference between the other electrocautery modes.

Combined direct and indirect estimates from the Bayesian NMA are demonstrated in Figure 3 and Table 2. PC was associated with an increased risk of bleeding compared with EC [RR = 4.29; 95% CI (1.53–12.87)]. However, the pooled results of the NMA showed no significant difference when comparing EC versus BC [RR = 0.44; 95% CI (0.34–3.95)], PC versus BC [RR = 1.92; 95% CI (0.71–2.91)], PC/BC versus BC [RR = 1.02; 95% CI (0.39–3.6)], PC/BC versus EC [RR = 2.72; 95% CI (0. 72–12.95)], or PC/BC versus PC [RR = 0.63; 95% CI (0.23–2.07)]. The results of

rank probability found EC was most likely to be ranked the best concerning risk of bleeding (rank probability: 98%), followed by BC, PC/BC, and PC (Supplemental File 5B).

Publication bias and network coherence. No evidence for publication bias was found, neither quantitatively based on visual inspection of funnel plots and quantitatively based on Egger's test, although the number of study for each comparison is small making these method less reliable. There were no significant differences between direct and indirect estimates when both were available, and the two methods had overlapping CIs for all interventions (Table 2 and Supplemental File 3).

Discussion

We hereby conducted the first NMA comparing the risks of adverse events associated with different




Figure 2. Network plot of relevant studies. Circles represent each electrocautery mode as a node and lines represent the direct comparisons. The extent of circle indicates the number of included participants in each electrocautery mode and the number on each blue line indicates the number of studies included in each comparison.

BC, blended cut; EC, endocut; PC, pure cut; PC/BC, pure cut followed by blended cut.




Figure 3. Forest plot of network estimate for the risk of post-ERCP acute pancreatitis with overall studies (a) and after excluding abstract (b). (c) Forest plot of network estimate for the risk of bleeding (c) and after excluding abstract (d). The summary effect estimate (risk ratio) is indicated by blue rectangles and lines representing 95% credible intervals.

BC, blended cut; EC, endocut; ERCP, endoscopic retrograde cholangiopancreatography; PC, pure cut; PC/BC, pure cut followed by blended cut.

First author,	Inclusion	Location	Setting	Current	Number of	Age, mean	Male	Indication for ES	S	
reference	period			mode	patients	(range)		Cholethiasis, N (%)	SOD, N [%]	Other, N [%]
Tanaka <i>et al.</i> ²⁰	June 2006 to May 2011	Japan	Multicenter	EC	179	73 (23–97) median	92	150 (83.8)	(0) 0	29 [16.2]
				BC	181	73 (30–101) median	108	143 [79]	(0) 0	38 (21)
Norton <i>et al.</i> ¹⁹	NR	The United	Multicenter	EC	133	59 [19–92]	70	63 (47.3)	7 (5.26)	73 (47.4)
		States		РС	134	59 (20–99)	67	62 (46.2)	5 (3.7)	67 (50.1)
MacIntosh <i>et al.</i> ²¹	1999–2002	The United States	Single center	РС	116	NR	NR	91 [78.4]	9 [7.7]	16 [13.8]
				BC	130	NR	NR	96 [73.8]	10 (7.7)	24 (18.5)
Stefanidis <i>et al.</i> ²²	October 1999 to	Greece	Single center	РС	62	62 (40–82)	35	62 [100]	0 (0)	0 (0)
	Uctober 2001			BC	62	66 [44–86]	36	62 [100]	(0) 0	0 (0)
				PC/BC	62	64 [42–84]	34	62 [100]	0 (0)	0 (0)
Gorelick <i>et al.</i> ²³	June 1997 to	The United	Single center	РС	75	NR	20	40 (53.3)	23 (30.6)	12 [16.1]
	November 1778	States		PC/BC	67	NR	27	39 [58.2]	18 [26.8]	10 (15)
Kohler <i>et al.</i> 7	June-November	Germany	Single center	EC	50	NR	NR	22 (44)	0 (0)	28 (56)
	1			BC	50	NR	NR	22 (44)	0 (0)	28 (56)
Elta <i>et al.</i> ²⁴	November 1994	The United	Single center	РС	86	51 [19–96]	30	55 (63.9)	18 (20.9)	13 (15.2)
	and June 1775	Slates		BC	84	55 [21–92]	30	56 [66.6]	18 (21.4)	13 (15.5)
Kida <i>et al.</i> ²⁵	From June 2002	Japan	Single center	EC	41	NR	NR	NR	NR	NR
(comerence proceeding)	10 2004			РС	43	NR	NR	NR	NR	NR
Mahadeva <i>et al.</i> ²⁶	November 1998	The United	Single center	РС	17	NR	NR	NR	NR	NR
(conterence proceeding)	to August 1777	Mingaom		BC	15	NR	NR	NR	NR	NR
				PC/BC	28	NR	NR	NR	NR	NR

Table 2. Network estimated	d risk ratios (RRs) of electro	Table 2. Network estimated risk ratios (RRs) of electrocautery modes on the risk of acute pancreatitis and risk of bleeding.	acute pancreatitis and risk o	ıf bleeding.	
Intervention		Acute pancreatitis		Bleeding	
		Direct estimate, RR [95% CI]	Network estimate, RR [95% CI]	Direct estimate, RR [95% CI]	Network estimate, RR [95% CI]
Blended					
Compared with	Endocut	0.72 [0.18–2.9]	0.81 [0.24–2.46]	2.54 [0.51–12.72]	2.25 [0.99-7.08]
	Pure cut	1.99 [0.65–6.06]	2.10 [0.92–5.69]	0.47 [0.34-0.64]	0.52 [0.26-1.41]
	PC/BC	1.04 [0.43-2.50]	1.19 [0.36–4.03]	1.12 [0.20-6.11]	0.83 [0.25–2.92]
Endocut					
Compared with	Blended	1.39 [0.34–5.59]	1.24 [0.41–4.21]	0.39 [0.08–1.97]	0.44 [0.14–1.01]
	Pure cut	3.59 [0.76-16.98]	2.62 [0.76–11.56]	0.35 [0.16-0.75]	0.23 [0.08-0.65]
	PC/BC	1	1.47 [0.31–7.83]	1	0.37 [0.08-1.38]
Pure cut					
Compared with	Blended	0.5 [0.16–1.54]	0.48 [0.18–1.09]	2.13 [1.57–2.91]	1.92 [0.71–2.91]
	Endocut	0.28 [0.06-1.32]	0.38 [0.09–1.32]	2.86 [1.33–6.14]	4.30 [1.53–12.87]
	PC/BC	0.58 [0.31-1.09]	0.57 [0.17–1.55]	1.59 [0.73–3.49]	1.58 [0.48-4.35]
PC/BC					
Compared with	Blended	1.04 [0.43-2.50]	0.84 [0.25–2.82]	0.90 (0.16–4.90	1.02 [0.34-3.95]
	Endocut	I	0.68 [0.13–3.18]	1	2.72 [0.72-12.95]
	Pure cut	1.73 [0.92–3.26]	1.76 [0.64–5.73]	0.63 [0.29–1.38]	0.63 [0.23-2.07]
Cl, confidence interval; PC/B(Risk ratios marked with bold	Cl, confidence interval; PC/BC, pure cut followed by blended cut; RR, relative risk. Risk ratios marked with bold indicate statistically significant differences between	cut; RR, relative risk. differences between the groups.			

journals.sagepub.com/home/cmg

A Hedjoudje, C Cheurfa et al.

electrical currents when performing sphincterotomy. Sphincterotomy during ERCP is a common procedure with significant risk of harm, the most feared being post-ES bleeding and PAP. The electrical current mode may be closely related with the occurrence of these adverse events as suggested for the first time by Sherman et al.27 in 1998. The papillary apparatus is a richly vascularized tissue with arterial branches of the retroduodenal artery, a gastroduodenal artery tributary. As such, there is a risk for severe hemorrhage when one of those is accidentally sectioned. HF current can diffuse inside the adjacent pancreatic tissue, hence the risk of pancreatic injury and systemic inflammatory response. A trade-off is essential because more coagulation may be advantageous to avoid bleeding and less may be better to avoid PAP. Several studies have been conducted to better understand the link between electrocautery and acute pancreatitis7,19-24 as well as the patient's and procedural characteristics associated with increased risk for post-ES adverse events. Risk factors for pancreatitis include suspected SOD, younger age, number of contrast injections into the pancreatic duct, precut sphincterotomy, and difficulty of cannulation.²⁷⁻²⁹ In close relationship with the latter, multiple cannulation attempts, perhaps related to excessive local trauma with subsequent tissue edema and transient pancreatic outflow obstruction, are also considered to increase the risk of PAP.³⁰ Theoretically, a pure cutting current is likely to induce less edema of the ampulla after sphincterotomy, which should decrease the risk of pancreatitis if one accepts the potentially greater risk of post-sphincterotomy hemorrhage. The theory of post-sphincterotomy pancreatitis focuses on the hypothesis of incidental temporary obstruction of the pancreatic duct, as the passage of electrical current through tissue and subsequent thermal injury cause a significant increase in the permeability of capillary beds, which in turn allows leakage of macromolecules and fluid into the extracellular space, causing localized edema which reduces the duct lumen. However and in contrast to the above quoted and well-established risk factors, the relative role of different electrocautery modes in triggering PAP remains contradictory. In a first meta-analysis conducted by Verma et al.31 in 2007, the type of current used for sphincterotomy (PC versus BC) did not influence the incidence or severity of pancreatitis. However, PC current was associated with a higher incidence of bleeding in this study. Subsequently, both the American Society for Gastrointestinal Endoscopy and the

European Society of Gastrointestinal Endoscopy recommended mixed current over PC in, respectively, their 2017 and 2019 biliary sphincterotomy guidelines.^{32,33}

In contrast, in an unpublished trial, Mitsuhiro Kida et al.25 concluded that the EC current tended to decrease the incidence of bleeding and hyperamylasemia after sphincterotomy, without clear evidence in favor of a decreased rate of clinically overt PAP. However, this trial has not been published and has therefore not been included in this NMA. Several further meta-analyses attempted to determine the superiority, if any, of one electrocautery mode over another.^{31,34,35} Li et al. compared EC with blended current modes and found that the rate of post-ERCP pancreatitis was not significantly different during ES. However, contrary to our meta-analysis, the authors chose to mix RCT and non-RCT.36 Another more recent meta-analysis conducted by Funari et al.34 compared all different electrocautery modes in a pairwise fashion, but did not make use of the indirect evidence available. Similar to our results, Funari et al. found that PC carries a higher risk of mild bleeding compared with other electrocautery modes.

By pooling results from both direct and indirect evidence, we found confirmative evidence that PC was associated with a statistically higher risk of bleeding compared with other electrocautery modes. Also, we found that EC presented the lowest risk of bleeding (SUCRA rank probability: 98%). We can also state from our analysis that the available evidence does not support any advantage of one electrocautery mode over another vis-à-vis the risk of PAP. Although we found that PC has the highest chance of being the best electrocautery mode regarding the risk of pancreatitis (SUCRA rank probability: 63%), difference with other electrocautery modes was not statistically significant. Basing one's assessment of efficacy on likelihood ranking can be misleading without an objective assessment of the strength of information in the network and the magnitude of absolute benefits that accompany rankings and can therefore be exaggerated, especially in the case of small differences in relative effects and limited information.11

We recognize our study has several limitations. First, if all patients analyzed were enrolled in an RCT, which can minimize the risk of bias, a total of 1471 patients remains a relatively small sample size. Second, the total number of ERCPs performed in the different institutions analyzed is also an important factor influencing the medical complication rates and this could not be taken into account from the available data. Inclusion period of the first trial started in June 1994, before the advent of EC and with outdated current generators, and prevention measures with nonsteroidal anti-inflammatory drugs or hyperhydration had not yet been developed at the time of the earliest trials. Third, all studies were at high risk of bias in at least one domain of the Cochrane risk of bias tool. Paucity and poor methodological quality of included studies may flaw overall results. Fourth, other preventive measures for acute pancreatitis such as the use of NSAIDs, intensive intravenous hydration, or prophylactic pancreatic stent were not reported. Fifth, bleeding definition was reported on a small number of study and definition was heterogenous. Authors did not distinguish between intraprocedural and delayed bleedings. Immediate bleeding might be just minor, nonsignificant adverse event, which can be self-limiting or managed easily, consequently clinically less significant compared with delayed bleedings, which can require hospital readmission, reintervention, and transfusion.

The main finding of our meta-analysis is that EC current is better at preventing bleeding than PC. Mild bleeding can be easily controlled in most patients, but not in high-risk patients that mandate resuming anticoagulant or antiplatelet agents post-ES. In light of our results, we recommend the use of EC because of its protective effect with regard to bleeding and because no other mode has demonstrated superiority in decreasing the risk of PAP. It is worth noting that the more recent HF generators offer the ability to fine tune the amount of coagulation current in EC modes (i.e. the socalled 'effect' in certain models), which can be a way to optimize the balance between hemostasis and current diffusion in pancreatic tissue and tailor this to the type of papilla and individual patient risk. Such possible improvements in generator settings deserve further study.

Author contributions

Abdellah Hedjoudje: Conceptualization; Data curation; Formal analysis; Methodology; Writingoriginal draft; Writing-review & editing. Chérifa Cheurfa: Conceptualization; Data curation; Formal analysis; Methodology. Jad Farha: Conceptualization; Data curation; Formal analysis; Methodology. **Bénédicte Jaïs:** Conceptualization; Data curation; Formal analysis; Investigation; Methodology; Writing-review & editing.

Alain Aubert: Conceptualization; Data curation; Formal analysis; Methodology; Writingreview & editing.

Diane Lorenzo: Conceptualization; Data curation; Formal analysis; Investigation; Methodology. **Frédérique Maire:** Conceptualization; Data curation; Formal analysis; Methodology.

Dilhana Badurdeen: Conceptualization; Data curation; Formal analysis; Investigation; Methodology.

Vivek Kumbhari: Conceptualization; Data curation; Formal analysis; Funding acquisition; Methodology; Writing-original draft; Writing-review & editing.

Frédéric Prat: Conceptualization; Data curation; Formal analysis; Investigation; Methodology; Supervision; Validation; Writing-review & editing.

Funding

The authors received no financial support for the research, authorship, and/or publication of this article.

Conflict of interest statement

The authors declared the following potential conflicts of interest with respect to the research, authorship, and/or publication of this article: VK is a consultant for Medtronic, Pentax Medical, Boston Scientific, FujiFilm, and Apollo Endosurgery. He receives research support from ERBE USA and Apollo Endosurgery. All other authors have no conflict of interest to disclose

Writing assistance

Nil writing assistance

ORCID iD

Abdellah Hedjoudje D https://orcid.org/0000-0002-6389-828X

Supplemental material

Supplemental material for this article is available online.

References

 Kawai K, Akasaka Y, Murakami K, et al. Endoscopic sphincterotomy of the ampulla of Vater. Gastrointest Endosc 1974; 20: 148–151.

- Cotton PB, Lehman G, Vennes J, et al. Endoscopic sphincterotomy complications and their management: an attempt at consensus. *Gastrointest Endosc* 1991; 37: 383–393.
- Freeman ML, Nelson DB, Sherman S, et al. Complications of endoscopic biliary sphincterotomy. N Engl J Med 1996; 335: 909–918.
- 4. Freeman ML. Understanding risk factors and avoiding complications with endoscopic retrograde cholangiopancreatography. *Curr Gastroenterol Rep* 2003; 5: 145–153.
- Lambert ME, Betts CD, Hill J, et al. Endoscopic sphincterotomy: the whole truth. Br J Surg 1991; 78: 473–476.
- Huibregtse K. Complications of endoscopic sphincterotomy and their prevention. N Engl J Med 1996; 335: 961–963.
- Kohler A, Maier M, Benz C, et al. A new HF current generator with automatically controlled system (endocut mode) for endoscopic sphincterotomy – preliminary experience. Endoscopy 1998; 30: 351–355.
- 8. Perini RF, Sadurski R, Cotton PB, *et al.* Postsphincterotomy bleeding after the introduction of microprocessor-controlled electrosurgery: does the new technology make the difference? *Gastrointest Endosc* 2005; 61: 53–57.
- 9. Hutton B, Salanti G, Chaimani A, *et al.* The quality of reporting methods and results in network meta-analyses: an overview of reviews and suggestions for improvement. *PLoS ONE* 2014; 9: e92508.
- Moher D, Liberati A, Tetzlaff J, et al. Preferred reporting items for systematic reviews and metaanalyses: the PRISMA statement. PLoS Med 2009; 6: e1000097.
- Hutton B, Salanti G, Caldwell DM, et al. The PRISMA extension statement for reporting of systematic reviews incorporating network metaanalyses of health care interventions: checklist and explanations. Ann Intern Med 2015; 162: 777–784.
- Cochrane handbook for systematic reviews of interventions, https://handbook-5-1.cochrane.org/ (2011, accessed 23 October 2019).
- DerSimonian R and Laird N. Meta-analysis in clinical trials. *Control Clin Trials* 1986; 7: 177–188.
- Higgins JPT, Thompson SG, Deeks JJ, et al. Measuring inconsistency in meta-analyses. BMJ 2003; 327: 557–560.

- van Valkenhoef G, Dias S, Ades AE, et al. Automated generation of node-splitting models for assessment of inconsistency in network meta-analysis. *Res Synth Methods* 2016; 7: 80–93.
- Shim SR, Kim S-J, Lee J, et al. Network metaanalysis: application and practice using R software. *Epidemiol Health* 2019; 41: e2019013.
- Béliveau A, Boyne DJ, Slater J, et al. BUGSnet: an R package to facilitate the conduct and reporting of Bayesian network Meta-analyses. BMC Med Res Methodol 2019; 19: 196.
- Sterne JAC, Sutton AJ, Ioannidis JPA, et al. Recommendations for examining and interpreting funnel plot asymmetry in meta-analyses of randomised controlled trials. *BMJ* 2011; 343: d4002.
- Norton ID, Petersen BT, Bosco J, et al. A randomized trial of endoscopic biliary sphincterotomy using pure-cut versus combined cut and coagulation waveforms. *Clin Gastroenterol Hepatol* 2005; 3: 1029–1033.
- 20. Tanaka Y, Sato K, Tsuchida H, *et al.* A prospective randomized controlled study of endoscopic sphincterotomy with the endocut mode or conventional blended cut mode. *J Clin Gastroenterol* 2015; 49: 127–131.
- 21. MacIntosh DG, Love J and Abraham NS. Endoscopic sphincterotomy by using pure-cut electrosurgical current and the risk of post-ERCP pancreatitis: a prospective randomized trial. *Gastrointest Endosc* 2004; 60: 551–556.
- 22. Stefanidis G, Karamanolis G, Viazis N, et al. A comparative study of postendoscopic sphincterotomy complications with various types of electrosurgical current in patients with choledocholithiasis. *Gastrointest Endosc* 2003; 57: 192–197.
- 23. Gorelick A, Cannon M, Barnett J, *et al.* First cut, then blend: an electrocautery technique affecting bleeding at sphincterotomy. *Endoscopy* 2001; 33: 976–980.
- 24. Elta GH, Barnett JL, Wille RT, *et al.* Pure cut electrocautery current for sphincterotomy causes less post- procedure pancreatitis than blended current. *Gastrointest Endosc* 1998; 47: 149–153.
- 25. Kida M, Kikuchi H, Araki M, *et al.* Randomized control trial of EST with either endocut mode or conventional pure cut mode. *Gastrointest Endosc* 2004; 59: P201.
- 26. Mahadeva S, Connelly J and Sahay P.
 Electrocautery in endoscopic sphincterotomy

 a randomised prospective trial comparing

combined current vs. cut or blend. *Gastrointest Endosc* 2000; 51: AB283.

- Sherman S, Ruffolo TA, Hawes RH, et al. Complications of endoscopic sphincterotomy. A prospective series with emphasis on the increased risk associated with sphincter of Oddi dysfunction and nondilated bile ducts. *Gastroenterology* 1991; 101: 1068–1075.
- Neoptolemos JP, Carr-Locke DL and Fossard DP. Prospective randomised study of preoperative endoscopic sphincterotomy versus surgery alone for common bile duct stones. Br Med J 1987; 294: 470–474.
- Masci E, Mariani A, Curioni S, *et al.* Risk factors for pancreatitis following endoscopic retrograde cholangiopancreatography: a meta-analysis. *Endoscopy* 2003; 35: 830–834.
- Testoni PA, Mariani A, Aabakken L, et al. Papillary cannulation and sphincterotomy techniques at ERCP: European Society of Gastrointestinal Endoscopy (ESGE) clinical guideline. *Endoscopy* 2016; 48: 657–683.
- 31. Verma D, Kapadia A and Adler DG. Pure versus mixed electrosurgical current for endoscopic

biliary sphincterotomy: a meta-analysis of adverse outcomes. *Gastrointest Endosc* 2007; 66: 283–290.

- Dumonceau J-M, Kapral C, Aabakken L, et al. ERCP-related adverse events: European Society of Gastrointestinal Endoscopy (ESGE) guideline. Endoscopy 2020; 52: 127–149.
- Chandrasekhara V, Khashab MA, Muthusamy VR, et al. Adverse events associated with ERCP. Gastrointest Endosc 2017; 85: 32–47.
- 34. Funari MP, Ribeiro IB, de Moura DTH, et al. Adverse events after biliary sphincterotomy: does the electric current mode make a difference? A systematic review and meta-analysis of randomized controlled trials. *Clin Res Hepatol Gastroenterol* 2020; 44: 739–752.
- Li DF, Yang MF, Chang X, et al. Endocut versus conventional blended electrosurgical current for endoscopic biliary sphincterotomy: a metaanalysis of complications. *Dig Dis Sci* 2019; 64: 2088–2094.
- Akiho H, Sumida Y, Akahoshi K, *et al.* Safety advantage of endocut mode over endoscopic sphincterotomy for choledocholithiasis. *World J Gastroenterol* 2006; 12: 2086–2088.

Visit SAGE journals online journals.sagepub.com/ home/cmg

SAGE journals