

MEETING ABSTRACT

Open Access

Automated body-lung-air material map segmentation from pre-clinical MRI images for PET attenuation correction in Tera-Tomo 3D PET reconstruction engine of nanoScan PET/MRI system

Peter Bandi^{1*}, Gabor Jakab¹, Norbert Zsoter¹, Domokos Mathe², Gabor Nemeth³, Kalman Nagy³, Sandor Hobor³, Laszlo Papp¹

From PSMR14: 3rd Conference in PET/MR and SPECT/MR
Kos Island, Greece. 19-21 May 2014

¹Department of Software Engineering, Mediso Ltd., Budapest, Hungary

The aim of current work was to automatically derive a body-lung-air three-valued material map from MRI images for pre-clinical PET attenuation correction. Our goal was also to measure the effect of attenuation on corresponding pre-clinical PET images.

12 mouse and 9 rat PET/MRI images were acquired by nanoScan PET/MRI (Mediso). Each MRI image was filtered with edge-preserving Non-Local Means filter.

Two segmentation stages were determined: the first stage separated the body and the air. The gravitational center of each disjunctive region determined by histogram-derived value ranges of the body and the air were inputs of two independent fuzzy-based affinity map generation steps. The fuzzy maps were merged with simple maximum into an initial binary body-air material map. In the second stage the gravitational center of disjunctive air regions laid inside the body mask were the seed points of the third fuzzy affinity map generation. Three fuzzy maps were merged with maximum operation into a three-valued mask.

The three-valued material map was given as input for the Tera-Tomo 3D PET reconstruction engine to serve as the base for attenuation correction.

Mouse and rat material map segmentation result were validated by two experienced pre-clinical researchers who manually drawn the ideal mask boundaries. The ratio of voxel misclassification was measured as 2.23% (mouse) and 1.87% (rat). Quantitative ROI measurements shown 1.3 - 2.7% AC and NAC PET differences.

An automated and robust segmentation method was designed to provide a material map for pre-clinical PET attenuation correction. Although current work operated with PET/MRI images, the same method can be applied for nanoScan SPECT/MRI acquisitions. Results show that on the level of real quantitative measurements, attenuation correction of PET makes sense. As the next step of our future work we will investigate the effects of AC on pre-clinical nanoScan SPECT/MRI images.

Authors' details

¹Department of Software Engineering, Mediso Ltd., Budapest, Hungary. ²Cromed Research Ltd, Budapest, Hungary.
³Department of Preclinical Development, Mediso Ltd., Budapest, Hungary.

Published: 29 July 2014

doi:10.1186/2197-7364-1-S1-A86

Cite this article as: Bandi *et al.*: Automated body-lung-air material map segmentation from pre-clinical MRI images for PET attenuation correction in Tera-Tomo 3D PET reconstruction engine of nanoScan PET/MRI system. *EJNM* Physics 2014 **1**(Suppl 1):A86.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- ▶ Convenient online submission
- ▶ Rigorous peer review
- ▶ Immediate publication on acceptance
- ▶ Open access: articles freely available online
- ▶ High visibility within the field
- ▶ Retaining the copyright to your article

Submit your next manuscript at ▶ springeropen.com
