Data in Brief 20 (2018) 503-509



Contents lists available at ScienceDirect

Data in Brief

journal homepage: www.elsevier.com/locate/dib

Data Article

Data set on Rapid Diagnostic Tests (RDTs) and microscopy for diagnosing plasmodium falciparum and plasmodium vivax



Grace I. Olasehinde^{*}, Uche C. Oyeka, Margaret I. Oniha, Olabode A. Onile-ere, Olayemi O. Ayepola, Adesola A. Ajayi, Louis O. Egwari

Department of Biological Sciences, Covenant University, Ota, Ogun State, Nigeria

ARTICLE INFO

Article history: Received 11 May 2018 Received in revised form 26 July 2018 Accepted 10 August 2018 Available online 15 August 2018

ABSTRACT

The World Heal Organization (WHO) has identified malaria diagnosis as being pivotal to eradicating the disease by 2030 as stipulated in the Sustainable Development Goals (SDG). The data presented here was obtained from outpatients of a hospital in the South Western Region of Nigeria from November 2016 to May 2017. The data contains malaria incidence amongst asymptomatic and symptomatic outpatients in the period under review. Malaria incidence was obtained using two diagnostic test kits, Bioline SD (HRP-2) and ACON (HRP-2/Aldolase) alongside Microscopy as gold standard. Specificity, Sensitivity and Kappa statistic of each test device is presented in the tables herewith. Data presented here could be used alongside other data sources to assess the state of malaria diagnostics.

© 2018 The Authors. Published by Elsevier Inc. This is an open access article under the CC BY license (http://creativecommons.org/licenses/by/4.0/).

* Corresponding author.

https://doi.org/10.1016/j.dib.2018.08.032

E-mail address: grace.olasehinde@covenantuniversity.edu.ng (G.I. Olasehinde).

^{2352-3409/© 2018} The Authors. Published by Elsevier Inc. This is an open access article under the CC BY license (http://creativecommons.org/licenses/by/4.0/).

| Subject area | Microbiology, Parasitology |
|----------------------------|--|
| More specific subject area | Malaria Diagnosis and Control |
| Type of data | Table |
| How data was acquired | Microscopy; Rapid Diagnostic Methods |
| Data format | Raw, Analyzed data in tables |
| Experimental factors | Ethical approval obtained from BIOSREC, Consent and accent sought |
| Experimental features | Blood samples obtained from febrile and non-febrile subjects visiting the selected health centers and tested for malaria using different |
| | methods |
| Data source location | Ado Odo/Ota, Ogun State, Nigeria |
| Data accessibility | Data is available in article |
| Related research article | Dozie and Chukwuocha [4] Comparative Evaluation of Malaria Rapid |
| | Diagnostic Test Kits Commercially Available in Parts of South Eastern |
| | Nigeria J Trop Dis 04(02) [4] |

Specifications Table

Value of the data

- Data provided here could inform the quality and choice of Rapid Diagnostic Test (RDT) in malaria endemic regions.
- Data presented here, when compared with data from other regions, could be used to measure the efficacy of RDTs vis-à-vis the malaria control agenda.
- Data provided may inform the development of cheap and non-invasive diagnostic method for malaria.

1. Data

Tables 1–4 present data from diagnosis using the RDT kits as well as microscopy among asymptomatic participants. The tables contain the counts for total number of participants alongside the number that tested positive. Tables 5–8 present results obtained for the RDTs and microscopy among participants who were symptomatic i.e. had a fever ≥ 37.5 °C. Tables 9–12 contains a summary assessment of both RDT kits tested, using microscopy as baseline).

Table 1

Incidence of P. falciparum using Bioline SD kits among asymptomatic carriers.

| Age group in years | No of sam | No of samples collected | | | No of positive samples on kit (BIOLINE SD) | | |
|--------------------|-----------|-------------------------|-----------|----------|--|---------|--|
| | Male | Female | Total | Male | Female | Total | |
| 0–5 | 35 | 43 | 78 | 9 | 13 | 22 | |
| 6-11 | 10 | 9 | 19 | 2 | 1 | 3 | |
| 12-17 | 6 | 11 | 17 | - | 3 | 3 | |
| 18-23 | 8 | 12 | 20 | 1 | 1 | 2 | |
| 24-29 | 12 | 19 | 31 | - | 2 | 2 | |
| 30 | 17 | 18 | 35 | 1 | 1 | 2 | |
| Total (%) | 88 (44) | 112 (66) | 200 (100) | 13 (6.5) | 21 (10.5) | 34 (17) | |

| Age group in years | No of samp | ples collected | | No of positive samples in kit (ACON) | | | |
|--------------------|------------|----------------|-----------|--------------------------------------|--------|--------|--|
| | Male | Female | Total | Male | female | Total | |
| 0–5 | 35 | 43 | 78 | 6 | 8 | 14 | |
| 6-11 | 10 | 9 | 19 | 3 | 3 | 6 | |
| 12-17 | 6 | 11 | 17 | - | 4 | 4 | |
| 18-23 | 8 | 12 | 20 | 2 | 2 | 4 | |
| 24-29 | 12 | 19 | 31 | 1 | 4 | 5 | |
| 30 | 17 | 18 | 35 | - | 3 | 3 | |
| Total (%) | 88 (44) | 112 (66) | 200 (100) | 12 (6) | 24(12) | 36(18) | |

 Table 2

 Incidence of P. falciparum using ACON kits among asymptomatic carriers.

 Table 3

 Incidence of P. falciparum using microscopic method among asymptomatic carriers.

| | No of sample | of samples collected | | | No of positive samples | | | |
|-----------|--------------|----------------------|-----------|---------|------------------------|-----------|--|--|
| | Male | Female | Total | Male | Female | Total | | |
| 0–5 | 35 | 43 | 78 | 12 | 16 | 28 | | |
| 6-11 | 10 | 9 | 19 | 2 | 3 | 5 | | |
| 12-17 | 6 | 11 | 17 | 4 | 3 | 7 | | |
| 18-23 | 8 | 12 | 20 | 3 | 4 | 7 | | |
| 24-29 | 12 | 19 | 31 | - | 4 | 4 | | |
| 30 | 17 | 18 | 35 | 1 | 1 | 2 | | |
| Total (%) | 88 (44) | 112 (66) | 200 (100) | 22 (11) | 31 (15.5) | 53 (26.5) | | |

Table 4

Comparative incidence rates of malaria among asymptomatic subjects using ACON, Bioline SD kits and microscopy.

| Sex | Number of sample collected (%) | Number of Positive samples (%) | | | | |
|--------|--------------------------------|--------------------------------|------------|----------|--|--|
| | | Microscopy | Bioline SD | ACON | | |
| Male | 88 (44) | 22 (25) | 13 (14.8) | 12(13.6) | | |
| Female | 112 (56) | 31 (27.7) | 21 (18.8) | 24(21.4) | | |
| Total | 200 (100) | 53 (26.5) | 34 (17) | 36(18) | | |

Table 5

Incidence of P. falciparum and P. vivax using ACON RDT kit among symptomatic subjects.

| Age group in years | Number of | samples colle | ected | Number of positive samples | | | | | |
|--------------------|-----------|---------------|----------|----------------------------|--------|----------|------|---------|-------|
| | | | | Male | | Female | | Total | |
| | Male | Female | Total | P.f | P.v | P.f | P.v | P.f | P.v |
| 0–5 | 1 | - | 1 | - | - | - | _ | - | _ |
| 6-10 | 1 | 1 | 2 | 1 | - | - | - | 1 | - |
| 11-15 | - | 4 | 4 | 1 | - | - | - | 1 | - |
| 16-20 | 11 | 16 | 27 | 7 | 2 | 9 | - | 16 | 2 |
| 21-25 | 3 | 4 | 7 | 2 | - | 1 | - | 3 | - |
| > 26 | 9 | 10 | 19 | - | - | 2 | - | 2 | - |
| Total | 25 (41.7) | 35 (58.3) | 60 (100) | 11 (47.8) | 2(100) | 12(52.2) | -(-) | 23(100) | 2(100 |

| Age | Number of sa | mples collected | | Number of positive samples | | | |
|-------------------|--------------|-----------------|---------|----------------------------|----------|--------|--|
| group in years | Male | Female | Total | Male | Female | Total | |
| 0–5 | 1 | _ | 1 | _ | _ | - | |
| 6-10 | 1 | 1 | 2 | 1 | - | 1 | |
| 11-15 | - | 4 | 4 | 1 | 1 | 2 | |
| 16-20 | 11 | 16 | 27 | 7 | 9 | 16 | |
| 21–25 | 3 | 4 | 7 | 2 | 1 | 3 | |
| >26 | 9 | 10 | 19 | - | 2 | 2 | |
| Total | 25(41.7) | 35(58.3) | 60(100) | 11 (45.8) | 13(54.2) | 24(100 | |

 Table 6

 Incidence of P. falciparum using Bioline SD kits among symptomatic subjects.

 Table 7

 Incidence of P. falciparum and P. vivax using microscopy among symptomatic subjects.

| Age group in years | Number of s | amples collected | | Number of | Number of Positive samples | | |
|--------------------|-------------|------------------|----------|-----------|----------------------------|----------|--|
| | Male | Female | Total | Male | Female | Total | |
| 0–5 | 1 | _ | 1 | 1 | _ | 1 | |
| 6-10 | 1 | 1 | 2 | 1 | - | 1 | |
| 11-15 | - | 4 | 4 | - | - | - | |
| 16-20 | 11 | 16 | 27 | 10 | 9 | 19 | |
| 21-25 | 3 | 4 | 7 | - | - | - | |
| >26 | 9 | 10 | 19 | 4 | - | 4 | |
| Total | 25(41.7) | 35(58.3) | 60 (100) | 16(64) | 9 (36) | 25 (100) | |

 Table 8

 Incidence of malaria using ACON kits, Bioline SD kits and microscopy among symptomatic subjects.

| Sex | Number of samples collected (%) | Number of Posit | | | |
|-------------------------|------------------------------------|----------------------------------|--------------------------------|-----------------------|---------------------------------|
| | | Microscopy (%) | RDTs | | |
| | | | ACON | | Bioline SD (HRP-2) (%) |
| | | | HRP-2 (%) | Pan-Aldolase (%) | |
| Male Female Total | 25 (41.7) 35 (58.3) 60 (100) | 16 (64) 9 (25.7) 25 (41.7) | 11 (44) 12 (34.3) 23(38) | 2 (8) - 2 (3.3) | 11 (44) 13 (37.1) 24 (40) |

2. Experimental design, materials, and methods

Data was obtained between November, 2016 and May, 2017 from 260 participants, 200 asymptomatic and 60 symptomatic subjects attending the University Health Centre.

Blood samples for analysis were obtained using either of two methods; direct sampling via finger prick or venous blood collected into EDTA bottles. Tests were performed using two RDT kits (ACON Malaria P.f/ Pan Rapid Test Device and SD BIOLINE Malaria Ag P.f test kits) [1-5]. Thick blood smears were prepared and stained with 10% Giemsa for 15 min to determine parasitemia which was estimated from the thick film by counting the number of parasites within 200 white blood cells (leukocyte) [6,7].

| | Symptomatic Cohort N=60 | | | | | Asympto | Asymptomatic Cohort N=200 | | | | | |
|--------------------|-------------------------|----------|---------------|---|----------------------|----------------|---------------------------|------------|---------------|--|----------------------|----------------|
| | Positive | Negative | Prevalence | Sensitivity (95% CI) | Fishers P value | к value | Positive | Negative | Prevalence | Sensitivity (95% CI) | Fishers P value | к value |
| SD Bioline ACON | 24 25 | 36 35 | 40% 41.67% | 96% (0.7965–0.990) 100% (0.8628–1.000) | < 0.0001 < 0.0001 | 0.966 1.000 | 34 36 | 166 164 | 64.15% 18% | 64.15% (0.4980–0.7686) 67.92% (0.5368–0.8008) | < 0.0001 < 0.0001 | 0.725 0.757 |
| Microscopy | | 35 | 41.67% | 100% (0.8028-1.000) | < 0.0001 | 1.000 | 53 | 147 | 26.5% | 07.52% (0.5508-0.0008) | < 0.0001 | 0.757 |

Table 9Performance of microscopy and RDTs across both cohorts.

Table 10

Test specifications.

| RDT | Specification |
|-----------------------------------|-----------------------------------|
| ACON Malaria Pf/Pan test kit | HRP-2 antigen Aldolase antigen |
| SD Bioline Malaria Ag Pf test Kit | HRP-2 Antigen |

Table 11

Contingency tables for symptomatic cohort.

| Microscopy | | |
|------------|--|--|
| | Positive | Negative |
| Positive | 19 | 5 |
| Negative | 6 | 30 |
| Microscopy | | |
| | Positive | Negative |
| Positive | 21 | 4 |
| Negative | 4 | 31 |
| | Positive Negative Microscopy Positive | Positive Positive 19 Negative 6 Microscopy Positive 21 |

Table 12

Contingency tables for asymptomatic cohort.

| | Microscopy | | |
|-------------------------|------------|----------|----------|
| SD Bioline Kit | | Positive | Negative |
| | Positive | 34 | 0 |
| | Negative | 19 | 147 |
| | Microscopy | | |
| ACON Malaria Pf/Pan Kit | | Positive | Negative |
| | Positive | 33 | 3 |
| | Negative | 20 | 144 |

The two-tailed Fisher's exact test (95% Confidence Interval) was used to check for significant differences in the sensitivities of the RDTs. Inter-test agreement for positive and negative results was expressed by the percentage of overall agreement. Kappa statistic (κ) was used to determine the agreement between malaria RDTs and the reference methods. κ -values 0.6–0.8 was considered as good while k-values > 0.8 were considered excellent.

Acknowledgments

The authors would like to appreciate the Covenant University Center for Research, Innovation and Discovery (CUCRID) (RG 009.11.14/FS) for its support.

Transparency document. Supporting information

Transparency data associated with this article can be found in the online version at https://doi.org/ 10.1016/j.dib.2018.08.032.

References

- C. Wongsrichanalai, M.J. Barcus, S. Muth, A. Sutamihardja, W.H. Wernsdorfer, A review of malaria diagnostic tools: microscopy and rapid diagnostic test (RDT), Am. J. Trop. Med Hyg. 77 (Suppl. 6) (2007) 119–127.
- [2] M.E. Parra, C.B. Evans, D.W. Taylor, Identification of Plasmodium falciparum histidine-rich protein 2 in the plasma of humans with malaria, J. Clin. Microbiol. 29 (8) (1991) 1629–1634.
- [3] WHO (2012) Malaria Rapid Diagnostic Test Performance. Summary results of WHO product testing of malaria RDTs: Round 1-5 (2008-2013) (Vol. 5). Geneva.
- [4] U. Dozie, U. Chukwuocha, Comparative evaluation of malaria rapid diagnostic test kits commercially available in parts of South Eastern Nigeria, J. Trop. Dis. 04 (02) (2016).
- [5] S. Sahu, K. Gunasekaran, P. Jambulingam, Field performance of malaria rapid diagnostic test for the detection of Plasmodium falciparum infection in Odisha State, India, Indian J. Med. Res. 142 (7) (2015) 52.
- [6] I.M. Ali, J.D. Bigoga, D.A. Forsah, F. Cho-Ngwa, V. Tchinda, V.A. Moor, , ...R.G.F. Leke, Field evaluation of the 22 rapid diagnostic tests for community management of malaria with artemisinin combination therapy in Cameroon, Malar. J. 15 (1) (2016) 31.
- [7] G.I. Olasehinde, D.O. Ojurongbe, O.J. Akinjogunla, L.O. Egwari, A.O. Adeyeba, Prevalence of malaria and predisposing factors to antimalarial drug resistance in Southwestern Nigeria, Res. J. Parasitol. 10 (2015) 92–101.