

ORAL PRESENTATION

Open Access

O066: Thermal disinfection of bedpans: European ISO 15883-3 guideline requirements are insufficient to ensure elimination of ARE and OXA-48 outbreak-strains

LB van der Velden^{1,2*}, MH Nabuurs-Franssen¹, A van Leeuwen¹, M Isken¹, A Voss^{1,2}

From 2nd International Conference on Prevention and Infection Control (ICPIC 2013)
Geneva, Switzerland. 25-28 June 2013

Introduction

During 2012, our hospital was faced with a vancomycin-resistant *Enterococcus faecium* (VRE) outbreak. Washer-disinfectors are used for the cleaning and thermal disinfection of bedpans. The European guideline NEN-ISO 15883-3 (ISO) states that washer-disinfectors have to achieve a minimum A_0 value of 60 for appropriate disinfection of bedpans ($\approx 80^\circ\text{C}$ for 60 seconds or 90°C for 6 seconds). However, previous data have shown that some *E. faecium* strains survive 60 seconds at 80°C . Following ISO, the A_0 measurement should include a cold (a minimum interval of 60 minutes since the machine was last used) and hot start.

Objectives

We determined the A_0 value of the VRE outbreak-strain and the outbreak-strain of the OXA-48 *K.pneumiae* outbreak in another Dutch hospital during 2011. Moreover the impact of a cold start measurement on the A_0 value was evaluated.

Methods

The minimum A_0 value that results in the killing of all isolates was determined for both strains. Bacterial suspensions were heated at 65, 75 and 80°C and samples for viable counts were obtained after 1,2,3 and 10 minutes at each temperature. VRE PCR and cultures were performed on bedpan swabs after disinfection; hot and cold start measurements were compared.

¹Medical Microbiology and Infectious Diseases, Canisius Wilhelmina Hospital Nijmegen, The Netherlands

Full list of author information is available at the end of the article

Results

Adequate killing required a minimum A_0 value of 180 for the VRE outbreak strain and 120 for the OXA-48 *K.pneumiae*. The cold start resulted in a 30% lower A_0 value, than the hot start. All washer-disinfectors in our hospital functioned in agreement with the European guideline, although the lowest A_0 value was 73, only just above 60. Swabs taken from bedpans processed in these washers with low A_0 values, were VRE-positive by PCR and cultures.

Conclusion

Both outbreak strains survived the A_0 value of 60 required in the ISO. VRE were identified by PCR as well as cultures of bedpans that had been disinfected by these washer-disinfectors. We suggest to increase the minimal acceptable A_0 value of washer-disinfectors to at least 180. Furthermore, the cold-start is needed for adequate A_0 -value measurement.

Disclosure of interest

None declared.

Author details

¹Medical Microbiology and Infectious Diseases, Canisius Wilhelmina Hospital Nijmegen, The Netherlands. ²Medical Microbiology, Radboud University Nijmegen Medical Center, Nijmegen, The Netherlands.

Published: 20 June 2013

doi:10.1186/2047-2994-2-S1-O66

Cite this article as: van der Velden et al.: O066: Thermal disinfection of bedpans: European ISO 15883-3 guideline requirements are insufficient to ensure elimination of ARE and OXA-48 outbreak-strains. *Antimicrobial Resistance and Infection Control* 2013 2(Suppl 1):O66.