








Research Article

Impact of Cold Weather on Setup Errors in Radiotherapy

Shuxue Zhao ¹, Xianfa Lu ¹, Jiasen Zou ², Zhouying Xu ¹, Siyu Wei ¹, Liping Li ¹, and Qingguo Fu ¹

¹Radiotherapy Center, Guangxi Medical University Cancer Hospital, Nanning, Guangxi 530021, China

²Pediatric Department, The First Affiliated Hospital of Guangxi Medical University, Nanning, Guangxi 530021, China

Correspondence should be addressed to Qingguo Fu; qingguofu@gxcahospital.cn

Received 8 September 2021; Revised 11 October 2021; Accepted 22 October 2021; Published 11 November 2021

Academic Editor: Kalidoss Rajakani

Copyright © 2021 Shuxue Zhao et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Objective. To investigate the influence of cold weather on setup errors of patients with chest and pelvic disease in radiotherapy. **Methods.** The image-guided data of the patients were collected from the Radiotherapy Center of Cancer Hospital Affiliated to Guangxi Medical University from October 2020 to February 2021. During this period, the cold weather days were December 15, 16, and 17, 2020, and January 7 and 8, 2021. For body fixation in radiotherapy, an integrated plate and a thermoplastic mold were employed in 18 patients with chest disease, while an integrated plate and a vacuum pad were applied in 19 patients with pelvic disease. All patients underwent cone beam computed tomography (CBCT) scans in the first five treatments and once a week thereafter. The obtained data were registered to the planning CT image to get the setup errors of the patient in the translational direction including X, Y, and Z axes and rotational direction including R_X , R_Y , and R_Z . Then, the Mann-Whitney U test was performed. The expansion boundary values of the chest and pelvis were calculated according to the formula $M_{PTV} = 2.5 \sum + 0.7\delta$. **Results.** A total of 286 eligible results of CBCT scans were collected. There were 138 chest CBCT scans, including 26 taken in cold weather and 112 in usual weather, and 148 pelvic CBCT scans, including 33 taken in cold weather and 115 in usual weather. The X-, Y-, and Z-axis translational setup errors of patients with chest disease in the cold weather group were 0.16 (0.06, 0.32) cm, 0.25 (0.17, 0.52) cm, and 0.35 (0.21, 0.47) cm, respectively, and those in the usual weather group were 0.14 (0.08, 0.29) cm, 0.23 (0.13, 0.37) cm, and 0.18 (0.1, 0.35) cm, respectively. The results indicated that there was a statistical difference in the Z-axis translational error between the cold weather group and the usual weather group ($U = 935.5$; $p = 0.005 < 0.05$), while there was no statistical difference in the rotational error between the two groups. The external boundary values of X, Y, and Z axes in the cold weather group were 0.57 cm, 0.92 cm, and 0.99 cm, respectively, and those in the usual weather group were 0.57 cm, 0.78 cm, and 0.68 cm, respectively. There was no significant difference in the translational and rotational errors of patients with pelvic disease between the cold weather group and the usual weather group ($p < 0.05$). The external boundary values of X, Y, and Z axes were 0.63 cm, 0.79 cm, and 0.68 cm in the cold weather group and 0.61 cm, 0.79 cm, and 0.61 cm in the usual weather group, respectively. **Conclusion.** The setup error of patients undergoing radiotherapy with their bodies fixed by an integrated plate and a thermoplastic mold was greater in cold weather than in usual weather, especially in the ventrodorsal direction.

1. Background

Roentgen first discovered X-ray in 1895 and Madame Curie discovered radium in 1898. After more than a hundred years of development, tumor radiotherapy has become one of the important means to treat malignant tumors. With the progress of science and technology, PET/CT, MRI, and other equipment have promoted the diagnosis of tumors and delineation of target areas to the molecular level. Radiotherapy planning has developed from the past forward

planning to reverse planning due to the rapid advances in computer technology. The radiotherapy implementation has developed from two-dimensional radiotherapy to three-dimensional conformal radiotherapy, intensity-modulated radiotherapy, and image-guided radiotherapy due to the development of materials science, engineering, imaging, and so on. The progress indicates that radiotherapy has entered the stage of precise diagnosis, precise planning, and precise implementation. With the arrival of the precise radiotherapy era, the transition in dose distribution between tumor target

area and normal tissues and organs needs to be very steep, because of the “win-win or lose-lose” relationship between tumor target area and normal tissues and organs in radiotherapy. There will be no benefit for any side if a tiny miscalculation occurs; thus, higher requirements for the setup in radiotherapy are put forward. Scholars have studied the effects of body weight [1–3], respiration [4–6], body position fixation, molds [7–9], registration mode between planning images and CT images [10–12], and whether to wear a mask during the epidemic of COVID-19 [13–15] on the setup repeatability. To our best knowledge, few studies considered the weather conditions. Therefore, this study aimed to investigate the influence of cold weather on setup errors in radiation therapy.

2. Materials and Methods

2.1. General Information. The clinical data of patients who underwent image-guided chest and pelvic radiotherapy in our radiotherapy center from October 2020 to February 2021 (December 15, 16, and 17, 2020; January 7 and 8, 2021) were collected. The data obtained on cold weather days (December 15, 16, and 17, 2020; January 7 and 8, 2021) and noncold weather days were assigned to the cold weather group and the usual weather group, respectively. There were 18 patients with chest disease, including 14 females and four males. Their ages ranged from 28 to 70 years with a mean age of 51.6 years. There were 19 patients with pelvic disease; they were all females, with an age range from 38 to 78 years and a mean age of 55.9 years. The detailed data are shown in Table 1.

2.2. Inclusion Criteria of Setup Error Data

- (i) The included data in the cold weather group should be the CBCT results on cold weather days.
- (ii) If the CBCT conducted in cold weather was the first five CBCT examinations of a patient and no setup error correction was ever implemented, all the CBCT data of the patient should be included. If setup error correction was conducted after the fifth time, only the data of the first five CBCT examinations were included.
- (iii) If the CBCT conducted in cold weather was not the first five CBCT examinations of a patient and no setup error correction was ever implemented, all the CBCT data of the patient should be included. If the setup error correction was conducted after the fifth time, the data of the first five CBCT examinations were excluded.

2.3. Molding and Positioning. Patients with chest disease were laid supine on an integrated plate (Klarity, China). The median sagittal plane of the body was vertical to the bed surface, and the patient’s upper limb was bent and lifted in a natural state to the hand support device. The thermoplastic mold (Create, China) that was able to fix the head, neck, and chest was used for fixation, as shown in Figure 1. Patients

with pelvic disease were placed in a supine position, and their bodies were fixed with an integrated plate (Klarity, China) and a vacuum pad (Klarity, China). The midline of the body was vertical to the bed surface, and the patient’s upper limb was bent and lifted in a natural state to the hand support device, with the toes forward and the feet together, as shown in Figure 2. The positioning of all patients with chest and pelvic disease was performed with a large-aperture dedicated CT simulator (GE LightSpeed™ RT spiral CT) produced by General Motors, USA, using a slice thickness of 5 mm. For patients with chest disease, the positioning mark point was delineated on the thermoplastic mold at the target area position using a pen; for patients with pelvic disease, the positioning mark point was made on the skin at the level of the tumor target area using a laser gun.

2.4. Planning and Design. The CT simulated positioning images obtained by planning design were transmitted to the planning system (Philips Pinnacle 9.10) through the DICOM local area network to develop the radiotherapy plan. The target area was delineated regarding the ICRU83 reporting standard. The gross tumor volume (GTV), clinical target volume (CTV), planning tumor volume (PTV), and organ at risk (OAR) were delineated. The physicist completed the design of the plan, and the superior physician and physicist jointly completed the evaluation of the plan. After completion of the radiotherapy treatment plan, it was transmitted to the MOSAIQ system, and the control accelerator (Elekta Synergy System) performed the treatment accordingly.

2.5. Image Guidance and Setup Error Data. All validations were performed by two experienced radiotherapists using Elekta X-ray volumetric imaging (XVI) software (version 4.5; Elekta Oncology Systems). Before each treatment, X-ray volumetric images were collected employing isocentric setups through airborne kilovoltage CBCT (XVI 4.5, Elekta Oncology Systems). Scanning parameters were 120 kV, collimator M20, filter F1, scanning angle 0°–180°, and clockwise direction. Each patient had a pretreatment CBCT scan in the first five treatments and weekly treatments thereafter. The treatment plan and planning CT images were transmitted to the XVI system through the network. In the obtained CBCT image, the three-dimensional volume containing the target area was set by selecting the corresponding three-dimensional matching frame. The setting range of the matching frame included the tumor and a few surrounding tissues, avoiding the tissues and organs with great movement changes including the heart as far as possible. Through the match between this three-dimensional volume and the corresponding three-dimensional volume in the planning CT, the three-dimensional setup error could be obtained. CBCT match for the patients with chest disease was performed using grayscale automatic matching and manual registration if necessary. CBCT match for patients with pelvic disease was performed using automatic bone matching and manual registration if necessary, and the setting range of the matching frame included the tumor and

TABLE 1: The collected data of the patients.

	Patients with chest disease	Patients with pelvic disease
Number of patients	18	19
Age	51.6 (28–70)	55.9 (38–79)
Gender	Male: 4 Female: 14	Male: 0 Female: 19
Tumor type	Breast cancer: 14 Esophageal cancer: 2 Lung cancer: 2	Uterine cancer: 7 Cervical cancer: 12
Fixation mode	Integrated plate + thermoplastic mold	Integrated plate + vacuum pad
Number of CBCT	138	148



FIGURE 1: Position for patients with thoracic malignancies.




FIGURE 2: Position for patients with pelvic tumors.

its nearby bone structures. The setup error data in cold weather and usual weather were obtained, including the values of left-right direction (X), head-foot direction (Y), anteroposterior direction (Z), rotational coronal position (R_X), sagittal position (R_Y), and transverse position (R_Z). The unit of translational direction was centimeter (cm), and the unit of rotational direction was degree ($^\circ$). The setup error of each patient was calculated according to the method described by Stroom et al. [16]. All data errors were derived from two aspects: system error and random error. The system error \sum (represented by the mean value of all setup errors) was mainly derived from the mechanical parameter errors of the simulated positioning machine and the treatment accelerator; the random error δ (represented by

the standard deviation of all setup errors) was mainly the difference in the repeatability of each setup during radiotherapy [17]. According to VanHerck's [18] study, the calculation formula of CTV to PTV expansion in X , Y , and Z directions was $M_{PTV} = 2.5 \sum + 0.7\delta$. The expansion range was calculated accordingly.

2.6. Statistical Analysis. Statistical analysis of the data was carried out using SPSS 17.0 statistical software. Since the setup error in all directions did not follow a normal distribution, it was expressed as median (25th percentile, 75th percentile), namely, $M (P25, P75)$. The Mann–Whitney U test was employed, and $P < 0.05$ was considered statistically significant.

3. Results

3.1. General Data. A total of 286 eligible results of CBCT scans were collected. There are 138 chest scans including 26 scans taken in cold weather and 112 in usual weather, and 148 pelvic cavity scans including 33 taken in cold weather and 115 in usual weather.

3.2. Comparison of Translational Setup Errors of Patients with Chest Disease in Different Groups. The translational and rotational setup errors of patients with chest disease in different groups are shown in Figure 3. The X -, Y -, and Z -axis translational setup errors of patients with chest disease in the cold weather group were 0.16 (0.06, 0.32) cm, 0.25 (0.17, 0.52) cm, and 0.35 (0.21, 0.47) cm, respectively. The X -, Y -, and Z -axis translational setup errors of patients with chest disease in the usual weather group were 0.14 (0.08, 0.29) cm, 0.23 (0.13, 0.37) cm, and 0.18 (0.1, 0.35) cm, respectively. There was a statistical difference in Z -axis translational error between the two groups ($U = 935.5$, $P = 0.005 < 0.05$). The P values of other directions between the two groups were all greater than 0.05, indicating no significant differences. The detailed data are shown in Table 2.

3.3. Comparison of Rotational Setup Errors of Patients with Chest Disease in Different Groups. The R_X , R_Y , and R_Z rotational setup errors of patients with chest disease in the cold weather group were $0.9 (0.38, 1.73)^\circ$, $0.75 (0.4, 1.2)^\circ$, and $0.75 (0.35, 1.1)^\circ$, respectively. The R_X , R_Y , and R_Z rotational setup




FIGURE 3: The translational and rotational setup errors of patients with chest disease.

TABLE 2: Comparison of translational errors between the two groups.

Group	Number of CBCT	Translational errors		
		X	Y	Z
Cold weather group	26	0.16 (0.06, 0.32)	0.25 (0.17, 0.52)	0.35 (0.21, 0.47)
Usual weather group	112	0.14 (0.08, 0.29)	0.23 (0.13, 0.37)	0.18 (0.1, 0.35)
U		1444.5	1264	935.5
P		0.95	0.296	0.005

There was a significant difference in Z-axis translational error between the two groups ($P < 0.05$).

errors in the usual weather group were $0.8 (0.4, 1.2)^\circ$, $0.7 (0.3, 1.4)^\circ$, and $0.7 (0.2, 1.1)^\circ$, respectively. There were no significant differences in rotational errors between the two groups ($P > 0.05$). The detailed data are shown in Table 3.

3.4. Comparison of Translational Setup Errors of Patients with Pelvic Disease in Different Groups. The translational and rotational setup errors of patients with pelvic disease in different groups are shown in Figure 4. The X-, Y-, and Z-axis translational errors of patients with pelvic disease in the cold weather group were $0.17 (0.09, 0.29)$ cm, $0.21 (0.1, 0.4)$ cm, and $0.17 (0.07, 0.31)$ cm, respectively. The X-, Y-, and Z-

axis translational errors in the usual weather group were $0.13 (0.06, 0.3)$ cm, $0.23 (0.11, 0.39)$ cm, and $0.22 (0.08, 0.32)$ cm, respectively. There were no significant differences in translational errors between the two groups. The detailed data are shown in Table 4.

3.5. Comparison of the Mean Rotational Setup Error of Patients with Pelvic Disease in Different Weather. The mean R_X , R_Y , and R_Z rotational setup errors of patients with pelvic disease were $1.1 (0.5, 1.8)^\circ$, $0.5 (0.2, 0.8)^\circ$, and $0.5 (0.3, 0.9)^\circ$ in the cold weather group and $0.8 (0.4, 1.55)^\circ$, $0.6 (0.25, 1.2)^\circ$, and $0.5 (0.2, 1.05)^\circ$ in the usual weather group, respectively. There were no

TABLE 3: Comparison of rotational errors between the two groups.

Group	<i>n</i>	Rotational errors		
		R_X	R_Y	R_Z
Cold weather group	26	0.9 (0.38, 1.73)	0.75 (0.4, 1.2)	0.75 (0.35, 1.1)
Usual weather group	112	0.8 (0.4, 1.2)	0.7 (0.3, 1.4)	0.7 (0.2, 1.1)
<i>U</i>		1378	1413	1357
<i>P</i>		0.671	0.815	0.589

There were no significant differences in rotational error between the two groups ($P > 0.05$).




FIGURE 4: The translational and rotational setup errors of patients with pelvic disease.

TABLE 4: Comparison of translational errors between the two groups.

Group	Number of CBCT	Translational errors		
		<i>X</i>	<i>Y</i>	<i>Z</i>
Cold weather group	33	0.17 (0.09, 0.29)	0.21 (0.10, 0.40)	0.17 (0.07, 0.31)
Usual weather group	115	0.13 (0.06, 0.3)	0.23 (0.11, 0.39)	0.22 (0.08, 0.32)
<i>t</i>		1732.5	1769.5	1701.5
<i>P</i>		0.447	0.555	0.366

There was no significant difference in X, Y, and Z translational errors between the two groups ($P > 0.05$).

significant differences in rotational errors between the two groups ($P > 0.05$). The detailed data are shown in Table 5.

3.6. Comparison of Expansion Boundary outside the Target Area of Patients with Chest Disease in the Two Groups. The expansion boundary outside the target area of the patients in the cold weather group and usual weather group was X (0.57, 0.57) cm, Y (0.92, 0.78) cm, and Z (0.99, 0.68) cm, respectively, as shown in Table 6.

3.7. Comparison of Expansion Boundary outside the Target Area of Patients with Pelvic Disease in the Two Groups. The expansion boundary outside the target area of patients with pelvic disease in the cold weather group and usual weather group was X (0.63, 0.61) cm, Y (0.79, 0.79) cm, and Z (0.68, 0.61) cm, respectively, as shown in Table 7.

4. Discussion

Most countries and regions around the world experience the change of seasons. Substances show thermal expansion and contraction due to the change of temperature, so as the human body. Radiation therapy for tumors has entered the era of precision, which can achieve complete coverage of the tumor target area with a high dose according to planning requirements while forming a steep dose drop area around the tumor to reduce the toxicity for normal tissues around the target area. Therefore, if the fixed position of cancer patients is slightly changed during radiotherapy, it will lead to the failure of cancer treatment and the aggravation of normal tissue toxicity. Our research institute is in the south of China, although the temperature is not as low as in the north in winter, the wet and cold weather in the south is more pungent compared with the dry and cold weather in the north. This study mainly investigated the influence of weather on the setup error of cancer patients undergoing radiotherapy, providing a reference for clinical practice and future development of artificial intelligence neural networks.

As illustrated previously, there were no significant differences in X - and Y -axis translational errors and R_X , R_Y , and R_Z rotational errors of patients with chest disease between the cold weather group and usual weather group. But the M value of the cold weather group was higher than that of the usual weather group. It is worth noting that there was a significant difference in Z -axis translational error between the two groups ($U = 935.5$, $P = 0.005 < 0.05$), which may result from the integrated plate + thermoplastic mold used for the fixation. When the patient takes off the upper body clothes and lay on the plate in cold weather, the contraction of the whole back muscle may be caused by the contraction of the capillaries on the back, which may change the human body thickness in the ventrodorsal direction, resulting in the increase of Z -axis error in the registration with the planning image.

There were no significant differences in both translational and rotational setup errors of patients with pelvic disease, which may be because of the vacuum pad used for body fixation in radiotherapy. In cold weather, patients did not feel

uncomfortable lying on the pad after undressing. However, the data showed that the R_X value in the cold weather group (2.5°) was higher than that in the usual weather group (2°). A previous study has demonstrated that the rotation angle $>2^\circ$ influences the distribution of planning dose [19]. Therefore, if the setup error angle was great than 2° , it was necessary to reset the position since our linear accelerator was not equipped with a six-dimensional treatment table. The reason for the greater R_X value in the cold weather group might be because the head pad pillow and the vacuum pad wrapped the chest, abdomen, and most of the thighs, while the lower leg and foot are not wrapped in the vacuum pad, which will directly contact the carbon fiber plate of the treatment bed. This made the patient uncomfortable and caused their legs to extend and contract, leading to the increased R_X .

As for the M_{PTV} value of CTV to PTV expansion, the external boundary values of X , Y , and Z were 0.57 cm, 0.92 cm, and 0.99 cm in the cold weather group and 0.57 cm, 0.78 cm, and 0.68 cm in usual weather group, respectively. The expansion boundary values in the left and right directions are almost equal. But the values in the head-foot direction and ventrodorsal direction in the cold weather group were 0.14 cm and 0.31 cm larger than that in the usual weather group, respectively. Shen et al. [20] designed a new breast vacuum bag to obtain the M_{PTV} boundary values of 0.44 cm, 0.31 cm, and 0.38 cm required for X , Y , and Z , respectively, which were smaller than those obtained in this study. Deseyne et al. [21] used a modified prone breastplate to obtain the M_{PTV} boundary values of 1.14 cm, 1.21 cm, and 1.0 cm required for X , Y , and Z , respectively, which were greater than those obtained in this study. Cravo Sá A et al. [22] employed 3D surface image guidance to achieve the expanded M_{PTV} boundary values of 0.89 cm, 10.4 cm, and 0.93 cm required for X , Y , and Z , respectively, which were greater than the values in the usual weather group and similar to those in the cold weather group in this study.

As for the M_{PTV} value of CTV to PTV of patients with pelvic disease, the external boundary values of X , Y , and Z were 0.63 cm, 0.79 cm, and 0.68 cm in the cold weather group and 0.61 cm, 0.79 cm, and 0.61 cm in the usual weather group, respectively. They were almost equal. Singh et al. [23] applied EPID image guidance to achieve the PTV boundary values required for X , Y , and Z , which were 0.9 cm, 1.0 cm, and 0.6 cm, respectively, when the alignment marker points were added without body fixation. Compared with the values in this study, X and Y values were larger, while Z value was the same. Wu et al. [1] used CBCT image-guided equipment, the M_{PTV} boundary values required for X , Y , and Z were 0.40 cm, 0.81 cm, and 0.45 cm, respectively. Compared with the values in this study, the expansion boundary values of X and Z were smaller, while the expansion boundary value of Y -axis was almost the same. Patni et al. [24] employed the CBCT image-guided equipment to achieve the M_{PTV} boundary values required for X , Y , and Z , which were 0.56 cm, 1.03 cm, and 0.58 cm, respectively. The expansion boundary values of X and Z were smaller, while that of Y -axis was larger compared with those in this study. In addition, the largest expansion boundary value was found in Y -axis among the three axes in the translational direction, as reported in other studies.

TABLE 5: Comparison of rotational errors between the two groups.

Group	Number of CBCT	Rotational errors		
		R_X	R_Y	R_Z
Cold weather group	33	1.1 (0.5, 1.8)	0.5 (0.20, 0.80)	0.5 (0.3, 0.9)
Usual weather group	115	0.8 (0.4, 1.55)	0.6 (0.25, 1.2)	0.5 (0.2, 1.05)
t		1607	1700.5	1853
P		0.18	0.363	0.837

TABLE 6: M_{PTV} of CTV to PTV expansion boundary of patients with chest disease in different groups.

Group	n	Σ			δ			M_{PTV}		
		X	Y	Z	X	Y	Z	X	Y	Z
Cold weather group	26	0.19	0.31	0.35	0.15	0.21	0.18	0.57	0.92	0.99
Usual weather group	112	0.15	0.21	0.18	0.19	0.26	0.23	0.57	0.78	0.68

TABLE 7: M_{PTV} of CTV to PTV expansion boundary of patients with pelvic disease in different groups.

Group	Number of CBCT	Σ			δ			M_{PTV}		
		X	Y	Z	X	Y	Z	X	Y	Z
Cold weather group	33	0.20	0.27	0.23	0.18	0.18	0.16	0.63	0.79	0.68
Usual weather group	115	0.18	0.18	0.16	0.20	0.26	0.20	0.61	0.79	0.61

There was no significant difference in R_X , R_Y , and R_Z rotational errors between the two groups ($P > 0.05$).

5. Conclusion

In conclusion, the setup error of patients treated with radiotherapy in cold weather using the integrated plate and thermoplastic mold fixation will be greater than that in usual weather, especially in the ventrodorsal direction. Therefore, attention must be paid to the possibility of this error in cold weather. The measures that should be taken are as follows. (i) The patient should be asked to arrive at the radiotherapy area about an hour in advance to avoid direct entry to the machine room for treatment. (ii) The air conditioning temperature in the radiotherapy machine room and radiotherapy waiting area can be appropriately increased to ensure the normal operation of radiotherapy equipment. (iii) The integrated plate can be heated to make the patient comfortable when touching it. (iv) In cold weather, the number of image guidance can be increased. (v) For patients with pelvic tumors who need the integrated plate and vacuum pad for fixation in radiotherapy, a large error in the rotation direction (R_X) may occur because the heel and lower leg are directly exposed to the ice-cold carbon brazing treatment bed surface. Thus, attention should be paid. A warming soft pad can be laid during positioning and treatment to improve comfort and reduce their movement.

Data Availability

The simulation experimental data used to support the findings of this study are available from the corresponding author upon request.

Conflicts of Interest

The authors declare that there are no conflicts of interest regarding the publication of this study.

Authors' Contributions

Shuxue Zhao and Xianfa Lu designed study; Jiasen Zou, Zhouying Xu, Siyu Wei and Liping Li analyzed and interpreted data; Qingguo Fu drafted manuscript; all authors reviewed and approved manuscript.

References

- [1] W.-C. Wu, Y.-R. Chang, Y.-L. Lai et al., "Impact of body-mass factors on setup displacement during pelvic irradiation in patients with lower abdominal cancer," *Radiology and Oncology*, vol. 53, no. 2, pp. 256–264, 2019.
- [2] L. L. Lin, L. Hertan, R. Rengan, and B.-K. K. Teo, "Effect of body mass index on magnitude of setup errors in patients treated with adjuvant radiotherapy for endometrial cancer with daily image guidance," *International Journal of Radiation Oncology, Biology, Physics*, vol. 83, no. 2, pp. 670–675, 2012.
- [3] J. Zhao, M. Zhang, F. Zhai, H. Wang, and X. Li, "Setup errors in radiation therapy for thoracic tumor patients of different body mass index," *Journal of Applied Clinical Medical Physics*, vol. 19, no. 3, pp. 27–31, 2018.
- [4] L. Conroy, S. Quirk, E. Watt et al., "Deep inspiration breath hold level variability and deformation in locoregional breast irradiation," *Practical Radiation Oncology*, vol. 8, no. 3, pp. e109–e116, 2018.

- [5] M. Aristophanous, P.-C. M. Chi, J. Kao et al., "Deep-Inspiration breath-hold intensity modulated radiation therapy to the mediastinum for lymphoma patients: setup uncertainties and margins," *International Journal of Radiation Oncology, Biology, Physics*, vol. 100, no. 1, pp. 254–262, 2018.
- [6] A. Arns, H. Wertz, J. Boda-Heggemann et al., "Ultrafast single breath-hold cone-beam CT lung cancer imaging with faster linac gantry rotation," *Radiotherapy & Oncology*, vol. 135, pp. 78–85, 2019.
- [7] Y. Lu, X. Ni, J. Yu et al., "Lower limb immobilization device induced small setup errors in the radiotherapy," *Medicine*, vol. 97, no. 15, p. e384, 2018.
- [8] H. Chai, Y. Narita, M. Takagi, M. Kudo, T. Kimura, and K. Kattou, "Technical note: improved positioning protocol for patient setup accuracy in conventional radiotherapy for lung cancer," *Radiological Physics and Technology*, vol. 12, no. 4, pp. 426–432, 2019.
- [9] J.-H. Pan, D.-D. Yu, Q.-B. Song et al., "Comparison of geometrical uncertainties in the radiotherapy for various treatment sites with two different immobilization marking methods," *Journal of X-Ray Science and Technology*, vol. 27, no. 4, pp. 703–714, 2019.
- [10] M. Dubec, S. Brown, R. Chuter et al., "MRI and CBCT for lymph node identification and registration in patients with NSCLC undergoing radical radiotherapy," *Radiotherapy & Oncology*, vol. 159, pp. 112–118, 2021.
- [11] C. Chen, C. Wu, Y. Zhong et al., "A new registration algorithm of electronic portal imaging devices images based on the automatic detection of bone edges during radiotherapy," *Scientific Reports*, vol. 10, no. 1, 2020.
- [12] M. Machiels, P. Jin, C. H. van Gurp, J. Hoof, T. Alderliesten, and M. Hulshof, "Comparison of carina-based versus bony anatomy-based registration for setup verification in esophageal cancer radiotherapy," *Radiation Oncology*, vol. 13, no. 1, 2018.
- [13] Z. Mulla, R. K. Alwassia, E. M. Senan et al., "A comparative study between open-face and closed-face masks for head and neck cancer (HNC) in radiation therapy," *Reports of Practical Oncology and Radiotherapy*, vol. 25, no. 3, pp. 382–388, 2020.
- [14] Y. Ding, P. Ma, W. Li et al., "Effect of surgical mask on setup error in head and neck radiotherapy," *Technology in Cancer Research and Treatment*, vol. 19, p. 1180566022, 2020.
- [15] S. Ohira, N. Kanayama, R. Komiyama et al., "Intra-fractional patient setup error during fractionated intracranial stereotactic irradiation treatment of patients wearing medical masks: comparison with and without bite block during COVID-19 pandemic," *Journal of Radiation Research*, vol. 62, no. 1, pp. 163–171, 2021.
- [16] J. C. Stroom, H. C. J. de Boer, H. Huizenga, and A. G. Visser, "Inclusion of geometrical uncertainties in radiotherapy treatment planning by means of coverage probability," *International Journal of Radiation Oncology, Biology, Physics*, vol. 43, no. 4, pp. 905–919, 1999.
- [17] D. Delishaj, S. Ursino, F. Pasqualetti et al., "Set-up errors in head and neck cancer treated with IMRT technique assessed by cone-beam computed tomography: a feasible protocol," *Radiation Oncology Journal*, vol. 36, no. 1, pp. 54–62, 2018.
- [18] M. Velec, J. N. Waldron, B. O'Sullivan et al., "Cone-beam CT assessment of ii setup error of two head-and-neck cancer thermoplastic masks," *International Journal of Radiation Oncology, Biology, Physics*, vol. 76, no. 3, pp. 949–955, 2010.
- [19] O. Gutfeld, A. E. Kretzler, R. Kashani, D. Tatro, and J. M. Balter, "Influence of rotations on dose distributions in spinal stereotactic body radiotherapy (SBRT)," *International Journal of Radiation Oncology, Biology, Physics*, vol. 73, no. 5, pp. 1596–1601, 2009.
- [20] K. Shen, J. Xiong, Z. Wang et al., "Design of a new breast vacuum bag to reduce the global and local setup errors and to reduce PTV margin in post-mastectomy radiation therapy," *Journal of Radiation Research*, vol. 61, no. 6, pp. 985–992, 2020.
- [21] P. Deseyne, B. Speleers, W. De Neve et al., "Crawl positioning improves set-up precision and patient comfort in prone whole breast irradiation," *Scientific Reports*, vol. 10, no. 1, 2020.
- [22] A. Cravo Sá, A. Fermento, D. Neves et al., "Radiotherapy setup displacements in breast cancer patients: 3D surface imaging experience," *Reports of Practical Oncology and Radiotherapy*, vol. 23, no. 1, pp. 61–67, 2018.
- [23] A. Singh, S. Singh, K. R. Mani, K. J. M. Das, and S. Kumar, "Comparing setup errors using portal imaging in patients with g cancers by two methods of alignment," *Journal of Medical Imaging and Radiation Sciences*, vol. 51, no. 3, pp. 394–403, 2020.
- [24] N. Patni, N. Burela, R. Pasricha et al., "Assessment of three-dimensional setup errors in image-guided pelvic radiotherapy for uterine and cervical cancer using kilovoltage cone-beam computed tomography and its effect on planning target volume margins," *Journal of Cancer Research and Therapeutics*, vol. 13, no. 1, pp. 131–136, 2017.