

# Robust contour propagation using deep learning and image registration for online adaptive proton therapy of prostate cancer

Mohamed S. Elmahdy<sup>a)</sup>

*Division of Image Processing, Department of Radiology, Leiden University Medical Center, 2300, RC Leiden, The Netherlands*

Thyrza Jagt

*Erasmus MC Cancer Institute, Rotterdam, The Netherlands*

Roel Th. Zinkstok

*Department of Radiation Oncology, Leiden University Medical Center, 2300, RC Leiden, The Netherlands*

Yuchuan Qiao, Rahil Shahzad, Hessam Sokooti, and Sahar Yousefi

*Division of Image Processing, Department of Radiology, Leiden University Medical Center, 2300, RC Leiden, The Netherlands*

Luca Incrocci

*Erasmus MC Cancer Institute, Rotterdam, The Netherlands*

C.A.M. Marijnen

*Department of Radiation Oncology, Leiden University Medical Center, 2300, RC Leiden, The Netherlands*

Mischa Hoogeman

*Erasmus MC Cancer Institute, Rotterdam, The Netherlands*

Marius Staring

*Division of Image Processing, Department of Radiology, Leiden University Medical Center, 2300, RC Leiden, The Netherlands*

*Department of Radiation Oncology, Leiden University Medical Center, 2300, RC Leiden, The Netherlands*

*Intelligent Systems Department, Delft University of Technology, 2600, GA Delft, The Netherlands*

(Received 22 February 2019; revised 13 May 2019; accepted for publication 13 May 2019; published 12 July 2019)

**Purpose:** To develop and validate a robust and accurate registration pipeline for automatic contour propagation for online adaptive Intensity-Modulated Proton Therapy (IMPT) of prostate cancer using `elastix` software and deep learning.

**Methods:** A three-dimensional (3D) Convolutional Neural Network was trained for automatic bladder segmentation of the computed tomography (CT) scans. The automatic bladder segmentation alongside the computed tomography (CT) scan is jointly optimized to add explicit knowledge about the underlying anatomy to the registration algorithm. We included three datasets from different institutes and CT manufacturers. The first was used for training and testing the ConvNet, where the second and the third were used for evaluation of the proposed pipeline. The system performance was quantified geometrically using the dice similarity coefficient (DSC), the mean surface distance (MSD), and the 95% Hausdorff distance (HD). The propagated contours were validated clinically through generating the associated IMPT plans and compare it with the IMPT plans based on the manual delineations. Propagated contours were considered clinically acceptable if their treatment plans met the dosimetric coverage constraints on the manual contours.

**Results:** The bladder segmentation network achieved a DSC of 88% and 82% on the test datasets. The proposed registration pipeline achieved a MSD of  $1.29 \pm 0.39$ ,  $1.48 \pm 1.16$ , and  $1.49 \pm 0.44$  mm for the prostate, seminal vesicles, and lymph nodes, respectively, on the second dataset and a MSD of  $2.31 \pm 1.92$  and  $1.76 \pm 1.39$  mm for the prostate and seminal vesicles on the third dataset. The automatically propagated contours met the dose coverage constraints in 86%, 91%, and 99% of the cases for the prostate, seminal vesicles, and lymph nodes, respectively. A Conservative Success Rate (CSR) of 80% was obtained, compared to 65% when only using intensity-based registration.

**Conclusion:** The proposed registration pipeline obtained highly promising results for generating treatment plans adapted to the daily anatomy. With 80% of the automatically generated treatment plans directly usable without manual correction, a substantial improvement in system robustness was reached compared to a previous approach. The proposed method therefore facilitates more precise proton therapy of prostate cancer, potentially leading to fewer treatment-related adverse side effects.

© 2019 The Authors. *Medical Physics* published by Wiley Periodicals, Inc. on behalf of American Association of Physicists in Medicine. [https://doi.org/10.1002/mp.13620]

Key words: convolutional neural networks (CNN), deformable image registration, distended rectum, generative adversarial network (GAN), prostate cancer, proton therapy

## 1. INTRODUCTION

Prostate cancer is one of the leading causes of mortality and the most common cancer among men. The National Cancer Society (NCS) estimates around 164 690 new cases and 24 430 deaths from prostate cancer in the United States only for 2018.<sup>1</sup> Due to its slow progress, individuals could develop prostate cancer for many years without explicit signs. There are treatment options for prostate cancer including surgical removal of the prostate, hormone therapy, and radiotherapy. Intensity-Modulated Proton Therapy (IMPT) is able to deliver a highly localized dose distribution to the target volume, while minimizing collateral damage to the surrounding healthy tissues.<sup>2</sup> IMPT is, however, more sensitive to daily changes than photon therapy, which may result in distortion of the delivered dose distribution.<sup>3,4</sup> These changes could arise from anatomical variations in the shape and position of both target volumes and organs-at-risk (OARs) or a misalignment in the patient setup. In order to compensate for these changes, a margin is added to the clinical target volume (CTV) to generate the planning target volume (PTV) in addition to robust treatment planning. These margins result in extra dose to the OARs, leading to an increase in the treatment-related toxicities that may prevent dose escalation. Traditionally, motion-induced variations are minimized by implanting fiducial markers in the prostate, subsequently compensating for the daily prostate motion using online imaging.<sup>5</sup> However, such correction strategies are invasive and only capable of correcting for translational motion and limited amount of rotational motion.<sup>6</sup> Online imaging and replanning should be able to handle this problem without using fiducial markers.<sup>7</sup> These online computed tomography (CT) scans have to be delineated first in order to update the treatment plan. Usually, this task is done by radiation oncologists according to certain guidelines.<sup>8,9</sup> However, intra- and interobserver inconsistency has been noted due to different preferences and experience among radiation oncologists.<sup>10,11</sup> Typically, daily manual recontouring is not performed because it is time consuming and new anatomical variations may be introduced in the time it takes to delineate the scan.<sup>12</sup> Automatic recontouring algorithms can alleviate these issues, but robust methods are required, because otherwise still time-consuming fallback strategies are needed.

Automatic recontouring could be accomplished effectively using deformable image registration (DIR) by deducing the correspondence between the daily CT and the planning CT. Using the generated deformation vector field (DVF), manual contours can be propagated from the planning CT to the daily CT. The automatically generated contours together with fast reoptimization of the treatment plan<sup>13</sup> could compensate for the daily variation and ensure the delivery of the prescribed dose distribution at small margins and robust settings. DIR is a crucial step toward developing online adaptive IMPT alongside replanning and personalized dose quality assurance (QA). Currently, these steps are time consuming, thus severely limiting online procedures.

There are commercially available applications for automatic recontouring such as atlas-based autosegmentation (ABAS), Mirada, and RayStation. These applications are, however, considered a black box for the end users and therefore limit the parameter choices and tuning. Open source DIR packages provide a high level of flexibility with a concrete scientific evidence and reproducibility. Qiao *et al.*<sup>14</sup> reported an MSD of  $1.36 \pm 0.30$  mm,  $1.75 \pm 0.84$  mm,  $1.49 \pm 0.44$  mm for the prostate, seminal vesicles, and lymph nodes, respectively, for 18 patients using the open source `elasticx` software. A clinical success rate of 69% was achieved, which means that 31% of the delineations have to be corrected, leading to increased costs and a suboptimal patient workflow. In 2011, Thor *et al.*<sup>15</sup> deployed DIR to propagate the contours of the prostate and OARs from CT to cone-beam CT. The system achieved a mean DSC of 0.80 for the prostate, 0.77 for the rectum, and 0.73 for the bladder with a relatively high variance. Moreover, the system was not qualitatively evaluated in terms of dosimetric coverage. Recently, Woerner *et al.*<sup>16</sup> investigated the error between different radiologists and both DIR and rigid registration in different body regions. They only reported the results for the prostate, which were 0.90, 0.99 mm, and 8.12 mm for the DSC, MSD, and Hausdorff Distance (HD), respectively. Thörnqvist *et al.*<sup>17</sup> used two different demons-based registration algorithms, with one more conservative than the other. They achieved an average DSC of 0.88, 0.85, 0.89, and 0.78 for the lymph nodes, prostate, bladder, and rectum, respectively.

In spite of the existence of quite accurate registration algorithms, they still suffer from a lack of robustness, which is a critical aspect for clinical application. Therefore, in this paper, we focus on the robustness aspect of the registration pipeline. The main challenges in Qiao *et al.* were the presence of gas pockets and large deformations surrounding the seminal vesicles, bladder, and rectum. Hence, we propose to tackle these challenges by inpainting the rectum gas pockets as well as embedding the bladder segmentation in the registration pipeline using deep learning to enhance the system's robustness. The proposed registration pipeline was evaluated geometrically and dosimetrically for generating clinically acceptable IMPT plans. Compared to our conference paper,<sup>18</sup> we made several improvements, such as the inclusion of more datasets, dealing with gas pockets, data normalization, and multistage registration. Moreover, we carried out an extensive dosimetric validation for the automatically generated contours to verify its clinical viability.

## 2. MATERIALS AND METHODS

The prostate and seminal vesicles are positioned between the bladder and the rectum; therefore, prostate motion is mainly influenced by the filling and motion of both the bladder and the rectum.<sup>19</sup> Hence, we hypothesize that embedding an explicit prior knowledge about the deformation of either organs to the intensity-based DIR method may improve the accuracy and robustness of the registration. Here, we considered the bladder because it has a well-defined shape that

could be more easily delineated in a fully automatic manner than the rectum. Since the registration is intensity based, the quality of the registration process is correlated to the quality of the input images. Hence, we introduced multiple data pre-processing steps to enhance the quality of the input images. These steps include rectum gas pocket detection and inpainting and contrast clipping as shown in Fig. 1.

### 2.A. Bladder segmentation using deep learning

In this study, we automatically segment the bladder using a three-dimensional U-Net convolutional neural network (3D-CNN) similar to the architecture introduced in Ref. [21]. The network consists of encoding and decoding branches connected with skip connections as shown in Fig. 2. In order to represent the volumetric information and tissue homogeneity of the CT volume, 3D convolution layers followed by non-linear leaky rectified linear units were used. The original maxpooling layers were replaced by strided convolution in both encoder and decoder branches. Negative Dice Similarity Coefficient (DSC)<sup>22</sup> is deployed as a cost function and the network is trained using the Adam optimizer<sup>23</sup> with a fixed learning rate of  $10^{-4}$ . The network has 64 320 trainable parameters which enable network inference of the entire CT image in approximately 2 s. The network was designed to output the same size as input; however, the input size should be divisible by 16. Largest connected component analysis was applied as a postprocessing step to eliminate irrelevant activations.

### 2.B. Gas pocket detection and inpainting

A problem that usually arises for intensity-based DIR of the pelvic region is the presence of gas pockets in the bowel and rectum. These pockets appear as dark areas surrounded by soft tissue. Usually, the size and position of these pockets are not the same in the planning and the daily CT. In such situations, physical correspondence between images at different sessions does not exist because of the insertion or occlusion of image content. Only few studies addressed this issue in the literature. Gao et al.<sup>25</sup> proposed introducing a virtual gas pocket to the planning CT that follows the pocket in the daily CT. They tested it on 15 prostate cancer patients with distended rectum. Foskey et al.<sup>26</sup> proposed to deflate the pocket to a virtual point. In both papers, the authors assumed no gas pockets in the planning CT, which is not usually the case. Recently, deep learning-based algorithms have revolutionized the medical image analysis field.<sup>27</sup> One category of deep learning architectures is Generative Adversarial Networks (GANs) introduced by Goodfellow et al.<sup>28</sup> in 2014. GANs have been growing since then in generating realistic natural and synthetic images. As for medical images, GANs have been used in image segmentation,<sup>29</sup> synthesis,<sup>30</sup> registration,<sup>31</sup> and denoising.<sup>32</sup> Recently, Yu et al.<sup>33</sup> proposed a two-dimensional (2D) GAN network with a contextual attention model to restore and inpaint occluded regions in natural images. The network also blends the restored region with the surrounding texture to make it look more realistic. The proposed model has two successive networks for image




FIG. 1. The proposed multistage registration process using elastix software and deep learning. The red boxes denote the contours finally used as output of the algorithm. [Color figure can be viewed at wileyonlinelibrary.com]




FIG. 2. The architecture for the three-dimensional-convolutional neural network, where the numbers on the blocks denote the number of feature maps. [Color figure can be viewed at wileyonlinelibrary.com]

generation in order to generate patches with fine quality. The first “generator” network generates a coarse result through a dilated convolution network. This result is then fed to the second network. The second “discriminator” network has two routes, one goes to a dilated convolution network while the other goes through a contextual attention model. Finally, the results from these two routes are concatenated and fed to a prediction network. This network has shown an improvement over a similar network proposed by Iizuka et al.<sup>34</sup> In this paper, we retrained this network so that it can inpaint (fill) gas pockets of different shapes and sizes with a more sophisticated and realistic content rather than a fixed value. The same implementation and hyperparameters were used as in the original paper. Alternatively, we also experimented with a simplified method for inpainting. Following the idea proposed by Rodriguez-Vila et al.,<sup>35</sup> we fill the gas pockets with a fixed value and smooth the output to blend it with the surrounding tissues. A threshold of  $-200$  is used to generate a binary mask of the gas pockets. This mask is then dilated with a kernel of size  $7 \times 7 \times 1$  voxels ( $M$ ) while the CT image is filled with a fixed HU number of 60 (the average HU number for faeces), and smoothed with a sigma of 4 mm ( $I_{smoothed}$ ). Eq. (1) shows the simple inpainting process:

$$I_{out} = I_{input} \times (1 - M) + I_{smoothed} \times M \quad (1)$$

Figure 3 shows a comparison between gas pocket inpainting using the GAN network and simple inpainting.

### 2.C. Contrast enhancement

To enhance the soft tissue contrast, the CT intensity was clipped to the range of  $[-300,300]$ . This clipping is similar to viewing the soft tissue with an appropriate window level. Moreover, such enhancement improves the registration convergence. Figure 4 shows the effect of intensity clipping.

### 2.D. Image registration

For carrying out the DIR experiments, we used the open software package `elastix`.<sup>36</sup> For more details, see the website <http://elastix.isi.uu.nl>. All the experiments were performed on a cluster of workstations operated on the Oracle Grid Engine (OGE), which has 500 nodes with a total of 800 cores. Testing time is reported using a PC with 16 GB memory, Windows 7 Professional 64 bit operation system, and an Intel Xeon E51620 CPU with 4 cores at 3.6 GHz, utilizing only the CPU. In this study, the planning CT scan (moving image) was aligned with the daily CT scan (fixed image) of each patient. The registrations were initialized based on the center of gravity of the bony anatomy defined by a Hounsfield number larger than 200. A mask of the body torso was generated using Pulmo software<sup>37</sup> to remove the effect of the CT table. The registration process is done in three stages. First, the moving and fixed images are registered using a single resolution affine transformation using 200 iterations as defined in Eq. (2):

$$\widehat{\mu}_1 = \arg \min_{\mu} C_1(I_F, I_M, M_F, M_M; T_{\mu_1}), \quad (2)$$

where  $I_F$  is the daily scan,  $I_M$  is the planning scan,  $M_F$  is the torso mask of the daily scan,  $M_M$  is the torso mask of the planning scan, and  $C_1$  is the mutual information cost function. The affine transformation aligns the bones and large structures. Second, a deformable registration is applied to tackle the local deformations of the organs. In this stage, the planning CT of each patient combined with the manual delineation of the bladder are considered the moving images, while the repeat CT of the same patient accompanied with the bladder segmentation resulting from the proposed 3D-CNN are the fixed images. Equation (3) defines the optimization problem for this stage:




FIG. 3. Different inpainting algorithms, where (a), (b), and (c) represent the original computed tomography, the result from simple-inpainting, and the result from Generative Adversarial network-inpainting, respectively.




FIG. 4. The effect of contrast clipping, where (a) and (b) represent the image before and after intensity clipping, respectively.

$$\widehat{\mu}_2 = \arg \min_{\mu} \{C_1(I_F, I_M, M_F, M_M, T_{\mu_1}; T_{\mu_2}) + \alpha C_2(DT(S_F), DT(S_M), T_{\mu_1}; T_{\mu_2})\}, \quad (3)$$

where  $C_2$  is the Mean Squared Difference (MSD) cost function,  $\alpha$  is a weight for balancing these two cost functions,  $DT(S_F)$  is the distance transform of the 3D-CNN bladder segmentation, and  $DT(S_M)$  is the distance transform of the manual annotation of the planning scan. The Distance Transform (DT) of the bladder segmentations is used instead of the binary segmentations themselves, to ensure a smooth and stable optimization process. The generated deformation vector field (DVF) from this step is then used to propagate the contours of the prostate, lymph nodes, bladder, and rectum from the planning CT to the repeat CT. Because the seminal vesicle is a small irregular structure, which is highly affected by the deformation in the rectum, we introduce a third stage to focus the registration on the rectum and seminal vesicle region. In this stage, the rectum contour of the planning CT and the rectum contour of the daily CT (from the previous stage) are dilated with a kernel of  $45 \times 45 \times 1$  voxels and used as a registration mask together with the fixed and moving CT scans. The contours of the rectum and seminal vesicles are then propagated using the generated DVF from the final stage. Eq. (4) defines the optimization problem for this stage:

$$\widehat{\mu}_3 = \arg \min_{\mu} C_1(I_F, I_M, \tilde{M}_F, \tilde{M}_M, T_{\mu_1}, T_{\mu_2}; T_{\mu_3}), \quad (4)$$

where  $\tilde{M}_M$  is the dilated rectum mask of the planning CT and  $\tilde{M}_F$  is the dilated rectum mask of the daily CT. A fast recursive implementation of the B-spline transformation was employed for DIR<sup>38</sup> in stages 2 and 3. Adaptive

stochastic gradient descent was used for optimization<sup>39</sup> in all three stages. For the DIR stage, we used a three-level Gaussian pyramid with smoothing factors of 4, 2, and 1 mm. Figure 1 illustrates the proposed registration pipeline in detail.

### 3. EXPERIMENTS AND RESULTS

#### 3.A. Dataset

This study includes three datasets representing three different institutes and CT scanners from three different vendors for patients who underwent intensity-modulated radiation therapy for prostate cancer. Table I shows detailed information about these datasets. The dataset from Leiden University Medical Center (LUMC) was used to train and validate the neural network for segmenting the bladder (Section 2.A) as well as the inpainting network (section 2.B), while the datasets from Erasmus Medical Center (EMC) and Haukeland Medical Center (HMC) were used as independent test sets for the complete registration pipeline. Geometric evaluation was performed on both the EMC and HMC dataset. Eleven of the 18 HMC patients were considered for dosimetric evaluation due to the availability of not only the manual delineations for the target organs (prostate, seminal vesicles, lymph nodes) and OARs (bladder, rectum) but also the manual delineations of the bowels and femoral heads needed for planning.

#### 3.B. Evaluation measures

The quality of the registration is quantified in terms of geometric aspects and dosimetric coverage. The geometric quality is measured by comparing the manual contours and the automatically propagated contours of the daily CT for the prostate, lymph nodes, seminal vesicles, rectum, and bladder. The Dice Similarity Coefficient (DSC) measures the overlap between the segmentations, while the Mean Surface Distance (MSD) and the 95% Hausdorff Distance (HD) measure the residual distance between the contours in 3D space.

$$DSC = \frac{2|F \cap M|}{|F| + |M|}, \quad (5)$$

where  $F$  and  $M$  are the propagated contour and the ground truth contour, respectively.

TABLE I. Details of the datasets reported in this study.

Institute	Scanner	#Patients	#Scans/ patient	Image size	Voxel spacing (mm)	Manual delineations
LUMC	Toshiba	418	1	512 × 512 × (68-240)	~1.0 × 1.0 × 3.0	bladder, rectum
EMC <sup>40</sup>	Siemens	14	4	512 × 512 × (91-218)	~0.9 × 0.9 × 1.5	prostate, SV bladder, rectum
HMC <sup>41</sup>	GE	18	8-11	512 × 512 × (90-180)	~0.9 × 0.9 × 2.0	prostate, SV, LN bladder, rectum

LUMC, EMC, and HMC are abbreviations for Leiden University Medical Center (Netherlands), Erasmus Medical Center (Netherlands), and Haukeland Medical Center (Norway), respectively. SV and LN denote seminal vesicles and lymph nodes, respectively.

$$MSD = \frac{1}{2} \left( \frac{1}{n} \sum_{i=1}^n d(a_i, M) + \frac{1}{m} \sum_{i=1}^m d(b_i, F) \right), \quad (6)$$

$$HD = \max \{ \max_i \{ d(a_i, M) \}, \max_j \{ d(b_j, F) \} \}, \quad (7)$$

where  $\{a_1, a_2, \dots, a_n\}$  and  $\{b_1, b_2, \dots, b_m\}$  are the surface mesh points of the fixed and moving contours, respectively, and  $d(a_i, M) = \min_j \|b_j - a_i\|$ . The geometrical success rate, as a marker for geometric robustness, is defined as the percentage of registrations with  $MSD < 2$  mm (slice thickness):  $\gamma = \frac{n}{N} \{MSD < 2\text{ mm}\}$ , where (N) is the total number of registrations performed.

IMPT plans were generated for 11 patients from the HMC dataset using both the manual and the automatic delineations. The plans were then evaluated on the manual delineations to investigate the clinical effect of the error between these two delineations. Erasmus-iCycle, an in-house developed treatment planning optimization system,<sup>42-46</sup> together with the Astroid dose engine was used to generate the IMPT plans. Erasmus-iCycle uses a multi-criteria optimization to generate a clinically desirable Pareto optimal treatment plan on the basis of a wish list consisting of hard constraints and objectives. A small margin of 2 mm around the prostate and 3.5 mm around the lymph nodes and seminal vesicles is used to compensate for the marginal error of the propagated contours and to account for intraobserver variations in the manual contouring. These margins alone cannot account for variations in shape and location of the target volumes. Dose was prescribed according to a simultaneously integrated boost scheme in which the high-dose PTV (prostate + 2 mm margin) was assigned 74 Gy and the low-dose PTV (seminal vesicles and lymph nodes + 3.5 mm margin) 55 Gy, to be delivered using two laterally opposed beams. In order to avoid underdose, the optimization ensures that at least 98% of the target volumes receive at least 95% of the prescribed dose ( $V_{95\%} \geq 98\%$ ). To avoid overdose, the optimization ensures that <2% of the target volumes receive more than 107% of the highest prescribed dose ( $V_{107\%} \leq 2\%$ ). To achieve a clinically acceptable result, automatically generated treatment plans from the propagated contours should still fulfill these goals. Hence, IMPT plans from the propagated contours are evaluated based on the manual contours. The clinical success rate, as a marker for geometric robustness, is defined as the percentage of registrations for which the prostate directly

meets the dose treatment criteria:  $\eta = \frac{n}{N} \{V_{95\%} \geq 98\%\}$ . Conservative Success Rate (CSR) is a more conservative measure of clinical success when all target volumes (the prostate, seminal vesicles, and lymph nodes) meet this dosimetric criterion. For dosimetric coverage calculation  $N = 99$ .

### 3.C. Network training and performance

We implemented the 3D-CNN and GAN-inpainting networks using Tensorflow.<sup>47</sup> For training these networks, we used the LUMC dataset. This dataset was a sufficiently large dataset to be able to train the neural networks. Since the LUMC dataset only had one CT scan per patient, it was not used for registration evaluation. From the 418 LUMC patients, 350 patients were used for network training and 68 patients for validation. The trained network was then applied without modification to the CT scans in the EMC and HMC datasets. In order to account for the variations in voxel size between datasets and scans, all scans were resampled to a fixed voxel size of 1.0 × 1.0 × 2.0 mm. For the 3D-CNN, 100 000 patches of size 96 × 96 × 96 voxels were randomly extracted from the training volumes, making sure that they are equally distributed between foreground and background. For the GAN-inpainting network, all the slices with gas pockets were eliminated from training. Moreover, all slices were resampled to a pixel size of 1.0 × 1.0 mm and centrally cropped to 256 × 256 pixels so that more patches could fit into memory as well as it would be beneficial for the network to learn the most relevant contextual information to the rectum. Randomly selected windows of size 64 × 64 pixels were occluded in order to train the network to inpaint these regions with a realistic content. Both the 3D-CNN and the 2D-GAN-inpainting networks were trained for 100 000 iterations on the raw CT patches without any preprocessing except for resampling. All the experiments were carried out using an NVIDIA GTX1080 Ti with 11 GB of GPU memory. The 3D-CNN bladder segmentation network obtained a DSC of 85.4% ± 1.4% on the validation scans. Moreover, the network was tested on the EMC and HMC datasets and achieved an average DSC of 82.3% ± 1.5% and 87.9% ± 1.2%, respectively. Using a single GPU, the average inference time of the segmentation and inpainting networks were approximately 2 and 3 s per volume depending on the number of slices per volume. Figure 5 shows examples of the network output.




FIG. 5. Examples of the automatic bladder segmentation using the three-dimensional-convolutional neural network alongside the dice similarity coefficient of the volume. First and second rows represent samples from HMC and EMC, respectively. (a) and (d) are suboptimal results and the rest are good results. The red line represents the ground truth and the blue line is the network output. [Color figure can be viewed at [wileyonlinelibrary.com](http://wileyonlinelibrary.com)]

### 3.D. Parameter optimization and preprocessing analysis

For a fair comparison, the same registration parameters as in Ref. [14] were used. For the weight  $\alpha$  that balances the contribution of the bladder segmentation in the cost function (3), we investigated multiple settings based on initial experiments on EMC and HMC datasets. The weight was set for the coarse (first) resolution only and was set to zero for the other two resolutions, in order to avoid overfitting issues. Here we compared four settings for  $\alpha$ : 0.2, 0.1, 0.05, and 0.01. For this experiment, we did not use inpainting. The results are shown in Table II for the HMC dataset where "Affine" refers to the affine registration defined in Eq. (2), which is considered a reference method. The weights 0.05 and 0.20 yielded very similar performance. We opted for a weight of 0.05 to avoid overfitting on the bladder. Since the target areas (prostate, lymph nodes, and seminal vesicles) obtained slightly better accuracy for a lower weight and these are important for radiotherapy planning, we selected 0.05. For the EMC dataset, a similar experiment gave a weight of 0.01 (not reported). Therefore, for the remainder of the paper, these weights have been used.

In order to investigate the difference between simple-inpainting and GAN-inpainting, we run the registration on HMC dataset using both techniques as shown in Table III. The results show a very similar performance for simple-inpainting and GAN-inpainting. Hence, the simple-inpainting is used for gas pocket inpainting for the remainder of the paper.

From the aforementioned experiments and analysis (Tables II and III), we noticed a similar performance between 100 and

500 iterations, and in order to reduce the registration time, we considered only the results from 100 iterations for the final experiments.

### 3.E. Registration performance

Since the LUMC dataset did not have any follow-up scans, we only consider the EMC and HMC datasets for evaluating the registration performance. Figure 6 shows example results of the automatically propagated contours. We compared the proposed method with the intensity-based registration approach of Qiao *et al.*<sup>14</sup> For the HMC data, we directly compare with the results reported in Ref. [14], as the same dataset was used. For the EMC data, we applied their algorithm, and compare with our results. The DSC overlap of the proposed algorithm is presented in Table IV. For the HMC dataset, the prostate, lymph nodes, and bladder performed similarly for the proposed method and Qiao *et al.*, while the seminal vesicles and rectum showed substantial improvements. The median DSC values of the prostate, seminal vesicles, lymph nodes, rectum, and bladder were 0.88, 0.70, 0.89, 0.78, and 0.91, respectively, for Qiao *et al.*, while they were 0.89, 0.73, 0.89, 0.85, and 0.94, respectively, for the proposed method. For the EMC dataset, the proposed algorithm showed consistent improvement for the seminal vesicles, rectum, and bladder. The median DSC values of the prostate, seminal vesicles, rectum, and bladder were 0.91, 0.80, 0.76, and 0.86, respectively, for Qiao *et al.* and 0.89, 0.81, 0.81, and 0.90, respectively, for the proposed method. For the MSD results shown in Table V, the proposed method outperformed Qiao *et al.* for all the target areas and OARs. The MSD of most

Method	$\alpha$	Prostate $\mu \pm \sigma$	Seminal vesicles $\mu \pm \sigma$	Lymph nodes $\mu \pm \sigma$	Rectum $\mu \pm \sigma$	Bladder $\mu \pm \sigma$
Affine, 200		1.63 ± 0.74	2.92 ± 1.74	1.23 ± 0.49	3.89 ± 1.62	4.37 ± 2.11
B-spline, 100	0.20	1.55 ± 0.90	1.70 ± 0.74	1.63 ± 0.58	2.70 ± 1.12	1.85 ± 1.85
	0.10	1.53 ± 0.82	1.72 ± 0.73	1.58 ± 0.50	2.72 ± 1.11	1.85 ± 1.71
	0.05	1.50 ± 0.75	1.74 ± 0.79	1.55 ± 0.46	2.75 ± 1.16	1.86 ± 1.56
	0.01	1.41 ± 0.36	1.75 ± 0.86	1.57 ± 0.38	2.76 ± 1.15	1.98 ± 1.19
B-spline, 500	0.20	1.49 ± 0.90	1.76 ± 0.80	1.65 ± 0.64	2.87 ± 1.39	1.74 ± 1.63
	0.10	1.45 ± 0.77	1.77 ± 0.93	1.59 ± 0.52	2.78 ± 1.19	1.77 ± 1.58
	0.05	1.43 ± 0.77	1.78 ± 0.90	1.55 ± 0.47	2.79 ± 1.19	1.81 ± 1.57
	0.01	1.36 ± 0.47	1.76 ± 0.82	1.56 ± 0.48	2.81 ± 1.18	1.84 ± 1.24

Registrations using 100 and 500 iterations were both tested.

# It.	Inpainting method	Prostate $\mu \pm \sigma$	Seminal vesicles $\mu \pm \sigma$	Lymph nodes $\mu \pm \sigma$	Rectum $\mu \pm \sigma$	Bladder $\mu \pm \sigma$
100	Simple	1.29 ± 0.39	1.48 ± 1.16	1.49 ± 0.44	2.39 ± 1.92	1.72 ± 1.17
	GAN	1.29 ± 0.41	1.70 ± 2.12	1.49 ± 0.44	2.65 ± 2.17	1.71 ± 1.16
500	Simple	1.28 ± 0.42	1.36 ± 0.40	1.49 ± 0.44	2.19 ± 1.03	1.67 ± 1.22
	GAN	1.28 ± 0.42	1.36 ± 0.38	1.48 ± 0.45	2.33 ± 0.95	1.67 ± 1.22

Registrations using 100 and 500 iterations were both tested.

of the targets and the OARs was less than one voxel (2 mm). The geometrical success rate was 97%, 93%, and 87% for the prostate, seminal vesicles, and lymph nodes, respectively, for the HMC dataset and 67% and 71% for the prostate and seminal vesicles for the EMC dataset. Table VI shows the 95% HD, yielding a significant improvement for the proposed method over Qiao et al. on the HMC dataset, but less improvement for the EMC dataset. Moreover, Qiao et al. and the proposed method show a significant improvement from the affine method except for the lymph nodes. Figure 7 shows a scatter plot depicting the effect of the bladder distension (volume difference between planning and daily CT) on the Mean Surface Distance (MSD) of different target organs of the HMC dataset. The figure shows that the MSD of the proposed method is less than the slice thickness (2 mm) for most of the cases, and that there is little correlation between registration performance and bladder distensibility. Figure 8 shows the comparison of the registration performance between Qiao et al. (intensity only) and the proposed method (intensity and bladder segmentation), both using 100 iterations for the HMC dataset. The comparison illustrates the performance in terms of DSC, MSD, and 95%HD for the target volumes and OARs. The figure shows a similar pattern between the proposed method using the manually annotated contours of the bladder and the contours from the 3D-CNN network. This pattern emphasizes that the proposed method achieved the upper limit of the system. The average runtime for the proposed pipeline is 98.3 s for each registration at 100 iterations.

TABLE II. MSD (mm) of the target volumes and organs-at-risks of the HMC dataset for different registration and weight settings after the third stage of registration.

TABLE III. MSD (mm) of the target volumes and organs-at-risks for different registration settings and inpainting methods at  $\alpha = 0.05$ .

### 3.F. Dosimetric performance

Figure 6 shows the Dose–Volume Histogram (DVH) of the target organs and OARs for some examples. The clinical constraints in terms of  $V_{95\%}$  and  $V_{107\%}$  were calculated for the prostate, seminal vesicles, and lymph nodes based on the manual contours. In order to monitor the accumulated dose for the OARs, we calculated  $V_{45Gy\%}$ ,  $V_{60Gy\%}$ ,  $V_{75Gy\%}$ , and  $D_{mean}$  for the rectum, as well as  $V_{45\%}$ ,  $V_{65Gy\%}$ , and  $D_{mean}$  for the bladder. Here,  $D_{mean}$  is the structure’s average dose and  $V_{xxGy\%}$  is the percentage of volume receiving a dose of xx Gy. Table VII shows a comparison between the propagated contours from Qiao et al. and the proposed algorithm in terms of the percentage of scans that achieved the clinical criteria of  $V_{95\%} \geq 98\%$  and  $V_{107\%} \leq 2\%$ . The table shows a significant improvement for the seminal vesicles, which is a small and difficult target organ, while the performance of the prostate and lymph nodes was very similar. The boxplot in Fig. 9 illustrates the difference between the dosimetric parameter values of the manual delineations, calculated by using either the treatment plan based on the automated delineations or the manual delineations. We can see that the difference for all dosimetric parameters of all the target organs and OARs is almost 0% or Gy except for the lymph nodes, which is approximately 1%.

## 4. DISCUSSION

In this study, we developed and evaluated an automatic contour propagation pipeline using DIR, while considering the






FIG. 6. Examples from the automatic contours propagation of the HMC dataset and the corresponding dose–volume histograms evaluated on the manual contours. The solid line represents the manual contouring results while the dotted line is the automatically propagated one. [Color figure can be viewed at wileyonlinelibrary.com]

robustness, accuracy, and clinical acceptance rate for the target organs and the OARs of prostate cancer. Online adaptive IMPT is a crucial step toward treatment with small margins for target organs. In this study, we used margins of 2 mm for the prostate and 3.5 mm for the seminal vesicles and lymph nodes, respectively. Such small margins are only viable when online and daily replanning is performed. This replanning procedure should be accurate as well as robust to avoid any subsequent adverse side effects. The automatically propagated contours were validated geometrically on the EMC and HMC datasets as well as dosimetrically on the HMC dataset in order to investigate whether or not the propagated contours meet the clinical

acceptance criteria for dose coverage. DSC, MSD, and 95% HD were chosen for geometric validation while  $V_{95\%} \geq 98\%$  and  $V_{107\%} \leq 2\%$  were used for dosimetric coverage validation. Here,  $V_{95\%} \geq 98\%$  ensures that at least 98% of the target volumes receive at least 95% of the prescribed dose and  $V_{107\%} \leq 2\%$  ensures that <2% of the target volumes receive more than 107% of the highest prescribed dose.

In order to enhance the registration robustness, the segmentation of the bladder was introduced to steer the optimization. Since the registration process is partially driven by the bladder segmentation, this segmentation should be as accurate and robust as possible. Hence, we chose a 3D-CNN




FIG. 6. Continued. [Color figure can be viewed at wileyonlinelibrary.com]

Method	# It.	Prostate $\mu \pm \sigma$	Seminal vesicles $\mu \pm \sigma$	Lymph nodes $\mu \pm \sigma$	Rectum $\mu \pm \sigma$	Bladder $\mu \pm \sigma$
HMC						
Affine	200	0.84 ± 0.11	0.46 ± 0.26	0.90 ± 0.08	0.71 ± 0.10	0.77 ± 0.11
Qiao et al.	100	0.87 ± 0.08	0.65 ± 0.18	0.88 ± 0.07	0.77 ± 0.09	0.88 ± 0.11
Proposed	100	0.87 ± 0.08	0.70 ± 0.13 <sup>a</sup>	0.87 ± 0.07	0.82 ± 0.12 <sup>a</sup>	0.89 ± 0.12
EMC						
Affine	200	0.78 ± 0.20	0.49 ± 0.32	–	0.62 ± 0.18	0.66 ± 0.25
iao et al.	100	0.87 ± 0.13	0.70 ± 0.26	–	0.72 ± 0.16	0.78 ± 0.22
Proposed	100	0.87 ± 0.12	0.75 ± 0.18 <sup>a</sup>	–	0.78 ± 0.15 <sup>a</sup>	0.83 ± 0.17 <sup>a</sup>

<sup>a</sup> Represents a significant difference (at  $P = 0.05$ ) between Qiao et al. and the proposed algorithm.

TABLE IV. Dice similarity coefficient value of the target volumes and the organs-at-risks of the HMC and EMC datasets for different registration methods.

for bladder segmentation and obtained a DSC of 87.9% and a Jaccard index of 80.2%, which is very comparable to the reported Jaccard index of 81.9% in Ref. [48], where the authors developed a CNN network alongside level sets to segment the bladder in CT urography. Moreover, our proposed network outperformed the 2D CNN network developed by Zhou et al.<sup>49</sup>, where the authors reported a DSC of 72%. The

high performance of the proposed network may be attributed to the use of a large receptive field as well as replacing the 2D convolutions with 3D convolutions, which helps the network to embed depth information.

Applying contrast clipping to the CT scans before registration was beneficial to the registration process, since the registration is intensity based, which is consistent with the

TABLE V. MSD (mm) of the target volumes and the organs-at-risks of the HMC and EMC datasets for different registration methods.

Method	# It.	Prostate $\mu \pm \sigma$	Seminal vesicles $\mu \pm \sigma$	Lymph nodes $\mu \pm \sigma$	Rectum $\mu \pm \sigma$	Bladder $\mu \pm \sigma$
<b>HMC</b>						
Affine	200	1.70 ± 0.96	3.02 ± 1.96	1.26 ± 0.51	3.92 ± 1.59	4.47 ± 2.27
Qiao et al.	100	1.40 ± 0.47	1.85 ± 1.26	1.51 ± 0.44	3.13 ± 1.38	2.38 ± 1.79
Proposed	100	1.29 ± 0.39	1.48 ± 1.16	1.49 ± 0.44	2.39 ± 1.92 <sup>a</sup>	1.72 ± 1.17 <sup>a</sup>
<b>EMC</b>						
Affine	200	2.82 ± 3.18	4.42 ± 6.03	–	4.63 ± 3.01	8.03 ± 6.46
Qiao et al.	100	1.41 ± 0.76	2.24 ± 3.14	–	3.21 ± 1.85	5.42 ± 5.84
Proposed	100	1.54 ± 0.67	1.67 ± 1.38 <sup>a</sup>	–	2.67 ± 1.76 <sup>a</sup>	3.89 ± 4.00 <sup>a</sup>

<sup>a</sup> Represents a significant difference (at  $P = 0.05$ ) between Qiao et al. and the proposed algorithm.

TABLE VI. %95HD (mm) of the target volumes and the organs-at-risks of the HMC and EMC datasets for different registration methods.

Method	# It.	Prostate $\mu \pm \sigma$	Seminal vesicles $\mu \pm \sigma$	Lymph nodes $\mu \pm \sigma$	Rectum $\mu \pm \sigma$	Bladder $\mu \pm \sigma$
<b>HMC</b>						
Affine	200	3.97 ± 1.96	6.61 ± 3.70	3.12 ± 1.27	11.8 ± 5.98	12.5 ± 7.06
Qiao et al.	100	3.31 ± 1.16	4.59 ± 2.95	3.73 ± 1.02	10.4 ± 5.99	7.41 ± 6.85
Proposed	100	3.07 ± 1.30	3.82 ± 3.19 <sup>a</sup>	3.74 ± 1.02	8.66 ± 6.92 <sup>a</sup>	5.11 ± 4.38 <sup>a</sup>
<b>EMC</b>						
Affine	200	5.98 ± 6.19	8.11 ± 7.66	–	13.2 ± 6.88	21.3 ± 16.3
Qiao et al.	100	3.65 ± 2.31	4.80 ± 5.09	–	11.3 ± 6.77	16.5 ± 17.2
Proposed	100	3.93 ± 2.24	4.92 ± 5.13	–	10.4 ± 7.77	11.5 ± 12.5 <sup>a</sup>

<sup>a</sup> Represents a significant difference (at  $P = 0.05$ ) between Qiao et al. and the proposed algorithm.




FIG. 7. Scatter plot showing the effect of the bladder volume change between planning and daily scans of the HMC dataset on the performance of the proposed method in terms of mean surface distance for prostate (a), lymph nodes (b), and seminal vesicles (c). Red line represents the slice thickness. [Color figure can be viewed at wileyonlinelibrary.com]

findings in Ref. [50]. Inpainting gas pockets in the rectum enhanced the registration of the rectum as well as the seminal vesicles. The presence of these pockets was challenging for the registration due to the physical noncorrespondence between the daily and planning CT scans. Although the inpainting results from the GAN-inpainting network were more realistic than the simple-inpainting procedure, a similar performance with respect to the registration was obtained. Our explanation for this finding is that the mutual information similarity metric pays more attention to the overall intensity distribution and since the results from the simple-

inpainting were blended and smoothed with respect to its neighbors, it produces a similar histogram distribution to the GAN-inpainting and subsequently gives a similar registration performance.

The initialization of the registration algorithm on the bony structures is a crucial step for optimal performance, which is consistent with the reported results in Ref. [14]. Moreover, masking out the couch using a torso mask removed its disrupting effect on the registration. Increasing the number of iterations had a minimal effect on the registration performance while increasing the registration time. We found that




FIG. 8. Boxplot comparison between Qiao et al. and the proposed algorithm for image registration on the HMC dataset vs the number of iterations. The columns show the DSC, MSD, and 95%HD from left to right. Prostate, seminal vesicles, lymph nodes, rectum, and bladder are shown from top to bottom rows, respectively. The red box is the method from Qiao et al., the blue box is the proposed method, while the green box is an upper bound of the proposed method using manual daily contours. [Color figure can be viewed at wileyonlinelibrary.com]

the effect of adding a third registration step focussing on the rectal area boosted the performance regarding the rectum and seminal vesicles while there was no detrimental effect for the prostate, lymph nodes, and bladder.

In this study, we focused on the generalizability and robustness of the registration represented by performance on different datasets and the number of failed registrations according to geometrical and dosimetric criteria. This target is achieved through several steps. First, inpainting the rectum

gas pockets. Second, enhancing the CT image contrast by contrast clipping. Third, introducing the bladder segmentation with an optimized weights ( $\alpha = 0.05$  and  $0.01$ ) to steer the optimization problem to a better local minimum while avoiding overfitting to the bladder. Fourth, using a third stage for registration to focus on the rectum and consequently the seminal vesicles by using a dilated mask for the rectum. Overall, these steps yielded a more robust registration and substantially decreased the number of registrations with

TABLE VII. Percentage of registrations that meets the dose constraints for different registration iterations.

	$V_{95\%} \geq 98\%$				$V_{107\%} \leq 2\%$		
	Prostate	SV	LN	CSR	Prostate	SV	LN
Qiao et al.	83.8%	75.7%	97.9%	65%	100%	100%	100%
Proposed	85.8%	90.9%	98.9%	80%	100%	100%	100%

Conservative Success Rate (CSR) refers to the percentage of registrations for which all target volumes (the prostate, seminal vesicles, and lymph nodes) meet the dose constraints.




FIG. 9. Boxplot depicting the difference in dosimetric parameters of the manual delineations, calculated by using either the treatment plan based on the automated delineations or the manual delineations for 99 scans of the HMC dataset. [Color figure can be viewed at wileyonlinelibrary.com]

insufficient quality, especially for the seminal vesicles, rectum, and bladder. Improving the MSD for the seminal vesicles, which is an important target volume, resulted in a more precise targeting with potential benefits in terms of local control (lower probability of recurrences). Moreover, both the rectum and the bladder improved in terms of MSD and 95% HD, thereby avoiding treatment-induced complications after the therapy, so a higher probability of better quality of life after treatment. For the bladder, 11 of the 18 registrations with an MSD larger than the top whisker in Fig. 8 were belonging to two patients. For these two patients, the 3D-CNN achieved an average DSC of 0.65, explaining the suboptimal performance of the proposed method on these cases. From the CT images, no apparent reason for this was found. In terms of the geometric success rate defined by the number of registrations that achieved an MSD lower than 2 mm (slice thickness), the system achieved 97%, 93%, and 87% for the prostate, seminal vesicles, and lymph nodes, respectively. This compares to a success rate of 95%, 78%, and 86% for

Qiao et al., that is, especially improving the performance for the seminal vesicles. Moreover, the proposed system showed robustness to the change in bladder distension between planning and daily CT as shown in Fig. 7. The proposed registration method achieved quite similar results on the EMC and HMC datasets, except for the bladder. We suspect that this is partially due to the difference in bladder segmentation performance of the neural network, which was 82% on the EMC data and 88% on the HMC data. It could also be related to the affine registration results for the EMC dataset (Table V) being slightly less than HMC dataset. We visually checked the affine results and noticed that the field of view for some cases was cropped or zoomed. The average runtime for the proposed pipeline is 98.3 s for each registration at 100 iterations, comparing to 13.5 s reported by Qiao et al. However, the pipeline could be further optimized and adapted for GPU acceleration. For validating the clinical acceptance of the proposed algorithm, we considered  $V_{95\%} \geq 98\%$ ,  $V_{107\%} \leq 2\%$ , and CSR for dosimetric coverage for 99 registrations. All the scans meet the  $V_{107\%} \leq 2\%$  constraint. Fourteen of the 99 registrations (14.1%) did not directly meet the  $V_{95\%} \geq 98\%$  constraint for the prostate. After visual inspection of these failure cases, we found inconsistencies between the manual delineations for the planning and daily CT scans for seven cases. These cases had a  $V_{95\%}$  of  $92.5\% \pm 0.1\%$ , meaning that these cases were still close to be dosimetrically acceptable. The proposed algorithm improved the contouring quality and robustness, especially for the seminal vesicles, which directly increased the percentage of acceptable scans from 75.5% to 90.9% for this important target organ. These success rates imply that the automatically generated contours have the potential to be employed for online adaptive IMPT. Moreover, the typical 7 mm margins<sup>51</sup> may be replaced with smaller daily margins, which means delivering an effective dose with potentially less adverse effects.

The reported performance of the proposed pipeline could be further improved by correcting the inconsistency present in the manual contouring. Also, the weighting parameter  $\alpha$  could be selected automatically by introducing it as a trainable parameter. Moreover, the current 3D-CNN was trained using CT scans without contrast material, and therefore is unlikely to perform well on scans acquired with contrast. In case the clinical protocol dictates contrast-enhanced CT acquisitions, the network could be easily retrained. We may further investigate the effect on segmentation performance of CT clipping as a preprocessing step for the 3D-CNN for bladder segmentation. We also consider developing an end-to-end neural network to jointly optimize the registration and segmentation tasks to further improve the system robustness and accuracy.

## 5. CONCLUSION

In this study, we proposed a registration pipeline for automatic contour propagation for online adaptive IMPT of prostate cancer using the open source package *elastix* software in combination with deep learning. The proposed

pipeline achieved a geometrical success rate of 97%, 93%, and 87% for the prostate, seminal vesicles, and lymph nodes, respectively, for HMC dataset as well as 67% and 71% for the prostate and seminal vesicles, respectively, for ECM dataset. The HMC automatically propagated contours meet the dose coverage constraints in 86%, 91%, and 99% of cases for these targets. A Conservative Success Rate (CSR) of 80% was achieved, meaning that 80% of the automatically generated treatment plans can be directly used without manual correction. This recontouring showed a promise for generating daily treatment plans. Moreover, it showed a substantial improvement in the system robustness compared to a previous open source method, which means that more treatment plans can be directly used without manual correction, which is a crucial factor for enabling online daily adaptation, and thus, the use of relatively small treatment margins. Therefore, the proposed method could facilitate online adaptive proton therapy of prostate cancer. The authors have no relevant conflicts of interest to disclose.

## ACKNOWLEDGMENTS

This study was financially supported by Varian Medical Systems and ZonMw, the Netherlands Organization for Health Research and Development, grant number 104003012. The HMC dataset with contours was collected at Haukeland University Hospital, Bergen, Norway, and was provided to us by responsible oncologist Svein Inge Helle and physicist Liv Bolstad Hysing; they are gratefully acknowledged.

<sup>a)</sup> Author to whom correspondence should be addressed. Electronic mail: m.s.e.elmahdy@lumc.nl.

## REFERENCES

- National Cancer Society. Cancer Stat Facts: Prostate Cancer? <https://seer.cancer.gov/statfacts/html/prost.html>
- Kooy HM, Grassberger C. Intensity modulated proton therapy. *Br J Radiol.* 2015;88:20150195.
- Zhang M, Westerly DC, Mackie TR. Introducing an on-line adaptive procedure for prostate image guided intensity modulate proton therapy. *Phys Med Biol.* 2011;56:4947–4965.
- Lomax AJ. Intensity modulated proton therapy and its sensitivity to treatment uncertainties 1: the potential effects of calculational uncertainties. *Phys Med Biol.* 2008;53:1027–1042.
- Van den Heuvel F, Fugazzi J, Seppi E, Forman JD. Clinical application of a repositioning scheme, using gold markers and electronic portal imaging. *Radiother Oncol.* 2006;79:94–100.
- Hoogeman MS, van Herk M, de Bois J, Lebesque JV. Strategies to reduce the systematic error due to tumor and rectum motion in radiotherapy of prostate cancer. *Radiother Oncol.* 2005;74:177–185.
- Hansen EK, Bucci MK, Quivey JM, Weinberg V, Xia P. Repeat CT imaging and replanning during the course of IMRT for head-and-neck cancer. *Int J Radiat Oncol Biol Phys.* 2006;64:355–362.
- Boehmer D, Maingon P, Poortmans P, et al. Guidelines for primary radiotherapy of patients with prostate cancer. *Radiother Oncol.* 2006;79:259–269.
- Salembier C, Villeirs G, De Bari B, et al. ESTRO ACROP consensus guideline on CT- and MRI-based target volume delineation for primary radiation therapy of localized prostate cancer. *Radiother Oncol.* 2018;127:49–61.
- Jensen NKG, Mulder D, Lock M, et al. Dynamic contrast enhanced CT aiding gross tumor volume delineation of liver tumors: An interobserver variability study. *Radiother Oncol.* 2014;111:153–157.
- Riegel AC, Antone JG, Zhang H, et al. Deformable image registration and interobserver variation in contour propagation for radiation therapy planning. *J Appl Clin Med Phys.* 2016;17:347–357.
- Kupelian P, Willoughby T, Mahadevan A, et al. Multi-institutional clinical experience with the Calypso System in localization and continuous, real-time monitoring of the prostate gland during external radiotherapy. *Int J Radiat Oncol Biol Phys.* 2007;67:1088–1098.
- Jagt T, Breedveld S, van Haveren R, Heijmen B, Hoogeman M. An automated planning strategy for near real-time adaptive proton therapy in prostate cancer. *Phys Med Biol.* 2018;63:135017.
- Qiao Y. Fast Optimization Methods For Image Registration In Adaptive Radiation Therapy; 2017. PhD thesis, chapter 5. Leiden University Medical Center. [http://elastix.isi.uu.nl/marius/downloads/2017\\_t\\_Qiao.pdf](http://elastix.isi.uu.nl/marius/downloads/2017_t_Qiao.pdf)
- Thor M, Petersen JBB, Bentzen L, Hyer M, Muren LP. Deformable image registration for contour propagation from CT to cone-beam CT scans in radiotherapy of prostate cancer. *Acta Oncol.* 2011;50:918–925.
- Woerner AJ, Choi M, Harkenrider MM, Roeske JC, Surucu M. Evaluation of deformable image registration-based contour propagation from planning CT to cone-beam CT. *Technol Cancer Res Treat.* 2017;16:801–810.
- Thörnqvist S, Petersen JBB, Hyer M, Bentzen LN, Muren LP. Propagation of target and organ at risk contours in radiotherapy of prostate cancer using deformable image registration. *Acta Oncol.* 2010;49:1023–1032.
- Elmahdy MS, Jagt T, Yousefi S, et al. Evaluation of multi-metric registration for online adaptive proton therapy of prostate cancer. In: Klein S, Staring M, Durrleman S, Sommer S, eds. *Biomedical Image Registration*. Cham:Springer International Publishing;2018:94–104.
- Mangar S, Coffey J, McNair H, et al. Prostate radiotherapy: evaluating the effect of bladder and rectal changes on prostate movement—a CT study. *Trends Med Res.* 2006;1:55–65.
- Wolterink JM, Leiner T, Viergever MA, Igum I. Dilated convolutional neural networks for cardiovascular MR segmentation in congenital heart disease. In: Zuluaga M, Bhatia K, Kainz B, Moghari M, Pace D, eds. *Reconstruction, Segmentation, and Analysis of Medical Images*. Cham: Springer International Publishing; 2017:95–102.
- Çiçek, Ö, Abdulkadir A, Lienkamp SS, Brox T, Ronneberger O. 3D U-Net: learning dense volumetric segmentation from sparse annotation. In: Zuluaga M, Bhatia K, Kainz B, Moghari M, Pace D, eds. *Medical Image Computing and Computer-Assisted Intervention MICCAI*. Cham: Springer International Publishing; 2016:424–432.
- Milletari F, Navab N, Ahmadi S-A. V-Net: Fully Convolutional Neural Networks for Volumetric Medical Image Segmentation. In: 2016 Fourth International Conference on 3D Vision (3DV); 2016.
- Kingma Diederik P, Jimmy BA. Adam: A Method for Stochastic Optimization. arXiv:1412.6980; 2014.
- Coltuc D, Bolon P, Chassery J-M. Exact histogram specification. *IEEE Trans Image Process.* 2006;15:1143–1152.
- Gao S, Zhang L, Wang H, et al. A deformable image registration method to handle distended rectums in prostate cancer radiotherapy. *Med Phys.* 2006;33:3304–3312.
- Foskey M, Davis B, Goyal L, et al. Large deformation three-dimensional image registration in image-guided radiation therapy. *Phys Med Biol.* 2005;50:5869–5892.
- Litjens G, Kooi T, Bejnordi BE, et al. A survey on deep learning in medical image analysis. *Med Image Anal.* 2017;42:60–88.
- Goodfellow I, Pouget-Abadie J, Mirza M, et al. *Generative Adversarial Networks*. arXiv:1406.2661; 2014.
- Xue Y, Xu T, Zhang H, Long LR, Huang X. SegAN: adversarial network with multi-scale L1 loss for medical image segmentation. *Neuroinformatics.* 2018;16:383–392.
- Nie D, Trullo R, Lian J, et al. Medical Image Synthesis with Context-Aware Generative Adversarial Networks. In: *Lecture Notes in Computer Science*. Cham: Springer International Publishing; 2017:417–425.
- Hu Y, Gibson E, Ghavami N, et al. Adversarial Deformation Regularization for Training Image Registration Neural Networks. arXiv:1805.10665; 2018.

32. Yang Q, Yan P, Zhang Y, et al. Low Dose CT Image Denoising Using a Generative Adversarial Network with Wasserstein Distance and Perceptual Loss. arXiv:1708.00961; 2017.
33. Yu J, Lin Z, Yang J, et al. Generative Image Inpainting with Contextual Attention. arXiv:1801.07892; 2018.
34. Iizuka S, Simo-Serra E, Ishikawa H. Globally and locally consistent image completion. *ACM Trans Graph*. 2017;36:1–14.
35. Rodriguez-Vila B, Garcia-Vicente F, Gomez EJ. Methodology for registration of distended rectums in pelvic CT studies. *Med Phys*. 2012;39:6351–6359.
36. Klein S, Staring M, Murphy K, Viergever MA, Pluim J. elastix: a toolbox for intensity-based medical image registration. *IEEE Trans Med Imaging*. 2010;29:196–205.
37. Staring M, Bakker ME, Stolk J, Shamonin DP, Reiber JHC, Stoel BC. Towards local progression estimation of pulmonary emphysema using CT. *Med Phys*. 2014;41:21905.
38. Huizinga W, Klein S, Poot DHJ. Fast multidimensional B-spline interpolation using template metaprogramming. In: Ourselin S, Modat M. eds. *Biomedical Image Registration*. Cham: Springer International Publishing; 2014:11–20.
39. Qiao Y, van Lew B, Lelieveldt BPF, Staring M. Fast automatic step size estimation for gradient descent optimization of image registration. *IEEE Trans Med Imaging*. 2016;35:391–403.
40. van der Wielen GJ, Mutanga TF, et al. Deformation of prostate and seminal vesicles relative to intraprostatic fiducial markers. *Int J Radiat Oncol Biol Phys*. 2008;72:1604–1611.
41. Muren LP, Wasb E, Helle SI, et al. Intensity-modulated radiotherapy of pelvic lymph nodes in locally advanced prostate cancer: planning procedures and early experiences. *Int J Radiat Oncol Biol Phys*. 2008;71:1034–1041.
42. Breedveld S, Storchi PRM, Voet PWJ, Heijmen BJM. iCycle: integrated, multicriterial beam angle, and profile optimization for generation of coplanar and noncoplanar IMRT plans. *Med Phys*. 2012;39:951–963.
43. van de Water S, Kraan AC, Breedveld S, et al. Improved efficiency of multi-criteria IMPT treatment planning using iterative resampling of randomly placed pencil beams. *Phys Med Biol*. 2013;58:6969–6983.
44. van de Water S, Kooy HM, Heijmen BJM, Hoogeman MS. Shortening delivery times of intensity modulated proton therapy by reducing proton energy layers during treatment plan optimization. *Int J Radiat Oncol Biol Phys*. 2015;92:460–468.
45. Breedveld S, Storchi PRM, Heijmen BJM. The equivalence of multi-criteria methods for radiotherapy plan optimization. *Phys Med Biol*. 2009;54:7199–7209.
46. Voet PWJ, Dirks MLP, Breedveld S, Fransen D, Levendag PC, Heijmen BJM. Toward fully automated multicriterial plan generation: a prospective clinical study. *Int J Radiat Oncol Biol Phys*. 2013;85:866–872.
47. Abadi M, Barham P, Chen J, et al. TensorFlow: Large-Scale Machine Learning on Heterogeneous Distributed Systems. arXiv:1603.04467; 2017.
48. Cha KH, Hadjiiski L, Samala RK, Chan H-P, Caoili EM, Cohan RH. Urinary bladder segmentation in CT urography using deep-learning convolutional neural network and level sets. *Med Phys*. 2016;43:1882–1896.
49. Zhou X, Ito T, Takayama R, Wang S, Hara T, Fujita H. Three-dimensional ct image segmentation by combining 2D fully convolutional network with 3D majority voting. In: Carneiro G. et al. eds. *Deep Learning and Data Labeling for Medical Applications*. Cham: Springer International Publishing; 2016:111–120.
50. Men K, Dai J, Li Y. Automatic segmentation of the clinical target volume and organs at risk in the planning CT for rectal cancer using deep dilated convolutional neural networks. *Med Phys*. 2017;44:6377–6389.
51. Thörnqvist S, Bentzen L, Petersen JBB, Hysing LB, Muren LP. Plan robustness of simultaneous integrated boost radiotherapy of prostate and lymph nodes for different image-guidance and delivery techniques. *Acta Oncol*. 2011;50:926–934.