

REVIEW

Open Access



Novel endoscopic optical diagnostic technologies in medical trial research: recent advancements and future prospects

Zhongyu He¹, Peng Wang¹ and Xuesong Ye^{1,2*}

*Correspondence:

yexuesong@zju.edu.cn

¹ Biosensor National Special Laboratory, College of Biomedical Engineering and Instrument Science, Zhejiang University, Hangzhou 310027, People's Republic of China

Full list of author information is available at the end of the article

Abstract

Novel endoscopic biophotonic diagnostic technologies have the potential to non-invasively detect the interior of a hollow organ or cavity of the human body with subcellular resolution or to obtain biochemical information about tissue in real time. With the capability to visualize or analyze the diagnostic target in vivo, these techniques gradually developed as potential candidates to challenge histopathology which remains the gold standard for diagnosis. Consequently, many innovative endoscopic diagnostic techniques have succeeded in detection, characterization, and confirmation: the three critical steps for routine endoscopic diagnosis. In this review, we mainly summarize researches on emerging endoscopic optical diagnostic techniques, with emphasis on recent advances. We also introduce the fundamental principles and the development of those techniques and compare their characteristics. Especially, we shed light on the merit of novel endoscopic imaging technologies in medical research. For example, hyperspectral imaging and Raman spectroscopy provide direct molecular information, while optical coherence tomography and multi-photo endomicroscopy offer a more extensive detection range and excellent spatial-temporal resolution. Furthermore, we summarize the unexplored application fields of these endoscopic optical techniques in major hospital departments for biomedical researchers. Finally, we provide a brief overview of the future perspectives, as well as bottlenecks of those endoscopic optical diagnostic technologies. We believe all these efforts will enrich the diagnostic toolbox for endoscopists, enhance diagnostic efficiency, and reduce the rate of missed diagnosis and misdiagnosis.

Keywords: Endoscopy, Optical diagnostic technologies, Endomicroscopy, Biomedical spectroscopy

Introduction

An endoscope is used in medicine to examine the interior of a hollow organ or cavity of the human body [1, 2]. Developments in optical imaging technology continue to promote the revolution of endoscopy. Currently, electronic chromoendoscopic techniques such as narrow-band imaging (NBI, Olympus, Japan) [3–5], linked color imaging (LCI, Fujifilm, Japan) [6–8] and i-scan (Pentax, Japan) [9–11] provide wide-field high-contrast



© The Author(s) 2021. This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

video images to help endoscopists find mucosal abnormalities especially precancers [12]. However, comparing with traditional histopathologic biopsy, electronic chromoendoscopy lacks the capacity to provide more detailed morphologic information about tissue and cell for accurately confirming or staging cancer [13]. Furthermore, although existing commercial endomicroscopy including endocytoscopy (EC, Olympus, Japan) [14] or probe-based confocal laser endomicroscopy (PCLE, Mauna KeaTech, France) [15] can image at cellular information, their detection range is limited to the surface of the mucosa (< 1 mm in depth) which may cause the missed diagnosis of the hidden lesions in deep tissue [16].

Therefore, many studies have been conducted in numerous research centers on novel endoscopic diagnostic technologies to provide more accurate diagnostic information. The advent of these technologies in medical trial research, such as photoacoustic endoscopy (PAE), Raman spectroscopy (RS), two-photon excited fluorescence (TPEF) imaging has opened a new era and created tremendous opportunities for the enhanced identification and biochemical characterization of diseases. These modalities also have the potential to allow non-invasive in vivo “optical biopsy” which differentiates areas of similar clinical characteristics, hence challenging the ex vivo histology which is the only way for definitive cancer diagnosis [17]. In addition, these endoscopic techniques own unprecedented temporal-spatial resolution of imaging with innovative mechanisms such as photoacoustics, optical coherent tomography, and multi-photo effect. By utilizing these novel tissue–photon interacting mechanisms, these techniques can guide biopsies by accessing the subtle mucosal/submucosal abnormalities with higher contrast, better resolution, and penetration depth. Hence, they cut down the number of biopsy times, costs, and risks for patients.


However, most of these technologies are not mature, so they have been neither officially launched into the market nor put into clinical use in the hospital so far. Thus, well-organized large-scale animal experiments and human clinical trials should be executed to further validate and standardize these techniques before approval of the clinical application.

For this review, the PubMed and Springer Link literature search was systematically performed for studies published from 1983 to 2019 about emerging endoscopic imaging and sensing technologies by using the search terms “Endoscopy,” “Endomicroscopy,” “Endoscopic medical imaging,” “Image-enhanced endoscopy,” “Hyperspectral imaging,” “Fluorescence lifetime imaging,” “Photoacoustic endoscopy,” “Optical coherent tomography,” “Coherent anti-Stokes Raman scattering,” “Multi-photon imaging,” “Diffuse reflectance spectroscopy,” “Light-scattering spectroscopy,” “Angle-resolved low coherence interferometry,” “Raman spectroscopy,” “Super-resolution microscopies” and “In vivo clinical trial research.” Further articles were obtained through the review of the quoted references from the selected reference articles. Only full manuscripts and case reports published in English were collected. We included studies that firstly proposed novel modalities of in vivo endoscopic optical detection or that significantly improved technical performances (such as spatiotemporal resolution, real-time performance, SNR, image contrast, etc.) of existing technologies in the last decade. Additionally, we specially presented some technologies that have been tested in human clinical trials in various medical specialties. However, for new or emerging technology where related pieces

of literature are scarce, articles with other study designs were also included. This review intends to provide an accurate and comprehensive synopsis, illustrated by a few examples with emphasis on their fundamental principles and technical parameters, of where and how the novel endoscopic optical diagnostic technologies in medical trial research have already contributed to, or probably will contribute to the way of endoscopic detection and treatment. Thus, this article could serve as a guideline for future researchers and promote cooperation between researchers in different disciplines to accelerate the development and clinical application of endoscopic technologies.

Classifications on novel technologies in medical trial research

We divide novel endoscopic optical diagnostic technologies in medical trial research into three categories: endoscopic biochemical optical imaging techniques, endomicroscopy, and optical sensing techniques (Fig. 1a). Label-free biochemical imaging technologies are mainly about mapping biochemical changes between healthy and abnormal tissue with high definition, while endomicroscopy reaches cellular/subcellular resolution. Optical sensing stands for detecting the nuclear size distributions or biomolecular information in tissue without the true representation of the targeted



tissue’s morphology. In this review, we will in turn introduce and summarize these technologies according to this classification.

As one of the novel endoscopic optical imaging techniques, hyperspectral imaging currently enhances image contrast by altering the illumination of conventional endoscopes through wavelength dispersion devices [18]. In contrast, other techniques are implemented in the form of a miniaturized probe, which is feasible for easy deployment in more specific medical scenarios. Probe-based imaging systems have two available operation modes:

(1) Through the working channel to the distal end of the therapeutic endoscope. In this way, endoscopists can simultaneously obtain information from both the wide-field imaging system and the probe.

(2) Working alone. Especially, when endoscopists detect the narrow-body cavities such as pancreatic duct and biliary tract.

Moreover, these techniques can also be categorized based on the way how light interacts with tissues while detecting, such as reflection, absorption, elastic scattering, inelastic scattering, and fluorescence [14] (Fig. 1b). Table 1 summarizes all the endoscopic diagnostic techniques in this review.

Endoscopic biochemical optical imaging

Hyperspectral imaging

Hyperspectral imaging (HSI) is a hybrid modality for optical diagnostics. It obtains spectroscopic data from an image and renders it in image form. With conventional spectroscopy, the spectral range can be recorded continuously, but for only a single analyte

Table 1 Summary of each endoscopic diagnostic technology

Technique	Excitation spectra	Frame rate (Hz)	Achievable spatial resolution (μm)	Penetration depth (mm)	Functional information	Real morphologic image	Whether the probe is in direct contact with the tissue surface while operating
HSI	Visible & Near-IR	> 30	< 250	1	✓	✓	–
FLIM	UV-Blue	20–120	< 50	< 0.1	✓	✓	Non-contact
PAE	Visible & Near-IR	2–5	10	> 1	✓	✓	Contact
OCT	Near-IR	< 50	< 3	2–4	✗	✓	Non-contact
CARS	Near-IR	0.01–0.2	< 3	> 1	✓	✓	Non-contact
TPEF/SHG	Near-IR	1–8	< 1	0.2–0.3	✓	✓	Contact or non-contact
3PEF/THG	Near-IR	0.05–0.1	1	0.2–0.7	✓	✓	Contact or non-contact
DRS	Visible	0.02–2	–	< 1	✗	✗	Contact
LSS	Visible	0.02–0.05	–	< 1	✗	✗	Non-contact
a/LCI	Near-IR	1–4	–	< 0.5	✗	✗	Contact
RS	Near-IR	0.1–10	–	> 1	✓	✗	Contact

HSI: hyperspectral imaging; FLIM: fluorescence lifetime imaging microscopy; PAE: photoacoustic endoscopy; OCT: optical coherent tomography; CARS: coherent anti-Stokes Raman scattering; TPEF: two-photon excited fluorescence; 3PEF: three-photon excited fluorescence; SHG: second-harmonic generations; THG: third-harmonic generations; DRS: diffuse reflectance spectroscopy; LSS: light-scattering spectroscopy; a/LCI: angle-resolved low coherence interferometry; RS: Raman spectroscopy

spot. By combining the concept of spectroscopy and digital imaging, the HSI allows the recording of the entire emission spectrum for every pixel on the entire image [19–21]. The HSI systems can provide a 3D “data tube” of spatial and spectral information of the whole image at each wavelength of interest.

On the other hand, HSI can be used for the spatial mapping of tissue morphology and physiology [22]. Likewise, it is possible to visualize the entire spatial information for a given wavelength. With spatial information, the source of each spectrum on samples can be located, which makes it possible to probe the light interactions with pathology more completely. Thus, the spectrum emitted from each pixel in the images enables HSI to identify various pathological conditions [18].

In the HSI system, visible and near-infrared (400–2500 nm) light [23, 24] is used to transmit information about molecular expression in healthy and diseased tissues. In this way, HSI generally covers a contiguous portion of the whole light spectrum with more spectral bands (up to a few hundred) and higher spectral resolution than multispectral imaging (such as RGB color cameras) [18]. Therefore, HSI may capture subtle spectral differences under different pathological conditions, while conventional multispectral imaging may miss important spectral information for diagnosis. Furthermore, a spectral signature can be constructed as a curve that links light with the target region. Then, the spectral signature is utilized as an indicator of the different biochemical components from various tissues which helps to discriminate healthy from abnormal tissue [25, 26]. In other words, HSI can identify various pathological states by extracting the spectral reflectance curve information of the pixels in the image and analyzing the tissue situation accordingly. Thus, HSI is promising as an ancillary and non-invasive method for distinguishing normal, precancerous, and cancerous cells [27].

To prove the applicability of hyperspectral imaging for the *in vivo* discrimination between healthy and dysplastic tissues, Arnold et al. [28] developed a hyperspectral video endoscopy system with electronically controlled acousto-optical tunable filters (AOTF) to differentiate larynx tumor through a combined classifier. Moreover, Gerstner et al. [29] implemented the HSI into endoscopy to visualize human larynx *in vivo* by utilizing a tunable light source for illumination. Moreover, Gu et al. [30] developed a flexible gastroscopy system capable of obtaining *in vivo* hyperspectral images of different types of stomach mucous diseases. Certain illumination bands, generated with a filter wheel, were assigned to color components of an enhanced image of the object with an appropriate band selection algorithm based on the dependent weight of information. Thus, images with higher color tone contrast could be attained to enhance visualization of gastric lesions. Furthermore, HSI can contribute to a more accurate determination of tumor boundaries, facilitating a complete resection of the tumor tissue [26, 31].

However, HSI is confronting notable challenges. First, the intra- and inter-patient variability of hyperspectral imaging data must be handled [24]. Second, the acquisition speed of high-resolution HSI datasets. Up to now, the HSI systems can not realize real-time detection. Third, more advanced classification algorithms are required to enable better discrimination among healthy, premalignant, and malignant tissue. Lastly, more precise delineation of cancer margins for image-guided biopsy and surgery is demanded [25].

Fluorescence lifetime imaging


The fluorescence intensity distribution is not only proportional to the concentration of the fluorophores in the tissue but also depends on fluorescence lifetime. The fluorescence lifetime refers to the average time that the molecules remain in the excited electronic energy state before returning to its ground state via the emission of fluorescent light, which is somewhat similar to the radioactive decay of an unstable atom. Thus, fluorescent light from a sample does not instantly cease when the excitation light is extinguished. Instead, it decays away over a period of several nanoseconds as excited fluorophore molecules in the sample return to their ground state. The fluorescent molecule in the excited state emits fluorescence. It releases energy during the de-excitation to the ground state. The attenuation of the fluorescence intensity of the excited state fluorophore can be expressed as a single exponential function by a mathematical expression as follows:

$$I(t) = I_0 e^{-\frac{t}{\tau}}. \quad (1)$$

In the Eq. (1), $I(t)$ is the intensity measured at time t after the sample is excited by the light pulse; I_0 is the intensity at $t=0$; τ is the average fluorescence lifetime and is a characteristic value of the molecule, and is defined as the time required for the fluorescence intensity to decay to $1/e$ (37%) of the initial. The emission of fluorescence is a statistical process; few fluorescent molecules emit fluorescence just at the moment of τ (fluorescence lifetime). Hence, the fluorescence lifetime only reflects the time required for the fluorescence intensity to decay to its initial value of $1/e$.

Fluorescence lifetimes are affected by the transfer of energy from a fluorophore to its surroundings (non-radiative decay), so FLIM is capable of mapping tissue biochemical changes, including pH, O_2 , Ca^{2+} , nicotinamide adenine dinucleotide (NAD), collagen, etc. [32, 33]. Therefore, fluorescence lifetime can be invoked as a contrast mechanism for imaging, and this is achieved in fluorescence lifetime imaging microscopy (FLIM), for which the fluorescence lifetime is determined for each pixel in a field of view [34]. The fluorescence lifetime can be measured in the time and frequency domain [35]. Time-domain measurements are the most intuitive because they involve direct measurement of the decay of fluorescence intensity over time. Moreover, time-domain fluorescence lifetime measurement is mainly implemented by four conventional methods, including time-correlated single-photon counting (TSCPC) [36–38] time-gated detection (Fig. 2) [39, 40], streak-FLIM [41] and pulse sampling techniques [42]. By contrast, frequency-domain FLIM uses modulation technique, which measures the phase shift and demodulation during excitation with sinusoidally modulated light to perform FLIM [43].

In the last few years, several endoscopic systems based on FLIM designs have been proposed [38, 39, 44–46]. Thereinto, McGinty et al. [47] developed a compact, clinically deployable instrumentation which achieved wide-field fluorescence lifetime images of unprecedented clarity to intrinsic tissue autofluorescence. Statistically, significant contrast was observed between cancerous and healthy colon tissue with excitation at 355 nm. Additionally, FLIM image was acquired ex vivo at near video rate, which was an essential step towards real-time FLIM for screening and image-guided biopsy applications. In 2013, Sun et al. [48] proposed an in vivo clinically compatible FLIM system, which consisted of a rigid fiber-bundle endoscope proving a field



of view of 4 mm. Intraoperative imaging of head and neck squamous cell carcinoma was performed in 10 patients at 337-nm excitation and fluorescence collected in the 435–485 nm range to facilitate the margin demarcation of the tumor. Shorter average lifetime (1210 ± 40 ps) was exhibited from head and neck squamous cell carcinoma than the surrounding healthy tissue (1490 ± 60 ps).

More recently, Ning et al. [49] proposed a wide-range pH-sensitive Yb^{3+} porphyrinate-based molecular probe, which utilized time-resolved FLIM in the near-infrared region of 900–1700 nm. The Yb^{3+} probe showed increasing NIR emission and lifetime with pKa values of ca. 6.6 from pH 9.0 and 5.0. Also, the system displayed an elongated lifetime from about 135 to 170 ns at lower pH values (5.0–1.0) due to aggregation and reduced exposure to water at low pH values. What is more, the probe could monitor a wide range of in vivo gastrointestinal pH values in mice models, which showed that lifetime contrast might be necessary for preclinical imaging. Furthermore, due to the long lifetime (millisecond scale) of lanthanide, the Yb^{3+} probe could quantitatively, dynamically, and in situ monitor real-time pH changes in gastric cells, with deep penetration, good reversibility, and high spatiotemporal resolution.

Further studies will be necessary to optimize and refine the FLIM technique to maximize the contrast available for a given tissue or tumor. And more promising early findings should be replicated in vivo to early approval of clinical applications [47].

Endoscopic imaging techniques

Unlike wide-field imaging techniques that detect morphologic changes visible to the unaided eye, novel endomicroscopy facilitates the visualization of cellular-scale or sub-cellular-scale objects in tissue, such as goblet cell, epithelium cells, organelles (nucleus and cytoplasm), etc. Compared with commercial endomicroscopy, these novel emerging techniques can provide high-definition three-dimensional images with better penetration. Thus, they have the potential to challenge traditional ex vivo histology by executing accurate in vivo “optical biopsy”. Table 2 summarizes the differences between wide-field technologies and endomicroscopy.

Photoacoustic endoscopy

Currently, endoscopic ultrasound (EUS) [50] is the only clinically available tomographic endoscopic tool used to diagnose various diseases [51, 52]. EUS cannot receive high-resolution vasculature information of the suspicious region, owing to the limited detectable depth and the operating mechanism.

Although photoacoustic spectroscopy and simple imaging were developed in the 1970s, photoacoustic imaging only recently became famous in biomedical research [53]. The advent of photoacoustic endoscopy (PAE) aims to overcome the EUS’s limitations by adapting the photoacoustic effect. Photoacoustics can be described as a laser-induced ultrasound [54]. Short light pulses (nanosecond range) are absorbed by tissue absorbers accordingly (such as endogenous contrast: hemoglobin, melanin and exogenous contrast: methylene blue [55]). Then a transient temperature increase is generated, resulting in local thermoelastic expansion effect, which gives rise to ultrasonic waves [56]. After that, the ultrasonic wave signal is recorded by nearby detectors [57]. Figure 3 illustrates the imaging process of PAE. Photoacoustic imaging has the primary advantages


Table 2 The comparison of properties between the endoscopic wide-field imaging and endomicroscopy

	Wide-field imaging	Endomicroscopy
Resolution	Around 100 μm	0.5–20 μm
Field of view	80–140°	∅0.3–∅3 mm
Detection target	Macroscopical objects in tissue	Tissue structures in cellular and subcellular scale
Corresponding techniques	Traditional white-light imaging Narrow-band imaging Linked color imaging i-scan	Endocytometry Confocal laser endomicroscopy Photoacoustic endoscopy Optical coherent tomography Coherent anti-Stokes Raman scattering Multi-photon endomicroscopy

of beneficial penetration (from millimeters to centimeters) and high spatial resolution because it overcomes the high degree scattering of optical photons in biological tissue.

Compared with NBI, CLE, and OCT [58], PAE provides functional optical contrast with high spatial resolution while maintaining the benefits of greater penetration depth of traditional ultrasound endoscopy [59]. Particularly, PAE embodies photoacoustic tomography (PAT), which is cross-sectional or three-dimensional imaging using the photoacoustic effect. Thus, PAE can offer in vivo 3D volumetric images of biological tissues with high spatial resolution and stark tissue-optical contrast [60]. Thus, characteristics of the targeted area, like architectural vascular features and dynamic changes of tumors, can be clearly observed from the 3D images [61].

The prototype PAE was based on a scanning mirror system that deflected both the light and the ultrasound. The mirror was integrated into a mini-probe that could be inserted into the working channel of the endoscope. Viator et al. [54] first developed a photoacoustic endoscopic probe for 1D sensing. Later, Lim et al. [62] conducted an ex vivo clinical trial using photoacoustic imaging (PAI) in patients undergoing endoscopic mucosal resection (EMR). The esophageal microvascular pattern could be characterized by creating a 3D reconstruction of the full ex vivo tissue volume. Moreover, many in vivo preclinical trials on animal tissue have been finished recently [63–67], which demonstrated the integrated in vivo imaging capability of the PAE-EUS system. He et al. [68] proposed a hybrid optical-resolution (OR) and acoustic-resolution (AR) photoacoustic endoscope, which was appropriate for improving the endoscopic depth and resolution range. In this system, the optical resolution achieved the order of 13 μm , which is feasible to distinguish the numerous smaller vessels ex vivo. However, the smaller vessels were not visible on the AR mode, with a reported lateral resolution 250 μm . In addition, Yang et al. [67] initially created an optical-resolution photoacoustic endomicroscopy (OR-PAEM) system. The transverse resolution and radial resolutions were 10 μm and 50 μm , respectively, with an imaging frame rate up to 2 Hz. The typical imaging depth of their system was 1 mm. However, the diameter of the probe was 3.4 mm, larger than



the work proposed by Bai et al. [69] whose probe-diameter was 1.1 mm with transverse resolution as fine as 19.6 μm . Furthermore, Lin et al. [70] used a flexible coil to transmit the rotational torque from the rotary stage, which enabled a more extensive 360° field-of-view imaging at a 5-Hz B-scan rate in vivo.

To conclude, PAE has great potential for in vivo endoscopic applications, such as pre-cancer detection, accurate diagnosis of submucosal abnormalities, and in situ characterization of diseased tissues. Endogenous or exogenous contrast agents may improve the ability of endoscopic imaging, resulting in the earlier and more accurate detection of malignant and premalignant lesions. The wealth of experience gained in preclinical studies as well as recent technological progress provides a solid foundation for photoacoustics to advance into mainstream use [55].

Nonetheless, further refinement of photoacoustic imaging is necessary for broader dissemination, including stable laser systems, user-friendly imaging devices, algorithms and compatible software that can robustly analyze imaging data to extract diagnostically relevant quantity. Moreover, the most apparent limitation remains the penetration depth of light. The near-infrared light is not sufficient to visualize targets beyond a few centimeters under the tissue surface [55]. Additionally, the imaging speed of PAE systems is relatively low, so there is still a long way to implement real-time detection.

Optical coherent tomography

Initially proposed in the early 1990s [71], optical coherent tomography (OCT) is a promising new tomography technology. OCT is another considerable breakthrough for medical imaging after computerized tomography (CT) and nuclear magnetic resonance (MRI). Similar in principle to ultrasonography [72], the OCT system based on the technique of low coherence interferometry. A light source emits low coherence broadband signals, while a beam splitter splits the light into two identical beams, one directed to the tissue and the other to a mirror. Then, the system utilizes an interferometer to combine the back reflection or scattering signal coming back from the tissue and the signal from the mirror [71]. The interference between the two beams is measured, then an image is created by analyzing single points in different depth layers within one axis [73].

One of the OCT's most significant advantage is the axial resolution (δ_z), which mainly depends on the coherence length of the light source (l_c). Higher axial resolution can be obtained when using a common objective lens with a small numerical aperture. For a Gaussian light source, it can be expressed as follows:

$$\delta_z = \frac{l_c}{2} = 0.44 \frac{\lambda_0^2}{\Delta\lambda} \quad (2)$$

According to formula (2), under the condition that the spectral bandwidth (λ_0) of the light source is constant, there is a simple inverse relationship between the axial resolution and the bandwidth of the light source ($\Delta\lambda$). But in practice, the axial resolution of OCT is ultimately limited by the available broadband light sources, detectors, and passive optical devices (such as couplers, circulators) and dispersion effects in optical systems. However, with the continuous progress of technology, it can be found that the axial resolution of OCT is also continuously improved.

Moreover, as long as the scattering effect decreases with increasing wavelength, longer wavelengths in the near-infrared region (from 800 to 2500 nm) can be utilized in OCT to realize high-resolution images with deeper penetration (2–4 mm) instead of the visible-range wavelength that everyday optical microscopy use. Similar to traditional optical microscopes, the lateral resolution of the OCT depends on the focus state of the detection beam, which is characterized by the numerical aperture (f) of the objective lens:

$$\delta_x = \frac{\lambda_0}{\pi} = \frac{f}{D} \quad (3)$$

where D is the diameter of the light spot at the focusing point of the sample arm. Higher numerical aperture microscope objectives can be used to improve the lateral resolution of the OCT system.

By rotationally scanning, the label-free two-dimensional (2D) or three-dimensional (3D) structure image of the biological tissue can be obtained. Thus, the non-contact, non-invasive tomography of living tissue is realized [74]. Any physical property that alters the amplitude, phase and polarization properties of the sample light can be used to extract diagnostic information.

Acted with low-numerical-aperture lenses, the mechanical components of OCT can be miniaturized. Furthermore, it is easy to combine with endoscopes, thanks to the no-touch and fiber-based feature of broadband low coherence light source and Michelson interferometers [75]. Therefore, the miniaturizing design of the scanning probe determines the application and development of the OCT-based endoscopy. Combining the advantages of both OCT and conventional endoscopy, OCT-based endoscopy eliminates the shortcomings of OCT: (1) cannot detect human internal organs in vivo, and (2) cannot diagnose small lesions as well as early carcinomatous change [76].

The first endoscopic OCT scanning probe was proposed by Tearney et al. [77]. In the subsequent works, the scan speed of the probes was improved (up to 50 frames per second by Wang et al. [78]), and the diameter of the probe decreased significantly (even up to 160 μm by Lee et al. [79]). In addition, the lateral and axial resolution of endoscopic OCT image is continuously improving [80]. Pahlevaninezhad et al. [81] recently developed an endoscopic OCT catheter that connecting to a Fourier-domain OCT system [82]. By integrating a metalens which can modify the phase of incident light at sub-wavelength level, near-diffraction-limited imaging (with a resolution of 6.37 μm in tangential and 6.53 μm in sagittal) was achieved through negating non-chromatic aberrations. Additionally, the trade-off between transverse resolution and depth of focus was eased (effective depth of focus was achieved as 211 μm in tangential and 315 μm in sagittal).

Furthermore, by effectively reducing the operating central wavelength in the OCT system to 800 nm, compared with the previous work at 1300 nm [83, 84], Yuan et al. [85] developed an ultracompact (520 μm in outer diameter and 5 mm in rigid length) and super-achromatic endoscopic OCT microprobe made with a built-in monolithic fiber-optic ball lens. Their system could image sheep small airways in vivo with 1.7- μm axial resolution and 6- μm transverse resolution in real-time (about 5 fps) (Fig. 4). Later on, Mavadia-Shukla et al. [86] presented a distal scanning OCT probe delivered through a conventional endoscope's working channel to perform in vivo imaging at




Fig. 4 In vivo endobronchial imaging of sheep small airways. **a** Cut-away view of a reconstructed three-dimensional (3D) image of an 8 mm-long sheep small airway with the en face projection view of blood vessels overlaid on the inner surface of the three-dimensional (3D) image. **b** Representative in vivo 2D image of the peripheral sheep airway corresponding to the cross section indicated with the green dashed line in **a**. **c** 3× enlarged view of the area boxed with red dashed lines in **b**. **d** Representative in vivo 2D image of the sheep small airway, which is relatively more towards the proximal side (indicated with the blue dashed line in **a**). **e** 3× enlarged view of the area boxed with red dashed line in **d**. **f** Corresponding hematoxylin and eosin (H&E) histology. A: alveoli; BV: blood vessel; C: cartilage; EP: epithelium; LP: lamina propria; S: smooth muscle. All scale bars are 250 μm . Adapted with permission from Ref. [85]. Copyright 2017 Springer Nature Publishing

50 frames per second while maintaining an ultra-high axial resolution (2.4 μm) with higher contrast.

At present, the endoscopic OCT system has only been adopted into clinical practice in the cardiovascular department. The existing clinically available OPTIS System (Abbott, USA) offers physicians an efficient way to optimize percutaneous coronary intervention (PCI) and to image coronary arteries. However, instead of characterizing lesions, this OCT system mainly focuses on helping physicians improve stent size and placement decisions according to the measured shape of coronary artery due to its relatively low imaging resolution. Ding et al. [87] proposed a lensless OCT probe that also had the potential to be used in cardiovascular investigations. The probe is only 340 μm in diameter and 6.37 mm in length. It was composed of a segment of large-core multimode fiber, a segment of tapered multimode fiber and a length of single-mode fiber. A controllable output beam can be designed by a simple adjustment of its probe structure parameters (PSPs), instead of the selection of fibers with different optical parameters. An effective imaging range of ~ 0.6 mm with a full width at half-maximum beam diameter of fewer than 30 μm was realized with high sensitivity by the probe.


Although OCT helped identify cellular structures of tissue, it was incapable of offering functional information on the region of interest and distinguishing different cytological lesions [88]. Additionally, it was challenging to affirm the correlation between morphological changes visualized in the OCT images and their corresponding histology [89], which could render endoscopic OCT ineffective in the provision of diagnosis [90]. Thus, further studies on the improvements in the resolution, frame rate, and depth of detection of endoscopic OCT imaging systems are warranted [80, 91–96].

Coherent anti-Stokes Raman scattering

Coherent anti-Stokes Raman scattering (CARS) microscopy has been demonstrated as a promising technology for label-free biomedical imaging [97–101]. The CARS microscopy offers many advantages, including (1) chemical contrast based on Raman vibrational activity, (2) high sensitivity and rapid acquisition rates due to the coherent nature of the CARS process, (3) and sub-wavelength spatial resolution [102]. Compared to Raman spectroscopy, CARS employs multiple photons to probe the molecular vibrations and produces a signal of coherent emitted waves. As a result, the CARS signal is orders of magnitude stronger than spontaneous Raman emission and can provide much faster imaging.

CARS requires two laser frequencies called “pump” (ω_p) and “Stokes” (ω_s). When the difference of pump and Stokes’s frequency, $\Delta\omega = \omega_p - \omega_s$, is set to that of vibrational resonance, Ω , a strong optical response occurs at the “anti-Stokes” frequency: $\omega_{as} = 2\omega_p - \omega_s$ (Fig. 5). The CARS signal is derived from the third-order nonlinear susceptibility, $\chi^{(3)}$, of a material, which is significantly enhanced when the different frequency is tuned into a resonance. In addition, CARS requires a high peak-power for efficient signal generation, and tunable picosecond laser sources are optimal for CARS imaging applications as they achieve high spectral resolution and optimal signal to background ratios [102].

Liu et al. [103] developed a CARS prototype endomicroscopy using a fiber bundle with the polarization four-wave mixing (FWM)-suppressing scheme. The highly improved collection efficiency and isolation of excitation laser dramatically enhance the image quality. Also, with a broadband dual-wavelength waveplate (DWW), spectroscopic CARS imaging capability could be achieved in the FWM-suppressing scheme. 3D scanning capability was also demonstrated. Recently, Hirose et al. [101] created a rigid endoscope toward nerve-sparing robot-assisted surgery which achieved 0.91% image distortion and 8.6% non-uniformity of CARS intensity in the whole field of view (650 μm diameter field of view (450 $\mu\text{m} \times 450 \mu\text{m}$ square image)). An image with a spatial resolution of 2.91 μm was obtained by scanning the beams with galvanometer mirrors and detecting the backscattered CARS signal with a photomultiplier tube. Furthermore, the imaging time was reduced to several seconds by increasing the CARS intensity. However,




although the above prototypes capable of collecting CARS signals have been developed, future expansion in delivering in vivo CARS imaging in various medical specialties are still awaited.

Multi-photon endomicroscopy

Multi-photon imaging is a promising technique for label-free imaging of biological samples [104]. The multi-photon effect refers to the phenomenon that a molecule simultaneously absorbs or scatters two or more nonlinear photons with the same frequency, relaxes to the ground state after reaching the high energy state, and emits shorter wavelengths which can be used for biomedical imaging. Compared to single-photon imaging, multi-photon imaging uses a near-IR light source that is able to create sufficient peak energy to produce multi-photon effect while keeping average energy low enough to minimize specimen damage [105]. There are several nonlinear imaging modalities, such as two-photon excited fluorescence (TPEF), three-photon excited fluorescence (3PEF), second-harmonic generation (SHG) and third-harmonic generation (THG) imaging (Fig. 6). They can provide valuable information about the biological samples' intrinsic properties, which can be employed in diagnosis applications [106, 107].

It is challenging to acquire high-resolution images from tissue in the deep layer, for the high degree scattering caused by discontinuous refractive index and heterogeneous constituents. Despite that, the near-IR owns deeper penetration depth and causes less damaging than shorter wavelengths [108]. Thus, the infrared light can be selected as the excitation light source for multi-photon imaging, which effectively reduces the scattering of excitation light and gets a high signal-to-noise ratio by avoiding the interference of single-photon autofluorescence.

Furthermore, no pinhole is required to prevent out-of-focus fluorescence from reaching the detector in MPE, because excitation occurs only at the excitation spot, which results in the optical sectioning inherent to nonlinear imaging techniques. MPE owns the advantages of strong penetrability, high resolution, low phototoxicity, and no need to stain the tissue [109]. MPE can ex vivo or in vivo image and detect early invasion of cancer in reconstructed 3D photos as well as 2D views [110]. The critical challenge of realizing MPE is how to miniaturize the system into a compact and lightweight distal probe while keeping the general performance. The following reviews research works on various types of MPE.



TPEF


The desktop two-photon microscopy was pioneered by Denk and Strickler in 1990 [111]. Subsequently, endoscopic applications of TPEF have been explored. Huland et al. [112] used a composite GRIN lens system with a length of 8 cm and a fast-moving robotic arm to create a portable rigid endoscopic system for multiphoton *in vivo* imaging of deep tissue. They used it for imaging the liver and kidney of live mice. Moreover, the inner surface of the colon was imaged.

Brown et al. [113] demonstrated the first work that used a compact and flexible TPEF endomicroscope to acquire *in vivo* fluorescence images of unstained tissue from a live subject (liver, kidney, and colon from an anesthetized rat). The device delivered femto-second pulsed 800-nm light from the core of a raster-scanned dual-clad fiber (DCF). The light was focused by a miniaturized gradient-index lens and projected onto the tissue. In this work, images under field-of-view of 115 μm by 115 μm with 4.1-Hz frame rate were obtained. In addition, lateral and axial two-photon resolutions measured 0.8 and 10 μm (full width at half maximum), respectively, in the image plane.

Later, a 2.2-mm-outer-diameter probe was developed by Ducourthial et al. [114], which also allowed simultaneous TPEF and SHG imaging at 8 frames per second. They successfully got images with penetration depth greater than 300 μm below the surface of an anesthetized mouse kidney. And the field (FOV) of the system was up to 450 μm , while transverse and axial resolutions reach 0.8 μm and 12 μm , respectively.

In 2017, Liang et al. [115] pushed out the first fiber-optic endomicroscopy platform, which achieved subcellular resolution. The platform was implemented to perform label-free two-photon metabolic imaging of living tissue *in vivo*. With a customized dispersion compensating fiber, the image quality of the system can compare favorably with a standard bench-top two-photon microscope (The spatial resolution of the endomicroscopy was 0.7 μm laterally and 6.5 μm axially, and a frame can be acquired within only 0.38 s). In their experiment, a model of acute mouse kidney ischemia–reperfusion was performed *in vivo*: the left renal artery was occluded in the anesthetized mouse, the blood was induced by the venous vein (4 min), and then reperfusion was performed. Changes in the redox rate of renal cortical tubules were monitored in histological resolution by the TPEF endomicroscopy platform. And the endogenous fluorescent substances: NADH, FAD were selected. In addition, the redox ratio was expressed as $\text{FAD}/(\text{FAD} + \text{NADH})$. At the time of ischemia, the concentration of NADH increased, while the redox ratio decreased. Thus, endoscopic metabolic imaging was successfully achieved in the normal mouse kidney model (Fig. 7). This device has made unprecedented detection sensitivity and was capable of capturing unlabeled subcellular living structures and dynamic functional metabolic information.

More recently, a 2.6-mm-diameter TPEF probe with a high numerical aperture gradient index (GRIN) lens was developed by Kim et al. [116]. A Lissajous fiber scanner which consisted of a piezoelectric (PZT) tube and a micro-tethered-silicon-oscillator (MTSO) was installed for the separation of biaxial resonant scanning frequencies. The lateral and axial resolutions of the endomicroscopy could achieve a higher level: 0.70 μm and 7.6 μm , respectively. TPEF images of a stained kidney section and miscellaneous *ex vivo* and *in vivo* organs from wild type and green fluorescent protein transgenic (GFP-TG) mice were successfully obtained at a frame rate of 5 Hz using their system. Figure 8




compares the spatiotemporal resolution of several TPEF endomicroscopes mentioned above.

3PEF

Comparing to TPEF, three photons are absorbed almost simultaneously during the excitation process in the 3PEF endomicroscopy. By utilizing an excitation light source with a longer wavelength, the scattering of biological tissues can be further overcome with reduced water absorption [117]. Thus, imaging depth could be increased, which is beneficial to 3D structural imaging.

Huland et al. [118] presented an endoscope system with a compact and portable three-photon gradient index (GRIN) lens working at 1040 nm, which was suitable for imaging tissues without staining. Its lateral and axial resolution in water were 1.0 μ m and 9.5 μ m,



respectively. The FOV was 200 μm in diameter, and the frame rate was 2 Hz for visualizing ex vivo unstained mouse lung.

In addition, Akhouni et al. [119] developed a compact fiber-based 3PEF which was capable of performing THG, SHG, and 3PEF imaging over a large field of view (900 μm) with a high lateral resolution (2.2 μm) and a frame rate of 15 s/frame. The image depth was increased to up to 200 μm. Furthermore, the author believed that by optimizing the laser pulse delivery, the penetration depth could be increased to 700 μm in the future prototype.

In future work, the increase of chromatic aberrations in the 3PEF optics due to the more considerable difference between the excitation and signal wavelengths should be further considered. And the fiber delivery of the energetic femtosecond pulses for three-photon excitation is needed to be concerned [118]. In addition, the imaging speed of 3PEF is too low to detect the targeted region in real time. Thus, the time-efficiency for 3D imaging is limited.

SHG and THG

As one of the most widely used contrast-enhancing mechanisms in optical microscopy, SHG and THG are nonlinear optical processes in which two (SHG) or three (THG) photons with the same frequency interact with a nonlinear material under investigation,

are “combined”, and generate a new photon with twice (SHG) or three times (THG) the energy of the initial photons. Thus, SHG is a second-order process while THG is a third-order process.

The most significant difference between SHG/THG and TPEF/3PEF is the way the light interacts with the targeted region: In SHG/THG the photons are scattered to produce a new photon with two (SHG) or three (THG) times the energy of the incident photons. Compared with TPEF/3PEF, two photons are absorbed to produce a single fluorescent photon. Thus, heating and damage of tissue are less likely to occur due to the absence of absorption. For these nonlinear optical effects are proportional to the second or third power of the fundamental light intensity, only the light at the focal plane of the optic efficiently drives the nonlinearity [107]. The second-order nonlinear susceptibility of a medium characterizes its tendency to cause SHG. Thus, SHG can reveal unique molecular features of microscopic objects that are otherwise obscured by the ensemble-averaged measurements, especially in noncentrosymmetric structures (such as collagen), with reported tissue penetration depths between 100 and 300 μm at laser excitation in the near-infrared range.

Zhuo et al. [120] analyzed the SHG images of 72 fresh colonic biopsy specimens *ex vivo* and were able to quantify significant changes in the circle length of the crypt basement membranes of normal, precancerous and cancerous colonic tissues. Furthermore, Zhang et al. [121] proposed a compact fiber-optic SHG scanning endomicroscopy and its application to visualizing cervical remodeling during pregnancy. Significant morphological changes in cervical collagen were evident over the course of pregnancy. Furthermore, the compact fiber-optic SHG endomicroscopy had a resolution and image quality comparable to a bench-top microscope ($0.76 \times 4.36 \mu\text{m}$ (lateral \times axial)).

Unlike SHG, the THG detects interfaces and inhomogeneities due to its nonlinear nature [122], and THG is generated only near the focal point. Therefore, high lateral resolution can be obtained, allowing THG microscopy to perform sectioning and to construct three-dimensional images of transparent samples.

Since all materials have non-vanishing third-order susceptibilities, THG microscopy can be utilized as a general-purpose microscopy technique [123]. Mehravar et al. [110] developed a compact label-free MPE system working at 1560 nm, which generated endoscopic images of cell nuclei and collagen. The system could acquire the SHG and THG images simultaneously. The backscattered THG, back-reflected forward THG and SHG signals were combined for detecting the tissue of Barrett's esophagus. Furthermore, the system shows the thickening of sub-epithelial (basement membrane) collagen with advancing stages of dysplasia. Additionally, the first flexible endoscopic THG imaging system was proposed by Akhondi et al. [119]. In their work, the lateral imaging resolution was 2.2 μm , while the axial resolution was 12.7 μm .

However, the application of THG was not mature because of the instability of specialized high-energy excitation light.

Optical sensing technologies

Diffuse reflectance spectroscopy

Spectroscopy is not easily affected by artifacts or sampling errors, so it can provide quantitative data that are free of subjective interpretation. The light delivered to

the tissue surface undergoes multiple elastic scattering and absorption, and part of it returns as diffuse reflectance carrying quantitative information about tissue structure and composition [124]. Moreover, primary scattering centers of tissue are the extracellular matrix consisting of a collagen fiber network and intracellular structures with sizes smaller than optical wavelengths [125]. And larger intracellular structures, such as the nuclei, also scatter light, with their relative contribution on increasing the backscatter direction [126].

Diffuse reflectance spectroscopy (DRS), also known as elastic scattering spectroscopy (ESS), analyzing mainly multiply scattered light, is sensitive to the bulk tissue scattering properties (the size and packing of dense subcellular components such as the nucleus, nucleolus, and mitochondria) as well as the absorption properties of hemoglobin [127, 128]. In the DRS system, a light source with broadband, typically the white light, is utilized to illuminate the sample. The reflected signal is collected and transmitted by optical fibers to the analyzing spectrometer [129]. Delivery and collection fibers are usually separated by a set distance, which allows the spectrometer to detect diffuse reflections preferentially [130]. Enlarged dense, crowded epithelial nuclei and an increased nucleus-to-cytoplasm ratio can be the primary pathological indicators of cancer, dysplasia, and cell regeneration [131, 132].

In the late 1990s, Mourant et al. [133] proposed the first DRS system to conduct clinical trials towards the gastrointestinal tract *in vivo*. Reflectance spectra were taken *in vivo* from the colon in 15 patients and the stomach in 17 patients. The slope of the spectrum in the 435 to 440 nm range was used to reasonably separate active colitis from quiescent colitis and normal colonic mucosa. When multiple linear regression was used for spectral classification, adenomatous polyps were distinguished from hyperplastic polyps with a sensitivity of 89% and a specificity of 75% [134]. Additionally, Zonios et al. [124] analyzed the data using an analytical light diffusion model, which was tested and validated on a physical tissue model composed of polystyrene beads and hemoglobin. Four parameters were obtained to tell healthy tissue from adenomatous tissue: hemoglobin concentration, hemoglobin oxygen saturation, effective scatterer density, and adequate scatterer size, which suggest that diffuse reflectance can be used to obtain tissue information about tissue structure and composition *in vivo*.

In addition, Lovat et al. [130] collected elastic scattering spectroscopy measurements *in vivo*, which were matched with histological specimens taken from identical sites within Barrett's esophagus. A total of 181 matched biopsy sites from 81 patients, where histopathological consensus was reached and analyzed. The results were that there was a good pathologist agreement in differentiating high-grade dysplasia and cancer from other pathology ($\kappa=0.72$). DRS detected high-risk sites with 92% sensitivity and 60% specificity and differentiated high-risk sites from inflammation with a sensitivity and specificity of 79%. When DRS was used to target biopsies during endoscopy, the number of low-risk biopsies taken would decrease by 60%, which could save significant endoscopist and pathologist time with consequent financial savings. A negative spectroscopy result would exclude high-grade dysplasia or cancer with an accuracy of 99.5%.

Like a variant of ESS, low coherence enhanced backscattering (LEBS) spectroscopy was recently proposed as an endoscopic prescreening method for colorectal cancer screening. Based on the concept of field carcinogenesis, micro-scale spectral changes

were detectable and were measured in biopsy specimens of the endoscopically normal rectum. Two hundred and nineteen patients took part in colonoscopy to predict if their colon harbored malignancy or advanced adenomas elsewhere. Alterations in the endoscopically normal rectum were seen with the help of LEBS to mirror neoplasia progression in patients with no neoplasia, 5-mm to 9-mm adenomas and advanced adenomas. For advanced adenomas, the signals from LEBS had a sensitivity of 100% with a specificity of 80% [135]. Further works with a fiber-optic probe compatible with conventional endoscopy are underway for validation of these results in vivo [129].


Light-scattering spectroscopy

Detecting singly scattered photons, rather than diffused photons from tissue, light-scattering spectroscopy (LSS) obtains quantitative information of nuclear morphology by extracting the singly backscattered light from epithelial layers of human tissue using modeling. Due to the small optical-path length of singly scattered photons in tissue, variations in the spectrum related to the concentrations of absorbers and scatterers in the tissue are minimized [136]. When the light propagates in tissue, multiple scattering randomizes information about the scatterers over the effective scattering length [126]. Thus, the spectrum of light-scattering of tissues in vivo consists of three main components: a Rayleigh component due to elastic scattering by small organelles, a broad background from submucosal tissue [137], and a relatively small backscattered component due to epithelial cell nuclei. It was reported that the submucosal background could be removed by one of the polarization [136, 138–141] or spatial [142] gating techniques.

The smaller organelles have a very different scattering spectral dependence than that of the nuclei. The combination of gating and difference in spectral behavior allows the epithelial nuclear scattering spectrum to be isolated in the processed LSS signal [142]. The shape of the backscattered spectrum over the wavelength range is a measure of nuclear size [143], and its amplitude is related to the number density of epithelial nuclei (nuclei per unit area), which is a measure of nuclear crowding [140].

In 2018, Zhang et al. [142] developed an endoscopic LSS-based fiber-optic probe that predicts the malignant potential of pancreatic cystic lesions during routine diagnostic EUS-FNA procedures. In a double-blind prospective study in 25 patients, 14 cysts were measured in vivo and 13 postoperatively. The technique achieved an overall accuracy of 95%, with a 95% confidence interval of 78–99%, in cysts with a definitive diagnosis. Qiu et al. [126] proposed a multispectral probe-based detecting system that combines light scattering spectroscopy, which analyzed light scattered from epithelial cells, thereby identifying otherwise invisible dysplastic sites [138]. A compatible fiber optic probe delivered a collimated broadband light beam through the working channel (Fig. 9). The system scanned the entire Barrett's esophagus segment in minutes, providing the endoscopist with real-time operator-independent information about the location of invisible high-grade dysplasia (HGD).


They developed a rapid semi-empirical algorithm to reconstruct the epithelial cell nuclei size distribution at the illuminated spot accurately. In their algorithm, a diagnostic parameter, Δ , was provided as a measure of the contribution from dysplastic cells. The nuclear size distributions were reconstructed from the LSS data for various spatial locations. From these distributions, they determined that a site should be considered



dysplastic if the diagnostic parameter exceeds 0.1. This $\Delta = 0.1$ threshold was equivalent to ~25% contribution from enlarged nuclei, defined as nuclei over 9 μm in diameter.

Based on the nuclear size distributions extracted from the backscattering spectra for each individual spatial location, they made pseudo-color maps where red and pink regions indicated areas suspicious for dysplasia (Δ below 0.05 colored dark green, 0.05–0.10 colored light green, 0.10–0.15 colored pink and above 0.15 as red).

In this way, the maps were displayed and presented to the endoscopists in cases where the multispectral endoscopic system was employed to guide biopsy (Fig. 10). In



the per-patient evaluation study, the multispectral light-scattering endoscopic system demonstrated a sensitivity of 96% and a specificity of 97% for high-grade dysplasia detection. Moreover, in the per-biopsy evaluation study, a specificity of 91%, a sensitivity of 88%, and a negative predictive value of 96% in detecting individual locations of HGD were demonstrated, which reached the PIVI thresholds established by American Society for Gastrointestinal Endoscopy (ASGE) Technology Committee [144].

Angle-resolved low coherence interferometry

Angle-resolved low coherence interferometry (a/LCI) is a novel scattering technique that combines the capabilities of light-scattering spectroscopy to detect morphological changes in cell nuclei with the depth-resolving power of optical coherence tomography. It obtains morphological information from subsurface sites by examining the depth-resolved angular distribution of elastically backscattered light [145]. A/LCI utilizes broadband light to separate light that is scattered from varied depths in tissue by mixing scattered light with that from a reference beam. By combining the low coherence interferometry with the capacity of light scattering methods, it can get structural information with subwavelength precision and accuracy [146]. Angular distribution of scattered light was also collected, then it was compared to various angular scattering solutions from a performed database to determine the average nuclear size and index of refraction from the scatters in question. There are numerous exciting applications of this technology.

Initially validated in a study measuring the size of polystyrene microspheres suspended in a neutrally buoyant medium [147], the a/LCI was also applied to measure cellular morphology in vitro [146]. Then, the a/LCI system was reported to assist in earlier detection of tissue dysplasia, resulting in more effective medical intervention and improved patient outcomes: Zhu et al. [148] proposed a probe-based endoscopic Fourier-domain a/LCI system to investigate the presence of dysplasia in vivo in Barrett's esophagus patients. Unique nuclear size and density information were obtained from each depth layer of the tissue, which indicated the ability of a/LCI system to retrieve depth-resolved in vivo morphological and optical information from the tissue in question. In addition, the fiber probe featured long length, small diameter, and sub-second data acquisition (0.4 s/frame). Obtained depth-resolved sizing patterns were closely associated with the pathological tissue conditions and can serve as a basis for potential assessment of tissue health. Moreover, Terry et al. [149] reported sensitivity as high as 100% and specificity of 84% for the discrimination of dysplasia with a/LCI fiber probe system in 46 patients who were undergoing routine surveillance in vivo for Barrett's esophagus. In addition, Ho et al. [150] found a strong relationship between nuclear enlargement at the basal/parabasal epithelial bin for clinical detection of cervical dysplasia and the presence of dysplasia in histological analysis.


Currently, the most significant issue is determining the state of tissue health based on average cell nuclei size [151]. Data collected will serve as a training set to establish the decision lines for future prospective studies of grading biopsy sites. Thus, more advanced hybrid algorithms, and the implementation of processing software, are still waiting to realize real-time processing and interpretation of a/LCI scan results [150].

Raman spectroscopy

Raman spectroscopy (RS) is an exciting analytical optical spectroscopic technique capable of probe vibrational and rotational modes of endogenous biomolecules of tissue or cells intrinsically. Photons in the incident laser light interact with the tissue and undergo inelastic collisions with molecules, resulting in an exchange of energy and a change in frequency. Moreover, in the Raman spectrum, the intensity of the peak of shifted light is directly proportional to the concentration of the sum of molecular constituents giving rise to that peak. Thus, the Raman spectrum is a direct function of the molecular composition of the targeted region within the tissue, giving us a complex molecular fingerprint. Most biological molecules have distinct vibrational energies in their molecular bonds. Thus, they have their specific fingerprint [152]. In addition, the Raman shift caused by frequency change is specific to the species of the molecule. The shifts in wavelengths are expressed as wave numbers, and they are independent of the wavelength of excitation, which means that the energy shift is constant for particular molecular bonds (Fig. 11).

Thus, by measuring the molecular specific inelastic scattering of light, RS enables the explanation of a tissue's biochemical fingerprint. Primarily, RS requires a monochromatic laser to illuminate the tissue sample. Subsequently, the intensity and wavelength of the scattered light are collected and analyzed. Therefore, RS can detect subtle biochemical and molecular changes that are vital in the discrimination of tissue samples, which makes RS can become a potential diagnostic tool [153]. However, the inherent weakness of Raman signals is one major limitation of RS. Thus, endoscopic RS systems need high laser powers and ultrasensitive detectors, which are enough to merit this technique with clinical value. Moreover, the sensitivity can also be increased with the Raman signals enhanced by ten orders of magnitude by a phenomenon known as Surface Enhanced Raman Spectroscopy (SERS) [90].

Aiming to develop as a histopathology tool, Shim et al. [154] pioneeringly developed a custom-built fiber-optic Raman spectroscopic system at 785-nm excitation. Raman spectra were obtained from various organs in the digestive tract during




gastrointestinal endoscopy, including the esophagus and colon, but because of the limited set of data, only subtle differences were observed in vivo between normal and pathologic states. Later on, distinctive Raman spectra were shown by the same group in a comparison between 10 adenomas and 9 hyperplastic colon polyps with 95% accuracy [155].

Furthermore, Huang et al. [156] employed a fiber-optic Raman endoscopic system for rapid Raman measurements on gastric tissue, which demonstrated for the first time that image-guided Raman endoscopy had a promising potential for the non-invasive in vivo diagnosis and detection of gastric cancer at the molecular level. In their system, wide-field images (white-light imaging/autofluorescence imaging/narrow-band imaging) and the corresponding real time in vivo Raman spectra of the tissue can be displayed and recorded simultaneously. Upon analysis of specific biochemical constituents and biomolecular differences from gastric Raman spectra, algorithms based on the Classification and Regression Tree (CART)-Raman model were developed for distinguishing neoplastic lesions from healthy tissue. The diagnostic sensitivity of 94.0% and specificity of 93.4% could be achieved.

In 2015, Wang et al. [158] developed a real-time fiber-optic Raman spectroscopy system capable of simultaneously acquiring both fingerprint (FP) ($800\text{--}1800\text{ cm}^{-1}$) and high-wavenumber (HW) ($2800\text{--}3600\text{ cm}^{-1}$) Raman spectra from gastric tissue in vivo. Multiple spectra for each tissue site were measured with a scanning time of 0.1–0.5 s, which permitted a rapid survey of large tissue areas. Based on the use of complementary biochemical/biomolecular information harvested through the integrated FP/HW Raman spectroscopy, the diagnosis of gastric dysplasia was enhanced as compared to either the FP or HW Raman techniques alone. Robust spectral diagnostic models were developed based on the implementation of the partial least squares discriminant analysis (PLS-DA) and leave-one-patient-out cross-validation (LOPCV). The system was finally able to provide diagnostic sensitivities of 96.0%, 81.8%, and 88.2%, and specificities of 86.7%, 95.3%, and 95.6%, respectively, for the classification of normal, dysplastic and cancerous gastric tissue, which was superior to WLE or FP or HW Raman techniques alone. Furthermore, the same group also demonstrated that their device could also improve real time in vivo diagnosis of esophageal squamous cell carcinoma (ESCC) during endoscopy [157]. In their clinical trial, a total of 1172 in vivo FP/HW Raman spectra were acquired from 48 esophageal patients. They split the total Raman dataset into two parts: 80% for training, 20% for testing (Fig. 12). The result showed that simultaneous FP/HW Raman spectroscopy on training dataset provided a diagnostic sensitivity of 97.0% and specificity of 97.4% for ESCC classification. The diagnostic algorithm applied to the independent testing dataset based on simultaneous FP/HW Raman technique gave a predictive diagnostic sensitivity of 92.7% and 93.6% specificity for ESCC identification, which is superior to either FP or HW Raman technique alone.

As far as we know, RS has excelled in the early detection of precancer and cancer in the interior of various organs with high diagnostic specificity (Table 3) [158–163]. Several probes are currently undergoing construction, and engineering tolerance levels are being carefully defined before a clinical trial [164]. Further evaluations of the clinical merits of Raman spectroscopy for prospective prediction of precancer and cancer in vivo are expected.



Discussion

Before any novel endoscopic optical diagnostic technology is successfully put into clinical use, it is necessary to conduct *ex vivo*/*in vivo* preclinical animal experiments and *ex vivo* human trials, which are followed by rigorous *in vivo* human clinical trials to further verify the effectiveness and safety of these techniques. We track the emerging *in vivo* human clinical trial researches on novel endoscopic optical diagnostic techniques and summarize their coverage on several primary medical specialties in Table 3, which also indicates unexplored speciality fields for clinical researchers. In addition, the challenges, as well as the future development orientation of endoscopic optical diagnostic techniques, are discussed below.

Future perspectives

Future work related to the advancements of endoscopic imaging and sensing may focus on:

- (1) A higher resolution with a wider field of view.
- (2) An increased frame rate with super-resolution.

Table 3 In vivo human clinical trial researches on novel endoscopic optical diagnostic techniques in primary medical specialities

Technique	References									
	Gastroenterology									
	Pancreatic duct	Biliary tract	Upper GI tract	Lower GI tract	Urology	Gynecology	Pneumology	Otorhinolaryngology	Laparoscopic surgery	
HSI	-	-	[30, 165, 166]	[23]	-	[167]	-	[29, 168]	-	
FLIM	-	-	[46, 49]	[169]	[170]	-	-	-	-	
PAE	-	-	-	-	-	-	-	-	-	
OCT	[171]	[172, 173]	[94, 96, 174, 175]	[91, 176, 177]	[95, 178, 179]	[93, 180, 181]	[174, 182, 183]	[184, 185]	[186]	
CASR	-	-	-	-	-	-	-	-	-	
MPE	-	-	-	-	-	-	-	-	-	
DRS	-	-	[187]	[188]	[189]	[190-192]	-	-	-	
LSS	[142]	-	[190, 193, 194]	[124, 195]	[195]	[190]	-	-	-	
a/LCI	-	-	[148, 149]	-	-	[150]	-	-	-	
RS	-	-	[158, 161, 162, 196]	[163, 197]	[198]	[199-201]	-	[202-204]	-	

- (3) Combining multiple modalities. Compared to one technique, several characteristic information can be obtained simultaneously, which cannot be acquired by these technologies individually. For instance, Lombardini et al. [205] combined CARS and SHG during endoscopy, which acquire SHG images immediately after the CARS images (Fig. 13).
- (4) Improving the image signal-to-noise ratio to highlight lesion characteristics.
- (5) Miniaturization of the probes. We prospect that further miniaturization of the flexible probes will benefit diagnosis by covering more specific medical scenarios. The birth of new materials has driven the miniaturization of detectors. For instance, in 2019, Yang et al. [206] developed miniaturized spectrometers based on the single, compositionally engineered nanowire. Their devices are capable of accurate, visible-range monochromatic and broadband light reconstruction, as well as spectral imaging from centimeter-scale focal planes down to lensless, single-cell-scale in situ mapping. In this way, the single-nanowire spectrometers can detect the spectral information of biological tissue in hyperspectral imaging.
- (6) Furthermore, we envision that future works on the advance of endoscopic modalities could shed light on transferring bench-top super-resolution microscopies (SRM) into miniature and flexible imaging devices such as endoscope-based probes or catheters. In this way, they could provide new mechanistic insights into biological processes.

The imaging resolution of SRM breaks the Abbe limit and is above 200 nm. Owning sub-micron or even nano-scale optical discrimination ability, SRM like photoactivated localization microscopy (PALM) [207–209], stochastic optical reconstruction microscopy (STORM) [210–212], stimulated emission depletion (STED) [213–216]






Fig. 13 Multimodal SHG (green, **a**) and CARS (red, **b**) images of fresh human colon tissues. Images were taken 50 μm below the sample surface. **c** The large images are overlaps of the SHG and CARS images. Adapted with permission from Ref. [205]. Copyright 2018 Springer Nature Publishing

and structured illumination microscopy (SIM) [217–224] can continuously monitor the evolution of macromolecules and organelles without affecting the biological activity of biological systems. Among various SRMs, SIM excelled in extending the spatiotemporal resolution and the imaging duration with a notably reduced illumination intensity. Moreover, SIMs are now the most widely used technique in super-resolution optical microscopy of living cells [225]. Thus, SRMs are vital tools to research life sciences. Figure 14 illustrates the spatiotemporal resolution of these SIM systems.

Challenges

- (1) As we know, the targeted objects detected by clinical endoscopy are all three-dimensional. Both surface and depth information are used to diagnose the disease. However, there is a significant barrier to quantitative characterization of the targeted tissue: imaging towards biomedical utilization has an inverse relationship between the volume that any imaging modality can cover (field of view and penetration depth) and the size of details it can visualize (spatial resolution) [226, 227]. Thus, emerged imaging modalities are often restricted to a single range for the combined resolution, field of view, and penetration depth [228]. In practical terms, those imaging techniques utilized in endoscopy usually acquire information on among organelle scale, subcellular, cellular scale, and tissue (multicellular) scale (Fig. 15). As for a particular modality, the higher spatial resolution is achieved, the deeper detectable depth it abandons.
- (2) The real-time performance of an endoscopic imaging system is a crucial indicator accounting for diagnostic efficiency. However, temporal resolution is limited by scanning speed, sample fluorescence signal intensity, detector sensitivity, and processing time. Several technologies such as THG, CARS, and PAE are not yet real





time, with frame rates below 5 Hz, thereby increasing the patient's suffering during clinical procedures.

- (3) The diagnostic capability of “optical biopsies” has not reached that of ex vivo biopsy. Because “optical biopsies” has some limitations: (i) it does not reach the detection depth as fine-needle aspiration (FNA); (ii) can not observe cells cultured in vitro for subsequential changes. Also, the optical biopsy result consistency should be evaluated because different endoscopists could lead to variable diagnostic results. Thus, the uniform standards of image data interpretation and technical specifications should be set clearly before clinical practice.
- (4) Another limitation is that it is difficult to judge whether the light projected on the tissue is safe. Thus, the power, wavelength, and duration of the light should be considered and calculated. In addition, studies on phototoxicity are required to identify the most discriminating wavelengths and lifetimes to create a relevant diagnostic model [129].
- (5) Meanwhile, there sorely lacks software tools for interpreting clinical data accurately and the guidance for the use of software for the different observers.
- (6) Furthermore, it is essential to develop enduring systems that are also easy to be sterilized.
- (7) Lastly, ethical acceptability needs to be considered and evaluated.

Conclusion

This review aims to help researchers to have a broad and clear understanding of emerging innovative endoscopic diagnostic technologies in medical trial researches, we mainly summarize studies on these promising techniques, with emphasis on their development processes and recent advances. We also contribute to cover both the basic principles and technical parameters of these technologies. We exploit the innovative optical imaging and sensing mechanisms of these prototype technologies, which promoted an unprecedented stride in basic scientific and clinical trial studies. We further explain how these technologies aids endoscopists to perform accurate and time-efficient optical diagnosis with real-time functional and spatial structural information on regions of interest.

Among all novel techniques in medical trial research we stated above, endoscopic OCT is relatively mature by now. A large number of successful clinical experiments have been conducted on different endoscopic OCT systems to validate their application in many medical specialties. Similar to OCT, PAE can also construct 3D stereo images in vivo with cellular resolution and millimeter-scale penetration depth. HSI acquires morphologic and physiologic information and provides a 3D “data tube” comprised of 2D spatial and 1D spectral details. Furthermore, FLIM differentiates carcinoma from healthy tissue by recording the fluorescence lifetimes which are recorded by each pixel in a particular field of view. Moreover, for prescreening, DRS, LSS, and a/LCI measure the average size and optical density of cell nuclei in epithelial tissue to identify potentially harboring neoplasia. RS reveals specific biomolecule fingerprint before morphologic changes by measuring the shift of wave numbers. In addition, promising endomicroscopic technologies such as MPE and CARS provide real-time subcellular imaging, but they have not been reported in any in vivo clinical experiments in any medical specialty (as shown in Table 3).

Furthermore, these emerging novel endoscopic optical diagnostic techniques will continue to develop with the cooperation of multi-disciplinary to improve the diagnostic capabilities of detection, characterization, and confirmation in the near future. To obtain certification of the supervisory authorities on medical devices in different regions all over the world, such as Food and Drug Administration (FDA) and European Conformity (CE), current efforts are directed to implement and validate these technologies for modern endoscopy before general clinical practice in hospitals. However, these emerging diagnostic devices have not obtained official approval of clinical use and vary across maturities. Some challenges lie in the results inconsistency, cost-effectiveness, time-efficiency, etc., which require further researches and developments to achieve a promising era of endoscopy.

Abbreviations

NBI: Narrow-band imaging; LCI: Linked color imaging; CLE: Confocal laser endomicroscopy; HSI: Hyperspectral imaging; FLIM: Fluorescence lifetime imaging microscopy; CLE: Confocal laser endomicroscopy; OCT: Optical coherent tomography; PAE: Photoacoustic endoscopy; DRS: Diffuse reflectance spectroscopy; LSS: Light-scattering spectroscopy; a/LCI: Angle-resolved low coherence interferometry; RS: Raman spectroscopy; CARS: Coherent anti-Stokes Raman scattering; TPEF: Two-photon excitation fluorescence; 3PEF: Three-photon excitation fluorescence; SHG: Second-harmonic generations; THG: Third-harmonic generations; SRM: Super-resolution microscopy; SIM: Structured illumination microscopy; PALM: Photoactivated localization microscopy; STORM: Stochastic optical reconstruction microscopy; STED: Stimulated emission depletion.

Acknowledgements

We thank Dr. Weiling Hu and Dr. Xiao Liang at Sir Run-Run Shaw Hospital, Hangzhou, PR China, for helping inspire the generation of this review article. We would also like to express our sincere gratitude to Professor Hai Lin from the State Key Lab of CAD&CG, Zhejiang University, Hangzhou, PR China, for his valuable insights.

Authors' contributions

ZH and PW contribute equally to the article. They performed the literature review and drafted the manuscript; XY gave suggestions and revised the manuscript. All authors read and approved the final manuscript.

Funding

This research was supported by the National Key Research and Development Project (Grant Nos. 2019YFC0117900, 2019YFC0117901, 2017YFB1302803, 2017YFC0109603 and 2017YFC0110802), the National Major Scientific Research Instrument Development Project (Grant No. 81827804), and the Robotics Institute of Zhejiang University (Grant Nos. K11806 and K11807), the Key Research and Development Plan of Zhejiang Province (Grant Nos. 2017C03036 and 2018C03064).

Availability of data and materials

Not applicable.

Ethics approval and consent to participate

Not applicable. **Consent for publication.** Not applicable.

Competing interests

The authors declare that they have no competing interests to disclose.

Author details

¹ Biosensor National Special Laboratory, College of Biomedical Engineering and Instrument Science, Zhejiang University, Hangzhou 310027, People's Republic of China. ² State Key Laboratory of CAD and CG, Zhejiang University, Hangzhou 310058, People's Republic of China.

Received: 17 January 2020 Accepted: 23 December 2020

Published online: 06 January 2021

References

1. Association BM, Smith T. The British Medical Association Complete Family Health Encyclopedia Complete Family Health Encyclopedia: Dorling Kindersley; 1990.
2. Groen PCD. History of the Endoscope Proceedings of the IEEE. 2017;105:1987–95.
3. Sabina B, Ana W, Krish R. The use of optical imaging techniques in the gastrointestinal tract. *Frontline Gastroenterology*. 2016;7:207–15.
4. Song LMWK, Adler DG, Conway JD, Diehl DL, Farraye FA, Kantsevov SV, et al. Narrow band imaging and multiband imaging. *Gastrointest Endosc*. 2008;67:581–9.
5. Kuznetsov K, Lambert R, Rey JF. Narrow-band imaging: potential and limitations. *Endoscopy*. 2006;38:76–81.
6. Sun X, Bi Y, Dong T, Min M, Shen W, Xu Y, et al. Linked colour imaging benefits the endoscopic diagnosis of distal gastric diseases. *Sci Rep*. 2017;7:5638.
7. Fukuda H, Miura Y, Hayashi Y, Takezawa T, Ino Y, Okada M, et al. Linked color imaging technology facilitates early detection of flat gastric cancers. *Clin J Gastroenterol*. 2015;8:385–9.
8. Kanzaki H, Takenaka R, Kawahara Y, Kawai D, Obayashi Y, Baba Y, et al. Linked color imaging (LCI), a novel image-enhanced endoscopy technology, emphasizes the color of early gastric cancer. *Endosc Int Open*. 2017;05:E1005–13.
9. Kodashima S, Fujishiro M. Novel image-enhanced endoscopy with i-scan technology. *World J Gastroenterol*. 2010;16:1043–9.
10. Sung Noh H, Won Hyeok C, Jung Hyun L, So-I K, Jeong Hwan K, Tae Yoon L, et al. Prospective, randomized, back-to-back trial evaluating the usefulness of i-SCAN in screening colonoscopy. *Gastrointest Endosc*. 2012;75(1011–1021):e1012.
11. Neumann H, Fujishiro M, Wilcox CM, Mönkemüller K. Present and future perspectives of virtual chromoendoscopy with i-scan and optical enhancement technology. *Dig Endosc*. 2014;26:43–51.
12. Manfredi MA, Dayyeh BKA, Bhat YM, Chauhan SS, Gottlieb KT, Hwang JH, et al. Electronic chromoendoscopy. *Gastrointest Endosc*. 2015;81:249.
13. Kiesslich R, Goetz M, Hoffman A, Galle PR. New imaging techniques and opportunities in endoscopy. *Nat Rev Gastroenterol Hepatol*. 2011;8:547–53.
14. Kwon RS, Song LMWK, Adler DG, Conway JD, Diehl DL, Farraye FA, et al. Endocytoscopy. *Gastrointest Endosc*. 2009;70:610–3.
15. Kiesslich R, Goetz M, Vieth M, Galle PR, Neurath MF. Confocal laser endomicroscopy. *Gastrointest Endosc*. 2014;2012:715.
16. Goetz M, Malek NP, Kiesslich R. Microscopic imaging in endoscopy: endomicroscopy and endocytoscopy. *Nat Rev Gastroenterol Hepatol*. 2014;11:11.
17. Mountney P, Giannarou S, Elson D, Yang G-Z. Optical biopsy mapping for minimally invasive cancer screening. In: *International Conference on Medical Image Computing and Computer-Assisted Intervention*: 2009. Springer: 483–490.
18. Lu G, Fei B. Medical hyperspectral imaging: a review. *J Biomed Optics*. 2014;19:010901.

19. Gebhart SC, Lin WC, Mahadevanjansen A. Characterization of a spectral imaging system. *Proc SPIE*. 2003;4959:34–45.
20. Martin ME, Wabuyele MB, Chen K, Kasili P, Panjehpour M, Phan M, et al. Development of an advanced hyperspectral imaging (HSI) system with applications for cancer detection. *Ann Biomed Eng*. 2006;34:1061–8.
21. Vo-Dinh T, Stokes DL, Wabuyele MB, Martin ME, Song JM, Jagannathan R, et al. A hyperspectral imaging system for in vivo optical diagnostics. *Hyperspectral imaging basic principles, instrumental systems, and applications of biomedical interest*. *Eng Med Biol Mag IEEE*. 2004;23:40–9.
22. Akbari H, Halig L, Schuster DM, Fei B, Osunkoya A, Master V, et al. Hyperspectral imaging and quantitative analysis for prostate cancer detection. *J Biomed Optics*. 2012;17:076005.
23. Joshi BP, Miller SJ, Lee CM, Seibel EJ, Wang TD. Multispectral endoscopic imaging of colorectal dysplasia in vivo. *Gastroenterology*. 2012;143:1435–7.
24. Ortega S, Fabelo H, Iakovidis DK, Koulaouzidis A, Callico GM. Use of hyperspectral/multispectral imaging in gastroenterology. Shedding some-different-light into the dark. *J Clin Med*. 2019;8:36.
25. Jacques SL. Optical properties of biological tissues: a review. *Phys Med Biol*. 2013;58:R37.
26. Ortega S, Fabelo H, Iakovidis D, Koulaouzidis A, Callico G. Use of hyperspectral/multispectral imaging in gastroenterology. Shedding some-different-light into the dark. *Journal of clinical medicine*. 2019;8:36.
27. Siddiqi AM, Li H, Faruque F, Williams W, Lai K, Hughson M, et al. Use of hyperspectral imaging to distinguish normal, precancerous, and cancerous cells. *Cancer Cytopathol*. 2010;114:13–21.
28. Arnold T, De Biasio M, Leitner R. High-sensitivity hyper-spectral video endoscopy system for intra-surgical tissue classification. In: *SENSORS*: 2010. IEEE: 2612–2615.
29. Gerstner AO, Laffers W, Bootz F, Farkas DL, Martin R, Bendix J, et al. Hyperspectral imaging of mucosal surfaces in patients. *J Biophotonics*. 2012;5:255–62.
30. Gu X, Han Z, Yao L, Zhong Y, Shi Q, Fu Y, et al. Image enhancement based on in vivo hyperspectral gastroscopic images: a case study. *J Biomed Optics*. 2016;21:101412.
31. Duann JR, Jan CI, Ou-Yang M, Lin CY, Mo JF, Lin YJ, et al. Separating spectral mixtures in hyperspectral image data using independent component analysis: validation with oral cancer tissue sections. *Journal of Biomedical Optics*. 2013;18:126005.
32. Lakowicz JR. *Principles of fluorescence spectroscopy*: Springer Science & Business Media; 2013.
33. Jo JA, Applegate BE, Park J, Shrestha S, Pande P, Gimenez-Conti IB, et al. In vivo simultaneous morphological and biochemical optical imaging of oral epithelial cancer. *IEEE Trans Biomed Eng*. 2010;57:2596–9.
34. R Cubeddu D, Valentini G. Time-resolved fluorescence imaging in biology and medicine. *Journal of Physics D Applied Physics*. 2002;35:R61.
35. Lakowicz JR. *Principles of fluorescence spectroscopy*, third edition. *Die Naturwissenschaften*. 1991;78:456.
36. Kennedy GT, Manning HB, Elson DS, Neil MAA, Stamp GW, Viellerobe B, et al. A fluorescence lifetime imaging scanning confocal endomicroscope. *J Biophotonics*. 2010;3:103–7.
37. O'Connor D. *Time-correlated single photon counting*: Academic Press; 2012.
38. Fruhwirth GO, Ameer-Beg S, Cook R, Watson T, Ng T, Festy F. Fluorescence lifetime endoscopy using TCSPC for the measurement of FRET in live cells. *Opt Express*. 2010;18:11148–58.
39. Requejo-Isidro J, McGinty J, Munro I, Elson DS, Galletly NP, Lever MJ, et al. High-speed wide-field time-gated endoscopic fluorescence-lifetime imaging. *Opt Lett*. 2004;29:2249–51.
40. Elson D, Requejoisidro J, Munro I, Reavell F, Siegel J, Suhling K, et al. Time-domain fluorescence lifetime imaging applied to biological tissue. *Photochem Photobiol Sci*. 2004;3:795–801.
41. Liu Lixin Qu, Junle LZ, Chen Danni Hu, Tao GB, et al. A High repetition rate picosecond streak camera for two-photon excited fluorescence lifetime microscopic imaging. *Acta Optica Sinica*. 2006;26:373–8.
42. James DR, Siemiarczuk A, Ware WR. Stroboscopic optical boxcar technique for the determination of fluorescence lifetimes. *Rev Sci Instrum*. 1992;63:1710–6.
43. Berberan-Santos M, Valeure B. *Molecular fluorescence: principles and applications*. In.: Wiley-VCH, Weinheim; 2013.
44. Munro I, McGinty J, Galletly N, Requejo-Isidro J, Lanigan PMP, Elson DS, et al. Toward the clinical application of time-domain fluorescence lifetime imaging. *J Biomed Optics*. 2005;10:051403.
45. Mizeret JRM, Stepinac T, Hansroul M, Studzinski A, van den Bergh H, Wagnieres G. Instrumentation for real-time fluorescence lifetime imaging in endoscopy. *Rev Sci Instrum*. 1999;70:4689.
46. Glanzmann T, Ballini J-P, van den Bergh H, Wagnieres G. Time-resolved spectrofluorometer for clinical tissue characterization during endoscopy. *Review of Scientific Instruments*. 1999;70:4067–4077.
47. McGinty J, Galletly NP, Dunsby C, Munro I, Elson DS, Requejo-Isidro J, et al. Wide-field fluorescence lifetime imaging of cancer. *Biomed Optics Express*. 2010;1:627–40.
48. Sun Y, Phipps JE, Meier J, Hatami N, Poirier B, Elson DS, et al. Endoscopic fluorescence lifetime imaging for in vivo intraoperative diagnosis of oral carcinoma. *Microsc Microanal*. 2013;19:791–8.
49. Ning Y, Cheng S, Wang J-X, Liu Y-W, Feng W, Li F, et al. Fluorescence lifetime imaging of upper gastrointestinal pH in vivo with a lanthanide based near-infrared τ probe. *Chemical science*. 2019;10:4227–35.
50. Anderson MA, Carpenter S, Thompson NW, Nostrant TT, Elta GH, Scheiman JM. Endoscopic ultrasound is highly accurate and directs management in patients with neuroendocrine tumors of the pancreas. *Am J Gastroenterol*. 2000;95:2271–7.
51. Hocke M, Schulze E, Theodor P, et al. Contrast-enhanced endoscopic ultrasound in discrimination between focal pancreatitis and pancreatic cancer. *World J Gastroenterol*. 2006;12:246–50.
52. Kelly S, Harris K, Berry E, Hutton J, Roderick P, Cullingworth J, et al. A systematic review of the staging performance of endoscopic ultrasound in gastro-oesophageal carcinoma. *Gut*. 2001;49:534–9.
53. Changhui L, Wang LV. Photoacoustic tomography and sensing in biomedicine. *Phys Med Biol*. 2009;54:R59.
54. Viator JA, Au G, Paltauf G, Jacques SL, Prael SA, Ren H, et al. Clinical testing of a photoacoustic probe for port wine stain depth determination. *Lasers Surg Med*. 2002;30:141–8.
55. Taruttis A, Ntziachristos V. Advances in real-time multispectral optoacoustic imaging and its applications. *Nat Photonics*. 2015;9:219–27.

56. Bézière N, Ntziachristos V. Photoacoustic imaging: an emerging modality for the gastrointestinal tract. *Gastroenterology*. 2011;141:1979–85.
57. Yoon TJ, Cho YS. Recent advances in photoacoustic endoscopy. *World J Gastrointest Endosc*. 2013;5:534–9.
58. Zhang JG, Liu HF. Functional imaging and endoscopy. *World J Gastroenterol*. 2011;17:4277.
59. Yang J-M, Maslov K, Yang H-C, Zhou Q, Shung KK, Wang LV. Photoacoustic endoscopy. *Opt Lett*. 2009;34:1591–3.
60. Wang LV, Song H. Photoacoustic tomography: in vivo imaging from organelles to organs. *Science*. 2012;335:1458–62.
61. Laufer J, Johnson P, Zhang E, Treeby B, Cox B, Pedley B, et al. In vivo preclinical photoacoustic imaging of tumor vasculature development and therapy. *Journal of Biomedical Optics*. 2012;17:056016.
62. Lim L, Streutker CJ, Marcon N, Cirocco M, Lakovlev VV, DaCosta R, et al. Clinical study of ex vivo photoacoustic imaging in endoscopic mucosal resection tissues. In: *Photons Plus Ultrasound: Imaging and Sensing 2015*: 2015. International Society for Optics and Photonics: 932307.
63. Yang J-M, Li C, Chen R, Zhou Q, Shung KK, Wang LV. Catheter-based photoacoustic endoscope for use in the instrument channel of a clinical video endoscope. In: *Photons Plus Ultrasound: Imaging and Sensing 2015*: 2015. International Society for Optics and Photonics: 93230Y.
64. Yang J-M, Chen R, Favazza C, Yao J, Li C, Hu Z, et al. A 2.5-mm diameter probe for photoacoustic and ultrasonic endoscopy. *Opt Express*. 2012;20:23944–53.
65. Yang JM, Li C, Chen R, Zhou Q, Shung KK, Wang LV. Catheter-based photoacoustic endoscope. *J Biomedical Optics*. 2014;19:066001.
66. Yang JM, Favazza C, Chen R, Yao J, Cai X, Maslov K, et al. Simultaneous functional photoacoustic and ultrasonic endoscopy of internal organs in vivo. *Nat Med*. 2012;18:1297–302.
67. Yang JM, Li C, Chen R, Rao B, Yao J, Yeh CH, et al. Optical-resolution photoacoustic endomicroscopy in vivo. *Biomedical Optics Express*. 2015;9:323:918–32.
68. He H, Wissmeyer G, Ovsepian SV, Buehler A, Ntziachristos V. Hybrid optical and acoustic resolution photoacoustic endoscopy. *Opt Lett*. 2016;41:2708–10.
69. Bai X, Gong X, Hau W, Lin R, Zheng J, Liu C, et al. Intravascular optical-resolution photoacoustic tomography with a 1.1 mm diameter catheter. *Sci Found China*. 2015;9:e92463.
70. Lin R, Li Y, Chen J, Song L. Full field-of-view photoacoustic endoscopy in vivo. In: *Photons Plus Ultrasound: Imaging and Sensing 2017*: 2017. International Society for Optics and Photonics: 100643Y.
71. Huang D, Swanson EA, Lin CP, Schuman JS, Stinson WG, Chang W, et al. Optical coherence tomography. *Science*. 1991;254:1178–81.
72. Tsai TH, Fujimoto JG, Mashimo H. Endoscopic optical coherence tomography for clinical gastroenterology. *Diagnosics*. 2014;4:57–93.
73. Testoni PA. Optical coherence tomography. *Sci World J*. 2007;7:87–108.
74. Ozaki N, Childs DT, Sarma J, Roberts TS, Yasuda T, Shibata H, et al. Superluminescent diode with a broadband gain based on self-assembled InAs quantum dots and segmented contacts for an optical coherence tomography light source. *J Appl Phys*. 2016;119:083107.
75. Zhao H, Gao F, Tanikawa Y, Homma K, Yamada Y. Time-resolved diffuse optical tomographic imaging for the provision of both anatomical and functional information about biological tissue. *Appl Opt*. 2005;44:1905–16.
76. Lindmark G, Bo AN. Intravascular atherosclerotic imaging with combined fluorescence and optical coherence tomography probe based on a double-clad fiber combiner. *J Biomed Optics*. 2012;17:070501.
77. Tearney GJ, Boppart SA, Bouma BE, Brezinski ME, Weissman NJ, Southern JF, et al. Scanning single-mode fiber optic catheter-endoscope for optical coherence tomography. *Opt Lett*. 1996;21:543–5.
78. Donglin W, Linlai F, Xin W, Zhongjian G, Sean S, Can D, et al. Endoscopic swept-source optical coherence tomography based on a two-axis microelectromechanical system mirror. *J Biomed Optics*. 2013;18:86005.
79. Lee J, Chae Y, Ahn Y-C, Moon S. Ultra-thin and flexible endoscopy probe for optical coherence tomography based on stepwise transitional core fiber. *Biomed Optics Express*. 2015;6:1782–96.
80. Jingyu L, Chunyu Z, Xiaoying T, Tianxin G. Research Status and Prospect of Endoscopic OCT. *Laser & Optoelectronics Progress*. 2015;6.
81. Pahlevaninezhad H, Khorasaninejad M, Huang Y-W, Shi Z, Hariri LP, Adams DC, et al. Nano-optic endoscope for high-resolution optical coherence tomography in vivo. *Nat Photonics*. 2018;12:540.
82. Yun S, Tearney G, Boer J, De BB. Removing the depth-degeneracy in optical frequency domain imaging with frequency shifting. *Opt Express*. 2004;12:4822–8.
83. Herz PR, Chen Y, Aguirre AD, Schneider K, Hsiung P, Fujimoto JG, et al. Micromotor endoscope catheter for in vivo, ultrahigh-resolution optical coherence tomography. *Opt Lett*. 2004;29:2261–3.
84. Jiefeng X, Li H, Yicong W, Cobb MJ, Joo Ha H, Xingde L. High-resolution OCT balloon imaging catheter with astigmatism correction. *Opt Lett*. 2009;34:1943–5.
85. Yuan W, Brown R, Mitzner W, Yarmus L, Li X. Super-achromatic monolithic microprobe for ultrahigh-resolution endoscopic optical coherence tomography at 800 nm. *Nat Commun*. 2017;8:1531.
86. Mavadiashukla J, Fathi P, Liang W, Wu S, Sears C, Li X. High-speed, ultrahigh-resolution distal scanning OCT endoscopy at 800 nm for in vivo imaging of colon tumorigenesis on murine models. *Biomed Optics Express*. 2018;9:3731.
87. Ding Z, Qiu J, Shen Y, Chen Z, Bao W. Lens-free all-fiber probe with an optimized output beam for optical coherence tomography. *Opt Lett*. 2017;42:2814–7.
88. Wilder-Smith P, Lee K, Guo S, Zhang J, Osann K, Chen Z, et al. In vivo diagnosis of oral dysplasia and malignancy using optical coherence tomography: preliminary studies in 50 patients. *Lasers Surg Med*. 2009;41:353–7.
89. Hamdoon Z, Jerjes W, Upile T, McKenzie G, Jay A, Hopper C. Optical coherence tomography in the assessment of suspicious oral lesions: an immediate ex vivo study. *Photodiagn Photodyn Ther*. 2013;10:17–27.
90. Davies K, Connolly J, Dockery P, Wheatley A, Olivo M, Keogh I. Point of care optical diagnostic technologies for the detection of oral and oropharyngeal squamous cell carcinoma. *The surgeon*. 2015;13:321–9.

91. Sivak MV Jr, Kobayashi K, Izatt JA, Rollins AM, Ung-Runya R. High-resolution endoscopic imaging of the GI tract using optical coherence tomography. *Gastrointest Endosc.* 2000;51:474–9.
92. Cobb MJ, Hwang JH, Upton MP, Chen Y, Oelschlager BK, Wood DE, et al. Imaging of subsquamous Barrett's epithelium with ultrahigh-resolution optical coherence tomography: a histologic correlation study. *Gastrointest Endosc.* 2010;71:223–30.
93. Zhou C, Kirtane T, Tsai TH, Lee HC, Adler DC, Schmitt J, et al. Three-dimensional endoscopic optical coherence tomography imaging of cervical inlet patch. *Gastrointest Endosc.* 2012;75:675–7.
94. Suter MJ, Jillella PA, Vakoc BJ, Halpern EF, Minokenudson M, Lauwers GY, et al. Image-guided biopsy in the esophagus through comprehensive optical frequency domain imaging and laser marking: a study in living swine. *Gastrointest Endosc.* 2010;71:346–53.
95. Kiseleva E, Kirillin M, Feldchtein F, Vitkin A, Sergeeva E, Zagaynova E, et al. Differential diagnosis of human bladder mucosa pathologies in vivo with cross-polarization optical coherence tomography. *Biomedical Optics Express.* 2015;6:1464–76.
96. Isenberg G Jr, SM, Chak A, Wong RC, Willis JE, Wolf B, et al. Accuracy of endoscopic optical coherence tomography in the detection of dysplasia in Barrett's esophagus: a prospective, double-blinded study. *Digest World Core Med J.* 2006;6:2:825–31.
97. Evans CL, Xie XS. Coherent anti-Stokes Raman scattering microscopy: chemical imaging for biology and medicine. *Annu Rev Anal Chem.* 2008;1:883–909.
98. Li F, Palapattu GS, Zhou H, Wong KK, Gao L, Thrall MJ, et al. Label-free high-resolution imaging of prostate glands and cavernous nerves using coherent anti-Stokes Raman scattering microscopy. *Biomed Optics Express.* 2011;2:915–26.
99. Cheng J-X. Coherent anti-Stokes Raman scattering microscopy. *Appl Spectrosc.* 2007;61:197A-208A.
100. Marocco M. Coherent anti-Stokes Raman scattering microscopy in a microcavity. In: *APS Division of Atomic, Molecular and Optical Physics Meeting Abstracts: 2007.*
101. Hirose K, Fukushima S, Furukawa T, Hashimoto M. Coherent anti-Stokes Raman scattering imaging of nerves under rigid endoscopy. In: *JSAP-OSA Joint Symposia: 2017.* Optical Society of America: 5p_A409_408.
102. Ji XC, Xie XS. Coherent anti-stokes Raman scattering microscopy: instrumentation, theory, and applications. *Jphyschemb.* 2004;108:827–40.
103. Liu Z, Wang Z, Satira ZA, Xu C, Chen S, Xin J, et al. Development of fibre bundle probe for coherent anti-Stokes Raman scattering microendoscopy. *Electron Lett.* 2013;49:522–4.
104. Aviles-Espinosa RA, Santos SICO, Brodschelm A, Kaenders WG, Alonso-Ortega C, Artigas-García D, et al. Third-harmonic generation for the study of *Caenorhabditis elegans* embryogenesis. *J Biomed Optics.* 2010;15:046020–7.
105. Wang BG, König K, Halbhuber KJ. Two-photon microscopy of deep intravital tissues and its merits in clinical research. *J Microsc.* 2010;238:1–20.
106. Harpel K, Baker RD, Amirsolaimani B, Mehravar S, Vagner J, Matsunaga TO, et al. Imaging of targeted lipid microbubbles to detect cancer cells using third harmonic generation microscopy. *Biomed Optics Express.* 2016;7:2849–60.
107. Carriles R, Schafer DN, Sheetz KE, Field JJ, Cisek R, Barzda V, et al. Invited review article: imaging techniques for harmonic and multiphoton absorption fluorescence microscopy. *Rev Sci Instrum.* 2009;80:2834–69.
108. Perry SW, Burke RM, Brown EB. Two-photon and second harmonic microscopy in clinical and translational cancer research. *Ann Biomed Eng.* 2012;40:277–91.
109. Flusberg BA, Cocker ED, Piyawattanametha W, Jung JC, Cheung ELM, Schnitzer MJ. Fiber-optic fluorescence imaging. *Nat Methods.* 2005;2:941–50.
110. Mehravar S, Banerjee B, Chatrath H, Amirsolaimani B, Patel K, Patel C, et al. Label-free multi-photon imaging of dysplasia in Barrett's esophagus. *Biomedical optics express.* 2016;7:148–57.
111. Denk W, Strickler JH, Webb WW. Two-photon laser scanning fluorescence microscopy. *Science.* 1990;248:73–6.
112. Huland DM, Brown CM, Howard SS, Ouzounov DG, Ina P, Ke W, et al. In vivo imaging of unstained tissues using long gradient index lens multiphoton endoscopic systems. *Biomed Optics Express.* 2012;3:1077–85.
113. Brown CM, Rivera DR, Pavlova I, Ouzounov DG, Williams WO, Mohanan S, et al. In vivo imaging of unstained tissues using a compact and flexible multiphoton microendoscope. *J Biomed Optics.* 2012;17:040505.
114. Ducourthial G, Leclerc P, Mansuryan T, Fabert M, Brevier J, Habert R, et al. Development of a real-time flexible multiphoton microendoscope for label-free imaging in a live animal. *Sci Rep.* 2015;5:18303.
115. Liang W, Hall G, Messerschmidt B, Li M-J, Li X. Nonlinear optical endomicroscopy for label-free functional histology in vivo. *Light Sci Appl.* 2017;6:e17082.
116. Kim DY, Hwang K, Ahn J, Seo Y-H, Kim J-B, Lee S, et al. Lissajous scanning two-photon endomicroscope for in vivo tissue imaging. *Sci Rep.* 2019;9:3560.
117. Kobat D, Durst ME, Nishimura N, Wong AW, Schaffer CB, Xu C. Deep tissue multiphoton microscopy using longer wavelength excitation. *Opt Express.* 2009;17:13354–64.
118. Huland DM, Kriti C, Ouzounov DG, Jones JS, Nozomi N, Chris X. Three-photon excited fluorescence imaging of unstained tissue using a GRIN lens endoscope. *Biomed Optics Express.* 2013;4:652–8.
119. Akhouni F, Qin Y, Peyghambarian N, Barton JK, Kieu K. Compact fiber-based multi-photon endoscope working at 1700 nm. *Biomedical optics express.* 2018;9:2326–35.
120. Zhuo S, Yan J, Chen G, Shi H, Zhu X, Lu J, et al. Label-free imaging of basement membranes differentiates normal, precancerous, and cancerous colonic tissues by second-harmonic generation microscopy. *PLoS ONE.* 2012;7:e38655.
121. Zhang Y, Akins ML, Murari K, Xi J, Li M-J, Luby-Phelps K, et al. A compact fiber-optic SHG scanning endomicroscope and its application to visualize cervical remodeling during pregnancy. *Proc Natl Acad Sci U S A.* 2012;109:12878–83.
122. Barad Y, Eisenberg H, Horowitz M, Silberberg Y. Nonlinear scanning laser microscopy by third harmonic generation. *Appl Phys Lett.* 1997;70:922–4.

123. Yelin D, Silberberg Y. Laser scanning third-harmonic-generation microscopy in biology. *Opt Express*. 1999;5:169–75.
124. Zonios G, Perelman LT, Backman V, Manoharan R, Fitzmaurice M, Van DJ, et al. Diffuse reflectance spectroscopy of human adenomatous colon polyps in vivo. *Appl Opt*. 1999;38:6628–37.
125. Yodh A, Chance B. Spectroscopy and imaging with diffusing Light. *Phys Today*. 1995;48:34–40.
126. Le Q, Chuttani R, Pleskow DK, Turzhitsky V, Khan U, Zakharov YN, et al. Multispectral light scattering endoscopic imaging of esophageal precancer. *Light Sci Appl*. 2018;7:17174.
127. Lovat L, Bown S. Elastic scattering spectroscopy for detection of dysplasia in Barrett's esophagus. *Gastrointest Endosc Clin N Am*. 2004;14:507–17.
128. Mourant JR, Canpolat M, Brocker C, Espondo-Ramos O, Johnson TM, Matanock A, et al. Light scattering from cells: the contribution of the nucleus and the effects of proliferative status. In: *Optical Biopsy III: 2000*. International Society for Optics and Photonics: 33–42.
129. Coda S, Siersema PD, Stamp GW, Thillainayagam AV. Biophotonic endoscopy: a review of clinical research techniques for optical imaging and sensing of early gastrointestinal cancer. *Endosc Int Open*. 2015;3:E380-392.
130. Lovat LB, Johnson K, Mackenzie GD, Clark BR, Novelli MR, Davies S, et al. Elastic scattering spectroscopy accurately detects high grade dysplasia and cancer in Barrett's oesophagus. *Gut*. 2006;55:1078–83.
131. Johansson A, Kromer K, Sroka R, Stepp H. Clinical optical diagnostics—status and perspectives. *Med Laser Appl*. 2008;23:155–74.
132. Riddell RH, Goldman H, Ransohoff DF, Appelman HD, Fenoglio CM, Haggitt RC, et al. Dysplasia in inflammatory bowel disease: standardized classification with provisional clinical applications. *Hum Pathol*. 1983;14:931–68.
133. Mourant JR, Bigio IJ, Boyer JD, Johnson TM, Lacey J, Bohorfoush AG, et al. Elastic scattering spectroscopy as a diagnostic tool for differentiating pathologies in the gastrointestinal tract: preliminary testing. *J Biomed Optics*. 1996;1:192–9.
134. Ge Z, Schomacker KT, Nishioka NS. Identification of colonic dysplasia and neoplasia by diffuse reflectance spectroscopy and pattern recognition techniques. *Appl Spectrosc*. 1998;52:833–9.
135. Roy HK, Gomes A, Turzhitsky V, Goldberg MJ, Rogers J, Ruderman S, et al. Spectroscopic Microvascular Blood Detection From the Endoscopically Normal Colonic Mucosa: Biomarker for Neoplasia Risk. *Gastroenterology*. 2008;135:1069–78.
136. Canpolat M, Denkceken T, Karaveli S, Pestereli E, Erdoğan G, Özel D, et al. Detection of precancerous cervical conditions using elastic light single-scattering spectroscopy. In: *Biomedical Applications of Light Scattering IV: 2010*. International Society for Optics and Photonics: 75730V.
137. Hui F, Ollero M, Vitkin E, Kimerer LM, Cipolloni PB, Zaman MM, et al. Noninvasive sizing of subcellular organelles with light scattering spectroscopy. *Select Topics in Quantum Electr IEEE J*. 2003;9:267–76.
138. Perelman LT, Backman V, Wallace M, Zonios G, Manoharan R, Nusrat A, et al. Observation of periodic fine structure in reflectance from biological tissue: a new technique for measuring nuclear size distribution. *Phys Rev Lett*. 1998;80:627–30.
139. Chung-Chieh Y, Condon L, Tunnell JW, Martin H, Maxim K, Christopher FY, et al. Assessing epithelial cell nuclear morphology by using azimuthal light scattering spectroscopy. *Opt Lett*. 2006;31:3119–21.
140. Qiu L, Pleskow DR, Vitkin E, Leyden J, Ozden N, Itani S, et al. Multispectral scanning during endoscopy guides biopsy of dysplasia in Barrett's esophagus. *Nat Med*. 2010;16:603.
141. Alexey M, Linda N, Lorenz W, Urs U, Rebecca RK, Konstantin S. Fiber optic probe for polarized reflectance spectroscopy in vivo: design and performance. *J Biomed Optics*. 2002;7:388.
142. Zhang L, Pleskow DK, Turzhitsky V, Yee EU, Berzin TM, Sawhney M, et al. Light scattering spectroscopy identifies the malignant potential of pancreatic cysts during endoscopy. *Nat Biomed Eng*. 2017;1:0040.
143. Gurjar RS, Backman V, Perelman LT, Georgakoudi I, Badizadegan K, Itzkan I, et al. Imaging human epithelial properties with polarized light-scattering spectroscopy. *Nat Med*. 2001;7:1245–8.
144. Thosani N, Dayyeh BKA, Sharma P, Aslanian HR, Enestvedt BK, Komanduri S, et al. ASGE Technology Committee systematic review and meta-analysis assessing the ASGE preservation and incorporation of valuable endoscopic innovations thresholds for adopting real-time imaging-assisted endoscopic targeted biopsy during endoscopic surveillance of Barrett's esophagus. *Gastrointest Endosc*. 2016;83(684–698):e687.
145. Pyhtila JW, Graf RN, Wax A. Determining nuclear morphology using an improved angle-resolved low coherence interferometry system. *Opt Express*. 2003;11:3473–84.
146. Wax A, Yang C, Backman V, Badizadegan K, Boone CW, Dasari RR, et al. Cellular organization and substructure measured using angle-resolved low-coherence interferometry. *Biophys J*. 2002;82:2256–64.
147. Adam W, Changhuei Y, Vadim B, Maxim K, Dasari RR, Feld MS. Determination of particle size by using the angular distribution of backscattered light as measured with low-coherence interferometry. *J Opt Soc Am A*. 2002;19:737.
148. Zhu Y, Terry NG, Woosley JT, Shaheen NJ, Wax A. Design and validation of an angle-resolved low-coherence interferometry fiber probe for in vivo clinical measurements of depth-resolved nuclear morphology. *J Biomed Optics*. 2011;16:011003.
149. Terry NG, Zhu Y, Rinehart MT, Brown WJ, Gebhart SC, Bright S, et al. Detection of dysplasia in Barrett's esophagus with in vivo depth-resolved nuclear morphology measurements. *Gastroenterology*. 2011;140:42–50.
150. Ho D, Drake TK, Smith-McCune KK, Darragh TM, Hwang LY, Wax A. Feasibility of clinical detection of cervical dysplasia using angle-resolved low coherence interferometry measurements of depth-resolved nuclear morphology. *Int J Cancer*. 2017;140:1447–56.
151. Brown WJ, Pyhtila JW, Terry NG, Chalut KJ, D'Amico TA, Sporn TA, et al. Review and recent development of angle-resolved low-coherence interferometry for detection of precancerous cells in human esophageal epithelium. *IEEE J Select Top Quantum Electr*. 2008;14:88–97.
152. Stallmach A, Schmidt C, Watson A, Kiesslich R. An unmet medical need: advances in endoscopic imaging of colorectal neoplasia. *J Biophotonics*. 2011;4:482–9.
153. Kallaway C, Almond LM, Barr H, Wood J, Hutchings J, Kendall C, et al. Advances in the clinical application of Raman spectroscopy for cancer diagnostics. *Photodiagn Photodyn Ther*. 2013;10:207–19.

154. Shim MG, Song LM, Marcon NE, Wilson BC. In vivo near-infrared Raman spectroscopy: demonstration of feasibility during clinical gastrointestinal endoscopy. *Photochem Photobiol.* 2010;72:146–50.
155. Andrea M, Song LMWK, Shim MG, Marcon NE, Wilson BC. Diagnostic potential of near-infrared Raman spectroscopy in the colon: differentiating adenomatous from hyperplastic polyps. *Gastrointest Endosc.* 2003;57:396–402.
156. Huang Z, Teh SK, Zheng W, Lin K, Ho KY, Teh M, et al. In vivo detection of epithelial neoplasia in the stomach using image-guided Raman endoscopy. *Biosens Bioelectron.* 2011;26:383–9.
157. Wang J, Lin K, Zheng W, Ho KY, Teh M, Yeoh KG, et al. Simultaneous fingerprint and high-wavenumber fiber-optic Raman spectroscopy improves in vivo diagnosis of esophageal squamous cell carcinoma at endoscopy. *Sci Rep.* 2015;5:12957.
158. Wang J, Lin K, Zheng W, Ho KY, Teh M, Yeoh KG, et al. Fiber-optic Raman spectroscopy for in vivo diagnosis of gastric dysplasia. *Faraday Discuss.* 2016;187:377–92.
159. Haka AS, Volynskaya ZI, Gardecki JA, Nazemi J, Shenk R, Wang N, et al. Diagnosing breast cancer using Raman spectroscopy: prospective analysis. *J Biomed Optics.* 2009;14:054023.
160. Huang Z, McWilliams A, Lui H, McLean DI, Lam S, Zeng H. Near-infrared Raman spectroscopy for optical diagnosis of lung cancer. *Int J Cancer.* 2003;41:550–550.
161. Wang J, Lin K, Zheng W, Ho KY, Teh M, Yeoh KG, et al. Comparative study of the endoscope-based bevelled and volume fiber-optic Raman probes for optical diagnosis of gastric dysplasia in vivo at endoscopy. *Anal Bioanal Chem.* 2015;407:8303–10.
162. Mads Sylvest B, Wei Z, Khok YuH, Ming T, Khay Guan Y, Jimmy Bok YS, et al. Fiberoptic confocal raman spectroscopy for real-time in vivo diagnosis of dysplasia in Barrett's esophagus. *Gastroenterology.* 2014;146:27–32.
163. Bergholt MS, Lin K, Wang J, Zheng W, Xu H, Huang Q, et al. Simultaneous fingerprint and high-wavenumber fiber-optic Raman spectroscopy enhances real-time in vivo diagnosis of adenomatous polyps during colonoscopy. *J Biophotonics.* 2016;9:333–42.
164. Almond LM, Hutchings J, Lloyd G, Barr H, Shepherd N, Day J, et al. Endoscopic Raman spectroscopy enables objective diagnosis of dysplasia in Barrett's esophagus. *Gastrointest Endosc.* 2014;79:37–45.
165. Han Z, Zhang A, Wang X, Sun Z, Wang MD, Xie T. In vivo use of hyperspectral imaging to develop a noncontact endoscopic diagnosis support system for malignant colorectal tumors. *J Biomed Optics.* 2016;21:16001.
166. Hohmann M, Kanawade R, Klämpfl F, Douplik A, Mudter J, Neurath MF, et al. In-vivo multispectral video endoscopy towards in-vivo hyperspectral video endoscopy. *J Biophotonics.* 2016;10:553–64.
167. Ferris DG, Lawhead RA, Dickman ED, Holtzapple N, Miller JA, Grogan S, et al. Multimodal hyperspectral imaging for the noninvasive diagnosis of cervical neoplasia. *J Lower Genital Tract Dis.* 2010;5:65–72.
168. Regeling B, Thies B, Gerstner AOH, Westermann S, Müller NA, Bendix J, et al. Hyperspectral imaging using flexible endoscopy for laryngeal cancer detection. *Sensors.* 2016;16:1288.
169. Mycek M, Schomacker K, Nishioka N. Colonic polyp differentiation using time-resolved autofluorescence spectroscopy. *Gastrointest Endosc.* 1998;48:390–4.
170. Glanzmann T, Ballini JP, Bergh HVD, Wagnières G. Time-resolved spectrofluorometer for clinical tissue characterization during endoscopy. *Rev Sci Instrum.* 1999;70:4067–77.
171. Testoni PA, Mariani A, Mangiavillano B, Arcidiacono PG, Di PS, Masci E. Intraductal optical coherence tomography for investigating main pancreatic duct strictures. *Gastrointestinal Endoscopy.* 2006;63:AB89–AB89.
172. Seitz U, Freund J, Jaecle S, Feldchtein F, Bohnacker S, Thonke F, et al. First in vivo optical coherence tomography in the human bile duct. *Endoscopy.* 2001;33:1018–21.
173. Poneris JM, Tearney GJ, Shiskov M, Kelsey PB, Lauwers GY, Nishioka NS, et al. Optical coherence tomography of the biliary tree during ERCP. *Gastrointest Endosc.* 2002;55:84–8.
174. Sergeev A, Gelikonov V. In vivo endoscopic OCT imaging of precancer and cancer states of human mucosa. *Opt Express.* 1997;1:432–40.
175. Zuccaro G, Gladkova N, Vargo J, Feldchtein F, Zagaynova E, Conwell D, et al. Optical coherence tomography of the esophagus and proximal stomach in health and disease. *Am J Gastroenterol.* 2001;96:2633–9.
176. Jäckle S, Gladkova N, Feldchtein F, Terentjeva A, Brand B, Gelikonov G, et al. In vivo endoscopic optical coherence tomography of the human gastrointestinal tract—toward optical biopsy. *Endoscopy.* 2000;32:743–9.
177. Sivak MV, Kobayashi K, Izatt JA, Rollins AM, Ung-Runyawee R, Chak A, et al. High-resolution endoscopic imaging of the GI tract using optical coherence tomography. *Gastrointest Endosc.* 2000;51:474–9.
178. Zagaynova EV, Streltsova OS, Gladkova ND, Snopova LB, Gelikonov GV, Feldchtein FI, et al. In vivo optical coherence tomography feasibility for bladder disease. *J Urol.* 2002;167:1492–6.
179. Lerner SP, Goh AC, Tresser NJ, Shen SS. Optical coherence tomography as an adjunct to white light cystoscopy for intravesical real-time imaging and staging of bladder cancer. *Urology.* 2008;72:133–7.
180. Escobar PF, Belinson JL, White A, Shakhova NM, Feldchtein FI, Kareta MV, et al. Diagnostic efficacy of optical coherence tomography in the management of preinvasive and invasive cancer of uterine cervix and vulva. *Int J Gynecol Cancer.* 2010;14:470–4.
181. Gallwas JK, Turk L, Stepp H, Mueller S, Ochsenkuehn R, Friese K, et al. Optical coherence tomography for the diagnosis of cervical intraepithelial neoplasia. *Lasers Surg Med.* 2011;43:206–12.
182. Julian A, Matthew L, Ian W, Andrei Z, Sergey A, Stefan S, et al. In vivo size and shape measurement of the human upper airway using endoscopic longrange optical coherence tomography. *Opt Express.* 2003;11:1817–26.
183. Lam S, Standish B, Baldwin C, McWilliams A, Leriche J, Gazdar A, et al. In vivo optical coherence tomography imaging of preinvasive bronchial lesions. *Clin Cancer Res.* 2008;14:2006.
184. Wong BJF, Ms RPJ, Guo S, Ridgway JM, Mahmood U, Su J, et al. In vivo optical coherence tomography of the human larynx: normative and benign pathology in 82 patients. *Laryngoscope.* 2010;115:1904–11.
185. Burns JA. Optical coherence tomography: imaging the larynx. *Curr Opin Otolaryngol Head Neck Surg.* 2012;20:477–81.
186. Harm LP, Bonnema GT, Kathy S, Kenneth H, Molly B, Banon JK. Laparoscopic optical coherence tomographic imaging of human ovarian cancer. *Gynecol Oncol.* 2009;114:188–94.

187. Lovat LB, Johnson K, Mackenzie GD, Clark BR, Novelli MR, Davies S, et al. Elastic scattering spectroscopy accurately detects high grade dysplasia and cancer in Barrett's oesophagus. *Gut*. 2006;55:1078.
188. Dhar A, Johnson KS, Novelli MR, Bown SG, Bigio IJ, Lovat LB, et al. Elastic scattering spectroscopy for the diagnosis of colonic lesions: initial results of a novel optical biopsy technique. *Gastrointest Endosc*. 2006;63:257–61.
189. Mourant JR, Bigio IJ, Boyer J, Conn RL, Johnson T, Shimada T. Spectroscopic diagnosis of bladder cancer with elastic light scattering. *Lasers Surg Med*. 2010;17:350–7.
190. Georgakoudi I, Sheets EE, Müller MG, Backman V, Crum CP, Badizadegan K, et al. Trimodal spectroscopy for the detection and characterization of cervical precancers in vivo. *Am J Obstet Gynecol*. 2002;186:374–82.
191. Chang VT, Cartwright PS, Bean SM, Palmer GM, Bentley RC, Ramanujam N. Quantitative physiology of the precancerous cervix in vivo through optical spectroscopy. *Neoplasia*. 2009;11:325–32.
192. Chang TCB, S. M, Cartwright PS, Ramanujam N. Visible light optical spectroscopy is sensitive to neovascularization in the dysplastic cervix. *Journal of Biomedical Optics*. 2010;15:057006.
193. Wallace MB, Perelman LT, Backman V, Crawford JM, Fitzmaurice M, Seiler M, et al. Endoscopic detection of dysplasia in patients with Barrett's esophagus using light-scattering spectroscopy. *Gastroenterology*. 2000;119:677–82.
194. Lau C, Šćepanović O, Mirkovic J, McGee S, Yu CC, Stephen Fulghum J, et al. Re-evaluation of model-based light-scattering spectroscopy for tissue spectroscopy. *J Biomed Optics*. 2009;14:024031.
195. Backman V, Wallace MB, Perelman LT, Arendt JT, Gurjar R, Müller MG, et al. Detection of preinvasive cancer cells. *Nature*. 2000;406:35–6.
196. Bergholt MS, Zheng W, Ho KY, Teh M, Yeoh KG, So JB, et al. Fiber-optic Raman spectroscopy probes gastric carcinogenesis in vivo at endoscopy. *J Biophotonics*. 2013;6:49–59.
197. Bergholt MS, Zheng W, Lin K, Wang J, Xu H, Ren JL, et al. Characterizing variability of in vivo Raman spectroscopic properties of different anatomical sites of normal colorectal tissue towards cancer diagnosis at colonoscopy. *Anal Chem*. 2015;87:960–6.
198. Draga RO, Grimbergen MC, Vijverberg PL, van Swol CF, Jonges TG, Kummer JA, et al. In vivo bladder cancer diagnosis by high-volume Raman spectroscopy. *Anal Chem*. 2010;82:5993–9.
199. Rashid N, Nawaz H, Poon KW, Bonnier F, Bakhiet S, Martin C, et al. Raman microspectroscopy for the early detection of pre-malignant changes in cervical tissue. *Exp Mol Pathol*. 2014;97:554–64.
200. Lyng FM, Traynor D, Ramos IR, Bonnier F, Byrne HJ. Raman spectroscopy for screening and diagnosis of cervical cancer. *Anal Bioanal Chem*. 2015;407:8279–89.
201. O'Brien CM, Vargis E, Rudin A, Slaughter JC, Thomas G, Newton JM, et al. In vivo Raman spectroscopy for biochemical monitoring of the cervix throughout pregnancy. *Am J Obstet Gynecol*. 2018;218:S0002937818300796.
202. Bergholt MS, Lin K, Zheng W, Lau DP, Huang Z. In vivo, real-time, transnasal, image-guided Raman endoscopy: defining spectral properties in the nasopharynx and larynx. *J Biomed Optics*. 2012;17:077002.
203. Kan L, Wei Z, Wang J, Lim CM, Huang Z. Simultaneous fingerprint and high-wavenumber fiber-optic Raman endoscopy for in vivo diagnosis of laryngeal cancer. In: *Photonic Therapeutics & Diagnostics XII*: 2016.
204. Lim CM, Lin K, Zheng W, Huang Z. Real-time in vivo diagnosis of laryngeal carcinoma with rapid fiber-optic Raman spectroscopy. *Biomed Optics Express*. 2016;7:3705.
205. Lombardini A, Mytskaniuk V, Sivankutty S, Andresen ER, Chen X, Wenger J, et al. High-resolution multimodal flexible coherent Raman endoscope. *Light: Science & Applications*. 2018;7:10.
206. Yang Z, Albrow-Owen T, Cui H, Alexander-Webber J, Gu F, Wang X, et al. Single-nanowire spectrometers. *Science*. 2019;365:1017–20.
207. Shroff H, Galbraith CG, Galbraith JA, Betzig E. Live-cell photoactivated localization microscopy of nanoscale adhesion dynamics. *Nat Methods*. 2008;5:417–23.
208. Patterson GH, Lippincottschwartz J. A Photoactivatable GFP for selective photolabeling of proteins and cells. *Science*. 2002;297:1873–7.
209. Ba TQ, Lenne PF. Superresolution measurements in vivo: imaging Drosophila embryo by photoactivated localization microscopy. *Methods Cell Biol*. 2015;125:119.
210. Bates M, Huang B, Dempsey GT, Zhuang X. Multicolor super-resolution imaging with photo-switchable fluorescent probes. *Science*. 2007;317:1749–53.
211. Rust MJ, Bates M, Zhuang X. Sub-diffraction-limit imaging by stochastic optical reconstruction microscopy (STORM). *Nat Methods*. 2006;3:793.
212. Huang B, Wang W, Bates M, Zhuang X. Three-dimensional super-resolution imaging by stochastic optical reconstruction microscopy. *Science*. 2008;319:810–3.
213. Meyer L, Wildanger D, Medda R, Punge A, Rizzoli SO, Donnert G, et al. Dual-Color STED microscopy at 30-nm focal-plane resolution. *Small*. 2010;4:1095–100.
214. Klar T, Jakobs S, Dyba M, Egner A, Hell S. Fluorescence microscopy with diffraction resolution barrier broken by stimulated emission. *Proc Natl Acad Sci USA*. 2000;97:8206–10.
215. Willig KI, Rizzoli SO, Westphal V, Jahn R, Hell SW. STED microscopy reveals that synaptotagmin remains clustered after synaptic vesicle exocytosis. *Nature*. 2006;440:935.
216. Dyba M, Jakobs S, Hell SW. Immunofluorescence stimulated emission depletion microscopy. *Nat Biotechnol*. 2003;21:1303–4.
217. Cogswell CJ. Doubling the lateral resolution of wide-field fluorescence microscopy using structured illumination. *Proc SPIE*. 2000;3919:141–50.
218. Huang X, Fan J, Li L, Liu H, Chen L. Fast, long-term, super-resolution imaging with Hessian structured illumination microscopy. *Nat Biotechnol*. 2018;36:451.
219. Nixon-Abell J, Obara CJ, Weigel AV, Li D, Legant WR, Xu CS, et al. Increased spatiotemporal resolution reveals highly dynamic dense tubular matrices in the peripheral ER. *Science*. 2016;354:3928.
220. York AG, Chandris P, Nogare DD, Head J, Wawrzusin P, Fischer RS, et al. Instant super-resolution imaging in live cells and embryos via analog image processing. *Nat Methods*. 2013;10:1122–6.
221. Guo M, Chandris P, Giannini JP, Trexler AJ, Fischer R, Chen J, et al. Single-shot super-resolution total internal reflection fluorescence microscopy. *Nat Methods*. 2018;15:425.

222. Schermelleh L, Carlton PM, Haase S, Shao L, Winoto L, Kner P, et al. Subdiffraction multicolor imaging of the nuclear periphery with 3D structured illumination microscopy. *Science*. 2008;320:1332–6.
223. Gustafsson MGL. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. Short communication. *J Microsc*. 2010;198:82–7.
224. Li D, Shao L, Chen B-C, Zhang X, Zhang M, Moses B, et al. Extended-resolution structured illumination imaging of endocytic and cytoskeletal dynamics. *Science*. 2015;349:aab3500.
225. Chen TA, Chen LC, Hui LI, Jia YU, Gao YF, Zheng W. Structured illumination super-resolution microscopy technology:review and prospect. *Chinese Optics*. 2018;11:307–28.
226. Lambin P, Rios-Velazquez E, Leijenaar R, Carvalho S, Aerts HJWL. Radiomics: Extracting more information from medical images using advanced feature analysis. *Eur J Cancer*. 2007;43:441–6.
227. Hsu W, Markey MK, Wang MD. Biomedical imaging informatics in the era of precision medicine: progress, challenges, and opportunities. *J Am Med Inf Assoc Jamia*. 2013;20:1010–3.
228. Pinkert MA, Salkowski LR, Keely PJ, Hall TJ, Block WF, Eliceiri KW. Review of quantitative multiscale imaging of breast cancer. *J Med Imaging*. 2018;5:010901.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

