



# Magnetic Resonance Imaging Safety Board for India

Rijo M. Choorakuttil<sup>1</sup> Hrishikesh A. Kale<sup>2</sup> Bejoy Thomas<sup>3</sup>

<sup>1</sup>Department of Preventive Radiology and Integrated Diagnostics, AMMA Scans, AMMA Center for Diagnosis and Preventive Medicine Pvt Ltd, Kochi, Kerala, India

<sup>2</sup>Department of Radiology, Fortis Hospital, Mumbai, Maharashtra, India

<sup>3</sup>Department of Imaging Sciences and Intervention Radiology, Sree Chitra Tirunal Institute for Medical Sciences and Technology, Thiruvananthapuram, Kerala, India

Address for correspondence Rijo M. Choorakuttil, MD, Department of Preventive Radiology and Integrated Diagnostics, AMMA Scans, AMMA Center for Diagnosis and Preventive Medicine Pvt Ltd, Kochi 682036, Kerala, India (e-mail: rijomc@gmail.com).

Indian J Radiol Imaging 2024;34:788–789.

Magnetic resonance imaging (MRI), a widely used imaging technique in healthcare diagnostics, has excellent soft tissue contrast with nonionizing radiation exposure.<sup>1</sup> The three electromagnetic fields in MRI, static magnetic field, time-varying fields, gradient and radio-frequency fields, cause different safety risks.<sup>2–5</sup> Vertigo, nausea, projectile, biomedical implant and device-related events, ferromagnetic translational forces, peripheral nerve stimulation, heat deposition, and acoustic noise are some adverse events related to electromagnetic fields.<sup>2,6–11</sup> The use of MRI scanners with higher capabilities in diagnostic and teaching centers can lead to an increase in magnitude of safety-related incidents.<sup>4</sup> The consequences of adverse events can be reduced by improvements in the reporting and learning from adverse incidents, understanding their causes, and taking prompt action to prevent similar incidents in the future.<sup>4,12–15</sup> MR safety incidents are grossly underreported and more measures are needed to address MR safety issues.<sup>16–19</sup>

Adverse events associated with static magnetic fields include interactions with human tissue, projectiles, and malfunction or movement of implants or monitoring devices.<sup>4</sup> The risks associated with radio-frequency fields include specific absorption rate issues, tissue heating, burns, implant heating, and implant interference effects.<sup>4,20</sup> Peripheral nerve stimulation and acoustic noise, including potential interference with implants or monitoring devices, are the major risks associated with time-varying gradients.<sup>4,20</sup> The American Society of Testing and Materials International Committee has identified three MRI safety categories: MR safe, MR conditional and MR unsafe, and labels each with an associated icon.<sup>21</sup> Acute sensory effects, including a metallic

taste, nausea and vertigo, are of particular concern as 7T systems are introduced into clinical practice.<sup>11</sup> Thermal injuries were a major contributor (59%) in the Food and Drug Administration MAUDE MRI adverse event database.<sup>22</sup> Advances in other industries, as in clothing manufactured with invisible silver-embedded microfibers, can also cause newer forms of thermal injuries.<sup>23</sup> The United Kingdom Medicines and Healthcare Products Regulatory Agency recommends the removal of medicinal patches that may contain metal if removal will not compromise patient treatment.<sup>24</sup>

The reporting of critical incidents is important to further improve or refine safety standards and processes. Kihlberg et al reported that only 38% of critical incidents were reported and that several of the unreported incidents could have turned catastrophic.<sup>17</sup> Hansson et al observed that 16% of MR safety incidents had the highest severity or worst-case scenario score, that severe adverse events still exist despite safety protocols, and critical incidents are poorly shared within the team and are preventable.<sup>25</sup>

We recommend the formation of a dedicated MR safety board in India with clear roles, responsibilities, and statutory power. The leadership of the MR safety board must include the state, regional and national level leadership of the respective Indian Radiological and Imaging Association (IRIA) chapters, and include other stakeholders like medical physicists, radiation specialists, and biomedical engineers, in coordination with the health ministry at the local and national level. Leadership by Radiology associations can bring in the necessary expertise on clinical imaging protocol guidelines. The MR safety board must focus on the development of protocols and processes, systems, communication

article published online  
May 3, 2024

DOI <https://doi.org/10.1055/s-0044-1785682>.  
ISSN 0971-3026.

© 2024. Indian Radiological Association. All rights reserved.

This is an open access article published by Thieme under the terms of the Creative Commons Attribution-NonDerivative-NonCommercial-License, permitting copying and reproduction so long as the original work is given appropriate credit. Contents may not be used for commercial purposes, or adapted, remixed, transformed or built upon. (<https://creativecommons.org/licenses/by-nc-nd/4.0/>)

Thieme Medical and Scientific Publishers Pvt. Ltd., A-12, 2nd Floor, Sector 2, Noida-201301 UP, India

channels and training with the demarcation of responsibilities, and develop processes for accreditation including certification and re-certification, audits, and compliance. At the institutional level, the MR safety committee must be led by the Director of Radiology services, with a dedicated MR safety officer and level 1 and 2 MR personnel. Several specific and interlinked actions include the identification and appropriate zone demarcation implementation plans with the help of industry partners, educational programs for professionals that work in MR sites, development of clear MR safety procedures including screening forms and protocols, and rigorous but easily manageable incident-reporting systems with focus on prevention and learning from mistakes.<sup>19</sup> The National Health Mission India guidelines for medical devices for radioimaging departments focus on the clinical purpose and technical characteristics of the device, and environmental and operating conditions and do not explicitly include safety practices and audits as part of the guidelines. The MR safety board, under the leadership of IRIA, can complete this important missing link and help to develop standards for optimal and safe care for patients who undergo an MR imaging procedure in India.

#### Conflict of Interest

None declared.

#### References

- Kim SJ, Kim KA. Safety issues and updates under MR environments. *Eur J Radiol* 2017;89:7–13
- Keevil S. Safety in magnetic resonance imaging. *Int J Med Phys* 2016;4:26–34
- Hartwig V, Giovannetti G, Vanello N, Lombardi M, Landini L, Simi S. Biological effects and safety in magnetic resonance imaging: a review. *Int J Environ Res Public Health* 2009;6(06):1778–1798
- De Wilde JP, Grainger D, Price DL, Renaud C. Magnetic resonance imaging safety issues including an analysis of recorded incidents within the UK. *Prog Nucl Magn Reson Spectrosc* 2007;51:37–48
- Tsai LL, Grant AK, Morteale KJ, Kung JW, Smith MP. A practical guide to MR imaging safety: what radiologists need to know. *Radiographics* 2015;35(06):1722–1737
- Chakeres DW, de Vocht F. Static magnetic field effects on human subjects related to magnetic resonance imaging systems. *Prog Biophys Mol Biol* 2005;87(2–3):255–265
- Shellock FG, Crues JV. MR procedures: biologic effects, safety, and patient care. *Radiology* 2004;232(03):635–652
- Glover PM. Magnetic field-induced Vertigo in the MRI environment. *Curr Radiol Rep* 2015;3(29):1–7
- Schenck JF. Safety of strong, static magnetic fields. *J Magn Reson Imaging* 2000;12(01):2–19
- Coskun O. Magnetic resonance imaging and safety aspects. *Toxicol Ind Health* 2011;27(04):307–313
- Panych LP, Madore B. The physics of MRI safety. *J Magn Reson Imaging* 2018;47(01):28–43
- Jones DN, Benveniste KA, Schultz TJ, Mandel CJ, Runciman WB. Establishing national medical imaging incident reporting systems: issues and challenges. *J Am Coll Radiol* 2010;7(08):582–592
- Soop M, Fryksmark U, Köster M, Haglund B. The incidence of adverse events in Swedish hospitals: a retrospective medical record review study. *Int J Qual Health Care* 2009;21(04):285–291
- Baker GR, Norton PG, Flintoft V, et al. The Canadian Adverse Events Study: the incidence of adverse events among hospital patients in Canada. *CMAJ* 2004;170(11):1678–1686
- Vincent C, Neale G, Woloshynowych M. Adverse events in British hospitals: preliminary retrospective record review. *BMJ* 2001;322(7285):517–519
- Azuma M, Kumamaru KK, Hirai T, et al. A national survey on safety management at MR imaging facilities in Japan. *Magn Reson Med Sci* 2021;20(04):347–358
- Kihlberg J, Hansson B, Hall A, Tisell A, Lundberg P. Magnetic resonance imaging incidents are severely underreported: a finding in a multicentre interview survey. *Eur Radiol* 2022;32(01):477–488
- Piersson AD, Gorleku PN. A national survey of MRI safety practices in Ghana. *Heliyon* 2017;3(12):e00480
- Blankholm AD, Hansson B. Incident reporting and level of MR safety education: a Danish national study. *Radiography* 2020;26(02):147–153
- Smith JA. Hazards safety, and anesthetic considerations for MRI. *Top Companion Anim Med* 2010;25:98–106
- Shellock FG, Woods TO, Crues JV III. MR labeling information for implants and devices: explanation of terminology. *Radiology* 2009;253(01):26–30
- Delfino JG, Krainak DM, Flesher SA, Miller DL. MRI-related FDA adverse event reports: a 10-yr review. *Med Phys* 2019;46(12):5562–5571
- Watari T, Tokuda Y. MRI thermal burn injury: an unrecognized consequence of wearing novel, high-tech undergarments. *QJM* 2018;111(07):495–496
- Medicines and Healthcare Products Regulatory Agency. Safety guidelines for magnetic resonance imaging equipment in clinical use. United Kingdom. February 2021. Accessed March 23, 2024 at: [https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment\\_data/file/476931/MRI\\_guidance\\_2015\\_-\\_4-02d1.pdf](https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment_data/file/476931/MRI_guidance_2015_-_4-02d1.pdf)
- Hansson B, Olsrud J, Wilén J, Owman T, Höglund P, Björkman-Burtscher IM. Swedish national survey on MR safety compared with CT: a false sense of security? *Eur Radiol* 2020;30(04):1918–1926