

Journal section: *Prosthetic Dentistry*Publication Types: *Review*

doi:10.4317/jced.57025

<https://doi.org/10.4317/jced.57025>

Precision and practical usefulness of intraoral scanners in implant dentistry: A systematic literature review

Ignacio García-Gil ¹, Jorge Cortés-Bretón-Brinkmann ², Jaime Jiménez-García ³, Jesus Peláez-Rico ⁴, María-Jesús Suárez-García ⁵

¹ DDS, MS. Master Program Advanced Oral Implantology Europea University of Madrid. Master Program Buccofacial Prostheses and Occlusion, Faculty of Dentistry, Complutense University of Madrid

² DDS, PhD, MS. Researcher/Assistant Professor of Oral Surgery and Buccofacial Prostheses, Faculty of Dentistry, Complutense University of Madrid

³ DDS, PhD, MS. Full profesor and Chairman. Implantology Department, European University of Madrid. Surgical Director of CIRO, Madrid

⁴ DDS, PhD, MS. Assistant Professor. Department Conservative Dentistry and Buccofacial Prostheses, Faculty of Dentistry, Complutense University of Madrid

⁵ MD, DDS, PhD, MS. Full Professor. Department Conservative Dentistry and Buccofacial Prostheses, Faculty of Dentistry, Complutense University of Madrid

Correspondence:

Department of Conservative Dentistry and Buccofacial Prostheses
Faculty of Dentistry, Complutense University of Madrid
Pza Ramon y Cajal s/n, 28040 Madrid, Spain
garciagil.ignacio@gmail.com

García-Gil I, Cortés-Bretón-Brinkmann J, Jiménez-García J, Peláez-Rico J, Suárez-García MJ. Precision and practical usefulness of intraoral scanners in implant dentistry: A systematic literature review. *J Clin Exp Dent.* 2020;12(8):e784-93.

Received: 11/03/2020

Accepted: 14/05/2020

Article Number: 57025 <http://www.medicinaoral.com/odo/indice.htm>
© Medicina Oral S. L. C.I.F. B 96689336 - eISSN: 1989-5488
eMail: jced@jced.es

Indexed in:

Pubmed
Pubmed Central® (PMC)
Scopus
DOI® System

Abstract

Background: This systematic review aimed to evaluate the efficiency and accuracy of digital impression techniques for implant-supported restorations, and to assess their economic feasibility.

Material and Methods: Two independent electronic database searches were conducted in the Pubmed/MedLine, Cochrane Library, and Lilacs databases complemented by a manual search, selecting relevant clinical and in vitro studies published between 1st January 2009 and 28st February 2019. All type of studies (*in vivo* and *in vitro*) were included in this systematic review.

Results: Twenty-seven studies (8 *in vivo* and 19 *in vitro* studies) fulfilled the inclusion criteria. No meta-analysis was performed due to a large heterogeneity of the study protocols. The passive fit of superstructures on dental implants presented similar results between digital and conventional impression techniques. The studies considered that several factors influence the accuracy of implant impression taking: distance and angulation between implants, depth of placement, type of scanner, scanning strategy, characteristics of scanbody, and operator experience. Regarding the economic viability of intraoral scanning systems, only one study reported any benefit in comparison with conventional techniques.

Conclusions: Digital impressions of dental implants can be considered a viable alternative in cases of one or two contiguous dental implants. However, more studies are needed to evaluate the accuracy of digital techniques in full-arch implant-supported restorations.

Key words: *Intraoral scanner, dental implant, prosthesis, misfit, systematic review.*

Introduction

It is many years since the long-term success of dental implants was confirmed by Branemark *et al.* and Albrektsson *et al.* (1,2) Since then, numerous studies have described new surgical and prosthodontic techniques that aim to improve the clinical outcomes of implant-based treatments (3,4). In cases of implant-supported restorations, treatment success depends on the superstructure's passive fit, as failure to achieve adequate passive fit can produce biological and mechanical complications (5). Fit depends on the accuracy of implant impression taking, which may be realised using long-established conventional techniques or more recently introduced digital techniques. The fabrication of an implant-supported prosthesis in a conventional workflow must start with the aid of an implant transfer post. Conventional impression taking can be classified as direct (pick-up) or indirect (transfer).

With the introduction of digital technologies in dentistry, intraoral scanners can now be used for digital impression taking. According to the manufacturers, the use of intraoral scanners are a key element in the digital workflow, providing greater comfort for the patient, decreased turnaround time, and even a better cost-benefit ratio when compared to conventional techniques (6). But to date, no systematic literature review has been conducted to confirm the advantages of digital impression taking. In this context, this systematic literature review aimed to: (a) to determine if it is possible to achieve an adequate level of accuracy and efficiency using intraoral digital impression systems and to compare them with various conventional techniques for implant-supported restorations and (b) to assess the economic feasibility of digital techniques.

Material and Methods

This systematic review was conducted following PRISMA guidelines (7) and was registered in the Prospero database (trial no. CRD42015029504). The systematic review focused question was based on the PICO format (Population, Intervention, Comparison, Outcome) as follows:

Population: healthy adult human patients.

Intervention: conventional impression techniques.

Comparison: digital impression taking with intra-oral scanners.

Outcome: accuracy of impression and efficiency for fixed implant-supported restorations.

-Study Selection Criteria

In order to identify relevant articles, the following inclusion criteria were applied: Clinical studies without language restriction that evaluated the accuracy of digital impressions taken with intraoral scanners or compared digital impression taking with conventional impression taking in treatment protocols leading to fixed implant-su-

ported restorations. As the initial search generated only a few articles, and so insufficient scientific evidence, the search was extended to include *in vitro* studies. Finally, due to the heterogeneity of different articles it was not possible to implement a meta-analysis.

-Search Strategy

An electronic search was conducted in the following databases: PubMed, Cochrane Library, Lilacs. Key search terms were applied, combined using MeSH terms, to locate relevant articles published between 1st January 2009 and 28th February 2019. A additional manual search was conducted in the following journals: Clinical Implant Dentistry and Related Research, International Journal of Oral & Maxillofacial Implants, Journal of Oral Implantology, Clinical Oral Implants Research, Journal of Dental Research, Clinical Oral Implants Research, European Journal of Oral Implantology, Implant Dentistry, International Journal of Oral and Maxillofacial Surgery, Journal of Oral Implantology, Journal of Dentistry, Clinical Oral Investigations, and Journal of Oral Rehabilitation. All the corresponding authors of the studies identified were contacted in order to ascertain if additional articles or unpublished data were available.

-Data Collection and Quality Assessment

The search was carried out by two independent reviewers. Any disagreement between the reviewers (IGG and JC-BB) regarding data collection or quality assessment was resolved by consensus. Inter-reviewer reliability was assessed obtained a Kappa coefficient of 0.88 (CI 95%), values above 0.8 being considered a good level of agreement (8). To assess the quality of *in vivo* articles, the Critical Appraisal Skills Program (CASP) proposed by the Public Health Resource Unit (2006) was used, and only studies with an overall score of at least 50% were included in the review. Due to the small number of *in vivo* studies available, a duplicate search was performed to obtain *in vitro* studies. Although *in vitro* research cannot reproduce the dynamic environment of the stomatognathic system or human variability, pre-clinical experiments can provide important information about the properties and characteristics of a new material or technique. It is therefore necessary to conduct *in vitro* research of the highest possible standard. Efforts have been made in recent years to improve the quality of reporting in scientific literature (9,10). Although the CASP consort checklist was not originally designed for analyzing *in vitro* trials, in 2012 a modified consort checklist was published of items selected to assess reporting *in vitro* studies of dental materials.¹⁸ The authors of the present review adapted this checklist for the purpose of comparing the accuracy of different dental implant impression-taking techniques. Only studies with an overall score of at least 50% were included in the review.

Results

-Included Studies

An electronic search of the PubMed/MedLine, Cochrane Library and Lilacs databases located 1358 articles, which were reduced to 40 following title, abstract and full text analysis (PubMed/MedLine n=29; Lilacs n=7; Cochrane Library n=4). The articles from the different databases were compared to identify any duplicates, and a further 11 articles were eliminated on the basis of duplication (n=11) (Fig. 1). The ten remaining *in vivo* articles were categorized as follows: systematic reviews (n=5), randomized clinical trials (RCT) (n=1), prospective cohort studies (n=1), case-control studies (n=2), and case reports (n=1).

The corresponding authors of the selected studies were contacted via email of whom four returned additional data. However, no additional data was included for

analysis as all proved to be either replicate information or failed to meet the inclusion criteria. Due to the small number of *in vivo* studies available, the search was extended to include *in vitro* studies, using the same method, selecting 20 additional *in vitro* studies. These authors were also contacted via email, generating further data in three cases (n=3), but these were not included in the review for the same reasons as before. A modified CONSORT checklist of items for reporting *in vitro* studies was used to evaluate the risk of bias in the *in vitro* studies included (Fig. 2). When applying this modified CONSORT checklist to *in vitro* articles, points 5-9 could not be applied as they were designed to evaluate sample standardization. In the *in vitro* studies, the master model was the same in each study group, and so always standard. Of the articles evaluated, only one 19 did not exceed the minimum score for inclusion in the review




Fig. 1: Numbers of articles in databases.

Section	Checklist item
Abstract	A Structured summary of trial design, methods, results, and conclusions
Introduction	B Scientific background
	C Objectives and/or hypotheses
Methods	D The intervention for each group, with sufficient detail
	E How and when the primary and secondary measures are evaluated
	F Statistical methods used to compare groups for primary and secondary outcomes
Results	G For each primary and secondary outcome, results for each group, and the estimated size of the effect and its precision
Discussion	H Trial limitations, addressing sources of potential bias, imprecision, and, if relevant, multiplicity of analyses
Other information	I Sources of funding and other support
	J Where the full trial protocol can be accessed

Fig. 2: Modified consort Checklist *in vitro* studies.

(5/10), obtaining a score of 0/10 and so was eliminated (Table 1). Finally, the review included eight *in vivo* and 19 *in vitro* studies. The reasons for exclusion of various articles are specified in (Table 2).

-Outcomes

Implant impressions can be obtained using open or closed tray, with or without splinting, using different impression materials (CI) or scanbody + an intraoral scanner system (DI). In order to carry out a complete analysis of the included articles, the outcomes were divided according to the technique(s) investigated: DI (17 studies), or CI vs. DI (12 studies) (Tables 3,4).

1. DI

Seventeen studies used DI to take impression of dental implants: five systematic reviews, one case report, and eleven *in vitro* studies.

In Vivo

This case report describes DI in a patient with a fully edentulous jaw rehabilitated with six dental implants; three clinical tests were carried out to evaluate the accu-

racy of the superstructure: saliva intrusion, the Sheffield test, and the screw resistance test, although the authors did not specify the fit values obtained (11).

In Vitro

Eleven *in vitro* studies were located that investigated the accuracy of IOS, divided into three subgroups: partially edentulous (PE), completely edentulous (CE), and partially and completely edentulous models (CE-PE).

***In Vitro* - PE**

Three *in vitro* studies used DI-PE models (12-14). In 2012, Van der Meer *et al.* (12) carried out a study using a PE model with the aim of evaluating the accuracy of three different IOS. The authors concluded that the Lava COS was more accurate than the other IOS. Flugge *et al.* (13) employed two models bearing dental implants to compare the precision of three IOS with a laboratory scanner, obtaining a decrease in precision of the IOS when the distance between scan bodies increased, whereas with the dental lab scanner this was not dependent. Koch *et al.* (14). compared volumetric

Table 1: Modified checklist used to assess quality and risk of bias of *in vitro* studies.

Author	Abstract	Introduction		Methods					
	1	2. A.	2. B.	3	4	5	6	7	8
Chia <i>et al.</i> (30)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Menini <i>et al.</i> (34)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Marghalini <i>et al.</i> (31)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Imburgia <i>et al.</i> (22)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Amin <i>et al.</i> (33)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Chew <i>et al.</i> (29)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Vandeweghe <i>et al.</i> (20)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Gimenez-Gonzalez <i>et al.</i> (19)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Mangano <i>et al.</i> (21)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Flugge <i>et al.</i> (13)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Koch <i>et al.</i> (14)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Papaspyridakos <i>et al.</i> (32)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Gimenez <i>et al.</i> (18)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Gimenez <i>et al.</i> (17)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Gimenez <i>et al.</i> (16)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Lin <i>et al.</i> (28)	YES	YES	YES	YES	NO	NO	NO	NO	NO
Lee <i>et al.</i> (27)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Rauscher <i>et al.</i> (40)	NO	NO	NO	NO	NO	NO	NO	NO	NO
Gimenez <i>et al.</i> (15)	YES	YES	YES	YES	YES	NO	NO	NO	NO
Van der Meer <i>et al.</i> (12)	YES	YES	YES	NO	NO	NO	NO	NO	NO

Table 2: Studies excluded and reasons for exclusion.

STUDIES	REASON FOR EXCLUSION
Eliasson and Ortorp, 2012; Ramsey and Ritter, 2012; Al-Abdullah, Zandparsa <i>et al.</i> ; 2013; Howell, McGlumphy <i>et al.</i> , 2013; Nayyar, Yilmaz <i>et al.</i> , 2013; Derhalli, 2013; Abdel-Azim, Zandinejad <i>et al.</i> , 2014; Ajioka, Kihara <i>et al.</i> , 2016.	Use of healing abutments.
Ortorp, Jemt <i>et al.</i> , 2005; Bergin, Rubenstein <i>et al.</i> , 2013.	Intraoral Scanner no longer available commercially
Lee and Gallucci, 2013; Lee, Macarthur <i>et al.</i> , 2013; Joda, Lenherr <i>et al.</i> , 2017.	Evaluated efficiency or difficulty of scanning operation or scanning learning curve.
Wismeijer, Mans <i>et al.</i> 2014; Joda and Bragger, 2015; Schepke, Meijer <i>et al.</i> 2015.	Evaluated patient preference for conventional or digital technique
Stimmelmayer, Erdelt <i>et al.</i> , 2012; Stimmelmayer, Guth <i>et al.</i> , 2013; Jokstad and Shokati, 2015.	Use of extraoral scanner, not intraoral scanner

deviations between single tessellation language (STL) datasets of a master model, and milled model, and IOS from a previous single implant model. The authors concluded that direct digitization using the IOS presented less systematic error than physical model fabrication by milling from IOS.

In Vitro - CE

Six *in vitro* studies used digital techniques to scan CE models (15-20).

In the studies carried out by Giménez *et al.*, (15-19) precision was assessed in an edentulous maxillary model with different implant angulations. The same authors (15)

Table 3: *In vivo* studies.

AUTHOR	STUDY DESIGN	MAX/MB	POSITION OF DENTAL IMPLANT	EDENTULISM	N° IMPLANTS/PATIENT	IMPLANT BRAND	CONNECTION
Flügge et al. (39)	Systematic Review	-	-	-	-	-	-
Muhlemann et al. (38)	Systematic Review	-	-	-	-	-	-
Rutkunas et al. (36)	Systematic Review	-	-	-	-	-	-
Alikhasi et al. (37)	Systematic Review	-	-	-	-	-	-
Gherlone et al. (26)	Randomized Clinical Trial	MAX / MB #35, #42, #45 or #15, #12, #22, #25	-	Compl.	120 impl. / 25 patients	No data	External
Gherlone et al. (25)	Pilot Study	MAX / MB #35, #42, #45 or #15, #12, #22, #25	-	Compl.	56 impl. / 14 patients	Winsix (BioSAFin)	External
Joda et al. (6)	Prospective Cohorts	T MAX / MB #33, #43	Premolar and Molar	Partial.	1 implant / 20 patient	Straumann TL RN/WN	Internal
Andriessen et al. (24)	Pilot Study	MB	-	Compl.	2 implant / 25 patient	Straumann TL RN/WN	Internal
Moreno et al. (11)	Case Report	MB	No data	Compl.	6 implant / 1 patient	Exfel External, Megagen Implant	External
AUTHOR	ANGULATION IMPL.	PLACEMENT DEPTH	CONVENTIONAL or DIGITAL IMPRESSION	SPT or NSPT	SPT MATERIAL	SPT METHOD	MATERIAL IMPRESSION
Flügge et al. (39)	-	-	-	-	-	-	-
Muhlemann et al. (38)	-	-	-	-	-	-	-
Rutkunas et al. (36)	-	-	-	-	-	-	-
Alikhasi et al. (37)	Parallel / 30- 35°	No data	CI immediate loading + DI Immediate Loading	-	-	OT / IOS	PE / SB
Gherlone et al. (26)	Parallel / 30- 35°	No data	CI immediate loading + DI final restoration	-	-	OT / IOS	PE / SB
Joda et al. (6)	Parallel	No data	CI + DI	-	-	OT / IOS	SB / no data
Andriessen et al. (24)	No data	No data	DI	-	-	IOS	SB
Moreno et al. (11)	No data	No data	DI	-	-	IOS	PEEK SB
AUTHOR	SCANNER SYSTEM / SOFTWARE	NUMBER OF IMPRESSION	IMPRESSION ACCURACY				
Flügge et al. (39)	-	-	-				
Muhlemann et al. (38)	-	-	-				
Rutkunas et al. (36)	-	-	-				
Alikhasi et al. (37)	TRIOS // No data	No data	No data				
Gherlone et al. (26)	LAVA C.O.S // Software 2.1	No data	No data				
Joda et al. (6)	Itero // No data	No data	No data				
Andriessen et al. (24)	Itero // Software 3.5.0.	5	Mean Distance Error 226.0µm and Distance error ranging (21-638µm); Mean Angulation Error 2.582° and Angulation error ranging (0.123-9.563°)				
Moreno et al. (11)	LAVA C.O.S // No data	1	No data				
AUTHOR	SCANNER SYSTEM / SOFTWARE	NUMBER OF IMPRESSION	IMPRESSION ACCURACY				
Flügge et al. (39)	-	-	-				
Muhlemann et al. (38)	-	-	-				
Rutkunas et al. (36)	-	-	-				
Alikhasi et al. (37)	TRIOS // No data	No data	No data				
Gherlone et al. (26)	LAVA C.O.S // Software 2.1	No data	No data				
Joda et al. (6)	Itero // No data	No data	No data				
Andriessen et al. (24)	Itero // Software 3.5.0.	5	Mean Distance Error 226.0µm and Distance error ranging (21-638µm); Mean Angulation Error 2.582° and Angulation error ranging (0.123-9.563°)				
Moreno et al. (11)	LAVA C.O.S // No data	1	No data				
AUTHOR	ACCURACY METHOD	e-Mail	CASP				
Flügge et al. (39)	-	-	7				
Muhlemann et al. (38)	-	-	9				
Rutkunas et al. (36)	-	-	8				
Alikhasi et al. (37)	-	yes	8				
Gherlone et al. (26)	X-Ray (panoramic radiographs) with prostheses and a bar-implant connection and Sheffield test.	yes	7				
Gherlone et al. (25)	X-Ray (panoramic radiographs) with prostheses and a bar-implant connection.	yes	7				
Joda et al. (6)	No data	-	6				
Andriessen et al. (24)	Software Geomagic Quality 12, 3D Systems	-	7				
Moreno et al. (11)	Rocking of the framework not be detected, Sheffield test, X-Ray.	-	8				
AUTHOR	ACCURACY METHOD	e-Mail	CASP				
Flügge et al. (39)	-	-	7				
Muhlemann et al. (38)	-	-	9				
Rutkunas et al. (36)	-	-	8				
Alikhasi et al. (37)	-	yes	8				
Gherlone et al. (26)	X-Ray (panoramic radiographs) with prostheses and a bar-implant connection and Sheffield test.	yes	7				
Gherlone et al. (25)	X-Ray (panoramic radiographs) with prostheses and a bar-implant connection.	yes	7				
Joda et al. (6)	No data	-	6				
Andriessen et al. (24)	Software Geomagic Quality 12, 3D Systems	-	7				
Moreno et al. (11)	Rocking of the framework not be detected, Sheffield test, X-Ray.	-	8				

Conventional Impressions (CI), Digital Impressions (DI), Intraoral Scanning (IOS), Maxilla (MAX), Mandible (MB), Coordinate measurement machine (CMM), PEEK Scan Bodies (PEEK), Partially edentulous (Part), Completely edentulous (Compl.), Bone Level (BL), Tissue Level (TL), Splinted (SPT), Not Splinted (NSPT), Open Tray (OT), Closed Tray (CT), Scan Body (SB), Vinyl polysiloxane (VPS), Polyether (PE), Plaster Impression (P), Acrylic Resin (AR), Light polymerizing acrylic resin (LAR), Auto-polymerizing acrylic resin (AAR), Resin bars (RB), SD (Standard Deviation).

Table 4: *In vitro* studies.

AUTHOR	STUDY	MAX/MB	POSITION OF DENTAL IMPLANT	EDENTULISM	N° IMPL.	IMPLANT BRAND	CONNECTION	ANGULATION IMPL.
Menini et al. (34)	In vitro	MAX	#16, #13, #23, #26, in 1 cast	Compl.	4	Biomet 3i	External	Parallel
Chia et al. (30)	In vitro	MB	#44, #46 in 3 cast	Part.	2	Straumann BL	Internal	0, 10, 20 degrees buccolingual
Marghalini et al. (31)	In vitro	MB	#34, #36, in 2 casts	Part.	2	Nobel Biocare & Straumann TL	Internal	0, 30 degrees
Imburgia et al. (22)	In vitro	MAX	Model 1 (PEM): #23, #24 Model 2 (FEM): #16, #14, #11, #21, #24, #26	Part. & Compl.	2, 6	BT Safe Int. BTK-Biotec Implants	Internal	No data
Amin et al. (33)	In vitro	MB	#31, #32, #35, #42, #45, in 1 cast	Compl.	5	Straumann BL	Internal	#31, #32, #42: 0° // #35: 10° distally // #45: 15° distally
Chew et al. (29)	In vitro	MB	#44, #45, in 1 cast	Compl.	2	Straumann BL and TL	Internal	Parallel
Vandeweghe et al. (20)	In vitro	MB	#46, #44, #42, #32, #34, #36, in 1 cast	Compl.	6	IBT (Southern Implants)	External	Parallel
Gimenez-Gonzalez et al. (19)	In vitro	MAX	#17, #15, #12, #22, #25, #27, in 1 cast	Compl.	6	Biomet 3i	Internal	#17, #12, #22, #27: 0° // #15: 30° distally // #25: 30° mesially
Mangano et al. (21)	In vitro	MAX	Model 1 (PEM): #21, #24, #26 Model 2 (FEM): #16, #14, #11, #21, #24, #26	Part. & Compl.	3, 6	BTK implants	No data	No data
Flugge et al. (13)	In vitro	MB	Model 1: #36, #35 Model 2: #36, #35, #33, #45, #47	Part.	2, 5	Straumann BL	Internal	No data
Koch et al. (14)	In vitro	MAX	#25, in 1 cast	Part.	1	Straumann BL	Internal	No data
Papapayridakos et al. (32)	In vitro	MB	#31, #32, #35, #42, #45, in 1 cast	Compl.	5	Straumann BL	Internal	#31, #32, #42: 0° // #35: 10° distally // #45: 15° distally
Gimenez et al. (18)	In vitro	MAX	#17, #15, #12, #22, #25, #27, in 1 cast	Compl.	6	Biomet 3i	Internal	#17, #12, #22, #27: 0° // #15: 30° distally // #25: 30° mesially
Gimenez et al. (17)	In vitro	MAX	#17, #15, #12, #22, #25, #27, in 1 cast	Compl.	6	Biomet 3i	Internal	#17, #12, #22, #27: 0° // #15: 30° distally // #25: 30° mesially
Gimenez et al. (16)	In vitro	MAX	#17, #15, #12, #22, #25, #27, in 1 cast	Compl.	6	Biomet 3i	Internal	#17, #12, #22, #27: 0° // #15: 30° distally // #25: 30° mesially
Lin et al. (28)	In vitro	MB	#35, #37, in 4 casts	Part.	2	Straumann TL	Internal	Model 1: 0°; model 2: 15°; model 3: 30°; model 4: 45°
Lee et al. (27)	In vitro	MAX	#17, #15, #12, #22, #25, #27, in 1 cast	Compl.	1	Straumann BL	Internal	No data
Gimenez et al. (15)	In vitro	MAX	#17, #15, #12, #22, #25, #27, in 1 cast	Compl.	6	Biomet 3i	Internal	#17, #12, #22, #27: 0° // #15: 30° distally // #25: 30° mesially
Van der Meer et al. (12)	In vitro	MB	#36, #41, #46, in 1 cast.	Compl.	3	No data	No data	No data
AUTHOR	PLACEMENT DEPTH	CI or DI	SPT or NSPT	SPT MATERIAL	SPT METHOD	METHOD IMPRESSION	MATERIAL IMPRESSION	SCANNER SYSTEM / SOFTWARE
Menini et al. (34)	No data	CI + DI	SPT and NSPT	AR	-	OT-SPT, OT-NSPT, CI, IOS	PE, P / SB	True Definition // no data
Chia et al. (30)	BL	CI + DI	SPT	AAR	-	OT / IOS	VPS / SB	Trios COLOR // v3.1.4
Marghalini et al. (31)	No data	CI + DI	SPT	LAR	-	OT-SPT / IOS	PE / SB	CEREC Omnicam, True Definition// no data
Imburgia et al. (22)	No data	DI	-	-	-	IOS	PEEK Scan Bodies	CS 3600, Trios 3, CEREC Omnicam, True Definition// no data
Amin et al. (33)	No data	CI + DI	SPT	LAR	RB	OT / IOS	PE / SB	CEREC Omnicam / 4.1.1; True Definition// 4.1
Chew et al. (29)	No data	CI + DI	SPT	AAR	-	OT / IOS	PE / SB	Trios COLOR/v3.1.4; iTero / vHD 2.9; True Definition// no data
Vandeweghe et al. (20)	No data	DI	-	-	-	IOS	PEEK SB	True Definition, LAVA COS, ., Planerica PlanScan, Trios//no data
Gimenez-Gonzalez et al. (19)	#17, #15, #25, #27: 0mm // #12: 4mm // #22: 2mm	DI	-	-	-	IOS	PEEK SB	True Definition//no data
Mangano et al. (21)	no data	DI	-	-	-	IOS	PEEK SB	Trios, CS 3500, Zfx Intracran, Planerica PlanScan, Richardson TX//no data
Flugge et al. (13)	No data	DI	-	-	-	IOS	SB	True Definition, Itero, Trios//no data
Koch et al. (14)	No data	DI	-	-	-	IOS	SB	iTero//no data
Papapayridakos et al. (32)	No data	CI + DI	SPT and NSPT	LAR	RB	OT-SPT, OT-NSPT, IOS	PE / SB	Trios//no data
Gimenez et al. (18)	#17, #15, #25, #27: 0mm // #12: 4mm // #22: 2mm	DI	-	-	-	IOS	PEEK SB	LAVA Chairside Oral Scanner// V 0.3.0.2
Gimenez et al. (17)	#17, #15, #25, #27: 0mm // #12: 4mm // #22: 2mm	DI	-	-	-	IOS	PEEK SB	3D Progress IO scan, ZFX Intracran // Exocan-mh-2012-12-19
Gimenez et al. (16)	#17, #15, #25, #27: 0mm // #12: 4mm // #22: 2mm	DI	-	-	-	IOS	PEEK SB	CEREC AC Bluecam/CEREC 4.0
Lin et al. (28)	1mm coronal	CI + DI	No data	-	-	OT	No data	iTero/Straumann Cares 8.0
Lee et al. (27)	No data	CI + DI	NSPT	-	-	OT / IOS	VPS / SB	iTero//no data
Gimenez et al. (15)	#17, #15, #25, #27: 0mm // #12: 4mm // #22: 2mm	DI	-	-	-	IOS	PEEK SB	iTero//v. 4.5.0.151
Van der Meer et al. (12)	Gingival level	DI	-	-	-	IOS	PEEK SB	iTero// 3.5.0; LAVA COS/2.1; CEREC Bluecam/ 3.85

Conventional Impressions (CI), Digital Impressions (DI), Intraoral Scanning (IOS), Maxilla (MAX), Mandible (MB), Coordinate measurement machine (CMM), PEEK Scan Bodies (PEEK), Partially edentulous (Part.), Completely edentulous (Compl.), Bone Level (BL), Tissue Level (TL), Splinted (SPT), Not Splinted (NSPT), Open Tray (OT), Closed Tray (CT), Scan Body (SB), Vinyl polysiloxane (VPS), Polyether (PE), Plaster Impression (P).

concluded that the accuracy of impressions with iTero® IOS (Cadent) decreased with the increased length of the scanned section but the angulation of dental implants did not affect scanning accuracy. In 2015, Gimenez et al. (18) performed a study to assess the accuracy of two different IOS: ZFX Intracran® (Zimmer Biomet, Dachu Germany) and 3D Progress® (MHT, Verona, Italy), concluding that neither IOS was suitable for taking impressions of dental implants in the full arch. In the same way, Gimenez et al. (17) concluded that augmented and deep implant placement did not seem to decrease the system's

accuracy with Lava COS® intraoral scanning system (3M ESPE), although accuracy was higher than among experienced operators. Also in 2015, the same authors published another *in vitro* study of the CEREC AC Bluecam (Sirona) intraoral scanner. They concluded that neither angulation nor implant depth significantly affected scanner accuracy but operator experience did, with a tendency for less experienced operators to commit lower levels of error (16). In 2017, Gimenez-Gonzalez et al. (19) concluded that 3M True Definition IOS (3M ESPE) allows impression taking within the clinically accepta-

ble range *in vitro*, and they identified certain factors that influence accuracy: the amount of visible scanbody, distance and angulation between scan bodies; and operator experience. Vandeweghe *et al.* (20) carried out a study to evaluate the accuracy (trueness and precision) of four IOS in a mandibular model. The authors concluded that the 3M True Definition (3M ESPE) and Trios (3Shape) scanners presented acceptable levels of trueness and precision for dental implant impression taking, but that LAVA COS (3M ESPE) failed to obtain the minimum level of accuracy.

In Vitro –PE-CE

Two *in vitro* studies used digital techniques in (PE) and completely (FE) models (21, 22).

Mangano *et al.* (21) used two models (PEM and FEM) and four IOS. No differences were found in trueness and precision between the IOS; however, differences were found between the PEM and FEM with different IOS. In 2017, Imburgia *et al.* (22) also carried out a study with PEM and FEM, concluding that scanning with IOS was more accurate on the PEM than the FEM, findings that could have important clinical implications.

2. CI vs DI

The twelve articles that compared (CI) with (DI) included four *in vivo* and eight *in vitro* studies.

In Vivo

Comparisons between CI and DI were analyzed in four *in vivo* studies: a randomized crossover trial (23), two pilot studies (24, 25), and one randomized clinical trial (26). Andriessen *et al.* (24) assessed the accuracy of IOS (iTero) in edentulous mandibles rehabilitated with overdentures compared with an extraoral laboratory scanner. They concluded that inter-implant distance and implant angulation were critical factors influencing the accuracy of intraoral scanning. Gherlone *et al.* (25) carried out two cases series studies with a similar design: CE rehabilitated with the “All on Four” protocol. In 2015, CI and DI (LAVA C.O.S scanner, 3M ESPE) were performed, assessing the accuracy of metallic structures through the use of an X-Ray (intraoral digital radiographs). In 2016, the patients were allocated either to the control group (CI) or test group (DI, using the Trios (3Shape). The authors concluded that it is possible to manufacture cobalt-chromium full-arch rehabilitations using computer-aided design/computer-assisted manufacturing (CAD/CAM) from DI with satisfactory accuracy (26). Joda *et al.* (23) concluded that in addition to the multiple benefits offered by digital technology, DI allows a more efficient workflow in terms of cost when compared with CI.

In Vitro

The present review included eight *in vitro* studies divided into two subgroups: PE (27-31) and CE (32-34).

In Vitro - PE

Lee *et al.* (27) compared the models obtained with CI and DI, using a PE customized maxillary model. The

authors reported that there were no statistically significant differences between DI and CI, although statistically significant differences were found with the reference model. Lin *et al.* (28) used four different models with dental implants placed with varying angulation, fabricating definitive casts, observing a decreasing linear trend in deviations for both distance and angle measurements, suggesting that DI was more accurate when the implants diverged more. Marghalini *et al.* (31) found, in their study, which compared CI and DI, that impression techniques could affect accuracy, although within clinically acceptable levels.

Chew *et al.* (29) also evaluated this parameter in two sectional mandibular arch master models with different implants (Straumann Bone Level (BL), and Standard Plus Tissue Level (TL) Straumann, Basel, Switzerland). The authors concluded that for the BL test groups, CI presented significantly lower distortion than DI. In a similar study, Chia *et al.* (30) compared the accuracy of CI versus DI. The authors concluded that CI with 0° angulation between implants was associated with the highest accuracy, although no significant differences were found between different angulations when comparing CI and DI

In Vitro - CE.

In 2016, Papaspyridakos *et al.* (32) did not find significant differences between CI and DI compared with the master cast, with exception of Group II [(Open-Tray non-splinted at implant level) (OPNSPT-BL)]. Menini *et al.* (34) used a CE model with four low-profile implant analogs to evaluate impression accuracy in four different groups: CI (open tray-splinted vs. open tray-no splinted vs. closed tray) and DI (PEEK scanbody, True Definition [3M ESPE]). The authors found that DI achieved higher accuracy than CI. Amin *et al.* (33) used a mandibular model with five inter-foramen analogs in a stone master cast to compare the accuracy of CI and DI, concluding that DI was significantly more accurate than CI.

Discussion

This systematic review was designed to evaluate the accuracy and efficiency of IOS for dental implant impression taking, compared with CI, and to assess the economic feasibility of introducing digital techniques.

The *in vivo* evidence located in the first search was scarce, further reduced by risk of bias determined by the CASP quality assessment (8 studies). So in order to expand the amount of information on the topic, an additional search was carried out expanding the criteria to include *in vitro* studies. In order to critically appraise the works identified, the authors adapted a previously published checklist¹⁸ for assessing the potential bias of *in vitro* studies. This checklist was initially designed to evaluate the quality of *in vitro* studies investigating dental materials. However, applying the checklist to the stu-

dies selected in the present review, none fulfilled points 5 to 9. Point 5 of this checklist analyzes sample size, while points 6-9 analyze randomization (sequence generation, allocation concealment mechanism, implementation, and blinding). An *in vitro* study which evaluates dental implant impression-taking employs a previously designed model, with replicas of dental implants from which impressions are taken. The choice of model does not alter the results, as the models are manufactured industrially in advance and so the rate of error from model to model is negligible. In turn, there is no need for randomization, and sample size does not affect the results obtained. In this way, the authors of the present review used a modified version of the checklist published in 2012 by Faggion *et al.* (35), removing questions 5-9. In this way, the risk of bias and the quality of the *in vitro* studies analyzed were assessed by an appropriate, simple, and practical method.

Because of the variability between the *in vivo* studies included and the fact that it was unclear how passive fit had been evaluated, comparisons of the results were not possible (11, 23-26). Likewise, the *in vitro* studies reviewed could not be compared because of the different methods and IOS employed in both partial (27-31) and completely edentulous model (32-34). Nevertheless, most of the studies analyzed obtained results indicating sufficient accuracy, precision or trueness to guarantee adequate passive fit; especially on partially edentulous models. Several authors concluded that dental implant angulation and depth did not influence outcomes in terms of passive fit (15-17). Regarding the economic feasibility of DI, in comparisons between DI and CI, only a single *in vivo* study found that DI allowed a more efficient workflow than CI (23).

Nevertheless, four systematic reviews have been conducted evaluating if there are any significant differences in accuracy between CI and DI (one *in vitro* study (36), two *in vivo* (37, 38) studies and one that analyzed both *in vivo* and *in vitro* (39) studies) and all authors have concluded that the quality and quantity of the articles analyzed were insufficient. The present systematic review studied the same issue, analyzing both *in vivo* and *in vitro* studies, and adding one further objective, to determine the economic feasibility of DI.

Conclusions

Based on the data extracted from the articles analyzed in this systematic review, objectives could not be clearly and objectively addressed. It was not possible to determine which implant impression technique leads to better passive fit of superstructures. Digital techniques with intraoral scan impressions offer promising results, although improvements are still needed, particularly in full-arch impression taking. The available *in vivo* evidence is scarce, mainly case reports, which only provided low

quality evidence. Randomized clinical studies comparing conventional and digital implant impression techniques are needed to generate decisive evidence. Finally, insufficient evidence was found regarding the economic feasibility of DI for implant-supported restorations, so additional research is needed to clarify this.

References

1. Branemark PI, Hansson BO, Adell R, Breine U, Lindstrom J, Hallen O, et al. Osseointegrated implants in the treatment of the edentulous jaw. Experience from a 10-year period. *Scand J Plast Reconstr Surg Suppl.* 1977;16:1-132.
2. Albrektsson T, Zarb G, Worthington P, Eriksson AR. The long-term efficacy of currently used dental implants: a review and proposed criteria of success. *Int J Oral Maxillofac Implants.* 1986;1:11-25.
3. Tarnow DP, Cho SC, Wallace SS. The effect of inter-implant distance on the height of inter-implant bone crest. *J Periodontol.* 2000;71:546-9.
4. Araujo MG, Lindhe J. Dimensional ridge alterations following tooth extraction. An experimental study in the dog. *J Clin Periodontol.* 2005;32:212-8.
5. Pjetursson BE, Thoma D, Jung R, Zwahlen M, Zembic A. A systematic review of the survival and complication rates of implant-supported fixed dental prostheses (FDPs) after a mean observation period of at least 5 years. *Clin Oral Implants Res.* 2012;23:22-38.
6. Joda T, Bragger U. Complete digital workflow for the production of implant-supported single-unit monolithic crowns. *Clin Oral Implants Res.* 2014;25:1304-6.
7. Moher D, Liberati A, Tetzlaff J, Altman DG, Group P. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *J Clin Epidemiol.* 2009;62:1006-12.
8. McHugh ML. Interrater reliability: the kappa statistic. *Biochem Med (Zagreb).* 2012;22:276-82.
9. Kilkenny C, Browne WJ, Cuthill IC, Emerson M, Altman DG. Improving bioscience research reporting: the ARRIVE guidelines for reporting animal research. *PLoS Biol.* 2010;8:e1000412.
10. Schulz KF, Altman DG, Moher D, Group C. CONSORT 2010 statement: updated guidelines for reporting parallel group randomised trials. *PLoS Med.* 2010;7:e1000251.
11. Moreno A, Gimenez B, Ozcan M, Pradies G. A clinical protocol for intraoral digital impression of screw-retained CAD/CAM framework on multiple implants based on wavefront sampling technology. *Implant Dent.* 2013;22:320-5.
12. van der Meer WJ, Andriessen FS, Wismeijer D, Ren Y. Application of intra-oral dental scanners in the digital workflow of implantology. *PLoS One.* 2012;7:e43312.
13. Flugge TV, Att W, Metzger MC, Nelson K. Precision of Dental Implant Digitization Using Intraoral Scanners. *Int J Prosthodont.* 2016;29:277-83.
14. Koch GK, Gallucci GO, Lee SJ. Accuracy in the digital workflow: From data acquisition to the digitally milled cast. *J Prosthet Dent.* 2016;115:749-54.
15. Gimenez B, Ozcan M, Martinez-Rus F, Pradies G. Accuracy of a digital impression system based on parallel confocal laser technology for implants with consideration of operator experience and implant angulation and depth. *Int J Oral Maxillofac Implants.* 2014;29:853-62.
16. Gimenez B, Ozcan M, Martinez-Rus F, Pradies G. Accuracy of a Digital Impression System Based on Active Triangulation Technology With Blue Light for Implants: Effect of Clinically Relevant Parameters. *Implant Dent.* 2015;24:498-504.
17. Gimenez B, Ozcan M, Martinez-Rus F, Pradies G. Accuracy of a digital impression system based on active wavefront sampling technology for implants considering operator experience, implant angulation, and depth. *Clin Implant Dent Relat Res.* 2015;17:e54-64.
18. Gimenez B, Pradies G, Martinez-Rus F, Ozcan M. Accuracy of two digital implant impression systems based on confocal microscopy with

- variations in customized software and clinical parameters. *Int J Oral Maxillofac Implants*. 2015;30:56-64.
19. Gimenez-Gonzalez B, Hassan B, Ozcan M, Pradies G. An In Vitro Study of Factors Influencing the Performance of Digital Intraoral Impressions Operating on Active Wavefront Sampling Technology with Multiple Implants in the Edentulous Maxilla. *J Prosthodont*. 2017;26:650-5.
 20. Vandeweghe S, Vervack V, Dierens M, De Bruyn H. Accuracy of digital impressions of multiple dental implants: an in vitro study. *Clin Oral Implants Res*. 2017;28:648-53.
 21. Mangano FG, Veronesi G, Hauschild U, Mijiritsky E, Mangano C. Trueness and Precision of Four Intraoral Scanners in Oral Implantology: A Comparative in Vitro Study. *PLoS One*. 2016;11:e0163107.
 22. Imburgia M, Logozzo S, Hauschild U, Veronesi G, Mangano C, Mangano FG. Accuracy of four intraoral scanners in oral implantology: a comparative in vitro study. *BMC Oral Health*. 2017;17:92.
 23. Joda T, Bragger U. Patient-centered outcomes comparing digital and conventional implant impression procedures: a randomized crossover trial. *Clin Oral Implants Res*. 2016;27:e185-e189.
 24. Andriessen FS, Rijkens DR, van der Meer WJ, Wismeijer DW. Applicability and accuracy of an intraoral scanner for scanning multiple implants in edentulous mandibles: a pilot study. *J Prosthet Dent*. 2014;111:186-94.
 25. Gherlone EF, Ferrini F, Crespi R, Gastaldi G, Cappare P. Digital impressions for fabrication of definitive “all-on-four” restorations. *Implant Dent*. 2015;24:125-9.
 26. Gherlone E, Cappare P, Vinci R, Ferrini F, Gastaldi G, Crespi R. Conventional Versus Digital Impressions for “All-on-Four” Restorations. *Int J Oral Maxillofac Implants*. 2016;31:324-30.
 27. Lee SJ, Betensky RA, Gianneschi GE, Gallucci GO. Accuracy of digital versus conventional implant impressions. *Clin Oral Implants Res*. 2015;26:715-9.
 28. Lin WS, Harris BT, Elathamna EN, Abdel-Azim T, Morton D. Effect of implant divergence on the accuracy of definitive casts created from traditional and digital implant-level impressions: an in vitro comparative study. *Int J Oral Maxillofac Implants*. 2015;30:102-9.
 29. Chew AA, Esguerra RJ, Teoh KH, Wong KM, Ng SD, Tan KB. Three-Dimensional Accuracy of Digital Implant Impressions: Effects of Different Scanners and Implant Level. *Int J Oral Maxillofac Implants*. 2017;32:70-80.
 30. Chia VA, Esguerra RJ, Teoh KH, Teo JW, Wong KM, Tan KB. In Vitro Three-Dimensional Accuracy of Digital Implant Impressions: The Effect of Implant Angulation. *Int J Oral Maxillofac Implants*. 2017;32:313-21.
 31. Marghalani A, Weber HP, Finkelman M, Kudara Y, El Rafie K, Papaspyridakos P. Digital versus conventional implant impressions for partially edentulous arches: An evaluation of accuracy. *J Prosthet Dent*. 2018;119:574-579.
 32. Papaspyridakos P, Gallucci GO, Chen CJ, Hanssen S, Naert I, Vandenberghe B. Digital versus conventional implant impressions for edentulous patients: accuracy outcomes. *Clin Oral Implants Res*. 2016;27:465-72.
 33. Amin S, Weber HP, Finkelman M, El Rafie K, Kudara Y, Papaspyridakos P. Digital vs. conventional full-arch implant impressions: a comparative study. *Clin Oral Implants Res*. 2017;28:1360-7.
 34. Menini M, Setti P, Pera F, Pera P, Pesce P. Accuracy of multi-unit implant impression: traditional techniques versus a digital procedure. *Clin Oral Investig*. 2018;22:1253-62.
 35. Faggion CM Jr. Guidelines for reporting pre-clinical in vitro studies on dental materials. *J Evid Based Dent Pract*. 2012;12:182-9.
 36. Rutkunas V, Geciauskaite A, Jegelevicius D, Vaitiekunas M. Accuracy of digital implant impressions with intraoral scanners. A systematic review. *Eur J Oral Implantol*. 2017;10:101-20.
 37. Alikhasi M, Alsharbaty MHM, Moharrami M. Digital Implant Impression Technique Accuracy: A Systematic Review. *Implant Dent*. 2017;26:929-35.
 38. Muhlemann S, Kraus RD, Hammerle CHF, Thoma DS. Is the use of digital technologies for the fabrication of implant-supported reconstructions more efficient and/or more effective than conventional techniques: A systematic review. *Clin Oral Implants Res*. 2018;29:184-95.
 39. Flugge T, van der Meer WJ, Gonzalez BG, Vach K, Wismeijer D, Wang P. The accuracy of different dental impression techniques for implant-supported dental prostheses: A systematic review and meta-analysis. *Clin Oral Implants Res*. 2018;29:374-92.
 40. Rauscher O. Impressionless implant-supported restorations with Cerec 4.2. *Int J Comput Dent*. 2014;17:159-68.

Conflict of Interest

Non declared.