



Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.

# Journal Pre-proof

The need for measurement science in digital pathology

Marina Romanchikova, Spencer Thomas, Alex Dexter, Mike Shaw, Ignacio Partarrieau, Nadia Smith, Jenny Venton, Michael Adegun, David Brettle, Robert James Turpin



PII: S2153-3539(22)00751-9

DOI: <https://doi.org/10.1016/j.jpi.2022.100157>

Reference: JPI 100157

To appear in:

Received date: 17 June 2022

Accepted date: 7 November 2022

Please cite this article as: M. Romanchikova, S. Thomas, A. Dexter, et al., The need for measurement science in digital pathology, (2022), <https://doi.org/10.1016/j.jpi.2022.100157>

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2022 The Author(s). Published by Elsevier Inc. on behalf of Association for Pathology Informatics.

**Title:** The need for measurement science in digital pathology

**Author names and affiliations:**

Marina Romanchikova<sup>a</sup>, Spencer Thomas<sup>a</sup>, Alex Dexter<sup>a</sup>, Mike Shaw<sup>a</sup>, Ignacio Partarrieau<sup>a</sup>,  
Nadia Smith<sup>a</sup>, Jenny Venton<sup>a</sup>, Michael Adeogun<sup>a</sup>, David Brettle<sup>b</sup>, Robert James Turpin<sup>c</sup>

<sup>a</sup>: National Physical Laboratory, Hampton Road, Teddington, Middlesex, TW11 0LW, United Kingdom

<sup>b</sup>: Leeds Teaching Hospitals NHS Trust, St. James's University Hospital, Beckett Street, Leeds, West Yorkshire, LS9 7TF, United Kingdom

<sup>c</sup>: British Standards Institution, 389 Chiswick High Road, London, W4 4AL, United Kingdom

**Corresponding author:** Marina Romanchikova, Marina.Romanchikova@npl.co.uk

## Abstract

### Background

Pathology services experienced a surge in demand during the COVID-19 pandemic. Digitalisation of pathology workflows can help to increase throughput, yet many existing digitalisation solutions use non-standardised workflows captured in proprietary data formats and processed by black-box software, yielding data of varying quality. This study presents the views of a UK-led expert group on the barriers to adoption and the required input of measurement science to improve current practices in digital pathology.

### Methods

With an aim to support the UK's efforts in digitalisation of pathology services, this study comprised 1) a review of existing evidence, 2) an online survey of domain experts and 3) a workshop with 42 representatives from healthcare, regulatory bodies, pharmaceutical industry, academia, equipment and software manufacturers. The discussion topics included sample processing, data interoperability, image analysis, equipment calibration, and use of novel imaging modalities.

### Findings

The lack of data interoperability within the digital pathology workflows hinders data lookup and navigation, according to 80 % of attendees. All participants stressed the importance of integrating imaging and non-imaging data for diagnosis, while 80 % saw data integration as a priority challenge. 90 % identified the benefits of artificial intelligence and machine learning, but identified the need for training and sound performance metrics.

Methods for calibration and providing traceability were seen as essential to establish harmonised, reproducible sample processing and image acquisition pipelines. Vendor-neutral data standards were seen as a "must-have" for providing meaningful data for downstream analysis. Users and vendors need good practice guidance or evaluation of uncertainty, fitness-for-purpose, and reproducibility of artificial intelligence/machine learning tools. All of the above needs to be accompanied by an upskilling of the pathology workforce.

### Conclusions

Digital pathology requires interoperable data formats, reproducible and comparable laboratory workflows, and trustworthy computer analysis software. Despite high interest in the use of novel imaging techniques and artificial intelligence tools, their adoption is slowed down by the lack of guidance and evaluation tools to assess the suitability of these techniques for specific clinical question. Measurement science expertise in uncertainty estimation, standardisation, reference materials and calibration can help establishing reproducibility and comparability between laboratory procedures, yielding high quality data and providing higher confidence in diagnosis.

## Glossary

Terms “calibration”, “primary standard”, “reference material”, “repeatability”, “reproducibility”, “traceability” and “uncertainty” follow the International Vocabulary of Metrology, 3<sup>rd</sup> edition.<sup>1</sup>

Term	Definition
<b>AI/ML</b>	Artificial Intelligence/Machine Learning
<b>calibration</b>	Operation that first establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and then uses this information to establish a relation for obtaining a measurement result from an indication
<b>DICOM</b>	Digital Imaging and Communications in Medicine: A standard for capturing and exchanging medical imaging data
<b>DP</b>	Digital Pathology
<b>H&amp;E</b>	Haematoxylin and Eosin stain: The most frequent stain type used in histopathology
<b>interoperability</b>	Ability of two or more systems or components to exchange information and to use the information that has been exchanged <sup>1</sup>
<b>LIS</b>	Laboratory Information System
<b>measurand</b>	quantity intended to be measured
<b>NHS</b>	National Health Service: the main UK healthcare provider
<b>NMI</b>	National Measurement Institute
<b>NPL</b>	National Physical Laboratory
<b>omics</b>	Collective term for biology disciplines that study various molecules including genomics, proteomics, transcriptomics metabolomics and many others
<b>primary standard</b>	Measurement standard established using a primary reference measurement procedure, or created as an artifact, chosen by convention
<b>reference material</b>	Material, sufficiently homogeneous and stable with reference to specified properties, which has been established to be fit for its intended use in measurement or in examination of nominal properties
<b>repeatability</b>	Measurement precision under a set of conditions that includes the same measurement procedure, same operators, same measuring system, same operating conditions and same location, and replicate measurements on the same or similar objects over a short period of time
<b>reproducibility</b>	Measurement precision under a set of conditions that includes different locations, operators, measuring systems, and replicate measurements on the same or similar objects
<b>SOP</b>	Standard Operating Procedure
<b>traceability</b>	Property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty
<b>uncertainty</b>	Non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used
<b>WSI</b>	Whole Slide Imaging

## 1 INTRODUCTION

Pathology is the backbone of diagnostic medicine and contributes to the majority of clinical pathways. As the pressure on health systems increases worldwide, the need for pathology services follows suit. In the UK, the Royal College of Pathologists reported a 4.5% year-on-year increase in demand for pathology services since 2007.<sup>2</sup> The COVID-19 pandemic has amplified already mounting pressures on pathologists and increased their willingness to adopt remote ways of working.<sup>2</sup> These pressures can be alleviated by the digitalisation of pathology services, and some early adopters have shown increased throughput of cases by 21% per year.<sup>3</sup>

Aside from increased throughput, digital pathology (DP) promises many benefits including improved diagnosis,<sup>4</sup> new insights into disease phenotypes and mechanisms<sup>5,6</sup>, validation of diagnosis,<sup>7</sup> and

<sup>1</sup>: <https://www.iso.org/obp/ui/#iso:std:iso:ts:27790:ed-1:v1:en:term:3.39>, assessed 14/03/2022.

<sup>2</sup>: <https://www.rcpath.org/discover-pathology/public-affairs/the-pathology-workforce.html>, assessed 20/02/2022.

the use of artificial intelligence (AI) to support image quality assurance, prioritisation, review and diagnosis. The rollout of DP poses a set of challenges including imaging data management, computational complexity, interoperability, image comparability, device and data quality assurance (QA), and deployment of artificial intelligence/machine learning (AI/ML) solutions. It is possible that the backlog of patients caused by the pandemic may force many services into premature adoption of DP solutions that are based on unstandardized processes, “black box” software, data of variable quality, and unknown quality.

Measurement science (metrology) concepts such as traceability, calibration, reproducibility, and uncertainty quantification can be used to address challenges in DP implementation and advance both clinical and research pathology to the next level. In the UK, metrological traceability underpins medical radiation dosimetry, where radiation dose measurements on therapy units are traceable to the national standard. The need for metrology has also been realised in quantitative imaging, where accurate estimates of pixel values and associated uncertainties are used as tissue or disease type biomarkers.<sup>8</sup>

This work presents the findings of an online survey and a workshop conducted by the UK's National Physical Laboratory in 2021. It collates the views of DP experts from clinical, pre-clinical, research, industry and regulatory authorities on the metrology support required to address the key challenges in the area.

### 1.1. DIGITAL PATHOLOGY WORKFLOW

DP workflow stages can be roughly grouped into sample preparation, image acquisition, data analysis, and diagnosis. All of these stages pose challenges some of which are unique to DP, while others are shared with conventional on-microscope histopathology.

During the sample preparation stage a biopsy may undergo dehydration, clearing, fixation, sectioning, embedding and staining of tissue slices on glass slides, as well as possible transport in between different steps. Variations in sample preparation result in high variability in the resulting images and thus can lead to differences in the diagnosis. Slide quality assurance (QA) is typically performed manually via visual assessment and its outcome is dependant upon the experience of the assessor, although AI/ML tools are being considered for early-stage QA.<sup>9</sup> Unsurprisingly, sample processing presents the largest source of the variability within the DP workflow,<sup>10</sup> with staining or “batch” variability being one of the major obstacles in producing consistent and comparable images.<sup>11</sup>

To digitise a glass slide, a whole slide imaging scanner (WSI) scans the stained tissue slides and produces high resolution images of between 1-4 Gigabyte in size. WSI scanners typically include a mechanical stage to feed the slides, a light source, optics, and a digital camera sensor. To date, the scope and frequency of WSI calibration and routine tests vary between vendors and laboratories. The 2018 report points out that, while WSI should undergo ongoing QA, since subjective perception of image quality cannot indicate the image is “fit-for-diagnosis”, research in the calibration area “is sparse”.<sup>12</sup> The main calibration areas include (1) dimensionality, (2) illumination and (3) colour, with the latter being the most widely recognised challenge. Colour calibration comprises (a) internal colour calibration which involves removing the variabilities arising from the scanning process itself, and (b) external colour calibration that focusses on the standardisation of the display, accounting for the monitor's effect on perceived colour and the viewing environment. Existing literature indicates that colour management in DP is challenging due to the lack of standards,<sup>13</sup> while the digitisation of the slide introduces further lack of colour control and compounds the issue.

In addition to conventional stains such as haematoxylin-eosin (H&E) and periodic-acid-Shiff (PAS) that focus on tissue morphology, pathologists can use a range of other techniques. These may include “omics” to provide genetic, proteomic or metabolomic information, using blood, urine, liquid biopsies,<sup>14</sup> or novel imaging techniques such as mass spectrometry imaging, imaging mass cytometry, immunohistochemistry, Raman microscopy, and many others. Combining multiple techniques offers advantages over histopathology, since a combination of measurements makes it possible to differentiate diseases where single measurands do not.<sup>15</sup> Such multi-dimensional data can improve accuracy of diagnosis, including tumour grading<sup>16</sup> and measuring intra-tumour heterogeneity.<sup>17</sup>

The major barriers to wide clinical uptake of these techniques can be broken down into three categories: the data collection and sample preparation differ between modalities and can be more difficult to compare;<sup>18,19</sup> the data themselves are much larger and more complex meaning data analysis is more challenging;<sup>20</sup> and, importantly, these methods are not yet fully understood within a clinical setting and are not yet approved by regulating bodies for clinical use.<sup>21</sup>

There is a growing body of research on AI/ML applications for various stages of the DP workflow, including diagnosis, pre-diagnostic slide QA,<sup>22,23</sup> and colour normalisation.<sup>24</sup> A PubMed search on “AI digital pathology” shows a fourfold growth from 43 publications in 2016 to 190 in 2021. At the same time, the uptake of AI/ML tools in clinical routine is relatively slow. The roadblocks to adoption include large image size, image artefacts, colour variations, regulatory approval, as well as lack of access to large well-annotated datasets, lack of protocols for training and validation of algorithms, and challenges regarding interpretation/explainability of results.<sup>4,25,26</sup> Attempts to increase the amount of data for algorithm development are being undertaken in several national and European projects that aim to create a large-scale AI/ML WSI data processing cloud for algorithm developers and clinical users.<sup>3,4</sup>

Data exchange connects all stages of the DP workflow. WSI data-related challenges can be subdivided into two categories: (1) lack of vendor-neutral data format and (2) lack of metadata standards to capture the knowledge about how the image was obtained. The lack of widespread adoption of vendor-neutral WSI formats makes data access, processing and exchange challenging and highly dependent on laboratory, equipment, and software setup. Slow adoption of vendor-neutral formats such as DICOM or OME-TIFF has negative impact on sharing of WSI data within and outside of a hospital.<sup>27</sup>

The heterogeneity of WSI formats causes poor integration of DP within the hospital IT infrastructure. A notable consequence is the inadequate data exchange between WSI scanners and a laboratory information system (LIS) that holds sample processing and patient metadata vital for image interpretation. Although proprietary WSI/LIS solutions exist, they do not solve the issue of inter-institution or inter-system data exchange that is the cornerstone of DP workflow. Efforts to create working specifications for WSI data exchange are being undertaken by the Integrating the Healthcare Enterprise (IHE) initiative in collaboration with DICOM working group 26.<sup>28</sup>

Poor interoperability of WSI data is compounded by the lack of consensus on minimum metadata on sample handling, scanner settings, and image processing that limits the diagnostic value of WSI data, its re-use and sharing. Failure to supply appropriate metadata and lack of reporting standards have been recognised as key contributors to the reproducibility crisis in digital medicine,<sup>29</sup> and open repositories such as BioArchive attempt to tackle the issue by specifying their own metadata requirements.<sup>30</sup>

<sup>3</sup>. <https://en.empaia.org/>, accessed 28/02/2022.

<sup>4</sup>. <https://bigpicture.eu/>, accessed 28/02/2022.

## 1.2. THE ROLE OF STANDARDS


The uptake of DP is greatly limited by the lack of systems interoperability and agreed good practice for data exchange.<sup>27</sup> Added to this is the complexity surrounding AI/ML solutions, and the need to ensure they are safe, secure, and perform in accordance with their intended use.

The adoption of standards - agreed organisational and technical best practices - can help to address the barrier to adoption of DP. Standards are a foundation for innovation, providing a key mechanism by which to diffuse knowledge of what good looks like. They provide a route to consensus between collaborating businesses and expert organizations, and enable policymakers to educate the market. Standards are voluntary agreements of good practice, commonly taking the form of guidance, codes of practice and specifications. They can inform both the technical and procedural aspects of data exchange, along with the conformity to regulations.

Standards play an important role in DP, from improving data management and interoperability, to ensuring the quality and competence of medical laboratories, and ensuring the safety and performance of medical diagnostic equipment and software. On a country level, standards sit within the National Quality Infrastructure, alongside testing and measurement to ensure validity and consistency, and certification and accreditation that ensures standards are applied correctly and with competency by the personnel who carry out testing and inspection.

## 1.3. ROLE FOR METROLOGY

Metrology is the science of measurement, embracing both experimental and theoretical determinations at any level of uncertainty in any field of science. Metrology methods help to obtain a reproducible measurement result with known uncertainty that is traceable to a reference and the International System of Units (SI).<sup>31</sup> Metrology principles that could be applied to DP workflow include traceability, calibration, uncertainty analysis, reproducibility, and comparison (Figure 1).



**Figure 1:** Critical metrology principles applied to a digital pathology workflow. The same principles apply for non-WSI modalities e.g. RAMAN, MSI, omics, radiology etc.

A well-established example of metrology applied in medicine is the traceability chain for UK radiation dosimetry, within which all hospitals delivering radiation therapy own a measurement device (secondary standard) that is regularly calibrated against the national primary standard. In the DP context, a similar traceability chain could be established for WSI image acquisition. Each WSI scanner would be regularly characterised using test objects with known physical properties to measure inter- and intra-device variability. These test objects could be traced to the national standard to assess variations between test objects. Having accurate and up-to-date information about WSI scanner performance, e.g., scanner calibration data, would aid image normalisation and colour re-scaling to match user preferences. Most importantly, normalised WSI datasets could be used to define reproducible quantitative imaging biomarkers with known uncertainties, and could be used to train AI/ML algorithms.

Efforts to apply metrology methods to obtain digital quantitative biomarkers are underway in medical imaging,<sup>32-34</sup> where they are led by the Quantitative Imaging Biomarkers Alliance.<sup>5</sup>

National Measurement Institutes such as NPL are well-placed to develop reference methods and standards for DP and biosciences. The following sections describe the landscape study carried out by NPL in 2021 to gain the views of DP stakeholders on what metrology interventions are most needed by the community.


## 2 METHODS

Participants for the study were selected through an established network of contacts, through word of mouth, and through identification of stakeholders via internet searches. We aimed to recruit at least four representatives from the professional backgrounds summarised in Figure 2, while ensuring there were at least two organisations in each category. The resulting cohort of participating organisations is presented in Table 1.

The online survey conducted in October-December 2020 contained 31 questions on participants background (n=6), imaging data acquisition and storage (n=10), QA and regulations (n=5), AI and ML applications (n=4), future directions in the area (n=4) and further comments (n=2). The survey was accompanied by a set of interviews with participants from healthcare (n=3), pre-clinical (n=2), software vendors (n=1) and WSI device manufacturers (n=3).

---

<sup>5</sup>: <https://www.rsna.org/research/quantitative-imaging-biomarkers-alliance>, accessed 28/02/2022.



**Figure 2:** Information collection process used in this study.

The survey and interview findings were used to set out five discussion groups for the workshop: (1) sample preparation and processing, (2) equipment calibration, (3) image processing and analysis, (4) omics and novel imaging techniques and (5) data integration.

The workshop attendees were allocated to the discussion groups so that 1) each group included at least one participant from healthcare, pre-clinical, device/software vendor and regulatory background and 2) participants in each group represented different organisations.

Two rounds of group discussions were used to identify the challenges in the area and to outline whether metrology support is required. Group discussions were led by moderators with expertise in the subject matter and captured by designated notetakers. The intermittent findings of both discussion sessions were shared with all workshop attendees during the joint session to ensure adequate coverage. Following the workshop, the notes from each group discussion were shared with the group members for review and collated to a single report.

### 3 RESULTS

Table 1 lists the organisations that took part in the workshop and survey. A comprehensive summary of the findings is presented in the NPL report “Metrology for Digital Pathology”.<sup>35</sup>

**Table 1:** Workshop participants background and their allocation to discussion groups. Asterisk (\*) denotes participants who took part in a phone interview before the workshop.

Organisation name	Organisation type	Workshop discussion groups	Survey
Ashford and St Peter's Hospitals NHS Foundation Trust	Healthcare provider	Equipment Calibration	Y
AstraZeneca	Industry (pharmaceutical)	Image Analysis, Novel Imaging, Sample Processing	Y
British Standards Institution	Standards bodies	Data Integration	N*
Cambridge University Hospitals	Healthcare provider	Sample Processing	Y
Charles River Laboratories	Industry (pharmaceutical)	Image Analysis	Y
DesAcc	Industry (software)	Data Integration	Y*
DICOM	Standards bodies	Data Integration	N
FFEI Ltd	Industry (equipment)	Equipment Calibration	N
GE Healthcare	Industry (software)	Image Analysis	N
GlaxoSmithKline	Industry (pharmaceutical)	Data Integration, Equipment Calibration, Sample Processing	Y*
IHE	Standards bodies	Data Integration	N
Leeds Teaching Hospitals NHS Trust	Healthcare provider	Equipment Calibration	N*
Leica Biosystems	Industry (equipment)	Image Analysis	Y
NHS Digital	Healthcare provider	Novel imaging	Y
Nikon Corporation	Industry (equipment)	Equipment Calibration	Y
Norfolk and Norwich University Hospitals NHS Foundation Trust	Healthcare provider	Data Integration	Y
Paige.AI	Industry (software)	Data Integration, Image Analysis	Y
Royal Brompton and Harefield Hospitals	Healthcare provider	Sample Processing	Y
Royal Surrey County Hospital	Healthcare provider	Equipment Calibration	N*
Royal Veterinary College	Academia	Sample Processing	Y*
Smith + Nephew	Industry (equipment)	Sample Processing	Y
Surrey Heartlands	Healthcare provider	Novel Imaging	Y
University College London	Academia	Image Analysis	Y
University of Cambridge	Academia	Image Analysis, Equipment Calibration	N*
University of Dundee	Academia	Data Integration	Y
University of Oslo	Academia	Image Analysis	N
University of Surrey	Academia	Image Analysis	Y*
Zeiss	Industry (equipment)	Equipment Calibration	N*

### 3.1 ONLINE SURVEY FINDINGS

Twenty-four respondents with backgrounds in the health sector (n=5), academia (n=3), regulatory bodies (n=2), drug development (n=7), veterinary medicine (n=2), WSI device manufacturing (n=2) and software development (n=3) completed the survey. Sixteen participants worked in pathology laboratories. The laboratories varied in the number of slides processed per year: 100-1000 (n=2, 13%), 1000-5000 (n=5, 32%) and over 5000 (n=9, 56%). All laboratories exchanged WSI data with other centres, while most also exchanged glass slides (n=12, 75%) and tissue samples (n=11, 69%).

#### 3.1.1 IMAGING DATA ACQUISITION, STORAGE AND PROCESSING

Most respondents used multiple imaging modalities, with WSI being the most common (n=20), followed by immunohistochemistry (n=14) and then immunofluorescence multiplex assays (n=11). WSI data was predominantly stored in a vendor-specific format (n=16), while purpose-built open source formats DICOM for WSI (n=7) and OME-TIFF (n=7) as well as generic image formats (n=10) were also in use. Two respondents used in-house developed formats.


Regarding ease of access to WSI data, most respondents reported it was easy to access and locate the images (n=23), while most (n=22) also noted that they needed to combine WSI data with other data sources such as electronic patient records (n=12), study protocols (n=12), laboratory results (n=10), radiology (n=10), omics (n=9) and other data (n=5) for diagnosis. Access to the non-WSI data sources has posed issues for some respondents (n=16), with key problems being segregated data storage (n=5), lack of interoperability between different systems hindering transfer of metadata and annotations (n=11), absence of necessary metadata (n=7), image registration (n=2) and large file sizes (n=3).

#### 3.1.2 QUALITY ASSURANCE AND REGULATIONS

75% of respondents (n=18) came from workplaces where good laboratory practice applied. Of these, 76% (n=13) also applied good clinical practice, while some laboratories also implemented current good manufacturing practice, relevant ISO standards and software-specific in-house practices. Quality assurance protocols were regularly applied to tissue management equipment, image analysis tools and imaging equipment, whereby calibration tools for assessment of uniformity of illumination within the field, magnification, and colour calibration were of equal interest.

#### 3.1.3 AI AND ML APPLICATIONS


From 79% of responders who were able to use AI/ML tools within their workflow (n=19), only 16% (n=3) showed no interest in AI applications (Figure 3). Cell type annotation and disease classification were most frequent use cases (n=8 and n=7, respectively), and AI-assisted image quality assurance was the least frequent case (n=2). At the same time, cell type annotation was the task where users showed least confidence (only 39% stated they are confident in the results), while automated quality assurance was the use case where most users (61%) declared high level of confidence.



**Figure 3:** Use of AI/ML applications in the digital pathology pipeline. There were zero “No interest” responses for “Registration” and “Image lookup and navigation” categories, and one response for each of the other categories.

Performance and reliability of AI tools were accessed by comparison to human operator(s) performance (n=6), or other software tools (n=1), while some respondents used multi-centre studies (n=2).

Use of novel imaging techniques and future work directions Novel imaging modalities including targeted and untargeted metabolic and protonic imaging, as well as genomic imaging, were used by relatively few laboratories (5%-16%, mean: 11%). Thereby high numbers of participants were interested in using all or some of these techniques in the future (48%-68%, mean: 58%). Consequently, nearly all participants were interested in software tools to combine data from multiple modalities (Figure 4). This finding is consistent with the lack of data integration stated in section 3.1.1.



**Figure 4:** Prioritised challenges in digital pathology (n=22).

### 3.2 WORKSHOP FINDINGS

Table 2 **Error! Reference source not found.** presents the selected findings of two discussion sessions on (1) sample preparation and processing, (2) equipment calibration, (3) image processing and analysis, (4) omics and novel imaging techniques and (5) data integration.

**Table 2:** Prioritised challenges and metrology support required in Digital Pathology grouped by domains. Priorities denote how urgently the issue should be addressed, with “high” meaning “as soon as possible” and “medium” meaning “within one to three years from now.”

Challenge	Priority	Metrology support required	Domain
Quality framework is not well developed or taken up by the community. Where quality frameworks are deployed, their implementation varies and is poorly documented	High	Develop standardised operation, quality assurance and quality control procedures for manual and computerised processes. Define framework for traceable capture of information.	 sample preparation
Lack of comparability of in-house sample processing methods with those used by Contract Research Organisation laboratories	Medium	Support the development of reference materials for comparison between staining protocols and sites to monitor and track performance. Prioritise standards and metrics to track stain intensity and to quantify variability.	
Differences in methods and frequency of instrument calibration across laboratories lead to significant variations in results and impact the downstream analysis	High	Develop good practice guides and robust calibration protocols to reduce interlaboratory variability, subjectivity and bias.	 equipment calibration
Lack of certified methods, instruments and calibration artifacts to assess WSI device performance	Medium	AI/MLs should develop vendor-agnostic SI-traceable physical calibration artifacts to quantify variations in colour, focus, luminance, depth of field and image quality.	
Issues in transfer of pathologists' annotations and WSI metadata limit the availability of datasets for training of AI/ML systems and junior pathologists	High	Introduce data standards for images, annotations and linking of both. Use vendor-neutral format to that can be transferred between systems.	 image processing and analysis
Lack of transparency and explainability in how the system arrived at the result. No mechanisms for calculating and communicating the uncertainty of the results	Medium	Provide mechanisms for standardised and trustworthy outputs from data analysis and training of pathologists in how to interpret the results. Develop DP-specific metrics for evaluation of AI-based system performance.	
Lack of knowledge and training in use of novel imaging techniques	Medium	Support the validation and standardisation of novel imaging techniques to ensure confidence in their capabilities and to understand their uses, advantages and limitations in DP	 omics and novel imaging techniques
Slow uptake of experimental techniques in clinical practice	Medium	Design a training program on good practices and the use of novel imaging techniques to aid their adoption in DP and to disseminate the interdisciplinary knowledge.	
Heterogeneity of image formats and metadata contents hinders the data exchange between devices, software systems and institutions	High	Encourage use of a vendor-neutral format for within DP pipeline. Agree on a set of minimum metadata to promote traceability, account for variations and enable meaningful analysis.	 data

Segregation of data between WSI and laboratory/clinical systems hinders data collection, costs time and increases potential for errors	High	Define tests for DP interoperability, including benchmarking tests to ensure that solutions work at scale, and guides on practical implementations to make the standards work for end users	integration
--	------	---	-------------

#### 4 DISCUSSION

This study describes the views of stakeholders on the metrology support required to address pressing challenges in digital pathology. The desired interventions and rationales behind them follow the order of the DP pipeline stages.

##### *Sample handling*

It has been recognised that sample handling and preparation present the largest source of variability. It is therefore crucially important to develop vendor-agnostic metrology methods to improve intra- and inter-laboratory reproducibility, e.g., calibration procedures in combination with standardised digital QA and QC. Working with end users, metrologists can develop fit-for-purpose reference materials, quality metrics and standardisation routines for H&E, immunohistochemical, in situ hybridization and other tissue staining protocols. The introduction of such metrics and computerised QA and QCs would benefit the community by improving system performance, maintaining reproducibility within and between laboratories and by accelerating the implementation of new techniques.

##### *WSI calibration*

The need for sharing and comparative analysis of digital images across sites and the advance of AI/ML systems place new demands on the consistency of image data sets. Such data sets can only be obtained if WSI instruments are subjected to regular and reliable calibration to improve image comparability and reproducibility. National Measurement Institutes (NMIs) should work together with vendors, AI developers, and users to develop calibration procedures and physical artifacts such as calibration slides to assess the imaging device performance including dimensionality, illumination, and colour. Some efforts to provide reliable calibration objects to assess colour reproduction have been undertaken, but need to be developed further to meet the area's needs. In the longer term, creating physical calibration objects that are a) validated through round-robin multi-NMI trials and b) traceable to primary standards and the SI will underpin quantitative imaging in pathology.

##### *Novel imaging*

Beyond WSI histopathology, there is a growing interest in novel imaging modalities such as super-resolution and light sheet microscopy, Raman spectroscopy, mass spectrometry imaging, and many others. Understanding the uses, advantages and limitations of these techniques in combination with histopathology requires validation and standardisation efforts from end users and metrologists. Good practice recommendations can then be developed for specific tasks such as diagnosis of a particular disease or stage, quantitative pathology, and integration of conventional H&E imaging with other experimental techniques.

##### *Image analysis*

AI/ML tools are increasingly used for image quality assurance, image analysis, annotation, review prioritisation, and disease classification. It is essential that such systems are based on explainable and trustworthy mechanisms that can account for uncertainties and provide visualisation tools to aid the interpretation of their results. The development of trustworthy AI systems requires large volumes of well annotated WSI data with consistent metadata that includes WSI device calibration and sample handling information. Lastly, AI systems should include clear guidelines of their potential usage scenarios or be "intelligent" enough to reject input data when it is not suitable.

*Data integration*

The intelligent use of DP data by clinicians and AI systems requires that the images are supplied with appropriate metadata and can be reliably linked to other data sources such as radiology or laboratory findings. Unlike radiology, where the Digital Imaging and Communications in Medicine standard (DICOM) enables data exchange with the hospital information system, DP systems tend to use proprietary storage and have limited interoperability. Using a vendor-agnostic image format such as DICOM for WSI will solve this issue and provide a common “language” to exchange data with other clinical systems. However, DICOM alone does not guarantee data quality. It is perfectly possible to have a well-shaped DICOM WSI file that has insufficient information about the image context and is clinically meaningless. Therefore, the community needs robust metadata standards to capture sample handling, imaging device setting, and image pre-processing steps. These metadata should be captured using consistent clinical terminologies, ontologies and units of measurement. DICOM WSI offers suitable mechanisms to encode metadata in widely used clinical nomenclatures such as SNOMED CT, LOINC, and UCUM. NMIs can develop minimum metadata guidelines in collaboration with DICOM working group 26. Agreeing a vendor-neutral format for annotations and defining the minimum metadata to be stored with the image will make the DP imaging pathway ready for AI/ML deployment, increasing the throughput, reliability and adoption of DP.

*Summary*

Across all challenge areas, there was a call for metrological guidance and shaping an appropriate training program to upskill the existing staff and to educate the new generation of pathologists on the role and value of metrology in creating reliable and trustworthy digital pathology solutions.

**5 ACKNOWLEDGEMENTS**

This work was funded by the UK Department of Business, Energy and Industrial Strategy through the National Measurement System programme. We would like to thank all workshop attendees and the industry co-chair of Integrating the Healthcare Enterprise Robin Breslin for their insights and comments.

**6 WEB RESOURCES**

Accessed between November 2021 – March 2022.

- Metrology for Digital Pathology: Digital pathology cross-theme project report <http://eprintspublications.npl.co.uk/9279/1/AS102.pdf>
- European Metrology Network for traceability in laboratory medicine: <https://www.euramet.org/european-metrology-networks/laboratory-medicine/>
- “Testing times to come? An evaluation of pathology capacity across the UK”, Cancer Research UK, 2016 [https://www.cancerresearchuk.org/sites/default/files/testing\\_times\\_to\\_come\\_nov\\_16\\_cruk.pdf](https://www.cancerresearchuk.org/sites/default/files/testing_times_to_come_nov_16_cruk.pdf)

## 7 REFERENCES

1. Köhler R. The International Vocabulary of Metrology, 3rd Edition: Basic and General Concepts and Associated Terms. Why? How? In: *Transverse Disciplines in Metrology* [Internet]. John Wiley & Sons, Ltd; 2009 [cited 2022 Mar 14]. p. 233–8. Available from: <https://onlinelibrary.wiley.com/doi/abs/10.1002/9780470611371.ch21>
2. Browning L, Fryer E, Roskell D, White K, Colling R, Rittscher J, et al. Role of digital pathology in diagnostic histopathology in the response to COVID-19: results from a survey of experience in a UK tertiary referral hospital. *J Clin Pathol*. 2021 Feb 1;74(2):129–32.
3. Retamero JA, Aneiros-Fernandez J, Del Moral RG. Complete Digital Pathology for Routine Histopathology Diagnosis in a Multicenter Hospital Network. *Arch Pathol Lab Med*. 2020 Feb;144(2):221–8.
4. Bera K, Schalper KA, Rimm DL, Velcheti V, Madabhushi A. Artificial intelligence in digital pathology - new tools for diagnosis and precision oncology. *Nat Rev Clin Oncol*. 2019 Nov;16(11):703–15.
5. Barisoni L, Lafata KJ, Hewitt SM, Madabhushi A, Basu UGJ. Digital pathology and computational image analysis in nephropathology. *Nat Rev Nephrol*. 2020 Nov;16(11):669–85.
6. Hipp AJ. The Revival of the H&E with Artificial Intelligence – *Journal of Clinical & Anatomic Pathology (JCAP)* [Internet]. 2020 [cited 2021 Apr 7]. Available from: <http://www.clinpathology.com/the-revival-of-the-he-with-artificial-intelligence/>
7. Snead DRJ, Tsang Y-W, Meskini A, Kimani PK, Crossman R, Rajpoot NM, et al. Validation of digital pathology imaging for primary histopathological diagnosis. *Histopathology*. 2016 Jun;68(7):1003–72.
8. Cashmore MT, McCann AJ, Wastling SJ, McGrath C, Thornton J, Hall MG. Clinical quantitative MRI and the need for metrology. *Br J Radiol*. 2021 Apr 1;94(1120):20201215.
9. Niazi MKK, Parwar A, Gurcan M. Digital Pathology and Artificial Intelligence. *Lancet Oncol*. 2019 May;20(5):e253–61.
10. Marée R. The need for careful data collection for pattern recognition in digital pathology. *J Pathol Inform*. 2017;8(1):19.
11. Aeffner F, Zarella MD, Buchbinder N, Bui MM, Goodman MR, Hartman DJ, et al. Introduction to Digital Image Analysis in Whole-slide Imaging: A White Paper from the Digital Pathology Association. *J Pathol Inform* [Internet]. 2019 Mar 8 [cited 2021 Apr 12];10. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC6437786/>
12. Cross S, Furness P, Igali L, Snead D, Treanor D. Best practice recommendations for implementing digital pathology January 2018. *R Coll Pathol*. 2018;
13. Clarke EL, Treanor D. Colour in digital pathology: a review. *Histopathology*. 2017 Jan;70(2):153–63.
14. Palmirotta R, Lovero D, Cafforio P, Felici C, Mannavola F, Pellè E, et al. Liquid biopsy of cancer: a multimodal diagnostic tool in clinical oncology. *Ther Adv Med Oncol*

- [Internet]. 2018 Aug 29 [cited 2021 Mar 25];10. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC6116068/>
15. Zhou Y, Liu C-H, Wu B, Yu X, Cheng G, Zhu K, et al. Optical biopsy identification and grading of gliomas using label-free visible resonance Raman spectroscopy. *J Biomed Opt.* 2019 Sep;24(9):1–12.
  16. Morais CLM, Lilo T, Ashton KM, Davis C, Dawson TP, Gurusinghe N, et al. Determination of meningioma brain tumour grades using Raman microspectroscopy imaging. *Analyst.* 2019;144(23):7024–31.
  17. Abdelmoula WM, Balluff B, Englert S, Dijkstra J, Reinders MJT, Walch A, et al. Data-driven identification of prognostic tumor subpopulations using spatially mapped t-SNE of mass spectrometry imaging data. *Proc Natl Acad Sci.* 2016 Oct 25;113(43):12244.
  18. Goodwin RJA. Sample preparation for mass spectrometry imaging: Small mistakes can lead to big consequences. *J Proteomics.* 2012 Aug;75(16):893–911.
  19. Everall NJ. Confocal Raman microscopy: common errors and artefacts. *The Analyst.* 2010;135(10):2512.
  20. Shakya R, Nguyen TH, Waterhouse N, Khanna R. Immune contexture analysis in immuno-oncology: applications and challenges of multiplex fluorescent immunohistochemistry. *Clin Transl Immunol*; [Internet]. 2020 Jan [cited 2021 Mar 31];9(10). Available from: <https://onlinelibrary.wiley.com/doi/10.1002/cti2.1183>
  21. Finlayson D, Rinaldi C, Baker MJ. Is Infrared Spectroscopy Ready for the Clinic? *Anal Chem.* 2019 Oct 1;91(19):12117–26.
  22. Wang M, Kothari S, Phan J. Eliminating tissue-fold artifacts in histopathological whole-slide images for improved image based prediction of cancer grade. *J Pathol Inform.* 2013;4(1):22.
  23. Senaras C, Niazi MKK, Lizarzoki G, Gurcan MN. DeepFocus: Detection of out-of-focus regions in whole slide digital images using deep learning. *PLOS ONE.* 2018 Oct 25;13(10):e0205387.
  24. Lafarge MW, Plim IPV, Eppenhof KAJ, Moeskops P, Veta M. Domain-Adversarial Neural Networks to Address the Appearance Variability of Histopathology Images. In: Cardoso MJ, Arbel T, Carneiro G, Syeda-Mahmood T, Tavares JMRS, Moradi M, et al., editors. *Deep Learning in Medical Image Analysis and Multimodal Learning for Clinical Decision Support.* Cham: Springer International Publishing; 2017. p. 83–91.
  25. Serag A, Ion-Margineanu A, Qureshi H, McMillan R, Saint Martin M-J, Diamond J, et al. Translational AI and Deep Learning in Diagnostic Pathology. *Front Med.* 2019 Oct 1;6:185.
  26. Komura D, Ishikawa S. Machine Learning Methods for Histopathological Image Analysis. *Comput Struct Biotechnol J.* 2018;16:34–42.
  27. Clunie DA. DICOM Format and Protocol Standardization—A Core Requirement for Digital Pathology Success. *Toxicol Pathol.* 2021 Jun 1;49(4):738–49.

28. Dash R, Jones N, Merrick R, Haroske G, Harrison J, Sayers C, et al. Integrating the health-care enterprise pathology and laboratory medicine guideline for digital pathology interoperability. *J Pathol Inform.* 2021;12(1):16.
29. Stuppel A, Singerman D, Celi LA. The reproducibility crisis in the age of digital medicine. *Npj Digit Med.* 2019 Jan 29;2(1):1–3.
30. Sarkans U, Chiu W, Collinson L, Darrow MC, Ellenberg J, Grunwald D, et al. REMBI: Recommended Metadata for Biological Images—enabling reuse of microscopy data in biology. *Nat Methods.* 2021 Dec;18(12):1418–22.
31. Czichos H, Saito T, Smith LE. Springer handbook of metrology and testing. Springer Science & Business Media; 2011.
32. Sullivan DC, Obuchowski NA, Kessler LG, Raunig DL, Gatsonis C, Huang EP, et al. Metrology Standards for Quantitative Imaging Biomarkers. *Radiology.* 2015 Dec;277(3):813–25.
33. Raunig DL, McShane LM, Pennello G, Gatsonis C, Carson PL, Voyvodic JT, et al. Quantitative imaging biomarkers: a review of statistical methods for technical performance assessment. *Stat Methods Med Res.* 2015 Feb;24(1):27–67.
34. European Society of Radiology (ESR). ESR Statement on the Validation of Imaging Biomarkers. *Insights Imaging.* 2020 Jun 4;11(1):76.
35. Adeogun M, Bunch J, Dexter A, Dondi C, Murta T, Nikula C, et al. Metrology for Digital Pathology. Digital pathology cross-theme project report [Internet]. 2021 [cited 2022 Mar 3]. Available from: <https://doi.org/10.11120/npl.AS102>

**Declaration of interests**

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

The authors declare the following financial interests/personal relationships which may be considered as potential competing interests:

Journal Pre-proof

The role for measurement science in digital pathology

Highlights

- Heterogeneity of image and metadata formats limits data (re-)use
- Lack of standardised data hinders development of AI/ML tools for pathology
- Open standard for images and annotations will help data sharing and re-use
- Frequency and scope of instrument calibration vary a lot between laboratories
- Standardised calibration tools are needed to yield consistent comparable images
- Pathology community needs own metrics to assess AI/ML performance

Journal Pre-proof