



Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.



Viral load quantitation of SARS-coronavirus RNA using a one-step real-time RT-PCR

Els Keyaerts^a, Leen Vijgen^a, Piet Maes^a, Griet Duson^a,
Johan Neyts^b, Marc Van Ranst^{a,c,*}

^aLaboratory of Clinical and Epidemiological Virology, Department of Microbiology and Immunology, Rega Institute for Medical Research, University of Leuven, Minderbroedersstraat 10, 3000 Leuven, Belgium

^bLaboratory of Virology and Chemotherapy, Department of Microbiology and Immunology, Rega Institute for Medical Research, University of Leuven, Belgium

^cLaboratory of Clinical Virology, University Hospital Gasthuisberg, University of Leuven, Belgium

Received 13 August 2004; received in revised form 20 January 2005; accepted 1 February 2005

Corresponding Editor: Jane Zuckerman, London, UK

KEYWORDS

TaqMan PCR;
Quantitative PCR;
Real-time PCR;
SARS-CoV;
cRNA standard;
RT-PCR

Summary

Introduction: Severe acute respiratory syndrome (SARS) is an emerging infectious disease that first occurred in humans in the People's Republic of China in November 2002 and has subsequently spread worldwide. A novel virus belonging to the *Coronaviridae* family has been identified as the cause of this pulmonary disease. The severity of the disease combined with its rapid spread requires the development of fast and sensitive diagnostic assays.

Results: A real-time quantitative RT-PCR was designed in the nsp11 region of the replicase 1B domain of the SARS-coronavirus (SARS-CoV) genome. To evaluate this quantitative RT-PCR, cRNA standards were constructed by in vitro transcription of SARS-CoV Frankfurt 1 RNA using T7 RNA polymerase, followed by real-time RT-PCR. The assay allowed quantitation over a range of 10^2 to 10^8 RNA copies per reaction. *Conclusions:* Extrapolated to clinical samples, this novel assay has a detection range of 10^4 to 10^{10} copies of viral genome equivalents per millilitre. In comparison to the current de facto cRNA Artus Biotech standard, the in-house cRNA standard gives a 100-fold higher absolute quantity, suggesting a possible underestimation of the viral load when using the Artus Biotech standard.

© 2005 International Society for Infectious Diseases. Published by Elsevier Ltd. All rights reserved.

* Corresponding author. Tel.: +32 16 347908; fax: +32 16 347900.
E-mail address: marc.vanranst@uz.kuleuven.ac.be (M. Van Ranst).

Introduction

Severe acute respiratory syndrome (SARS) has recently emerged as a new severe human disease that resulted in 774 deaths and more than 8000 reported probable cases worldwide.¹ The SARS epidemic began in the Guangdong province in southern China, where several cases of atypical pneumonia of unknown etiology were reported in November 2002.

A novel member of the *Coronaviridae* family has been identified as the causative agent of this pulmonary disease.²⁻⁶ The SARS-coronavirus (SARS-CoV) is a new coronavirus type, of which the morphology and genome organization are very similar to those of other coronaviruses. Phylogenetic analyses and sequence comparisons showed that SARS-CoV is not closely related to any of the previously characterized coronaviruses.^{6,7}

The clinical case definition of SARS is essentially one of fever and pneumonia, with or without a contact history. In the absence of a definite history of contact with other SARS patients, it is not easy to differentiate SARS from other causes of atypical pneumonia. Laboratory SARS-CoV confirmation tests are therefore a critical clinical need. Detection of specific SARS-CoV antibodies has been a reliable way to confirm the diagnosis. Due to the fact that the antibody response appears only around day ten of the illness, IgM or other subtype assays have not been useful for closing the diagnostic window within the first week of illness.

Virus culture is relatively insensitive, and takes too much time to be clinically relevant. During the epidemic, a variety of reverse transcription-polymerase chain reaction (RT-PCR) assays were developed^{3,8} including a commercial ready-to-use RT-PCR kit (Artus Biotech, Hamburg, Germany). Detection of viral RNA by RT-PCR in clinical samples offers the opportunity of diagnosis in an earlier stage of the disease.⁹

In contrast to many other acute respiratory infections, viral loads in clinical samples are low in the first few days of illness and peak around day ten of the disease. Recently Ng and colleagues demonstrated the role of serum SARS-CoV viral load measurement as a prognostic marker, even on the first day of hospital admission. A high systemic viral load may directly lead to severe tissue damage, or indirectly through the activation of a potentially damaging immune reaction.¹⁰ For SARS-CoV diagnosis, plasma or serum RT-PCR can be performed with high sensitivity during the first week of the disease. Stool and respiratory samples can be used during the second week, and serologic testing for antibodies against SARS-CoV can be used from day 21 onward.¹¹ A major advantage of real-time PCR is

that amplification and analysis are completed in a closed system. The risk of contamination, which can confound conventional (nested) RT-PCR protocols, is markedly reduced. This paper presents a very sensitive and specific quantitative RT-PCR assay, which can contribute to a rapid and efficient diagnosis of SARS in case of a new outbreak.

Materials and methods

Viral propagation

The SARS-CoV Frankfurt 1 (FFM-1) strain was kindly provided by Professor Dr H.F. Rabenau and Professor Dr H.W. Doerr from the Johann Wolfgang Goethe University, Frankfurt, Germany. SARS-CoV was propagated in African green monkey kidney cells (Vero E6 cells), obtained from the American Type Culture Collection (ATCC C1008). Supernatant was harvested after four days incubation at 37 °C in a 5% CO₂ atmosphere. Human coronaviruses OC43 and 229E were also obtained from the American Type Culture Collection (VR-759 and VR-740). Viral RNA was isolated using the QIAamp viral RNA kit (Qiagen, Westburg, Leusden, The Netherlands).

Primer and probe design

The nsp11 region of the replicase 1B domain of the SARS-CoV genome was screened for primer and probe target sites that would be compatible with TaqMan PCR requirements (ABI 7700 Users Manual), using the Primer express 2.0 software. Primers spanning a target region of 68 bp were selected, with matched dissociation temperatures and a minimal likelihood for duplex or hairpin formation: forward primer SARS-FP (5'-CACCCGCGAAG-AAGCTATTC-3'), MGB probe SARS-TP (FAM 5'-TGCGTGGATTGGCTT-3'/NFQ-MGB), and reverse primer SARS-RP (5'-TTGC-ATGACAGCCCTCTACATC-3').

Construction of cRNA standards

The TaqMan SARS-CoV forward primer was modified with a T7-promoter sequence at the 5'-end (SARS-FPT7 5'-TAATACGACTCACTATAGGGAGGCACCCGCGA-AGAAGCTA-TTC-3'). PCR products amplified with the modified primer pairs were quantified spectrophotometrically at 260 nm. Two hundred nanograms of PCR product were used for in vitro transcription (MEGashortscript T7 kit, Ambion, Austin, TX, USA) performed at 37 °C overnight in a 20 µl reaction mix containing 2 µl of reaction buffer, 2 µl of each dNTP and 2 µl enzyme mix. The cDNA was removed by digestion with two units of RNase-free DNase I for

15 min at 37 °C. The cRNA was precipitated by adding 3 μ l 3 M NaOAc and 60 μ l of 96% ethanol and a subsequent incubation at -20 °C for 30 min. After centrifugation at 13000 rpm for 15 min, the supernatant was removed and 500 μ l 70% ethanol was added. After another 5 min of centrifugation at 13000 rpm, the supernatant was removed and the pellet was dissolved in 200 μ l RNase free H₂O (Sigma-Aldrich NV, Bornem, Belgium) and stored at -80 °C. Quantitation of cRNA was performed spectrophotometrically at 260 nm.¹²

Real-time quantitative RT-PCR

A 25 μ l RT-PCR was carried out using 5 μ l of extracted RNA or standard cRNA, 12.5 μ l of a real-time One-Step RT-PCR Master Mix containing ROX as a passive reference (Applied Biosystems, Foster City, CA, USA), 900 nM forward and reverse primer, and 150 nM MGB probe. Amplification and detection were performed in a ABI PRISM 7700 Sequence Detection System (Applied Biosystems, Foster City, CA, USA) under the following conditions: an initial reverse transcription at 48 °C for 30 min, followed by PCR activation at 95 °C for 10 min and 45 cycles of amplification (15 seconds at 95 °C and 1 min at 60 °C). During amplification, the ABI PRISM sequence detector monitored real-time PCR amplification by quantitative analysis of fluorescence emissions. The reporter dye (FAM) signal was measured against the internal reference dye (ROX) signal to normalize for non-PCR-related fluorescence fluctuations occurring from well to well. The threshold cycle represented the refraction cycle number at which a positive amplification was measured, and was set at ten times the standard deviation of the mean baseline emission calculated for PCR cycles 3 to 15. Analysis of the copy numbers and linear regression curve was performed using the Sequence Detector v1.9 software (Applied Biosystems, Foster City, CA, USA).

One-step RT-PCR

To compare the sensitivity of the real-time RT-PCR with a conventional RT-PCR, a cRNA sample was used (kindly provided by Professor Dr C. Drosten from the Bernard Nocht Institute for Tropical Medicine, Hamburg, Germany). A 195 bp fragment was amplified by a one-step RT-PCR using the BNI outer primers (5'-ATGAATTACCAAGTCAATGGTTAC-3' as the sense primer and 5'-CATAACCAGTCGGTACAGCTAC-3' as the antisense primer). The RT-PCR was carried out on a Geneamp PCR System 9700 thermal cycler (Applied Biosystems, Foster City, CA, USA). Each reaction

contained 0.6 μ M of sense and antisense primers, 0.4 mM of each dNTP and 2 μ l Qiagen OneStep RT-PCR enzyme mix. RT-PCR conditions were as follows: an initial reverse transcription for 30 min at 50 °C followed by a PCR activation for 15 min at 95 °C, 45 cycles of amplification (30 sec at 94 °C, 30 seconds at 50 °C and 1 min at 72 °C) and a final extension step at 72 °C for 10 min.

Results

Absolute viral RNA load quantitation

To determine the amount of genome equivalents/ml, i.e. the viral RNA load, in cell culture samples, SARS-CoV cRNA standards were used for the construction of a standard curve (Figure 1). Ten-fold serial dilutions of cRNA transcripts were used corresponding to copy numbers of 10² to 10⁸. Equal volumes of standard and sample were used for PCR amplification. SARS-CoV RNA, extracted from the supernatant of infected Vero E6 cells, which contained a virus titer of 10³ CCID₅₀/ml was analysed and was found to correspond to a viral RNA load of 1.7 \times 10¹² RNA copies/ml.

Dynamic range

To investigate the dynamic range of the assay, 10-fold serial dilutions of the cRNA standard, ranging from 10 to 10¹² molecules in the reaction, were tested. The results were analysed in terms of C_T value (the cycle in which a target sequence is detected). As shown in Figure 2, the dynamic range of the assay spanned 6 logs ranging from 10² to 10⁸ molecules SARS-CoV per reaction, corresponding to a range in C_T values from 37.45 \pm 0.04 to 17.40 \pm 0.04.




Figure 1 Ten-fold serial dilutions ranging from 10⁸ to 10² copies of cRNA were tested in duplicate in real-time RT-PCR. A standard curve graph was generated by plotting the C_T values on the y-axis and the log of the input amounts on the x-axis. The slope of this standard curve was -3.43 and the correlation coefficient was 0.997.




Figure 2 The dynamic range of the assay was tested running 10-fold serial dilutions of the authors' cRNA standard ranging from 10^2 to 10^{12} molecules in the reaction. The amplification plot shows delta Rn on the y-axis (Rn: fluorescence emission intensity of the reporter dye normalized to a passive reference, delta Rn: Rn of an unreacted sample – Rn value of the reaction) against the cycle number displayed on the x-axis. The dynamic range of the assay spans 6 logs ranging from 10^2 to 10^8 log molecules of SARS-CoV RNA per reaction, corresponding to C_T values ranging from 37.45 ± 0.04 for 10^2 copies to 17.40 ± 0.04 for 10^8 copies. cRNA standards with a copy number above or below the dynamic range are not shown.

Specificity, sensitivity and reproducibility

Experiments were undertaken to assess diagnostic criteria such as specificity, sensitivity and reproducibility. Specificity of the primers and probe was tested using other human coronaviruses OC43 and 229E. No detectable non-specific amplification of these coronaviruses took place. The sensitivity of the real-time RT-PCR was investigated by assaying ten replicates of samples containing 10 , 10^2 , 10^3 or 10^4 copies of the cRNA standard per reaction. One hundred copies of the standard cRNA were detected in all ten replicates. For ten genome equivalents per reaction, only one out of ten replicates was found to be positive. The standard deviation of the mean threshold cycle varied between 0.34 and 0.98 cycles, resulting in coefficients of variation ranging from 0.97 to 2.57%. The intra-assay reproducibility of the SARS-CoV real-time RT-PCR was assessed by running dilutions of the cRNA standards ranging from 10^7 to 10^2 copies per reaction. Each of these dilutions was assayed four times per run. The mean values of C_T , the standard deviations and the coefficients of variation were determined. Table 1 shows that the SARS-CoV cRNA standards can be detected with a coefficient of variation of less than 0.7%. The interassay reproducibility of the SARS-CoV real-time RT-PCR was assessed by running dilutions of the cRNA standards ranging from 10^8 to 10^5 copies per reaction in several assay runs. The mean values of the C_T , the standard deviations and the coefficients of variation were

Table 1 Intra-assay reproducibility of the SARS-CoV quantitative RT-PCR.

Sample	Dilution	cRNA copies	Mean C_T	S.D.	CV (%)
cRNA 1	1×10^{-6}	1.86×10^7	18.50	0.07	0.36
	1×10^{-7}	1.86×10^6	21.66	0.08	0.39
	1×10^{-8}	1.86×10^5	25.33	0.12	0.48
	1×10^{-9}	1.86×10^4	28.28	0.12	0.41
	1×10^{-10}	1.86×10^3	32.61	0.22	0.66
	1×10^{-11}	1.86×10^2	35.31	0.17	0.48

The intra-assay reproducibility was assessed by running dilutions of the cRNA standards ranging from 10^2 to 10^7 copies per reaction. Each of these dilutions was assayed four times per run. The mean values of the C_T , the standard deviations (S.D.) and the coefficients of variation (CV) were determined.

determined (Table 2). The results show that the SARS-CoV cRNA standard can be detected with a coefficient of variation of less than 4%.

Comparison of the in-house cRNA standard with the cRNA standard of Artus Biotech

To validate the in-house cRNA standard, serial ten-fold dilutions of the in-house standard together with equivalent dilutions of the commercially available Artus Biotech standard were assessed in the same run. Analysing a specific C_T value with the two standard curves resulted in a 100-fold difference in copy number.

Table 2 Interassay reproducibility of the SARS-CoV quantitative RT-PCR.

Sample	Dilution	cRNA copies	Mean C_T	S.D.	CV (%)
cRNA 1	1×10^{-5}	1.86×10^8	14.80	0.53	3.61
	1×10^{-6}	1.86×10^7	17.62	0.23	1.32
	1×10^{-7}	1.86×10^6	21.16	0.18	0.87
	1×10^{-8}	1.86×10^5	24.22	0.06	0.23
cRNA 2	1×10^{-5}	1.94×10^8	15.39	0.36	2.34
	1×10^{-6}	1.94×10^7	18.89	0.40	2.10
	1×10^{-7}	1.94×10^6	21.77	0.15	0.71
	1×10^{-8}	1.94×10^5	25.00	0.09	0.38
cRNA 3	1×10^{-5}	2.08×10^8	15.43	0.41	2.69
	1×10^{-6}	2.08×10^7	18.45	0.27	1.49
	1×10^{-7}	2.08×10^6	21.57	0.15	0.68
	1×10^{-8}	2.08×10^5	25.23	0.10	0.38
cRNA 4	1×10^{-5}	1.85×10^8	16.22	0.19	1.15
	1×10^{-6}	1.85×10^7	19.49	0.09	0.46
	1×10^{-7}	1.85×10^6	23.18	0.08	0.36
	1×10^{-8}	1.85×10^5	26.95	0.10	0.39

The interassay reproducibility was assessed by running dilutions of the cRNA standards ranging from 10^5 to 10^8 copies per reaction in three assay runs. The mean values of the C_T , the standard deviations (S.D.) and the coefficients of variation (CV) were determined.

Comparison of real-time RT-PCR to RT-PCR

To demonstrate the feasibility of broad usage of real-time RT-PCR for both research and clinical studies, the sensitivity of the real-time RT-PCR was compared to that of the RT-PCR using the BNI-outer primer pair. For both assays, ten-fold dilutions of a sample containing 10^9 copies per μl of the BNI cRNA fragment were used. Whereas the conventional RT-PCR could detect up to 10^4 copies of cRNA per reaction, the detection rate of this real-time RT-PCR assay was 100-fold higher, with a sensitivity up to 10^2 cRNA copies per reaction.

Discussion

Although the initial global outbreak of SARS has been successfully contained, SARS will remain a serious concern. The severity of the disease combined with its rapid spread requires a diagnostic test that can detect the SARS-CoV infection early in the course of illness. The SARS-CoV can be detected by conventional and real-time RT-PCR in specimens obtained from the upper and lower respiratory tracts, in plasma or serum, and in feces.^{9,10,13–17}

A conventional RT-PCR, although performed as a one-step reaction in a thermocycler, still requires the use of PAGE analysis for detection and it takes four to five hours to complete the test. This real-time RT-PCR, on the other hand, can generate complete results within a period of three hours. The speed of this assay combined with a minimum of manipulations that need to be performed, make this real-time RT-PCR an ideal routine protocol for high-throughput screening of possible SARS-CoV-infected samples.

The assay described here has been shown to be highly sensitive, with a detection limit of 10^2 copies of SARS-CoV per reaction, which is 2 logs more sensitive than a conventional SARS-CoV RT-PCR. Furthermore, this assay has a sensitivity that is similar or even better than that reported for other existing real-time SARS-CoV assays. This sensitivity is achieved without the need for a nested step in the RT-PCR reaction.¹⁸

During the first week of illness, the viral load in the upper respiratory tract specimens is very low, but will reach a peak around day ten of illness. Since early diagnosis is important, highly sensitive diagnostic methods are necessary to avoid false-negative results in samples that contain only trace amounts of the virus in the early phase of the infection. This study presents a diagnostic real-time quantitative assay that is sensitive enough to detect the presence of SARS-CoV even at an early stage of the illness. The potential for quantitation over a

wide dynamic range (at least 6 logs) was demonstrated with low intra- and inter-assay variation. Furthermore, this method was highly specific for the SARS-CoV and gave no false-positive results with HCoV-OC43 and HCoV-229E.

Comparison of the in-house cRNA standard with the existing cRNA standard of Artus Biotech^{2,18} showed a 100-fold difference in absolute quantity. Using the in-house standard, this assay can detect 100 copies per reaction, whereas with the Artus standard, the detection limit of our assay is one RNA copy per reaction.

Drosten and co-workers were the first to develop a standard to quantify the SARS-CoV, and this standard is also used in the commercial ready-to-use RT-PCR kit (Artus Biotech). Even after repeating the cRNA construction several times, the same results were obtained. At the moment the correct absolute quantities cannot be predicted. The Artus Biotech standard is currently the *de facto* standard, and it is necessary to point out that there is a possible 100-fold underestimation of the copy numbers. This underestimation of the viral load can be important for in vitro and in vivo antiviral testing, where real-time RT-PCR is used for virus yield assays and time-of-drug-addition experiments.^{19,20}

Acknowledgements

The authors wish to thank their colleagues at the laboratory of Clinical and Epidemiological Virology, Department of Microbiology & Immunology, Rega Institute for Medical Research, University of Leuven, Belgium for their helpful comments and discussion. This work was supported by a fellowship from the Flemish Fonds voor Wetenschappelijk Onderzoek (FWO) to Leen Vijgen, and by FWO-grant G.0288.01.

Conflict of interest: No conflict of interest to declare.

References

1. Peiris JS, Guan Y, Yuen KY. Severe acute respiratory syndrome. *Nat Med* 2004;**10**:S88–97.
2. Drosten C, Gunther S, Preiser W, van der Werf S, Brodt HR, Becker S, et al. Identification of a novel coronavirus in patients with severe acute respiratory syndrome. *N Engl J Med* 2003;**348**:1967–76.
3. Ksiazek TG, Erdman D, Goldsmith CS, Zaki SR, Peret T, Emery S, et al. SARS Working Group A novel coronavirus associated with severe acute respiratory syndrome. *N Engl J Med* 2003;**348**:1953–66.
4. Kuiken T, Fouchier RA, Schutten M, Rimmelzwaan GF, van Amerongen G, van Riel D, et al. Newly discovered coronavirus

- as the primary cause of severe acute respiratory syndrome. *Lancet* 2003;**362**:263–70.
5. Peiris JS, Lai ST, Poon LL, Guan Y, Yam LY, Lim W, et al. SARS study group Coronavirus as a possible cause of severe acute respiratory syndrome. *Lancet* 2003;**361**:1319–25.
 6. Rota PA, Oberste MS, Monroe SS, Nix WA, Campagnoli R, Icenogle JP, et al. Characterization of a novel coronavirus associated with severe acute respiratory syndrome. *Science* 2003;**300**:1394–9.
 7. Marra MA, Jones SJ, Astell CR, Holt RA, Brooks-Wilson A, Butterfield YS, et al. The genome sequence of the SARS-associated coronavirus. *Science* 2003;**300**:1399–404.
 8. Poutanen SM, Low DE, Henry B, Finkelstein S, Rose D, Green K, et al. National Microbiology Laboratory, Canada, Canadian Severe Acute Respiratory Syndrome Study Team Identification of severe acute respiratory syndrome in Canada. *N Engl J Med* 2003;**348**:1995–2005.
 9. Poon LL, Chan KH, Wong OK, Yam WC, Yuen KY, Guan Y, et al. Early diagnosis of SARS-coronavirus infection by real-time RT-PCR. *J Clin Virol* 2003;**28**:233–8.
 10. Ng EK, Hui DS, Chan KC, Hung EC, Chiu RW, Lee N, et al. Quantitative analysis and prognostic implication of SARS-coronavirus RNA in the plasma and serum of patients with severe acute respiratory syndrome. *Clin Chem* 2003;**49**:1976–80.
 11. Peiris JS, Chu CM, Cheng VC, Chan KS, Hung IF, Poon LL, et al. HKU/UCH SARS Study Group. Clinical progression and viral load in a community outbreak of coronavirus-associated SARS pneumonia: a prospective study. *Lancet* 2003;**361**:1767–72.
 12. Fronhoffs S, Totzke G, Stier S, Wernert N, Rothe M, Bruning T, et al. A method for the rapid construction of cRNA standard curves in quantitative real-time reverse transcription polymerase chain reaction. *Mol Cell Probes* 2002;**16**:99–110.
 13. Lau LT, Fung YW, Wong FP, Lin SS, Wang CR, Li HL, et al. A real-time PCR for SARS-coronavirus incorporating target gene pre-amplification. *Biochem Biophys Res Commun* 2003;**312**:1290–6.
 14. Poon LL, Chan KH, Wong OK, Cheung TK, Ng I, Zheng B, et al. Detection of SARS-coronavirus in patients with severe acute respiratory syndrome by conventional and real-time quantitative reverse transcription-PCR assays. *Clin Chem* 2004;**50**:67–72.
 15. Grant PR, Garson JA, Tedder RS, Chan PK, Tam JS, Sung JJ. Detection of SARS-coronavirus in plasma by real-time RT-PCR. *N Engl J Med* 2003;**349**:2468–9.
 16. Ng LF, Wong M, Koh S, Ooi EE, Tang KF, Leong HN, et al. Detection of severe acute respiratory syndrome coronavirus in blood of infected patients. *J Clin Microbiol* 2004;**42**:347–50.
 17. Zhai J, Briesse T, Dai E, Wang X, Pang X, Du Z, et al. Real-time polymerase chain reaction for detecting SARS coronavirus, Beijing, 2003. *Emerg Infect Dis* 2004;**10**:300–3.
 18. Jiang SS, Chen TC, Yang JY, Hsiung CA, Su IJ, Liu YL, et al. Sensitive and quantitative detection of severe acute respiratory syndrome coronavirus infection by real-time nested polymerase chain reaction. *Clin Infect Dis* 2004;**38**:293–6.
 19. Keyaerts E, Vijgen L, Maes P, Neyts J, Van Ranst M. In vitro inhibition of severe acute respiratory syndrome coronavirus by chloroquine. *Biochem Biophys Res Commun* 2004;**323**:264–8.
 20. Yamamoto N, Yang R, Yoshinaka Y, Amari S, Nakano T, Cinatl J, et al. HIV protease inhibitor nelfinavir inhibits replication of SARS-associated coronavirus. *Biochem Biophys Res Commun* 2004;**318**:719–25.

Available online at www.sciencedirect.com

