

Overview of Robotic Technology in Spine Surgery

Ram K. Alluri, MD¹, Fedan Avrumova, BS¹,
Ahilan Sivaganesan, MD¹, Avani S. Vaishnav, MBBS¹,
Darren R. Lebl, MD, MBA^{1,2}, and Sheeraz A. Qureshi, MD, MBA^{1,2}

HSS Journal®: The Musculoskeletal
Journal of Hospital for Special Surgery
2021, Vol. 17(3) 308–316
© The Author(s) 2021



Article reuse guidelines:
sagepub.com/journals-permissions
DOI: 10.1177/15563316211026647
journals.sagepub.com/home/hss



Abstract

As robotics in spine surgery has progressed over the past 2 decades, studies have shown mixed results on its clinical outcomes and economic impact. In this review, we highlight the evolution of robotic technology over the past 30 years, discussing early limitations and failures. We provide an overview of the history and evolution of currently available spinal robotic platforms and compare and contrast the available features of each. We conclude by summarizing the literature on robotic instrumentation accuracy in pedicle screw placement and clinical outcomes such as complication rates and briefly discuss the future of robotic spine surgery.

Keywords

robot, robotic navigation, spine, body sites, robotic spine surgery, spinal robotics

History of Robotic Technology

Spine surgery has made significant advancements over the past 3 decades, largely driven by innovation in operative techniques, instrumentation, implant design, biomaterials, biologics, and intraoperative imaging. Many of these advances allowed for smaller incisions, higher fusion rates, and improved accuracy, all of which may contribute to decreased complications and improved patient-reported outcomes [13,16,26,46]. Perhaps the most significant progress occurred in the field of intraoperative imaging, with developments such as real-time navigation and 3-dimensional (3D) spinal mapping. In addition to facilitating some spine procedures, these innovations have directly driven the advancement of spinal robotic technology.

Robotic-assisted surgery has been used in urology, gynecology, and general surgery since the early 1990s. However, spine surgery has been a relatively late adopter. In general, surgical robots can be classified into 3 categories: telesurgical, supervisory-controlled, and shared-control. Telesurgical systems allow the surgeon to directly control the robotic instruments from a remote location. Supervisory-controlled systems allow the surgeon to preoperatively plan the surgery; once the plan is set, the robot executes it under close supervision by the surgeon. Shared-control systems allow the surgeon and robot to function simultaneously—both the robot and the surgeon control instruments and motions. Historically and currently, available spine robots are shared-control systems.

Early attempts to use robotic technology in spine surgery occurred in the 1990s but were met with varying degrees of success and significant workflow challenges. The Miro system (German Aerospace Center, Cologne, Germany) used an optical tracking system with central control to position a drill guide [39]. The SpineBot (Centre for Intelligent Surgery Systems, Hanyang University, South Korea) used preoperative computed tomography (CT) to plan pedicle screw instrumentation, and an intraoperative robotic arm and holder were used to localize the trajectory [28]. Two initial telesurgical systems—one made by Georgetown University that used joystick control and fluoroscopic-based imaging and a second, the Spine Bull's-Eye Robot (First Affiliated Hospital, Zhengzhou University, Henan Province, China)—demonstrated initial promise with regard to accurate placement of needles or guidewires in targeted spine locations, but neither reached significant commercial success [5,56].

In the early 2000s, with improvements in intraoperative CT imaging and navigation, there was a resurgence of interest in robotic-assisted spine surgery. Ultimately,

¹Weill Cornell Medicine, New York, NY, USA

²Hospital for Special Surgery, New York, NY, USA

Corresponding Author:

Sheeraz A. Qureshi, MD, MBA, Hospital for Special Surgery, 535 E. 70th St., New York, NY 10021, USA.
Email: sheerazqureshimd@gmail.com

Table 1. Comparison of spine robot platforms.

Spine robot platform	Integrated navigation	K-wire based screw placement	Independent instrument navigation	Spine-specific platform
Mazor SpineAssist	No	Yes	No	Yes
Mazor Renaissance	No	Yes	No	Yes
Mazor X	Yes	No	No	Yes
Mazor X Stealth Edition	Yes	No	Yes	Yes
ExcelsiusGPS	Yes	No	Yes	Yes
ROSA	No	Yes	No	No
ROSA ONE	Yes	Yes	Yes	No
TiRobot	Yes	Yes	No	No
Cirq	Yes	Yes	Yes	Yes

improved intraoperative navigation led to the development of several spine robots, now commercially available. However, early iterations of these robots were fraught with challenges, including software crashes, errors in synchronizing intraoperative fluoroscopic images with preoperative 3D imaging, problems deflecting the robotic arm resulting in decreased accuracy, and challenges with intraoperative user interfaces [48]. Technical or clinical errors were noted in more than 50% of spine procedures performed using robotic assistance during this time [2]. For example, pedicle screws had to be placed over guidewires, resulting in decreased workflow efficiency. In addition, platforms did not allow for real-time visual feedback for instrument localization, which meant surgeons were forced to trust the robot's accuracy and precision. Considering these challenges, this technology was not widely adopted.

Currently Available Robots

An overview of spinal robot platforms is provided in Table 1.

Mazor

The Mazor robotic platform (Medtronic Navigation, Louisville, CO, USA; Medtronic Spine, Memphis, TN, USA) was the first robotic platform cleared by the US Food and Drug Administration (FDA) for use in spine surgery. The SpineAssist, the first-generation Mazor robot, obtained FDA clearance in 2004 and continues to be one of the most widely used systems worldwide. It uses a patient-mounted track attached to an anatomical landmark and relies on preoperative and/or intraoperative CT imaging to allow for planning screw trajectories. The SpineAssist also allows for the pairing of preoperative CT with intraoperative fluoroscopy, as opposed to intraoperative CT imaging. The Renaissance, the second-generation Mazor robot, was released commercially in 2011. The Renaissance is smaller and lighter than the SpineAssist, and software improvements allow for 10 times faster processing speeds and

enhanced image recognition algorithms. It also provides surgeons the ability to flatten bone at desired pedicle screw start points to minimize the risk of skiving. Both the SpineAssist and Renaissance Mazor robots require pedicle screws to be placed over guidewires.

The Mazor X, the third-generation Mazor spine robot, launched in 2016, and upgrades include a fully automated robotic arm that no longer requires a patient-mounted track (as seen with first- and second-generation Mazor robots). Software upgrades allow for the registration of each vertebral body independently, providing multiplanar imaging datasets to assist with surgical planning. Last, an optic camera allows for real-time volumetric assessment of the operating room environment to self-detect its location within space and avoid collision intraoperatively. The acquisition of Mazor Robotics by Medtronic resulted in the integration of Medtronic StealthStation navigation technology, and the 2019 release of the Mazor X Stealth Edition, the latest generation Mazor robot (Fig. 1). The Mazor X Stealth Edition provides for fully independent navigation capability and real-time instrumentation position feedback.

Excelsius

The ExcelsiusGPS (Globus Medical, Inc., Audubon, PA, USA) was released in 2017 and is marketed as the first spine robot with a fully integrated navigation platform, allowing for real-time instrument tracking and pedicle screw placement without guidewires (Fig. 2). The robot is anchored to a floor-mounted base station, as opposed to the operative table. The navigation platform is compatible with multiple intraoperative 2-dimensional (2D) and 3D imaging modalities and can also merge a preoperative CT scan with intraoperative fluoroscopic imaging, similar to the latest generation Mazor robots. The ExcelsiusGPS robotic arm is rigid and capable of maintaining less than 1 mm of deflection when exposed to lateral forces up to 200 N, and excessive lateral force can be detected with a unique sensor [6]. Its independent navigation capability allows for localization



Fig. 1. (a) Mazor X Stealth Edition Spinal Robotic Platform (Medtronic Navigation Louisville, CO, USA; Medtronic Spine, Memphis, TN, USA) and (b) Intraoperative image demonstrating utilization of the Mazor Robotic platform to place lumbar pedicle screws in the prone position.




Fig. 2. (a) ExcelsiusGPS Spinal Robotic Platform (Globus Medical, Inc.) and (b) intraoperative image demonstrating utilization of the ExcelsiusGPS spine robot to place lumbar pedicle screws in the lateral position.

of manufacturer-specific instrumentation, similar to the Mazor X Stealth Edition, and its modular robotic arm allows for additional instruments to be mounted to the robot.

ROSA

The ROSA Spine Robot (Zimmer Biomet, Warsaw, IN, USA), FDA clearance in 2016, is a freestanding robotic system that uses preoperative or intraoperative CT imaging to plan screw trajectories and has a fully automated robotic arm, allowing for screw placement over guidewires. A recent upgrade, known as the ROSA ONE, received FDA clearance in 2019 (Fig. 3) and features a fully integrated

navigation system compatible with manufacturer-specific instrumentation, similar to the Mazor X Stealth Edition and Globus ExcelsiusGPS. The ROSA platform allows for navigation and instrumentation across spine, cranial, and knee arthroplasty procedures. This unique capability makes it a potentially more attractive option for hospitals or ambulatory surgery centers that have needs beyond spine-specific robotic platforms.

Recent Spine Robots

The Cirq spinal robot (Brainlab AG, Munich, Germany) received FDA clearance for spine procedures in September



Fig. 3. The ROSA ONE Spinal Robotic Platform (Zimmer Biomet).



Fig. 4. The Cirq spinal robotic platform (Brainlab AG).

2019, and its accuracy for placing cervical instrumentation has recently been assessed (Fig. 4) [10,21]. More recently, the Fusion Robotics spinal navigation and robotics system (Boulder, CO) received FDA clearance in February 2021 (Fig. 5). Finally, the TiRobot (TINAVI Medical Technologies, Beijing, China) was approved in China in 2016 but not in the United States. Like the ROSA platform, the TiRobot can also be used for other orthopedic applications outside of spine surgery.

Advantages of Robotic Technology

Safe and efficient spine surgery relies upon meticulous fine motor skills to navigate narrow corridors, particularly when placing pedicle screw instrumentation. These procedures

can be arduous; surgeons are susceptible to fatigue, and this may compromise accuracy and operative efficiency. Spinal robotic platforms offer a number of benefits, including reducing surgeon fatigue and tremors, while providing stability for instrumentation through a fixed working angle that may increase accuracy and precision. Ultimately, it is hoped that robotic platforms will increase operating room efficiency, result in more consistent and improved outcomes, decrease complications, and decrease overall health care costs. While much of the evidenced-based benefit of robotic technology in spine surgery has yet to be established, there is a growing body of evidence on accuracy and radiation exposure.

Pedicle Screw Placement

Several studies investigating the accuracy of robotic-assisted pedicle screw placement have demonstrated 91% to 98% accuracy [7,22,23,29,40,49,50]. Studies comparing robotic-assisted pedicle screw placement versus free-hand techniques have demonstrated varying results. Two studies demonstrated significantly higher accuracy with robot-assisted placement [9,54], while a prospective study demonstrated lower accuracy of robotic-assisted screws versus fluoroscopic-guided screws [42]. Given the inconsistencies in the literature, high-quality meta-analyses have been performed, 2 of which demonstrated equivalent accuracy between the 2 techniques [12,41], and a third demonstrated superior accuracy with robotic assistance [32]. Compared with 3D navigation techniques, robotic-assisted pedicle screw placement has demonstrated a slightly higher accuracy in some retrospective studies [30,43]. A recent meta-analysis demonstrated that compared with freehand techniques, both robotic-assisted and navigated screw placement resulted in lower intraoperative and postoperative screw revision risk [47].

Aside from potentially improved accuracy secondary to minimizing pedicle cortical breaches, robotic-guided screws may allow for less proximal facet violation, thereby potentially decreasing the risk of adjacent segment disease that can compromise clinical outcomes [1,44,53]. Several prospective studies and 2 meta-analyses have demonstrated a reduced risk of proximal facet joint violation when using robotic technology as compared with free-hand or fluoroscopically based pedicle screw placement [15,19,27,32,57,58].

There are currently no available peer-reviewed studies directly comparing the accuracy of the available spinal robotic systems. The majority of the published literature on instrumentation accuracy is based on the Mazor SpineAssist and Renaissance platforms, as these were the earliest platforms with advanced navigation technology to come to market. Table 2 highlights the accuracy results based on the literature for the Mazor SpineAssist, Mazor Renaissance,



Fig. 5. Fusion Robotics LLC spinal robotic platform.

Table 2. Peer-reviewed accuracy results across spine robot platforms.

Spine robot platform	Accuracy
Mazor SpineAssist	93%–98.5% [7,34,42]
Mazor Renaissance	98.9%–100% [17,19]
Mazor X	100% [24,25]
Mazor X Stealth Edition	100% [38]
ExcelsiusGPS	96.6%–100% [3,8,14,18,20,51,52]
ROSA	96.3%–98.3% [4,31,35]
TiRobot	95.3%–98.7%[15,54]

Mazor X, Mazor X Stealth Edition, ExcelsiusGPS, ROSA, and TiRobot robotic platforms.

Radiation

When comparing robotic-assisted surgery to freehand pedicle screw placement, many studies have demonstrated significantly decreased radiation exposure using the former approach [22,23,34,43], while 2 studies demonstrated no significant difference in radiation exposure between the 2 techniques [30,42]. However, some studies suggest that with robot-assisted pedicle screw placement, radiation exposure may be higher to the patient [42] and lower to the surgeon [54]. We need studies comparing radiation exposure resulting from the use of both robotic platforms and traditional 3D navigation-based techniques in pedicle screw placement.

Clinical Outcomes

Studies examining clinical outcomes of robotic technology in spine surgery are difficult to interpret due to study design. Most of the available literature compares traditional open procedures to robot-assisted minimally invasive spine surgery (MISS). Therefore, it is not clear if it is the robotic component of the procedure or the less invasive aspect of MISS that is contributing to lower complication rates, decreased length of stay, less postoperative opioid consumption, and faster postoperative recovery [19,22,23,36,54]. To better understand the additive benefit of robotic technology in a similar cohort of spine procedures, well-designed studies are needed to compare clinical outcomes in traditional open versus robot-assisted open spine surgery or MISS using fluoroscopic or real-time navigation versus robot-assisted MISS.

Complications of Robotic Technology

We have focused mostly on the potential benefits of robotic technology in spine surgery, but several papers have demonstrated contrasting results. For example, when compared with freehand techniques of pedicle screw placement, the accuracy of robotic-assisted pedicle screw placement has not been found to be consistently higher. In fact, 2 of the meta-analyses mentioned earlier failed to demonstrate higher accuracy of robotic-assisted pedicle screw placement [12,41], while Ringel et al found the approach demonstrated lower accuracy [42].

Two recent database studies that compared robot-assisted lumbar spinal fusion with conventional lumbar spinal fusion demonstrated increased short-term complications and revisions associated with robotic assistance. Yang et al found that higher risk of revision, 30-day readmission, and total 30-day complications were associated with robotic assistance during lumbar spinal fusion [55]. A study by Lieber et al also demonstrated a higher inpatient complication rate associated with robotic assistance during lumbar spinal fusion [33]. One possible cause of such increased complications may be the longer operative time associated with robotic-assisted surgery—a longer operative time can lead to increases in perioperative complications, particularly infection. Another explanation may be the learning curve associated with adopting robotic technology, which may result in higher initial rates of screw malposition that ultimately require early revision. However, the limitation of both studies is the use of large national databases that limit patient identification to procedure codes, not allowing for the further stratification of MISS versus open procedures. In addition, both studies are subject to the standard limitations associated with any database study, such as selection bias and coding error.

Costs

The use of robotics in spine surgery has increased 169% over the past decade and has linearly increased each year [55]. This has led to significant concerns about expenditures associated with robotic technology, as it can require significant capital acquisition costs—for some platforms, costs near \$1,000,000. In order for robotic technology in spine surgery to be economically pragmatic, there needs to be a proven benefit demonstrated through decreased operative time, lower postoperative lengths of stay and revision rates, and improved postoperative patient-reported outcomes.

While there is no current consensus on the economic practicality of robotic technology in spine surgery, studies citing decreased revision rates suggest that this may be the most likely justification for robotic technology in spine surgery from a cost standpoint; lower revision rates decrease direct health care expenditure related to additional surgery and also minimize the indirect societal cost of prolonged disability [11,45]. In examining retrospective data from 1 academic center, Menger et al calculated a cost benefit of approximately \$600,000 over 1 year due to shorter operative times and lower revision rates associated with robotic spine surgery [37]. Further prospective studies are needed to evaluate the overall costs of robotic technology in spine surgery, but it may become more favorable as acquisition and maintenance costs continue to decrease and robotic technology advances.

Future Advances

The latest generation of spine robot platforms have made significant advances in real-time navigation, improved intraoperative user interface, and workflow efficiency. More research is needed to increase the use of robotics in spine surgery. Studies are needed to assess whether the potentially improved outcomes, reduced complications, and decreased health care costs justify the capital expense of a robotic platform. In addition, robotic technology needs to evolve beyond just facilitating pedicle screw placement. Some platforms currently allow surgeons to plan tubular retractor trajectories and can stabilize retractors with the robotic arm. In addition, some surgeons are using the integrated navigation platform to plan interbody placement and execute navigated disk preparation. However, a minority of surgeons make use of these relatively novel advancements. As these features become more streamlined, more surgeons will likely integrate them into the surgical workflow.

Future spine robots may have burrs that are compatible with the robotic platform and allow for precise intraoperative execution of preoperatively planned bony decompression. Haptic feedback must also improve for this feature to evolve. Finally, it is possible that pedicle screw placement becomes fully automated and evolves from the current shared-controlled platform. As spinal robotic technology continues to advance, capital cost decreases, and further research is conducted that demonstrates the potential clinical and economic benefits of robotics in spine surgery, it is more likely that spine surgeons will adopt it in their operating rooms.

In conclusion, spinal robotic technology has made marked advancement over the past 2 decades, and the latest generation of robots is just starting to penetrate the spine market. Along with expansions in the coming years will come increased competition and innovation.

CME Credit

Please go to HSS eAcademy at <https://bit.ly/HSSJCME> to find all journal-related CME, complete the online post-test, and claim CME credit.

Declaration of Conflicting Interests

The author(s) declared the following potential conflicts of interest with respect to the research, authorship, and/or publication of this article: Darren R. Lebl, MD, reports relationships with Medtronic, Nuvasive, Stryker, Depuy Synthes, Guidepoint, Remedy Logic, ISPH II, Vestia Ventures, MiRUS, Woven Orthopedic Technologies, Integrity Implants, and HS2. Sheeraz A. Qureshi, MD, MBA, reports relationships with Stryker K2M, Globus Medical, Paradigm Spine, AMOpportunities, RTI Surgical, Integrity Implants, Medical Device Business Services, Medtronic USA, Nuvasive, Avaz Surgical, Simplify Medical, Tissue Differentiation Intelligence, Vital 5, Spinal Simplicity, LifeLink.

com, Healthgrades, Society of Minimally Invasive Spine Surgery, Simplify Medical, North American Spine Society, Minimally Invasive Spine Study Group, Lumbar Spine Research Society, International Society for the Advancement of Spine Surgery, *Contemporary Spine Surgery*, Cervical Spine Research Society, Association of Bone and Joint Surgeons, and *Annals of Translational Medicine*. Ram K. Alluri, MD, Fedan Avrumova, BS, Ahilan Sivaganesan, MD, and Avani S. Vaishnav, MBBS, declare no potential conflicts of interest.

Funding

The author(s) received no financial support for the research, authorship, and/or publication of this article.

Human/Animal Rights

All procedures followed were in accordance with the ethical standards of the responsible committee on human experimentation (institutional and national) and with the Helsinki Declaration of 1975, as revised in 2013.

Required Author Forms

Disclosure forms provided by the authors are available with the online version of this article as supplemental material.

References

1. Bagheri SR, Alimohammadi E, Froushani AZ, Abdi A. Adjacent segment disease after posterior lumbar instrumentation surgery for degenerative disease: incidence and risk factors. *J Orthop Surg (Hong Kong)*. 2019;27:1–6.
2. Barzilay Y, Liebergall M, Fridlander A, Knoller N. Miniature robotic guidance for spine surgery—introduction of a novel system and analysis of challenges encountered during the clinical development phase at two spine centres. *Int J Med Robot*. 2006;2(2):146–153.
3. Benech CA, Perez R, Benech F, Greeley SL, Crawford N, Ledonio C. Navigated robotic assistance results in improved screw accuracy and positive clinical outcomes: an evaluation of the first 54 cases. *J Robot Surg*. 2019;14(3):431–437.
4. Chenin L, Capel C, Fichten A, Peltier J, Lefranc M. Evaluation of screw placement accuracy in circumferential lumbar arthrodesis using robotic assistance and intraoperative flat-panel computed tomography. *World Neurosurg*. 2017;105:86–94.
5. Cleary K, Watson V, Lindisch D, et al. Precision placement of instruments for minimally invasive procedures using a “needle driver” robot. *Int J Med Robot Comput Assist Surg*. 2005;1(2):40–47.
6. Crawford N, Johnson N, Theodore N. Ensuring navigation integrity using robotics in spine surgery. *J Robot Surg*. 2020;14:177–183.
7. Devito DP, Kaplan L, Dietl R, et al. Clinical acceptance and accuracy assessment of spinal implants guided with SpineAssist surgical robot: retrospective study. *Spine (Phila Pa 1976)*. 2010;35:2109–2115.
8. Elswick CM, Strong MJ, Joseph JR, Saadeh Y, Oppenlander M, Park P. Robotic-assisted spinal surgery: current generation instrumentation and new applications. *Neurosurg Clin N Am*. 2020;31:103–110.
9. Fan Y, Du J, Zhang J, et al. Comparison of accuracy of pedicle screw insertion among 4 guided technologies in spine surgery. *Med Sci Monit*. 2017;23:5960–5968.
10. Farah K, Meyer M, Prost S, Dufour H, Blondel B, Fuentes S. Cirq® robotic assistance for minimally invasive c1-c2 posterior instrumentation: report on feasibility and safety [published online ahead of print July 25, 2020]. *Oper Neurosurg*. <https://doi.org/10.1093/ons/opaa208>.
11. Fiani B, Quadri S, Farooqui M, et al. Impact of robot-assisted spine surgery on health care quality and neurosurgical economics: a systemic review. *Neurosurg Rev*. 2020;43(1):17–25.
12. Gao S, Lv Z, Fang H. Robot-assisted and conventional freehand pedicle screw placement: a systematic review and meta-analysis of randomized controlled trials. *Eur Spine J*. 2018;27:920–930.
13. Ge DH, Stekas ND, Varlotta CG, et al. Comparative analysis of two transforaminal lumbar interbody fusion techniques: open TLIF versus Wiltse MIS TLIF. *Spine (Phila Pa 1976)*. 2019;44(9):E555–E560.
14. Godzik J, Walker CT, Hartman C, et al. A quantitative assessment of the accuracy and reliability of robotically guided percutaneous pedicle screw placement: technique and application accuracy. *Oper Neurosurg*. 2019;17:389–395.
15. Han X, Tian W, Liu Y, et al. Safety and accuracy of robot-assisted versus fluoroscopy-assisted pedicle screw insertion in thoracolumbar spinal surgery: a prospective randomized controlled trial [published online ahead of print February 8, 2019]. *J Neurosurg Spine*. <https://doi.org/10.3171/2018.10.SPINE18487>.
16. Hockley A, Ge D, Vasquez-Montes D, et al. Minimally invasive versus open transforaminal lumbar interbody fusion surgery: an analysis of opioids, nonopioid analgesics, and perioperative characteristics. *Glob Spine J*. 2019;9(6):624–629.
17. Hu X, Ohmmeiss DD, Lieberman IH. Robotic-assisted pedicle screw placement: lessons learned from the first 102 patients. *Eur Spine J*. 2013;22:661–666.
18. Huntsman KT, Ahrendtsen LA, Riggleman JR, Ledonio CG. Robotic-assisted navigated minimally invasive pedicle screw placement in the first 100 cases at a single institution. *J Robot Surg*. 2020;14(1):199–203.
19. Hyun SJ, Kim KJ, Jahng TA, Kim HJ. Minimally invasive robotic versus open fluoroscopic-guided spinal instrumented fusions: a randomized controlled trial. *Spine (Phila Pa 1976)*. 2017;42(6):353–358.
20. Jain D, Manning J, Lord E, et al. Initial single-institution experience with a novel robotic-navigation system for thoracolumbar pedicle screw and pelvic screw placement with 643 screws. *Int J Spine Surg*. 2019;13:459–463.
21. Kaissar F, Mikael M, Solene P, et al. Robotic assistance for minimally invasive cervical pedicle instrumentation: report on feasibility and safety. *World Neurosurg*. 2021;150:e777–e782.
22. Kantelhardt SR, Martinez R, Baerwinkel S, Burger R, Giese A, Rohde V. Perioperative course and accuracy of screw positioning in conventional, open robotic-guided and percutaneous robotic-guided, pedicle screw placement. *Eur Spine J*. 2011;20:860–868.

23. Keric N, Doenitz C, Haj A, et al. Evaluation of robot-guided minimally invasive implantation of 2067 pedicle screws. *Neurosurg Focus*. 2017;42:E11.
24. Khan A, Meyers JE, Siasios I, Pollina J. Next-generation robotic spine surgery: first report on feasibility, safety, and learning curve. *Oper Neurosurg*. 2019;17:61–69.
25. Khan A, Meyers JE, Yavorek S, et al. Comparing next-generation robotic technology with 3-dimensional computed tomography navigation technology for the insertion of posterior pedicle screws. *World Neurosurg*. 2019;123:e474–e481.
26. Kim CH, Easley K, Lee JS, et al. Comparison of minimally invasive versus open transforaminal interbody lumbar fusion. *Glob Spine J*. 2020;10(suppl 2):143S–150S.
27. Kim HJ, Jung WI, Chang BS, Lee CK, Kang KT, Yeom JS. A prospective, randomized, controlled trial of robot-assisted vs freehand pedicle screw fixation in spine surgery. *Int J Med Robot*. 2017;13:1–7.
28. Kim S, Chung J, Yi BJ, Kim YS. An assistive image-guided surgical robot system using O-arm fluoroscopy for pedicle screw insertion: preliminary and cadaveric study. *Neurosurgery*. 2010;67(6):1757–1767.
29. Kuo KL, Su YF, Wu CH, et al. Assessing the intraoperative accuracy of pedicle screw placement by using a bone-mounted miniature robot system through secondary registration. *PLoS ONE*. 2016;11:e0153235.
30. Laudato PA, Pierzchala K, Schizas C. Pedicle screw insertion accuracy using O-arm, robotic guidance, or freehand technique: a comparative study. *Spine (Phila Pa 1976)*. 2018;43(6):E373–E378.
31. Lefranc M, Peltier J. Evaluation of the ROSATM Spine robot for minimally invasive surgical procedures. *Expert Rev Med Devices*. 2016;13(10):899–906.
32. Li HM, Zhang RJ, Shen CL. Accuracy of pedicle screw placement and clinical outcomes of robot-assisted technique versus conventional freehand technique in spine surgery from nine randomized controlled trials: a meta-analysis. *Spine (Phila Pa 1976)*. 2020;45:E111–E119.
33. Lieber AM, Kirchner GJ, Kerbel YE, Khalsa AS. Robotic-assisted pedicle screw placement fails to reduce overall postoperative complications in fusion surgery. *Spine J*. 2019;19(2):212–217.
34. Lieberman IH, Hardenbrook MA, Wang JC, Guyer RD. Assessment of pedicle screw placement accuracy, procedure time, and radiation exposure using a miniature robotic guidance system. *J Spinal Disord Tech*. 2012;25(5):241–248.
35. Lonjon N, Chan-Seng E, Costalat V, Bonnafoux B, Vassal M, Boetto J. Robot-assisted spine surgery: feasibility study through a prospective case-matched analysis. *Eur Spine J*. 2016;25:947–955.
36. Lucio JC, Vanconia RB, Deluzio KJ, Lehman JA, Rodgers JA, Rodgers W. Economics of less invasive spinal surgery: an analysis of hospital cost differences between open and minimally invasive instrumented spinal fusion procedures during the perioperative period. *Risk Manag Heal Policy*. 2012;5:65–74.
37. Menger RP, Savardekar AR, Farokhi F, Sin A. A cost-effectiveness analysis of the integration of robotic spine technology in spine surgery. *Neurospine*. 2018;15(3):216–224.
38. O'Connor TE, O'Hehir MM, Khan A, et al. Mazor X stealth robotic technology: a technical note. *World Neurosurg*. 2021;145:435–442.
39. Ortmaier T, Weiss H, Hagn U, et al. A hands-on-robot for accurate placement of pedicle screws. In: *IEEE International Conference on Robotics and Automation, May 2006*. IEEE; 2006:4179–4186. Available at: <https://ieeexplore.ieee.org/document/1642345>.
40. Pechlivanis I, Kiriyanthan G, Engelhardt M, et al. Percutaneous placement of pedicle screws in the lumbar spine using a bone mounted miniature robotic system: first experiences and accuracy of screw placement. *Spine (Phila Pa 1976)*. 2009;34:392–398.
41. Peng YN, Tsai LC, Hsu HC, Kao CH. Accuracy of robot-assisted versus conventional freehand pedicle screw placement in spine surgery: a systematic review and meta-analysis of randomized controlled trials. *Ann Transl Med*. 2020;8(13):824.
42. Ringel F, Stuer C, Reinke A, et al. Accuracy of robot-assisted placement of lumbar and sacral pedicle screws: a prospective randomized comparison to conventional freehand screw implantation. *Spine (Phila Pa 1976)*. 2012;37(8):E496–E501.
43. Roser F, Tatagiba M, Maier G. Spinal robotics: current applications and future perspectives. *Neurosurgery*. 2013;72(suppl 1):12–18.
44. Sakaura H, Miwa T, Yamashita T, Kuroda Y, Ohwada T. Cortical bone trajectory screw fixation versus traditional pedicle screw fixation for 2-level posterior lumbar interbody fusion: comparison of surgical outcomes for 2-level degenerative lumbar spondylolisthesis. *J Neurosurg Spine*. 2018;28:57–62.
45. Schroerlucke SR, Good CR, Wang MY. A prospective, comparative study of robotic-guidance versus freehand in minimally invasive spinal fusion surgery: first report from MIS ReFRESH. *Spine J*. 2016;16:S253.
46. Shin DA, Kim KN, Shin HC, Yoon DH. The efficacy of microendoscopic discectomy in reducing iatrogenic muscle injury. *J Neurosurg Spine*. 2008;8:39–43.
47. Staartjes VE, Klukowska AM, Schroder ML. Pedicle screw revision in robot-guided, navigated, and freehand thoracolumbar instrumentation: a systematic review and meta-analysis. *World Neurosurg*. 2018;116:433–443.
48. Sukovich W, Brink-Danan S, Hardenbrook M. Miniature robotic guidance for pedicle screw placement in posterior spinal fusion: early clinical experience with the SpineAssist. *Int J Med Robot*. 2006;2:114–122.
49. Tsai TH, Wu DS, Su YF, Wu CH, Lin CL. A retrospective study to validate an intraoperative robotic classification system for assessing the accuracy of Kirschner wire (K-wire) placements with postoperative computed tomography classification system for assessing the accuracy of pedicle screw placements. *Medicine (Baltimore)*. 2016;95:e4834.
50. van Dijk JD, van den Ende RP, Stramigioli S, Kochling M, Hoss N. Clinical pedicle screw accuracy and deviation from planning in robot-guided spine surgery: robot-guided pedicle screw accuracy. *Spine (Phila Pa 1976)*. 2015;40:E986–E991.

51. Vardiman AB, Wallace DJ, Crawford NR, Riggleman JR, Ahrendsten LA, Ledonio CG. Pedicle screw accuracy in clinical utilization of minimally invasive navigated robot-assisted spine surgery. *J Robot Surg*. 2020;14(3):409–413.
52. Wallace DJ, Vardiman AB, Booher GA, et al. Navigated robotic assistance improves pedicle screw accuracy in minimally invasive surgery of the lumbosacral spine: 600 pedicle screws in a single institution. *Int J Med Robot*. 2020;16:e2054.
53. Wang H, Ma L, Yang D, et al. Incidence and risk factors of adjacent segment disease following posterior decompression and instrumented fusion for degenerative lumbar disorders. *Medicine (Baltimore)*. 2017;96:e6032.
54. Xiaofeng L, Tian W, Shi Z, et al. Robot-assisted versus fluoroscopy-assisted cortical bone trajectory screw instrumentation in lumbar spinal surgery: a matched-cohort comparison. *World Neurosurg*. 2018;120:e745–e751.
55. Yang DS, Li NY, Kleinhenz DT, Patel S, Daniels AH. Risk of postoperative complications and revision surgery following robot-assisted posterior lumbar spinal fusion. *Spine (Phila Pa 1976)*. 2020;45(24):E1692–E1698.
56. Zhang C, Wang Z, Zhang C, Chen F, Zhang H, Yan X. Spine Bull's-Eye Robot guidewire placement with pedicle standard axis view for thoracic and lumbar pedicle screw fixation. *J Spinal Disord Tech*. 2012;25(7):E191–E198.
57. Zhang Q, Xu YF, Tian W, et al. Comparison of superior-level facet joint violations between robot-assisted percutaneous pedicle screw placement and conventional open fluoroscopic-guided pedicle screw placement. *Orthop Surg*. 2019;11(5):850–856.
58. Zhou LP, Zhang RJ, Li HM, Shen CL. Comparison of cranial facet joint violation rate and four other clinical indexes between robot-assisted and freehand pedicle screw placement in spine surgery: a meta-analysis. *Spine (Phila Pa 1976)*. 2020;45(22):E1532–E1540.