

Computer-Aided Diagnostic Systems for Osteoporotic Vertebral Fracture Detection: Opportunities and Challenges

In the current issue of *JBMR*, Kolanu and colleagues evaluate a computer-aided diagnostic (CAD) system designed to identify osteoporotic vertebral fractures (VFs) visualized opportunistically from computed tomography (CT) images.⁽¹⁾ The system, developed by Zebra Medical Vision (Shefayim, Israel; www.zebra-med.com), extracts a virtual sagittal section visualizing the spinal midplane and identifies VFs using machine-learning algorithms. It outputs the probability that the volume contains a VF and a heat map indicating the probable locations of VFs in the sagittal image. In a single-site study involving thoracic CT scans from 1696 patients with a VF prevalence of 24%, the system achieved a sensitivity, specificity, and accuracy of 54%, 92%, and 83%, respectively.

Osteoporosis is underdiagnosed and undertreated in clinical practice.⁽²⁾ Although VFs are the most common osteoporotic fracture, they can be clinically “silent” despite being a crucial early clinical manifestation of the disease. This creates a potential role for opportunistic assessment for VFs in clinical images acquired for other indications, with CT arguably presenting the most important target. A large and increasing number of procedures are performed, and they frequently visualize part of the spine, offering the possibility of identifying a subset of the currently undiagnosed osteoporosis patients at little additional cost. However, current diagnosis rates for CT are significantly worse than other modalities; a recent, UK-wide audit found that VFs visualized opportunistically in CT images were accurately reported in only 26.2% of patients, and only 2.6% were referred for further management.⁽³⁾ The feasibility of improvement through existing radiology service provision is questionable. CT imaging activity has increased rapidly in recent years, rising by 110% between 2000 and 2016 in the US and by 69% in the UK NHS between 2012–13 and 2018–19.^(4,5) In contrast, the NHS currently has a shortfall of 1876 radiologists, 33% of posts, forecast to rise to 43% by 2024, and only 1% of trusts were able to meet their reporting requirements within contracted hours in 2019.⁽⁶⁾ Similarly, the American Association of Medical Colleges has predicted a shortfall of between 17,100 and 41,900 physicians in specialties including radiology by 2033.⁽⁷⁾

CAD systems may provide a solution, with potential benefits for the patients in whom VFs are identified. However, identification of the fracture is only part of the jigsaw puzzle, and integration into existing clinical workflows presents considerable challenges. The current shortfall in radiologists militates against any solution that increases clinical workload. Kolanu and

colleagues show that Zebra’s CAD can identify images containing VFs and highlight the probable location of the VF within the image, but the results are neither perfectly accurate nor complete. The radiologist cannot yet *rely* on the system without thoroughly assessing the sagittal reformats, which takes precious reporting time. The “elephant in the room” is that manual input from radiologists is still needed to confirm the diagnosis, identify the vertebral level, grade the severity of any fracture, and make recommendations for future management in line with local service provision in order to complete a radiology report that complies with clinical guidelines.⁽⁸⁾

Other CAD-based VF detection methods for clinical CT images are emerging, including those of Infervision (Beijing, China; global.infervision.com), O.N. Diagnostics (Berkeley, CA, USA; www.ondiagnostics.com), and Optasia Medical Ltd. (Manchester, UK; www.optasiamedical.com). Systems that estimate bone mineral density (BMD) from CT, such as CliniQCT (Mindways Inc, Austin, TX, USA; www.qct.com) or that combine BMD assessment with bone strength analysis, such as VirtuOst (O.N. Diagnostics), are also available. In contrast to the Zebra and Infervision systems, Optasia Medical’s ASPIRE operates as a tele-radiology service, outsourcing the opportunistic reporting of VFs with an in-house radiologist confirming the diagnoses, thus removing workload from radiology departments. Nevertheless, these systems all have the potential to significantly increase the number of VFs identified and thus impact downstream services, a factor that must be considered at the point of adoption.

CAD systems will have utility only if they can be incorporated into clinical workflows without imposing unrealistic burdens. This requires consideration of multiple issues and potential technological interventions, of which improved VF diagnostic rates are only one. Integration into existing radiology services must avoid increasing workload, and subsequent patient management must similarly avoid overwhelming the capacity of downstream services. CAD systems should therefore produce actionable radiology reports that minimize the requirement for manual input or further diagnostic testing. Computerized patient management systems that automate routine tasks, such as scheduling of patient contact and communications with primary care, might mitigate the impact of increased patient numbers. It is also important to remember the ultimate aim: VF detection has clinical utility only if it enables prevention of future fractures through appropriate prescription of medications to reduce fracture risk and high enough patient adherence with medication

Received in original form September 13, 2020; revised form September 29, 2020; accepted October 7, 2020.

Journal of Bone and Mineral Research, Vol. 35, No. 12, December 2020, pp 2305–2306.
DOI: 10.1002/jbmr.4205

© 2020 The Authors. *Journal of Bone and Mineral Research* published by Wiley Periodicals LLC on behalf of American Society for Bone and Mineral Research (ASBMR).

This is an open access article under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

regimes. A retrospective study of Zebra's CAD system on a cohort of 48,227 patients found that fracture risk prediction using CT-based BMD estimation and VF detection was only slightly better than that provided by FRAX (www.shef.ac.uk/FRAX), which estimates fracture risk based on clinical factors, without BMD estimation.⁽⁹⁾

CAD systems may have clinical utility, even without increasing diagnostic accuracy, if they automate fracture risk estimation while reducing clinical workload. However, careful attention to health economics is required. Studies have been performed for specific interventions, such as bisphosphonate drugs.⁽¹⁰⁾ However, these are insufficient to model the full impact of appropriate patient management and will underestimate patient benefit unless the full range of standard interventions, such as falls risk assessments, is also considered. Finally, adherence and persistence for first-line osteoporosis drugs are relatively poor.⁽¹⁰⁾ Long-term patient support and monitoring are required to ensure persistence rates sufficient to reduce fracture risk. Simple technologies such as mobile telephone apps that facilitate communication between patients, clinicians, and patient support groups could therefore be highly effective.

In conclusion, a future where patients can benefit from machine-learning-based diagnostic systems in the diagnosis and management of osteoporosis seems increasingly likely. CAD systems with a high sensitivity and specificity for VFs are one facet of a completely novel clinical pathway, where technological support at multiple points will be required.

Disclosures

PAB has collaborated with Optasia Medical Ltd. on the development of the ASPIRE service but has no financial interest in the company or service. EMC and KEP state that they have no conflicts of interest.

Acknowledgments

Authors' roles: PAB, EMC, and KEP discussed and agreed upon the contents of the article. PAB wrote the article with substantial input from EMC and KEP. All authors reviewed and agreed upon the final version.

Author contributions: PAB: Writing-original draft; writing-review and editing. EMC: Writing-original draft; writing-review and editing. KEP: Writing-original draft; writing-review and editing.

Paul A Bromiley PhD 

Division of Informatics, Imaging, and Data Sciences, The University of Manchester, Manchester, UK

Emma M Clark MBBS, PhD 

Translational Health Sciences, Bristol Medical School, University of Bristol, Bristol, UK

Kenneth E Poole BM, PhD 

Cambridge NIHR Biomedical Research Centre, Department of Medicine, University of Cambridge, Cambridge, UK

References

1. Kolanu N, Silverstone EJ, Ho BH, et al. Clinical utility of computer-aided diagnosis of vertebral fractures from computed tomography images. *J Bone Miner Res*. 2020. Epub ahead of print. Available from: <https://doi.org/10.1002/jbmr.4146>.
2. Akesson K, Marsh D, Mitchell PJ, et al. Capture the fracture: a best practice framework and global campaign to break the fragility fracture cycle. *Osteoporos Int*. 2013;24(8):2135–52.
3. Howlett DC, Drinkwater KJ, Mahmood N, Illes J, Griffin J, Javaid K. Radiology reporting of osteoporotic vertebral fragility fractures on computed tomography studies: results of a UK national audit. *Eur Radiol*. 2020;30(9):4713–23.
4. Smith-Bindman R, Kwan ML, Marlow EC, et al. Trends in use of medical imaging in US health care systems and in Ontario, Canada, 2000–2016. *JAMA*. 2019;322(9):843–56.
5. The National Health Service. Diagnostic imaging dataset statistical release 2018/2019. 2019. Available at: www.england.nhs.uk/statistics/wp-content/uploads/sites/2/2019/12/Annual-Statistical-Release-2018-19-PDF-1.9MB.pdf. Accessed 2020 Sept 11.
6. The Royal College of Radiologists. Clinical radiology UK workforce census 2019 report. 2020. Available at: www.rcr.ac.uk/system/files/publication/field_publication_files/clinical-radiology-uk-workforce-census-2019-report.pdf. Accessed 2020 Sept 11.
7. Association of American Medical Colleges. The complexities of physician supply and demand: projections from 2018 to 2033. 2020. Available at: www.aamc.org/system/files/2020-06/stratcomm-aamc-physician-workforce-projections-june-2020.pdf. Accessed 2020 Sept 11.
8. Compston J, Cooper A, Cooper C, et al. UK clinical guideline for the prevention and treatment of osteoporosis. *Arch Osteoporos*. 2017;12(1):43.
9. Dagan N, Elnekave E, Barda N, et al. Automated opportunistic osteoporotic fracture risk assessment using computed tomography scans to aid in FRAX underutilization. *Nat Med*. 2020;26:77–82.
10. Davis S, Martyn-St James M, Sanderson J, et al. A systematic review and economic evaluation of bisphosphonates for the prevention of fragility fractures. *Health Technol Assess*. 2016;20(78):1–406.