


Review

How to evaluate a flexible ureterorenoscope? Systematic mapping of existing evaluation methods

Nora Hendriks^{1,2} , Michaël M.E.L. Henderickx^{1,3}, Barbara M.A. Schout², Joyce Baard¹, Faridi S. van Etten-Jamaludin⁴, Harrie P. Beerlage¹, Rob C.M. Pelger⁵ and Guido M. Kamphuis¹

¹Department of Urology, Amsterdam UMC, University of Amsterdam, Amsterdam, ²Department of Urology, Alrijne Hospital, Leiderdorp, the Netherlands, ³Department of Urology, GZA Hospitals, Antwerp, Belgium, ⁴Research Support, Medical library location AMC, Amsterdam UMC, University of Amsterdam, Amsterdam, and ⁵Department of Urology, Leids UMC, University of Leiden, Leiden, the Netherlands

N.H. and M.M.E.L.H contributed equally to this paper.

Objectives

The objective of this study was to identify, map and review scope-related and user-related parameters used to evaluate the quality of flexible ureterorenoscopes. Thereby identifying key items and variability in grading systems.

Methods

A literature search of four databases (MEDLINE [Ovid], EMBASE [Ovid], Web of Science, Google scholar and the Cochrane Library) was performed according to the Preferred Reporting Items for Systematic Reviews and Meta-Analyses guidelines encompassing articles published up to August 2020. A total of 2386 articles were screened.

Results

A total of 48 articles were included in this systematic scoping review. All studies had a prospective design. Five key items in the assessment of flexible ureterorenoscopy were distinguished: ‘Manoeuvrability’ (87.5%), ‘Optics’ (64.6%), ‘Irrigation’ (56.3%), ‘Handling’ (39.6%) and ‘Durability’ (35.4%). After regrouping, every key item could be divided into specific subcategories. However, the quality assessment showed a wide variation in denomination, method of measurement, circumstances of measurement, tools used during measurements, number of measurements performed, number of observers, and units of outcomes.

Conclusion

The research field regarding quality assessment of ureterorenoscopes is heterogeneous. In this systematic scoping review we identified five key parameters: Manoeuvrability, Optics, Irrigation, Handling and Durability, used to grade flexible ureterorenoscopes. However, within these categories we found a wide variety in terms of method of measurements. A standardised, uniform grading tool is required to assess and compare the quality of flexible ureterorenoscopes in the future.

Keywords

evaluation, ureterorenoscope, ureteroscopy

Introduction

The first ureteroscopy was performed by Young in 1912 using a cystoscope in an infant with dilated ureters that advanced easily to the renal pelvis [1]. Marshall was the first to describe the use of a flexible fibre optic ureteroscope in 1964 and the first intended ureterorenoscopy (URS) was reported in 1977 [2]. This continuous evolution progressed rapidly in the 1980s, when flexible ureterorenoscopes were increasingly used to treat stones.

Since the introduction of flexible ureterorenoscopes in the 1970s, efforts have been made to reduce their size whilst optimising the working channel for irrigation flow and introducing numerous accessory instruments. Meanwhile, retaining optimal bidirectional deflecting properties and image quality [3,4]. Attempts to serve these seemingly conflicting interests resulted in flexible ureterorenoscopes which have become more fragile and prone to damage with loss of functionality [5]. Overcoming these problems of fragility, may result in ureterorenoscopes which are less

ergonomic or user-friendly. This loss of functionality could have an impact on procedure time, perioperative tissue damage, and surgical outcomes of URS.

Over the years, a wide range of flexible scopes have been introduced for clinical use. Functionality as well as characteristics related to the loss of functionality of these flexible ureterorenoscopes have been studied abundantly [5–52]. Until now, studies used different approaches to evaluate the quality of flexible ureterorenoscopes. To objectively evaluate and compare parameters of different flexible ureterorenoscopes, a standardised way of assessment is needed to assess quality pre-, peri-, and postoperatively. The first step towards such an assessment tool is the identification of key parameters used to assess quality. Therefore, the aim of the present study was to systematically identify, map and review scope-related and user-related parameters used to evaluate flexible ureterorenoscopes in the current literature.

Evidence Acquisition

This review was conducted in accordance with the Preferred Reporting Items for Systematic Reviews and Meta-Analyses (PRISMA) guidelines [53]. However, for this study we systematically reviewed the different evaluation methods and not the actual measurement outcomes of these studies. Therefore, a meta-analysis and a risk of bias was not performed.

Search Strategy

A literature search was performed in collaboration with a medical librarian in August 2020, covering four different databases: MEDLINE (Ovid), EMBASE (Ovid), Web of Science, Google scholar, and the Cochrane library. The Medical Subject Headings (MeSH) term ‘Ureteroscope/ureteroscopy’ was used in combination (AND, OR) with ‘flexible’, ‘evaluation’, ‘urolithiasis’, ‘equipment design’, ‘optic/image’, ‘manoeuvrability’ and ‘durability’. In addition, derivatives and terms with no available MeSH terms were used. The search strategy for the four databases is shown in Appendix S1.

Inclusion Criteria

1. Peer-reviewed articles
2. Articles written in English
3. Studies evaluating scope-related parameters of flexible ureterorenoscopes
4. Studies evaluating user parameters of flexible ureterorenoscopes

Exclusion Criteria

1. Editorials/Letters to the editor

2. Case reports/Abstracts
3. Reviews
4. Ongoing trials
5. Studies outside the specialism of urology
6. Studies without any description on method of measurement (e.g. ‘this scope is excellent’)
7. Studies not assessing scope or user-related parameters

Search Strategy and Outcome

A total of 2386 articles were left to screen after removing duplicates and screening the reference lists of the included articles for potentially relevant studies. All abstracts were reviewed by two independent reviewers (N.H. and M.H.), following pre-defined inclusion and exclusion criteria to retrieve relevant articles. Subsequently, full-text versions of the relevant articles were retrieved and screened for final inclusion. Any disagreement regarding the inclusion of an article was resolved based on consensus. If consensus could not be reached, a third reviewer (G.K.) was asked to make a final judgment. The results of the search strategy and the selection process are shown in a PRISMA flow chart (Figure 1). All articles screened for full text are appended in Appendix S3. A summary of the included articles is presented in Table 1 [5–52].

Types of Study Designs Included

For this systematic review, all peer-reviewed articles describing original data (case series, cohort studies, randomised controlled trials [RCTs], quasi-RCTs, and non-randomised studies) were eligible for inclusion. Additionally, all comparative and non-comparative studies were eligible for inclusion. As the aim of this study was to map all types of grading systems, *in vivo* clinical studies, *ex vivo* bench top evaluations as well as *ex vivo* simulations were included as presented in Table 1.

Types of Ureterorenoscopes Included


The flexible scopes that were assessed in the included studies consisted of both single-use and re-usable flexible ureterorenoscopes.

Evidence Synthesis

Included Studies

A total of 48 articles were included in this review [5–52]. The main characteristics of these 48 articles are presented in Table 1. All studies had a prospective design and were published between 1997 and 2020.

Studies did not distinguish different methods of evaluation for different types of scopes (reusable vs single-use; fibre-optic vs digital etc.).

Fig. 1 Literature search and study selection process according to the PRISMA statement criteria.

The included articles evaluated different parameters of flexible ureterorenoscopes, which can be divided into five key parameters: 'Manoeuvrability', 'Optics', 'Irrigation', 'Handling', and 'Durability'.

Additionally, manufacturer's characteristics were specifically described in 42 papers as presented with an asterisk in Table 1. Expanded tables on type of manufacturer specification mentioned in articles are presented in Appendix S2.

Table 1 Summary of included articles.

#	Author	Year	Journal	Scope types, n	Fibre optic /digital, n	reusable/ single-use, n	Outcomes	Ex vivo	Simulation	In vivo	LOE
1	Abdelshheid <i>et al.</i> [6]*	2005	Journal of Urology	5	5/0	5/0	manoeuvrability, optics, irrigation	Yes	No	No	2B
2	Afane <i>et al.</i> [5]*	2000	Journal of Urology	4	4/0	4/0	manoeuvrability, optics, irrigation, durability, handling	Yes	No	Yes	2B
3	Al Qahtani <i>et al.</i> [7]	2020	World Journal of Urology	1	0/1	1/0	manoeuvrability, optics, durability	Yes	No	Yes	2B
4	Al Qahtani <i>et al.</i> [8]	2011	Urology Annals	1	0/1	1/0	manoeuvrability, optics, durability, handling	Yes	No	Yes	2B
5	Ames <i>et al.</i> [9]	2006	Journal of Endourology	3	3/0	3/0	manoeuvrability, irrigation	Yes	No	No	2B
6	Andonian <i>et al.</i> [47]*	2008	Journal of Endourology	1	0/1	1/0	optics	Yes	No	No	4
7	Bach <i>et al.</i> [10]*	2008	Journal of Endourology	5	5/0	5/0	manoeuvrability, irrigation	Yes	No	No	2B
8	Bader <i>et al.</i> [11]*	2010	Journal of Endourology	1	1/0	0/1	manoeuvrability, optics, irrigation, durability	Yes	No	Yes	2B
9	Baghdadi <i>et al.</i> [12]	2017	World Journal of Urology	8	3/5	7/1	manoeuvrability, irrigation, handling	Yes	No	No	2B
10	Bedke <i>et al.</i> [13]*	2013	Urolithiasis	1	1/0	1/0	manoeuvrability, irrigation, breaking strength basket	Yes	No	No	2B
11	Binbay <i>et al.</i> [14]*	2010	Journal of Endourology	2	1/1	2/0	manoeuvrability, irrigation	Yes	No	Yes	2B
12	Boylu <i>et al.</i> [15]*	2009	Journal of Urology	7	6/1	6/1	manoeuvrability, optics, irrigation	Yes	No	No	2B
13	Cho <i>et al.</i> [16]	2018	Scientific reports	1	0/1	0/1	manoeuvrability, optics, irrigation, handling, durability	Yes	No	Yes	4
14	Dale <i>et al.</i> [17]*	2017	Journal of Endourology	3	0/3	2/1	manoeuvrability, optics, irrigation	Yes	No	No	2B
15	Deininger <i>et al.</i> [18]	2018	World Journal of Urology	3	0/3	1/2	manoeuvrability, optics, irrigation	Yes	Yes	No	2B
16	Doizi <i>et al.</i> [19]*	2017	World Journal of Urology	1	0/1	0/1	manoeuvrability, optics, irrigation, handling, durability	Yes	No	Yes	2B
17	Dragos <i>et al.</i> [20]*	2017	Journal of Endourology	9	4/5	6/3	manoeuvrability	No	Yes	No	2B
18	Dragos <i>et al.</i> [21]*	2019	Translational Andrology and Urology	8	0/8	4/4	manoeuvrability, optics, irrigation	Yes	No	No	2B
19	EHusseiny <i>et al.</i> [22]*	2009	Future Medicine Ltd	1	0/1	1/0	manoeuvrability, optics, durability	Yes	No	No	2B
20	Emiliani <i>et al.</i> [23]*	2018	Central European Journal of Urology	1	0/1	0/1	manoeuvrability, optics, irrigation, handling, durability	No	No	Yes	2B
21	Emiliani <i>et al.</i> [48]*	2017	International Brazilian Journal of Urology	1	0/1	1/0	optics	No	Yes	No	2B
22	Hennessey <i>et al.</i> [24]*	2018	BJU International	1	0/1	0/1	manoeuvrability, irrigation, costs	Yes	No	No	2B
23	Inoue <i>et al.</i> [25]*	2021	World Journal of Urology	4	3/1	3/1	manoeuvrability	No	Yes	No	2B
24	Johnson and Grasso [26]*	2004	BJU International	5	NA	NA	manoeuvrability	Yes	No	Yes	2B
25	Johnston <i>et al.</i> [27]*	2018	Central European Journal of Urology	1	0/1	0/1	manoeuvrability, optics, handling, durability	Yes	No	Yes	2B
26	Kam <i>et al.</i> [28]*	2019	International Journal of Urology	3	0/3	1/2	manoeuvrability, optics, durability	Yes	No	Yes	2B

Table 1 (continued)

#	Author	Year	Journal	Scope types, n	Fibre optic /digital, n	reusable/ single-use, n	Outcomes	Ex vivo	Simulation	In vivo	LOE
27	Kim <i>et al.</i> [29]*	2018	<i>Investigative and Clinical Urology</i>	3	2/1	3/0	manoeuvrability, optics, irrigation, handling, durability	Yes	Yes	Yes	2B
28	Kruck <i>et al.</i> [49]*	2011	<i>Urological Research</i>	5	2/3	5/0	irrigation	Yes	No	No	2B
29	Legemate <i>et al.</i> [30]*	2018	<i>European Urology Focus</i>	4	2/2	4/0	manoeuvrability, optics, handling, durability	No	No	Yes	2B
30	Ludwig <i>et al.</i> [50]*	2017	<i>Journal of Endourology</i>	3	1/2	2/1	handling	No	Yes	No	2B
31	Lusch <i>et al.</i> [31]*	2013	<i>Journal of Endourology</i>	4	4/0	4/0	manoeuvrability, optics, irrigation	Yes	Yes	No	2B
32	Marchini <i>et al.</i> [32]*	2018	<i>Journal of Endourology</i>	3	1/2	1/2	manoeuvrability, optics, irrigation, handling	No	Yes	No	2B
33	Muftescu <i>et al.</i> [33]*	2010	<i>Journal of Endourology</i>	2	2/0	1/1	manoeuvrability, optics, irrigation	Yes	No	Yes	2B
34	Muftescu <i>et al.</i> [34]*	2013	<i>Journal of Urology</i>	3	1/2	3/0	manoeuvrability, optics, irrigation, handling, durability	Yes	No	Yes	2B
35	Paffen <i>et al.</i> [35]*	2008	<i>Journal of Endourology</i>	4	4/0	4/0	manoeuvrability, optics, irrigation, handling	Yes	No	No	2B
36	Poon <i>et al.</i> [36]*	1997	<i>Journal of Endourology</i>	4	4/0	4/0	manoeuvrability, handling	Yes	No	No	2B
37	Proietti <i>et al.</i> [37]*	2016	<i>Journal of Endourology</i>	1	0/1	0/1	manoeuvrability, optics	No	Yes	No	2B
38	Proietti <i>et al.</i> [51]*	2017	<i>Journal of Endourology</i>	12	4/8	11/1	handling	Yes	No	No	2B
39	Schlager <i>et al.</i> [38]*	2017	<i>Journal of Endourology</i>	2	1/1	1/1	manoeuvrability, optics, handling	No	Yes	No	2B
40	Schlager <i>et al.</i> [39]*	2020	<i>Journal of Endourology</i>	4	1/3	1/3	manoeuvrability, optics, handling	No	Yes	No	2B
41	Shvarts <i>et al.</i> [40]*	2004	<i>Journal of Endourology</i>	2	2/0	2/0	manoeuvrability	Yes	No	No	2B
42	Talso <i>et al.</i> [52]*	2018	<i>Journal of Endourology</i>	7	2/5	6/1	optics	No	Yes	No	2B
43	Tambo <i>et al.</i> [41]*	2020	<i>Journal of Endourology</i>	2	2/0	2/0	manoeuvrability, handling	No	Yes	No	2B
44	Tom <i>et al.</i> [42]*	2017	<i>Journal of Endourology</i>	5	1/4	2/3	manoeuvrability, optics, irrigation, durability	Yes	No	No	2B
45	Traxer <i>et al.</i> [43]*	2006	<i>Journal of Urology</i>	1	1/0	1/0	manoeuvrability, irrigation, handling, durability	No	No	Yes	2B
46	Villa <i>et al.</i> [44]*	2020	<i>BJU International</i>	2	1/1	2/0	manoeuvrability	No	Yes	No	2B
47	Wendt-Nordahl <i>et al.</i> [45]*	2007	<i>Journal of Endourology</i>	2	2/0	2/0	manoeuvrability, irrigation, durability	Yes	Yes	Yes	2B
48	Winship <i>et al.</i> [46]*	2019	<i>Journal of Endourology</i>	2	0/2	0/2	manoeuvrability, optics, irrigation, durability	Yes	No	No	2B

*Mention of manufacturer specifications in article. LOE, Level of Evidence.

The following scopes were investigated in the included articles: ACMI DUR-8, ACMI DUR-8 Elite, ACMI DUR-8 prototype, ACMI Invicio DUR-D, Boston Scientific Lithovue, Circon-ACMI AUR-7, Circon-ACMI AUR-9, Cook Flexor Vue HF-EH, MaxiFlex SemiFlex Scope, Neoscope NeoFlex, Olympus URF-P3, Olympus URF-P5, Olympus URF-P6, Olympus URF-V, Olympus URF-Y0016, Olympus URF-V2, PolyScope, Pusen Uscope, Storz FlexX, Storz FlexX2, Storz FlexXc (SPIES), Storz 11274AA, Storz prototype, UscopePU3022, Wolf Cobra (Vision), Wolf 7331.001, Wolf

7325.076, Wolf7325.172, Wolf7330.072, Wolf 7331.001, Wolf BOA vision, Wolf Viper, YouCare ShaoGang, You Care YC-FR-A.

Key Parameters

Manoeuvrability

This parameter was evaluated in 42 of the 48 included studies [5–46]. Table 2 [5–52] shows an overview of the studies

Table 2 Manoeuvrability.

#	Author (year)	Manoeuvrability	Deflection	Access	Flexibility
1	Abdelshehid <i>et al.</i> (2005) [6]	Yes	Degrees		
2	Afane <i>et al.</i> (2000) [5]	Yes	Degrees		
3	Al Qahtani <i>et al.</i> (2020) [7]	Yes	Degrees		Likert scale
4	Al Qahtani <i>et al.</i> (2011) [8]	Yes	Degrees		Likert scale
5	Ames <i>et al.</i> (2006) [9]	Yes	Degrees		
6	Andonian <i>et al.</i> (2008) [47]	No			
7	Bach <i>et al.</i> (2008) [10]	Yes	Degrees		
8	Bader <i>et al.</i> (2010) [11]	Yes	Degrees		
9	Baghdadi <i>et al.</i> (2017) [12]	Yes	Degrees		
10	Bedke <i>et al.</i> (2013) [13]	Yes	Degrees		
11	Binbay <i>et al.</i> (2010) [14]	Yes	Degrees		
12	Boylu <i>et al.</i> (2009) [15]	Yes	Degrees		
13	Cho <i>et al.</i> (2018) [16]	Yes	Degrees		Likert scale
14	Dale <i>et al.</i> (2017) [17]	Yes	Degrees		
15	Deininger <i>et al.</i> (2018) [18]	Yes	Degrees		
16	Doizi <i>et al.</i> (2017) [19]	Yes	Degrees		Likert scale
17	Dragos <i>et al.</i> (2017) [20]	Yes	Degrees	Dichotomous	
18	Dragos <i>et al.</i> (2019) [21]	Yes	Degrees		
19	El-Husseiny <i>et al.</i> (2010) [22]	Yes	Degrees		
20	Emiliani <i>et al.</i> (2018) [23]	Yes	Likert scale		
21	Emiliani <i>et al.</i> (2017) [48]	No			
22	Hennessey <i>et al.</i> (2018) [24]	Yes	Degrees		
23	Inoue <i>et al.</i> (2021) [25]	Yes		Degrees	
24	Johnson and Grasso (2004) [26]	Yes	Degrees	Dichotomous	
25	Johnston <i>et al.</i> (2018) [27]	Yes	Degrees		Likert scale
26	Kam <i>et al.</i> (2019) [28]	Yes			Likert scale
27	Kim <i>et al.</i> (2018) [29]	Yes	Degrees		VAS
28	Kruck <i>et al.</i> (2011) [49]	No			
29	Legemate <i>et al.</i> (2018) [30]	Yes	Likert scale		
30	Ludwig <i>et al.</i> (2017) [50]	No			
31	Lusch <i>et al.</i> (2013) [31]	Yes	Degrees		
32	Marchini <i>et al.</i> (2018) [32]	Yes	mm	Dichotomous	
33	Multescu <i>et al.</i> (2010) [33]	Yes	Degrees	Dichotomous	Likert scale
34	Multescu <i>et al.</i> (2013) [34]	Yes	Degrees		Likert scale
35	Paffen <i>et al.</i> (2008) [35]	Yes	Degrees and mm		
36	Poon <i>et al.</i> (1997) [36]	Yes	Degrees		
37	Proietti <i>et al.</i> (2016) [37]	Yes	Degrees	Dichotomous	Frequencies
38	Proietti <i>et al.</i> (2017) [51]	No			
39	Schlager <i>et al.</i> (2017) [38]	Yes		Identification calices, Stone retrieval and time required	
40	Schlager <i>et al.</i> (2020) [39]	Yes		Dichotomous	Likert scale
41	Shvarts <i>et al.</i> (2004) [40]	Yes	Degrees		
42	Taiso <i>et al.</i> (2018) [52]	No			
43	Tambo <i>et al.</i> (2020) [41]	Yes		Time required	
44	Tom <i>et al.</i> (2017) [42]	Yes	Degrees		
45	Traxer <i>et al.</i> (2006) [43]	Yes	Degrees		Likert scale
46	Villa <i>et al.</i> (2020) [44]	Yes		%	
47	Wendt-Nordahl <i>et al.</i> (2007) [45]	Yes	Degrees		
48	Winship <i>et al.</i> (2019) [46]	Yes	Degrees		

assessing Manoeuvrability in three subcategories: 'Deflection', 'Access', and 'Flexibility'.

Deflection The subcategory deflection included studies that evaluated deflection defined as the maximum upward and/or downward deflection of the distal tip of the flexible ureterorenoscope; 35 studies met this definition (Table 2).

In all, 13 of these 35 papers, performed the deflection measurements on a photocopy or picture of the ureterorenoscope [5–9,19,26,27,31,32,35,40,46]. In the other 22 papers, deflection was directly measured from the ureterorenoscope in maximal deflection.

Measurements were performed with a protractor in 14/35 papers [6,10,12,19–22,27,31,34,36,43,45,46]. Dragos *et al.* [20], Johnson and Grasso [26], Tom *et al.* [42] and Deininger *et al.* [18] respectively used a ruler, a clockwork with degrees, SolidWorks angle measuring software and an Aristo goniometer to measure deflection. There was a broad range in the number of performed measurements (range 1–198 times), as well as in the number of observers evaluating access (range 1–7 observers).

The outcome measurement was described in degrees in 33/35 papers [5–22,24,26,27,29,31,33–37,40,42,43,45,46], and in two of the 35 papers using a Likert scale [23,30]. Two of the 35 studies described the radius of the curve of a scope in maximum deflection in millimetres [32,35]. Paffen *et al.* [35] described two measurement outcomes: maximal deflection in degrees and radius of the curve in millimetres.

In all, 25 of the 35 papers assessing deflection, evaluated maximum deflection with and without instruments inserted in the scope [6,9–15,17,18,21–24,26,31–37,40,42,45,46]. Three papers only evaluated deflection without instruments in the flexible ureterorenoscopes [7,8,43]. The remaining seven papers did not specifically describe whether they measured deflection with or without instruments inserted [5,16,19,20,27,29,30].

Access A total of 11 papers described the assessment of access in an *ex vivo* or simulated setting (K-Box, cadaver kidney or artificial kidney model) (Table 2). Access was defined as the feasibility of the flexible ureterorenoscope to access calices. Furthermore, Schlager *et al.* also [38,39] evaluated the extraction of artificial calculi in two studies.

There was a broad range in the number of performed measurements (range 1–115 times), as well as in the number of observers evaluating access (range 1–25 observers).

The measurement outcome within the access subcategory was scored as the success or failure (dichotomous) to access the calyces or extract artificial stones in seven of 11 studies [8,20,26,32,33,37,39]. Whereas degrees (range of reach within

a calyx), the number or percentage of correctly identified calyces, the number of stones extracted and the time needed to do so, was used as an outcome measurement in the remaining four studies [25,38,41,44].

Flexibility A total of 12 articles evaluated the flexibility of flexible ureterorenoscopes in an *ex vivo* or simulated setting (porcine kidney, cadaver kidney or artificial kidney model) (Table 2). This subcategory was defined as the ability of the scope to perform all necessary manoeuvres or movements.

In accordance with previous subcategories, there was a broad range in the number of performed measurements (range 1–150 times), as well as in the number of observers evaluating access (range 1–6 observers).

Kim *et al.* [29] used a visual analogue scale to assess flexibility. In the study of Proietti *et al.* [37] the surgeons' preference of one scope over other available scopes was used as subjective parameter reflecting flexibility. The remaining 10 papers used Likert scales scoring flexibility, almost all with a different rating system [7,8,16,19,27,28,33,34,39,43].

Optics

Optics was studied in 31/48 articles as shown in Table 3 [5–52]. Six subcategories were assessed in the different papers: 'resolution', 'distortion', 'luminosity', 'colour/greyscale', 'view' (depth, direction, angle and field of view) and 'visibility' as assessed by the surgeon.

Resolution Resolution was defined as the potential detail of an image. Resolution was studied in 11 papers, all using the 1951 USAF pattern card in an *ex vivo* setting (Table 3). The outcome was defined as line pairs per millimetre in all studies. However, there was a variation in the distance between the flexible ureterorenoscope and the pattern card (3–50 mm), the number of repeated measurements (1–3 times) and the number of observers (2–4 observers).

Distortion Distortion, defined as the deviation from a rectilinear projection, was evaluated in seven papers using different types of distortion grid target cards and described in percentages in an *ex vivo* setting (Table 2). The distance to the test card varied between 3 and 10 mm. Measurements of distortion were performed one to three times by two to four observers.

Luminosity Six articles assessed luminosity or brightness, defined as the intensity of the light produced by the flexible ureterorenoscope. Five different types of outcomes (mV [oscilloscope], Lumen [photometer], Lux [Lux meter/Optical spectrometer], mW [power meter], and points [questionnaire]) in an *ex vivo* setting (Table 3). The intensity

Table 3 Optics.

#	Author (Year)	Optics	Resolution	Distortion	Luminosity	Colour and grey scale	View	Visibility
1	Abdelshahid <i>et al.</i> (2005) [6]	Yes	LP/mm	%	mV			Likert scale
2	Afane <i>et al.</i> (2000) [5]	Yes			Lm			Likert scale
3	Al Qahtani <i>et al.</i> (2020) [7]	Yes						NR
4	Al Qahtani <i>et al.</i> (2011) [8]	Yes						
5	Ames <i>et al.</i> (2006) [9]	No	NR					
6	Andonian <i>et al.</i> (2008) [47]	Yes					Degrees	
7	Bach <i>et al.</i> (2008) [10]	No						
8	Bader <i>et al.</i> (2010) [11]	Yes	LP/mm	%	mW		Degrees	
9	Baghdadi <i>et al.</i> (2017) [12]	No						
10	Bedke <i>et al.</i> (2013) [13]	No						
11	Binbay <i>et al.</i> (2010) [14]	No						
12	Boylu <i>et al.</i> (2009) [15]	Yes						
13	Cho <i>et al.</i> (2017) [16]	Yes						
14	Dale <i>et al.</i> (2017) [17]	Yes						
15	Deininger <i>et al.</i> (2018) [18]	Yes	LP/mm	%	Lux and questionnaire	Reproducibility Questionnaire	Degrees and mm	Questionnaire Likert scale
16	Doizi <i>et al.</i> (2017) [19]	Yes						
17	Dragos <i>et al.</i> (2017) [20]	Yes						
18	Dragos <i>et al.</i> (2019) [21]	Yes	LP/mm	%	Lux	NR	mm	Likert scale Likert scale
19	El-Husseiny <i>et al.</i> (2010) [22]	Yes						
20	Emiliani <i>et al.</i> (2018) [23]	Yes						
21	Emiliani <i>et al.</i> (2017) [48]	Yes						
22	Hennessey <i>et al.</i> (2018) [24]	No						
23	Inoue <i>et al.</i> (2021) [25]	No						
24	Johnson and Grasso (2004) [26]	No						
25	Johnston <i>et al.</i> (2018) [27]	Yes						
26	Kam <i>et al.</i> (2019) [28]	Yes						
27	Kim <i>et al.</i> (2018) [29]	Yes	LP/mm	%				Likert scale Likert scale VAS
28	Kruck <i>et al.</i> (2011) [49]	No						Likert scale
29	Legemate <i>et al.</i> (2018) [30]	Yes						
30	Ludwig <i>et al.</i> (2017) [50]	No						
31	Lusch <i>et al.</i> (2013) [31]	Yes	LP/mm			NR Likert scale	Degrees	Likert scale Likert scale
32	Marchini <i>et al.</i> (2018) [32]	Yes	LP/mm					
33	Multescu <i>et al.</i> (2010) [33]	Yes	LP/mm					
34	Multescu <i>et al.</i> (2013) [34]	Yes						
35	Paffen <i>et al.</i> (2008) [35]	Yes	LP/mm	%	Lux		Degrees	Frequencies Questionnaire Likert scale
36	Poon <i>et al.</i> (1997) [36]	No						
37	Proietti <i>et al.</i> (2016) [37]	Yes						
38	Proietti <i>et al.</i> (2017) [51]	No						
39	Schlager <i>et al.</i> (2017) [38]	Yes						
40	Schlager <i>et al.</i> (2020) [39]	Yes						
41	Shvaris <i>et al.</i> (2004) [40]	No						
42	Talao <i>et al.</i> (2018) [52]	Yes						
43	Tambo <i>et al.</i> (2020) [41]	No						
44	Tom <i>et al.</i> (2017) [42]	Yes	LP/mm	%		NR	mm	Likert scale
45	Traxer <i>et al.</i> (2006) [43]	No						
46	Villa <i>et al.</i> (2020) [44]	No						
47	Wendt-Nordahl <i>et al.</i> (2007) [45]	No						
48	Winship <i>et al.</i> (2019) [46]	Yes	LP/mm					Likert scale

of the light source while measuring luminosity varied between 50% and 100%. Except for Deininger *et al.* [18], who described using seven observers, no specific number of observers was mentioned. Four studies mentioned the number of measurements that was performed (range 1–5) [11,18,22,35].

Colour/greyscale Colour assessment was done using a Gretag Macbeth colour checker target card or Greyscale camera contrast chart in an *ex vivo* setting in six studies (Table 3). Deininger *et al.* [18] used a questionnaire to assess colour/greyscale of high-quality pictures. All studies that reported the number of measurements (two of six), only performed the measurement once [17,21]. Similar to luminosity, Deininger *et al.* [18] were the only ones describing the number of observers (seven).

View Depth, direction, angle, and field of view were studied in an *ex vivo* setting in seven studies (Table 3). A protractor was used as a measurement-tool in two studies [11,32]. Two different test cards (a Multifrequency grid target card or the Edmund optics depth of field test target card) were used as a tool to perform measurements in three of the seven studies [17,21,42].

Outcomes were described in degrees or millimetres. None of the included studies described the number of observers. Four studies described the number of measurements performed (range 1–20 times) [11,17,21,35].

Visibility Visibility of the flexible ureterorenoscopes was assessed in 20 papers (Table 3). This subcategory was defined as the quality of the image as perceived by the surgeon.

A Likert scale was used most of the studies (14/20). A visual analogue scale, a questionnaire, and the frequency of which a scope was preferred over other scopes were used as an alternative in four studies as presented in Table 3.

Two studies did not mention which tool was used to evaluate visibility [8,16]. The Likert scales used differed between studies and all had different rating systems. Assessment of visibility was performed in an *in vivo*, simulated and *ex vivo* setting. Great diversity was observed in the number of observers (range 1–103) and the number of assessments performed (range 1–90 times).

Irrigation

Irrigation was studied in 27/48 articles. It was defined as the flow of fluid exiting the scope at the distal end. The outcome is presented in Table 4 [5–52].

Deininger *et al.* [18] also used intrapelvic pressure besides irrigation flow to describe irrigation. The majority of the studies (24/27) evaluating irrigation, performed their

measurements in an *ex vivo* setting [5,6,9–15,17,18,21,24,31–35,42,43,45,46,49].

Pressure during measurements of irrigation flow differed substantially (range 40–339.9 cmH₂O). Measuring irrigation flow whilst applying a pressure of 100 cmH₂O was mentioned most frequently (12/27) [9,10,13,14,17,18,24,32,33,42,43,49]. However, four articles did not describe the pressure used while measuring irrigation flow [16,19,23,29]. The unit used to describe pressure while measuring irrigation also differed between studies; 18 studies used cmH₂O [9,10,12–14,17,18,21,24,32–35,42,43,45,46,49], five studies used mmHg as a unit for pressure [5,6,11,29,31].

In all, 23 of the 27 papers evaluated irrigation flow with and without instruments inserted in the scope [6,9–19,23,24,28,29,32–35,42,45,49]. Afane *et al.* [5] and Dragos *et al.* [21] only evaluated irrigation without inserted instruments. The remaining papers did not specifically describe whether irrigation measurements were performed with or without inserted instruments.

Overall, 10 papers measured irrigation with an undeflected and deflected ureterorenoscope [10,13,16,18,19,21,23,29,34,49]. The remaining 19 papers did not describe state of deflection while measuring irrigation.

A minority of papers (five of 27) reported flushing the scope for a certain amount of time before measurements were performed [6,13,18,46,49].

When described, irrigation flow was measured for a minimum of 30 s up to a maximum of 5 min [5,9,11–13,18,29,31,35,46,49]. The measurements were repeated between one and five times depending on the study. Furthermore, the number of observers varied between one and four.

Four studies used a Likert scale ranging from 0 to 5 to 0–10, all with different rating systems, to describe irrigation flow. All other studies (24/27) evaluating irrigation choose to use mL/min as a unit of measurement as shown in Table 4.

Handling

Handling was mentioned in 19 studies. The results for this key parameter are shown in Table 5 [5–52]. Handling was defined as the amount of gain or strain experienced by the surgeon when using the flexible ureterorenoscopes.

It was the most heterogeneous group with 26 subcategories after regrouping. Of the 19 studies describing a form of Handling, 15 defined two or more subcategories [16,19,23,27,29,30,32,34,35,38,39,41,43,50,51]. The subcategories included up to 12 items per study to define a form of Handling.

Table 4 Irrigation.

#	Author (year)	Irrigation	Flow	Intrapelvic pressure
1	Abdelshahid <i>et al.</i> (2005) [6]	Yes	mL/min	
2	Afane <i>et al.</i> (2000) [5]	Yes	mL/min	
3	Al Qahtani <i>et al.</i> (2020) [7]	No		
4	Al Qahtani <i>et al.</i> (2011) [8]	No		
5	Ames <i>et al.</i> (2006) [9]	Yes	mL/min	
6	Andonian <i>et al.</i> (2008) [47]	No		
7	Bach <i>et al.</i> (2008) [10]	Yes	mL/min	
8	Bader <i>et al.</i> (2010) [11]	Yes	mL/min	
9	Baghdadi <i>et al.</i> (2017) [12]	Yes	mL/min	
10	Bedke <i>et al.</i> (2013) [13]	Yes	mL/min	
11	Binbay <i>et al.</i> (2010) [14]	Yes	mL/min	
12	Boylu <i>et al.</i> (2009) [15]	Yes	mL/min	
13	Cho <i>et al.</i> (2018) [16]	Yes	Likert scale	
14	Dale <i>et al.</i> (2017) [17]	Yes	mL/min	
15	Deininger <i>et al.</i> (2018) [18]	Yes	mL/min	mmHg
16	Doizi <i>et al.</i> (2017) [19]	Yes	Likert scale	
17	Dragos <i>et al.</i> (2017) [20]	No		
18	Dragos <i>et al.</i> (2019) [21]	Yes	mL/min	
19	El-Husseiny <i>et al.</i> (2010) [22]	No		
20	Emiliani <i>et al.</i> (2018) [23]	Yes	Likert scale	
21	Emiliani <i>et al.</i> (2017) [48]	No		
22	Hennessey <i>et al.</i> (2018) [24]	Yes	mL/min	
23	Inoue <i>et al.</i> (2021) [25]	No		
24	Johnson and Grasso (2004) [26]	No		
25	Johnston <i>et al.</i> (2018) [27]	No		
26	Kam <i>et al.</i> (2019) [28]	No		
27	Kim <i>et al.</i> (2018) [29]	Yes	mL/min & likert scale	
28	Kruck <i>et al.</i> (2011) [49]	Yes	mL/min	
29	Legemate <i>et al.</i> (2018) [30]	No		
30	Ludwig <i>et al.</i> (2017) [50]	No		
31	Lusch <i>et al.</i> (2013) [31]	Yes	mL/min	
32	Marchini <i>et al.</i> (2018) [32]	Yes	mL/min	
33	Mulfescu <i>et al.</i> (2010) [33]	Yes	mL/min	
34	Mulfescu <i>et al.</i> (2013) [34]	Yes	mL/min	
35	Paffen <i>et al.</i> (2008) [35]	Yes	mL/min	
36	Poon <i>et al.</i> (1997) [36]	No		
37	Proietti <i>et al.</i> (2016) [37]	No		
38	Proietti <i>et al.</i> (2017) [51]	No		
39	Schlager <i>et al.</i> (2017) [38]	No		
40	Schlager <i>et al.</i> (2020) [39]	No		
41	Shvarts <i>et al.</i> (2004) [40]	No		
42	Talso <i>et al.</i> (2018) [52]	No		
43	Tambo <i>et al.</i> (2020) [41]	No		
44	Tom <i>et al.</i> (2017) [42]	Yes	mL/min	
45	Traxer <i>et al.</i> (2006) [43]	Yes	mL/min	
46	Villa <i>et al.</i> (2020) [44]	No		
47	Wendt-Nordahl <i>et al.</i> (2007) [45]	Yes	mL/min	
48	Winship <i>et al.</i> (2019) [46]	Yes	mL/min	

After categorisation, the following groups were formed: 'Difficulty', 'Control', 'Ergonomics', and 'Satisfaction'. The vast majority (14/19) of all measurements were evaluated on a Likert scale or visual analogue scale [5,8,12,16,19,23,27,29,30,32,34,38,39,43]. Nonetheless, 17 different scales were used to estimate Handling.

Difficulty Difficulty was evaluated in 12 out of 19 papers describing Handling (Table 5). Difficulty was defined as the ease of insertion and handling of a flexible ureterorenoscope. This subcategory included the ease of insertion of a scope or

instruments, ease of handling, rigidity of the flexible ureterorenoscopes, torsion stiffness of the flexible ureterorenoscopes, the workload on the surgeon to perform a task, and the difficulty of the procedure as experienced by the surgeon.

The ease of insertion was evaluated for the flexible ureterorenoscope, as well as for instruments in the flexible ureterorenoscope. All these measurements were assessed on different Likert scales or a visual analogue scale. Baghdadi *et al.* [12] additionally defined ease of insertion as a dichotomous variable (failure vs success). Attempts were

Table 5 Handling.

#	Author (year)	Handling	Difficulty	Control	Ergonomics	Satisfaction
1	Abdelshehid <i>et al.</i> (2005) [6]	No				
2	Afane <i>et al.</i> (2000) [5]	Yes		Likert scale		
3	Al Qahtani <i>et al.</i> (2020) [7]	No				
4	Al Qahtani <i>et al.</i> (2011) [8]	Yes	Likert scale			
5	Ames <i>et al.</i> (2006) [9]	No				
6	Andonian <i>et al.</i> (2008) [47]	No				
7	Bach <i>et al.</i> (2008) [10]	No				
8	Bader <i>et al.</i> (2010) [11]	No				
9	Baghdadi <i>et al.</i> (2017) [12]	Yes	Dichotomous			
10	Bedke <i>et al.</i> (2013) [13]	No				
11	Binbay <i>et al.</i> (2010) [14]	No				
12	Boylu <i>et al.</i> (2009) [15]	No				
13	Cho <i>et al.</i> (2018) [16]	Yes		Likert scale	Likert scale	
14	Dale <i>et al.</i> (2017) [17]	No				
15	Deininger <i>et al.</i> (2018) [18]	No				
16	Doizi <i>et al.</i> (2017) [19]	Yes	Likert scale	Likert scale		Likert scale
17	Dragos <i>et al.</i> (2017) [20]	No				
18	Dragos <i>et al.</i> (2019) [21]	No				
19	El-Husseiny <i>et al.</i> (2010) [22]	No				
20	Emiliani <i>et al.</i> (2018) [23]	Yes	Likert scale	Likert scale		Likert scale
21	Emiliani <i>et al.</i> (2017) [48]	No				
22	Hennessey <i>et al.</i> (2018) [24]	No				
23	Inoue <i>et al.</i> (2021) [25]	No				
24	Johnson and Grasso (2004) [26]	No				
25	Johnston <i>et al.</i> (2018) [27]	Yes	Likert scale		Likert scale	Likert scale
26	Kam <i>et al.</i> (2019) [28]	No				
27	Kim <i>et al.</i> (2018) [29]	Yes	VAS	VAS		Likert scale
28	Kruck <i>et al.</i> (2011) [49]	No				
29	Legemate <i>et al.</i> (2018) [30]	Yes	Likert scale			
30	Ludwig <i>et al.</i> (2017) [50]	Yes			Average and cumulative muscular workload	
31	Lusch <i>et al.</i> (2013) [31]	No				
32	Marchini <i>et al.</i> (2018) [32]	Yes	Likert scale			
33	Multescu <i>et al.</i> (2010) [33]	No				
34	Multescu <i>et al.</i> (2013) [34]	Yes	Likert scale	Likert scale		
35	Paffen <i>et al.</i> (2008) [35]	Yes				
36	Poon <i>et al.</i> (1997) [36]	Yes	Torsion stiffness meter and dichotomous			
37	Proietti <i>et al.</i> (2016) [37]	No				
38	Proietti <i>et al.</i> (2017) [51]	Yes			Grammes and cm	
39	Schlager <i>et al.</i> (2017) [38]	Yes		Likert scale		
40	Schlager <i>et al.</i> (2020) [39]	Yes	NASA TLX	Likert scale		
41	Shvarts <i>et al.</i> (2004) [40]	No				
42	Talso <i>et al.</i> (2018) [52]	No				
43	Tambo <i>et al.</i> (2020) [41]	Yes			Questionnaire	
44	Tom <i>et al.</i> (2017) [42]	No				
45	Traxer <i>et al.</i> (2006) [43]	Yes	Likert scale	Likert scale		
46	Villa <i>et al.</i> (2020) [44]	No				
47	Wendt-Nordahl <i>et al.</i> (2007) [45]	No				
48	Winship <i>et al.</i> (2019) [46]	No				

made to pass fibres at different states of deflection and irrigation. Kim *et al.* [29] assessed the rigidity of the flexible ureterorenoscope on a visual analogue scale. Legemate *et al.* [30] evaluated the ease of handling during URS on a Likert scale. Torsion stiffness was assessed by Paffen *et al.* [35] and defined as the resistance of the shaft to rotate at its longitudinal axis. It was measured with a torsion stiffness meter and expressed in Newton meter. Traxer *et al.* [43] assessed the difficulty of the procedure on a Likert scale. Finally, Schlager *et al.* [39] evaluated the workload during a

URS with the National Aeronautics and Space Administration Task Load Index (NASA-TLX).

Difficulty was assessed by one, two or three observers, depending on the study. Only Paffen *et al.* [35] assessed difficulty multiple times (eight). The other 10 studies performed one assessment.

Control This subcategory was assessed in nine of 19 papers evaluating this key parameter as found in Table 5. Control

was defined as the influence of the surgeon on the manoeuvrability of the flexible ureterorenoscope. However, all the studies used a slightly different definition. Emiliani *et al.* [23] for instance, mentioned torque: the degree to which movement in the handle was transmitted and precisely reproduced at the tip of the scope. As opposed to Kim *et al.* [29] who used the word control or Traxer *et al.* [43] who used manoeuvrability of the scope.

Furthermore, all studies used a different Likert scale or visual analogue scale to express control, leading to a great heterogeneity. There was no difference in the number of assessments, as all 19 studies only performed one evaluation of control. The number of observers, ranged from one to six observers.

Ergonomics Ergonomics was studied in five of 19 studies (Table 5). Ergonomics was defined as the physical comfort and efficiency in use of a flexible ureterorenoscope. It mostly concerned the extent to which a muscle group or body part was used, strained, or in pain [16,27,41,50]. One study focussed on the body mass index of the scope as an influencing factor on ergonomics [51]. All studies used different tools such as a Likert scale, a questionnaire or surface electromyography (EMG) electrodes to evaluate ergonomics. Likert scales and scales in questionnaires varied from 0-5, 1-5 to 1-10 all with different rating systems. Ludwig *et al.* [50] used surface EMG electrodes and expressed the outcome in average and cumulative workload.

Ludwig *et al.* [50] evaluated ergonomics three times. All other studies conducted measurements only once. With the exception of Tambo *et al.* [41] who asked 25 observers to assess ergonomics, all papers used one observer.

Satisfaction Satisfaction was described in four papers and was defined as overall satisfaction of the surgeon or convenience of use of the flexible ureterorenoscopes as assessed by the surgeon during a procedure as presented in Table 5. All four papers used a variation of a Likert scale to evaluate satisfaction.

Only Emiliani *et al.* [23] had more than one observer, namely two, to evaluate satisfaction. Furthermore, all four papers only performed one evaluation to assess overall satisfaction with the flexible ureterorenoscope.

Durability

Durability was defined as the property of the scope to withstand wear, pressure, or damage. Seventeen studies described the durability of flexible ureterorenoscopes, as presented in Table 6 [5–52]. All measurements were performed in an *ex vivo* setting. Emiliani *et al.* [23] and

Legemate *et al.* [30] evaluated loss of deflection, visibility and image quality in an *in vivo* setting as well.

The interval of measurements varied substantially between studies. Wendt-Nordahl *et al.* [45] evaluated ureterorenoscopes in an *ex vivo* setting after the first and 100th procedure in a pig cadaver. Whereas Winship *et al.* [46] performed the evaluation on never-used single-use ureterorenoscopes and then repeated the same measurements after 200 cycles of maximum deflection. El-Husseiny *et al.* [22] evaluated durability after the first, tenth, 30th and 44th procedure. The remainder of studies evaluated durability before and after each procedure or did not report on the timing of evaluation.

Wear A total of 16 studies presented wear of functionality over a certain amount of time (Table 6). Wear was defined as the loss of quality of a specific characteristic (deflection, irrigation, luminosity, image quality, or visibility) over a period of time.

Of these 16 studies, 14 studies within the wear subcategory, focussed on loss of deflection [5,7,8,19,22,23,27,29,30,34,42,43,45,46]. Loss of deflection was mostly (13/14 papers) expressed in degrees (Table 6). Emiliani *et al.* [23] and Legemate *et al.* [30] expressed loss of deflection in the form of a Likert scale.

Three papers studied the loss of irrigation expressed in mL/min as can be seen in Table 6. Afane *et al.* [5] and El-Husseiny *et al.* [22] assessed loss of luminosity measured by a photo- or spectrometer and expressed in lumen or lux. Three studies evaluated loss of image quality or visibility measured with a Likert scale (Table 6).

Except for Tom *et al.* [42], who performed the measurement three times, all other studies performed the measurement once [5,7,8,11,19,22,23,27–30,34,42,43,45,46]. The number of observers varied between one and five observers.

Damage Seven studies presented damage within the durability group (Table 6). Damage was defined as the breakdown of a characteristic of the flexible ureterorenoscope. The main topic of interest within the damage subcategory was broken optical fibres defined as the number of dark dots on the monitor. This type of damage was studied in four studies (Table 6). Leakage, damage to sheath and working channel and damage to the scope in general, were characteristics studied to evaluate damage in the remaining three papers. These characteristics were all scored as dichotomous variables.

Measurements were only performed once in all seven studies. Furthermore, six of the seven studies only used one observer.

Table 6 Durability.

#	Author (year)	Durability	Wear (loss of *)	Damage
1	Abdelshahid <i>et al.</i> (2005) [6]	No		
2	Afane <i>et al.</i> (2000) [5]	Yes	Degrees, mL/min and Lm	# dark dots on monitor
3	Al Qahtani <i>et al.</i> (2020) [7]	Yes	Degrees	Dichotomous
4	Al Qahtani <i>et al.</i> (2011) [8]	Yes	Degrees	Dichotomous
5	Ames <i>et al.</i> (2006) [9]	No		
6	Andonian <i>et al.</i> (2008) [47]	No		
7	Bach <i>et al.</i> (2008) [10]	No		
8	Bader <i>et al.</i> (2010) [11]	Yes	mL/min	
9	Baghdadi <i>et al.</i> (2017) [12]	No		
10	Bedke <i>et al.</i> (2013) [13]	No		
11	Binbay <i>et al.</i> (2010) [14]	No		
12	Boylu <i>et al.</i> (2009) [15]	No		
13	Cho <i>et al.</i> (2018) [16]	Yes		Dichotomous
14	Dale <i>et al.</i> (2017) [17]	No		
15	Deiningner <i>et al.</i> (2018) [18]	No		
16	Doizi <i>et al.</i> (2017) [19]	Yes	Degrees and Likert scale	
17	Dragos <i>et al.</i> (2017) [20]	No		
18	Dragos <i>et al.</i> (2019) [21]	No		
19	El-Husseiny <i>et al.</i> (2010) [22]	Yes	Degrees and Lux	
20	Emiliani <i>et al.</i> (2018) [23]	Yes	Likert scale	
21	Emiliani <i>et al.</i> (2017) [48]	No		
22	Hennessey <i>et al.</i> (2018) [24]	No		
23	Inoue <i>et al.</i> (2021) [25]	No		
24	Johnson and Grasso (2004) [26]	No		
25	Johnston <i>et al.</i> (2018) [27]	Yes	Degrees	
26	Kam <i>et al.</i> (2019) [28]	Yes	Dichotomous	
27	Kim <i>et al.</i> (2018) [29]	Yes	Degrees	
28	Kruck <i>et al.</i> (2011) [49]	No		
29	Legemate <i>et al.</i> (2018) [30]	Yes	Degrees and Likert scale	
30	Ludwig <i>et al.</i> (2017) [50]	No		
31	Lusch <i>et al.</i> (2013) [31]	No		
32	Marchini <i>et al.</i> (2018) [32]	No		
33	Multescu <i>et al.</i> (2010) [33]	No		
34	Multescu <i>et al.</i> (2013) [34]	Yes	Degrees	# dark dots on monitor
35	Paffen <i>et al.</i> (2008) [35]	No		
36	Poon <i>et al.</i> (1997) [36]	No		
37	Proietti <i>et al.</i> (2016) [37]	No		
38	Proietti <i>et al.</i> (2017) [51]	No		
39	Schlager <i>et al.</i> (2017) [38]	No		
40	Schlager <i>et al.</i> (2020) [39]	No		
41	Shvarts <i>et al.</i> (2004) [40]	No		
42	Talso <i>et al.</i> (2018) [52]	No		
43	Tambo <i>et al.</i> (2020) [41]	No		
44	Tom <i>et al.</i> (2017) [42]	Yes	Degrees	
45	Traxer <i>et al.</i> (2006) [43]	Yes	Degrees and mL/min	# dark dots on monitor
46	Villa <i>et al.</i> (2020) [44]	No		
47	Wendt-Nordahl <i>et al.</i> (2007) [45]	Yes	Degrees	# dark dots on monitor
48	Winship <i>et al.</i> (2019) [46]	Yes	Degrees	

Only Afane *et al.* [5] used four different observers to assess damage.

Manufacturer's Specifications

A wide range of manufacturer's specifications was described in 42 of the 48 included papers (Table 1). There was a wide variety of specifications described in the different studies. Finally, a total of 29 different characteristics were mentioned in the included papers. A full overview of the specifications can be found in Appendix S2.

Discussion

The evolution in flexible ureteroscopes have enabled urologists to expand the indication for endoscopic retrograde procedures. However, the instruments are fragile, and the purchase and repairs comes at high costs. Therefore, there is a great interest in evaluation and comparison of the quality of flexible ureterorenoscopes.

The present review systematically evaluated different approaches to assess the quality of flexible

ureterorenoscopes. To the best of our knowledge, the present review is the first on this topic. The systematic literature search resulted in 48 articles [5–52]. Five key parameters were identified: Manoeuvrability (87.5%), Optics (64.6%), Irrigation (56.3%), Handling (39.6%) and Durability (35.4%). These parameters were partially described in previous systematic reviews, which focussed on comparing clinical results of different flexible ureterorenoscopes [54–57]. Within these parameters there is a great heterogeneity in terms of method of measurement, construct, definitions, and measurement outcomes. This variety makes it difficult to cluster or compare study results evaluating flexible ureterorenoscopes.

One explanation for this variety could be the range of year of publication of included articles (1997–2020) and the concomitant technological advances during this time frame. Although the importance of the evaluation of flexible ureterorenoscopes is illustrated by the number of included studies in the present review, uniformity in method of appraisal has yet to be reached.

Grading Systems in Alternate Fields of Expertise

Originally, flexible ureterorenoscopes were derived from the first flexible endoscopic instrument used in the field of gastroenterology, the gastroscope. The gastroscope was developed in the 1950s, when a physics student and a gastroenterologist joined forces to develop this instrument [2]. The latter performed the first gastroscopy on himself and treated a stomach ulcer in a patient the next day. Although the gastroscope can be seen as a precursor of the first flexible URS (which had its first purposeful use in a patient in the 1970s), the gastroscope does not seem to be assessed in a uniform way as well.

Yet, in other fields such as physics and engineering, research is done to create uniform grading systems, for instance to compare images [58,59]. Moreover, institutions such as NASA, have found ways to systematically evaluate workload of certain tasks on a person in relation to the use of a tool (NASA-TLX) [60]. Finally, in the field of laparoscopy, validated questionnaires to estimate ergonomics are more common. As well as recommendations on improving ergonomics when performing laparoscopy, with the first articles on this subject already widely published in the 1990s [61]. As developments are made in different fields of expertise within medicine, as well as outside the working field of medicine, it is time to join forces and combine knowledge to create a reproducible and uniform assessment tool. A standardised assessment tool to evaluate the quality of flexible ureterorenoscopes is needed to structure outcomes and facilitate comparison in future studies. As the validity of a measurement can only accurately be graded if the characteristics measured and their applications are clearly defined [62].

Future Perspectives

As we deem the heterogeneity of all aspects of analysed grading systems to be of too great an extent, we feel that it is not fitted to select preferred subcategories and methods based solely on our expertise. However, we do believe that with the present systematic scoping review, we identified main categories that should be used as a basis when creating a uniform assessment tool, namely: Manoeuvrability, Optics, Irrigation, Handling and Durability. Therefore, we will use these main categories to explore subcategories, measurement methods, construct, and measurement outcomes to establish consensus by means of a Delphi Consensus Project with active members in the research field of evaluation of ureterorenoscopes. In our opinion this is the only appropriate step after establishing such a vast non-uniformity. Finally, we would not like to venture possible influence on the results of such a project and therefor limit expressions of opinions on preferred methods.

Limitations

The major limitation of the present review is the possible publication bias, which is inherent to all non-randomised controlled cohort studies. However, we do believe, that even if there was a possible bias, this would have had an influence on the outcomes reported and not on the characteristics investigated. Thus, making it irrelevant for the present review. Furthermore, regrouping items in order to form categories is prone to be influenced by interpretation of the reviewers.

Conclusion

In the present systematic scoping review, we reviewed and identified parameters and methods used to evaluate flexible ureterorenoscopes. This review showed that there are five key parameters in the quality assessment of flexible ureterorenoscopes: Manoeuvrability, Optics, Irrigation, Handling, and Durability. Nevertheless, within the different studies there is great heterogeneity in terms of measurement methods, construct, definitions, and measurement outcomes, which complicates comparison of study outcomes. Therefore, a standardised assessment tool to evaluate the quality of flexible ureterorenoscopes is mandatory to structure outcomes and facilitate comparison in future studies.

Acknowledgements

None.

Conflict of Interest

G.M. Kamphuis received payment for the EMS Webinar on Endoscopic Combined Intrarenal Surgery, an online course for Boston scientific on decision making in Endourology and

an online course for Olympus on the Thulium Fiber Laser usage in treatment of kidney stones. Furthermore he attends advisory board meetings for Olympus and Boston Scientific.

Author Contributions

Both Nora Hendriks and Michaël M.E.L. Henderickx contributed equally and share first authorship. N. Hendriks: protocol/project development, data collection or management, data analysis, manuscript writing/editing. M.M.E.L. Henderickx: protocol/project development, data collection or management, data analysis, manuscript writing/editing. B.M.A. Schout: Protocol/project development, manuscript writing/editing. J. Baard: manuscript writing/editing. F.S. van Etten-Jamaludin: data collection or management, manuscript writing/editing. H.P. Beerlage: manuscript writing/editing. R.C.M. Pelger: manuscript writing/editing. G.M. Kamphuis: protocol/project development, data collection or management, manuscript writing/editing.

Funding

None.

References

- Meldrum KK, Mathews R, Gearhart JP. Hugh Hampton Young: A pioneer in pediatric urology. *J Urol* 2001; 1415–7
- Marshall VF. Fiber optics in urology. *J Urol* 1964; 91: 110–4
- Georgescu D, Multescu R, Mirciulescu V, Geavlete PA, Geavlete B. Chapter 3, Instruments. In: Geavlete PA, ed., *Retrograde Ureteroscopy*. Cambridge, MA: Academic Press, 2016: p. 21–52
- Gridley CM, Knudsen BE. Digital ureteroscopes: Technology update. *Res Reports Urol* 2017; 9: 19–25
- Afane JS, Olweny EO, Bercowsky E et al. Flexible ureteroscopes: A single center evaluation of the durability and function of the new endoscopes smaller than 9Fr. *J Urol* 2000; 1164–8
- Abdelshahid C, Ahlering MT, Chou D et al. Comparison of flexible ureteroscopes: Deflection, irrigant flow and optical characteristics. *J Urol* 2005; 173: 2017–21
- Al Qahtani SH, Abdelhamied MH, AlMuhrij AH et al. Prospective comparative study between the effect of CIDEX[®] OPA and STERRAD NX on the durability of digital flexible ureteroscope. *World J Urol* 2020; 38: 775–81
- Al-Qahtani SM, Geavlete B, Geavlette BP, Gil-Diez S, de Medina SG, Traxer OP. The new Olympus digital flexible ureteroscope (URF-V): Initial experience. *Urol Ann* 2011; 3: 133–7
- Ames CD, Perrone JM, Weld KJ et al. Alteration in irrigant flow and deflection of flexible ureteroscopes with Nitinol baskets. *J Endourol* 2006; 20: 74–7
- Bach T, Geavlete B, Herrmann TR, Gross AJ. Working tools in flexible ureterorenoscopy - Influence on flow and deflection: What does matter? *J Endourol* 2008; 22: 1639–44
- Bader MJ, Gratzke C, Walther S et al. The polyScope: A modular design, semidisposable flexible ureterorenoscopy system. *J Endourol* 2010; 24: 1061–6
- Baghdadi M, Emiliani E, Talso M et al. Comparison of laser fiber passage in ureteroscopic maximum deflection and their influence on deflection and irrigation: Do we really need the ball tip concept? *World J Urol* 2017; 35: 313–8
- Bedke J, Leichtle U, Lorenz A, Nagele U, Stenzl A, Kruck S. 1.2 French stone retrieval baskets further enhance irrigation flow in flexible ureterorenoscopy. *Urol Res* 2013; 41: 153–7
- Binbay M, Yuruk E, Akman T et al. Is there a difference in outcomes between digital and fiberoptic flexible ureterorenoscopy procedures? *J Endourol* 2010; 24: 1929–34
- Boylu U, Oommen M, Thomas R, Lee BR. In Vitro Comparison of a Disposable Flexible Ureteroscope and Conventional Flexible Ureteroscopes. *J Urol* 2009; 182: 2347–51
- Cho SY, Lee JY, Shin DG, Seo IY, Yoo S, Park HK. Evaluation of Performance Parameters of the Disposable Flexible Ureterorenoscope (LITHOVUE) in Patients with Renal Stones: A Prospective, Observational, Single-arm, Multicenter Study. *Sci Rep* 2018; 8: 9795
- Dale J, Kaplan AG, Radvak D et al. Evaluation of a novel single-use flexible ureteroscope. *J Endourol* 2017; 35: 903–7
- Deininger S, Haberstock L, Kruck S et al. Single-use versus reusable ureterorenoscopes for retrograde intrarenal surgery (RIRS): systematic comparative analysis of physical and optical properties in three different devices. *World J Urol* 2018; 36(12): 2059–63
- Doizi S, Kamphuis G, Giusti G et al. First clinical evaluation of a new single-use flexible ureteroscope (LithoVueTM): a European prospective multicentric feasibility study. *World J Urol* 2017; 35: 809–18
- Dragos LB, Somani BK, Sener ET et al. Which flexible ureteroscopes (digital vs. fiber-optic) can easily reach the difficult lower pole calices and have better end-tip deflection. in vitro study on K-Box. A PETRA evaluation. *J Endourol* 2017; 31: 630–7
- Dragos LB, Somani BK, Keller EX et al. Characteristics of current digital single-use flexible ureteroscopes versus their reusable counterparts: An in-vitro comparative analysis. *Transl Androl Urol* 2019; 8(S4): S359–70
- El-Husseiny T, Masood J, Karamanolakis D, Birch M, Buchholz N. The new Invisio[®] DUR-D 'chip-on-the-tip' digital flexible ureterorenoscopes: Evaluation of changes in physical properties over time. *Therapy* 2010; 7: 423–7
- Emiliani E, Mercadé A, Millan F, Sánchez-Martín F, Konstantinidis CA, Angerri O. First clinical evaluation of the new single-use flexible and semirigid PUSEN ureteroscopes. *J Cent Eur J Urol* 2018; 71: 208–13
- Hennessey DB, Fojecki GL, Papa NP, Lawrentschuk N, Bolton D. Single-use disposable digital flexible ureteroscopes: an ex vivo assessment and cost analysis. *BJU Int* 2018; 121 (Suppl. 3): 55–61
- Inoue T, Okada S, Hamamoto S et al. Evaluation of flexible ureteroscope with an omni-directional bending tip, using a JOYSTICK unit (URF-Y0016): an ex-vivo study. *World J Urol* 2021; 39: 209–15
- Johnson GB, Grasso M. Exaggerated primary endoscope deflection: Initial clinical experience with prototype flexible ureteroscopes. *BJU Int* 2004; 93: 109–14
- Johnston TJ, Baard J, de la Rosette J et al. A clinical evaluation of the new digital single-use flexible ureteroscope (UscopePU3022): An international prospective multicentered study. *Cent Eur J Urol* 2018; 71: 453–61
- Kam J, Yuminaga Y, Beattie K et al. Single use versus reusable digital flexible ureteroscopes: A prospective comparative study. *Int J Urol* 2019; 26: 999–1005
- Kim DH, Shin JH, Choi SY, You D, Kim CS, Park HK. In vitro, in vivo, and clinical tests of a novel flexible ureteroscope for the diagnosis and treatment of kidney and ureteral diseases. *Investig Clin Urol* 2018; 59: 328–34
- Legemate JD, Kamphuis GM, Freund JE et al. Durability of flexible ureteroscopes: a prospective evaluation of longevity, the factors that affect it, and damage mechanisms. *Eur Urol Focus* 2019; 5: 1105–11
- Lusch A, Abdelshahid C, Hidas G et al. In vitro and in vivo comparison of optics and performance of a distal sensor ureteroscope versus a standard fiberoptic ureteroscope. *J Endourol* 2013; 27: 896–902

- 32 Marchini GS, Batagello CA, Monga M et al. In vitro evaluation of single-use digital flexible ureteroscopes: A practical comparison for a patient-centered approach. *J Endourol* 2018; 32: 184–91
- 33 Multescu R, Geavlete B, Georgescu D, Geavlete P. Conventional fiberoptic flexible ureteroscope versus fourth generation digital flexible ureteroscope: A critical comparison. *J Endourol* 2010; 24: 17–21
- 34 Multescu R, Geavlete B, Geavlete P. A new era: Performance and limitations of the latest models of flexible ureteroscopes. *Urology* 2013; 82: 1236–9
- 35 Paffen ML, Keizer JG, De Winter GV, Arends AJ, Hendriks AJ. A comparison of the physical properties of four new generation flexible ureteroscopes: (de)flexion, flow properties, torsion stiffness, and optical characteristics. *J Endourol* 2008; 22: 2227–34
- 36 Poon M, Beagler M, Baldwin D. Flexible endoscope deflectability: Changes using a variety of working instruments and laser fibers. *J Endourol* 1997
- 37 Proietti S, Dragos L, Molina W, Doizi S, Giusti G, Traxer O. Comparison of New Single-Use Digital Flexible Ureteroscope Versus Nondisposable Fiber Optic and Digital Ureteroscope in a Cadaveric Model. *J Endourol* 2016; 30: 655–9
- 38 Schlager D, Hein S, Obaid MA, Wilhelm K, Miernik A, Schoenthaler M. Performance of Single-Use FlexorVue vs Reusable BoaVision Ureteroscope for Visualization of Calices and Stone Extraction in an Artificial Kidney Model. *J Endourol* 2017; 31: 1139–44
- 39 Schlager D, Obaid MA, Hein S et al. Current disposable ureteroscopes: Performance and limitations in a standardized kidney model. *J Endourol* 2020; 34: 1015–20
- 40 Shvarts O, Perry KT, Goff B, Schulam PG. Improved functional deflection with a dual-deflection flexible ureteroscope. *J Endourol* 2004; 18: 141–4
- 41 Tambo M, Inoue T, Miura H et al. A novel flexible ureteroscope with omnidirectional bending tip using joystick-type control unit (URF-Y0016): initial validation study in bench models. *J Endourol* 2020; 34: 676–81
- 42 Tom WR, Wollin DA, Jiang R et al. Next-Generation Single-Use Ureteroscopes: An in Vitro Comparison. *J Endourol* 2017
- 43 Traxer O, Dubosq F, Jamali K, Gattegno B, Thibault P. New-generation flexible ureterorenoscopes are more durable than previous ones. *Urology* 2006; 68: 276–9
- 44 Villa L, Ventimiglia E, Proietti S et al. Does working channel position influence the effectiveness of flexible ureteroscopy? Results from an in vitro study. *BJU Int* 2020; 125: 449–56
- 45 Wendt-Nordahl G, Trojan L, Alken P, Michel MS, Knoll T. Ureteroscopy for stone treatment using new 270° semiflexible endoscope: In vitro, ex vivo, and clinical application. *J Endourol* 2007; 21: 1439–44
- 46 Winship B, Wollin D, Carlos E, Li J, Preminger GM, Lipkin ME. Avoiding a lemon: Performance consistency of single-use ureteroscopes. *J Endourol* 2019; 33: 127–31
- 47 Andonian S, Okeke Z, Smith AD. Digital ureteroscopy: The next step. *J Endourol* 2008; 22: 603–6
- 48 Emiliani E, Talso M, Baghdadi M et al. Evaluation of the Spies™ modalities image quality. *Int Braz J Urol* 2017; 43: 476–80
- 49 Kruck S, Anastasiadis AG, Gakis G et al. Flow matters: Irrigation flow differs in flexible ureteroscopes of the newest generation. *Urol Res* 2011; 39: 483–6
- 50 Ludwig WW, Lee G, Ziembra JB, Ko JS, Matlaga BR. Evaluating the Ergonomics of Flexible Ureteroscopy. *J Endourol* 2017; 31: 1062–6
- 51 Proietti S, Somani B, Sofer M et al. The “body mass index” of flexible ureteroscopes. *J Endourol* 2017; 31: 1090–5
- 52 Talso M, Proietti S, Emiliani E et al. Comparison of Flexible Ureterorenoscope Quality of Vision: An in Vitro Study. *J Endourol* 2018; 17: e1800
- 53 Liberati A, Altman DG, Tetzlaff J et al. The PRISMA statement for reporting systematic reviews and meta-analyses of studies that evaluate healthcare interventions: explanation and elaboration. *BMJ* 2009; 339: b2700
- 54 Scotland KB, Chan JYH, Chew BH. Single-use flexible ureteroscopes: How do they compare with reusable ureteroscopes? *J Endourol* 2019; 33: 71–8
- 55 Traxer O. Flexible ureterorenoscopic management of lower-pole stone: does the scope make the difference? *J Endourol* 2008; 22: 1847–50
- 56 Moore B, Proietti S, Giusti G, Eisner BH. Single-Use Ureteroscopes. *Urol Clin North Am* 2019; 46(2): 165–74
- 57 Şener TE, Pietropaolo A, Talso M, Somani B, Tanidir Y, “Endourology and Urolithiasis” Working Party YAU (YAU). An Overview of the Advantages of Digital Flexible Ureteroscopes. A Review by Young Academic Urologists Endourology and Urolithiasis Working Party of the European Association of Urology. *J Urol Surg* 2020; 7: 76–82
- 58 Wang Z, Bovik AC. A universal image quality index. *IEEE Signal Process Lett* 2002; 9: 81–4
- 59 Wang Z, Bovik AC. Modern image quality assessment. *Synth Lect Image Video Multimed Process* 2006; 2: 1–156
- 60 Hart SG. NASA-Task Load Index (NASA-TLX); 20 Years Later. *Proc Hum Factors Ergon Soc Annu Meet.* 2006; 50: 904–8
- 61 van Veelen MA, Jakimowicz JJ, Kazemier G. Improved physical ergonomics of palarscopic surgery. *Minim Invasive Ther Allied Technol* 2004; 13: 161–6
- 62 Johnston MV, Graves D, Greene M. The uniform Postacute assessment tool: systematically evaluating the quality of measurement evidence. *Arch Phys Med Rehab* 2007; 88: 1505–12

Correspondence: Nora Hendriks, Department of Urology, University of Amsterdam, Amsterdam UMC, De Boelelaan 1117, NL-1081 HV Amsterdam, the Netherlands.

e-mail: n.hendriks2@amsterdamumc.nl

Abbreviations: EMG, electromyography; MeSH, Medical Subject Headings; NASA(-TLX), National Aeronautics and Space Administration (Task Load Index); PRISMA, Preferred Reporting Items for Systematic Reviews and Meta-Analyses; RCT, randomised controlled trials; URS, ureterorenoscopy.

Supporting Information

Additional Supporting Information may be found in the online version of this article:

Appendix S1. Literature searches.

Appendix S2. Manufacturer specifications 1.

Appendix S3. List of references screened for inclusion based on full text.