



Corrigendum

Corrigendum to “Ultra-trace determination of oxyhalides in ozonated aquacultural marine waters by direct injection ion chromatography coupled with triple-quadrupole mass spectrometry” [Heliyon 7 (4) (April 2021) e06885]

Estrella Sanz Rodriguez ^{a,*}, Shing Lam ^a, Gregory G. Smith ^b, Paul R. Haddad ^a, Brett Paull ^a^a Australian Centre for Research on Separation Science (ACROSS), School of Natural Sciences, University of Tasmania, Sandy Bay, Hobart, Tasmania, 7001, Australia^b Institute for Marine and Antarctic Studies (IMAS), University of Tasmania, Private Bag 49, Hobart, Tasmania, 7001, Australia

In the original published version of this article, there is a mistake in Table 3, where values for most of the parameters were omitted. This has now been corrected. The correct version of Table 3 can be found below.

The authors apologize for the errors. Both the HTML and PDF versions of the article have been updated to correct the errors.

Table 3. MS parameters and settings.

Make-up solvent	Acetonitrile		
Make-up solvent flow:	0.120 mL min ⁻¹		
Ion source Type:	ESI		
Ionisation mode:	Negative		
Spray voltage:	2500 V		
Sheath gas:	50		
Aux gas:	10		
Sweep gas:	1		
Ion transfer tube temp:	325 °C		
Vaporiser temp:	350 °C		
Acquisition mode:	Selected reaction monitoring (SRM)		
Compound	Precursor (Q1) (m/z)	Product (Q3) (m/z)	Collision Energy (eV)
chloride	35	35	0
nitrite	46	46	0
nitrate	62	62	0
chlorite	67	67	0
bromide	79	79	0
chlorate	83	67	20
sulfate	97	79	24
phosphate	97	79	15
perchlorate	99	83	20
bromate	127	111	20
iodide	127	127	0
iodate	175	159	21
Q1 Resolution (FWHM)	0.7		
Q3 Resolution (FWHM)	1.2		
CID Gas (mTorr)	1.5		

DOI of original article: <https://doi.org/10.1016/j.heliyon.2021.e06885>.

* Corresponding author.

E-mail address: Estrella.SanzRodriguez@utas.edu.au (E. Sanz Rodriguez).<https://doi.org/10.1016/j.heliyon.2021.e07426>

Received 24 June 2021; Accepted 24 June 2021

2405-8440/© 2021 Published by Elsevier Ltd.