



Novel imaging phantom for accurate and robust measurement of brain atrophy rates using clinical MRI



Houshang Amiri^{a,*}, Iman Brouwer^a, Joost P.A. Kuijter^a, Jan C. de Munck^a, Frederik Barkhof^{a,b}, Hugo Vrenken^a

^a Department of Radiology and Nuclear Medicine, VU University Medical Center, Amsterdam, the Netherlands.

^b Institutes of Neurology and Healthcare Engineering, UCL, London, UK.

ARTICLE INFO

Keywords:

Brain atrophy
MRI
Standardization
Phantom
Segmentation

ABSTRACT

Brain volume loss, or atrophy, has been proven to be an important characteristic of neurological diseases such as Alzheimer's disease and multiple sclerosis. To use atrophy rate as a reliable clinical biomarker and to increase statistical power in clinical treatment trials, measurement variability needs to be minimized. Among other sources, systematic differences between different MR scanners are suspected to contribute to this variability. In this study we developed and performed initial validation tests of an MR-compatible phantom and analysis software for robust and reliable evaluation of the brain volume loss. The phantom contained three inflatable models of brain structures, i.e. cerebral hemisphere, putamen, and caudate nucleus. Software to reliably quantify volumes from the phantom images was also developed. To validate the method, the phantom was imaged using 3D T1-weighted protocols at three clinical 3T MR scanners from different vendors. Calculated volume change from MRI was compared with the known applied volume change using ICC and mean absolute difference. As assessed by the ICC, the agreement between our developed software and the applied volume change for different structures ranged from 0.999–1 for hemisphere, 0.976–0.998 for putamen, and 0.985–0.999 for caudate nucleus. The mean absolute differences between measured and applied volume change were 109–332 μL for hemisphere, 2.9–11.9 μL for putamen, and 2.2–10.1 μL for caudate nucleus. This method offers a reliable and robust measurement of volume change using MR images and could potentially be used to standardize clinical measurement of atrophy rates.

1. Introduction

Neurodegeneration is a persistent problem in multiple sclerosis (MS) and some other diseases like Alzheimer's disease (AD). In MS, new treatments are being developed to slow down or prevent this neurodegeneration. An accessible in vivo surrogate of neurodegeneration is brain atrophy, i.e. brain volume loss, which can be measured using magnetic resonance imaging (MRI). However, measurement variability is substantial; among other factors, differences between MRI scanners introduce unwanted variability of atrophy measures. Studies in MS observed different volumes and volume changes for two different scanners using similar or even identical acquisition parameters (Durand-Dubief et al., 2012), differences in global brain tissue volumes and cortical thickness were found between three different 3-Tesla MR scanners (Biberacher et al., 2016), and even different whole-brain volumes between two scanners of the exact same type with identical imaging protocols (Takao et al., 2011).

Unwanted scanner or acquisition protocol dependent variability of atrophy rates directly affects the statistical power of multi-center treatment trials. Moreover, it hampers the use of monitoring of atrophy progression in individual patients in the clinical setting. This is especially problematic in diseases like MS in which the typical annual rate of brain volume loss is only about 0.5–1% (De Stefano et al., 2016).

In MRI, imaging phantoms have been widely used for quality control as well as for detecting geometrical distortions (Chen et al., 2004; Ihalainen et al., 2011; Vassiliou et al., 2016; Wang et al., 2004a, 2004b). To our knowledge, no application has yet been published in which imaging phantoms are used to standardize measurements of brain atrophy rates between scanners and acquisition protocols.

In this study we aimed to develop a phantom for standardizing measurements of volume change between different scanners. This works by scanning the phantom at each scanner/site independently, and derive a calibration function specific to the scanner and acquisition protocol used. Therefore, the resulting calibration is supposed to be

* Corresponding author.

E-mail address: h.amiri@vumc.nl (H. Amiri).

<https://doi.org/10.1016/j.nicl.2019.101667>

Received 18 January 2018; Received in revised form 26 November 2018; Accepted 4 January 2019

Available online 04 January 2019

2213-1582/ © 2019 The Authors. Published by Elsevier Inc. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).




Fig. 1. (a) Side-view and (b) top-view of the phantom, with (c) the volume change system and (d) the complete device. Each structure had inlet and outlet valves to fill them up with water and to make them air-free, respectively. Rigid tubes (VYGON) of inner and outer diameter of 1 mm and 2 mm, respectively, were used to connect structures to the volume change system. All structures are mounted in an oval-shaped 3D-printed housing (PA2200, EOS SLS).

scanner-specific which takes into account a couple of sources of variability such as hardware, field strength and pulse sequence type. The method makes use of an imaging phantom containing three inflatable structures to which volume change can be applied in a highly controlled way. The phantom structures resemble three brain structures, i.e. hemisphere, putamen and caudate nucleus. In this initial validation study, the phantom is imaged at various known set volume changes, and then for each structure the volume change is quantified from the phantom MR images using dedicated segmentation software.

2. Materials and methods

2.1. Phantom development

The MRI phantom contains inflatable structures equipped with a precise volume change system (Fig. 1). The phantom contains three brain structures: a cerebral hemisphere, putamen, and caudate nucleus. The hemisphere model was created by scanning a post-mortem healthy brain using a surface 3D scanner (Artec Spider, Arc 3D). The models of putamen and caudate nucleus were generated from MR images of healthy control subjects using FSL-FIRST (Patenaude et al., 2011). Using these models, molds of all three structures were 3D-printed by ZP151 powder (3D systems). Then, the molds were repeatedly dipped into pre-vulcanized latex rubber (pvs-elastica, FromX) to reach a homogenized thickness. Finally, the molds were dissolved in water and phantoms remained. All three structures were attached on a nylon plate for stabilization and embedded in a 3D-printed cylinder made by plexiglass. Syringes (Hamilton®, Switzerland) of 100 mL and 100 μ L were incorporated with a custom-made stainless steel plunger with a

finely bored syringe barrel to change the volume in the hemisphere and the small structures, respectively.

2.2. MR imaging

To validate the method, the phantom was imaged at three different clinical 3 T MR scanners (GE, Toshiba and Philips) using a standard clinical 3D gradient echo sequence available at each scanner with voxel volumes 1–1.2 μ L. To eliminate any signal originating from the latex material constituting the inflatable structures, the clinical protocol was adjusted by a suitable choice of echo time (TE) and repetition time (TR). Other parameters remained unchanged. For example, at each scanner, we used the k-space filters (Fermi at GE, NL2:7* k-space LSI (linear space invariant) only at Toshiba, and no filters at Philips) which were already used in the clinic to get the best image qualities. Scanner and pulse sequence parameters are given in Table 1. Typical phantom MR images on Toshiba are shown in Fig. 2.

2.3. Volume change

In each scanning session, five volume changes were applied to each of the three structures as indicated in Table 2. The volume changes were applied as integer multiples of the full rotation of the precision screw controlling the syringe. They were chosen to range between about 1% - 5% of the volume of each structure. Because the caudate nucleus and putamen are of comparable size, the same amount of volume changes were applied to them. For the hemisphere, the applied volume changes were chosen much larger than for the two small structures. Note that the volumes of all three structures were increased

Table 1
Scanner and pulse sequence parameters for the three scanners used.

Vendor	GE	Philips	Toshiba
Model	MR750	Ingenuity	Titan
Field strength (T)	3	3	3
Coil type	8-channel head coil	8-channel head coil	32-channel head coil
Sequence type	FSPGR	FEE	FEE
TR (ms)	8.21	7.9	5.7
TE (ms)	3.2	4.5	2.4
Flip angle (degrees)	12	8	5
Acquisition matrix	256*256	256*256	256*256
Voxel size (mm)	1 × 1 × 1	1 × 1 × 1	1 × 1.2 × 1
Slab orientation	sagittal	sagittal	sagittal
Number of slices	140	140	140
Parallel imaging?	Yes	Yes	Yes

simultaneously as indicated in Table 2, to maximize efficiency of scanning. For the analysis, step number five was taken as a reference volume to calculate the volume decrease towards baseline.

2.4. Image analysis

We developed a segmentation software to measure the volume changes from the MR images of the phantom. Note that the software operates on a single 3D image, and quantifies the water volume inside the structure of interest for that image. Volume changes between steps in a phantom measurement series are then quantified by subtracting the calculated volumes, rather than being quantified directly by our software.

Fig. 3 shows a diagram of the image analysis pipeline. The final output of the software is the volume of water inside the structure of interest, which is quantified by integrating the voxel-wise partial volume estimates of water across the interior of the structure, including its internal latex-water boundary voxels. First, the images were corrected for intensity non-uniformity using nonparametric non-uniform intensity normalization (N3) (Sled et al., 1998). A spline distance of 50 mm and a full width at half maximum (FWHM) value of 0.05 were used. The small FWHM value was chosen to favour accuracy over speed. Because some Gibbs ringing is present in our images, peak values were adjusted as follows. First, the mean and standard deviation of the water inside the hemisphere was calculated. Then inside the area that is considered to be surely water, voxels with intensity higher than the median plus 3 times the standard deviation were reduced to the square root above this cut-off level. Because latex and nylon give very low signal, we improved their contrast with the background by adding to the intensity of every voxel inside the housing (i.e., also including the water voxels) 0.5 times the mean intensity of the water inside the core of the hemisphere.

Next, images were prepared to perform FSL-FAST partial volume

Table 2
Applied volume changes of different structures at different steps.

Step	Caudate nucleus volume-change (μL)	Putamen volume-change (μL)	Hemisphere volume-change (mL)
Baseline	0	0	0
1	26.46	26.46	3.332
2	52.92	52.92	6.665
3	79.38	79.38	9.998
4	105.84	105.84	13.330
5	132.30	132.30	16.663

estimation. FSL-FAST models voxel intensities with 3 Gaussian distributions and noise. Without additional pre-processing, FSL-FAST may treat the nylon and latex material as noise because (i) only a small portion of voxels in the image contain nylon or latex and (ii) the intensity distribution of these voxels does not match a Gaussian distribution due to the relatively high number of partial volume voxels. Therefore, we added areas of voxels with the same intensity as the latex (“virtual latex”). In order not to influence the partial volume estimates (PVEs) in the boundary between water and real latex, we added this (virtual) extra latex well away from the real latex. Subsequently, FSL-FAST was run without bias field correction. Finally, to correctly estimate volumes, the class of the virtual latex was replaced with the water class.

2.5. Statistical analysis

To determine the reliability of our measured volume changes, measured volume changed were compared with the applied volume changes using interclass correlation coefficient (ICC) with the two-way mixed model for absolute agreement. Furthermore, mean absolute differences (MAD) between measured and applied volume changes were calculated across each phantom measurement series. For both statistics, SPSS software version 22 (IBM) was used.

3. Results

3.1. Phantom validation

A representative MR image of the phantom for a given volume is shown in Fig. 2. To validate the accuracy of the volume change system and the developed segmentation software, the true applied volume changes versus measured volume changes are plotted in Fig. 4. As seen in this figure, the measured data showed a perfect agreement with the applied volume changes.




Fig. 2. Representative MR images of the hemisphere, putamen, and caudate nucleus phantoms in (a) axial, (b) coronal and (c) sagittal plane as obtained on Toshiba Titan.




Fig. 3. Diagram of the image analysis pipeline to estimate the volumes of the hemisphere, putamen and caudate nucleus. Briefly, the software includes correction for intensity non-uniformity using N3, Gibbs ringing artifact removal, introduction of an artificial (“virtual”) latex component to balance the classes in the subsequent FSL-FAST segmentation, and quantification of total internal volume of each structure from the PVE values obtained from FSL-FAST. Red-to-yellow color bar indicates the water PVE.




Fig. 4. Applied versus measured volume change at three different scanners for (top row) hemisphere, (middle row) caudate nucleus, and (bottom row) putamen. The green dotted line represents the identity line.

Table 3
ICC and MAD for all three brain structures.

	Hemisphere		Putamen		Caudate nucleus	
	ICC ^a	MAD ^b (mL)	ICC	MAD (μL)	ICC	MAD (μL)
GE	0.999	0.3322	0.998	2.9	0.999	2.2
Toshiba	1.000	0.1089	0.994	5.5	0.997	3.7
Philips	1.000	0.2612	0.976	11.9	0.985	10.1

^a ICC = Interclass correlation coefficient.

^b MAD = Mean absolute difference; voxel sizes: 1 μL for GE and Philips and 1.2 μL for Toshiba.

3.2. ICC and mean absolute differences

Table 3 reports the ICC for absolute agreement and MAD, for the three brain structures. ICC values indicate that agreement between applied and measured volume changes is excellent. The MAD data also shows that deviation of the measured volume changes from the mean change are relatively negligible compared to the applied volume changes.

4. Discussion

In this study we developed and validated a phantom and analysis software to induce and measure volume change of models of brain structures. The ICCs revealed a very strong agreement between our developed method and the applied volume change. Therefore, our method can potentially offer a reliable and robust tool to standardize measurements of volume change in clinical trials as well as in clinical settings.

Despite the importance of brain atrophy measurement in neurological conditions such as MS and AD, its clinical use at the individual subject level is hampered by the large measurement variability (Rocca et al., 2017). Among other sources of variability such as physiological effects, between-scanner and between-acquisition protocol differences are suspected to contribute importantly (Amiri et al., 2018). For example, volume differences were seen when using two different scanners with similar or even identical acquisition parameters (Durand-Dubief et al., 2012). In another study performed at three different 3-Tesla MR scanners, significant differences in total lesion volume, global brain tissue volumes and cortical thickness measures between the scanners were observed (Biberacher et al., 2016). Even when using identical imaging protocols, it was found that the whole brain volumes were significantly different between two scanners of the exact same type (Takao et al., 2011). Therefore, to increase statistical power in clinical trials and to establish atrophy measures as reliable clinical biomarkers, such scanner-induced effects should be minimized.

To date, some anthropomorphic brain phantoms have been described, e.g. by using agar gel and manganese chloride (Altermatt et al.,

2017; Fujimoto et al., 2015), or using polyvinyl alcohol (PVA) (Chen et al., 2012). The advantage of the method presented in this paper is that it would allow standardization of the volume change measurements within and between scanners and acquisition protocols. Moreover, this method can accurately and robustly provide a selected volume change to mimic e.g. the change typically observed in a certain disease.

A limitation of this study was that we did not test our method in vivo. Therefore, an important next step in the development of the present method would be to investigate translation of this method to a clinical setting to validate it in patients. In our view, as schematically illustrated in Fig. 5, the translation to an in vivo setting requires additional steps such as (i) the phantom contrast-to-noise ratio (CNR) should mimic the in vivo CNR and (ii) the image analysis software used to obtain calibration functions from the phantom images should be similar to that used in the patient studies. This implementation includes acquiring scans of the phantom on the same day as patient scans while applying known volume changes to the phantom. This would lead to calibration factors not only per scanner and acquisition protocol, but also per day of patient visits. Future studies should investigate whether scanning the phantom on the same day is necessary. Furthermore, it would be interesting to use this method to test widely used brain volume analysis software to assess their accuracy, in a manner comparable to the work done by Chen et al. on their PVA phantom (Chen et al., 2012), and to possibly find standardization functions to correct for them. This would facilitate use of this method to standardize volume change measurements obtained on different scanners and analysed using specific analysis methods.

It is known that analysis methods are another source of variability in volume measurements. A study tested reproducibility of the commonly used software package SIENA (Cover et al., 2011). They showed that the variability of atrophy measurement, obtained from scan and re-scan of the subjects without repositioning, was of considerable magnitude compared to the typical atrophy rates. Interestingly, use of 3-Tesla instead of 1.5-Tesla MRI did not reduce the error (Cover et al., 2014). It also has been shown that, for both FSL-FIRST and FreeSurfer, within-session reproducibility error for measuring 1-year hippocampal volume change was substantial (Mulder et al., 2014). A recent study (Cover et al., 2016) showed that reproducibility of 1-year volume change highly depends on the image analysis method, which affects the statistical power and required sample sizes in clinical trials. Therefore, next to the scanner-induced variabilities, the analysis software should also be considered as an important source of variability in measuring brain volume change in vivo. Even for in vivo measurements that are standardized according to the final proposed clinical method, such variability of the in vivo analysis software would remain an issue. Of note, the analysis software that we developed provides high accuracy in the measurement of the phantom volume change, thus adding only minimal error.

In conclusion, we demonstrated the feasibility of producing




Fig. 5. Scheme of the phantom clinical implementation for “true volume change” calculation.

inflatable brain phantoms, and were able to adjust the volume of such a phantom in a controlled and precise way. Future studies should implement and validate the clinical application of this phantom for standardizing atrophy measurements.

Acknowledgments

This study was kindly supported by Amsterdam Neuroscience (formerly Neuroscience Campus Amsterdam) (proof-of-concept grant PoC-2014-BIT-03), and Novartis (grant SP037.15/432282). F. Barkhof was supported by the NIHR UCLH biomedical research centre. We would like to thank the 3D Innovation Lab at the VU University Medical Center for their technical support, especially D. Koops, F. Verver, and S. te Slaa.

References

- Altermatt, A., Santini, F., Deligianni, X., Magon, S., Sprenger, T., Kappos, L., Cattin, P., Wuerfel, L., 2017. On the Construction of a 3D-Printed Brain Phantom as Gold Standard for the Validation of Brain Segmentations. ECRIMS, Paris, France.
- Amiri, Houshang, Sitter, Alexandra de, Bendfeldt, Kerstin, Battaglini, Marco, Wheeler-Kingshott, Claudia A.M. Gandini, Calabrese, Massimiliano, Geurts, Jeroen J.G., Rocca, Maria A., Sastre-Garriga, Jaume, Enzinger, Christian, Stefano, Nicola de, Filippi, Massimo, Rovira, Alex, Barkhof, Frederik, Vrenken, Hugo, MAGNIMS Study Group, 2018. Urgent challenges in quantification and interpretation of brain grey matter atrophy in individual MS patients using MRI. *NeuroImage: Clinical* 19, 466–475.
- Biberacher, V., Schmidt, P., Keshavan, A., Boucard, C.C., Righart, R., Samann, P., Preibisch, C., Frobel, D., Aly, L., Hemmer, B., Zimmer, C., Henry, R.G., Muhlau, M., 2016. Intra- and interscanner variability of magnetic resonance imaging based volumetry in multiple sclerosis. *NeuroImage* 142, 188–197.
- Chen, C.C., Wan, Y.L., Wai, Y.Y., Liu, H.L., 2004. Quality assurance of clinical MRI scanners using ACR MRI phantom: preliminary results. *J. Digit. Imaging* 17, 279–284.
- Chen, S.J., Hellier, P., Marchal, M., Gauvrit, J.Y., Carpentier, R., Morandi, X., Collins, D.L., 2012. An anthropomorphic polyvinyl alcohol brain phantom based on Colin27 for use in multimodal imaging. *Med. Phys.* 39, 554–561.
- Cover, K.S., van Schijndel, R.A., van Dijk, B.W., Redolfi, A., Knol, D.L., Frisoni, G.B., Barkhof, F., Vrenken, H., neuGrid, Alzheimer's Disease Neuroimaging, I, 2011. Assessing the reproducibility of the SienaX and Siena brain atrophy measures using the ADNI back-to-back MP-RAGE MRI scans. *Psychiatry Res.* 193, 182–190.
- Cover, K.S., van Schijndel, R.A., Popescu, V., van Dijk, B.W., Redolfi, A., Knol, D.L., Frisoni, G.B., Barkhof, F., Vrenken, H., neuGrid, Alzheimers Disease Neuroimaging, I, 2014. The SIENA/FSL whole brain atrophy algorithm is no more reproducible at 3T than 1.5 T for Alzheimer's disease. *Psychiatry Res.* 224, 14–21.
- Cover, K.S., van Schijndel, R.A., Versteeg, A., Leung, K.K., Mulder, E.R., Jong, R.A., Visser, P.J., Redolfi, A., Revillard, J., Grenier, B., Manset, D., Damangir, S., Bosco, P., Vrenken, H., van Dijk, B.W., Frisoni, G.B., Barkhof, F., Alzheimer's Disease Neuroimaging Initiative, n, 2016. Reproducibility of hippocampal atrophy rates measured with manual, FreeSurfer, AdaBoost, FSL/FIRST and the MAPS-HBSI methods in Alzheimer's disease. *Psychiatry Res.* 252, 26–35.
- De Stefano, N., Stromillo, M.L., Giorgio, A., Bartolozzi, M.L., Battaglini, M., Baldini, M., Portaccio, E., Amato, M.P., Sormani, M.P., 2016. Establishing pathological cut-offs of brain atrophy rates in multiple sclerosis. *J. Neurol. Neurosurg. Psychiatry* 87, 93–99.
- Durand-Dubief, F., Belaroussi, B., Armspach, J.P., Dufour, M., Roggerone, S., Vukusic, S., Hannoun, S., Sappey-Mariniere, D., Confavreux, C., Cotton, F., 2012. Reliability of longitudinal brain volume loss measurements between 2 sites in patients with multiple sclerosis: comparison of 7 quantification techniques. *AJNR Am. J. Neuroradiol.* 33, 1918–1924.
- Fujimoto, K., Robertson, T.V., Douet, V., Garmire, D., Stenger, V.A., 2015. A Structurally Anthropomorphic Brain Phantom. ISMRM, Toronto, Canada.
- Ihalainen, T.M., Lonnroth, N.T., Peltonen, J.I., Uusi-Simola, J.K., Timonen, M.H., Kuusela, L.J., Savolainen, S.E., Sipila, O.E., 2011. MRI quality assurance using the ACR phantom in a multi-unit imaging center. *Acta Oncol.* 50, 966–972.
- Mulder, E.R., de Jong, R.A., Knol, D.L., van Schijndel, R.A., Cover, K.S., Visser, P.J., Barkhof, F., Vrenken, H., Alzheimer's Disease Neuroimaging, I, 2014. Hippocampal volume change measurement: quantitative assessment of the reproducibility of expert manual outlining and the automated methods FreeSurfer and FIRST. *NeuroImage* 92, 169–181.
- Patenaude, B., Smith, S.M., Kennedy, D.N., Jenkinson, M., 2011. A Bayesian model of shape and appearance for subcortical brain segmentation. *NeuroImage* 56, 907–922.
- Rocca, M.A., Battaglini, M., Benedict, R.H., De Stefano, N., Geurts, J.J., Henry, R.G., Horsfield, M.A., Jenkinson, M., Pagani, E., Filippi, M., 2017. Brain MRI atrophy quantification in MS: from methods to clinical application. *Neurology* 88, 403–413.
- Sled, J.G., Zijdenbos, A.P., Evans, A.C., 1998. A nonparametric method for automatic correction of intensity nonuniformity in MRI data. *IEEE Trans. Med. Imaging* 17, 87–97.
- Takao, H., Hayashi, N., Ohtomo, K., 2011. Effect of scanner in longitudinal studies of brain volume changes. *J. Magn. Reson. Imaging* 34, 438–444.
- Vassiliou, V.S., Heng, E.L., Gatehouse, P.D., Donovan, J., Raphael, C.E., Giri, S., Babu-Narayan, S.V., Gatzoulis, M.A., Pennell, D.J., Prasad, S.K., Firmin, D.N., 2016. Magnetic resonance imaging phantoms for quality-control of myocardial T1 and ECV mapping: specific formulation, long-term stability and variation with heart rate and temperature. *J. Cardiovasc. Magn. Reson.* 18, 62.
- Wang, D., Strugnell, W., Cowin, G., Doddrell, D.M., Slaughter, R., 2004a. Geometric distortion in clinical MRI systems part I: evaluation using a 3D phantom. *Magn. Reson. Imaging* 22, 1211–1221.
- Wang, D., Strugnell, W., Cowin, G., Doddrell, D.M., Slaughter, R., 2004b. Geometric distortion in clinical MRI systems Part II: correction using a 3D phantom. *Magn. Reson. Imaging* 22, 1223–1232.