

Received 7 July 2016; revised 26 September 2016 and 16 November 2016; accepted 27 November 2016.
Date of publication 9 January 2017; date of current version 6 February 2017.

Digital Object Identifier 10.1109/JTEHM.2017.2649491

Detection of Paroxysms in Long-Term, Single-Channel EEG-Monitoring of Patients with Typical Absence Seizures

TROELS W. KJAER¹, HELGE B. D. SORENSEN², (Member, IEEE), SABINE GROENBORG³, CHARLOTTE R. PEDERSEN⁴, AND JONAS DUUN-HENRIKSEN⁵

¹Center of Neurophysiology, Department of Neurology, Zealand University Hospital, Roskilde, Denmark

²Department of Electrical Engineering, Technical University of Denmark, Kgs. Lyngby, Denmark

³Juliane Marie Centre, Rigshospitalet University Hospital, Copenhagen, Denmark

⁴Department of Children and Youth, Nordzealand Hospital Hillerød, Hillerød, Denmark

⁵Hypo-Safe A/S, Lyngby, Denmark

CORRESPONDING AUTHOR: J. DUUN-HENRIKSEN (jdh@hyposafe.com)

This work was supported in part by The Ministry of Higher Education and Science, Denmark, under Grant 08-041638-U and in part by the Elsass Foundation.

ABSTRACT Absence seizures are associated with generalized 2.5–5 Hz spike-wave discharges in the electroencephalogram (EEG). Rarely are patients, parents, or physicians aware of the duration or incidence of seizures. Six patients were monitored with a portable EEG-device over four times 24 h to evaluate how easily outpatients are monitored and how well an automatic seizure detection algorithm can identify the absences. Based on patient-specific modeling, we achieved a sensitivity of 98.4% with only 0.23 false detections per hour. This yields a clinically satisfying performance with a positive predictive value of 87.1%. Portable EEG-recorders identifying paroxysmic events in epilepsy outpatients are a promising tool for patients and physicians dealing with absence epilepsy. Albeit the small size of the EEG-device, some children still complained about the obtrusive nature of the device. We aim at developing less obtrusive though still very efficient devices, e.g., hidden in the ear canal or below the skin.

INDEX TERMS Absence seizures, automatic seizure detection, epilepsy, single channel EEG, SVM.

I. INTRODUCTION

A dilemma when treating patients with epilepsy is that the attending physician does not know the number of seizures, and often the patient does not know it either [1]. The physician works with an otherwise normal patient with seizures happening elsewhere and the patient might be unaware of the episodes. Devices that record the seizures during normal everyday life conditions may reduce this problem.

Use of outpatient ambulatory electroencephalogram (EEG) began 50 years ago. At first the recording systems were large and heavy, the data quality poor and time to mount considerable. Since then the data quality has improved and the number of channels increased [2]. Now numerous portable recording systems of up to 36 channels with sampling rates of up to 400 Hz are commercially available. We currently work on unobtrusive solutions like a small earplug that communicates with a smartphone and is easily mounted by

the patient in a few seconds [3]. This allows for access to real time registration as well as intervention and alarm systems.

Clinical investigations have documented that ambulatory EEG can record focal and generalized epileptiform activity [4], [5]. Several studies of long term (up to a few days) EEG recording has proven useful when validating a diagnosis of epilepsy, do seizure classification or when identifying seizure onset zone in patients undergoing epilepsy surgery workup [6]. Several studies have also found that ambulatory EEG monitoring is a useful tool in patients with generalized discharges [7], [8]. Furthermore, the number and location of electrodes is likely to be less important than in focal epilepsies [9]. When EEG is performed in the clinic in a short time slot at a specific time of day, much clinical important information is likely to be lost compared to repeated 24-h measures in the natural environment.

Especially within many idiopathic generalized epilepsies, there is a clear advantage of repeated EEG recordings. There exist a correlation between the amount of anti-epileptic drug intake and the number of seizures they have [10]. But the drugs entail considerable adverse effects [11] that should be minimized by optimization of treatment dosage.

In order to explore this central schism in epilepsy treatment we chose to investigate patients with typical Absence Seizures with repeated 24-h EEG recording at home. Typical Absence Seizures are seen in childhood absence epilepsy (CAE) and juvenile absence epilepsy (JAE) which are relatively common epileptic syndromes. Absences are characterized by rhythmic generalized 2.5-5 Hz spike-polyspike-wave paroxysms in the EEG [12]. Clinically, absences are characterized by sudden loss of awareness, a blank stare often with upward eye deviation and cessation of normal activities. On average the ictal duration is 9.4 ± 7 s and due to their subtle nature they are often overlooked by the surroundings [13]. In a study with ambulatory cassette EEG it was found that parents only registered 6% of paroxysms lasting more than 3 s [14] which is the generally believed threshold for minimum length of a paroxysm with a clinical correlate [15], [16].

The generalized paroxysms of absence epilepsy are usually not truly generalized, i.e. the signal is present with certain topographic preferences especially in the frontal lobes [17]. This suggests that electrode placement for EEG-monitoring influence signal detection.

Multiple research groups have investigated the feasibility of ambulatory monitoring of absence epilepsy patients. Generally, the studies are small, aged and based on relatively short period (up to 12 hours) of EEG-recording. Early publication are based on combined analogue and digital analysis and demonstrate good agreement between visual inspection and the automated detection of sharp transients in one fronto-central EEG-channel [18], [19].

Later studies have used microcomputer based detectors of sharp transients with loose temporal coupling allowing for detection of spike-waves of varying frequencies [20]. It was found that automatic detection using a single channel did not perform as well as human specialists with the methods and signal quality available at that time. The authors state that multichannel EEG signal processing is required for comprehensive quantitative detection of clinical significant spike wave patterns. A similar microcomputer based study focused on the frequency of the wave component and found on the other hand acceptable level of detection with a single channel [21].

Recent studies have used comprehensive digital signal processing methods based on mixed spike and slow wave characteristics extracted by methods like wavelet analysis [9], [22]. However even the most recent of these studies have a limited amount of data with only few hours of recording in awake patients. Since delta activity during sleep shares some characteristics with the EEG during absences there is a pressing

need for ultra-long term recording like the ones presented here.

Based on 20 standard-EEG recordings, we have previously found that the Fp1-F7 channel is the most sensitive channel for reliable detection of seizures in a group of patients with absence seizures [9]. That study showed it was possible to detect 97.2% of all seizures lasting more than 2 seconds without any false detections. However, it was limited to data recorded in the outpatient clinic of the hospital, thus not showing the type and extends of artifacts expected in ambulatory measurements of children. Furthermore, this location aligns well with an EEG monitoring application where parents can mount pre-gelled disposable electrodes themselves whenever there is indication for a prolonged ambulatory EEG. We thus set up an experiment with multiple day ambulatory EEG recordings with electrodes at Fp1 and F7.

II. METHODS

A. PARTICIPANTS

Children aged 5-18 years with suspected or diagnosed absence epilepsy were enrolled after informed consent by the custody holder (approved by the local ethics committee, H-3-2011-054). Two pediatric outpatient clinics (Rigshospitalet and Northzealand Hospital) screened for participants by the attending physicians (SG, CRP). The selection of patients is probably biased towards cases that are more complicated since simple cases are handled out of hospital and the motivation for enrollment was based on suspicion of unrecognized epileptic episodes.

15 children aged 5-16 were monitored in this protocol. Six patients were suspected of having epilepsy based on teacher or parent observations but turned out not to have epilepsy. Two patients had epilepsy but not typical absences. One patient had previously had juvenile absence epilepsy and participated successfully to unmount the diagnosis. Therefore, we present data on six patients with absence seizures, see Table 1. Five of the six patients were tapered up in antiepileptic drug (AED) during the trial. In only one patient (#5) the increase was partly founded in the EEG-monitoring-results during this study. The therapeutic approach to the rest of the patients did not consider the EEG-monitoring-results.

Data from the two first patients have previously been mentioned in [23] although without the extensive analysis done in this paper.

B. STUDY PROTOCOL

The study protocol is schematized in Fig. 1. At the first experimental day, patients came sleep deprived to the department of clinical neurophysiology to have a standard-EEG performed as part of normal workup.

Three electrodes (Ambu® Neuroline 700, Ballerup, Denmark) were placed near Fp1 (Ref), F7 (Active 1) and TP7 (Active 2), see Fig. 2. Electrodes were connected to a tiny, portable EEG-recorder (Actiwave, CamNtech Ltd., Cambridge, UK) measuring only 3.7x2.7x8.5 cm and

TABLE 1. Participants.

#	Gender	Age [yr]	Body weight [kg]	Epileptic syndrome	AED at day 1	AED at day 30	# of paroxysms	Paroxysm length [s]	Time analyzed [h]
1	M	11	44	JAE	Nihil	VPA 600 mg	26	10.9±11.13	40.3
2	F	11	43	Complex CAE	VPA 900 mg LTG 150 mg ZNS 200 mg	VPA 900 mg LTG 150 mg ZNS 200 mg	56	9.8±5.5	65.1
3	F	12	40	JAE	Nihil	VPA 400 mg	26	4.8±1.5	45.0
4	F	8	31	CAE	LTG 20 mg	LTG 75 mg	301	4.2±1.1	94.5
5	F	11	29	CAE	Nihil	LTG 400 mg	13	4.8±2.4	43.4
6	F	7	31	CAE	LTG 25 mg	LTG 75 mg	171	8.7±6.7	93.2

M is male and F is female. JAE is Juvenile Absence Epilepsy and CAE is Childhood Absence Epilepsy. AED is Anti-Epileptic Drug, VPA is Valproate, LTG is Lamotrigine and ZNS is Zonisamide.

of paroxysms are only those lasting more than three seconds and only included on days with at least five paroxysms recorded. Time analyzed is also based on days with at least five paroxysms.




FIGURE 1. Study protocol overview. Patients are monitored 24 hours on four independent days. At the first recording day, a standard scalp-EEG is performed at the hospital.

weighing 8.5 g which was taped to their neck for the following 24 hours of recording. Then Ag/AgCl cup electrodes (Ambu® Neuroline Cup) for the standard-EEG were mounted with Ten20 conductive paste (Weaver and Company, Aurora, CO, USA) without interfering with the ambulatory electrodes. Patients were examined for 30 minutes in supine position with subdued light and eyes closed. Provocations with hyperventilation and photo-stimulation were performed. Half of the subjects dozed or fell asleep.

When the standard-EEG was completed, participants left the department with the portable EEG device still recording. They kept this device on for the following 24 hours until the parents dismantled it as the recording period was over. At approximately day 4, 8, and 30, JDH came to the patients’ home to mount the EEG device for another 24 hours of recording. To accommodate the plans for the families, we were very unrestricted on the exact timing of when to do the recordings.

If electrodes became loose, parents and children were instructed to simply push them back on, and perhaps apply extra tape over the electrodes if they found that necessary.

After each 24 hours recording the EEG was visually inspected and scored for paroxysms by an expert trained




FIGURE 2. Experimental setup. Three electrodes were attached to the skin at approximate location of Fp1 (Ref), F7 and TP7. VD is the voltage divider that doubles the dynamic input range to 800 μVpp.

technician and the number was reported immediately to the treating physician, who could choose to include the information in decision on medical action or not.

Parents were asked for some time during the four 24 hours recording sessions to be 100% attentive to whether their child had clinical episodes and note any of them on a report sheet. Before and after the experiments parents and children underwent a semi-structured interview regarding the effect of the absences on their child and how obtruded the children had felt wearing the portable EEG device.

C. DATA ACQUISITION

The portable EEG device sampled with a frequency of 128 Hz. The hardware frequency band spans from 0.3 to 50 Hz and the dynamic range of the EEG-signal in




FIGURE 3. Data and signal processing diagram. The processes within the dashed line are repeated in a 5-fold cross-validation scheme.

the portable recorder was $400 \mu\text{V}$ peak to peak. Initial measurements showed this was inadequate so a voltage divider was constructed to double the maximum amplitude. It was based on $220 \text{ k}\Omega$ resistors. This level of resistance is much lower than the input impedance of $10 \text{ M}\Omega$ and much higher than usual skin impedance of approximately $5 \text{ k}\Omega$ [23]. The voltage division will thus only occur over the resistors. In [23] it was found that the thermal noise from the resistors only amounted to $0.30 \mu\text{V}_{\text{rms}}$ which is well below the discrete voltage resolution of $0.78 \mu\text{V}$ meaning that the resistor noise contribution is small.

As the portable device was only able to store 14 hours and 33 minutes of EEG, the parents had to change devices after approximately 12 hours.

D. FEATURE EXTRACTION AND TRANSFORMATION

Special attention has been paid towards development of features that were descriptive of the spike-wave pattern as well as using state-of-the-art digital signal processing methods. The signal processing diagram can be seen in Fig. 3.

The EEG signal was divided into 2 s windows with an overlap of 1 s. Each window was weighted with a Tukey tapered cosine window with a ratio of taper to constant sections of

$1/3$ [24]. Based on a previous study [9] multiple features were extracted for each window using MATLAB R2011a. With use of the Mahalanobis distance criterion, the 10 most orthogonal features were selected for further analysis [25]. Only the method of extraction of these features will be presented here.

1) LOG-SUM OF WAVELET TRANSFORM

The popular wavelet transform has previously proved to be a strong method for EEG characterization [26], [27] especially using the Daubechies 4 mother wavelet. With a sampling frequency of 128 Hz, the multilevel wavelet detail decompositions contain the following frequencies: $d1$: 32-64 Hz, $d2$: 16-32 Hz, $d3$: 8-16 Hz, $d4$: 4-8 Hz and $d5$: 2-4 Hz. For each band the absolute sum was calculated for all windows [9]. The features were finally log-transformed to obtain normal distribution. Only $d1$, $d2$, $d3$ and $d5$ showed to be most valuable for the automatic seizure detection based on the feature selection criterion.

2) POWER MEASURES

During an absence seizure, the EEG amplitude is greatly increased. From an background activity of $\pm 15 \mu\text{V}$ the amplitude can easily increase to $\pm 200 \mu\text{V}$ depending on the electrode positions [7]. The high amplitudes are especially pronounced in the frontal channels. Two power measures were thus used as features: One that computed the power in the signal in a frequency band of 1 to 30 Hz, representing most of the physiological EEG power in normal EEG, and one that computed the relative power between the 3 to 12 Hz band and 1 to 30 Hz, being close to 1 if signal is paroxysmic and lower if signal is due to wide band artifact. The pass-bands were computed by filtering with a FIR equiripple design of order 467.

The first feature was transformed by an exponentiation of $1/10$ to ensure normal distribution. The latter needed no transformation.

3) CROSS-CORRELATION MEASURES

As a typical absence seizure is highly regular with its characteristic spike-wave pattern, two cross-correlation measures were developed. The first calculated the cross-correlation between two on each other following windows and the second was the cross-correlation between the same time signal filtered in the frequency band of 3-12 Hz and 1-30 Hz based on the same filtering as above. Both features were normalized to lag 0 and the features were extracted as the highest absolute value at any lag. To obtain normal distribution the exponentiation was set to $1/2$ and 2 respectively.

4) MEAN PHASE VARIANCE

A typical absence seizure consists of a spike and a wave repeatedly. This might imply that the phase variation is higher than normal background EEG. To measure the phase variance we removed the offset from each window and calculated the imaginary part of the Hilbert transform. The relative phase between the signal window and the Hilbert transformed signal

was then calculated and finally the variance constituted the mean phase variance and the feature. No transformation was needed to obtain normal distribution.

5) MAHALANOBIS VARIANCE

Another way to compare the similarity of two signals is by the Mahalanobis distance. With the same filters as previously described, the signal was split into a 3-12 Hz frequency band and a 1-30 Hz band. The Mahalanobis distance was then calculated for each point in the 3-12 Hz band to the distribution of the 1-30 Hz band [25]. The variance of this result constituted the feature. To obtain normal distribution the exponentiation was set to 1/4.

E. TRAIN/TEST DATA SPLIT

As new electrodes were mounted on every new recording day in the trial, data from different days had to be analyzed separately. This requires a stringent split of data into train and test sets. Only days with at least 5 recorded paroxysms were included to ensure a robust 5-fold cross-validation with 4/5 of data used for training and the last 1/5 for testing. As EEG is non-stationary, the non-paroxysmic EEG was split into 10 min epochs, each representing five 2min epochs that were either considered as training or test data. This ensured that data from all EEG states were included in the model without overfitting the model. To create a random stratified division, *cvpartition* was used in MATLAB for both the paroxysms and non-paroxysmic data.

F. NORMALIZATION, SUBSET SELECTION AND MODELLING

To ensure equal weight between the features in the classifier, they were normalized to zero mean, unit standard deviation based on the training data.

Data were strongly unbalanced with huge amounts of normal EEG even in patients with many paroxysms. There was typically up to 1.000 times more normal data than paroxysmic. Therefore, before modelling the training set the amount of normal data had to be limited. It was chosen to do this by random under-sampling where the number of normal training windows is chosen as a factor of the number of paroxysmic windows. This factor was set to 25 as a reasonable choice to ensure a broad representation of the heterogeneous background EEG without dismissing the rather homogeneous paroxysmic windows. More advanced alternatives to under-sampling has been suggested [28] which would most likely lower the factor 25 and still represent the normal EEG satisfactory, but this was not investigated in the present study.

Design of the classifier was based on modelling of data using a support vector machine (SVM) from the LibSVM package, vers. 3.18 [29] as a toolbox for MATLAB R2011a. SVM is a supervised, binary, quadratic minimization problem. It is applied in combination with a radial basis kernel to formulate nonlinear extensions of the linear algorithm whenever nonlinear trends in the data are present. By mapping of data into a high-dimensional kernel induced feature space,

it is possible to obtain a global optimum with respect to classification performance.

To find the optimal settings of the SVM, gamma (the parameter deciding the width of the kernel) and C (the parameter deciding the cost of choosing normal samples over paroxysmic) were varied.

A radial basis kernel was used with a gamma varying between 0.005, 0.01, 0.02, 0.05, 0.1, and 0.2 and C was varied between 0.5, 1, 2, 4, and 8.

G. PERFORMANCE CALCULATION

Four measures were found suitable for this study; sensitivity, Se , false detection rate, FDR , Specificity, Sp , and positive predictive ration, PPV . All measures are based on the test data only. Se was calculated as the number of truly detected paroxysms divided by the number of all registered paroxysms longer than 3 s. FDR was calculated as the number of detected continuous windows in groups of at least two (corresponding to three seconds) divided by the number of hours of normal data tested upon. Sp was calculated as the number of false positive windows divided by the total number of normal windows. PPV was calculated as the number of truly detected paroxysms divided by the total number of detected windows in groups of at least two.

H. MODEL SELECTION

With gamma taking 6 different values and C taking 5, a total of 30 models were generated. To choose the optimal model a performance measure, P , weighting the tradeoff between false detections and sensitivity was developed. The optimization was treated as a minimization problem as the sensitivity was subtracted from 101. As the goal was to obtain a sensitivity of at least 90% and a FDR below 1/h the $(101-Se)$ would have a dynamic range of 11 while the FDR would only have a dynamic range of 1. To obtain an approximate equal weight we added the value of five to the FDR to reach equivalent means. This resulted in the following metric:

$$P = (FDR + 5) \cdot (101 - Se) \quad (1)$$

III. RESULTS

A. PATIENT COMPLIANCE

Six highly compliant subjects were identified for this study. They exhibited between 8 and 84 paroxysms on the first day of recording, but all were having fewer paroxysms on the last day of recording, most likely due to increase in AED during the experiment for five of the patients, see Fig. 4 and Table 1. Two of the subjects experienced more than 40% of their maximum number of paroxysms at the final recording day. They could most likely have benefitted from continued monitoring to improve their treatment.

B. AUTOMATIC PAROXYSM DETECTION

Based on the optimal setup we were able to detect 98.4% of all paroxysms with only 0.23 false detections per hour corresponding to 5.5 false detections per 24 hours, see Table 2.




FIGURE 4. Relative number of paroxysms per 24-hours (normalized by maximum day) in relation to study day. The number of paroxysms decline during the study for all subject. For five out of six patients this is most likely explained by an increase in AED during the trial. By multiple recordings on the patients, the physician has a better chance of getting the AED dosage right.

TABLE 2. Results.

Chan	Pt. #	Sensitivity %	FDR /h	Specificity %	PPV %
F7-Fp1	1	96.2	0.225	100.0	81.4
	2	96.4	0.015	100.0	98.3
	3	87.5	0.000	100.0	100.0
	4	98.7	0.203	100.0	94.3
	5	100.0	0.139	100.0	71.7
	6	99.4	0.509	100.0	76.7
	Mean	98.4	0.230	100.0	87.1
TP7-Fp1	1	88.5	0.225	100.0	82.0
	2	98.2	0.278	100.0	78.5
	3	87.5	0.000	100.0	100.0
	4	98.0	0.544	100.0	85.6
	5	100.0	0.232	100.0	72.9
	6	93.0	0.324	100.0	81.9
	Mean	96.0	0.330	100.0	83.5

FDR: False Detection Rate
PPV: Positive Predictive Value

The double inter-electrode distance TP7-Fp1 performed with a sensitivity of 96.0% and FDR of 0.33/h corresponding to 7.9 per 24 hours. Due to the large amount of data and only few windows being classified as false positives; the specificity was 100.0% for all patients on both channels.

An analysis of the false detection showed various artifacts primarily with low frequency contents. These were rather simple for a trained expert to distinguish from true paroxysms in a visual evaluation.

C. PATIENT AND PARENT PERCEPTION

For every hour of the four 24h-EEG-recording parents were asked to note the number of minutes they were attentive to a degree which would let them detect all seizures. According to these estimates parents should have observed a median

of 8.75 seizures (range: 2-22) in each patient. The fraction they actually observed was only 4.7% of the possible (range: 0-100%). The patient in which all seizures were detected was only observed intensely for 30 minutes in which 2 out of 2 seizures were observed.

The semi structured interviews revealed a number of issues. One parent wanted to know if the child had seizures the parents did not see. Another parent reported that she was concerned about how seizures affected the brain. One family stated “Seizures are associated with many conflicts and emotions. Our child has lost all self-confidence due to the diagnosis”. A mother said “I lost my husband in a traffic accident 5 years ago. The fear of losing again is huge”. Generally, patients and parents were happy that they chose to participate and felt they had obtained a better understanding of their child’s condition during the study. Some patients were uncomfortable wearing the device in public places free for everyone to watch. One parent told us that now the child is doing much better and does not have any seizures (which was only partly true).

IV. DISCUSSION

We have acquired a novel dataset of EEG from six children with typical absence epilepsy with four 24-h recordings over one month. Both the duration of the whole period of investigation and the duration of the individual recordings are much longer than what is normally used in the clinic.

When patients, relatives, and healthcare personnel choose a device for seizure detection a number of issues matter. Size, obtrusiveness, ease of use and price are what usually matters most to the patient when acquiring the device. However, the sensitivity and false detection rate are paramount. Parameters of 100% sensitivity and false detection rate of 0/h are difficult to obtain. What is acceptable depends on the clinical situation. If the application is a system made for alarming relatives or health care personal, a system giving more false alarms than true would be annoying and the patient would probably stop using it. On the other hand, if the system is made as a decision support system, where the physician is pointed to important parts of the signal where there might be ictal activity, it is of no worry that there are some false positives as long as the sensitivity is also high. In discussion with two pediatricians (SG and CRP), we agreed that a sensitivity of 90% and false detection rate of 1/h seems clinically acceptable.

Data were analyzed retrospectively in a patient-specific manner only possible offline. This means that the procedure is not directly applicable to the clinic. However, with this study we have demonstrated that automatic paroxysm detection is feasible. We furthermore believe that it generalizes well as the amount of data is exceptionally high. As the pathological topography is quite well defined for absence seizures and the patient group is in a confined age group, the inter-subject variability is expected to be relatively low. Before it can be used as a decision support system in the clinic a more generic method has to be developed. This can either be done by

development of a one-fits-all algorithm, or by development of a patient-specific algorithm generic across time. The latter would require the expert reviewer to score a certain amount of EEG and number of paroxysms before the algorithm fits the specific patient. For future work we will obtain a much larger dataset to investigate the possibility to train a generic algorithm as well as focus on the performance if only the first day is used for training in a patient-specific setting.

Clinically satisfying sensitivity and false detection rate could be obtained in 5 out of 6 patients. This implies that for the majority of patients the method could be useful as a clinical tool.

Thirty different models were constructed for each patient to find the optimal balance between sensitivity and FDR. A large proportion of the models showed similar performance levels. This makes us confident that the optimized model is robust. When comparing the performance to that of similar models in the literature our models seem quite promising [9], [22].

At first glance, 5.5 false detections per 24 hours might sound like a lot, but considering the high frequency of the absences; the positive predictive value is 87.1%. This means that the vast majority of detections are true. If a clinician should use the algorithm as a decision support system, it would be possible to present all the detected paroxysms while only 12.9% of displays would show false detections which could be rejected manually.

When the distance between two EEG-electrodes is increased, the amplitude of the recorded signal is generally increased; however, the noise level may also be affected. We therefore recorded two bipolar channels allowing for normal and double inter-electrode distance within the left fronto-temporal region based on the 10-20 system.

Sensitivity and FDR of the short inter-electrode distance channel, F7-Fp1, were superior to the long inter-electrode distance of TP7-Fp1. This might be due to multiple reasons: First of all, the TP7 electrode is placed near the temporalis muscle which generates EMG artifacts. Chewing artifacts are noticeable in the temporal region showing a high frequency burst followed by a slow frequency pattern resembling the spike-wave pattern of the absences. Furthermore, it was observed on channel TP7-Fp1 that some of the absence paroxysms actually showed amplitudes above the 800 μ Vpp dynamic range of the recording system. This resulted in signal cutting with a putative effect on the computed features.

The quality of single channel EEG-recordings has improved significantly in recent years. Several researchers have looked into the use of single channel EEG-devices for various clinical purposes. Generally there are positive results both in sleep monitoring with detection of sleep stages [30] and in seizure detection [9]. Thus development of less obtrusive single-channel EEG monitoring devices for clinical use seems realistic.

Clinical epileptologists often state “*we treat the patient not the EEG*”. However in some syndromes like CAE and JAE it probably does make sense to reduce the number and duration of paroxysms, i.e. treating the EEG. We have demonstrated

a strong correlation between level of AED and the number of paroxysms. Furthermore, when the number of paroxysms decreased, parents also reported that the child was doing better. Thus treating the EEG in this case may be beneficial.

REFERENCES

- [1] K. Heo, S.-D. Han, S. R. Lim, M. A. Kim, and B. I. Lee, “Patient awareness of complex partial seizures,” *Epilepsia*, vol. 47, no. 11, pp. 1931–1935, Nov. 2006.
- [2] J. S. Ebersole and R. F. Leroy, “A direct comparison of ambulatory cassette and intensive inpatients EEG monitoring in detecting interictal abnormalities,” *Electroencephalogr. Clin. Neurophysiol.*, vol. 53, no. P21, 1982.
- [3] D. Looney *et al.*, “The in-the-ear recording concept: User-centered and wearable brain monitoring,” *IEEE Pulse*, vol. 3, no. 6, pp. 32–42, Nov. 2012.
- [4] J. R. Ives and J. F. Woods, “4-Channel 24 hour cassette recorder for long-term EEG monitoring of ambulatory patients,” *Electroencephalogr. Clin. Neurophysiol.*, vol. 39, no. 1, pp. 88–92, Jul. 1975.
- [5] J. R. Ives and J. F. Woods, “A study of 100 patients with focal epilepsy using a 4-channel ambulatory cassette recorder,” in *Proc. Third Int. Symp. Ambulatory Monitor.*, 1980, pp. 383–392.
- [6] H. J. Faulkner, H. Arima, and A. Mohamed, “The utility of prolonged outpatient ambulatory EEG,” *Seizure*, vol. 21, no. 7, pp. 491–495, Sep. 2012.
- [7] M. R. de Feo, O. Mecarelli, G. Ricci, and M. F. Rina, “The utility of ambulatory EEG monitoring in typical absence seizures,” *Brain Develop.*, vol. 13, no. 4, pp. 223–227, Jul. 1991.
- [8] J. E. Sullivan and D. J. Dlugos, “Idiopathic generalized epilepsy,” *Current Treat. Opt. Neurol.*, vol. 6, no. 3, pp. 231–242, May 2004.
- [9] J. Duun-Henriksen, R. E. Madsen, L. S. Remvig, C. E. Thomsen, H. B. D. Sorensen, and T. W. Kjaer, “Automatic detection of childhood absence epilepsy seizures: Toward a monitoring device,” *Pediatric Neurol.*, vol. 46, no. 5, pp. 287–292, 2012.
- [10] A. J. Rowan, J. W. A. Meijer, N. de Beer-Pawlikowski, P. van der Geest, and H. Meinardi, “Valproate-ethosuximide combination therapy for refractory absence seizures,” *Arch. Neurol.*, vol. 40, no. 13, pp. 797–802, Dec. 1983.
- [11] E. Perucca and K. J. Meador, “Adverse effects of antiepileptic drugs,” *Acta Neurol. Scand.*, vol. 121, no. 181, pp. 30–35, 2005.
- [12] J. R. Tenney and T. A. Glauser, “The current state of absence epilepsy: Can we have your attention?” *Epilepsy Currents*, vol. 13, no. 3, pp. 40–135, May 2013.
- [13] L. G. Sadleir, K. Farrell, S. Smith, M. B. Connolly, and I. E. Scheffer, “Electroclinical features of absence seizures in childhood absence epilepsy,” *Neurology*, vol. 67, no. 3, pp. 413–418, Aug. 2006.
- [14] M. J. Keilson, W. A. Hauser, J. P. Magrill, and J. Tepperberg, “Ambulatory cassette EEG in absence epilepsy,” *Pediatric Neurol.*, vol. 3, no. 5, pp. 273–276, Oct. 1987.
- [15] J. E. Mendizabal and W. J. Nowack, “Transitory cognitive impairment with brief generalized spike-wave paroxysm: A clinical counterexample to the three-second rule,” *Clin. Electroencephalogr.*, vol. 27, no. 4, pp. 215–217, Oct. 1996.
- [16] G. L. Holmes, M. McKeever, and M. Adamson, “Absence seizures in children: Clinical and electroencephalographic features,” *Ann. Neurol.*, vol. 21, no. 3, pp. 268–273, Mar. 1987.
- [17] I. Westmijse, P. Ossenblok, B. Gunning, and G. Van Luijckelaar, “Onset and propagation of spike and slow wave discharges in human absence epilepsy: A MEG study,” *Epilepsia*, vol. 50, no. 12, pp. 2538–2548, Dec. 2009.
- [18] J. R. Carrie and J. D. Frost, Jr., “Clinical evaluation of a method for quantification of generalized spike-wave EEG patterns by computer during prolonged recordings,” *Comput. Biomed. Res.*, vol. 10, no. 5, pp. 449–457, Oct. 1977.
- [19] B. L. Ehrenberg and J. K. Penry, “Computer recognition of generalized spike-wave discharges,” *Electroencephalogr. Clin. Neurophysiol.*, vol. 41, no. 1, pp. 25–36, Jul. 1976.
- [20] D. J. Koffler and J. Gotman, “Automatic detection of spike-and-wave bursts in ambulatory EEG recordings,” *Electroencephalogr. Clin. Neurophysiol.*, vol. 61, no. 2, pp. 165–180, Aug. 1985.
- [21] J. Principe and J. R. Smith, “Microcomputer-based system for the detection and quantification of petit mal epilepsy,” *Comput. Biol. Med.*, vol. 12, no. 2, pp. 87–95, Dec. 1982.

- [22] P. Xanthopoulos *et al.*, "A novel wavelet based algorithm for spike and wave detection in absence epilepsy," in *Proc. IEEE Int. Conf. Bioinf. Bioeng.*, May 2010, pp. 14–19.
- [23] J. Duun-Henriksen, H. B. D. Sørensen, T. W. Kjaer, and C. E. Thomsen, "Detection and prediction of epileptic seizures," Ph.D. dissertation, Dept. Elect. Eng., Tech. Univ. Denmark, Kongens Lyngby, Denmark, 2012.
- [24] P. Bloomfield, *Fourier Analysis of Time Series: An Introduction*, 2nd ed. Hoboken, NJ, USA: Wiley, 2000.
- [25] Z. Yongli, Z. Yungui, T. Weiming, and C. Hongzhi, "An improved feature selection algorithm based on MAHALANOBIS distance for network intrusion detection," in *Proc. Int. Conf. Sensor Netw. Secur. Technol. Privacy Commun. Syst.*, May 2013, pp. 69–73.
- [26] A. Shoeb, H. Edwards, J. Connolly, B. Bourgeois, S. T. Treves, and J. Guttag, "Patient-specific seizure onset detection," *Epilepsy Behavior*, vol. 5, no. 4, pp. 483–498, Aug. 2004.
- [27] J. Henriksen *et al.*, "Automatic seizure detection: Going from sEEG to iEEG," in *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc.*, Aug. 2010, pp. 2431–2434.
- [28] S.-J. Yen and Y.-S. Lee, "Cluster-based under-sampling approaches for imbalanced data distributions," *Expert Syst. Appl.*, vol. 36, no. 3, pp. 5718–5727, Apr. 2009.
- [29] C.-C. Chang and C.-J. Lin, "LIBSVM: A library for support vector machines," *ACM Trans. Intell. Syst. Technol.*, vol. 2, no. 3, pp. 27:1–27:27, 2011.
- [30] Y. Wang, K. A. Loparo, M. R. Kelly, and R. F. Kaplan, "Evaluation of an automated single-channel sleep staging algorithm," *Nature Sci. Sleep*, vol. 7, pp. 101–111, Jan. 2015.



HELGE B. D. SØRENSEN (M'90) received the M.S.E.E. and Ph.D. degrees in electrical engineering from the Institute of Electronic Systems, Aalborg University, in 1985 and 1992, respectively. He was a Research Assistant with Aalborg University. From 1989 to 1993, he was an Assistant Professor with the Institute of Electronic Engineering, Aalborg University. From 1993 to 1995, he was an Associate Professor with the Engineering Academy Denmark. Since 1995, he has been an Associate Professor, and the Head of the Biomedical Signal Processing Research Group under Biomedical Engineering with the Department of Electrical Engineering, Technical University of Denmark.

SABINE GROENBORG was born in 1971. She received the Medical degree in 1999. She is a Board Certified Pediatrician and Pediatric Neurologist. She is currently a Consultant with the Center for Rare Disease, Department of Clinical Genetics, Juliane Marie Center, Rigshospitalet University Hospital Copenhagen.



CHARLOTTE R. PEDERSEN was born in 1964. She received the M.D. and Ph.D. degrees in medicine from the University of Copenhagen in 1992 and 2004, respectively. She is a Board certified Pediatrician. From 2005 to 2008, was a Resident in pediatrics, and from 2009 to 2011 as a Trainee in child neurology with different hospitals in the Region of Copenhagen. Since 2012, she has been a Consultant with The Department for Children and Youth, Nordzealand Hospital, Hillerød.



JONAS DUUN-HENRIKSEN received the M.S. degree in biomedical engineering and industrial Ph.D. degrees from DTU Electrical Engineering, Technical University of Denmark, in 2008 and 2013, respectively. His Ph.D. was in collaboration with the company Hypo-Safe A/S. He is currently with Hypo-Safe A/S as a Scientific Researcher, with special responsibility within epilepsy research and EEG data quality.



TROELS W. KJÆR was born in 1965. He received the Medical Degree and the Ph.D. degree, on single unit recording of monkey brain primary visual cortex, from the University of Copenhagen, in 1992 and 1996, respectively. He is a Board certified Clinical Neurophysiologist. He was an IRTA fellow with the NIMH, Bethesda Maryland, as a Senior Researcher with the John F. Kennedy Institute, Denmark, and as a Chief Physician with Rigshospitalet University Hospital. He is currently a Professor in charge of Clinical Neurophysiology and the Head of Research with Zealand University Hospital, Denmark. He is also an External Associate Professor with the University of Copenhagen. His main research interest is brain monitoring and plasticity in normals and patients, and the aspects of consciousness and full time range brain monitoring from milliseconds to years.