

Letter to the Editor

Response to: Comment on “The Effect of Cochlear Size on Cochlear Implantation Outcomes”

Jafri Kuthubutheen ¹, **Amandeep Grewal**², **Sean Symons**^{2,3}, **Julian Nedzelski**^{2,3},
David Shipp^{2,3}, **Vincent Lin**^{2,3} and **Joseph Chen**^{2,3}

¹School of Surgery, University of Western Australia, Perth, Australia

²Sunnybrook Health Sciences Centre, Toronto, Canada M4N3M5

³University of Toronto, Toronto, Canada M5S 1A1

Correspondence should be addressed to Jafri Kuthubutheen; Jafri.Kuthubutheen@health.wa.gov.au

Received 21 June 2020; Accepted 15 July 2020; Published 27 March 2021

Academic Editor: Johan H.M. Frijns

Copyright © 2021 Jafri Kuthubutheen et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

We thank the authors Eser et al. for their comments [1] on our manuscript entitled “The Effect of Cochlear Size on Cochlear Implantation Outcomes” [2]. We agree that the effect of cochlear size on cochlear outcomes is very topical and that much research still needs to be done in order to determine the relationship between the two variables.

As we discussed in our paper, there are a number of possible explanations for the correlation between cochlear size and speech outcomes only seen with the shorter electrode. These include factors related to insertion trauma and postoperative electrode movement. We also do not believe that absolute spiral ganglion counts or tonotopy is the reason for this observed correlation. However, this does not mean that these are not factors contributing to this correlation. For example, spiral ganglion *distribution* rather than absolute spiral ganglion counts may be important, particularly in those patients with better residual hearing function (preoperative consonant-nucleus-consonant (CNC) word score) which was the case in our group who received the shorter electrode.

We decided in our study not to measure the apical turn of the cochlea because we felt that the challenges of obtain-

ing an accurate result using our described technique would outweigh the benefits of obtaining this data. We agree that understanding the apical anatomy of the cochlea is important, particularly when using lateral wall electrodes.

The authors correctly pointed out the error in Table 3 where the values of the outer wall length to 360 degrees should actually be those for the outer wall length to 720 degrees and vice versa. This has now been corrected [3].

The p values can be derived statistically from the correlation coefficient and the number of pairs. In this case, there are 21 pairs of data for the Flex 28 group. Therefore, the correlation between Ac and CNC (%) is 0.64 which has a p value of 0.00178, and the correlation between outer wall length and CNC (%) is 0.71 which has a p value of 0.000311. Similarly, the correlation between Ac and AZBIO (%) is 0.46 which has a p value of 0.03589, and the correlation between outer wall length and AZBIO (%) is 0.47 which has a p value of 0.031561.

Conflicts of Interest

The authors have no conflicts of interest to disclose.

References

- [1] M. Bilgin Eser, B. Atalay, and M. T. Kalcioğlu, “Comment on “The Effect of Cochlear Size on Cochlear Implantation Outcomes”,” *BioMed Research International*, vol. 2020, Article ID 5674547, 2 pages, 2020.
- [2] J. Kuthubutheen, A. Grewal, S. Symons et al., “The effect of cochlear size on cochlear implantation outcomes,” *BioMed Research International*, vol. 2019, Article ID 5849871, 8 pages, 2019.
- [3] J. Kuthubutheen, A. Grewal, S. Symons et al., “Corrigendum to “The Effect of Cochlear Size on Cochlear Implantation Outcomes.”” *BioMed Research International*,” vol. 2020, Article ID 6407456, 2 pages, 2020.