

POSTER PRESENTATION

Open Access


# Decoding motor intent from simulated multiple longitudinal intrafascicular electrode recordings

Mohamed Abdelghani<sup>1\*</sup>, James Abbas<sup>2</sup>, Kenneth Horch<sup>1</sup>, Ranu Jung<sup>1</sup>

From Twenty Second Annual Computational Neuroscience Meeting: CNS\*2013  
Paris, France. 13-18 July 2013

Signals recorded from peripheral nerves may provide an effective and reliable means of controlling powered prosthetic limbs. Longitudinal intrafascicular electrodes (LIFEs) have been used to record extracellular motor activity from peripheral nerves in upper-limb amputees for periods up to several weeks and the ability to decode the activity and use it for single degree-of-freedom (DOF) control of a prosthetic arm has been demonstrated [1]. However, simultaneous control of multiple DOFs of the prosthesis, which is important for many daily tasks, presents additional challenges. Recently we

developed a platform to simulate recording of extracellular motor activity from multiple LIFE electrodes [2]. We have also designed and tested an online decoding algorithm that utilizes these simulated recordings. Figure 1A&B shows the schematic of the decoder structure. The decoder is composed of multiple single channel decoders (SCDs) and a demixer. The SCD decodes motor intent from a LIFE recording. It is composed of a bandpass filter to attenuate noise and sharpen spikes, a clipping function to identify spikes and a half-Gaussian smoothing kernel to get a smoothed real-time estimate



\* Correspondence: mabdelgh@fiu.edu

<sup>1</sup>Department of Biomedical Engineering, Florida International University, Miami, FL, 33174, USA

Full list of author information is available at the end of the article

of motor intent. The demixer identifies the motor intent signals as corresponding to a particular motion class, such as wrist flexion, supination etc. The demixer requires a learning stage, where recordings from LIFEs are correlated to motion classes. A simple batched LMS algorithm is used to train the parameters of the demixer. Figure 1C, display the result of a single channel decoding of a sinusoidal motor intent. Figure 1D shows results of demixing of two overlapped motor intents recorded by two LIFEs: the first electrode records activity from the two motor pools while the second electrode records activity from only one (not shown). During the learning stage, the demixer learns to account for common motor intent and provide good estimates of the two different motor intent signals. The decoder developed here could be readily implemented in a real-time, portable, low-power configuration to translate multiple LIFE recordings to motor intent signals that enable multi-DOF control of a powered prosthesis.

#### Acknowledgements

Supported by DARPA (MTO) under the auspices of Dr. Jack Judy, Grant/Contract N66001-12-C-4195.

#### Author details

<sup>1</sup>Department of Biomedical Engineering, Florida International University, Miami, FL, 33174, USA. <sup>2</sup>School of Biological and Health Systems Engineering, Arizona State University, Tempe, AZ, 85287, USA.

Published: 8 July 2013

#### References

1. Dhillon GS, Lawrence SM, Hutchinson DT, Horch KW: Residual function in peripheral nerve stumps of amputees: implications for neural control of artificial limbs. *J Hand Surg Am* 2004, **29**(4):605-615.
2. Abdelghani MN, Abbas JJ, Horch K, Jung R: A computational model to simulate neural recordings from longitudinal intrafascicular electrodes. *Society for Neuroscience Conference* 2012, 584, 20/SS19.

doi:10.1186/1471-2202-14-S1-P201

Cite this article as: Abdelghani et al.: Decoding motor intent from simulated multiple longitudinal intrafascicular electrode recordings. *BMC Neuroscience* 2013 14(Suppl 1):P201.

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at  
[www.biomedcentral.com/submit](http://www.biomedcentral.com/submit)

