

POSTER PRESENTATIONS

Open Access

Evaluation of the Clearview[®] Malaria pLDH as malaria rapid diagnostic test in travellers

Sandrine Houzé^{1*}, Véronique Hubert¹, Pressler D Cohen², Baruch Rivetz², Jacques Le Bras¹

From Parasite to Prevention: Advances in the understanding of malaria
Edinburgh, UK. 20-22 October 2010

Background

Malaria Rapid Diagnostic Tests (RDTs) are widely used to diagnose malaria. The present prospective study evaluated a new RDT, the Clearview[®] Malaria pLDH test targeting the pan-*Plasmodium* antigen lactate dehydrogenase (pLDH).

Methods

The Clearview[®] Malaria pLDH test was evaluated on fresh samples obtained in returned international travellers using microscopy corrected by PCR as the reference method. Included samples were *P. falciparum* (139), *P. vivax* (22), *P. ovale* (20), *P. malariae* (7), and 102 negative.

Results

Overall sensitivity for the detection of *Plasmodium sp.* was 93.2%. For *P. falciparum*, the sensitivity was 98.6%; for *P. vivax*, *P. ovale* and *P. malariae*, overall sensitivities were 90.9%, 60.0% and 85.7% respectively. For *P. falciparum* and for *P. vivax*, the sensitivities increased to 100% at parasite densities above 100/μl. The specificity was 100%. The test was easy to perform and the result was stable for at least 1 hour.

Conclusion

The Clearview[®] Malaria pLDH was efficient for the diagnosis of malaria. The test was very sensitive for *P. falciparum* and *P. vivax* detection. The sensitivities for *P. ovale* and *P. malariae* were better than other RDTs [1].

Author details

¹Laboratory of Parasitology-French Malaria Reference Center, AP-HP Bichat Hospital, Paris, 75018, France. ²Organics Ltd, Yavne, 70650, Israel.

¹Laboratory of Parasitology-French Malaria Reference Center, AP-HP Bichat Hospital, Paris, 75018, France

Full list of author information is available at the end of the article

Published: 20 October 2010

Reference

1. Maltha J, Gillet P, Bottieau E, Cnops L, van Esbroeck M, Jacobs J: Evaluation of a rapid diagnostic test (CareStrat TM Malaria HRP2/pLDH (Pf/pan) Combo Test) for the diagnosis of malaria in a reference setting. *Malaria Journal* 2010, **9**:171.

doi:10.1186/1475-2875-9-S2-P15

Cite this article as: Houzé *et al.*: Evaluation of the Clearview[®] Malaria pLDH as malaria rapid diagnostic test in travellers. *Malaria Journal* 2010 **9**(Suppl 2):P15.

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit

