



Application of Artificial Intelligence in the Detection and Characterization of Colorectal Neoplasm

Kyeong Ok Kim¹ and Eun Young Kim²

¹Division of Gastroenterology and Hepatology, Department of Internal Medicine, Yeungnam University College of Medicine, and
²Division of Gastroenterology and Hepatology, Department of Internal Medicine, Daegu Catholic University School of Medicine, Daegu, Korea

Article Info

Received June 12, 2020

Accepted June 28, 2020

Published online August 12, 2020

Corresponding Author

Eun Young Kim

ORCID <https://orcid.org/0000-0003-3965-9964>

E-mail kimey@cu.ac.kr

Endoscopists always have tried to pursue a perfect colonoscopy, and application of artificial intelligence (AI) using deep-learning algorithms is one of the promising supportive options for detection and characterization of colorectal polyps during colonoscopy. Many retrospective studies conducted with real-time application of AI using convolutional neural networks have shown improved colorectal polyp detection. Moreover, a recent randomized clinical trial reported additional polyp detection with shorter analysis time. Studies conducted regarding polyp characterization provided additional promising results. Application of AI with narrow band imaging in real-time prediction of the pathology of diminutive polyps resulted in high diagnostic accuracy. In addition, application of AI with endocytoscopy or confocal laser endomicroscopy was investigated for real-time cellular diagnosis, and the diagnostic accuracy of some studies was comparable to that of pathologists. With AI technology, we can expect a higher polyp detection rate with reduced time and cost by avoiding unnecessary procedures, resulting in enhanced colonoscopy efficiency. However, for AI application in actual daily clinical practice, more prospective studies with minimized selection bias, consensus on standardized utilization, and regulatory approval are needed. (*Gut Liver* 2021;15:346-353)

Key Words: Polyp; Colon; Colonoscopy; Artificial intelligence; Convolutional neural network

INTRODUCTION

The incidence of colorectal cancer (CRC), the third most common cancer worldwide,¹ has been steadily increasing in the Republic of Korea recently. Because most of the CRC arise from adenomas, detection and complete removal of these precancerous lesions can reduce the incidence and mortality associated with CRC.^{2,3} However, for the effective prevention of CRC, high-quality colonoscopy that detects all the polyps, is a prerequisite. Adenoma detection rate (ADR) has been considered as one of the important quality indicators of colonoscopy, and the inverse association between ADR and incidence of interval CRC has been established.^{4,5} Because colonoscopy may not be perfect, many efforts to enhance ADR have been made.⁶

Artificial intelligence (AI), one of the promising technologies, can mitigate the shortcomings of colonoscopy. As it is easy to obtain polyp images to provide enough data for AI training, many studies using this technology have been reported lately.⁷ Well-trained AI modelling can increase polyp identification and optical diagnosis. Especially, its use can increase the detection rate of polyps in the right colon, where a higher miss rate is understandably anticipated with conventional colonoscopy.⁸

In this review, we aim to contemplate the current status and future directions of AI applications, including technologies, applications, efficacy, and unmet needs, regarding colorectal neoplasm.



COMPUTER-AIDED DIAGNOSIS AND DEEP-LEARNING FOR DETECTION AND CHARACTERIZATION OF THE COLORECTAL POLYPS

AI, having machine intelligence that is different from the natural intelligence displayed by humans and other animals,⁹ can learn and solve problems.¹⁰ Earlier, machine learning (ML) had been the main focus of AI. A computer-vision algorithm, developed by computer scientists based on earlier ML research for the detection of colorectal polyps was “hand-crafted,” based on the features adopted by the designers.¹¹ In other words, human efforts or instructions were needed for extracting features, such as color, shape, or texture of polyps, because the earlier ML could learn only the classification of the extracted images.¹² Although the “hand-crafted” algorithm showed high accuracy, the risk of missing the lesions without those extracted features or obtaining false positive results was of concern because it was designed to detect the lesions with certain features chosen by designers. In addition, actual clinical application was limited due to the difference in image quality and also slow processing time.¹¹

Deep-learning (DL) algorithm, one of the subtype of ML introduced in the 1980s, could overcome the limitation of earlier ML by combining both the extraction and classification of image features using deep neural networks (DNN).¹³ The innovative method of DNN gained significant attention because of self-learning capability of DL that could automatically identify polyp and non-polyp features from the huge dataset, instead of capturing specific features of a polyp using several networks. Self-extraction, the key feature, was achieved using backpropagation algorithm and changing the internal parameters of each neural network layer.¹⁴

Among the variable classes of DNN for image and video applications, convolutional neural network (CNN), the most popular method, can carry out layers of convolutions and can completely connect layers to unite all features in the final outcome.^{15,16} Provided with sufficient annotated data, CNNs can be trained to describe, in detail, what they see and discriminate polyps from non-polypoid lesions.¹¹ Currently, AI can be trained with enough input data, thanks to the easy accessibility of big data, aiding rapid progress in the research and application of AI in colonoscopic polyp detection and characterization.¹⁷

STUDIES RELATED TO APPLICATION OF AI IN COLONOSCOPY

1. AI for detection of colorectal polyp

Since the initial study regarding computer-aided detection based on “hand-crafted” data,¹⁸ the performance of AI has improved, especially with the introduction of DL, for colorectal neoplasm detection.^{19,20} Table 1 shows the summary of the clinical studies of AI for the detection of colorectal polyps.¹⁸⁻²⁶

Urban *et al.*²⁰ reported the first real-time application. They initially pretrained AI using ImageNet and then trained the deep CNNs. The algorithm was tested with multiple sets of colonoscopic images and 11 challenging videos sets. The result was very promising with 97% sensitivity, 95% specificity, and 96% overall accuracy. For the practical utilization in real-world, the CNN-assisted video was reviewed by experts. During the index colonoscopy, 28 polyps were noted in nine standard colonoscopic videos. Experts could identify 36 polyps without and 45 polyps with CNN assistance. The additional 17 polyps identified with CNN assistance were not larger than 10 mm in size. Further, the algorithm was faster than the real-time analysis by endoscopists (10 ms/frame vs 33–40 ms/frame). Meanwhile, Yu *et al.*²⁴ developed a novel three-dimensional CNN algorithm that could learn more representative spatiotemporal features and improve performance of automated polyp detection.

Wang *et al.*²¹ developed a CNN system that processed data in real-time with a 77 millisecond delay on monitor, in which, a blue box appeared around the area of the polyp upon its detection, along with an alarm. They randomized the colonoscopy of about 1,058 patients into those using this system and the standard one and proved an increased ADR in AI (29.1% vs 20.3%, respectively) with minimal delay in examining time. However, the robust detection rate was limited to polyps smaller than 10 mm in size, and 43.6% of the polyps were hyperplastic polyps with low malignant potential.^{27,28} The most recent meta-analysis of three randomized controlled trial of AI-assisted colonoscopy reported 32.9% of ADR (20.8% in standard colonoscopy, risk ratio [RR]=1.58, $p<0.001$) and 43.0% of polyp detection rate (27.8% in standard colonoscopy, RR=1.55, $p<0.001$).²⁹ AI utility in screening colonoscopy to improve ADR looks optimistic.

2. AI for characterization of colorectal neoplasm

Accurate histologic diagnosis of the colorectal polyp before resection is desired because endoscopic resection and pathologic evaluation of lesions with very low risk of malignant potential may result in waste of time and cost. On

Table 1. Clinical Studies of Artificial Intelligence for the Detection of Colorectal Polyps

Author (year)	Study design	Algorithm type	Dataset	Processing time	Results
Wang <i>et al.</i> (2019) ²¹	Randomized controlled study	Convolutional neural network	5,545 Images	25 fps with 77 ms latency	9% Increase of ADR
Klare <i>et al.</i> (2019) ²²	Prospective <i>In vivo</i>	Convolutional neural network	55 Live colonoscopies	50 ms latency	Sensitivity 75%/polyp ADR 29% [31% in endoscopist]
Urban <i>et al.</i> (2018) ²⁰	Retrospective <i>Ex vivo</i>	Convolutional neural network	Image dataset: 8,641 polyps Video: 20 colonoscopies	10 ms/frame (real-time)	Image dataset: accuracy 96.4% AUROC 0.991
Misawa <i>et al.</i> (2018) ¹⁹	Retrospective <i>Ex vivo</i>	Convolutional neural network	135 Video clips	No description	Sensitivity 90% Specificity 63.3% Accuracy 76.5%
Zhang <i>et al.</i> (2017) ²³	Retrospective <i>Ex vivo</i>	Convolutional neural network	150 Random+30 NBI images	No description	Sensitivity 98% PPV 99% AUROC 1.00
Yu <i>et al.</i> (2017) ²⁴	Retrospective <i>Ex vivo</i>	Convolutional neural network	ASU-Mayo 20 videos	1.23 s/frame	Sensitivity 7% PPV 88%
Angermann <i>et al.</i> (2017) ²⁵	Retrospective <i>Ex vivo</i>	Hand-crafted	No description	20–185 ms 0.3–1.8 s delay	Sensitivity 100%/polyp PPV50%
Tajbakhsh <i>et al.</i> (2015) ²⁶	Retrospective <i>Ex vivo</i>	Hand-crafted	No description	2.6 s/frame	Sensitivity 48% on proprietary database Sensitivity 88% in CVC-colon DB
Karkanis <i>et al.</i> (2003) ¹⁸	Retrospective <i>Ex vivo</i>	Hand-crafted	180 Still images	1.5 m/video	Sensitivity 94% Specificity 99%

ADR, adenoma detection rate; AUROC, area under the receiver operating characteristics; NBI, narrow band imaging; PPV, positive predictive value; ASU, Arizona State University; CVC, computer vision center; DB, data base.

the other hand, incomplete resection of lesions with high risk of malignancy should be avoided. Therefore, various advanced endoscopic techniques with image enhancement for optical diagnosis have been introduced, and various studies concerning the application of AI to these endoscopic systems have also been steadily reported to date (Table 2).

Narrow band imaging (NBI; Olympus Corp., Tokyo, Japan), an image-enhanced endoscopy used for the observation of microstructure or capillaries of colorectal neoplasm, could discriminate each polyp well, based on the pit and vascular pattern.³⁰ Tischendorf *et al.*³¹ and Gross *et al.*³² investigated AI application to NBI (excluding its performance enhancement) for the characterization of colon polyps. They both extracted nine vessel features from the NBI images in a similar manner and sorted them into neoplastic and non-neoplastic lesions using a support vector machine (SVM), a discriminated ML model.¹⁵ The diagnostic accuracies of the two studies were 85.3% and 93.1%, respectively. In addition, Gross *et al.*³² proved superior accuracy of AI-assisted diagnosis compared to that of non-experts in differentiating colorectal polyps smaller than 10 mm in size, suggesting that AI could be a good support for the beginners of endoscopy.

Another Japanese group developed a real-time image recognition system that could achieve 93.2% accuracy in real-time prediction of diminutive polyp pathology.³³ In

addition, the follow-up recommendation based on the prediction by that model showed 92.7% consistency with Preservation and Incorporation of Valuable Endoscopic Innovations (PIVI)-2 criteria of American Society for Gastrointestinal Endoscopy for “resect-and-discard” strategy.³⁴ Resect and discard strategy can bring a substantial economic benefit.

Two groups carried out retrospective studies based on DL. Byrne *et al.*³⁵ developed CNN model using NBI video frames. Although the model could not make adequate credence of histology prediction in 15% of polyps, it could differentiate diminutive adenomas from hyperplastic polyps with 94% accuracy. The sensitivity and specificity for detecting adenoma were 98% and 83%, respectively. Chen *et al.*³⁶ also developed a similar model, a DNN-computer aided diagnosis (CAD), with 2,157 images for identification of about 284 neoplastic or hyperplastic polyps smaller than 5 mm in size, with 89.6% positive predictive value (PPV) and 91.5% negative predictive value (NPV). Consequently, among 117 tubular adenomas, DNN-CAD could diagnose high grade dysplasia with 100% sensitivity and 94% specificity. The performance of DNN-CAD, when compared to that of four endoscopists with less than 1 year of experience, was superior with a shorter procedure time and perfect intra-observer agreement (kappa score of 1). The result of the above-mentioned two studies could satisfy the PIVI-2 threshold (90% NPV for adenoma detection)

Table 2. Clinical Studies of Artificial Intelligence for Characterization of Colorectal Polyps

Author (year)	Study design	Classification target and base	Algorithm type	Image modality	Dataset	Results
Byrne <i>et al.</i> (2019) ³⁵	Retrospective	Histology of diminutive polyp	Convolutional neural network	NBI video frames	125 Diminutive polyp videos	Sensitivity 98% Specificity 83% Accuracy 94%
Chen <i>et al.</i> (2018) ³⁶	Retrospective	Neoplastic or hyperplastic polyp <5 mm	Convolutional neural network	Magnifying NBI	284 Diminutive polyps image	Sensitivity 96.3% Specificity 78.1% Accuracy 90.1%
Mori <i>et al.</i> (2018) ⁴⁴	Prospective	Diagnosis of neoplastic diminutive polyp	SVM	Endocytoscopy with NBI and stained images	466 Diminutive polyps from 325 patients	Prediction rate 98.1%
Takeda <i>et al.</i> (2017) ⁴³	Retrospective	Invasive CRC	SVM	Endocytoscopy with NBI and stained images	200 Images	Sensitivity 89.4% Specificity 98.9% Accuracy 94.1%
Kominami <i>et al.</i> (2016) ³³	Prospective	Histology	SVM with logistic regression	Magnifying NBI	118 Colorectal lesions	Sensitivity 95.9% Specificity 93.3% Accuracy 94.9%
Misawa <i>et al.</i> (2016) ⁴⁶	Retrospective	Microvascular findings	SVM	Endocytoscopy with NBI	100 Images	Sensitivity 84.5% Specificity 97.6% Accuracy 90.0%
Mori <i>et al.</i> (2015) ⁴²	Retrospective	Neoplastic changes in small polyps	Multivariate regression analysis	Endocytoscopy	176 Polyps from 152 patients	Sensitivity 92% Specificity 79.5% Accuracy 89.2%
Takemura <i>et al.</i> (2012) ⁵⁷	Retrospective	Pit pattern	SVM	Magnifying NBI	371 Images	Sensitivity 97.8% Specificity 97.9% Accuracy 97.8%
Gross <i>et al.</i> (2011) ³²	Prospective	Small colonic polyp <10 mm	SVM	Magnifying NBI	434 Polyps from 214 patients	Sensitivity 95% Specificity 90.3% Accuracy 93.1%
Tischendorf <i>et al.</i> (2010) ³¹	Prospective pilot	Vascularization features	SVM	Magnifying NBI	209 Polyps from 128 patients	Sensitivity 90% Specificity 70.2% Accuracy 85.3%
Takemura <i>et al.</i> (2010) ³⁹	Retrospective	Pit pattern	HuPAS software version 1.3	Magnifying NBI with chromoendoscopy [crystal violet]	134 Images	Accuracy 98.5%

NBI, narrow band imaging; SVM, support vector machine; CRC, colorectal cancer.

for “leave-in-place” strategy for diminutive hyperplastic polyps.³⁴

Although there’s a lack of recent further studies, CAD of pit pattern by magnifying chromoendoscopy was done by quantitative analysis of pit structure or texture analysis of endoscopic images.^{37,38} Takemura *et al.*³⁹ reported 98.5% diagnostic accuracy after automatically evaluating the area, perimeter, major/minor fit ellipse and circularity of the pit using software.

Recently, novel introduction of *in vivo* contact microscopic imaging modalities, such as endocytoscopy (H290ECI; Olympus Corp.) and confocal laser endomicroscopy (Cellvizio; Mauna Kea Technologies Inc, Paris, France), enabled real-time diagnosis of cellular images.^{40,41} Because both endocytoscopy and endomicroscopy could enhance image analysis with focused fixed-size images, they were ideal to be used in combination with the AI system. They magnified the image with 500- or 1,000-

fold power, respectively, during colonoscopy and showed diagnostic accuracy comparable to that of pathologists. The first application of AI in endocytoscopy was assessed by quantitative analysis of six nuclear features, and the accuracy for the detection of neoplastic change was 89.2%.⁴² Takeda *et al.*⁴³ trained the computer-aided ultrahigh (approximately ×400) magnification endoscopy system for the diagnosis of invasive CRC with 5,543 endocytoscopic images. This system, when assessed using 200 endocytoscopic test images, could discriminate invasive cancers with 89.4% sensitivity, 98.9% specificity, and 94.1% accuracy. Mori *et al.*⁴⁴ assessed the efficacy of an endocytoscopy-based CAD with 466 cases of diminutive polyps, and the NPV for diminutive rectosigmoid adenomas was 93.7%. They also proved that polyp diagnosis with AI, an add-on analysis, could reduce the cost of annual reimbursement for colonoscopy by 18.9%, by leaving 145 rectosigmoid diminutive polyps based on the AI support.⁴⁵ However, AI endocytos-

copy had a limitation—it needed pre-staining with crystal violet and methylene blue before the extraction of images. Misawa *et al.*⁴⁶ upgraded the system by combining endocytoscopy with NBI, eliminating the pre-staining step and resulting in 90.0% overall accuracy and 84.5% sensitivity within 0.3 seconds.

AI application with confocal endomicroscopy was assessed in several studies, in experimental setting, with promising accuracy.⁴⁷⁻⁵⁰ AI in other advanced endoscopies, including laser-induced fluorescence spectroscopy and autofluorescence endoscopy, was also evaluated retrospectively or prospectively.⁵¹⁻⁵⁷ We can expect that the use of AI in these advanced endoscopies for optical diagnosis could aid the real-time decision making of endoscopists.

3. AI for combination of detection and characterization of colorectal polyps

For an endoscopist, both polyp detection and characterization are essential in clinical practice. An ideal scenario would be an AI-assisted immediate detection and characterization of the colorectal polyp. A Japanese group developed novel technologies that included two algorithm systems—one, based on DL algorithm, for the detection of polyps in white light images, and the other, for the prediction of pathology by endocytoscopic images generated by a photograph.⁵⁸ According to the most recent study using CNN, AI system (Single Shot Multibox Detector) could detect 1,246 polyps with 92% sensitivity, 86% PPV, and 83% accuracy in polyp classification.⁵⁹ Although more studies are needed, an ideal colonoscopy for the detection and characterization of colorectal neoplasm seems to be achieved with AI assistance.

4. Prediction of prognosis

Prior to DL, SVM was a highly efficient computational tool that classified and regressed best by optimizing a hyperplane with largest functional margin.¹⁵ Ichimasa *et al.*⁶⁰ evaluated the predictive factors for lymph node metastasis from endoscopically resected T1 CRC using SVM. Their result showed better sensitivity (100%), specificity (66%), and accuracy (69%) of the AI model than most of the current guidelines. In addition, this model could reduce the unnecessary additional colectomy after endoscopic resection of T1 CRC compared to the current guidelines that lead to misdiagnosis.

UNMET NEEDS AND FUTURE PERSPECTIVES OF AI FOR COLORECTAL NEOPLASM

Although AI-assisted detection and diagnosis of

colorectal neoplasm are promising, most of the studies are retrospective and covered lesions which might have been selected with bias. Well-designed prospective studies that present more reliable data compared to the previous retrospective studies are needed. Most of the studies regarding efficacy of AI deal with polypoid lesions. However, for practical utilization, the efficacy of AI needs to be consistent irrespective of the shape of the polyps. Therefore, more studies with all types of polyps, including polypoid, depressed, or flat type, are needed because these non-polypoid type lesions are often more aggressive.⁶¹

In addition, as AI is trained with high-quality images, the system has to overcome the blurry vision, inadequate preparation status, and variable unpredictable hurdles observed in actual practice. There is a need for real-time application of CAD and randomized controlled comparative study between usage and avoidance of AI.

Previous studies, conducted in a variable design, resulted in different primary outcomes. Because AI development needed both engineers, who dealt with the software, and clinicians, who contributed to the clinical use, the outcomes and study design could be different depending on the study conductors.¹¹ Communication and collaboration between these different groups are also needed.

To apply AI in real clinical practice, a regulatory approval of AI-based decision making, the rules of which vary for each country, is an essential step. For that, we need to prove the minimal risk of treatment failure due to the misdiagnosis by AI.^{22,62} So, we need the evaluation of risk stratification for AI system through well-designed randomized clinical trials.

CONCLUSION

Because colonoscopy cannot completely prevent CRC, AI application in the field of colorectal neoplasm could be one of the promising options for enhancing the efficiency of colonoscopy. Owing to easy accessibility of big data and computer science, the AI technologies for the detection and classification of colorectal polyps have developed rapidly, with studies supporting the advantage of AI use in colonoscopy. However, obstacles, such as insufficient evidence of practical clinical usefulness, lack of consensus on standardized utilization, and need for regulatory approval, exist. For the ideal implementation of AI in actual clinical practice, comprehensive understanding of the strengths and weaknesses of the technology, qualified real-time studies, and accumulation of experience are warranted.

CONFLICTS OF INTEREST

E.Y.K. is an editorial board member of the journal but did not involve in the peer reviewer selection, evaluation, or decision process of this article. No other potential conflicts of interest relevant to this article were reported.

ORCID

Kyeong Ok Kim <https://orcid.org/0000-0001-5799-7436>
Eun Young Kim <https://orcid.org/0000-0003-3965-9964>

REFERENCES

- American Cancer Society. Cancer Facts & Figures 2019. Atlanta: American Cancer Society, 2019.
- Winawer SJ, Zauber AG, Ho MN, et al. Prevention of colorectal cancer by colonoscopic polypectomy. The National Polyp Study Workgroup. *N Engl J Med* 1993;329:1977-1981.
- Nishihara R, Wu K, Lochhead P, et al. Long-term colorectal-cancer incidence and mortality after lower endoscopy. *N Engl J Med* 2013;369:1095-1105.
- Rex DK, Schoenfeld PS, Cohen J, et al. Quality indicators for colonoscopy. *Gastrointest Endosc* 2015;81:31-53.
- Kaminski MF, Regula J, Kraszewska E, et al. Quality indicators for colonoscopy and the risk of interval cancer. *N Engl J Med* 2010;362:1795-1803.
- Castaneda D, Popov VB, Verheyen E, Wander P, Gross SA. New technologies improve adenoma detection rate, adenoma miss rate, and polyp detection rate: a systematic review and meta-analysis. *Gastrointest Endosc* 2018;88:209-222.
- Liedlgruber M, Uhl A. Computer-aided decision support systems for endoscopy in the gastrointestinal tract: a review. *IEEE Rev Biomed Eng* 2011;4:73-88.
- Kudo SE, Mori Y, Misawa M, et al. Artificial intelligence and colonoscopy: current status and future perspectives. *Dig Endosc* 2019;31:363-371.
- Wikipedia. Artificial intelligence [Internet]. San Francisco: Wikipedia Foundation, Inc.; c2020 [cited 2020 Mar 30]. Available from: https://en.wikipedia.org/wiki/Artificial_intelligence.
- Russell S, Norvig P. Artificial intelligence: a modern approach. 3rd ed. Harlow: Pearson Education Limited, 2009.
- Hoerter N, Gross SA, Liang PS. Artificial intelligence and polyp detection. *Curr Treat Options Gastroenterol* 2020;18:120-136.
- Ahmad OF, Soares AS, Mazomenos E, et al. Artificial intelligence and computer-aided diagnosis in colonoscopy: current evidence and future directions. *Lancet Gastroenterol Hepatol* 2019;4:71-80.
- LeCun Y, Bengio Y, Hinton G. Deep learning. *Nature* 2015;521:436-444.
- Takiyama H, Ozawa T, Ishihara S, et al. Automatic anatomical classification of esophagogastroduodenoscopy images using deep convolutional neural networks. *Sci Rep* 2018;8:7497.
- Yang YJ, Bang CS. Application of artificial intelligence in gastroenterology. *World J Gastroenterol* 2019;25:1666-1683.
- Yoon HJ, Kim JH. Lesion-based convolutional neural network in diagnosis of early gastric cancer. *Clin Endosc* 2020;53:127-131.
- Choi J, Shin K, Jung J, et al. Convolutional neural network technology in endoscopic imaging: artificial intelligence for endoscopy. *Clin Endosc* 2020;53:117-126.
- Karkanis SA, Iakovidis DK, Maroulis DE, Karras DA, Tzivras M. Computer-aided tumor detection in endoscopic video using color wavelet features. *IEEE Trans Inf Technol Biomed* 2003;7:141-152.
- Misawa M, Kudo SE, Mori Y, et al. Artificial intelligence-assisted polyp detection for colonoscopy: initial experience. *Gastroenterology* 2018;154:2027-2029.
- Urban G, Tripathi P, Alkayali T, et al. Deep learning localizes and identifies polyps in real time with 96% accuracy in screening colonoscopy. *Gastroenterology* 2018;155:1069-1078.
- Wang P, Berzin TM, Glissen Brown JR, et al. Real-time automatic detection system increases colonoscopic polyp and adenoma detection rates: a prospective randomised controlled study. *Gut* 2019;68:1813-1819.
- Klare P, Sander C, Prinzen M, et al. Automated polyp detection in the colorectum: a prospective study (with videos). *Gastrointest Endosc* 2019;89:576-582.
- Zhang R, Zheng Y, Mak TW, et al. Automatic detection and classification of colorectal polyps by transferring low-level CNN features from nonmedical domain. *IEEE J Biomed Health Inform* 2017;21:41-47.
- Yu L, Chen H, Dou Q, Qin J, Heng PA. Integrating online and offline three-dimensional deep learning for automated polyp detection in colonoscopy videos. *IEEE J Biomed Health Inform* 2017;21:65-75.
- Angermann Q, Bernal J, Sánchez-Montes C, et al. Towards real-time polyp detection in colonoscopy videos: adapting still frame-based methodologies for video sequences analysis. In: Cardoso MJ, Arbel T, Luo X, eds. Computer assisted and robotic endoscopy and clinical image-based procedures. Cham: Springer, 2017:29-41.
- Tajbakhsh N, Gurudu SR, Liang J. Automated polyp detection in colonoscopy videos using shape and context information. *IEEE Trans Med Imaging* 2016;35:630-644.

27. Noshirwani KC, van Stolk RU, Rybicki LA, Beck GJ. Adenoma size and number are predictive of adenoma recurrence: implications for surveillance colonoscopy. *Gastrointest Endosc* 2000;51(4 Pt 1):433-437.
28. Martínez ME, Baron JA, Lieberman DA, et al. A pooled analysis of advanced colorectal neoplasia diagnoses after colonoscopic polypectomy. *Gastroenterology* 2009;136:832-841.
29. Aziz M, Fatima R, Dong C, Lee-Smith W, Nawras A. The impact of deep convolutional neural network-based artificial intelligence on colonoscopy outcomes: a systematic review with meta-analysis. *J Gastroenterol Hepatol* 2020;35:1676-1683.
30. Tanaka S, Sano Y. Aim to unify the narrow band imaging (NBI) magnifying classification for colorectal tumors: current status in Japan from a summary of the consensus symposium in the 79th Annual Meeting of the Japan Gastroenterological Endoscopy Society. *Dig Endosc* 2011;23 Suppl 1: 131-139.
31. Tischendorf JJ, Gross S, Winograd R, et al. Computer-aided classification of colorectal polyps based on vascular patterns: a pilot study. *Endoscopy* 2010;42:203-207.
32. Gross S, Trautwein C, Behrens A, et al. Computer-based classification of small colorectal polyps by using narrow-band imaging with optical magnification. *Gastrointest Endosc* 2011;74:1354-1359.
33. Kominami Y, Yoshida S, Tanaka S, et al. Computer-aided diagnosis of colorectal polyp histology by using a real-time image recognition system and narrow-band imaging magnifying colonoscopy. *Gastrointest Endosc* 2016;83:643-649.
34. Rex DK, Kahi C, O'Brien M, et al. The American Society for Gastrointestinal Endoscopy PIVI (Preservation and Incorporation of Valuable Endoscopic Innovations) on real-time endoscopic assessment of the histology of diminutive colorectal polyps. *Gastrointest Endosc* 2011;73:419-422.
35. Byrne MF, Chapados N, Soudan F, et al. Real-time differentiation of adenomatous and hyperplastic diminutive colorectal polyps during analysis of unaltered videos of standard colonoscopy using a deep learning model. *Gut* 2019;68:94-100.
36. Chen PJ, Lin MC, Lai MJ, Lin JC, Lu HH, Tseng VS. Accurate classification of diminutive colorectal polyps using computer-aided analysis. *Gastroenterology* 2018;154:568-575.
37. Kudo S, Rubio CA, Teixeira CR, Kashida H, Kogure E. Pit pattern in colorectal neoplasia: endoscopic magnifying view. *Endoscopy* 2001;33:367-373.
38. Kudo SE, Mori Y, Wakamura K, et al. Endocytoscopy can provide additional diagnostic ability to magnifying chromoendoscopy for colorectal neoplasms. *J Gastroenterol Hepatol* 2014;29:83-90.
39. Takemura Y, Yoshida S, Tanaka S, et al. Quantitative analysis and development of a computer-aided system for identification of regular pit patterns of colorectal lesions. *Gastrointest Endosc* 2010;72:1047-1051.
40. Kiesslich R, Burg J, Vieth M, et al. Confocal laser endoscopy for diagnosing intraepithelial neoplasias and colorectal cancer in vivo. *Gastroenterology* 2004;127:706-713.
41. Mori Y, Kudo S, Ikehara N, et al. Comprehensive diagnostic ability of endocytoscopy compared with biopsy for colorectal neoplasms: a prospective randomized noninferiority trial. *Endoscopy* 2013;45:98-105.
42. Mori Y, Kudo SE, Wakamura K, et al. Novel computer-aided diagnostic system for colorectal lesions by using endocytoscopy (with videos). *Gastrointest Endosc* 2015;81:621-629.
43. Takeda K, Kudo SE, Mori Y, et al. Accuracy of diagnosing invasive colorectal cancer using computer-aided endocytoscopy. *Endoscopy* 2017;49:798-802.
44. Mori Y, Kudo SE, Misawa M, et al. Real-time use of artificial intelligence in identification of diminutive polyps during colonoscopy: a prospective study. *Ann Intern Med* 2018;169: 357-366.
45. Mori Y, Kudo SE, East JE, et al. Cost savings in colonoscopy with artificial intelligence-aided polyp diagnosis: an add-on analysis of a clinical trial (with video). *Gastrointest Endosc* 2020;92:905-911.e1.
46. Misawa M, Kudo SE, Mori Y, et al. Characterization of colorectal lesions using a computer-aided diagnostic system for narrow-band imaging endocytoscopy. *Gastroenterology* 2016;150:1531-1532.
47. André B, Vercauteren T, Buchner AM, Krishna M, Ayache N, Wallace MB. Software for automated classification of probe-based confocal laser endomicroscopy videos of colorectal polyps. *World J Gastroenterol* 2012;18:5560-5569.
48. Ștefănescu D, Streba C, Cârțână ET, Săftoiu A, Gruionu G, Gruionu LG. Computer aided diagnosis for confocal laser endomicroscopy in advanced colorectal adenocarcinoma. *PLoS One* 2016;11:e0154863.
49. Tafreshi MK, Linard N, André B, Ayache N, Vercauteren T. Semi-automated query construction for content-based endomicroscopy video retrieval. *Med Image Comput Comput Assist Interv* 2014;17(Pt 1):89-96.
50. Prieto SP, Lai KK, Laryea JA, Mizell JS, Muldoon TJ. Quantitative analysis of ex vivo colorectal epithelium using an automated feature extraction algorithm for microendoscopy image data. *J Med Imaging (Bellingham)* 2016;3:024502.
51. Kuiper T, Alderlieste YA, Tytgat KM, et al. Automatic optical diagnosis of small colorectal lesions by laser-induced autofluorescence. *Endoscopy* 2015;47:56-62.
52. Rath T, Tontini GE, Vieth M, Nägel A, Neurath MF, Neumann H. In vivo real-time assessment of colorectal polyp histology using an optical biopsy forceps system based on la-

- ser-induced fluorescence spectroscopy. *Endoscopy* 2016;48:557-562.
53. Renkoski TE, Banerjee B, Graves LR, et al. Ratio images and ultraviolet C excitation in autofluorescence imaging of neoplasms of the human colon. *J Biomed Opt* 2013;18:16005.
54. Arita K, Mitsuyama K, Kawano H, et al. Quantitative analysis of colorectal mucosal lesions by autofluorescence endoscopy: discrimination of carcinomas from other lesions. *Oncol Rep* 2011;26:43-48.
55. Aihara H, Saito S, Inomata H, et al. Computer-aided diagnosis of neoplastic colorectal lesions using 'real-time' numerical color analysis during autofluorescence endoscopy. *Eur J Gastroenterol Hepatol* 2013;25:488-494.
56. Inomata H, Tamai N, Aihara H, et al. Efficacy of a novel auto-fluorescence imaging system with computer-assisted color analysis for assessment of colorectal lesions. *World J Gastroenterol* 2013;19:7146-7153.
57. Takemura Y, Yoshida S, Tanaka S, et al. Computer-aided system for predicting the histology of colorectal tumors by using narrow-band imaging magnifying colonoscopy (with video). *Gastrointest Endosc* 2012;75:179-185.
58. Mori Y, Kudo SE, Misawa M, Mori K. Simultaneous detection and characterization of diminutive polyps with the use of artificial intelligence during colonoscopy. *VideoGIE* 2019;4:7-10.
59. Ozawa T, Ishihara S, Fujishiro M, Kumagai Y, Shichijo S, Tada T. Automated endoscopic detection and classification of colorectal polyps using convolutional neural networks. *Therap Adv Gastroenterol* 2020;13:1756284820910659.
60. Ichimasa K, Kudo SE, Mori Y, et al. Artificial intelligence may help in predicting the need for additional surgery after endoscopic resection of T1 colorectal cancer. *Endoscopy* 2018;50:230-240.
61. Kudo SE, Takemura O, Ohtsuka K. Flat and depressed types of early colorectal cancers: from East to West. *Gastrointest Endosc Clin N Am* 2008;18:581-593.
62. Chinzei K, Shimizu K, Mori K, et al. Regulatory science on AI-based medical devices and systems. *Adv Biomed Eng* 2018;7:118-123.