

Poster presentation

Open Access

Quantitation of cyclic dinucleotides by reversed-phase LC-MS/MS

Christian Spangler*, Roland Seifert and Volkhard Kaefer

Address: Institute of Pharmacology, Hannover Medical School, Hannover, Germany

Email: Christian Spangler* - spangler.christian@mh-hannover.de

* Corresponding author

from 4th International Conference of cGMP Generators, Effectors and Therapeutic Implications
Regensburg, Germany. 19–21 June 2009

Published: 11 August 2009

BMC Pharmacology 2009, 9(Suppl 1):P65 doi:10.1186/1471-2210-9-S1-P65

This abstract is available from: <http://www.biomedcentral.com/1471-2210/9/S1/P65>

© 2009 Spangler et al; licensee BioMed Central Ltd.

Cyclic nucleotides function as second messengers in all kingdoms of life. Recently, cyclic dinucleotides (c-di-NMPs) have been identified as bacterial signaling molecules. Cyclic diguanosine monophosphate (c-di-GMP) plays a major role in the modulation of bacterial surface components. Synthesis and degradation of c-di-GMP is controlled by so-called GGDEF (diguanylate cyclase activity) and EAL (phosphodiesterase activity) domains, respectively. These domains influence the sessility and motility of bacteria *via* regulation of biosynthesis of exopolysaccharides (e.g. cellulose) and fimbriae. Thus, the c-di-GMP level affects the formation of biofilms that represent the predominant form in which bacteria exist in the natural environment. The progress and persistence of many infectious diseases is based on biofilm formation by bacterial pathogens.

Hence, there is a great need for the establishment of a robust and sensitive quantitation method for cyclic dinucleotides. Recently, a matrix-assisted laser desorption/ionization-time-of-flight mass spectrometry (MALDI-TOF) method for the detection of c-di-GMP has been reported [1]. However, quantitation of low molecular weight molecules by MALDI-TOF is rather problematic and the described method requires a tedious separate chromatographic work-up step prior to the sample preparation for the final MALDI-TOF analysis. We have established a much more simple and sensitive quantitation method for c-di-NMPs using a reversed-phase liquid chromatography-coupled tandem mass spectrometry (LC-MS/MS) system with xanthosine 3',5'-cyclic monophosphate (cXMP) as internal standard (LOD = 4 ng/mL, see Figure 1). Cyclic

diadenosine monophosphate (c-di-AMP) would be the most suitable internal standard in terms of close structural relation and comparable chromatographic behavior. However, c-di-AMP may also be present in bacteria and thus might interfere with the determination of c-di-GMP levels.




Figure 1
Calibration curve for c-di-GMP obtained by LC-MS/MS with a reversed-phase column and cXMP as internal standard.

We will apply the established method for the quantitation of c-di-NMPs in various *Pseudomonas aeruginosa* strains and aim at determining the enzymatic activity of diguanilate cyclase.

Acknowledgements

The authors thank Dr. H.-G. Genieser and Dr. F. Schwede (BioLog, Bremen, Germany) for providing cyclic dinucleotide standards.

References

1. Simm R, Morr M, Remminghorst U, Andersson M, Römling U: **Quantitative determination of cyclic diguanosine monophosphate concentrations in nucleotide extracts of bacteria by matrix-assisted laser desorption/ionization-time-of-flight mass spectrometry.** *Anal Biochem* 2009, **386**:53-58.

Publish with **BioMed Central** and every scientist can read your work free of charge

"BioMed Central will be the most significant development for disseminating the results of biomedical research in our lifetime."

Sir Paul Nurse, Cancer Research UK

Your research papers will be:

- available free of charge to the entire biomedical community
- peer reviewed and published immediately upon acceptance
- cited in PubMed and archived on PubMed Central
- yours — you keep the copyright

Submit your manuscript here:
http://www.biomedcentral.com/info/publishing_adv.asp

