



MEETING ABSTRACT

Open Access

Development of a molecular platform for HTLV confirmatory diagnosis: importance of the internal amplification control

Maurício C Rocha-Junior^{1,2*}, Rodrigo Haddad^{1,3}, Evandra S Rodrigues^{1,2}, Osvaldo M Takayanagui³, Dimas T Covas^{1,3}, Simone Kashima^{1,2}

From 15th International Conference on Human Retroviruses: HTLV and Related Viruses
Leuven and Gembloux, Belgium. 5-8 June 2011

Background

Brazil may harbor the largest absolute number of HTLV-1-infected individuals worldwide. The current HTLV diagnosis is based mainly on antibodies detection. However, these tests exhibit high proportion of indeterminate results. Several real-time PCR techniques have been developed for the detection of HTLV-1/2. However, up to day the major drawbacks of HTLV-1/2 molecular diagnosis are the lack of standard molecular tests and the absence of suitable internal amplification control (IAC). The aim of this study was to develop a multiplex qualitative real-time PCR for the simultaneous detection and discrimination of HTLV-1/2 and to design an IAC for reaction monitoring.

Methods

After multiple sequence alignments of the full genomes of HTLV-1/2 subtypes, a conserved tax region was chosen for the design of specific primers and probes. The IAC was generated after the annealing of synthetic nucleotide sequences and cloned into TOPO TA[®] vector. MT-2 and Gu cell lines were used as positive controls for HTLV-1 and HTLV-2, respectively.

Results

The developed multiplex real-time reaction detected both HTLV-1 and HTLV-2 (105 to 101 copies/reaction) at the presence of IAC in the same reaction. Analytical sensitivity was 1.2 copies/reaction for HTLV-1 and 19.1 copies/reaction for HTLV-2. The analytical sensitivity analysis was performed in singleplex format.

Conclusion

The detection of HTLV-1/2 and IAC by multiplex real-time PCR was efficient. The developed IAC is suitable for the molecular diagnosis and its presence ensures that the negative results are not due to failure in pre-PCR procedures.

Acknowledgements

Financial Support: CTC/INCTC, FAPESP, FUNDHERP, FINEP.

Author details

¹Regional Blood Center of Ribeirão Preto, Ribeirão Preto, Brazil. ²Faculty of Pharmaceutical Sciences of Ribeirão Preto, University of São Paulo, Ribeirão Preto, Brazil. ³Faculty of Medicine of Ribeirão Preto, University of São Paulo, Ribeirão Preto, Brazil.

Published: 6 June 2011

doi:10.1186/1742-4690-8-S1-A251

Cite this article as: Rocha-Junior *et al.*: Development of a molecular platform for HTLV confirmatory diagnosis: importance of the internal amplification control. *Retrovirology* 2011 **8**(Suppl 1):A251.

Submit your next manuscript to BioMed Central
and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit



* Correspondence: mauricio_crj@hemocentro.fmrp.usp.br

¹Regional Blood Center of Ribeirão Preto, Ribeirão Preto, Brazil

Full list of author information is available at the end of the article