

LETTER

Pulse pressure variation, stroke volume variation and dynamic arterial elastance

Raphael Giraud, Nils Siegenthaler and Karim Bendjelid*

See related research by Monge Garcia *et al.*, <http://ccforum.com/content/15/1/R15>

We read with interest the recent article by Monge Garcia and colleagues [1]. We have two comments regarding this interesting physiological study.

First, although we support their observation [2], their explanations may lead to some confusion. Indeed, the Edwards Vigileo™ system-FloTrac™ sensor calculates stroke volume (SV) using the equation:

$$SV = Khi \times \sigma AP$$

where σAP is the standard deviation of the arterial pressure (AP) curve and Khi a constant quantifying arterial elastance and vascular resistance [3]. As stroke volume variation (SVV):

$$SVV \text{ (in \%)} = (SV_{\max} - SV_{\min}) / SV_{\text{mean}}$$

$$SVV \text{ (\%)} = (Khi \times \sigma AP_{\max} - Khi \times \sigma AP_{\min}) / Khi \times \sigma AP_{\text{mean}}$$

$$SVV \text{ (\%)} = Khi \times (\sigma AP_{\max} - \sigma AP_{\min}) / Khi \times \sigma AP_{\text{mean}}$$

$$SVV \text{ (\%)} = (\sigma AP_{\max} - \sigma AP_{\min}) / \sigma AP_{\text{mean}}$$

$$SVV \text{ (\%)} \approx (PP_{\max} - PP_{\min}) / PP_{\text{mean}}$$

Which means that $SVV \text{ (\%)} \approx$ pulse pressure variation (PPV; %) for FloTrac™.

With the present mathematical equation, it can be demonstrated that the SVV calculated by FloTrac™ (SVV_{FT}) is not influenced by Khi , which means that SVV_{FT} calculation does not include effective arterial elastance (Ea). The originality of Garcia and colleagues' finding is that they realized that when measuring PPV

conventionally [4], the ratio PPV/SVV_{FT} becomes a mirror of a dynamic Ea, as the conventional PPV selected integrates Khi (Figure 1). However, when the present innovative method is used, PPV and SVV should be sampled during the same period of time.

In conclusion, we believe that dynamic Ea predicting arterial pressure response to volume loading in preload-dependent patients is an interesting physiological concept. However, this demonstration cannot be achieved without taking into consideration the impact of PPV/SVV sampling.

Abbreviations

AP, arterial pressure; Ea, arterial elastance; PP, pulse pressure; PPV, pulse pressure variation; SV, stroke volume; SVV, stroke volume variation.

Competing interests

The authors declare that they have no competing interests.

Published: 23 March 2011

References

1. Monge Garcia MI, Gil Cano A, Gracia Romero M: Dynamic arterial elastance to predict arterial pressure response to volume loading in preload-dependent patients. *Crit Care* 2011, **15**:R15.
2. Bendjelid K: When to recalibrate the PiCCO? From a physiological point of view, the answer is simple. *Acta Anaesthesiol Scand* 2009, **53**:689-690.
3. Manecke GR: Edwards FloTrac sensor and Vigileo monitor: easy, accurate, reliable cardiac output assessment using the arterial pulse wave. *Expert Rev Med Devices* 2005, **2**:523-527.
4. Michard F, Chemla D, Richard C, Wysocki M, Pinsky MR, Lecarpentier Y, Teboul JL: Clinical use of respiratory changes in arterial pulse pressure to monitor the hemodynamic effects of PEEP. *Am J Respir Crit Care Med* 1999, **159**:935-939.

doi:10.1186/cc10088

Cite this article as: Giraud R, *et al.*: Pulse pressure variation, stroke volume variation and dynamic arterial elastance. *Critical Care* 2011, **15**:414.

*Correspondence: Karim.Bendjelid@hcuge.ch

Medecin Adjoint Agrégé, Division des Soins Intensifs, Hôpitaux Universitaires de Genève, CH-1211 Genève 14, Switzerland

