

SCIENTIFIC REPORTS

OPEN

Estimation of intraocular lens position from full crystalline lens geometry: towards a new generation of intraocular lens power calculation formulas

Eduardo Martínez-Enriquez¹, Pablo Pérez-Merino¹, Sonia Durán-Poveda², Ignacio Jiménez-Alfaro² & Susana Marcos¹

In a cataract surgery, the opacified crystalline lens is replaced by an artificial intraocular lens (IOL). To optimize the visual quality after surgery, the intraocular lens to be implanted must be selected preoperatively for every individual patient. Different generations of formulas have been proposed for selecting the intraocular lens dioptric power as a function of its estimated postoperative position. However, very few formulas include crystalline lens information, in most cases only one-dimensional. The present study proposes a new formula to preoperatively estimate the postoperative IOL position (ELP) based on information of the 3-dimensional full shape of the crystalline lens, obtained from quantitative eye anterior segment optical coherence tomography imaging. Real patients were measured before and after cataract surgery (IOL implantation). The IOL position and the postoperative refraction estimation errors were calculated by subtracting the preoperative estimations from the actual values measured after surgery. The proposed ELP formula produced lower estimation errors for both parameters -ELP and refraction- than the predictions obtained with standard state-of-the-art methods, and opens new avenues to the development of new generation IOL power calculation formulas that improve refractive and visual outcomes.

Age-related cataract is a cause of blindness on a global scale (43% of worldwide blindness) due to biological aging, genetic, and environmental factors of the crystalline lens, which loses its transparency. Cataract surgery has become a routine surgical procedure, with a treatment rate of 22 million annual surgeries worldwide¹. Typically, the opacified crystalline lens is removed and replaced by a foldable intraocular lens (IOL). To optimize the visual quality after surgery, it is crucial to individually select the optimal IOL power to be implanted². In current practice, the IOL power is usually calculated by using statistical regression formulas obtained from a large population sample^{3,4} or, more accurately, using theoretical formulas^{2,5-13}.

Deviations in the preoperative estimation of postoperative IOL position, i.e., the estimated lens position (ELP), represent the largest contribution to error in modern IOL power calculation formulas^{2,13-15}. Therefore, improvements in the ELP will provide better IOL power selection and thus refractive and visual outcomes.

Different generations of theoretical formulas have been proposed since 1970s, which vary in the way in which they estimate the lens position (ELP): the first generation assumed a constant value for the ELP^{5,7}; the second generation individualized the prediction by replacing the constant ELP by one variable dependent on the axial length (AL) measured for every patient^{6,11}; the third generation formulas (i.e. SRK/T formula⁹, Holladay formula⁸, or the Hoffer Q formula¹²) used axial length and anterior corneal curvature to predict ELP. Fourth and fifth generations included the preoperative anterior chamber depth (ACD) to improve the prediction (Olsen¹⁶ and Haigis formulas¹⁰, respectively).

¹Instituto de Óptica "Daza de Valdés", Consejo Superior de Investigaciones Científicas (CSIC), C/Serrano, 121, 28006, Madrid, Spain. ²Fundación Jiménez Díaz, Madrid, Av. Reyes Católicos, 2, 28040, Madrid, Spain. Correspondence and requests for materials should be addressed to E.M.-E. (email: eduardo.martinez@io.cfmac.csic.es)

	Male/Female	Age (y.o)	Axial length (mm)	IOL power (D)	Postoperative objective refraction (SE) (D)	Clinical grade of Cataract (BCN10)
S#1 (OD)	Female	80	23.10	22.0	0.750	N5
S#2 (OS)	Female	82	22.37	22.5	-1.125	N5
S#3 (OD)	Female	87	22.15	23.5	0.375	N6
S#3 (OS)	Female	87	22.08	23.5	0.125	N6
S#4 (OD)	Female	74	22.41	22.0	-0.370	N5
S#4 (OS)	Female	74	22.37	22.0	-0.625	N5
S#5 (OD)	Male	78	22.71	22.0	0	N6
S#5 (OS)	Male	78	22.62	22.0	0.250	N5
S#6 (OD)	Male	75	23.75	19.5	-0.250	N5
S#6 (OS)	Male	75	23.83	19.5	-0.125	N5
S#7 (OD)	Female	78	23.22	24.0	0.250	N4
S#7 (OS)	Female	78	23.35	24.0	0.625	N4

Table 1. Patients clinical data. D = diopters; IOL = intraocular lens; SE = spherical equivalent.

In all of these approaches, the ELP is estimated from parameters unrelated to the shape of the crystalline lens. Recent approaches included crystalline lens thickness (e.g., the Olsen formula¹³, or more recently, the Olsen *et al.* C-constant¹⁷) or 2-dimensional measurements of the central area of the crystalline lens (e.g., the so-called intersection approach¹⁸).

To our knowledge, 3-dimensional (3-D) quantitative estimations of the preoperative full crystalline lens geometry and position have never been used, in part due to the unavailability of full crystalline lens quantitative data. Nevertheless, it seems reasonable that the estimation of where the IOL will be placed within the capsular bag will largely benefit from accurate 3-D information of the crystalline lens geometry.

Optical coherence tomography (OCT) is an excellent technique to image the anterior segment of the eye due to its high-resolution and high-speed, and provided with fan and optical distortion correction algorithms^{19,20}, allows an accurate 3-D quantification of the geometry of the anterior segment of the eye²¹ including the central part of the crystalline lens²². We have recently presented and validated a method to accurately estimate and quantify the full shape of the crystalline lens from OCT data visible through the eye's natural pupil aperture²³.

In this study we propose a new formula to estimate the ELP based on preoperative 3-D full crystalline lens parameters (lens volume, surface area, diameter and equatorial plane position) measured in pre-cataract surgery patients. We compare the predictions (estimated lens position and refractive error) of the new formula with those from the SRK/T formula, the Olsen's C constant and the intersection approach.

To our knowledge, this is the first work reporting a formula that relates 3-D full shape of the crystalline lens parameters and ELP.

Material and Methods

Patients, Surgery and Clinical Measurements. Twelve eyes from seven patients (79 ± 5 years old) were measured before and after cataract surgery (2 to 4 months from surgery). All patients received Carl Zeiss CT Asphina 409 M (Zeiss, Jena, Germany) IOL. IOL power ranged from +19.50 D to +24 D. The SRK/T formula was used to calculate the IOL power, selecting the first myopic value closest to emmetropia. Preoperative refractive errors ranged between -2 to +3.25 D sphere and -2.25 to -0.5 D cylinder. The surgery was performed at Fundación Jiménez Díaz by a single surgeon (SD-P) using phacoemulsification through a 2-mm incision at 90°. A 6-mm continuous curvilinear capsulorhexis was made under viscoelastic material. Phacoemulsification of the lens was performed with a commercial microsurgical system (Stellaris Microsurgical System; Bausch & Lomb, Rochester, NY).

The study met the tenets of the Declaration of Helsinki. Ethical approval was granted by the Ethics Committee of Consejo Superior de Investigaciones Científicas (CSIC). Written informed consent was obtained from the patients after detailed explanation of the procedure.

Table 1 shows the patient's clinical profile, including the preoperative axial length (AL) measured with IOLMaster (Carl Zeiss AG, Jena, Germany), the actual IOL power implanted for each patient, and the postoperative final objective refraction (spherical equivalent, SE) obtained from auto-refractometry (Topcon, Tokyo, Japan).

The A-constant used for the IOL power calculations was 118.3, based on manufacturer recommendations. Table 1 also shows the clinical grade of cataract following the grading system BCN 10²⁴, which ranges between clear lens (N0) to fully opacified nuclear cataract (N10).

Optical Coherence Tomography Imaging. A custom-developed spectral OCT system, described in detail in previous publications^{20,25}, was used to acquire the OCT images. The light source is a superluminescent diode ($\lambda_0 = 840$ nm, $\Delta\lambda = 50$ nm). The set-up is based on a fiber-optics Michelson interferometer configuration, and the detector is composed of a spectrometer and a CMOS camera. The effective acquisition speed is 25000 A-Scans/s and the axial range is 7 mm in depth in air, resulting in a theoretical pixel resolution of 3.4 μ m. The axial resolution is 6.9 μ m in tissue.




Figure 1. Raw optical coherence tomography (OCT) images for subject S#1 OD, including the definition of some biometric parameters. **(a)** Preoperative measurement. **(b)** Postoperative measurement. ACD_{pre} = anterior chamber depth, EPP = equatorial plane position, CT = corneal thickness.

One 3-D volume was composed of 300 A-scans, and 50 B-scans on a 10×10 mm lateral area, acquired in 0.6 seconds, which led to a good trade-off between time acquisition and resolution. Measurements were acquired after inducing mydriasis with one drop of tropicamide. The OCT axis was aligned with the pupillary axis by asking the patient to follow a fixation stimulus until the iris appeared flat in the preview OCT horizontal and vertical cross-sections^{21,26}.

The axial range is not sufficiently large to capture all anterior segment surfaces in a single acquisition. In the preoperative measurements three sets of 3-D images were captured sequentially 5 seconds after blinking: (1) cornea, (2) anterior lens and (3) posterior lens. In the postoperative measurements two sets were needed (1) cornea and (2) IOL, shifting axially the plane of focus; all 3-D sets of data contained the iris. At least seven repeated measurements were collected for each plane of focus. Not useful measurements (e.g., measurements with blinking) were discarded from the analysis.

Figure 1 shows examples of raw OCT central B-scans for subject S#1 OD for preoperative (Fig. 1a) and postoperative (Fig. 1b) measurements. B-scans taken at the different foci have been manually merged for representation purposes. Automatic registration is performed in 3-D using the iris as a reference, as explained below.

3-Dimensional Eye Model Construction. In previous works^{22,26,27} we described the automatic construction of accurate 3-D eye models from OCT images, which involves three steps: (1) automatic surfaces detection and segmentation, (2) registration, and (3) distortion correction. The custom-developed segmentation approaches are described in detail in previous work^{22,26,27}. Specific routines have been designed in this study for the automatic detection of the IOL, including new segmentation functions. As in previous work, the automatically detected iris is used as a reference for image registration²⁶. Due to thicker crystalline lenses in some older eyes –which prevented the iris to be captured in the posterior lens images– the specular reflex of the anterior lens was used for registration in selected cases. Fan and optical distortion correction algorithms were applied on the registered volumes by using 3-D ray tracing routines¹⁹. The corneal and aqueous humor group refractive indexes were taken as 1.385²⁸, and 1.345, respectively, and the crystalline lens refractive index was obtained from the age-dependent expression derived by Uhlhorn *et al.*²⁹. The Asphina IOL refractive index was 1.460²¹.

Entire Lens Shape Estimation. In a recently published paper²³ we proposed a method to estimate the entire lens geometry from the information of the lens visible through the pupil in OCT images. The entire crystalline lens is described by means of a parametric model with two parameters that were trained and validated using 27 *ex vivo* lenses, in which the information of the full lens was available. Using the trained parametric model and the information within the pupil, very realistic lenses are constructed *in vivo* with a small estimation error. This method was applied to study the changes of lens full shape parameters with accommodation²⁷, in older patients with cataract²³, and in animal models of myopia³⁰.

Figure 2 shows constructed 3-D models from the preoperative measurement (Fig. 2a), including the part of the crystalline lens visible throughout the pupil (green) and its full shape estimation, as well as the postoperative measurement (Fig. 2b), with the IOL in purple for S#1 (OD). Both models superimposed are also shown (Fig. 2c).

Quantification. Once the 3-D models were constructed, various biometric parameters were quantified. From the preoperative measurements, we obtained: (1) anterior chamber depth (ACD_{pre}), defined as the distance between posterior cornea apex and anterior lens apex; (2) corneal thickness (CT); (3) lens thickness (LT); (4) crystalline lens anterior and posterior radius of curvature (RAL, RPL); (5) lens equatorial diameter (DIA); (6) lens equatorial plane position (EPP); (7) lens surface area (LSA); and (8) lens volume (VOL). From the postoperative measurements: (9) ACD_{post} , defined as the distance from the posterior cornea to the IOL apex; (10) IOL thickness




Figure 2. 3-D models for S#1 (OD). (a) Preoperative measurement, including the part of the crystalline lens visible through the pupil (green) and its full shape estimation. (b) Postoperative measurement, including the intraocular lens (IOL) in purple. (c) Both models superimposed. IOL = intraocular lens.

(IOLT); and (11) actual IOL position (LP), defined as the distance from the posterior cornea to the middle of the IOL: $LP = ACD_{post} + IOLT/2$. Figure 1 illustrates the definition of some of these parameters. In particular, the EPP is the 3-D distance between the anterior lens apex and the position of the equatorial plane. More details of the different parameters definition are given in previous publications^{23,27}. The crystalline lens anterior and posterior radii of curvature were calculated, after 6th order Zernike approximation and tilt removal of the lens, by fitting the best sphere in a 5-mm of diameter optical zone with respect to their apexes.

The analyzed crystalline lens parameters combinations included: (1) LSA/VOL ratio, which for a given volume reaches its minimum when the crystalline lens shape approaches a sphere, and (2) EPP/LT ratio, which gives the equator position in relation with the lens thickness.

ELP and Estimation Error. We obtained the ELP as a function of the preoperative geometry of the anterior segment, including the full shape of the lens. In addition to the proposed ELP estimation, the ELP was also obtained from state-of-the-art methods, namely:

- Using the C constant for the corresponding lens (Asphina 409) recently proposed by Olsen and Hoffmann:¹⁷ 0.39.
- Using the C constant proposed by Olsen and Hoffmann optimized with our data set: 0.397.
- Using the ELP from the SRK/T formula⁹ using the A-constant recommended by the manufacturer: 118.3.
- Using the ELP from the SRK/T formula optimizing the A constant with our data set: $A_{opt} = 118.5$.
- Using the intersection approach, that estimates the ELP from the intersection of the two parametric surfaces that best fit the anterior and posterior crystalline lens surfaces within the pupil¹⁸.

The estimation error (EE) was calculated as the difference between the prediction using the different methods mentioned above and the actual lens position (LP, obtained from the postoperative measurements), i.e., *estimation error (EE) = estimated lens position (ELP) – measured lens position (LP)*.

Mean arithmetic EE (ME) across the 12 eyes and standard deviation of the ME were calculated. A positive value of ME indicates that the method predicts a position further from the anterior lens than the actual position of the lens, and vice versa for a negative value. Mean of the *absolute* value of the EE (MAE) and standard deviation of the MAE were also calculated.

The optimization of the C and A constants was done by finding those values that minimized the MAE.

Postoperative Refraction Estimates and Error. To evaluate the improvements in terms of postoperative refraction estimation, the actual measured postoperative refraction (spherical equivalent, SE, Table 1) was

	LT (mm)	RAL (mm)	RPL (mm)	VOL (mm ³)	DIA (mm)	LSA (mm ²)	EPP (mm)	EPP/LT	LSA/VOL (mm ⁻¹)
S#1 (OD)	4.62 ± 0.005	7.14 ± 0.21	5.57 ± 0.07	178 ± 3.7	8.83 ± 0.12	165 ± 3.7	2.16 ± 0.016	0.467	0.927
S#2 (OS)	4.35 ± 0.005	8.94 ± 0.49	5.80 ± 0.14	177 ± 5.0	9.00 ± 0.13	168 ± 4.1	1.92 ± 0.025	0.441	0.949
S#3 (OD)	5.67 ± 0.001	7.77 ± 0.18	5.46 ± 0.20	264 ± 2.5	9.51 ± 0.10	209 ± 1.0	2.63 ± 0.003	0.463	0.792
S#3 (OS)	5.61 ± 0.002	7.62 ± 0.06	5.05 ± 0.02	246 ± 1.6	9.36 ± 0.05	198 ± 1.0	2.57 ± 0.007	0.457	0.805
S#4 (OD)	4.79 ± 0.019	9.70 ± 0.21	5.84 ± 0.12	218 ± 0.4	9.54 ± 0.03	191 ± 0.8	2.10 ± 0.020	0.438	0.876
S#4 (OS)	4.75 ± 0.01	9.53 ± 0.19	5.90 ± 0.03	224 ± 2.5	9.77 ± 0.08	195 ± 1.5	2.14 ± 0.021	0.450	0.871
S#5 (OD)	4.41 ± 0.003	9.15 ± 0.56	5.41 ± 0.17	176 ± 5.1	8.92 ± 0.17	166 ± 4.2	1.94 ± 0.041	0.440	0.943
S#5 (OS)	4.55 ± 0.003	8.75 ± 0.23	5.88 ± 0.002	199 ± 0.4	9.26 ± 0.05	182 ± 1.3	2.05 ± 0.004	0.451	0.915
S#6 (OD)	4.18 ± 0.004	11.70 ± 0.63	6.54 ± 0.23	193 ± 4.5	9.60 ± 0.11	185 ± 3.5	1.84 ± 0.026	0.442	0.958
S#6 (OS)	4.17 ± 0.001	11.25 ± 0.70	6.37 ± 0.17	186 ± 5.0	9.43 ± 0.15	178 ± 1.8	1.84 ± 0.002	0.440	0.957
S#7 (OD)	5.23 ± 0.005	8.75 ± 0.06	5.43 ± 0.02	237 ± 4.2	9.60 ± 0.03	198 ± 2.2	2.32 ± 0.001	0.444	0.835
S#7 (OS)	4.78 ± 0.005	8.45 ± 0.29	5.74 ± 0.12	197 ± 4.4	9.09 ± 0.10	176 ± 3.2	2.20 ± 0.008	0.459	0.893

Table 2. Full shape of the crystalline lens biometric parameters for each patient (mean ± standard deviation across measurements). LT = lens thickness; RAL = crystalline lens anterior radius of curvature; RPL = crystalline lens posterior radius of curvature; VOL = lens volume; DIA = lens equatorial diameter; LSA = lens surface area; EPP = equatorial plane position.

compared in retrospect with the expected (predicted) refraction obtained with the SRK/T IOL power calculation formula, using the actual IOL power implanted and the different estimations of the IOL position (ELP). Thus, the refractive error (RE) is defined as: $RE = \text{Measured_postoperative_refraction} - \text{estimated_refraction}$.

The mean arithmetic RE (MRE) across the 12 eyes and the standard deviation of the RE were calculated. A positive value of MRE indicates that the method predicts a refraction lower than the postoperative measured refraction, and vice versa for a negative value. The mean of the *absolute* values of the RE (MARE) and standard deviation of the MARE across eyes were also calculated.

Data Analysis. For all analyses, statistical significance was defined as a p-value lower than 0.05.

The normality of the data distribution was assessed for all the features and for the estimation errors with the Shapiro-Wilk test. When the data were not normally distributed, a square-root transformation was applied and the test was conducted over the transformed variable.

Linear regression analysis between pair of geometrical parameters of the crystalline lens and the lens position was performed, obtaining the Pearson correlation coefficient (r) and the p-value for testing the hypothesis of no correlation. Normality of the residuals was tested. Autocorrelation of the residuals was also tested with the Durbin-Watson test.

To assess if the ME and MRE produced by various methods were significantly different to zero, 1-sample t-test were used.

To assess if the MAE and the MARE produced by state-of-the-art methods were significantly different from the proposed, linear mixed-effects models (that handle repeated measures allowing an unequal number of repetitions) were used. As fixed effect, we entered methods into the model. As random effects, we had subjects, and as repeated measurements methods*eye.

The Bonferroni correction was applied for multiple tests. The SPSS 24.0 for Windows (SPSS Inc., Chicago, IL) was used for statistical analysis.

Data availability. The datasets generated and/or analysed during the current study are available in the “Estimation-of-Intraocular-Lens-Position-from-Full-Crystalline-Lens-Geometry” repository, [<https://github.com/EduardoMartinezEnriquez/Estimation-of-Intraocular-Lens-Position-from-Full-Crystalline-Lens-Geometry>].

Results

Full Shape Crystalline Lens Geometry. Table 2 shows crystalline lens biometric parameters for each patient (mean ± standard deviation across measurements).

We found a significant correlation between Lens Thickness (LT) and: Volume, VOL ($r = 0.905$, $p = 5.1 \cdot 10^{-5}$); Lens Surface Area, LSA ($r = 0.76$, $p = 0.004$); Equatorial Plane Position, EPP ($r = 0.98$, $p = 2.5 \cdot 10^{-9}$); Radius of the Anterior Lens, RAL ($r = -0.66$, $p = 0.019$); Radius of the Posterior Lens, RPL ($r = -0.75$, $p = 0.0046$), LSA/VOL ($r = -0.98$, $p = 1.5 \cdot 10^{-8}$); Preoperative Anterior Chamber Depth, ACD_{pre} ($r = -0.83$, $p = 8.5 \cdot 10^{-4}$) and age ($r = 0.69$, $p = 0.013$). The correlation was not significant with Lens Diameter, DIA ($r = 0.27$, $p = 0.39$) or EPP/LT ratio ($r = 0.53$, $p = 0.07$). Higher LT was associated with higher VOL and LSA, lower RAL and RPL (i.e., more curved surfaces), lower LSA/VOL ratio (more rounded lenses), lower ACD_{pre}, and higher EPP/LT (the equatorial plane position tends to be closer to LT/2). Figure 3 shows a graphical comparison of a thick crystalline lens (S#3 OD, purple) and a thin crystalline lens (S#1 OD, black), using the real geometries obtained from the analysis. In this particular example, the cornea, captured in both models, was used for registration.

Proposed ELP Estimation from the Preoperative Full Shape of the Lens Geometry. ELP was estimated using the EPP calculated from the full shape of the lens. Linear regressions between the available data




Figure 3. Graphical comparison of 3-D models for a thin lens (S#1 OD) and a thick lens (S#3 OD) using the real geometries obtained from the analysis. Cornea was used for registration.




Figure 4. Linear regression between $EPP/(LT \cdot ACD_{pre})$ and ELP/ACD_{pre} . $r = 0.98$, $p = 3.6 \cdot 10^{-9}$;
 $\frac{LP}{ACD_{pre}} = 0.250(\pm 0.07) + 8.497(\pm 0.43) \cdot \frac{EPP}{(LT \cdot ACD_{pre})}$. EPP = equatorial plane position, LT = lens thickness, ACD_{pre} = anterior chamber depth preoperative, ELP = estimated lens position, LP = lens position.

revealed that the highest correlation occurred between the ratio $EPP/(LT \cdot ACD_{pre})$ and the ratio ELP/ACD_{pre} ($r = 0.98$, $p = 3.6 \cdot 10^{-9}$). Figure 4 shows a scatter plot and the regression line given by

$$\frac{LP}{ACD_{pre}} = 0.250(\pm 0.07) + 8.497(\pm 0.43) \cdot \frac{EPP}{(LT \cdot ACD_{pre})}$$

Reordering the equation of the best linear fitting (solving for LP), the proposed estimation linearly relates the LP with the EPP/LT ratio and the ACD preoperative (ACD_{pre}) as follows:

$$ELP = 8.497 \cdot \frac{EPP}{LT} + 0.250 \cdot ACD_{pre}, \tag{1}$$

Method	ME \pm SD (μm)	P value (t-test)	MAE \pm SD (μm)	EE range (μm)	P value (multiple test, Bonferroni correction)
Proposed	0 \pm 81	1	65 \pm 45	−102 to 144	—
Olsen Constant	−26 \pm 186	0.64	158 \pm 91*	−297 to 289	0.02
Olsen C optimized	1 \pm 181	0.97	145 \pm 101	−263 to 312	0.34
SRK/T	−187 \pm 211 [†]	0.01	238 \pm 145*	−501 to 125	0.02
SRK/T A optimized	0 \pm 182	0.99	151 \pm 91	−350 to 250	0.65
Intersection approach	−79 \pm 201	0.20	177 \pm 114*	−352 to 380	0.04

Table 3. ELP estimation error: mean arithmetic error (ME), mean absolute error (MAE), estimation error (EE) range, and p values of the statistical tests performed, for the different methods. ME = Mean arithmetic estimation error; MAE = Mean absolute estimation error; SD = Standard deviation; EE = Estimation error. [†]ME significantly different from zero (t-test). *MAE significantly different than the proposed method (with Bonferroni correction).

Method	MRE \pm SD (D)	P value (t-test)	MARE \pm SD (D)	RE range (D)	P value (multiple test, Bonferroni correction)
Proposed	−0.09 \pm 0.37	0.41	0.24 \pm 0.29	−1.08 to 0.33	—
Olsen Constant	−0.04 \pm 0.46	0.75	0.36 \pm 0.27	−0.95 to 0.63	1
Olsen C optimized	−0.08 \pm 0.46	0.54	0.37 \pm 0.26	−1.00 to 0.58	1
SRK/T	0.20 \pm 0.53	0.23	0.47 \pm 0.28	−0.73 to 0.93	0.16
SRK/T A optimized	−0.08 \pm 0.53	0.61	0.41 \pm 0.31	−1.20 to 0.62	1
Intersection approach	0.03 \pm 0.44	0.81	0.37 \pm 0.23	−0.83 to 0.53	0.12

Table 4. Postoperative refraction estimation: mean arithmetic refractive error (MRE), mean absolute refractive error (MARE), refractive error (RE) range, and p values of the statistical tests performed, for the different methods. MRE = mean arithmetic refractive error; MARE = mean absolute refractive error; SD = Standard deviation; EE = Estimation error; D = diopters. [†]RE significantly different from zero (t-test). *MARE significantly different than the proposed method (with Bonferroni correction).

where LT is the measured lens thickness and ACD_{pre} is the preoperative anterior chamber depth (Fig. 1). As expected, the lens position increases with the EPP/LT ratio (i.e., as the equatorial plane position is closer to the half of the lens thickness) and with the ACD_{pre} .

Alternatively, another relationship with high correlation was found between the ratios ELP/LT and LSA/VOL ($r = 0.97$, $p = 1.4 \cdot 10^{-7}$), leading to the formula:

$$\text{ELP} = 1.72 \cdot \text{LT} \cdot \frac{\text{LSA}}{\text{VOL}} - 0.59 \cdot \text{LT}. \quad (2)$$

In the following we will employ equation (1) as the proposed method to obtain the ELP.

ELP Estimation Error. Table 3 shows the ME, MAE, EE range and p value of the statistical tests. The ME was close to zero for all methods, except for the SRK/T, in which the ME was significantly different to zero. For the proposed method, the standard deviation of the ME (81 μm) and the EE range (−102 to 144 μm) were smaller than those obtained by the other methods.

The MAE with the proposed method was statistically significantly lower (multiple comparisons, Bonferroni correction) than with the C constant by Olsen, the SRKT and the intersection approach. The EE for every eye when using the proposed method, the optimized C constant, the optimized SRK/T, and the intersection approach can be found as Supplementary Fig. S1.

Postoperative Refraction Estimation. Table 4 shows the MRE, MARE, RE range and p value of the statistical tests. The MRE was not significantly different to zero for any of the compared methods. For the proposed method, the standard deviation of the MRE (0.37 D) was smaller than for the other methods. The RE range (−1.08 to 0.33 D) was also smaller with the proposed method than for the other methods except for the intersection approach (−0.83 to 0.53 D). The MARE was not significantly different across the different methods (multiple comparisons, Bonferroni correction).

Figure 5 shows the MAE (color bars) and the MARE (black dotted line) for all the compared estimation methods.

The RE for every eye when using the proposed method, the optimized C constant, the optimized SRK/T, and the intersection approach can be found as Supplementary Fig. S2.

Discussion

We proposed a new method to estimate the postoperative IOL position (ELP) from preoperative 3-D geometry of the crystalline lens. A more accurate estimation of the IOL position leads to improved formulas for IOL power




Figure 5. Mean of the *absolute* value of the estimation error (MAE, color bars) and of the refractive error (MARE, black dotted line) across the 12 eyes, for the compared estimation methods.

estimation in which ELP is a primary source of error, and opens new avenues to the development of new strategies for IOL power estimation based on ray tracing.

The mean absolute estimation error (MAE) of the LP obtained in this study with the proposed method was >2 times lower than the error using the C constant by Olsen (even if the constant was optimized with our data set) or the SRK/T with the A constant optimized, and around 3 times lower than using the ELP estimate of the SRK/T or the intersection approach (Table 3, Fig. 5). These differences were statistically significant (multiple comparisons, Bonferroni correction) between the proposed method and the Olsen constant, the SRK/T, and the intersection approach.

Optimizing the C and the A constants reduced the error by 1.09 and 1.57 times with respect to the constants recommended by Olsen and the manufacturer respectively.

The deviation in the ELP estimation error across patients was lower with the proposed formula than using other methods ($\pm 81 \mu\text{m}$ with the proposed vs ± 181 to $\pm 211 \mu\text{m}$ with the other methods), and the estimation error range was also lower (-102 to $144 \mu\text{m}$ with the proposed vs -263 to 312 and -352 to $380 \mu\text{m}$ with other methods), indicating that the proposed method provides a more individualized account of the lens position than compared methods in the studied sample.

Figure 5 shows that the MAE and the MARE were well correlated ($r = 0.94$, $p = 0.005$), indicating that an improvement in the ELP estimation should result in an improvement in the postoperative refraction. Comparing the absolute RE obtained subject by subject (see Supplementary Fig. S2) with the proposed method and with all the other methods (12 eyes*5 methods = 60 data), the improvement in refraction produced by the proposed approach was greater than 0.15 D in 32 cases, greater than 0.25 D in 18 cases and greater than 0.5 D in 3 cases; the estimation was >0.15 D worse in 4 cases and >0.25 D worse just in 2 cases. As with the ELP, the standard deviation of the refraction estimation error across eyes using the proposed ELP (± 0.37 D) was lower than with the other methods (from ± 0.44 D to ± 0.53 D), indicating a more customized performance for the different subjects of the study.

Although improvement in RE was found in most eyes, MARE differences between the proposed method and the compared methods were not proven to be significantly different (multiple comparisons, Bonferroni correction), likely due to the high data dispersion and the low number of eyes ($N = 12$). Clinical relevance may be achieved when combining this methodology with further methods that use additional anatomical information of the patient (i.e. patient-based custom eye models using anterior and posterior corneal topography, IOL geometry, and ray tracing). A larger sample population may be required to generalize the formula for clinical use.

Note that the proposed estimations were compared to those provided by SRK/T because it is one of the most commonly used formulas in clinic and because it was the method used by the surgeon for IOL power selection in all the patients in the present work. We also compared with Olsen's C constant and the intersection approach because, as similarly to our proposed method, these methods use crystalline lens parameters to estimate the ELP. There are several anterior segment OCT instruments clinically available (corneal topographers; 2-D optical biometers and OCT units to guide femtosecond laser cataract surgery) that could be envisioned to be modified and coupled with the quantification tools described in this study to be used in the optimization of the selection of IOLs.

In a clinical application, typical IOLs are sold in 0.5 D increments. Comparing with SRK/T, the proposed approach would result in a different preoperative selection of IOL power in 5 of the 12 eyes, following the same selection criteria by the same surgeon. Specifically, IOL power would change in 0.5 D for 2 eyes and 1 D for 3 eyes.

Advanced cataracts may complicate measuring the posterior lens surface or the lens thickness. We did not find any difficulty to capture the posterior lens in cataracts of grade up to N6 (over N10) in our study. Besides, recent works show cataract crystalline lens images up to grade 6 in LOCS III grading³¹ and excellent results (no failures in a sample of 188 eyes of different cataract grades) in axial measurements (ACD, AL) in the OCT based IOLMaster 700³². The changes in the refractive index with the cataract grade are not well known. We did not find

any relation between the grade of the cataract and the ELP estimation error in our study. Although changes in the refractive index may affect the estimation of the EPP, ELP estimation formula (1) is quite robust because it does not directly depend on the considered refractive index of the crystalline lens (the ratio EPP/LT does not depend on the refractive index).

The use of 3-D full shape information of the crystalline lens goes beyond the state-of-the-art, as current approaches use, at most, lens thickness (1-D) or lens cross-sectional data of the central part. In this study, we proposed a specific formula that estimates the ELP from 3-D full shape of the crystalline lens, in particular the equatorial plane position (EPP), the thickness of the crystalline lens (LT) and the ACD_{pre} . Nevertheless, other relationships (such as that shown in equation (2)), can be obtained that use other lens 3-D parameters (i.e. LSA and VOL) to derive ELP. We did not attempt to use a larger number of features in the equation (1) for the ELP estimation to avoid over-fitting, given the small data set (12 eyes), and prevent an unfair comparison with other methods. However, the fact that we obtained lower MAE over other ELP approaches also optimized with the same dataset (such as the C-Constant formula of Olsen or the A constant in the SRK/T) suggests that it is the use of a more comprehensive description of the crystalline lens, and not the optimization to the specific eyes, the reason behind the improvement.

This study demonstrates the potential opened by image-based 3-D quantification of the preoperative crystalline lens to improve cataract surgery. The retrieved formulas (equations (1)-(2)) or even other alternatives can be generalized using a larger clinical population. The availability of larger dataset will allow to use powerful tools to systematically retrieve the best set of features³³ (i.e. those most informative, such as, lens geometrical parameters, age³⁴, preoperative refraction, ...) as well as a mathematical expression (N-dimensional, linear or non-linear) depending on these features to estimate the ELP.

Improving the generalization will allow obtaining formulas that more accurately estimate the IOL position of outliers (e.g., extreme axial lengths³⁵) or after refractive surgery, which creates difficulties in IOL power calculation^{36,37}.

The refraction predictions have been obtained using the proposed ELP in the SRK/T formula for IOL power selection. However, our ELP estimation can be used in different current IOL power calculation formulas to improve the refraction prediction and thus the IOL selection.

While most current formulas assume some simplifications (e.g., the posterior cornea radius of curvature is not considered, assumption of a keratometric index), ray tracing-based IOL power estimations are not subject to these assumptions. Computer eye models for ray tracing customized to the patient's anatomy²¹ can be constructed using all inputs from quantitative 3-D OCT (anterior and posterior corneal topography, 3-D biometry, IOL tilt and decentration and the proposed ELP based on the full shape of the crystalline lens), further increasing the accuracy of the IOL power selection, particularly in non-standard cases such as irregular corneas or post-refractive surgery eyes.

In the current study we obtained all anterior segment parameters from the same spectral OCT system, and axial length from a commercial low coherence interferometry system. However, the larger range of swept source OCT systems will allow acquiring all the information from a single instrument.

It should be noted that the proposed formula has been obtained for the specific implanted IOL (Asphina 409 M). Different constants in the formula are likely to result from different IOL platforms¹⁷.

The proposed estimate of ELP relies on a reliable estimation of the crystalline lens parameters. We found high repeatability in the parameter estimation from repeated measurements on the same eye: the mean across eyes of the standard deviation of repeated measurements was 0.014 mm for EPP, 3.30 mm³ for VOL, 3.24 mm² for LSA, 0.092 mm for DIA, and 0.005 mm for LT.

The technique for full lens shape estimation has been validated in eyes *ex vivo*, and further supported *in vivo* in young and presbyopic eyes^{23,27}. In previous studies we found that the assumption of an equivalent homogeneous refractive index in the lens (instead of the GRIN) resulted in <1.5% inaccuracies in VOL, and the effect of OCT lateral sampling, axial resolution, and registration errors resulted in <1% inaccuracies in EPP, VOL and DIA, which translates into <1% in the ELP. Furthermore, the application in older lenses reduces potential criticisms of the technique which may play some role in younger lenses. The training set of *ex vivo* lenses used to develop the parametric full lens model mainly corresponded to nearly presbyopic or presbyopic eyes, where one expects minimal changes with accommodation and thus *ex vivo* and *in vivo* lenses are expected to have approximately the same shape. On the other hand, GRIN is expected to be much flatter in older lenses, thus minimizing the error related to the assumption of constant index of refraction in the calculations of the current paper.

Postoperative measurements were obtained 2 to 4 months after surgery. Long-term (1 day to 1 year) changes in refractive outcomes in single-piece IOLs (as the IOL of the present study) due to IOL position shifts have been reported to be of negligible clinical relevance^{38,39}. Further analysis will be needed to potentially incorporate IOL postoperative shifts in other types of IOLs (e.g., multipiece IOLs).

Our results show an improvement in the estimation of the lens position over approaches based on limited anatomical information (e.g., only the radius of curvature of anterior cornea and the axial length of the eye in the SRK/T ELP). These results further improve over previous proposals which included for the first time lens data (e.g., Olsen constant or the intersection approach). Thus, knowledge of the full lens shape, and in particular EPP, is very valuable to estimate ELP.

In summary, 3-D OCT with image processing, distortion correction, and full-shape estimation of the crystalline lens algorithms allows accurate quantification of the anterior segment of the eye, which is very useful for improving the preoperative estimation of the postoperative IOL position (ELP).

Patient-specific eye models that include the information on lens volume, surface area, diameter, and equatorial plane position, open the possibility for development a new generation of IOL power estimation formulas.

References

- Foster, A. Cataract and “Vision 2020—the right to sight” initiative. *British Journal of Ophthalmology* **85**, 635–637 (2001).
- Olsen, T. Calculation of intraocular lens power: a review. *Acta Ophthalmol. Scand.* **85**, 472–485 (2007).
- Sanders, D. *et al.* Comparison of the accuracy of the Binkhorst, Colenbrander, and SRK implant power prediction formulas. *J. Am. Intraocul. Implant. Soc.* **7**, 337–340 (1981).
- Sanders, D. R., Retzlaff, J. & Kraff, M. C. Comparison of the SRK II formula and other second generation formulas. *J. Cataract Refract. Surg.* **14**, 136–141 (1988).
- Binkhorst, R. D. The optical design of intraocular lens implants. *Ophthalmic Surg.* **6**, 17–31 (1975).
- Binkhorst, R. D. Biometric A-scan ultrasonography and intraocular lens power calculation. *Selected proceedings of the fifth biennial cataract surgical congress, St. Louis, Mosby* 175–182 (1978).
- Colenbrander, M. C. Calculation of the power of an iris clip lens for distant vision. *Br. J. Ophthalmol.* **57**, 735–740 (1973).
- Holladay, J. T. *et al.* A three-part system for refining intraocular lens power calculations. *J. Cataract Refract. Surg.* **14**, 17–24 (1988).
- Retzlaff, J. A., Sanders, D. R. & Kraff, M. C. Development of the SRK/T intraocular lens implant power calculation formula. *J. Cataract Refract. Surg.* **16**, 333–340 (1990).
- Haigis, W. In *Intraocular Lens Power Calculations* (ed Shammas H.) 41–57 (Slack, 2004).
- Hoffer, K. J. The effect of axial length on posterior chamber lens and posterior capsule position. *Current concepts Ophthalmic Surg.* **1**, 20–22 (1984).
- Hoffer, K. J. The Hoffer Q formula: a comparison of theoretic and regression formulas. *J. Cataract Refract. Surg.* **19**, 700–712 (1993).
- Olsen, T. Prediction of the effective postoperative (intraocular lens) anterior chamber depth. *J. Cataract Refract. Surg.* **32**, 419–424 (2006).
- Norrby, S. Sources of error in intraocular lens power calculation. *J. Cataract Refract. Surg.* **34**, 368–376 (2008).
- Wang, W. *et al.* Precision of a new ocular biometer in eyes with cataract using swept source optical coherence tomography combined with Placido-disk corneal topography. *Sci. Rep.* **7**, 13736 (2017).
- Olsen, T., Olesen, H., Thim, K. & Corydon, L. Prediction of postoperative intraocular lens chamber depth. *J. Cataract Refract. Surg.* **16**, 587–590 (1990).
- Olsen, T. & Hoffmann, P. C constant: new concept for ray tracing-assisted intraocular lens power calculation. *J. Cataract. Refract. Surg.* **40**, 764–773 (2014).
- Hwang, K. Y., Yoo, Y. S., Joo, C. K. & Yoon, G. New parameter for predicting the postoperative IOL position: preoperative lens equator depth measured by three-dimensional anterior segment optical coherence tomography. *Investigative Ophthalmology & Visual Science* **56**, 1356–1356 (2015).
- Ortiz, S. *et al.* Optical distortion correction in optical coherence tomography for quantitative ocular anterior segment by three-dimensional imaging. *Opt. Express* **18**, 2782–2796 (2010).
- Ortiz, S., Siedlecki, D., Remon, L. & Marcos, S. Optical coherence tomography for quantitative surface topography. *Appl. Opt.* **48**, 6708–6715 (2009).
- Sun, M., Pérez-Merino, P., Martínez-Enriquez, E., Velasco-Ocana, M. & Marcos, S. Full 3-D OCT-based pseudophakic custom computer eye model. *Biomed. Opt. Express* **7**, 1074–1088 (2016).
- Ortiz, S., Perez-Merino, P., Gamba, E., de Castro, A. & Marcos, S. *In vivo* human crystalline lens topography. *Biomed. Opt. Express* **3**, 2471–2488 (2012).
- Martínez-Enriquez, E. *et al.* Optical coherence tomography based estimates of crystalline lens volume, equatorial diameter, and plane position. *Invest. Ophthalmol. Vis. Sci.* **57**, OCT600–610 (2016).
- Barraquer, R. I. *et al.* Validation of the Nuclear Cataract Grading System BCN 10. *Ophthalmic Res.* **57**, 247–251 (2017).
- Grulkowski, I. *et al.* Anterior segment imaging with Spectral OCT system using a high-speed CMOS camera. *Opt. Express* **17**, 4842–4858 (2009).
- Pérez-Merino, P., Velasco-Ocana, M., Martínez-Enriquez, E. & Marcos, S. OCT-based crystalline lens topography in accommodating eyes. *Biomed. Opt. Express* **6**, 5039–5054 (2015).
- Martínez-Enriquez, E., Pérez-Merino, P., Velasco-Ocana, M. & Marcos, S. OCT-based full crystalline lens shape change during accommodation *in vivo*. *Biomed. Opt. Express* **8**, 918–933 (2017).
- Drexler, W. *et al.* Submicrometer precision biometry of the anterior segment of the human eye. *Invest. Ophthalmol. Vis. Sci.* **38**, 1304–1313 (1997).
- Uhlhorn, S. R., Borja, D., Manns, F. & Parel, J. M. Refractive index measurement of the isolated crystalline lens using optical coherence tomography. *Vision Res.* **48**, 2732–2738 (2008).
- Pérez-Merino, P. *et al.* Three-dimensional OCT based guinea pig eye model: relating morphology and optics. *Biomed. Opt. Express* **8**, 2173–2184 (2017).
- de Castro, A. *et al.* Three-dimensional cataract crystalline lens imaging with swept-source optical coherence tomography. *Invest. Ophthalmol. Vis. Sci.* **59**, 897–903 (2018).
- Akman, A., Asena, L. & Gungor, S. G. Evaluation and comparison of the new swept source OCT-based IOLMaster 700 with the IOLMaster 500. *Br. J. Ophthalmol.* **100**, 1201–1205 (2016).
- Blum, A. L. & Langley, P. Selection of relevant features and examples in machine learning. *Artificial Intelligence* **97**, 245–271 (1997).
- Hayashi, K., Ogawa, S., Yoshida, M. & Yoshimura, K. Influence of patient age on intraocular lens power prediction error. *Am. J. Ophthalmol.* **170**, 232–237 (2016).
- Carifi, G., Aiello, F., Zygoura, V., Kopsachilis, N. & Maurino, V. Accuracy of the refractive prediction determined by multiple currently available intraocular lens power calculation formulas in small eyes. *Am. J. Ophthalmol.* **159**, 577–583 (2015).
- Seitz, B., Langenbucher, A., Nguyen, N. X., Kus, M. M. & Kuchle, M. Underestimation of intraocular lens power for cataract surgery after myopic photorefractive keratectomy. *Ophthalmology* **106**, 693–702 (1999).
- McCarthy, M., Gavanski, G. M., Paton, K. E. & Holland, S. P. Intraocular lens power calculations after myopic laser refractive surgery: a comparison of methods in 173 eyes. *Ophthalmology* **118**, 940–944 (2011).
- Wirtitsch, M. G. *et al.* Effect of haptic design on change in axial lens position after cataract surgery. *J. Cataract Refract. Surg.* **30**, 45–51 (2004).
- Klijn, S., Sicam, V. A. & Reus, N. J. Long-term changes in intraocular lens position and corneal curvature after cataract surgery and their effect on refraction. *J. Cataract Refract. Surg.* **42**, 35–43 (2016).

Acknowledgements

The research leading to these results has received funding from the European Research Council under the European Union’s Seventh Framework Programme (FP7/2007-2013)/ERC Grant Agreement ERC-2011-AdG-294099. This study has been also supported by Spanish government grants FIS2011-25637 and FIS2014-56643-R to S. Marcos. The funding organizations had no role in the design or conduct of this research.

Author Contributions

Concept and design: E.M.E., P.P.M., S.M.; Data acquisition: P.P.M., S.D.P., I.J.A.; Data analysis/interpretation: E.M.E., P.P.M., S.M.; Drafting manuscript: E.M.E., S.M.; Critical revision of manuscript: all authors; Statistical analysis: E.M.E.; Admin, technical or material support: P.P.M.; Supervision: S.M.; Final approval: all authors.

Additional Information

Supplementary information accompanies this paper at <https://doi.org/10.1038/s41598-018-28272-6>.

Competing Interests: The authors declare no competing interests.

Publisher's note: Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2018