

# Diabetes patient preferences for glucose-monitoring technologies: results from a discrete choice experiment in Poland and the Netherlands

Ian P Smith <sup>1</sup>, Chiara L Whichello,<sup>2,3,4</sup> Jorien Veldwijk,<sup>1,3,4</sup> Maureen P M H Rutten-van Mölken,<sup>3</sup> C G M Groothuis-Oudshoorn,<sup>5</sup> Rimke C Vos <sup>6</sup>, Esther W de Bekker-Grob,<sup>3,4</sup> G Ardine de Wit<sup>1</sup>

**To cite:** Smith IP, Whichello CL, Veldwijk J, *et al.* Diabetes patient preferences for glucose-monitoring technologies: results from a discrete choice experiment in Poland and the Netherlands. *BMJ Open Diab Res Care* 2023;**11**:e003025. doi:10.1136/bmjdr-2022-003025

► Additional supplemental material is published online only. To view, please visit the journal online (<http://dx.doi.org/10.1136/bmjdr-2022-003025>).

Received 4 July 2022  
Accepted 10 December 2022



© Author(s) (or their employer(s)) 2023. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

For numbered affiliations see end of article.

## Correspondence to

Ian P Smith;  
i.p.smith@umcutrecht.nl

## ABSTRACT

**Introduction** New glucose-monitoring technologies have different cost–benefit profiles compared with traditional finger-prick tests, resulting in a preference-sensitive situation for patients. This study aimed to assess the relative value adults with diabetes assign to device attributes in two countries.

**Research design and methods** Adults with type 1 or 2 diabetes from the Netherlands (n=226) and Poland (n=261) completed an online discrete choice experiment. Respondents choose between hypothetical glucose monitors described using seven attributes: precision, effort to check, number of finger pricks required, risk of skin irritation, information provided, alarm function and out-of-pocket costs. Panel mixed logit models were used to determine attribute relative importance and to calculate expected uptake rates and willingness to pay (WTP).

**Results** The most important attribute for both countries was monthly out-of-pocket costs. Polish respondents were more likely than Dutch respondents to choose a glucose-monitoring device over a standard finger prick and had higher WTP for a device. Dutch respondents had higher WTP for device improvements in an effort to check and reduce the number of finger pricks a device requires.

**Conclusion** Costs are the primary concern of patients in both countries when choosing a glucose monitor and would likely hamper real-world uptake. The costs-benefit profiles of such devices should be critically reviewed.

## INTRODUCTION

Diabetes is a chronic disease characterized by the body's inability to maintain healthy levels of blood glucose, which is associated with long-term health problems including an increased risk of mortality with an estimated global prevalence of 10.5% in 2021.<sup>1–3</sup> Diabetes care is centered around the cornerstone of metabolic control, specifically keeping glucose levels as close to normal as possible through medication, a careful diet, physical activity and self-monitoring of blood

## WHAT IS ALREADY KNOWN ON THIS TOPIC

⇒ Previous research identified factors that patients with diabetes find important when choosing a glucose monitor for self-management of their diabetes. These studies did not assess the relative importance of these factors or were only conducted in samples of type 1 diabetes mellitus (T1DM), limiting how these findings can be applied.

## WHAT THIS STUDY ADDS

⇒ This study identified out-of-pocket costs as the most important factor for both patients with T1DM and type 2 diabetes mellitus in two countries and quantified that these costs are 5–50 times more important than any other factor.

## HOW THIS STUDY MIGHT AFFECT RESEARCH, PRACTICE OR POLICY

⇒ Regulators and policy makers can incorporate this patient preference information into decision making to develop diabetes care management strategies that are not only cost-efficient but also based on patient values.

glucose (SMBG).<sup>4 5</sup> SMBG has traditionally been done using the finger-prick test and is associated with improvements in glycemic control.<sup>6</sup> While highly accurate,<sup>7</sup> this technique represents a large burden to patients, which can result in non-compliance to medical treatment advice.<sup>8–10</sup> Studies examining the adherence of patients to SMBG regimens report adherence rates ranging from 88% in Australia<sup>11</sup> to as low as 44% in Sweden,<sup>12</sup> 26% in the USA<sup>13</sup> and 20% in Hungary.<sup>14</sup> These low adherence rates are related to barriers to the practice of SMBG including low socioeconomic status (SES), fear of testing and fingertip pain, distressing

emotions and thoughts, frustration about 'poor' blood glucose reading, lack of awareness of hypoglycemia and hyperglycemia, lack of social support and difficulty in interpreting SMBG results.<sup>15</sup>

Recent technological developments have resulted in commercially available medical devices which can (semi-) continuously monitor blood glucose levels (or proxies thereof).<sup>16 17</sup> These devices are often less invasive, quicker and easier to use, and can give more detailed daily blood glucose-level information by showing trends over time compared with SMBG with finger pricking.<sup>18–20</sup> However, these devices vary in regard to functionality and features including (but not limited to) differences in accuracy, size, battery requirements, range of transmitter, calibration requirements, scanning procedures and longevity (replacement time). Further, these devices are often not reimbursed through insurance plans and can have high out-of-pocket costs for the patient.<sup>21</sup> The differences in function, features and costs have resulted in a situation where personal preferences may guide the choice of device used for SMBG.

Despite growing interest in patient preference assessment, limited research has been done quantifying patient preferences for glucose monitors. Hannah *et al* found that for patients with type 1 diabetes (T1DM), the most important factors for choosing a continuous glucose monitor (CGM) were method of data retrieval, longer sensor wear time with more adhesive durability, and personalized alerts and alarms.<sup>22</sup> Engler *et al* found that the reasons related to stopping CGM usage for patients with T1DM were poor accuracy due to lag times, insurance reimbursement or cost, comfort and false alarms.<sup>18</sup> They also found that for patients with T1DM without CGM experience cost, having a device attached to the body and expectations of discomfort in wearing were primary reasons for not using a CGM for SMBG.<sup>18</sup> Both studies highlight the preference-sensitive nature of these devices; however, neither included patients with type 2 diabetes mellitus (T2DM), a growing population of patients who may need to monitor their blood glucose.<sup>19 20 23</sup> Further, only Hannah *et al*<sup>22</sup> used a method of relative valuation to show how important these attributes were in regard to each other but did not include a cost attribute, which is a major concern for many patients. There is thus a gap in knowledge regarding the relative valuation information that regulatory authorities and decision makers use to guide policies for medical treatments.<sup>24</sup> This study aimed to fill that gap by quantitatively assessing the factors that patients with T1DM and T2DM consider important when choosing a glucose-monitoring device for SMBG and identify willingness to pay (WTP) and expected device uptake rates.

## MATERIALS AND METHODS

### Subjects

Participants were recruited from the Netherlands and Poland through a professional panel provider

(SurveyEngine). These countries were chosen as costs were expected to play an important role in deciding between devices, and these two countries had partial and no reimbursement of glucose monitors for SMBG at the time of data collection (respectively). To be eligible to complete the survey, patients had to have a self-reported diagnosis of T1DM or T2DM, reside in the Netherlands or Poland, be over 18 years of age, be able to read and understand Dutch or Polish, and have access to a computer.

### Discrete choice experiment (DCE)

A DCE was used to quantify patient preferences.<sup>25 26</sup> DCEs are based on random utility theory, which assumes that the utility or value of a healthcare alternative can be derived through the compound valuation of the different attributes and attribute levels used to describe the treatment alternative.<sup>27–29</sup> In a DCE, respondents are presented with choice tasks in which they chose their preferred option from two or more alternative treatment profiles. These alternative profiles describe treatments using a set of characteristics (called attributes) with varying levels, representing realistic values of these attributes.<sup>30 31</sup> Patients choose the alternative which represents the highest personal utility based on the personal value they attach to the different levels of attributes used to describe the alternative. After a patient completes the DCE, attribute estimates can be generated using econometric models and the relative importance of the included attributes can be inferred from these estimates.<sup>32–34</sup>

### Attributes and levels

The attributes and levels used in this DCE were developed according to best practices using a stepwise, qualitative approach from April 2019 to October 2019.<sup>35 36</sup> This approach started with a scoping literature review of articles describing aspects relevant to patients in using glucose-monitoring devices. The results of this review were used to create an interview guide (see online supplemental material) which was used in semistructured interviews with patients with T1DM and T2DM from the Netherlands (n=9), clinical diabetes experts (n=5), patient organization representatives (n=2) and pharmaceutical industry representatives involved in glucose-monitoring device development (n=4), as well as a focus group with patients with T1DM and T2DM in Poland (n=10). This process generated a list of 12 potentially relevant attributes which was reviewed and reduced by the research team to ensure relevance according to the interviewees, non-redundancy and operationality to a final list of 7 attributes for use in the DCE. The levels used to describe the attributes were developed based on the literature review and interviews and were chosen to be realistic and reflect the most common types of commercially available glucose monitors, including CGMs and flash glucose monitors (FGMs).<sup>37–40</sup> One attribute ('out-of-pocket costs') was standardized between the two countries using

**Table 1** Attributes and levels for the discrete choice experiment

Attributes	Level 1	Level 2	Level 3	Level 4
Precision compared with finger pricking‡	Less accurate than finger pricking (higher or lower by 0.6 mmol/L (*10.8 mg/dL‡))	Less accurate than finger pricking (higher or lower by 0.3 (*5.4 mg/dL‡))	Accurate as finger pricking†	–
Average number of finger pricks per day§	4‡	2	0	–
Effort to check¶	High effort: you need to measure your glucose levels yourself	Moderate effort: you scan a sensor to check glucose levels	Low effort: glucose levels automatically sent to you†	–
Probability of getting skin irritation or redness**	35% chance of skin irritation or redness	20% chance of skin irritation or redness	5% chance of skin irritation or redness	No chance of skin irritation or redness‡
Monthly costs††	€250 (*550zł)	€175 (*390zł)	€100 (*220zł)	€25 (*55zł)
Glucose information‡‡	Current glucose level†	Current Glucose level and arrow	Current glucose level and a graphic of your level trends over the day	–
Alarms§§	No‡	Yes	–	–

\*Unit equivalents shown for Polish survey.

†Reference level.

‡Attribute explanation as presented to patients: Some glucose monitors are more precise than others. Finger pricking is generally regarded as the most accurate way to measure glucose levels. Measurements from devices that use sensors can be just as accurate but can also be less accurate than finger pricking, especially if your glucose levels are very high or very low. For example, if your glucose level is 6 mmol/L and you measure it with a device that is off by 0.6 mmol/L, then this device can say your glucose is anywhere from 5.4 to 6.6 mmol/L

§Attribute explanation as presented to patients: This is how many times you would need to do a finger-prick test each day on an average day. This number could be higher on days when you feel the need to test more often like when you're sick, but we want you to picture an average day. Sometimes, this is your only method of measuring your glucose levels or you might need to do finger-prick tests to confirm the levels from another device.

¶Attribute explanation as presented to patients: This means how much effort you need to give to check your blood glucose levels. High effort checking means you need to stop what you're doing and concentrate on measuring your levels. You need to wash your hands, get out your device equipment, prick your finger, put blood on a strip, check the results and then clean everything up. Moderate effort checking means you need to get out a small device and use it to scan the sensor on your body to obtain your glucose levels. Low-effort checking means your glucose levels are automatically sent to a device which you can view at any time. This could be a dedicated glucose device, your phone or a smartwatch. You do not need to do anything to have your blood glucose levels sent through, just look at the device to check.

\*\*Attribute explanation as presented to patients: A chance of skin irritation or redness around a sensor means a redness or itchy rash on the skin around or under the sensor. This is similar to having an itchy allergic reaction and can be rather uncomfortable or irritating. The sensor will need to be removed and replaced in a different spot. This skin irritation and redness usually last until after the sensor is replaced. Not all sensors have this side effect, so chances of getting the side effect can differ per device. If a device gives you a 15% chance, this means that 15 out of a 100 people who get this device experience skin irritation and redness, while 85 out of a 100 people do not experience this.

††Attribute explanation as presented to patients: This means how much money you need to pay out-of-pocket per month in order to check your blood glucose. Please note that this is money that is not reimbursed by your insurance. This could be money needed to pay for devices, sensors, or strips used.

‡‡Attribute explanation as presented to patients: This means how your glucose levels are presented to you. This information could be only your current glucose level (you only see a digital number like 8.3 mmol/L). This could be your current glucose level with an arrow showing how your blood glucose is changing as compared to your previous measurement (increasing, decreasing or stable). Or it could show your current glucose level with a graphic of your blood glucose levels over the day.

§§Attribute explanation as presented to patients: Your device will give you a beeping alarm (like a phone notification) any time your blood glucose levels are (getting) too high or too low.

purchasing power parity weights to ensure that the relative value of the levels was similar, given the differences in wealth between the two countries.<sup>41</sup> The final list of attributes and levels used in the DCE can be found in [table 1](#).

### Experimental design

The DCE was developed using an efficient design (Bayesian D-efficient design<sup>42 43</sup>) generated in NGene V.1.0 software. This allows for participants to complete a minimal amount of choice tasks (3 blocks of n=12 choice tasks each) while

maximizing the amount of information each task generates. Available literature, interviews, and researcher knowledge were used to generate the initial design. The design was updated after a pilot of n=99 Dutch participants. In each choice task, patients were instructed to imagine that their doctor told them to check their blood glucose levels at least four times per day and gave them options of devices to choose from to do this. The choice tasks were presented using a dual response 'best–best' set-up where participants

Imagine that your doctor told you to check your blood glucose levels at least four times per day. To do this, the doctor offers you different hypothetical devices to choose from.

	Device A	Device B
Precision compared to fingerpricking	Less accurate than fingerpricking (higher or lower by 0.3)	Less accurate than fingerpricking (higher or lower by 0.6)
Average number of fingerpricks per day	0	0
Effort to check	Low effort	Moderate effort
Probability of getting skin irritation or redness	5% chance of skin irritation or redness (5 out of 100)	35% chance of skin irritation or redness (35 out of 100)
Glucose information	Current Glucose level	Current Glucose level and arrow
Alarms	Yes	No
Monthly costs	€25	€175
I prefer:	<input type="radio"/>	<input type="radio"/>

If you have to choose between the device you have chosen above and the traditional fingerprick-test to check your glucose levels, which one would you prefer? (Please note that a fingerprick-test should be done four times a day, requires high effort to check, does not result in skin irritation or redness, will show your glucose levels, doesn't have an alarm and costs €25 per month).

Select only one answer

I prefer the device I have selected above

I prefer the fingerprick-test

**Figure 1** Example discrete choice experiment choice task.

first chose between two hypothetical glucose monitors (device A or device B) and then they chose between that choice and a standard finger-prick test for their care (see figure 1).<sup>44,45</sup> This method mimics realistic choice scenarios while also ensuring data quality. The finger-prick test was always described as requiring four finger pricks per day, a high amount of effort to check blood glucose, no skin irritation or redness associated with a device on skin, showing glucose level only at time of measurement, no alarm and with out-of-pocket costs of €25 (or 55zl) per month. Participants were given two 'warm-up' DCE choice tasks before the main exercise started to ensure comprehension.

### Questionnaire

Prior to completing the DCE, the participants were given information describing glucose monitoring as a part of diabetes self-management, including the impact of uncontrolled blood glucose on health outcomes and a description of the attributes used in the DCE. Participants were asked to answer sociodemographic questions and disease-related questions including diabetes type, years since diagnosis, use of medication and questions related to their current diabetes self-care regimen. Two brief measures assessing subjective numeracy (the Shortened Subjective Numeracy Scale-3)<sup>46</sup> and health literacy (Brief Health Literacy Screener (Chew Items))<sup>47</sup> were included in the survey. The final survey was pretested in think-aloud interviews with n=6 patients with diabetes from the Netherlands. The outcomes of this pretest were used to reword the survey for understandability.

### Analysis

#### Data quality

Respondents were required to answer all questions, and only surveys that included all necessary questions for the final analysis were included. Completed responses were checked for flatlining (only choosing device A or device B) and speeding (respondents completing the survey faster than 70% of the mean response time based on log data) as data quality checks. Differences in sample demographics were assessed using  $\chi^2$  tests or t-tests where applicable. A significance of  $p < 0.05$  was used for all analyses.

#### Preferences

Data from the DCE was analyzed by combining the two questions from each choice task as a single observation (device A vs device B vs the finger-prick test). Preference estimates in each country were assessed independently using a panel mixed-effects logit regression to account for heterogeneity of preferences within patient populations.<sup>32</sup> Effects-coding was used for all variables except for cost, which was assumed to be linear.<sup>48</sup> Effects-coding allows for a calculation of the reference category coefficient, which can be used for comparison to other attributes and a clear interpretation of a constant term (reflecting the utility of a status-quo finger-prick test).<sup>48</sup> Robust outcomes were generated by applying 1000 Halton draws.<sup>49</sup> The analysis was conducted in Stata V.14.<sup>34</sup> The optimal model was identified based on log likelihood. Attributes with significant SD for at least one level were included as random effects in the final model. The following value functions were used for the final analyses (Equations 1–3):

$$V_{\text{Device A } i} = \beta_0 + \beta_{1i} \times \text{precision}_{0.3} + \beta_{2i} \times \text{precision}_{0.6} + \beta_{3i} \times \text{pricks per day}_{2x} + \beta_{4i} \times \text{effort}_{\text{moderate}} + \beta_{5i} \times \text{skin irritation}_{20\%} + \beta_{6i} \times \text{skin irritation}_{35\%} + \beta_{7i} \times \text{monthly costs} + \beta_{8i} \times \text{information}_{\text{arrow}} + \beta_{9i} \times \text{information}_{\text{trendline}} + \beta_{10i} \times \text{alarms}_{\text{none}} \quad (1)$$

$$V_{\text{Device B } i} = \beta_{1i} * \text{precision}_{0.3} + \beta_{2i} * \text{precision}_{0.6} + \beta_{3i} \times \text{pricks per day}_{2x} + \beta_{4i} \times \text{effort}_{\text{moderate}} + \beta_{5i} \times \text{skin irritation}_{20\%} + \beta_{6i} \times \text{skin irritation}_{35\%} + \beta_{7i} \times \text{monthly costs} + \beta_{8i} \times \text{information}_{\text{arrow}} + \beta_{9i} \times \text{information}_{\text{trendline}} + \beta_{10i} \times \text{alarms}_{\text{none}} \quad (2)$$

$$V_{\text{Fingerprick } i} = \beta_{11} \quad (3)$$

In these equations, the value of an alternative for individual  $i$  is calculated based on the coefficients reflecting the relative importance of each attribute or attribute level ( $\beta_1 - \beta_{10}$ ).  $\beta_{11}$  is an alternative specific constant reflecting the individual's preference for the fixed alternative of the finger-prick test over device B.  $\beta_0$  is a constant term which identifies the respondent's preferences for device A over device B, reflecting a left-right bias in case participants had a tendency to favor the left option. All attributes and

attribute levels were included as random parameters, with a normal distribution to identify heterogeneity in the preferences for those attributes.

The mixed logit model preference estimates were used to calculate the attribute relative importance score (RIS).<sup>50</sup> The RIS reflects how important one attribute is compared with another. These were calculated by identifying the attribute with the greatest absolute difference between highest and lowest valued levels and using this as a reference (RIS=1). The RIS for each attribute was then calculated as the quotient of the absolute difference of the most and least valued levels of that attribute and the reference value. This results in a normalized scale for comparison.

### WTP estimates and uptake rates

Individual attribute coefficient estimates were extracted from the mixed-effects models to calculate individual WTP estimates and expected uptake rates. WTP estimates were generated by calculating the utility difference between attribute levels and dividing this by the negative linear cost coefficient resulting in the estimated amount that each participant would be willing to pay for the change in attribute level. Very small cost coefficients for some participants led to extreme WTP outliers so the median and IQR are reported rather than the mean. Differences in median WTP estimates were assessed using Mood's test for equality of medians.<sup>51</sup>

Expected uptake rate estimates were calculated using the individual attribute coefficient estimates. Three device profiles represent potential glucose-monitoring devices were used to calculate uptake rates compared with a standard finger-prick test. The first profile represented the most desired device according to the outcomes of the mixed logit model: high precision, zero finger pricks per day, low effort to check, low chance of skin irritation, €25 per month out-of-pocket costs, glucose information with a daily trendline, and an alarm. The second profile was similar to a generic FGM: moderate precision, zero finger pricks, moderate effort, moderate chance of skin irritation, €100 per month out-of-pocket costs, glucose information with an arrow indicating glucose direction, and no alarm. The last profile used the generic FGM profile but changed the monthly out-of-pocket costs to €25. The uptake estimate was calculated at the individual level by taking the proportion of the individuals' (*i*) total utility, which was accountable to a device (*V*) in a scenario containing both this device and a finger-prick alternative (*W*) using the following equation:

Equation 4:

$$\sum_{i=1}^n \frac{e^{V_i}}{e^{V_i} + e^{Fingerprick_i}} \quad (4)$$

The mean of these expected uptake rate estimates was interpreted as the expected population uptake rate.

## RESULTS

In total, n=521 respondents completed the surveys. Of those, n=487 responses were included in the final analysis

**Table 2** Respondent characteristics (N=487)

Characteristics	Dutch respondents	Polish respondents
	n=226	n=261
Age (years)† (mean±SD)	51.6±17.2	39.4±13.4
Sex, n (%)		
Female	116 (51.3)	125 (47.9)
Male	110 (48.7)	134 (51.3)
Type of diabetes, n (%)		
Type 1	65 (28.8)	83 (31.8)
Type 2	158 (69.9)	167 (64.0)
Other	3 (1.3)	11 (4.2)
Number of years having diabetes, † mean±SD (median, range)	9.5±9.1 (6.5, 0–60)	6.1±7.1 (3, 0–53)
Current glucose monitor used as part of diabetes care†		
CGM or FGM	38 (16.8)	39 (14.9)
Finger-prick testing only	128 (56.6)	211 (80.8)
None	60 (26.5)	11 (4.2)
Checks glucose more than two times per day*	83 (31.8)	161 (71.2)
Uses insulin, n (%)	120 (53.1)	140 (53.6)
Health literacy, n (%)		
High	102 (45.1)	113 (43.3)
Low	124 (54.9)	148 (56.7)
Numeracy*		
High	195 (86.3)	243 (93.1)
Low	31 (13.7)	18 (6.9)
OECD educational level, † n (%)		
Tertiary	100 (44.2)	134 (51.3)
Upper secondary/vocational	114 (50.4)	127 (48.7)
Secondary or lower	12 (20.8)	0 (0.0)

\*Significant differences between countries at p<0.05.  
 †Significant differences between countries at p<0.001.  
 CGM, continuous glucose monitor; FGM, flash glucose monitor; OECD, Organisation for Economic Co-operation and Development; SD, standard deviation.

after n=34 (6.5%) respondents were excluded following a check of data quality. Participant demographic information can be found in [table 2](#). Compared with the Polish sample, the Dutch sample was significantly older (51.6 years vs 39.4 years), had lived with diabetes for more years, were less educated, had lower levels of health numeracy, and were less likely to monitor their blood glucose than the Polish sample. No other significant differences were found between the samples.

### Preferences for glucose monitors

All attributes were found to be significant for patients in at least one of the countries. Significant heterogeneity of preferences was found for all attributes except for type

of glucose information. High costs were associated with a lower likelihood of choosing a device. Increased precision was preferred over lower precision, and decreased number of finger pricks and chance of skin irritation were consistently favored over increases in these attributes. Samples from both countries favored a device with an alarm over one without an alarm. Improving a device's effort to check from moderate to low and improving glucose information to show more than only current levels were only important for the Dutch respondents. Both samples preferred glucose-monitoring devices over a finger-prick test. The complete results of the mixed logit model can be found in [table 3](#).

Regarding the RIS of the attributes, costs were found to be the most important factor when choosing a device by a factor of 5, compared with the next most important attribute, and a factor of approximately 50, compared with the least important attribute (online supplemental figure 2). For the Dutch sample, after costs, the most important attributes were number of finger pricks, followed by precision and chance of skin irritation, all of which were comparably valued. For the Polish population, after costs, precision of device was the second most important attribute followed by chance of skin irritation. These were also comparably valued. Polish patients were not as averse to additional finger pricks as Dutch respondents and found this approximately half as important as Dutch respondents. However, Polish respondents valued switching to a device from a finger-prick test more than Dutch respondents. Having an alarm and improving glucose information were both relatively unimportant in a device. Only the Dutch sample viewed improved effort to check and the type of glucose information as important when deciding on a device.

#### WTP for a glucose monitor and expected uptake rates

WTP results can be found in [table 4](#). It was estimated that Polish patients would pay significantly more to switch from a standard finger prick to a device than Dutch patients (€65.01 vs €27.74 per month). The median WTP for improvements in glucose monitors ranged from €2.58 (for the Dutch respondents to improve glucose information) to €33.64 (for the Polish respondents to increase precision from low to high). Significant differences were found between the two countries with Dutch respondents having higher WTP for device improvements in precision from low to medium, improving effort to check, and improving glucose information. Dutch patients were also willing to pay significantly more for a reduction in the number of finger pricks per day in conjunction with a device compared with Poland (€32.71 vs €13.35).

These differences were also reflected in the expected uptake rates for devices. Polish patients were significantly more likely to choose a device over finger prick ([table 4](#)) compared with Dutch patients. These differences were most pronounced in patients aged 18–50, patients with T2DM, and current finger prick-only users.

#### DISCUSSION

To the best of our knowledge, this was the first study to investigate the relative importance of different attributes describing glucose-monitoring technologies which involved cost as an attribute. As expected, cost was found to be the most important factor for patients when deciding on glucose monitors in both the Netherlands and Poland. Increased device precision, reduction in skin irritation, and required number of finger pricks per day were the next most important attributes when choosing between glucose monitors.

The findings from this study replicate some of the findings of earlier studies,<sup>18 22</sup> but the current study enables us to show that costs were at least more than five times more important for patients when choosing a glucose monitor than any other attribute. As costs are the primary consideration for patients when deciding to use a glucose-monitoring device or a standard finger prick, it may not be a question of WTP, but the ability to pay that is determining glucose monitor choice.<sup>18 52–54</sup> This is unfortunate as the improvements in diabetes outcomes, patient quality of life, and healthcare expenditures in connection with using these devices are increasingly documented.<sup>17 54–60</sup>

Beyond costs, the relative importance of the other attributes differed to some degree between the two countries. Specifically, Dutch respondents valued reducing the number of daily finger pricks to zero more than twice as much as Polish respondents. The acceptance of additional finger pricks to verify blood glucose levels may reflect the greater importance that Polish respondents assigned to precision as these finger pricks are the most accurate reading and can be used for calibration of devices or verification of device glucose information. For both populations, precision was mainly significant when the device was described as having higher levels of imprecision. Lower levels of imprecision were not important for choosing a device, indicating that there is an acceptable amount of device imprecision. This was also reported by patients during the qualitative phase.

These preference differences resulted in different in WTP for glucose devices and expected uptake rates for the two countries. Both samples reflected an overall desire to move away from finger-prick tests for SMBG, although this was more pronounced in the Polish population compared with the Dutch population. Patients were consistently willing to pay for device improvements that resulted in devices that more closely represented FGMs or CGMs regarding functionality.

While we found type of information to be relatively less important based on the model outcomes, this conflicts with the findings from the qualitative phase of this study. During the interviews, stakeholders from every area including the patients indicated that only having the current glucose level was insufficient for proper glucose management. In the preference study outcomes, improvements in this area were not nearly as important for choosing a device as the interviews would have led

**Table 3** Attribute-level estimates for the panel mixed logit model

Attribute	Levels		The Netherlands			Poland		
			Estimate	SE	P sig.	Estimate	SE	P sig.
Precision compared with finger pricking	Accurate as finger pricking (ref)	Mean	0.343	0.075	***	0.457	0.071	***
		SD						
	±0.3 mmol/L	Mean	0.000	0.061		−0.081	0.051	
		SD	0.036	0.101		0.047	0.112	
		Mean	−0.343	0.079	***	−0.376	0.073	***
±0.6 mmol/L	SD	0.536	0.093	***	0.655	0.082	***	
	Mean							
Average number of finger pricks per day	0 times per day (ref)	Mean	0.352	0.059	***	0.172	0.044	***
		SD						
	2 times per day	Mean	−0.352	0.059	***	−0.172	0.044	***
		SD	0.479	0.070	***	0.349	0.059	***
		Mean						
Effort to check	Low (ref)	Mean	0.120	0.039	**	0.042	0.033	
		SD						
	Moderate	Mean	−0.120	0.039	**	−0.042	0.033	
		SD						
		Mean						
Probability of getting skin irritation or redness	5% (ref)	Mean	0.336	0.076	***	0.377	0.064	***
		SD						
	20%	Mean	−0.059	0.066		−0.018	0.059	
		SD	0.061	0.127		0.015	0.166	
		Mean	−0.277	0.076	***	−0.359	0.066	
35%	SD	0.450	0.097	***	0.402	0.084	***	
	Mean							
Monthly costs	per €1 increase	Mean	−0.017	0.002	***	−0.016	0.001	***
		SD	0.015	0.001	***	0.019	0.001	***
		Mean						
Glucose information	Current glucose level only (ref)	Mean	−0.142	0.063	*	−0.056	0.054	
		SD						
	Current glucose level with arrow	Mean	0.068	0.063		0.004	0.055	
		SD						
	Current glucose level with daily trendline	Mean	0.074	0.063		0.052	0.053	
		SD						
Alarms	Yes (ref)	Mean	0.152	0.044	***	0.148	0.035	***
		SD						
	No	Mean	−0.152	0.044	***	−0.148	0.035	***
		SD	0.252	0.076	***	0.151	0.063	*
		Mean						
Alternative specific constant for device instead of finger-prick test	Mean	−0.982	0.502		−2.770	0.336	***	
	SD	4.527	0.386	***	4.767	0.371	***	
Alternative specific constant indicating left–right bias	Mean	0.376	0.085	***	0.346	0.074	***	
	SD	0.446	0.140	***	0.540	0.098	***	

Higher estimates represent increasing levels of importance for the patient in choosing a device.

All attributes were effects-coded, enabling the direct comparison of the estimates. The sum of the effect coded attributes is zero, and therefore the coefficient of the reference category can be easily calculated, and the relative importance of the reference categories of the attributes can be compared with one another, and so that the alternative specific constants have independent interpretation signifying the average utility for that alternative. SDs are given where parameters were found to have a significant random parameter estimate. The significant alternative specific constant indicates that patients tended to choose the alternative on the left side. A normal distribution using 1000 Halton draws was used in model development.

\*P<0.05, \*\*P<0.01, \*\*\*P<0.001.

ref, reference level; SD, standard deviation; SE, standard error.

**Table 4** Median WTP estimates and average uptake rates compared with traditional finger prick

	The Netherlands WTP		Poland WTP	
	Median	IQR	Median	IQR
Increase precision from				
Low to medium	15.94	27.73–3.93	10.18	–3.19 to 21.23***
Medium to high	15.87	3.72–26.43	22.17	4.54–41.84
Low to high	31.82	7.60–54.73	33.64	1.02–59.45
Reduce daily finger-pricks: 2–0	32.71	14.34–63.41	13.35	4.28–30.34***
Improve chance of skin irritation				
High to medium	10.39	18.86–5.56	13.8	6.38–27.87**
Medium to low	18.7	12.52–32.79	16.06	8.20–34.52*
High to low	28.97	16.55–52.5	29.83	14.47–61.74
Improve effort from medium to low	11.32	8.78–22.24	3.55	2.31–7.4***
Improve glucose information with				
An arrow showing blood glucose is changing	10.22	7.70–19.49	2.58	1.68–5.38***
Daily trend information	14.19	7.92–20.07	4.6	2.99–9.59***
Get a glucose alarm	14.19	7.15–25.74	12.67	7.44–24.02
WTP to not use finger-prick test	27.74	–231.85 to 278.23	65.01	–183.76 to 295.5**
<b>Estimated Uptake Rates</b>				
		<b>Most preferred device (%)†</b>	<b>FGM proxy device (%)‡</b>	<b>FGM proxy with reduced cost (%)§</b>
Total samples				
The Netherlands (n=226)		63.6	44.4	54.8
Poland (n=261)		77.1***	56.1**	67.6***
Age 18–50				
Netherlands (n=88)		69.1	51.7	59.9
Poland (n=202)		78.9*	59.5***	69.5*
Age 50 and over				
Netherlands (n=137)		60.4	39.9	51.9
Poland (n=59)		70.9	44.3	61.1
FP only users				
Netherlands (n=128)		57.5	37.8	48
Poland (n=211)		75***	53.6***	65.8***
CGM/FGM users				
Netherlands (n=38)		78.2	66.7	73
Poland (n=39)		85.3	70.2	76.7
Type 1				
Netherlands (n=65)		70.8	53.1	62
Poland (n=83)		81.7	63.2	71.7
Type 2				
Netherlands (n=157)		61	41	52.3
Poland (n=167)		74.8**	52.8**	65.6**

Note for WTP estimates: estimates are only presented for attribute improvement where level increases were found to be significant in the mixed logit model; costs presented in euros; Mood's test of equality of median values was used to assess difference between countries.

Significant differences between countries: \* $p < 0.05$ , \*\* $p < 0.01$ , \*\*\* $p < 0.001$ .

†High precision, 0 finger pricks, low effort, low chance of skin irritation, €25 /month, glucose information with trendline, alarm.

‡Moderate precision, 0 finger pricks, moderate effort, moderate chance of skin irritation, €100/month, arrow information, no alarm.

§Moderate precision, 0 finger pricks, moderate effort, moderate chance of skin irritation, €25/month, arrow information, no alarm.

CGM, continuous glucose monitor; FGM, flash glucose monitor; WTP, willingness to pay.



us to believe. In addition to this, industry interviewees and patients reported that connectivity to devices which the patients normally carried around (eg, smartphone and smartwatch) was a desirable feature as it reduces effort to check and the stigma of checking blood glucose levels. The preference outcomes indicated that while the Dutch patients significantly preferred a device with low burden, the added benefit of accessing this information on a smartphone or watch instead of a dedicated device was relatively limited compared with other features or costs. This indicates that connectivity is something that is a want but not a must in a device. Exceptions to this may be in specific instances, such as parents who want to be able to monitor a child's glucose level at a distance.<sup>61 62</sup>

Our case study focused on two countries, the Netherlands and Poland, which are examples of 'Western' and 'Eastern' European countries with partial and no reimbursement for glucose-monitoring devices supporting the transferability of these findings to other countries with out-of-pocket costs for SMBG. At the time of designing the study, the reimbursement for CGMs was limited in the Netherlands with FGMs not fully reimbursed.<sup>21</sup> The reimbursement policy of Dutch insurance companies changed while the study was being conducted to allow patients with T1DM, patients with T2DM with intense insulin regimens, and patients with T2DM who are pregnant or trying to become pregnant to be eligible for FGMs through their health insurance. CGMs and FGMs were not reimbursed in Poland at the time that the study was conducted, and to the best of our knowledge are still not reimbursed.<sup>63 64</sup> Respondent awareness of the change in reimbursement in the Netherlands may have resulted in lower WTP estimates. It would be interesting to study how improved access to these devices for some patients has changed preferences in Dutch patient populations and if the removal of cost as an attribute impacts their preferences compared with Poland without a change in reimbursement. The removal of cost as a barrier would likely have a large impact on patient preferences and expected uptake rates of these devices with a greater focus on how the device fits into the patient's lifestyle as reflected in the study by Hannah *et al.*<sup>22</sup>

The strengths of this study include the extensive qualitative phase used to identify the relevant attributes for use in the DCE. This process was more extensive than what is commonly done to generate attributes in preference studies. Interviewees were internationally diverse with a broad range of backgrounds and contributed to the identification of a set of attributes relevant to a broad sample of patients. Another strength of this study is the multi-country sampling, which allows for a better understanding of the transferability of these findings to diabetes patient populations in other countries. This study did have some limitations. First, the study used data that relied on self-reports of diabetes diagnosis and no quotas based on SES were imposed. This limited exploring subgroup analyses of SES group preferences which may be relevant as SES has previously been associated with adherence to SMBG.

Second, patients were recruited through an online panel only and not through clinical partners or patient organizations due to COVID-19-related restrictions on all non-vital research. This resulted in a sample of respondents that had generally higher levels of education and were younger than we would expect from the general diabetes population.<sup>65–68</sup> The results of a more representative sample may produce different relative preference outcomes as we found differences in expected uptake rates based on age stratifications.

## CONCLUSION

While patients value many aspects of glucose monitors, out-of-pocket costs are the primary concern of patients when deciding on devices to self-monitor blood glucose. Even when different welfare levels between the two countries were accounted for, differences in estimated WTP were found between the countries. This study shows that uptake of modern glucose-monitoring devices is dependent on out-of-pocket costs. In light of these clear preferences to switch from glucose measurement by finger pricks to more modern equipment, a critical review of the costs and benefits of such devices is needed to see if removing the cost barrier is justified by the potential improvements in blood glucose monitoring.

## Author affiliations

<sup>1</sup>Healthcare Innovation and Evaluation, University Medical Center Utrecht, Julius Center for Health Sciences and Primary Care, Utrecht, The Netherlands

<sup>2</sup>Evidera, London, UK

<sup>3</sup>Erasmus School of Health Policy and Management, Erasmus Universiteit Rotterdam, Rotterdam, The Netherlands

<sup>4</sup>Erasmus Choice Modelling Centre, Erasmus University Rotterdam, Rotterdam, The Netherlands

<sup>5</sup>Health Technology and Services Research (HTSR), Faculty of Behavioural Management and Social Sciences, University of Twente, Enschede, The Netherlands

<sup>6</sup>Public Health and Primary Care/LUMC Campus The Hague, Leiden University Medical Center, Leiden, The Netherlands

**Acknowledgements** The authors of this article acknowledge the following people for their contributions to the study: the patients for participating and sharing their valuable insights on the interviews and survey, Angela de Rooij and Anna Sliwinska for their insights into patient values and experiences in the Netherlands and Poland, Maria van Dijk-Okla for her help in translating and reviewing the Polish survey, and the PREFER consortium for their advice and guidance throughout the study.

**Contributors** IPS, CLW, JV, MPMHR-vM, RCV, EWdB-G and GAdW were responsible for the initial development, set-up and design of the study. IPS and CLW were responsible for the qualitative aspects of the study as well as quantitative data collection under the direct supervision of GAdW, JV, EWdB-G, and MPMHR-vM. The primary analysis was conducted by IPS with assistance from JV and CGMG-O. All authors reviewed the final manuscript and contributed to the revisions. IPS is the guarantor for this article and accepts full responsibility for the work and/or the conduct of the study, had access to the data, and controlled the decision to publish.

**Funding** This study formed part of the Patient Preferences in Benefit-Risk Assessments during the Drug Life Cycle (PREFER) project. The PREFER project has received funding from the Innovative Medicines Initiative 2 Joint Undertaking (grant agreement number 115966). This joint undertaking receives support from the European Union's Horizon 2020 research and innovation programme and the European Federation of Pharmaceutical Industries and Associations.

**Competing interests** None declared.

**Patient consent for publication** Not applicable.

**Ethics approval** This study involves human participants and was reviewed by the medical research ethics committee of the UMC Utrecht in 2019 and found to present limited risk to participants according to the Medical Research Involving Human Subjects Act 1998 (Wet Medisch-Wetenschappelijk Onderzoek met Mensen). Thus, the study was thus waived from full medical ethical review and was approved to be conducted (reference number WAG/mb/19/045208). The study was conducted according to the principles of the Declaration of Helsinki. All participants were informed about the study through written materials and provided written informed consent prior to participating in the study.

**Provenance and peer review** Not commissioned; externally peer reviewed.

**Data availability statement** Data are available upon reasonable request.

**Supplemental material** This content has been supplied by the author(s). It has not been vetted by BMJ Publishing Group Limited (BMJ) and may not have been peer-reviewed. Any opinions or recommendations discussed are solely those of the author(s) and are not endorsed by BMJ. BMJ disclaims all liability and responsibility arising from any reliance placed on the content. Where the content includes any translated material, BMJ does not warrant the accuracy and reliability of the translations (including but not limited to local regulations, clinical guidelines, terminology, drug names and drug dosages), and is not responsible for any error and/or omissions arising from translation and adaptation or otherwise.

**Open access** This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

#### ORCID iDs

Ian P Smith <http://orcid.org/0000-0002-1236-2872>

Rimke C Vos <http://orcid.org/0000-0003-1074-6255>

#### REFERENCES

- American Diabetes Association. Diagnosis and classification of diabetes mellitus. *Diabetes Care* 2012;35 Suppl 1:S64–71.
- Sun H, Saeedi P, Karuranga S, et al. IDF diabetes atlas: global, regional and country-level diabetes prevalence estimates for 2021 and projections for 2045. *Diabetes Res Clin Pract* 2022;183:109119.
- Saeedi P, Salpea P, Karuranga S, et al. Mortality attributable to diabetes in 20–79 years old adults, 2019 estimates: Results from the International Diabetes Federation Diabetes Atlas, 9<sup>th</sup> edition. *Diabetes Res Clin Pract* 2020;162:108086.
- Evert AB, Boucher JL, Cypress M, et al. Nutrition therapy recommendations for the management of adults with diabetes. *Diabetes Care* 2014;37 Suppl 1:S120–43.
- Rubin RR, Peyrot M. Psychological issues and treatments for people with diabetes. *J Clin Psychol* 2001;57:457–78.
- Canadian Agency for Drugs and Technologies in Health (CADTH). Systematic review of use of blood glucose test strips for the management of diabetes mellitus. *CADTH Technol Overv* 2010;1:e0101.
- Benjamin EM. Self-Monitoring of blood glucose: the basics. *Clinical Diabetes* 2002;20:45–7.
- Ahola AJ, Groop P-H. Barriers to self-management of diabetes. *Diabet Med* 2013;30:413–20.
- Dalewitz J, Khan N, Hershey CO. Barriers to control of blood glucose in diabetes mellitus. *Am J Med Qual* 2000;15:16–25.
- Peyrot M, Rubin RR, Lauritzen T, et al. Psychosocial problems and barriers to improved diabetes management: results of the cross-national diabetes attitudes, wishes and needs (dawn) study. *Diabet Med* 2005;22:1379–85.
- Bruce DG, Davis WA, Cull CA, et al. Diabetes education and knowledge in patients with type 2 diabetes from the community: the Fremantle diabetes study. *J Diabetes Complications* 2003;17:82–9.
- Moström P, Ahlén E, Imberg H, et al. Adherence of self-monitoring of blood glucose in persons with type 1 diabetes in Sweden. *BMJ Open Diabetes Res Care* 2017;5:e000342.
- Harris MI, Cowie CC, Howie LJ. Self-monitoring of blood glucose by adults with diabetes in the United States population. *Diabetes Care* 1993;16:1116–23.
- Hankó B, Kázmér M, Kumli P, et al. Self-reported medication and lifestyle adherence in Hungarian patients with type 2 diabetes. *Pharm World Sci* 2007;29:58–66.
- Ong WM, Chua SS, Ng CJ. Barriers and facilitators to self-monitoring of blood glucose in people with type 2 diabetes using insulin: a qualitative study. *Patient Prefer Adherence* 2014;8:237.
- Garg SK, Akturk HK. *The future of continuous glucose monitoring*. New York: Mary Ann Liebert, Inc, 2017: 19. S-1–S-2.
- Ajjan RA. How can we realize the clinical benefits of continuous glucose monitoring? *Diabetes Technol Ther* 2017;19:S-27–S-36.
- Engler R, Routh TL, Lucisano JY. Adoption barriers for continuous glucose monitoring and their potential reduction with a fully implanted system: results from patient preference surveys. *Clin Diabetes* 2018;36:50–8.
- Carlson AL, Mullen DM, Bergenstal RM. Clinical use of continuous glucose monitoring in adults with type 2 diabetes. *Diabetes Technol Ther* 2017;19:S-4–S-11.
- Taylor PJ, Thompson CH, Brinkworth GD. Effectiveness and acceptability of continuous glucose monitoring for type 2 diabetes management: a narrative review. *J Diabetes Investig* 2018;9:713–25.
- Heinemann L, Franc S, Phillip M, et al. Reimbursement for continuous glucose monitoring: a European view. *J Diabetes Sci Technol* 2012;6:1498–502.
- Hannah K, Lich R, Nair K, et al. PDB40 eliciting patient preferences for continuous glucose monitoring devices in type 1 diabetes. *Value in Health* 2021;24:S85.
- Harris MI, National Health and Nutrition Examination Survey (NHANES III). Frequency of blood glucose monitoring in relation to glycemic control in patients with type 2 diabetes. *Diabetes Care* 2001;24:979–82.
- Zhang Y, Coello PA, Brozek J, et al. Using patient values and preferences to inform the importance of health outcomes in practice Guideline development following the grade approach. *Health Qual Life Outcomes* 2017;15:52.
- de Bekker-Grob EW, Ryan M, Gerard K. Discrete choice experiments in health economics: a review of the literature. *Health Econ* 2012;21:145–72.
- Soekhai V, de Bekker-Grob EW, Ellis AR, et al. Discrete choice experiments in health economics: past, present and future. *Pharmacoeconomics* 2019;37:201–26.
- Lancaster KJ. A new approach to consumer theory. *J Polit Econ* 1966;74:132–57.
- Manski CF. The structure of random utility models. *Theory Decis* 1977;8:229–54.
- McFadden D. Econometric models of probabilistic choice. *The Journal of Business* 1981;53:S13–29.
- Gerard K, Shanahan M, Louviere J. Using Discrete Choice Modelling to Investigate Breast Screening Participation. In: Ryan M, Gerard K, Amaya-Amaya M, eds. *Using discrete choice experiments to value health and health care*. Springer Netherlands: Dordrecht, 2008: 117–37.
- Ryan M, Gerard K, Amaya-Amaya M. *Using discrete choice experiments to value health and health care*. Springer Science & Business Media, 2007.
- Hensher DA, Greene WH. The mixed Logit model: the state of practice. *Transportation* 2003;30:133–76.
- Hensher DA, Rose JM, Rose JM, et al. *Applied choice analysis: a primer*. Cambridge university press, 2005.
- Louviere JJ, Hensher DA, Swait JD. *Stated choice methods: analysis and applications*. Cambridge university press, 2000.
- Reed Johnson F, Lancsar E, Marshall D, et al. Constructing experimental designs for discrete-choice experiments: report of the ISPOR conjoint analysis experimental design good research practices Task force. *Value Health* 2013;16:3–13.
- Bridges JFP, Hauber AB, Marshall D, et al. Conjoint analysis applications in health—a checklist: a report of the ISPOR Good Research Practices for Conjoint Analysis Task Force. *Value Health* 2011;14:403–13.
- FDA. *Blood glucose monitoring test systems for prescription point-of-care use*. Food and Drug Administration (FDA), 2018.
- Carlson AL, Mullen DM, Bergenstal RM. Clinical use of continuous glucose monitoring in adults with type 2 diabetes. *Diabetes Technol Ther* 2017;19:S4–S11.
- Heinemann L, DeVries JH. Reimbursement for continuous glucose monitoring. *Diabetes Technol Ther* 2016;18:S2-48–S2-52.
- Allen NA, Fain JA, Braun B, et al. Continuous glucose monitoring in non-insulin-using individuals with type 2 diabetes: acceptability, feasibility, and teaching opportunities. *Diabetes Technol Ther* 2009;11:151-8.
- OECD. *Purchasing power parities (PPP) (indicator)*, 2022.
- Kanninen BJ. Optimal design for multinomial choice experiments. *Journal of Marketing Research* 2002;39:214–27.
- Rose JM, Bliemer MCJ. Constructing efficient stated choice experimental designs. *Transport Reviews* 2009;29:587–617.
- Ghijben P, Lancsar E, Zavarsek S. Preferences for oral anticoagulants in atrial fibrillation: a best-best discrete choice experiment. *Pharmacoeconomics* 2014;32:1115–27.

- 45 Veldwijk J, Lambooi MS, de Bekker-Grob EW, *et al.* The effect of including an opt-out option in discrete choice experiments. *PLoS One* 2014;9:e111805.
- 46 Fagerlin A, Zikmund-Fisher BJ, Ubel PA, *et al.* Measuring numeracy without a math test: development of the subjective Numeracy scale. *Med Decis Making* 2007;27:672–80.
- 47 Chew LD, Bradley KA, Boyko EJ. Brief questions to identify patients with inadequate health literacy. *Fam Med* 2004;36:12.
- 48 Bech M, Gyrd-Hansen D. Effects coding in discrete choice experiments. *Health Econ* 2005;14:1079–83.
- 49 Train KE. Drawing from Densities. In: Train KE, ed. *Discrete choice methods with simulation*. 2 ed. Cambridge: Cambridge University Press, 2009: 205–36.
- 50 Gonzalez JM. A guide to measuring and interpreting attribute importance. *Patient* 2019;12:287–95.
- 51 Mood AM. *Introduction to the theory of statistics*, 1950.
- 52 Zgibor JC, Simmons D. Barriers to blood glucose monitoring in a multiethnic community. *Diabetes Care* 2002;25:1772–7.
- 53 Divan V, Greenfield M, Morley CP, *et al.* Perceived burdens and benefits associated with continuous glucose monitor use in type 1 diabetes across the lifespan. *J Diabetes Sci Technol* 2022;16:88–96.
- 54 Halford J, Harris C. Determining clinical and psychological benefits and barriers with continuous glucose monitoring therapy. *Diabetes Technol Ther* 2010;12:201–5.
- 55 Bonora B, Maran A, Ciciliot S, *et al.* Head-To-Head comparison between flash and continuous glucose monitoring systems in outpatients with type 1 diabetes. *J Endocrinol Invest* 2016;39:1391–9.
- 56 Dover AR, Stimson RH, Zammit NN, *et al.* Flash glucose monitoring improves outcomes in a type 1 diabetes clinic. *J Diabetes Sci Technol* 2017;11:442–3.
- 57 Oyagüez I, Merino-Torres JF, Brito M, *et al.* Cost analysis of the flash monitoring system (FreeStyle Libre 2) in adults with type 1 diabetes mellitus. *BMJ Open Diabetes Res Care* 2020;8:e001330.
- 58 Leelarathna L, Wilmot EG. Flash forward: a review of flash glucose monitoring. *Diabet Med* 2018;35:472–82.
- 59 Kelly B. FreeStyle Libre use in a real-world population: the Southampton City experience. *Journal of Diabetes Nursing* 2019;23:71.
- 60 Leinung M, Thompson S, Nardacci E. Benefits of continuous glucose monitor use in clinical practice. *Endocr Pract* 2010;16:371–5.
- 61 Elbalshy M, Boucher S, Crocket H, *et al.* Exploring parental experiences of using a do-it-yourself solution for continuous glucose monitoring among children and adolescents with type 1 diabetes: a qualitative study. *J Diabetes Sci Technol* 2020;14:844–53.
- 62 Gammon D, Årsand E, Walseth OA, *et al.* Parent-Child interaction using a mobile and wireless system for blood glucose monitoring. *J Med Internet Res* 2005;7:e57.
- 63 Hohendorff J, Gumprecht J, Mysliwiec M, *et al.* Intermittently Scanned continuous glucose monitoring data of Polish patients from real-life conditions: more scanning and better glycemic control compared to worldwide data. *Diabetes Technol Ther* 2021;23:577–85.
- 64 Głowińska-Olszewska B, Tobiaszewska M, Łuczyński W, *et al.* Monthly use of a real-time continuous glucose monitoring system as an educational and motivational tool for poorly controlled type 1 diabetes adolescents. *Adv Med Sci* 2013;58:344–52.
- 65 OECD. Education at a glance: OECD indicators Poland, 2019. Available: [https://www.oecd.org/education/education-at-a-glance/EAG2019\\_CN\\_POL.pdf](https://www.oecd.org/education/education-at-a-glance/EAG2019_CN_POL.pdf) [Accessed 19 Jan 2021].
- 66 Rijksinstituut voor Volksgezondheid en Milieu. "Diabetes mellitus: Opleiding." Rijksinstituut voor Volksgezondheid en Milieu, 2021. Available: <https://www.vzinfo.nl/diabetes-mellitus#node-prevalentie-diabetes-huisartsenpraktijk-naar-opleiding> [Accessed 19 Jan 2021].
- 67 Rijksinstituut voor Volksgezondheid en Milieu. Diabetes mellitus: Leef tijd en geslacht. Available: <https://www.vzinfo.nl/diabetes-mellitus/leef-tijd-en-geslacht> [Accessed 19 Jan 2021].
- 68 Polakowska M, Piotrowski W. Incidence of diabetes in the Polish population: results of the Multicenter Polish Population Health Status Study--WOBASZ. *Pol Arch Med Wewn* 2011;121:156–63.