

RESEARCH ARTICLE

3D printed microchannels for sub-nL NMR spectroscopy

E. Montinaro¹, M. Grisi¹, M. C. Letizia¹, L. Pethö², M. A. M. Gijs¹, R. Guidetti³, J. Michler², J. Brugger¹, G. Boero^{1*}

1 Ecole Polytechnique Fédérale de Lausanne (EPFL), Laboratory for Microsystems, Lausanne, Switzerland, **2** Swiss Federal Laboratories for Materials Science and Technology (EMPA), Laboratory for Mechanics of Materials and Nanostructures, Thun, Switzerland, **3** University of Modena and Reggio Emilia, Department of Life Sciences, Modena, Italy

* giovanni.boero@epfl.ch



Abstract

Nuclear magnetic resonance (NMR) experiments on subnanoliter (sub-nL) volumes are hindered by the limited sensitivity of the detector and the difficulties in positioning and holding such small samples in proximity of the detector. In this work, we report on NMR experiments on liquid and biological entities immersed in liquids having volumes down to 100 pL. These measurements are enabled by the fabrication of high spatial resolution 3D printed microfluidic structures, specifically conceived to guide and confine sub-nL samples in the sub-nL most sensitive volume of a single-chip integrated NMR probe. The microfluidic structures are fabricated using a two-photon polymerization 3D printing technique having a resolution better than 1 μm^3 . The high spatial resolution 3D printing approach adopted here allows to rapidly fabricate complex microfluidic structures tailored to position, hold, and feed biological samples, with a design that maximizes the NMR signals amplitude and minimizes the static magnetic field inhomogeneities. The layer separating the sample from the microcoil, crucial to exploit the volume of maximum sensitivity of the detector, has a thickness of 10 μm . To demonstrate the potential of this approach, we report NMR experiments on sub-nL intact biological entities in liquid media, specifically ova of the tardigrade *Richtersius coronifer* and sections of *Caenorhabditis elegans* nematodes. We show a sensitivity of 2.5×10^{13} spins/ $\text{Hz}^{1/2}$ on ^1H nuclei at 7 T, sufficient to detect 6 pmol of ^1H nuclei of endogenous compounds in active volumes down to 100 pL and in a measurement time of 3 hours. Spectral resolutions of 0.01 ppm in liquid samples and of 0.1 ppm in the investigated biological entities are also demonstrated. The obtained results may indicate a route for NMR studies at the single unit level of important biological entities having sub-nL volumes, such as living microscopic organisms and eggs of several mammals, humans included.

OPEN ACCESS

Citation: Montinaro E, Grisi M, Letizia MC, Pethö L, Gijs MAM, Guidetti R, et al. (2018) 3D printed microchannels for sub-nL NMR spectroscopy. PLoS ONE 13(5): e0192780. <https://doi.org/10.1371/journal.pone.0192780>

Editor: David T. Eddington, University of Illinois at Chicago, UNITED STATES

Received: October 17, 2017

Accepted: January 30, 2018

Published: May 9, 2018

Copyright: © 2018 Montinaro et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: All relevant data are within the paper and its Supporting Information files.

Funding: EPFL (Laboratoire de microsystemes 1 École polytechnique fédérale de Lausanne) funded this research to EM. The funder had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Competing interests: The authors have declared that no competing interests exist.

Introduction

Methods based on the nuclear magnetic resonance (NMR) phenomenon are widely used in physics, chemistry, medicine and biology.[1–3] During an NMR experiment, the sample is

placed in a static magnetic field and excited with electromagnetic fields at frequencies and strengths which have no biological effects, if the pulse sequence is chosen to avoid heating issues. Due to this property and to its resolving power, NMR is successfully applied, e.g., to diagnostic imaging[4,5] and in-vivo spectroscopy[6–8] of large living animals. The use of NMR methodologies for the study of sub- μL volumes is hindered by sensitivity limitations. The search for methods enabling the extensions of this powerful technique to the study of smaller volumes is an active research domain. These efforts include the miniaturization of inductive methods[9–42] as well as the use of more sensitive but less versatile non-inductive approaches.[36,43,44] Various techniques permitted the optimization of NMR inductive detectors for volume ranging from 1 μL down to a few nL. Some of these were used to perform pioneering studies of small collections of microorganisms,[21,45–47] perfused tumor spheroids,[33] and large single cells and embryos.[48–54] NMR-based studies of intact single biological entities were, until recently, demonstrated down to volumes of 5 nL[47,50,51] whereas typical volumes of most cells and microorganisms are below the nL scale.[55] Recently, we reported the use of ultra-compact single-chip NMR probes as convenient tools to deliver state-of-art spin sensitivity for sub-nL volumes.[32,40] Such probes, entirely realized on a single 1 mm^2 complementary-metal-oxide-semiconductor (CMOS) microchip, consist of a multilayer microcoil and a co-integrated low-noise electronic transceiver. Thanks to the achieved sensitivity, we were able to perform the first NMR spectroscopy studies of single untouched sub-nL ova of microorganisms having active volumes down to 0.1 nL.[40] These experiments were conducted by manually placing the sample on top of the microcoil using a polystyrene cup filled with agarose gel.[40] This study demonstrated the detection of 10 pmol of ^1H nuclei in a measurement time of 3 hours, but improvements concerning sample manipulation are required to enable long-lasting experiments in more biocompatible environments. In particular, the study of biological samples in liquids would facilitate a non-invasive sample handling and it would provide a more biocompatible and controllable environment during the experiments. To perform NMR spectroscopy studies on small biological entities in a liquid environment, relatively complex microfabrication techniques are required, in particular for the placing and holding of the samples in the sensitive region of the miniaturized detector. Several approaches to combine microfluidic structures with microsolenoids[10,21,28,31,56,57] and planar microcoils[16,56,58] have been reported to date. Various techniques were developed to shrink the size of solenoids and pattern them on capillaries[42,59–61] or around hollow pillars via wire bonding.[28,39] At the sub-nL volume scales, an efficient combination of microsolenoids with microfluidic structures would require significantly more complex microfabrication processes than the one for planar microcoils with planar microfluidic structures.

In this work, we report on the design and fabrication of high spatial resolution microfluidic structures capable to position and hold sub-nL biological entities in the sub-nL sensitive volume of a planar microcoil, used for NMR signal excitation/detection, co-integrated on the same silicon chip with the transceiver electronics (Fig 1A). The single-chip CMOS detector offer a robust planar working surface, but its very small detection volume (about 0.2 nL) sets challenging fabrication constraints for the microfluidic design, which must hold the sample in close proximity of the microcoil without introducing significant static magnetic field inhomogeneities. To overcome these challenges, we fabricated microfluidic structures using a high spatial resolution 3D printer (Photonic Professionals GT, Nanoscribe GmbH, Germany), based on a two-photon polymerization process [62,63] and having a resolution better than 1 μm^3 . With this approach, we managed to fabricate microchannels capable to confine the samples under investigation at distances of about 10 μm from the microcoil surface. The proximity between the microcoil and the sample confined in the microfluidic channel is crucial to preserve a high effective spin sensitivity. To demonstrate the validity of our approach, we




Fig 1. NMR single chip detector, maps of sensitivity and illustration of the sample placement. (a) Microphotograph of the NMR single-chip detector, fabricated using a 130 nm CMOS technology from STMicroelectronics. The integrated planar microcoil is realized using four copper metal layers. The total number of turns is 22 and the outer diameter is 150 μm (see inset). The details of the microcoil and of the co-integrated electronics are reported in Ref. (33). (b) Sensitivity maps of the microcoil (the sensitivity is defined as $B_{xy}(\mathbf{r})\sin(\theta)$, see main text). The microcoil lies in the yz plane and the static magnetic field B_0 lies along the z axis. The sensitivity in the xz (top) and xy (bottom) planes are computed for a pulse length $\tau = 3.7 \mu\text{s}$ and an excitation current $I = 9 \text{ mA}$. Dark red color identifies the region of maximum sensitivity. (c) Rendered image of the single chip integrated detector combined with the 3D printed microfluidic structure. (d) Illustration of the approach used to place the sample onto the most sensitive area of the excitation/detection microcoil. The flow drives the sample in proximity of the integrated microcoil. The dashed arrow indicates the direction of insertion of the ovum, which is trapped by two 10- μm -high pillars. A 50- μm -high pillar is employed to block the sample in case of accidental overpressure. A 10- μm -thick crosslinked IP-S photoresist layer defines the spacing between the sample and the chip surface (see inset).

<https://doi.org/10.1371/journal.pone.0192780.g001>

report NMR measurements on sub-nL samples (and sub-nL portion of larger samples) having significantly different size, geometry, and nature. We report experiments on liquids, where we show spectral resolutions down to 0.007 ppm full width at half maximum (i.e., about 2 Hz at 300 MHz) in liquid samples of 100 pL. Additionally, we report experiments on two radically

different biological entities, i.e. tardigrade *Richtersius coronifer* ova and nematode *Caenorhabditis elegans* worms. Despite the tiny size (about 100 pL) and the broad intrinsic linewidth (about 30 Hz at 300 MHz) of these samples, the achieved sensitivity (2.5×10^{13} spins/Hz^{1/2}) is sufficient to detect highly concentrated endogenous compounds. The microfluidic channels are connected to a robust fluidic interface that tolerates the application of flows as strong as 7 μ L/s and guarantees an efficient sealing for at least two weeks. The results of the reported experiments indicate that the approach proposed here allows for the non-invasive and efficient handling and trapping of living entities for NMR investigations at the sub-nL volume scale, in conditions of high sensitivity and sample limited spectral resolution.

Materials and methods

Sensitive volume of the microcoil and related microfabrication constraints

The single chip NMR detector, described in details in Ref. [32], has an integrated excitation/detection multilayer microcoil having a diameter of 150 μ m (Fig 1A). The planar geometry and the localized sensitive volume (the distance between the microcoil and the co-integrated RF preamplifier is less than 300 μ m) allow for an eased approach and assembly with microfluidic structures. The integrated planar microcoil has a high sensitivity in a distorted spherical volume having a diameter of about 100 μ m. In the following, we quantify the dependence of the local sensitivity as a function of the distance from the microcoil center. As a convention throughout the article, the planar microcoil lies in the yz plane, with the static magnetic field B_0 along the z axis. Using the principle of reciprocity,[64] the signal contribution per elementary volume $dS(\mathbf{r})$ (i.e., the local sensitivity) is proportional to $B_{uxy}(\mathbf{r})\sin(\vartheta)$, where $B_{uxy}(\mathbf{r}) = ((B_{ux}(\mathbf{r}))^2 + B_{uy}(\mathbf{r})^2)^{1/2}$ is the field component in the xy plane (i.e., perpendicular to B_0) produced by the microcoil carrying a unitary current, $\vartheta = -\gamma B_{lxy}(\mathbf{r})\tau$ is the flip angle, γ is the nuclear gyromagnetic ratio, τ is the pulse length, $B_{lxy}(\mathbf{r}) = (B_{uxy}(\mathbf{r})/2)I$ and I is the current carried by the coil during the excitation pulse.[1] The maps of sensitivity of the integrated microcoil shown in Fig 1B are obtained calculating $B_{uxy}(\mathbf{r})\sin(\vartheta)$, starting from the computation of $B_{uxy}(\mathbf{r})$ via a Biot-Savart code implemented in Matlab. As shown in Fig 1B, the sensitivity of the microcoil decreases rapidly with the distance along the x axis. As a result, the microfluidic system used to hold the sample on top of the microcoil should have a thin separation layer. Variation of the pulse length τ implies a modification of the local flip angle ϑ , thus a variation of the spatial distribution of the sensitivity. As described in details in the next subsections, we fabricated two microchannels having different height and width. The integrated NMR signal S is proportional to the integral of $dS(\mathbf{r})$ over the total volume of the microchannel. Fig 1B shows the sensitivity maps for $\tau = 3.7$ μ s, which is the pulse length that maximizes the integrated signal for both microchannels, assuming a separation d between the microcoil and the sample of 10 μ m (the position of the samples with respect to the active volume of the detector microcoil are shown in section A in S1 File). Upon normalization of the integrated NMR signal to $S_{(d=0\ \mu\text{m})} = 1$ for the case of the channel volume placed in direct contact with the chip, the signals $S_{(d=10\ \mu\text{m})}$ and $S_{(d=20\ \mu\text{m})}$, evaluated in conditions of optimal τ , are respectively of about 0.7 and 0.5 for both microchannels. Hence, a separation layer of 10 μ m (Fig 1C and 1D) is an acceptable compromise between the loss of sensitivity and the robustness of the structure.

Design of the microchannels

The microfluidic device has to gently handle the biological entity, and to accurately position and hold the sample in proximity of the detector microcoil, while providing a stable sealing over the duration of the experiments. In order to allow an easy alignment of the sample with the microcoil, the structural material of the microchannel has to be optically transparent. The geometry of the

channel is designed to minimize static magnetic field inhomogeneities (i.e., smooth long channels with walls parallel to the static magnetic field direction) and maximize the detected NMR signal amplitude by holding the samples into the most sensitive volume of the detector.

We realized two microfluidic devices, designed for driving and trapping two different types of microscopic living entities in the most sensitive volume of the NMR integrated detector. These two designs are conceived, respectively, for NMR investigations of single tardigrade *Richtersius coronifer* (*Rc*) ova and of single *C. elegans* worms.

The design of the microchannel for the trapping of a single *Rc* ovum consists of a cylindrical inlet and outlet with a diameter of 240 μm . The microchannel gradually narrows to form a rectangular cross section with a height of 130 μm and a width of 140 μm at its centre. In correspondence to this constriction, two 10- μm -high pillars are designed to trap the ovum over the duration of the experiment (see Fig 2A). A second, larger and taller pillar is implemented to prevent the loss of the ovum in case of accidental overpressure during the loading of the sample.




Fig 2. Design and fabrication results of the *Rc* ovum and *C. elegans* worm-dedicated microchannels. (a) Schematic of the microchannel used to trap a single *Rc* ovum. A close up view of the trapping pillars is shown. Two 10- μm -high pillars trap the ovum, while a second pillar is implemented to prevent the loss of the ovum in case of accidental overpressure. (b) Top view optical microscope picture of the central part of the microchannel. (c) SEM picture of a microchannel section appropriately 3D printed to better visualize the trapping system. (d) Schematic of the microchannel used to trap a single *C. elegans* worm. Close up views of the trapping pillars and the central part of the microfluidic structure are shown in the insets. (e) Top view optical microscope picture of the pillars used to trap the head or tail of the *C. elegans*. (f) Top view optical microscope picture of the three microfluidic channels. The central microchannel traps the worm while the two lateral ones provide food to the trapped *C. elegans*. (g) SEM picture of a 3D printed channel section showing the trapping pillars. (h) SEM picture of a 3D printed channel section showing the central trapping microchannel and two lateral feeding microchannels.

<https://doi.org/10.1371/journal.pone.0192780.g002>

These dimensions of the channel allow to achieve a trapping efficiency of 100%. The investigated population of *Rc* ova is characterized by a width of the diameter distribution of about 10 μm , which is smaller than their capability to deform and much larger than the tolerances in the microchannel fabrication. The trapping of the ovum in the location of interest does not prohibit the flow in the channel, which is essential if we want to apply reverse flow to recover the sample or direct flow to refresh or change the medium. Fig 2B shows an optical microscope picture of the top view of the central part of the microchannel. To better visualize the cross section of the microchannel and the system of pillars used to trap the ovum, a section of the microfluidic structure is 3D printed and inspected by scanning electron microscopy (SEM) (Fig 2C).

The microfluidic chip designed for *C. elegans* is configured with one inlet and one outlet, featuring a main entrance that splits into three channels that re-join at the outlet (Fig 2D). The central microchannel is designed to match tightly the size of an adult worm, being 60 μm wide, 50 μm high and 1.3 mm long. 50- μm -high pillars at the end of this channel are placed in a V-shaped arrangement at a distance of 12 μm , and used to block either the tail or the head of the *C. elegans* worm. The other two 55- μm -wide lateral channels are designed to deliver nutrients to the living trapped worm, being in fluidic connection with the two extremities of the central channel. To prevent the undesired trapping of other worms in the lateral feeding channels, one pillar is placed at their entrance (Fig 2D). Top view optical microscope pictures of the trapping pillars and the central part of the microfluidic structure are shown respectively in Fig 2E and 2F. Sections of the microfluidic device are 3D printed and inspected by SEM to better visualize the V-shaped arrangement of the pillars (Fig 2G) and the part of the microfluidic structure formed of a central trapping microchannel and two lateral feeding microchannels (Fig 2H). The size variation among the investigated population of *C. elegans* is larger than their capabilities to deform and much larger than the geometrical tolerances in the fabrication process. Therefore, an appropriate selection under the microscope of a worm having a size that fits into the dedicated microchannel is necessary. This selection allows for a trapping efficiency of 100%.

In both designs the gradual narrowing of the channels facilitates the hydrodynamic trapping of the samples. The size and shape of the inlet and outlet parts match the capillaries used for the fluidic connection with the external pumps, limiting the presence of dead volumes (see details in section B in S1 File,). Moreover, the thickness of the separation layer gradually changes over the microchannel length, with a thickness of only 10 μm in proximity of the most sensitive region of the detector (i.e., above the excitation/detection microcoil). This strategy simultaneously implements the requirements of proximity for the maximization of the NMR signal amplitude and robustness of the structure. The implementation of all these essential 3D features (i.e. gradual narrowing of the channel in both directions of the cross-section, inlet and outlet shape and dimension, change of the thickness over the microchannel length) would be complicated with conventional microfabrication techniques.

Fabrication of the microchannels

The microchannels are fabricated via a two-photon polymerization technique using a high resolution 3D printer (Photonic Professional GT, Nanoscribe GmbH, Germany). This additive manufacturing technique provides a resolution better than 1 μm^3 in a more versatile manner compared to traditional microfabrication methods. The structures are patterned into IP-S photoresist, a negative tone cross-linking polymer proprietary to Nanoscribe GmbH, which is optically transparent and gives an extraordinary geometrical freedom in the design of the microchannels. An indium tin oxide (ITO) coated glass slide serves as a substrate. The refractive index difference between the two materials guarantees a facilitated detection of the interface. For the release of the microchannels, a 500-nm-thick dextran layer (Sigma-Aldrich, 31390-

25G) is spin coated onto the substrate. This sacrificial layer is subsequently dissolved in water, enabling the release of the structures at the end of the fabrication process. A 25x objective (LCI Plan-Neofluar 25/0.8 Imm Korr DIC M27, Zeiss, Germany) is used in direct immersion mode into the photoresist. The refractive indices of the objective and the photoresist are matched to enable high spatial resolution. The voxel diameter is determined by a preceding test exposure, and the results are fed into the modelling software which imports Solidworks STL files to generate machine specific data. Using the 25x objective and 50 m/s writing speed, the voxel size is about $(0.3 \times 0.3 \times 2) \mu\text{m}^3$ (with a 63x objective the voxel size is about $(0.15 \times 0.15 \times 0.45) \mu\text{m}^3$). A droplet of IP-S resist is placed onto the substrate, in which the objective is immersed. The objective is fixed in space, while the positioning of the substrate is given by the combination of galvanometric MEMS mirrors and a piezoelectric unit. In the xy plane, the galvanometric mirrors travel within a 200 μm radius at each fixed piezo position. A slicing distance of 1 μm is defined to split the structure into equal distance horizontal planes which are set by the piezoelectric stage. This parameter is set to ensure proper overlapping and adhesion in between the horizontal planes. To decrease writing time, a scaffolding technique is implemented within the bulk volumes. A triangular support structure is used, with a 20 μm spacing between planes and a scaffolding wall thickness of 3 μm . The writing time for a single microfluidic structure is approximately 5 hours. To provide control over the resist-air interface, and to maintain mechanical resistance over an extended period of development, an 18 μm thick outer shell is defined, which is patterned as a bulk area. The accessible area of the galvanometric mirror is limited by the beam deflection and it is confined in a 200 μm radius circle. The block is a volume which can be written at a single xy piezoelectric stage position, only by moving the galvanometric mirrors and the piezoelectric stage in the z direction. The block shape has hexagonal shape to optimize the volume accessible from the galvanometric mirror and to facilitate stitching by having large neighbouring block surfaces. The blocks are written in a consecutive manner: a piezoelectric stage xy position is chosen, which gets exposed by the combination of the galvanometric mirrors and the piezoelectric z stage. When the block writing is finished, the piezoelectric stage moves to the next, neighbouring position and restarts exposure as before. A block shear angle of 13° is used in the z direction to avoid a shadowing effect, which occurs at overlapping block edges due to increased exposure. A block overlap of 2.5 μm enhances the stitching by reinforcing the adhesion between blocks. The block size is chosen to be x : 220 μm , y : 190 μm , z : 250 μm . Applying the shear angle and overlap, the total block dimension becomes x : 259.5 μm , y : 259 μm , z : 250 μm .

The laser power is 65 mW at the point of entering the objective. For writing the shell of the structure, the laser power is reduced to 42%. For the internal scaffolding, the laser power is increased to 50%, which increases the robustness at the cost of the spatial resolution. Following exposure, the objective is removed from the photoresist, and the substrate is placed into PGMEA (Sigma-Aldrich, 484431) for development. The substrate is positioned in a way that the channels stand vertically. A continuous stirring in the beaker enhances the removal of developed photoresist residues from inside the channel. The development duration is typically 6 hours. When development is finished, the substrate is immersed into isopropanol in an identical configuration for one hour. A second rinse step, with ultra-high purity isopropanol (99.99%+) is used to further clean the inside of the channels, with a duration of 30 minutes. Finally, the substrate is left to naturally dry in vertical position.

Assembly of the 3D printed microchannels with the single chip integrated CMOS detector

In order to tightly hold the high resolution 3D printed microfluidic chips in contact with the single-chip integrated NMR detector and to connect its microchannels to external computer

controlled syringe pumps, we fabricated an interfacing microfluidic structure (see details in sections B and C in [S1 File](#)). The positioning of the 3D printed microfluidic chip over the integrated microcoil is performed under a standard optical microscope (MZ8, Leica, Germany).

Single chip integrated detector and NMR experiments

The single chip integrated NMR detector, realized with a 130 nm CMOS technology from STMicroelectronics, consists of a RF power amplifier, a RF low-noise preamplifier, a frequency mixer, an audio-frequency (AF) amplifier, transmit/receive switches, and an excitation/detection microcoil with an external diameter of 150 μm . A detailed description of the single chip integrated NMR detector is reported in Ref. [32]. All NMR experiments performed in this work consist of a RF pulsed excitation (about 300 MHz, pulse duration τ of a few μs) immediately followed by a detection time T_D (50 ms to 10 s). This simple excitation/detection sequence is repeated n times (1 to 10^6) with a repetition time T_R (50 ms to 10 s). A schematic of the NMR experimental set-up is shown in section D in [S1 File](#).

C. elegans preparation and microfluidic manipulation

The *C. elegans* wild type worms are cultured at 20°C on nematode growth media (NGM) 90 mm Petri dishes seeded with *Escherichia coli* strain OP50. Worms and HT115 *E. coli* bacteria are provided by the Caenorhabditis Genetics Center (University of Minnesota). HT115 *E. coli* bacteria are grown in Luria Broth (LB) with 100 $\mu\text{g}/\text{mL}$ ampicillin and 12.5 $\mu\text{g}/\text{mL}$ tetracycline overnight in a thermal shaker at 37°C. The following day, 50 μL of the confluent bacterial cultures are used to inoculate freshly prepared LB medium containing only ampicillin. The new cultures are grown until reaching an optical density between 0.6 and 0.8, measured at a wavelength of 600 nm. 90 μL are used for seeding the experimental plates.

The microfluidic chip and tubes are first filled with S medium (prepared following the protocol reported in Ref. [65]) using the fluidic setup described in section C in [S1 File](#). A worm is transferred, using a worm picker, from the agarose plate to an S medium reservoir, from which it is sucked up in a tube connected to the device. Afterwards, with a flow of 500 nL/s, the worm is injected in the microfluidic chip from the inlet. In order to insert the animal in the trapping channel, a flow of 1 $\mu\text{L}/\text{s}$ is applied. Once trapped, a gentle flow of 50 nL/s of *E. coli* bacteria is used to replace the S medium in the chip and therefore provide nutrients to the worm through the lateral channels. After observing the pharyngeal pumping in the worm, which confirms that it is eating properly, the flow is stopped and the tubes are clamped. The whole operation, and the following NMR experiments, are performed at 20°C.

Rc ova preparation and microfluidic manipulation

Rc ova were extracted from a moss sample (*Orthotricum cupulatum*) collected in Möckelmosen (Öland Island, Sweden) by washing the substrate, previously submerged in water for 30 min, on sieves under tap water and then individually picking up eggs with a glass pipette under a dissecting microscope. The ova were shipped within 24 hours in sealed Eppendorf tubes with water and subsequently stored at -20°C before use. The ova are first transferred from the Eppendorf into a Petri dish prepared with 3% H₂O-based agarose gel.

The microfluidic chip and tubes are first filled with the chosen medium (D₂O or H₂O according to the experiment). A single *Rc* ovum is transferred from the agarose plate to the medium reservoir using a manipulation pipette (Vitrolife, Sweden). The single ovum is then released into the tube connected to the microfluidic system directly from the pipette. Later, with a flow of 500 nL/s, the ovum is injected in the microfluidic chip through the inlet. In order to place the ovum on top of the two trapping pillars, a flow of 3 $\mu\text{L}/\text{s}$ is applied. Once

trapped, the flow is stopped and the tubes are clamped. The whole operation, and the following NMR experiments, are performed at 20 °C.

Results

Spectroscopy of liquid samples

In Fig 3, we show a collection of ^1H spectra obtained from liquid samples (water and lactic acid in water) at 7.05 T (300 MHz) contained in the *C. elegans*-dedicated microchannels. These measurements are performed to characterize the spectral resolution limits of the setup and to measure the spin sensitivity of the detector. All chemical shifts are expressed in ppm deviation from the resonance frequency of tetramethylsilane (TMS). Since this standard reference compound is not present in our samples, we assigned a chemical shift of 4.8 ppm to the peak of water (present in all the investigated samples). The excitation pulse length used in the reported experiments ($\tau = 3.5 \mu\text{s}$) corresponds to the experimental condition of maximum sensitivity, in good agreement with the value of $3.7 \mu\text{s}$ computed with simulations via sensitivity maps.

Fig 3A shows the ^1H NMR spectrum of pure water (Sigma-Aldrich, 270733) after averaging 1000 scans. The lorentzian fit of the data indicates a linewidth of 2 Hz FWHM. The baseline width, defined as the peak width at 0.55% height of the peak of water, is 24 Hz. These spectral resolutions are systematically achieved in six separated experiments employing six different microchannels with the same nominal design. Fig 3B shows the ^1H NMR spectrum of 1.3 M of lactic acid (Sigma-Aldrich, L1750) in pure water after averaging over 10800 scans. The FWHM in this spectrum is also of 2 Hz. As shown in the inset, the achieved spectral resolution is sufficient to show the 7 Hz J-splitting within each of the two chemically shifted signals of the lactic acid. The two peaks at about 1.3 ppm arise from the ^1H nuclei in the CH_3 group, J-split by the ^1H nucleus of the nearby CH group. The four peaks at about 4.08 ppm arise from the ^1H nucleus of the CH group, and show the 1:3:3:1 J-split due to the ^1H spins in the CH_3 group. [66] Better spectral resolutions (down to 0.6 Hz) have been reported in literature,[9,27,36,42] but with probes having a worse spin sensitivity. In Fig E in S1 File, we report the spectrum of pure water obtained in the Rc ova-dedicated microchannel. Contrarily to the case of *C. elegans*-dedicated microchannels, in this microchannel we inserted pillars in close proximity to




Fig 3. ^1H NMR measurements on pure water and lactic acid in water performed at 7.05 T. The spectra are the real parts of the Fast Fourier Transform (FFT) of the time-domain NMR signals. Notations: V is the active volume, Avg number of averaged measurements, T_R is the repetition time, τ is the pulse length, T_m is the matching filter decay time constant. (a) ^1H spectrum of water: $V \cong 100 \text{ pL}$, $\text{Avg} = 1000$, $T_R = 4 \text{ s}$, $\tau = 3.5 \mu\text{s}$, $T_m = \infty$ (b) ^1H spectrum of 1.3 M lactic acid in H_2O : $V \cong 100 \text{ pL}$, $\text{Avg} = 10800$, $T_R = 2 \text{ s}$, $\tau = 3.5 \mu\text{s}$, $T_m = 500 \text{ ms}$.

<https://doi.org/10.1371/journal.pone.0192780.g003>

the sensed volume, which are necessary to trap the *Rc* ova. The mismatches in susceptibility are larger than in the microchannel designed for the trapping of the *C. elegans*, resulting in an experimentally measured spectral resolution of 10 Hz (i.e., 5 times worse with respect to the other design). The approach proposed in this work (i.e., high spatial resolution 3D printed microchannels combined with single-chip CMOS integrated detectors) allows for sub-nL NMR spectroscopy with spectral resolutions consistently limited by the specific design of the microfluidic structure, with an experimentally demonstrated spectral resolution of 0.007 ppm for one of the two implemented designs. We used the NMR spectra obtained with the water sample also to calibrate the sensitivity of the detector in combination with the microfluidic channel. The time-domain sensitivity N_{\min} can be defined as $(N_S V_{\text{SN}_{\text{SD}}})/S_0$ where n_{SD} represents the noise spectral density, S_0 is the signal amplitude at $t = 0$, N_S is the spin density and V_S is the sample volume. Following this definition, a value of 2.5×10^{13} spins/Hz^{1/2} is experimentally obtained for both microchannels. This value of sensitivity differs by a factor of 1.7 with respect to the experimental value found when the samples are placed in direct contact with the microcoil.[32,40] Overall, this is in good agreement with what is computed via the sensitivity maps, where we estimate that a 10 μm separation layer reduces the sensitivity of the experiment by a factor of 1.4. The measured time-domain spin sensitivity corresponds to a limit of detection (LOD) at 300 MHz of approximately 600 pmol s^{1/2} of 1H nuclei in pure water ($T_1 \cong T_2 \cong 3$ s) with an effective transversal relaxation time $T_2^* \cong 0.15$ s (i.e., a 2 Hz FWHM line) and repetition time $TR = 4$ s.

Spectroscopy of single intact biological samples

Fig 4 shows ¹H NMR spectra obtained at 7.05 T (300 MHz) from intact biological samples. Fig 4A and 4B show NMR spectra obtained from a single *Rc* ovum. Fig 4C and 4D show the NMR spectrum of a subsection of a single *C. elegans* worm. *Rc* ova have typical volumes of 500 pL, whereas adult *C. elegans* worms have typical volumes of 5 nL. The most sensitive region of the NMR integrated detector corresponds to a deformed ellipsoid of about 300 pL. In the experiments, the sensed portion of the microorganisms is given by the intersection of the sensitive region with the volume of the sample. Defining the active volume as the fraction of the sample that contributes to the 70% of the total signal and considering the geometries in play and their position with respect to the microcoil, we estimate active volumes of about 250 pL in the case of the *Rc* ovum and 100 pL in the case of the *C. elegans* worm. The chemical shifts scale is obtained assigning a chemical shift of 4.8 ppm to the peak of water contained in the samples under investigation.

Fig 4A shows the NMR spectrum obtained from a single *Rc* ovum in H₂O after averaging over 75 hours. With a FWHM linewidth of about 0.1 ppm (35 Hz) and a baseline width of about 3 ppm (1 kHz), the water peak can overlap significantly with eventual small nearby signals. To suppress the water peak which originates from the carrying medium, we measured the *Rc* ova also in D₂O. The same *Rc* ovum is used to perform the measurements in H₂O and D₂O shown in Fig 4A and 4B. In D₂O at least one additional peak is more clearly visible. Thanks to the fluidic interface (see section C in S1 File), the liquid is easily exchanged in between the two experiments without repeating the trapping procedure. Due to the relatively low SNR (caused by the very small number of spins contained in the sample) and the relatively poor spectral resolution, a detailed proton peak assignment is currently impossible. Nevertheless, some information can be obtained from previous studies on *Xenopus laevis*[52,67] and hen eggs[68]. The spectra shown in Fig 4A and 4B indicate that we can associate the nature of the dominant peaks in the *Rc* ovum (i.e. the peaks at 0.9, 1.3, 2.1, 2.8 ppm) to the yolk lipid content of the sample.




Fig 4. ^1H NMR measurements of a single *Rc* ovum and a *C. elegans* subsection at 7.05 T. See definition of notations in Fig 3. The dashed circles indicate the 150 μm outer diameter of the integrated microcoil. (a) ^1H spectrum of a single *Rc* tardigrade ovum in H_2O : $V \cong 150$ pL, $\text{Avg} = 710400$, $T_R = 200$ ms, $\tau = 3.5$ μs , $T_m = 30$ ms. (b) ^1H spectrum of a single *Rc* tardigrade ovum in D_2O : $V \cong 150$ pL, $\text{Avg} = 28416$, $T_R = 200$ ms, $\tau = 3.5$ μs , $T_m = 30$ ms. (c) ^1H spectrum of a single *C. elegans* subsection: $V \cong 100$ pL, $\text{Avg} = 568320$, $T_R = 200$ ms, $\tau = 3.5$ μs , $T_m = 30$ ms. (d) ^1H spectrum of a single *C. elegans* subsection: $V \cong 100$ pL, $\text{Avg} = 113664$, $T_R = 200$ ms, $\tau = 3.5$ μs , $T_m = 30$ ms.

<https://doi.org/10.1371/journal.pone.0192780.g004>

Fig 4C shows the NMR spectrum obtained from a single *C. elegans* subsection, after averaging over 60 hours. After trapping the worm using S medium, a gentle flow of S medium and *E. coli* (1.5×10^9 cell/mL) is applied to provide food to the *C. elegans*. At this point the flow is stopped and the fluidic connection clamped for sealing. The *C. elegans* goes through a squeezing process when entering the microchannel, whose size tightly matches the worm. It follows that, in good approximation, only the worm is present in the active area of the detector. For this experiment, we decided to align the subsection of the worm containing ova with the microcoil. As shown in Fig 4C (inset), most of the sensitive region is indeed occupied by about eight eggs contained in *C. elegans* abdomen. The resulting spectrum has features which are similar to those measured in *Rc* ova. The peaks assignment is challenging due to the reduced size of the samples combined with the relatively poor spectral resolution. The peaks observed at about 0.9, 1.3, 2.1, 2.8 ppm may be mainly associated to yolk lipids. Some qualitative information can be obtained from previous ^1H HRMAS NMR studies on a collection of *C. elegans*[45] and on a single worm[47]. A prominent signal arises at a chemical shift of about 3.9 ppm. As shown in the Fig G in S1 File, the same resonance appears in the spectrum of *E. coli* in S medium, which is used as a base to feed the worm during the experiment. Therefore, we suspect that such resonance may result from the ingestion of nutrients operated by the microorganism.

In both *Rc* and *C. elegans* spectra, the linewidths are broader than the ones measured with liquid samples (as shown in Fig 3 for pure water and lactic acid). Previous studies suggest that the spectral resolution may be limited by microscopic constituents in the samples, which introduce susceptibility mismatches which are difficult to compensate by field shimming. As shown above, the linewidths observed in this study are of 35 Hz FWHM for the *Rc* ovum and 30 Hz FWHM for the *C. elegans*. For these biological entities, measured with a repetition time $T_R = 200$ ms and having linewidths of 30 Hz FWHM (i.e., $T_2^* \cong 0.01$ s), the LOD is about 600 pmol $s^{1/2}$ of ^1H nuclei, which corresponds to 6 pmol of ^1H nuclei in an averaging time of 3 hours (see further experimental details in section F in S1 File).

Thanks to the freedom in the design of the microchannels, the stable trapping, the efficient sealing (tested up to two weeks) and the robustness of the fluidic interface, the duration of our experiments depends only on the viability of the biological sample under investigation. The successful implementation of long measurements allows to improve the signal-to-noise ratio and to perform continuous observations on the same single entity on a time scale that can be comparable to the biological development of the sample.

Conclusions and outlook

NMR experiments on sub-nL volumes are hindered by the limited sensitivity of the detector and the difficulties in positioning and holding such small samples in proximity of the excitation/detection microcoils. In this work, we show the first examples of NMR experiments on liquids and biological entities immersed in liquids having active volumes down to 100 pL. These measurements are enabled by the successful design and fabrication of high spatial resolution 3D printed microfluidic structures, specifically conceived to guide and confine sub-nL samples in the sub-nL most sensitive volume of a single-chip integrated NMR detector. The obtained results may indicate a promising route for NMR studies at the single unit level of important biological entities having sub-nL volumes, which include, e.g., the eggs of several mammals.[55]

The microfluidic structures are fabricated using a two-photon polymerization 3D printing technique having a resolution better than better than $1 \mu\text{m}^3$. Compared to conventional clean-room microfabrication techniques, high resolution 3D printing techniques are currently not appropriate for an efficient large scale production, but represent a very promising solution for the rapid prototyping of simple and complex microfluidic systems[69–71] for the handling, feeding and trapping of biological samples having different geometry, size, and fluidic behaviour. More specifically, in this work we show that high resolution 3D printing is well suited for the microfabrication of microfluidic structures dedicated to NMR spectroscopy studies of sub-nL samples. Thanks to the fabrication of a robust separation layer of only 10 μm between the sample and the excitation/detection microcoil, the sample can be trapped in a liquid environment for several days very close to the most sensitive volume of the detector. The obtained effective spin sensitivity of $2.5 \cdot 10^{13}$ spin/Hz $^{1/2}$ is sufficient for the detection of highly concentrated endogenous compounds in the sub-nL volumes of the investigated biological entities (i.e., single *Rc* ova and portion of *C. elegans* worms) in experimental times of a few hours. In the measurements on the intact biological samples we achieved spectral resolutions of approximately 0.1 ppm, due to mismatches in the susceptibility caused by the sample itself, which we have not managed to shim. In the microchannels designed to minimize magnetic field non-uniformities we obtained spectral resolutions down to 0.007 ppm (i.e., about 2 Hz at 7 T) in 100 pL liquid samples.

The combination of CMOS integrated NMR detectors with high spatial resolution 3D printed microfluidic structures is compatible with the implementation of arrays of

miniaturized probes, which would enable simultaneous studies on a large number of single biological entities in the same magnet. This could provide an enabling technology for systematic investigation of the heterogeneity among individuals or among different subsections of a single organism (such as a *C. elegans* worm), as well as of their response to different environmental conditions and drugs.

Supporting information

S1 File. Fig A. Sensitivity map for the NMR microcoil computed via a Biot-Savart based code implemented in Matlab. For the NMR microcoil, the local sensitivity is defined as $B_{uxy}(\mathbf{r})\sin(\gamma B_{lxy}(\mathbf{r})\tau)$. The coil is in the yz plane and the static magnetic field B_0 is along the z axis. Sensitivity distribution in the xz (a and c) and xy (b and d) planes for NMR coil with excitation parameters $\tau = 3.7 \mu\text{s}$ and $I = 9 \text{ mA}$. The dashed lines indicate the position of the *Rc* ovum (a and b) and the *C. elegans* worm (c and d). **Fig B. Illustration of the fluidic interface.** (a) The microfluidic channel is attached to a PMMA rod and connected to PMMA capillaries to create a fluidic assembly. (b) The fluidic assembly is mounted in the holder. (c) The Micro-to-macro interface is completed by connecting tubes to the PMMA capillaries and performing the casting of epoxy resin to create the sealing and give robustness to the fluidic system. **Fig C. Schematic of the fluidic set-up used for the loading and trapping of biological samples.** (1) Microfluidic chip and its fluidic interface. (2) Inverted microscope (Axio Observer, Zeiss, Germany), High-Power LED Illumination system (precisExcite, VisiTron, Germany) for bright-field imaging. (3) Liquid waste. (4) High resolution digital camera (ORCA-ER C4742-80, Hamamatsu, Japan). (5) Syringe pumps (Nemesys, Cetoni GmbH, Germany). (6) Desktop computer. **Fig D. Schematic of the electronics setup used for the NMR measurements.** (1) Integrated excitation/detection coil interfaced with the 3D printed microchannel through a micro-to-macro fluidic interface. (2) Single chip NMR detector (see details in Ref. [72]). (3) Printed circuit board (PCB). (4) Superconductive magnet (Bruker, 7 T). (5) RF source (MG3633A, Anritsu; Japan). (6) AF amplifier (SRS560, Stanford Research Systems, USA) (7) Multifunctional board (PCIe-6259, National Instruments, USA) for the generation of Tx/Rx switching pulse and signal acquisition. **Fig E. ^1H spectrum of H_2O in the *Rc* ovum-dedicated microchannels, performed at 7.05 T.** The spectra are the real parts of the Fast Fourier Transform (FFT) of the time-domain NMR signals. Notations: V is the active volume, Avg number of averaged measurements, T_R is the repetition time, τ is the pulse length, T_m is the matching filter decay time constant. $V \cong 250 \text{ pL}$, $Avg = 1800$, $T_R = 2 \text{ s}$, $\tau = 3.5 \mu\text{s}$, $T_m = \infty$. **Fig F. NMR measurements of a single *Rc* tardigrade ovum in H_2O at different repetition times.** NMR measurements performed at 7 T. See definition of notation in Fig E. (a): $V \cong 210 \text{ pL}$; $Avg = 21600$; $T_R = 2 \text{ s}$, $\tau = 3.5 \mu\text{s}$, $T_m = 30 \text{ ms}$. (b): $V \cong 210 \text{ pL}$; $Avg = 216000$; $T_R = 200 \text{ ms}$, $\tau = 3.5 \mu\text{s}$, $T_m = 30 \text{ ms}$. (c): $V \cong 210 \text{ pL}$; $Avg = 864000$; $T_R = 50 \text{ ms}$, $\tau = 3.5 \mu\text{s}$, $T_m = 30 \text{ ms}$. **Fig G. ^1H spectrum of *E.coli* in S medium.** NMR measurement performed at 7 T, using the 3D oriented microchannel designed for the trapping of a single *C. elegans*. See definition of notation in Fig E. $V \cong 100 \text{ pL}$; $Avg = 19800$; $T_R = 2 \text{ s}$, $\tau = 3.5 \mu\text{s}$, $T_m = 60 \text{ ms}$. (DOCX)

Acknowledgments

We thank Vincenzo Sorrentino for kindly providing the *C. elegans* worm plates.

Author Contributions

Conceptualization: E. Montinaro, G. Boero.

Data curation: E. Montinaro, M. Grisi, G. Boero.

Formal analysis: E. Montinaro, G. Boero.

Funding acquisition: J. Brugger.

Investigation: E. Montinaro, M. Grisi, M. C. Letizia, L. Pethö, J. Michler, G. Boero.

Methodology: E. Montinaro, M. Grisi, M. C. Letizia, L. Pethö, M. A. M. Gijs, R. Guidetti, J. Michler, G. Boero.

Project administration: E. Montinaro.

Resources: J. Brugger, G. Boero.

Software: G. Boero.

Supervision: J. Brugger, G. Boero.

Validation: E. Montinaro, J. Brugger, G. Boero.

Visualization: E. Montinaro, G. Boero.

Writing – original draft: E. Montinaro, G. Boero.

Writing – review & editing: E. Montinaro, G. Boero.

References

1. Abragam A (1961) *The Principles of Nuclear Magnetic Resonance*. Oxford: Clarendon Press.
2. Callaghan PT (1991) *Principles of nuclear magnetic resonance microscopy*: Clarendon Press Oxford.
3. Ernst RR, Bodenhausen G, Wokaun A (1991) *Principles of nuclear magnetic resonance in one and two dimensions*.
4. Pykett IL (1982) NMR Imaging in Medicine. *Scientific American* 246: 78–88. PMID: [7079720](#)
5. McRobbie DW, Moore EA, Graves MJ, Prince MR (2017) *MRI from Picture to Proton*: Cambridge university press.
6. Tkac I, Starcuk Z, Choi IY, Gruetter R (1999) In vivo H-1 NMR spectroscopy of rat brain at 1 ms echo time. *Magnetic Resonance in Medicine* 41: 649–656. PMID: [10332839](#)
7. Tkac I, Andersen P, Adriany G, Merkle H, Ugurbil K, Gruetter R (2001) In vivo H-1 NMR spectroscopy of the human brain at 7 T. *Magnetic Resonance in Medicine* 46: 451–456. PMID: [11550235](#)
8. Pfeuffer J, Tkac I, Provencher SW, Gruetter R (1999) Toward an in vivo neurochemical profile: Quantification of 18 metabolites in short-echo-time H-1 NMR spectra of the rat brain. *Journal of Magnetic Resonance* 141: 104–120. <https://doi.org/10.1006/jmre.1999.1895> PMID: [10527748](#)
9. Olson DL, Peck TL, Webb AG, Magin RL, Sweedler JV (1995) High-Resolution Microcoil H-1-Nmr for Mass-Limited, Nanoliter-Volume Samples. *Science* 270: 1967–1970.
10. Peck TL, Magin RL, Lauterbur PC (1995) Design and Analysis of Microcoils for Nmr Microscopy. *Journal of Magnetic Resonance Series B* 108: 114–124. PMID: [7648010](#)
11. Webb AG (1997) Radiofrequency microcoils in magnetic resonance. *Progress in Nuclear Magnetic Resonance Spectroscopy* 31: 1–42.
12. Lacey ME, Subramanian R, Olson DL, Webb AG, Sweedler JV (1999) High-resolution NMR spectroscopy of sample volumes from 1 nL to 10 μ L. *Chemical Reviews* 99: 3133–+. PMID: [11749512](#)
13. Artzi Y, Twig Y, Blank A (2015) Induction-detection electron spin resonance with spin sensitivity of a few tens of spins. *Applied Physics Letters* 106: 084104.
14. Twig Y, Dikarov E, Hutchison WD, Blank A (2011) Note: High sensitivity pulsed electron spin resonance spectroscopy with induction detection. *Review of Scientific Instruments* 82: 076105. <https://doi.org/10.1063/1.3611003> PMID: [21806239](#)
15. Minard KR, Wind RA (2002) Picoliter H-1 NMR spectroscopy. *Journal of Magnetic Resonance* 154: 336–343. <https://doi.org/10.1006/jmre.2001.2494> PMID: [11846593](#)
16. Massin C, Vincent F, Homsy A, Ehrmann K, Boero G, Besse PA, et al. (2003) Planar microcoil-based microfluidic NMR probes. *Journal of Magnetic Resonance* 164: 242–255. PMID: [14511593](#)

17. Yamauchi K, Janssen JWG, Kentgens APM (2004) Implementing solenoid microcoils for wide-line solid-state NMR. *Journal of Magnetic Resonance* 167: 87–96. <https://doi.org/10.1016/j.jmr.2003.12.003> PMID: 14987602
18. Brecker L, Weber H, Griengl H, Ribbons DW (1999) In situ proton-NMR analyses of *Escherichia coli* HB101 fermentations in $^1\text{H}_2\text{O}$ and in D_2O . *Microbiology* 145: 3389–3397. <https://doi.org/10.1099/00221287-145-12-3389> PMID: 10627037
19. Maguire Y, Chuang IL, Zhang SG, Gershensfeld N (2007) Ultra-small-sample molecular structure detection using microslot waveguide nuclear spin resonance. *Proceedings of the National Academy of Sciences of the United States of America* 104: 9198–9203. <https://doi.org/10.1073/pnas.0703001104> PMID: 17517654
20. van Bentum PJM, Janssen JWG, Kentgens APM, Bart J, Gardeniers JGE (2007) Stripline probes for nuclear magnetic resonance. *Journal of Magnetic Resonance* 189: 104–113. <https://doi.org/10.1016/j.jmr.2007.08.019> PMID: 17890116
21. Ehrmann K, Saillen N, Vincent F, Stettler M, Jordan M, Wurm FM, et al. (2007) Microfabricated solenoids and Helmholtz coils for NMR spectroscopy of mammalian cells. *Lab on a Chip* 7: 373–380. <https://doi.org/10.1039/b614044k> PMID: 17330169
22. Krojanski HG, Lambert J, Gerikalan Y, Suter D, Hergenroder R (2008) Microslot NMR Probe for Metabolomics Studies. *Analytical Chemistry* 80: 8668–8672. <https://doi.org/10.1021/ac801636a> PMID: 18939856
23. Weiger M, Schmidig D, Denoth S, Massin C, Vincent F, Schenkel M, et al. (2008) NMR microscopy with isotropic resolution of $3.0\ \mu\text{m}$ using dedicated hardware and optimized methods. *Concepts in Magnetic Resonance Part B: Magnetic Resonance Engineering* 33: 84–93.
24. Anders J, Chiaramonte G, SanGiorgio P, Boero G (2009) A single-chip array of NMR receivers. *Journal of Magnetic Resonance* 201: 239–249. <https://doi.org/10.1016/j.jmr.2009.09.019> PMID: 19836280
25. Leidich S, Braun M, Gessner T, Riemer T (2009) Silicon cylinder spiral coil for nuclear magnetic resonance spectroscopy of nanoliter samples. *Concepts in Magnetic Resonance Engineering* 35: 11–22.
26. Lam MH, Homenuke MA, Michal CA, Hansen CL (2009) Sub-nanoliter nuclear magnetic resonance coils fabricated with multilayer soft lithography. *Journal of Micromechanics and Microengineering* 19: 095001.
27. Bart J, Janssen JWG, van Bentum PJM, Kentgens APM, Gardeniers JGE (2009) Optimization of stripline-based microfluidic chips for high-resolution NMR. *Journal of Magnetic Resonance* 201: 175–185. <https://doi.org/10.1016/j.jmr.2009.09.007> PMID: 19786359
28. Kratt K, Badilita V, Burger T, Korvink JG, Wallrabe U (2010) A fully MEMS-compatible process for 3D high aspect ratio micro coils obtained with an automatic wire bonder. *Journal of Micromechanics and Microengineering* 20.
29. Anders J, SanGiorgio P, Boero G (2011) A fully integrated IQ-receiver for NMR microscopy. *Journal of Magnetic Resonance* 209: 1–7. <https://doi.org/10.1016/j.jmr.2010.12.005> PMID: 21257327
30. Badilita V, Meier RC, Spengler N, Wallrabe U, Utz M, Korvink JG (2012) Microscale nuclear magnetic resonance: a tool for soft matter research. *Soft Matter* 8: 10583–10597.
31. Meier RC, Hofflin J, Badilita V, Wallrabe U, Korvink JG (2014) Microfluidic integration of wirebonded microcoils for on-chip applications in nuclear magnetic resonance. *Journal of Micromechanics and Microengineering* 24.
32. Grisi M, Gualco G, Boero G (2015) A broadband single-chip transceiver for multi-nuclear NMR probes. *Review of Scientific Instruments* 86.
33. Kalfe A, Telfah A, Lambert J, Hergenroder R (2015) Looking into Living Cell Systems: Planar Waveguide Microfluidic NMR Detector for in Vitro Metabolomics of Tumor Spheroids. *Analytical Chemistry* 87: 7402–7410. <https://doi.org/10.1021/acs.analchem.5b01603> PMID: 26121119
34. Finch G, Yilmaz A, Utz M (2016) An optimised detector for in-situ high-resolution NMR in microfluidic devices. *Journal of Magnetic Resonance* 262: 73–80. <https://doi.org/10.1016/j.jmr.2015.11.011> PMID: 26754062
35. Anders J, Handwerker J, Ortmanns M, Boero G (2016) A low-power high-sensitivity single-chip receiver for NMR microscopy. *Journal of Magnetic Resonance* 266: 41–50. <https://doi.org/10.1016/j.jmr.2016.03.004> PMID: 27011023
36. Spengler N, Höfflin J, Moazenzadeh A, Mager D, MacKinnon N, Badilita V, et al. (2016) Heteronuclear micro-Helmholtz coil facilitates μm -range spatial and sub-Hz spectral resolution NMR of nL-volume samples on customisable microfluidic chips. *PLoS one* 11: e0146384. <https://doi.org/10.1371/journal.pone.0146384> PMID: 26730968

37. Tijssen KC, Bart J, Tiggelaar RM, Janssen JH, Kentgens AP, van Bentum PJM (2016) Spatially resolved spectroscopy using tapered stripline NMR. *Journal of Magnetic Resonance* 263: 136–146. <https://doi.org/10.1016/j.jmr.2015.12.021> PMID: 26796112
38. Duong NT, Endo Y, Nemoto T, Kato H, Bouzier-Sore AK, Nishiyama Y, et al. (2016) Evaluation of a high-resolution micro-sized magic angle spinning (HR μ MAS) probe for NMR-based metabolomic studies of nanoliter samples. *Analytical Methods* 8: 6815–6820.
39. Kamberger R, Moazenzadeh A, Korvink JG, Gruschke OG (2016) Hollow microcoils made possible with external support structures manufactured with a two-solvent process. *Journal of Micromechanics and Microengineering* 26.
40. Grisi M, Vincent F, Volpe B, Guidetti R, Harris N, Beck A, et al. (2017) NMR spectroscopy of single sub-nL ova with inductive ultra-compact single-chip probes. *Scientific Reports* 7.
41. Bart J, Kolkman AJ, Oosthoek-de Vries AJ, Koch K, Nieuwland PJ, Janssen H, et al. (2009) A Microfluidic High-Resolution NMR Flow Probe. *Journal of the American Chemical Society* 131: 5014+. <https://doi.org/10.1021/ja900389x> PMID: 19320484
42. Rogers JA, Jackman RJ, Whitesides GM, Olson DL, Sweedler JV (1997) Using microcontact printing to fabricate microcoils on capillaries for high resolution proton nuclear magnetic resonance on nanoliter volumes. *Applied Physics Letters* 70: 2464–2466.
43. Blank A, Twig Y, Ishay Y (2017) Recent trends in high spin sensitivity magnetic resonance. *Journal of Magnetic Resonance* 280: 20–29. <https://doi.org/10.1016/j.jmr.2017.02.019> PMID: 28545918
44. Spencer RG (2010) Equivalence of the time-domain matched filter and the spectral-domain matched filter in one-dimensional NMR spectroscopy. *Concepts in Magnetic Resonance Part A* 36: 255–265.
45. Blaise BJ, Giacomotto J, Elena B, Dumas M-E, Toulhoat P, Ségalat L, et al. (2007) Metabotyping of *Caenorhabditis elegans* reveals latent phenotypes. *Proceedings of the National Academy of Sciences* 104: 19808–19812.
46. Blaise BJ, Giacomotto J, Triba MN, Toulhoat P, Piotto M, Emsley L, et al. (2009) Metabolic profiling strategy of *Caenorhabditis elegans* by whole-organism nuclear magnetic resonance. *Journal of proteome research* 8: 2542–2550. <https://doi.org/10.1021/pr900012d> PMID: 19267476
47. Wong A, Li X, Molin L, Solari F, Elena-Herrmann Bnd, Sakellariou D (2014) μ High resolution-magic-angle spinning NMR spectroscopy for metabolic phenotyping of *Caenorhabditis elegans*. *Analytical chemistry* 86: 6064–6070. <https://doi.org/10.1021/ac501208z> PMID: 24897622
48. Aguayo JB, Blackband SJ, Schoeniger J, Mattingly MA, Hintermann M (1986) Nuclear magnetic resonance imaging of a single cell. *Nature* 322: 190–191. <https://doi.org/10.1038/322190a0> PMID: 3724861
49. Schoeniger JS, Aiken N, Hsu E, Blackband SJ (1994) Relaxation-time and diffusion NMR microscopy of single neurons. *Journal of Magnetic Resonance Series B* 103: 261–273. PMID: 8019778
50. Grant SC, Aiken NR, Plant HD, Gibbs S, Mareci TH, Webb AG, et al. (2000) NMR spectroscopy of single neurons. *Magnetic Resonance in Medicine* 44: 19–22. PMID: 10893516
51. Grant SC, Buckley DL, Gibb S, Webb AG, Blackband SJ (2001) MR microscopy of multicomponent diffusion in single neurons. *Magnetic Resonance in Medicine* 46: 1107–1112. PMID: 11746576
52. Lee SC, Cho JH, Mietchen D, Kim YS, Hong KS, Lee C, et al. (2006) Subcellular in vivo ^1H MR spectroscopy of *Xenopus laevis* oocytes. *Biophysical Journal* 90: 1797–1803. <https://doi.org/10.1529/biophysj.105.073502> PMID: 16361348
53. Lee SC, Mietchen D, Cho JH, Kim YS, Kim C, Hong KS, et al. (2007) In vivo magnetic resonance microscopy of differentiation in *Xenopus laevis* embryos from the first cleavage onwards. *Differentiation* 75: 84–92. <https://doi.org/10.1111/j.1432-0436.2006.00114.x> PMID: 17244024
54. Lee CH, Flint JJ, Hansen B, Blackband SJ (2015) Investigation of the subcellular architecture of L7 neurons of *Aplysia californica* using magnetic resonance microscopy (MRM) at 7.8 microns. *Scientific Reports* 5: 11147. <https://doi.org/10.1038/srep11147> PMID: 26059695
55. Flindt R (2006) *Amazing numbers in biology*: Springer Science & Business Media.
56. Sorli B, Chateaux JF, Pitaval M, Chahboune H, Favre B, Briguet A, et al. (2004) Micro-spectrometer for NMR: analysis of small quantities in vitro. *Measurement Science and Technology* 15: 877–880.
57. Seeber DA, Cooper RL, Ciobanu L, Pennington CH (2001) Design and testing of high sensitivity micro-receiver coil apparatus for nuclear magnetic resonance and imaging. *Review of Scientific Instruments* 72: 2171–2179.
58. Massin C, Boero C, Vincent F, Abenheim J, Besse PA, Popovic RS (2002) High-Q factor RF planar microcoils for micro-scale NMR spectroscopy. *Sensors and Actuators a-Physical* 97–8: 280–288.
59. Demas V, Bernhardt A, Malba V, Adams KL, Evans L, Harvey C, et al. (2009) Electronic characterization of lithographically patterned microcoils for high sensitivity NMR detection. *Journal of Magnetic Resonance* 200: 56–63. <https://doi.org/10.1016/j.jmr.2009.06.003> PMID: 19581116

60. Malba V, Maxwell R, Evans LB, Bernhardt AE, Cosman M, Yan K (2003) Laser-lathe lithography—a novel method for manufacturing nuclear magnetic resonance microcoils. *Biomedical Microdevices* 5: 21–27.
61. Sillerud LO, McDowell AF, Adolphi NL, Serda RE, Adams DP, Vasile MJ, et al. (2006) H-1 NMR Detection of superparamagnetic nanoparticles at 1 T using a microcoil and novel tuning circuit. *Journal of Magnetic Resonance* 181: 181–190. <https://doi.org/10.1016/j.jmr.2006.04.005> PMID: 16698297
62. Matheoud AV, Gualco G, Jeong M, Zivkovic I, Brugger J, Rønnow HM, et al. (2017) Single-chip electron spin resonance detectors operating at 50GHz, 92GHz, and 146GHz. *Journal of Magnetic Resonance* 278: 113–121. <https://doi.org/10.1016/j.jmr.2017.03.013> PMID: 28388496
63. Gijs MA (2004) Magnetic bead handling on-chip: new opportunities for analytical applications. *Microfluidics and Nanofluidics* 1: 22–40.
64. Hoult DI, Richards R (1976) The signal-to-noise ratio of the nuclear magnetic resonance experiment. *Journal of Magnetic Resonance* (1969) 24: 71–85.
65. Stiernagle T (1999) Maintenance of *C. elegans*. *C elegans* 2: 51–67.
66. Lloyd SG, Zeng HD, Wang PP, Chatham JC (2004) Lactate isotopomer analysis by H-1 NMR spectroscopy: Consideration of long-range nuclear spin-spin interactions. *Magnetic Resonance in Medicine* 51: 1279–1282. <https://doi.org/10.1002/mrm.20075> PMID: 15170850
67. Sehy JV, Ackerman JJH, Neil JJ (2001) Water and lipid MRI of the *Xenopus* oocyte. *Magnetic Resonance in Medicine* 46: 900–906. PMID: 11675641
68. Jayasundar R, Ayyar S, Raghunathan P (1997) Proton resonance imaging and relaxation in raw and cooked hen eggs. *Magnetic Resonance Imaging* 15: 709–717. PMID: 9285811
69. Waheed S, Cabot JM, Macdonald NP, Lewis T, Guijt RM, Paull B, et al. (2016) 3D printed microfluidic devices: enablers and barriers. *Lab on a Chip* 16: 1993–2013. <https://doi.org/10.1039/c6lc00284f> PMID: 27146365
70. Au AK, Huynh W, Horowitz LF, Folch A (2016) 3D-printed microfluidics. *Angewandte Chemie International Edition* 55: 3862–3881. <https://doi.org/10.1002/anie.201504382> PMID: 26854878
71. Ho CMB, Ng SH, Li KHH, Yoon Y-J (2015) 3D printed microfluidics for biological applications. *Lab on a Chip* 15: 3627–3637. <https://doi.org/10.1039/c5lc00685f> PMID: 26237523
72. Grisi M, Gualco G, Boero G (2015) A broadband single-chip transceiver for multi-nuclear NMR probes. *Review of Scientific Instruments* 86: 044703. <https://doi.org/10.1063/1.4916206> PMID: 25933876